

No.	Sample Name	Time	Dil.Fac.	Amount	Comment	Analyst:	wbh
				CRVI-LOW			
				UV VIS 1		Criteria	
1,	Standard 1 - 0.1ppb,	02/24/06 09:56,	1.0,	0.105	Standard Stock		
2,	Standard 2 - 0.2 ppb,	02/24/06 10:04,	1.0,	0.204	LMR050904-1		
3,	Standard 3 - 2.0 ppb,	02/24/06 10:12,	1.0,	2.038	exp 25 Feb 06		
4,	Standard 4 - 10 ppb,	02/24/06 10:20,	1.0,	10.075			
5,	Standard 5 - 20 ppb,	02/24/06 10:29,	1.0,	20.158			
6,	Standard 6 - 50ppb,	02/24/06 10:37,	1.0,	49.920			
7,	IPC 20,	03/15/06 08:42,	1.0,	20.888	20.89 104%		
8,	IPC 10,	03/15/06 08:50,	1.0,	10.476	10.47 105%		
9,	LRB,	03/15/06 08:58,	1.0,	n.a.			
10,	LRB + Buffer,	03/15/06 09:07,	1.0,	0.003		MRL 50%-150%	
11,	MRL 0.1ppb,	03/15/06 09:15,	1.0,	0.110	0.110 110%	0.05 - 0.15ppb	
12,	LCS 2.0ppb,	03/15/06 09:23,	1.0,	2.127	2.13 107%	LCS 90%-110%	
13,	LCSD 2.0ppb,	03/15/06 09:31,	1.0,	2.084	2.08 104%	Range: 1.80 - 2.20	
14,	2603140584,	03/15/06 09:39,	1.0,	0.041		MS/MSD 90%-110%	
15,	2603140584MS,	03/15/06 09:47,	1.0,	2.098	2.06 - 103% recovery	True Value = 2.0	
16,	2603140584MSD,	03/15/06 09:55,	1.0,	2.086	2.05 - 102% recovery	IPC 95%-105%	
17,	IPC 20,	03/15/06 10:03,	1.0,	20.861	20.86 104%	20ppb - 19-21ppb	
18,	LRB,	03/15/06 10:11,	1.0,	n.a.		10ppb - 9.5-10.5 ppb	
19,	2603150120,	03/15/06 11:33,	1.0,	12.892	7199		
20,	LRB,	03/15/06 11:41,	1.0,	0.001			
21,	IPC 20,	03/15/06 11:49,	1.0,	20.929	20.92 105%		