

No.	Sample Name	Time	Dil.Fac.	Amount	Comment	Analyst:	Imr
				CRVI-LOW			
				UV VIS 1		Criteria	
1,	Standard 1 - 0.1 ppb,	03/22/06 15:39,	1.0,	0.116	Standard Stock		
2,	Standard 2 - 0.2ppb,	03/22/06 15:47,	1.0,	0.226	LMR050904-1		
3,	Standard 3 - 2.0 ppb,	03/22/06 15:56,	1.0,	2.000	exp 20 Apr 06		
4,	Standard 4 - 10 ppb,	03/22/06 16:04,	1.0,	10.007			
5,	Standard 5 - 20 ppb,	03/22/06 16:12,	1.0,	20.019			
6,	Standard 6 - 50ppb,	03/22/06 16:20,	1.0,	49.991			
7,	IPC 20,	03/24/06 08:00,	1.0,	20.061	20.1 100%		
8,	IPC 10,	03/24/06 08:09,	1.0,	10.063	10.1 101%		
9,	LRB,	03/24/06 08:17,	1.0,	n.a			
10,	LRB + Buffer,	03/24/06 08:25,	1.0,	0.011			
11,	MRL 0.1ppb,	03/24/06 08:33,	1.0,	0.108	0.108 108%		
12,	LCS 2.0ppb,	03/24/06 08:41,	1.0,	2.016	2.02 101%		
13,	LCSD 2.0ppb,	03/24/06 08:49,	1.0,	1.997	2.00 100%		
14,	2603240267,	03/24/06 10:04,	1.0,	0.002			
15,	2603240267_MS,	03/24/06 10:12,	1.0,	1.985	1.99 - 99.3% recovery		
16,	2603240267_MSD,	03/24/06 10:20,	1.0,	1.980	1.98 - 99.0% recovery		
17,	2603240268,	03/24/06 10:28,	1.0,	4.024			
18,	LRB,	03/24/06 10:36,	1.0,	n.a			
19,	IPC20,	03/24/06 10:44,	1.0,	20.010	20.0 100%		
20,	2603240114,	03/24/06 11:53,	1.0,	12.685		MRL 50%-150%	
21,	2603240115,	03/24/06 12:01,	1.0,	51.915	over range	0.05 - 0.15ppb	
22,	2603240117,	03/24/06 12:09,	1.0,	16.933		LCS 90%-110%	
23,	2603240136,	03/24/06 12:18,	1.0,	22.588		Range: 1.80 - 2.20	
24,	2603240115 (1:2),	03/24/06 12:26,	2.0,	51.835		MS/MSD 90%-110%	
25,	2603240124_MS,	03/24/06 12:34,	1.0,	24.685	2.10 - 105% recovery	True Value = 2.0	
26,	2603240125_MSD,	03/24/06 12:42,	1.0,	24.751	2.16 - 108% recovery	IPC 95%-105%	
27,	2603240122,	03/24/06 12:50,	1.0,	5.324		20ppb - 19-21ppb	
28,	2603240116,	03/24/06 12:58,	1.0,	9.167		10ppb - 9.5-10.5 ppb	
29,	2603240118,	03/24/06 13:06,	1.0,	n.a	green/filtr 0.2um		
30,	LRB,	03/24/06 13:14,	1.0,	n.a			
31,	IPC20,	03/24/06 13:22,	1.0,	19.602	19.6 98%		