



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3829  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)


## Laboratory Report

for

Tronox LLC - Henderson  
PO Box 55

Henderson , NV 89009

Attention: Susan Crowley  
Fax: (405) 302-4607



ADE Andy Eaton  
Project Manager



Report#: 194600  
Project: CLO4  
PO#: Susan Crowle

Laboratory certifies that the test results meet all **NELAC** requirements unless noted in the Comments section or the Case Narrative. Following the cover page are Comments, QC Report, QC Summary, Data Report, Hits Report, totaling 22 page[s].

194600  
*Good*



MONTGOMERY WATSON LABORATORIES

CHAIN OF CUSTODY RECORD

760 Royal Oaks Ave, Suite 100, Montrovia, CA 91016  
 (626) 386-1100 (800) 566-5227

MVLABS USE ONLY:

LOG IN COMMENTS:

SAMPLES CHECKED/LOGGED IN BY: *HL*

SAMPLE TEMP, RECEIPT AT LAB: *76*

BLUE ICE: FROZEN  PARTIALLY FROZEN  THAWED

TO BE COMPLETED BY SAMPLER:

COMPANY / PROJECT NAME: **KERRMCGEE-MP**  
 PROJECT JOB # / P.O.#: **Quarterly Groundwater Sampling**  
 SCHEDULE: **Schedule B**  
 Sampler: **Michele Brown**  
 Location: **Tronox LLC - Henderson Plant**  
 Address: **PO Box 55, Henderson, NV 89009**  
 Contact: **Susan Crowley (702) 651-2234**

REFER TO ATTACHED BOTTLE ORDER FOR ANALYSES

(check for yes)

ANALYSES REQUIRED (mark an 'X' in all tests required for each sample line)

TIME	DATE	LOCATION	IDENTIFIER, STATE ID#	MATRIX	GRAB	COMP	CR 6010	DH 9040	TDS	ClO4	CRVI 7196	NO3 9056	ClO3 9056	See Bottle Order	SAMPLER Comments
6:55	1-29-07		PC-123	RGW	X		X	X	X	X					2 Bottles
7:11	1-29-07		PC-124	RGW	X		X	X	X	X					2 Bottles
7:24	1-29-07		PC-125	RGW	X		X	X	X	X					2 Bottles
7:37	1-29-07		PC-126	RGW	X		X	X	X	X					2 Bottles
7:51	1-29-07		PC-127	RGW	X		X	X	X	X					2 Bottles
8:02	1-29-07		PC-128	RGW	X		X	X	X	X					2 Bottles
8:17	1-29-07		PC-129	RGW	X		X	X	X	X					2 Bottles
8:33	1-29-07		PC-130	RGW	X		X	X	X	X					2 Bottles
8:46	1-29-07		PC-131	RGW	X		X	X	X	X					2 Bottles
9:01	1-29-07		PC-132	RGW	X		X	X	X	X					2 Bottles
9:26	1-29-07		M-96	RGW	X		X	X	X	X					2 Bottles
9:38	1-29-07		M-95	RGW	X		X	X	X	X					2 Bottles

Reported by Volume:

CFW = Chlor(am)inated Finished Water  
 FW = Other Finished Water

Reported by Weight:

SO = Soil  
 SL = Sludge

RGW = Raw Ground Water  
 RSW = Raw Surface Water

CWW = Chlorinated Waste Water  
 WW = Other Waste Water  
 SW = Storm Water

RELINQUISHED BY: *Michele Brown* SIGNATURE  
 RECEIVED BY: *[Signature]* PRINT NAME: Michele Brown  
 COMPANY/TITLE: Veolia Water NA for Tronox LLC - Henderson Plant  
 DATE: 1-29-07 TIME: 12:00 PM  
 RECEIVED BY: *[Signature]*  
 RECEIVED BY: *[Signature]*



MONTGOMERY WATSON LABORATORIES

# CHAIN OF CUSTODY RECORD

750 Royal Oaks Ave, Suite 100, Monrovia, CA 91016

(626) 386-1100 (800) 566-5227

MVLABS USE ONLY:

LOG IN COMMENTS:

SAMPLES CHECKED/LOGGED IN BY: *[Signature]*

SAMPLE TEMP, RECEIPT AT LAB: *[Signature]*

BLUE ICE:  FROZEN  PARTIALLY FROZEN  THAWED

TO BE COMPLETED BY SAMPLER:

COMPANY / PROJECT NAME: **KERRMCGEE-MP**  
 PROJECT JOB # / P.O.#: **Quarterly Groundwater Sampling**  
 Schedule B

SAMPLER: **Michelle Brown**  
 Susan Crowley (702) 651-2234

PROJECT LOCATION: **Tronox LLC - Henderson Plant**  
 PO Box 55  
 Henderson, NV 89009

REFER TO ATTACHED BOTTLE ORDER FOR ANALYSES  (check for yes)

TIME	DATE	LOCATION	IDENTIFIER, STATE ID#	MATRIX *	GRAB	COMP	ANALYSES REQUIRED (mark an 'X' in all tests required for each sample line)							SAMPLER Comments	
							CR 6010	PH 9040	TDS	CLO4	CRVI 7196	NO3 9056	CLO3 9056		See Bottle Order
1031	1-29-07		PC-54	RGW	X		X	X	X	X					2 Bottles
1047	1-29-07		M-48	RGW	X		X	X	X	X					2 Bottles
1119	1-29-07		PC-41	RGW	X		X	X	X	X					2 Bottles
1130	1-29-07		PC-42	RGW	X		X	X	X	X					2 Bottles
1140	1-29-07		PC-43	RGW	X		X	X	X	X					2 Bottles
1215	1-29-07		M-23	RGW	X		X	X	X	X					2 Bottles
1156	1-29-07		PC-37	RGW	X		X	X	X	X					2 Bottles
1229	1-29-07		M-44	RGW	X		X	X	X	X					3 Bottles
1246	1-29-07		M-94	RGW	X		X	X	X	X					3 Bottles
1230	1-29-07		FB-1	RGW	X		X	X	X	X					3 Bottles
	1-29-07		MD-3	RGW	X		X	X	X	X					2 Bottles

\* MATRIX TYPES: **Reported by Volume:**  
 CFW = Chlor(amine)ated Finished Water  
 FW = Other Finished Water

**Reported by Weight:**  
 SO = Soil  
 SL = Sludge

**RGW = Raw Ground Water**  
**RSW = Raw Surface Water**  
**CWW = Chlorinated Waste Water**  
**OWW = Other Waste Water**  
**SW = Storm Water**

RELINQUISHED BY: *Michelle Brown* SIGNATURE

RECEIVED BY: *[Signature]* PRINT NAME: **Michelle Brown**

RELINQUISHED BY: *[Signature]* COMPANY/TITLE: **Veolia Water NA for Tronox LLC - Henderson Plant**

RECEIVED BY: *[Signature]* DATE: **1-29-07** TIME: **12:00PM**

RECEIVED BY: *[Signature]* **SCANNED**



MWH Laboratories, a Division of MWH Americas, Inc.  
750 Royal Oaks Avenue Suite 100  
Monrovia CA 91016 (626) 386-1100 FAX (626) 386-1124

**Bottle Order for Ironox LLC- Henderson**  
Standing

Andrew Eaton..... Your MWL Project Manager  
(626) 386-1125..... Direct Phone/Voice Mail

Client Code KERRMCGEE-MP  
Project Code CLO4  
PO# / Job#  
Blanket PO

SO# 33751 26973 RS

**Sampler: Please Return this Paper with your samples**

Created by 0  
Order Date 12/12/06  
Date Needed by Client  
Date Samples to Arrive at MWL  
SHIP LOCATION

Ship Sample Kits to  
Tronox LLC-Veolia Water  
Gate 1  
8000 West Lake Mead Drive  
Henderson, NV 89015

Send Report to  
Tronox LLC Henderson Plant  
P.O. Box 55  
Henderson, NV 89009

Billing Address  
Ironox LLC  
PO Box 3049  
Livonia, MI 48150

ATTN: Susan Crowley  
PHONE: 702-651-2234

ATTN: Susan Crowley  
PHONE: 702-651-2234  
FAX: 702-651-2310

Quote#

# of Samples Tests

Bottles-Qty for each sample, type & preservative if any

UN#

Important Comments

# of Samples	Tests	Bottles-Qty for each sample, type & preservative if any	UN#	Important Comments
101	CR6010	1 250ml poly acid rinsed + 1ml HNO3 (18%)	UN 2031	QUARTERLY SAMPLING - PLEASE PUT LABELS ON BOTTLES; PLEASE PUT IN 4 COOLERS SINCE SAMPLING TAKES 3-4 DAYS
101	CLO4, TDS, PH9040	1 500 ml poly /no preservative		Q1, Q3, Q4
15	CRV17196	1 125ml poly acid rinsed/ no preservative		NOTIFY LAB AS SOON AS CR-VI COMES IN.- 24HR Int
				TDS count increased to 101 effective 6/16/06.; deleted EC as of 7-14-06

ActiveCode Status Date Shipped Carrier Qty of Coolers Tracking Number Prepared By

SCANNED

THIS MEMORANDUM is an acknowledgement that a Bill of Lading has been issued and is not the Original Bill of Lading, nor a copy or duplicate, covering the property named herein, and is intended solely for filing or record.

SHIPPER'S NUMBER: 142430

RECEIVED subject to the classifications and lawfully filed tariffs in effect on the date of the issue of this Bill of Lading.

**From: TRONOX LLC**

the property described below, in apparent good order, except as noted (contents and condition of contents of packages unknown), marked, consigned, and destined as indicated below, which said Carrier (the word Carrier being understood throughout this contract as meaning any person or corporation in possession of the property under the contract) agrees to carry to its usual place of delivery at said destination, if on its route, otherwise to deliver to another Carrier on the route to said destination. It is mutually agreed, as to each Carrier of all or any said property over all or any portion of said route to destination, and as to each party at any time interested in all or any of said property, that every service to be performed hereunder shall be subject to all the terms and conditions of the Uniform Domestic Straight Bill of Lading set forth (1) in Uniform Freight Classification in effect on the date hereof, if this is a rail or a rail-water shipment, or (2) in the applicable motor carrier classification or tariff if this is a motor carrier shipment.

Shipper hereby certifies that he is familiar with all the terms and conditions of the said Bill of Lading, including those on the back thereof, set forth in the classification or tariff which governs the transportation of this shipment, and the said terms and conditions are hereby agreed to by the Shipper and accepted for himself and his assigns.

CARRIER <b>Federal Express</b>		Date <b>1/29/07</b>	FROM NO. STATION: STATE <b>Henderson, NV 89015</b>	
MWH LABORATORIES 750 Royal Oaks Avenue, Suite #100 Monrovia, CA 91016-3629 Phone: 626-568-6400			Authorization <b>S. CROWLEY</b>	
FREIGHT CHARGES <input checked="" type="checkbox"/> Prepaid <input type="checkbox"/> Collect		FULL NAME OF SHIPPER <b>TRONOX LLC</b>		
N/AR		CUSTOMER PO OR REQ'N NO.	CODE NO. WCN IS <b>1321.10400</b> If it moves between two ports by water, the law requires that the Bill of Lading shall state whether it is Carrier's or Shippers weight.	
SHIPPED FROM <b>Henderson, NV</b>		Subject to Section 7 of Conditions of applicable Bill of Lading, if this shipment is to be delivered to the Consignee without recourse on the Consignor, the Consignor shall sign the following statement: The Carrier shall not make delivery of this shipment without payment of freight and all other lawful charges. <b>TRONOX LLC</b>		
LINE NO.	DESCRIPTION AND CLASSIFICATION	STOCK NO.	TOTAL QUANTITY	The description and weight indicated on this Bill of Lading are correct. Subject to verification by the Governing Weighing and Inspection Bureau according to Agreement.
	<b>Quarterly Monitoring Well Samples</b>  <b>One ice chest @ 54 lbs</b> <b>One ice chest @ 51 lbs</b>  <b>Not Regulated</b>		<b>2 COOLERS</b>	
TRUCK SHIPMENTS		FOR CHEMICAL EMERGENCY-SPILL, LEAK, FIRE, EXPOSURE OR ACCIDENT CALL CHEMTREC - DAY OR NIGHT 800-424-9300 483-7616 IN DISTRICT OF COLUMBIA 202-483-7616 FROM OUTSIDE THE CONTINENTAL US.		
PLACARDS OFFERED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		PLACARDS ACCEPTED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		
NUMBER OF PACKAGES	GROSS WEIGHT	TARE WEIGHT	NET WEIGHT	
		<b>0</b>		
<b>2</b>	<b>TOTAL GROSS WEIGHT 105</b>	<b>TOTAL TARE WEIGHT 0</b>	<b>105</b>	
NOTE: Where the rate is dependent on value, shippers are required to state specifically in writing the agreed or declared value of the property. The agreed or declared value of the property is hereby specifically stated by the shipper to be not exceeding _____ per				"Shippers imprint in lieu of stamp; not a part of Bill of Lading approved by the Interstate Commerce Commission"

THIS IS TO CERTIFY THAT THE ABOVE-NAMED MATERIALS ARE PROPERLY CLASSIFIED, DESCRIBED, PACKAGED, MARKED AND LABELED AND ARE IN PROPER CONDITION FOR TRANSPORTATION ACCORDING TO THE APPLICABLE REGULATIONS OF THE DEPARTMENT OF TRANSPORTATION.

TRONOX LLC Shipper permanent post office address of shipper, PO Box 268857, Oklahoma City, OK 73126-8857	PER <b>Judi Durkin</b>	AGENT	PER
---	---------------------------	-------	-----

From: Origin ID: LASA (702)651-2230  
TRONOX LLC  
TRONOX LLC  
8000 LAKE MEAD PARKWAY  
SHIPPING DEPARTMENT  
HENDERSON, NV 89015



Ship Date: 29JAN07  
ActWgt: 54 LB  
System#: 2274147/INET2600  
Account#: S \*\*\*\*\*


Delivery Address Bar Code



Ref # MSO #142430 - 54 lbs.  
Invoice #  
PO #  
Dept #

SHIP TO: (626)568-6400 **BILL SENDER**  
**ATTN: MONTGOMERY LABS RECEIVING**  
**MONTGOMERY WATSON LABS**  
**750 ROYAL OAKS DR # 100**

**MONROVIA, CA 910163629**



**PRIORITY OVERNIGHT**

**TUE**

TRK# 7929 2380 3670


FORM 0201

Deliver By:  
30JAN07

**BUR**

91016 -CA-US

**QZ WHPA**



Shipping Label: Your shipment is complete

1. Use the 'Print' feature from your browser to send this page to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

**Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.**

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$500, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

SCANNED

**MWH Laboratories**  
 750 Royal Oaks Drive, Monrovia, CA 91016  
 PHONE: 626-386-1100/FAX: 626-386-1101

ACKNOWLEDGMENT OF SAMPLES RECEIVED

Tronox LLC - Henderson  
 PO Box 55  
 Henderson, NV 89009  
 Attn: Susan Crowley  
 Phone: 702-651-2234

Customer Code: KERRMCGEE-MP  
 PO#: Susan Crowley PO  
 Group#: 194600  
 Project#: CLO4  
 Proj Mgr: Andrew Eaton  
 Phone: (626) 386-1125

The following samples were received from you on **01/30/07**. They have been scheduled for the tests listed beside each sample. If this information is incorrect, please contact your service representative. Thank you for using MWH Laboratories.

Sample#	Sample Id	Tests Scheduled	Matrix	Sample Date
2701300116	PC-123	CLO4 CR6010	Water P PH9040	29-jan-2007 06:55:00 TDS
2701300118	PC-124	CLO4 CR6010	Water P PH9040	29-jan-2007 07:11:00 TDS
2701300119	PC-125	CLO4 CR6010	Water P PH9040	29-jan-2007 07:24:00 TDS
2701300120	PC-126	CLO4 CR6010	Water P PH9040	29-jan-2007 07:37:00 TDS
2701300121	PC-127	CLO4 CR6010	Water P PH9040	29-jan-2007 07:51:00 TDS
2701300122	PC-128	CLO4 CR6010	Water P PH9040	29-jan-2007 08:02:00 TDS
2701300123	PC-129	CLO4 CR6010	Water P PH9040	29-jan-2007 08:17:00 TDS
2701300124	PC-130	CLO4 CR6010	Water P PH9040	29-jan-2007 08:33:00 TDS
2701300125	PC-131	CLO4 CR6010	Water P PH9040	29-jan-2007 08:46:00 TDS
2701300126	PC-132	CLO4 CR6010	Water P PH9040	29-jan-2007 09:01:00 TDS
2701300127	M-96	CLO4 CR6010	Water P PH9040	29-jan-2007 09:26:00 TDS
2701300128	M-95	CLO4 CR6010	Water P PH9040	29-jan-2007 09:38:00 TDS
2701300129	PC-54	CLO4 CR6010	Water P PH9040	29-jan-2007 10:31:00 TDS
2701300130	M-48	CLO4 CR6010	Water P PH9040	29-jan-2007 10:47:00 TDS
2701300131	PC-71	CLO4 CR6010	Water P PH9040	29-jan-2007 11:19:00 TDS
2701300132	PC-72	CLO4 CR6010	Water P PH9040	29-jan-2007 11:30:00 TDS
2701300133	PC-73	CLO4 CR6010	Water P PH9040	29-jan-2007 11:40:00 TDS

Tronox LLC - Henderson  
 PO Box 55  
 Henderson, NV 89009  
 Attn: Susan Crowley  
 Phone: 702-651-2234

Customer Code: KERRMCGEE-MP  
 PO#: Susan Crowley PO  
 Group#: 194600  
 Project#: CLO4  
 Proj Mgr: Andrew Eaton  
 Phone: (626) 386-1125

Sample#	Sample Id	Tests Scheduled	Matrix	Sample Date
2701300134	M-23	CLO4 CR6010	Water P	29-jan-2007 12:15:00
2701300135	PC-37	CLO4 CR6010	Water P	29-jan-2007 11:56:00
2701300137	M-44	CLO4 CR6010	Water P	29-jan-2007 12:29:00
2701300138	M-94	CLO4 CR6010	Water P	29-jan-2007 12:46:00
2701300139	FB-1	CLO4 CR6010	Water P	29-jan-2007 12:30:00
2701300140	MD-3	CLO4 CR6010	Water P	29-jan-2007 00:00:00

Test Acronym Description

Test Acronym	Description
CLO4	Perchlorate
CR6010	Chromium, Total, ICAP
CRVI7196	Hexavalent chromium (Cr VI)
P	Metals sample pH
PH9040	PH (H3=past HT, not compliant)
T	Metals Turbidity
TDS	Total Dissolved Solid (TDS)





**Client Specific Comments**

I hereby certify that all laboratory analytical data was generated by a laboratory certified by the NDEP for each constituent and media presented herein.

Client Signature: \_\_\_\_\_

(QC Ref#: 2701300127)

Test: Total Dissolved Solid (TDS) (SM 2540C)

H2- Initial analysis within holding time. Reanalysis for the required dilution was past holding time.

(QC Ref#: 2701300128)

Test: Total Dissolved Solid (TDS) (SM 2540C)

H2- Initial analysis within holding time. Reanalysis for the required dilution was past holding time.

(QC Ref#: 2701300130)

Test: Total Dissolved Solid (TDS) (SM 2540C)

H2- Initial analysis within holding time. Reanalysis for the required dilution was past holding time.

(QC Ref#: 2701300132)

Test: Total Dissolved Solid (TDS) (SM 2540C)

H2- Initial analysis within holding time. Reanalysis for the required dilution was past holding time.

(QC Ref#: 2701300137)

Test: Hexavalent chromium (Cr VI) (ML/EPS 7196)

H3- Sample was received and analyzed past holding time. Data not acceptable for regulatory compliance.

(QC Ref#: 2701300138)

Test: Hexavalent chromium (Cr VI) (ML/EPS 7196)

H3- Sample was received and analyzed past holding time. Data not acceptable for regulatory compliance.

(QC Ref#: 2701300139)



**MWH Laboratories**

*A Division of MWH Americas, Inc.*

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

---

Report  
Comments  
#194600

**Test: Hexavalent chromium (Cr VI) (ML/EPS 7196)**

H3- Sample was received and analyzed past holding time. Data not acceptable for regulatory compliance.



750 Royal Oaks Drive, Suite 100  
 Monrovia, California 91016-3829  
 Tel: 626 386 1100  
 Fax: 626 386 1101  
 1 800 566 LABS (1 800 566 5227)

Laboratory  
 Hits Report  
 #194600

Tronox LLC - Henderson  
 Susan Crowley  
 PO Box 55  
 Henderson , NV 89009

Samples Received  
 30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	<b>2701300116</b>	<b>PC-123</b>				
02/02/07		Chromium, Total, ICAP	1.3		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
01/31/07		Perchlorate	360000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	6640	500	mg/l	10
	<b>2701300118</b>	<b>PC-124</b>				
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
01/31/07		Perchlorate	4450		ug/l	400
02/01/07		Total Dissolved Solid (TDS)	5650	500	mg/l	10
	<b>2701300119</b>	<b>PC-125</b>				
02/02/07		Chromium, Total, ICAP	0.022		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
01/31/07		Perchlorate	5520		ug/l	400
02/01/07		Total Dissolved Solid (TDS)	6100	500	mg/l	10
	<b>2701300120</b>	<b>PC-126</b>				
02/02/07		Chromium, Total, ICAP	0.081		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.8		Units	0.0010
01/31/07		Perchlorate	10500		ug/l	2000
02/01/07		Total Dissolved Solid (TDS)	7240	500	mg/l	10
	<b>2701300121</b>	<b>PC-127</b>				

SUMMARY OF POSITIVE DATA ONLY.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#194600

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	2701300121	PC-127				
02/02/07		Chromium, Total, ICAP	1.2		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
01/31/07		Perchlorate	337000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	6940	500	mg/l	10
	2701300122	PC-128				
02/02/07		Chromium, Total, ICAP	0.091		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.1		Units	0.0010
01/31/07		Perchlorate	157000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	5050	500	mg/l	10
	2701300123	PC-129				
02/02/07		Chromium, Total, ICAP	0.57		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.0		Units	0.0010
01/31/07		Perchlorate	407000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	7110	500	mg/l	10
	2701300124	PC-130				
02/02/07		Chromium, Total, ICAP	0.62		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.0		Units	0.0010
01/31/07		Perchlorate	434000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	6470	500	mg/l	10

SUMMARY OF POSITIVE DATA ONLY.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#194600

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	2701300124	PC-130				
	2701300125	PC-131				
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.2		Units	0.0010
01/31/07		Perchlorate	16700		ug/l	2000
02/01/07		Total Dissolved Solid (TDS)	9440	500	mg/l	10
	2701300126	PC-132				
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.2		Units	0.0010
01/31/07		Perchlorate	4880		ug/l	800
02/01/07		Total Dissolved Solid (TDS)	8800	500	mg/l	10
	2701300127	M-96				
02/02/07		Chromium, Total, ICAP	1.9		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.0		Units	0.0010
01/31/07		Perchlorate	446000		ug/l	40000
02/15/07		Total Dissolved Solid (TDS)	7310	500	mg/l	10
	2701300128	M-95				
02/02/07		Chromium, Total, ICAP	1.7		mg/l	0.020
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	8.0		Units	0.0010
01/31/07		Perchlorate	623000		ug/l	40000
02/15/07		Total Dissolved Solid (TDS)	7830	500	mg/l	10

SUMMARY OF POSITIVE DATA ONLY.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#194600

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	2701300128	M-95				
	2701300129	PC-54				
02/02/07	Chromium, Total, ICAP		2.8		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
02/02/07	PH (H3=past HT, not compliant)		8.1		Units	0.0010
01/31/07	Perchlorate		337000		ug/l	20000
02/02/07	Total Dissolved Solid (TDS)		6790	500	mg/l	10
	2701300130	M-48				
02/02/07	Chromium, Total, ICAP		1.1		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		8.1		Units	0.0010
01/31/07	Perchlorate		144000		ug/l	20000
02/15/07	Total Dissolved Solid (TDS)		7260	500	mg/l	10
	2701300131	PC-71				
02/02/07	Chromium, Total, ICAP		0.50		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		7.9		Units	0.0010
01/31/07	Perchlorate		577000		ug/l	40000
02/02/07	Total Dissolved Solid (TDS)		8130	500	mg/l	10
	2701300132	PC-72				
02/02/07	Chromium, Total, ICAP		0.33		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		7.9		Units	0.0010

SUMMARY OF POSITIVE DATA ONLY.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#194600

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	<b>2701300132</b>	<b>PC-72</b>				
01/31/07	Perchlorate		376000		ug/l	20000
02/15/07	Total Dissolved Solid (TDS)		7240	500	mg/l	10
	<b>2701300133</b>	<b>PC-73</b>				
02/02/07	Chromium, Total, ICAP		0.38		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		7.9		Units	0.0010
01/31/07	Perchlorate		397000		ug/l	20000
02/02/07	Total Dissolved Solid (TDS)		6910	500	mg/l	10
	<b>2701300134</b>	<b>M-23</b>				
02/02/07	Chromium, Total, ICAP		0.82		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		8.0		Units	0.0010
01/31/07	Perchlorate		449000		ug/l	40000
02/02/07	Total Dissolved Solid (TDS)		5710	500	mg/l	10
	<b>2701300135</b>	<b>PC-37</b>				
02/02/07	Chromium, Total, ICAP		0.15		mg/l	0.020
01/31/07	Metals digestion performed.		Y		Yes/No	
01/31/07	PH (H3=past HT, not compliant)		7.9		Units	0.0010
01/31/07	Perchlorate		284000		ug/l	40000
02/02/07	Total Dissolved Solid (TDS)		7040	500	mg/l	10
	<b>2701300137</b>	<b>M-44</b>				
02/02/07	Chromium, Total, ICAP		0.91		mg/l	0.020

SUMMARY OF POSITIVE DATA ONLY.



**MWH Laboratories**  
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#194600

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
30-jan-2007 13:44:27

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	<b>2701300137</b>	<b>M-44</b>				
01/30/07		Hexavalent chromium (Cr VI)	1.0		mg/l	0.050
01/31/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
01/31/07		Perchlorate	826000		ug/l	40000
02/02/07		Total Dissolved Solid (TDS)	8370	500	mg/l	10
	<b>2701300138</b>	<b>M-94</b>				
02/02/07		Chromium, Total, ICAP	0.76		mg/l	0.020
01/30/07		Hexavalent chromium (Cr VI)	0.89		mg/l	0.050
02/01/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	7.9		Units	0.0010
02/02/07		Perchlorate	698000		ug/l	40000
02/02/07		Total Dissolved Solid (TDS)	7640	500	mg/l	10
	<b>2701300139</b>	<b>FB-1</b>				
02/01/07		Metals digestion performed.	Y		Yes/No	
01/31/07		PH (H3=past HT, not compliant)	6.2		Units	0.0010
02/02/07		Total Dissolved Solid (TDS)	8010	500	mg/l	10
	<b>2701300140</b>	<b>MD-3</b>				
02/02/07		Chromium, Total, ICAP	0.57		mg/l	0.020
02/01/07		Metals digestion performed.	Y		Yes/No	
02/02/07		PH (H3=past HT, not compliant)	8.1		Units	0.0010
02/01/07		Perchlorate	397000		ug/l	20000
02/01/07		Total Dissolved Solid (TDS)	6220	500	mg/l	10

SUMMARY OF POSITIVE DATA ONLY.





750 Royal Oaks Drive, Suite 100  
 Monrovia, California 91016-3629  
 Tel: 626 386 1100  
 Fax: 626 386 1101  
 1 800 566 LABS (1 800 566 5227)

Laboratory  
 Data Report  
 #194600

Tronox LLC - Henderson  
 Susan Crowley  
 PO Box 55  
 Henderson, NV 89009

Samples Received  
 01/30/07

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>PC-123 (2701300116)                      Sampled on 01/29/07 06:55</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	360000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	1.3	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350722	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	6640	mg/l	10	1
<b>PC-124 (2701300118)                      Sampled on 01/29/07 07:11</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	4450	ug/l	400	100
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	ND	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350722	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	5650	mg/l	10	1
<b>PC-125 (2701300119)                      Sampled on 01/29/07 07:24</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	5520	ug/l	400	100
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.022	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350722	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	6100	mg/l	10	1
<b>PC-126 (2701300120)                      Sampled on 01/29/07 07:37</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	10500	ug/l	2000	500
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.081	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350723	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.8	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	7240	mg/l	10	1



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Data Report  
#194600

Tronox LLC - Henderson  
(continued)

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>PC-127 (2701300121)                    Sampled on 01/29/07 07:51</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	337000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	1.2	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	6940	mg/l	10	1
<b>PC-128 (2701300122)                    Sampled on 01/29/07 08:02</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	157000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.091	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.1	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	5050	mg/l	10	1
<b>PC-129 (2701300123)                    Sampled on 01/29/07 08:17</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	407000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.57	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.0	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	7110	mg/l	10	1
<b>PC-130 (2701300124)                    Sampled on 01/29/07 08:33</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	434000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.62	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350723	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.0	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	6470	mg/l	10	1
<b>PC-131 (2701300125)                    Sampled on 01/29/07 08:46</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	16700	ug/l	2000	500
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	ND	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350722	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.2	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	9440	mg/l	10	1



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Data Report  
#194600

Tronox LLC - Henderson  
(continued)

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>PC-132 (2701300126)</b>		<b>Sampled on 01/29/07 09:01</b>						
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	4880	ug/l	800	200
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	ND	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350723	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.2	Units	0.0010	1
02/01/07	02/01/07 18:02	351213	( SM 2540C )	Total Dissolved Solid (TDS)	8800	mg/l	10	1
<b>M-96 (2701300127)</b>		<b>Sampled on 01/29/07 09:26</b>						
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	446000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	1.9	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.0	Units	0.0010	1
02/02/07	02/15/07 17:03	353634	( SM 2540C )	Total Dissolved Solid (TDS)	7310 (H2)	mg/l	10	1
<b>M-95 (2701300128)</b>		<b>Sampled on 01/29/07 09:38</b>						
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	623000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	1.7	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350723	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.0	Units	0.0010	1
02/02/07	02/15/07 17:03	353634	( SM 2540C )	Total Dissolved Solid (TDS)	7830 (H2)	mg/l	10	1
<b>PC-54 (2701300129)</b>		<b>Sampled on 01/29/07 10:31</b>						
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	337000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	2.8	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	02/02/07 00:00	351090	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.1	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	6790	mg/l	10	1
<b>M-48 (2701300130)</b>		<b>Sampled on 01/29/07 10:47</b>						
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	144000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	1.1	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350722	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.1	Units	0.0010	1
02/02/07	02/15/07 17:03	353634	( SM 2540C )	Total Dissolved Solid (TDS)	7260 (H2)	mg/l	10	1



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
 Monrovia, California 91016-3629  
 Tel: 626 386 1100  
 Fax: 626 386 1101  
 1 800 566 LABS (1 800 566 5227)

Laboratory  
 Data Report  
 #194600

Tronox LLC - Henderson  
 (continued)

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>PC-71 (2701300131)                    Sampled on 01/29/07 11:19</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	577000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.50	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	8130	mg/l	10	1
<b>PC-72 (2701300132)                    Sampled on 01/29/07 11:30</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	376000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.33	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/15/07 17:03	353634	( SM 2540C )	Total Dissolved Solid (TDS)	7240 (H2)	mg/l	10	1
<b>PC-73 (2701300133)                    Sampled on 01/29/07 11:40</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	397000	ug/l	20000	5000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.38	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	6910	mg/l	10	1
<b>M-23 (2701300134)                    Sampled on 01/29/07 12:15</b>								
	01/31/07 00:00	350724	( EPA 314 )	Perchlorate	449000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.82	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.0	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	5710	mg/l	10	1
<b>PC-37 (2701300135)                    Sampled on 01/29/07 11:56</b>								
	01/31/07 00:00	350795	( EPA 314 )	Perchlorate	284000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.15	mg/l	0.020	2
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	7040	mg/l	10	1



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
 Monrovia, California 91016-3829  
 Tel: 626 386 1100  
 Fax: 626 386 1101  
 1 800 566 LABS (1 800 566 5227)

Laboratory  
 Data Report  
 #194600

Tronox LLC - Henderson  
 (continued)

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>M-44 (2701300137)</b>		<b>Sampled on 01/29/07 12:29</b>						
	01/31/07 00:00	350795	( EPA 314 )	Perchlorate	826000	ug/l	40000	10000
	02/02/07 00:00	351230	( ML/EPA 6010B )	Chromium, Total, ICAP	0.91	mg/l	0.020	2
	01/30/07 19:05	350518	( ML/EPS 7196 )	Hexavalent chromium (Cr VI)	1.0 (H3)	mg/l	0.050	10
	01/31/07 16:46		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	8370	mg/l	10	1
<b>M-94 (2701300138)</b>		<b>Sampled on 01/29/07 12:46</b>						
	02/02/07 00:00	351057	( EPA 314 )	Perchlorate	698000	ug/l	40000	10000
	02/02/07 00:00	351232	( ML/EPA 6010B )	Chromium, Total, ICAP	0.76	mg/l	0.020	2
	01/30/07 19:02	350521	( ML/EPS 7196 )	Hexavalent chromium (Cr VI)	0.89 (H3)	mg/l	0.050	10
	02/01/07 10:22		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350720	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	7.9	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	7640	mg/l	10	1
<b>FB-1 (2701300139)</b>		<b>Sampled on 01/29/07 12:30</b>						
	02/01/07 00:00	350795	( EPA 314 )	Perchlorate	ND	ug/l	4.0	2
	02/02/07 00:00	351232	( ML/EPA 6010B )	Chromium, Total, ICAP	ND	mg/l	0.020	2
	01/30/07 19:01	350521	( ML/EPS 7196 )	Hexavalent chromium (Cr VI)	ND (H3)	mg/l	0.0050	1
	02/01/07 10:22		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	01/31/07 00:00	350723	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	6.2	Units	0.0010	1
02/02/07	02/02/07 15:37	351498	( SM 2540C )	Total Dissolved Solid (TDS)	8010	mg/l	10	1
<b>MD-3 (2701300140)</b>		<b>Sampled on 01/29/07 00:00</b>						
	02/01/07 00:00	350795	( EPA 314 )	Perchlorate	397000	ug/l	20000	5000
	02/02/07 00:00	351232	( ML/EPA 6010B )	Chromium, Total, ICAP	0.57	mg/l	0.020	2
	02/01/07 10:22		( EPA 200 Prep )	Metals digestion performed.	Y	Yes/No	0	1
	02/02/07 00:00	351090	( ML/EPA 9040B )	PH (H3=past HT, not compliant)	8.1	Units	0.0010	1
02/01/07	02/01/07 16:45	351211	( SM 2540C )	Total Dissolved Solid (TDS)	6220	mg/l	10	1



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Summary  
#194600

Tronox LLC - Henderson

QC Ref #350518 - Hexavalent chromium (Cr VI) Analysis Date: 01/30/2007

2701300137 M-44 Analyzed by: wbh

QC Ref #350521 - Hexavalent chromium (Cr VI) Analysis Date: 01/30/2007

2701300138 M-94 Analyzed by: wbh  
2701300139 FB-1 Analyzed by: wbh

QC Ref #350720 - PH (H3=past HT, not compliant) Analysis Date: 01/31/2007

2701300121 PC-127 Analyzed by: jxl  
2701300122 PC-128 Analyzed by: jxl  
2701300123 PC-129 Analyzed by: jxl  
2701300127 M-96 Analyzed by: jxl  
2701300131 PC-71 Analyzed by: jxl  
2701300132 PC-72 Analyzed by: jxl  
2701300133 PC-73 Analyzed by: jxl  
2701300134 M-23 Analyzed by: jxl  
2701300135 PC-37 Analyzed by: jxl  
2701300137 M-44 Analyzed by: jxl  
2701300138 M-94 Analyzed by: jxl

QC Ref #350722 - PH (H3=past HT, not compliant) Analysis Date: 01/31/2007

2701300116 PC-123 Analyzed by: jxl  
2701300118 PC-124 Analyzed by: jxl  
2701300119 PC-125 Analyzed by: jxl  
2701300125 PC-131 Analyzed by: jxl  
2701300130 M-48 Analyzed by: jxl

QC Ref #350723 - PH (H3=past HT, not compliant) Analysis Date: 01/31/2007

2701300120 PC-126 Analyzed by: jxl  
2701300124 PC-130 Analyzed by: jxl  
2701300126 PC-132 Analyzed by: jxl  
2701300128 M-95 Analyzed by: jxl  
2701300139 FB-1 Analyzed by: jxl



**MWH Laboratories**  
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Summary  
#194600

Tronox LLC - Henderson  
(continued)

QC Ref #350724 - Perchlorate

Analysis Date: 01/31/2007

2701300116	PC-123	Analyzed by: raja
2701300118	PC-124	Analyzed by: raja
2701300119	PC-125	Analyzed by: raja
2701300120	PC-126	Analyzed by: raja
2701300121	PC-127	Analyzed by: raja
2701300122	PC-128	Analyzed by: raja
2701300123	PC-129	Analyzed by: raja
2701300124	PC-130	Analyzed by: raja
2701300125	PC-131	Analyzed by: raja
2701300126	PC-132	Analyzed by: raja
2701300127	M-96	Analyzed by: raja
2701300128	M-95	Analyzed by: raja
2701300129	PC-54	Analyzed by: raja
2701300130	M-48	Analyzed by: raja
2701300131	PC-71	Analyzed by: raja
2701300132	PC-72	Analyzed by: raja
2701300133	PC-73	Analyzed by: raja
2701300134	M-23	Analyzed by: raja

QC Ref #350795 - Perchlorate

Analysis Date: 01/31/2007

2701300135	PC-37	Analyzed by: raja
2701300137	M-44	Analyzed by: raja
2701300139	FB-1	Analyzed by: raja
2701300140	MD-3	Analyzed by: raja

QC Ref #351057 - Perchlorate

Analysis Date: 02/02/2007

2701300138	M-94	Analyzed by: raja
------------	------	-------------------

QC Ref #351090 - PH (H3=past HT, not compliant) Analysis Date: 02/02/2007

2701300129	PC-54	Analyzed by: jxl
2701300140	MD-3	Analyzed by: jxl



750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Summary  
#194600

Tronox LLC - Henderson  
(continued)

QC Ref #351211 - Total Dissolved Solid (TDS) Analysis Date: 02/01/2007

2701300140 MD-3 Analyzed by: anh

QC Ref #351213 - Total Dissolved Solid (TDS) Analysis Date: 02/01/2007

2701300116	PC-123	Analyzed by: anh
2701300118	PC-124	Analyzed by: anh
2701300119	PC-125	Analyzed by: anh
2701300120	PC-126	Analyzed by: anh
2701300121	PC-127	Analyzed by: anh
2701300122	PC-128	Analyzed by: anh
2701300123	PC-129	Analyzed by: anh
2701300124	PC-130	Analyzed by: anh
2701300125	PC-131	Analyzed by: anh
2701300126	PC-132	Analyzed by: anh

QC Ref #351230 - Chromium, Total, ICAP Analysis Date: 02/02/2007

2701300116	PC-123	Analyzed by: wbh
2701300118	PC-124	Analyzed by: wbh
2701300119	PC-125	Analyzed by: wbh
2701300120	PC-126	Analyzed by: wbh
2701300121	PC-127	Analyzed by: wbh
2701300122	PC-128	Analyzed by: wbh
2701300123	PC-129	Analyzed by: wbh
2701300124	PC-130	Analyzed by: wbh
2701300125	PC-131	Analyzed by: wbh
2701300126	PC-132	Analyzed by: wbh
2701300127	M-96	Analyzed by: wbh
2701300128	M-95	Analyzed by: wbh
2701300129	PC-54	Analyzed by: wbh
2701300130	M-48	Analyzed by: wbh
2701300131	PC-71	Analyzed by: wbh
2701300132	PC-72	Analyzed by: wbh
2701300133	PC-73	Analyzed by: wbh
2701300134	M-23	Analyzed by: wbh
2701300135	PC-37	Analyzed by: wbh
2701300137	M-44	Analyzed by: wbh





# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Summary  
#194600

Tronox LLC - Henderson  
(continued)

---

**QC Ref #351232 - Chromium, Total, ICAP**

**Analysis Date: 02/02/2007**

2701300138	M-94	Analyzed by: wbh
2701300139	FB-1	Analyzed by: wbh
2701300140	MD-3	Analyzed by: wbh

**QC Ref #351498 - Total Dissolved Solid (TDS)**

**Analysis Date: 02/02/2007**

2701300129	PC-54	Analyzed by: anh
2701300131	PC-71	Analyzed by: anh
2701300133	PC-73	Analyzed by: anh
2701300134	M-23	Analyzed by: anh
2701300135	PC-37	Analyzed by: anh
2701300137	M-44	Analyzed by: anh
2701300138	M-94	Analyzed by: anh
2701300139	FB-1	Analyzed by: anh

**QC Ref #353634 - Total Dissolved Solid (TDS)**

**Analysis Date: 02/15/2007**

2701300127	M-96	Analyzed by: anh
2701300128	M-95	Analyzed by: anh
2701300130	M-48	Analyzed by: anh
2701300132	PC-72	Analyzed by: anh



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#194600

Tronox LLC - Henderson

## QC Ref #350518 Hexavalent chromium (Cr VI)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300138	MGL		( 0-0 )	
LCS1	Hexavalent chromium (Cr VI)	0.050	0.053	MGL	106.0	( 85-115 )	
LCS2	Hexavalent chromium (Cr VI)	0.050	0.050	MGL	100.0	( 85-115 )	
MBLK	Hexavalent chromium (Cr VI)	ND	<0.0050	MGL			
MRL_CHK	Hexavalent chromium (Cr VI)	0.005	0.005	MGL	100.0	( 50-150 )	
MS	Hexavalent chromium (Cr VI)	0.50	0.49	MGL	98.0	( 70-130 )	
MSD	Hexavalent chromium (Cr VI)	0.50	0.48	MGL	96.0	( 70-130 )	

## QC Ref #350521 Hexavalent chromium (Cr VI)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300138	MGL		( 0-0 )	
LCS1	Hexavalent chromium (Cr VI)	0.050	0.053	MGL	106.0	( 85-115 )	
LCS2	Hexavalent chromium (Cr VI)	0.050	0.050	MGL	100.0	( 85-115 )	
MBLK	Hexavalent chromium (Cr VI)	ND	<0.0050	MGL			
MRL_CHK	Hexavalent chromium (Cr VI)	0.005	0.005	MGL	100.0	( 50-150 )	
MS	Hexavalent chromium (Cr VI)	0.50	0.49	MGL	98.0	( 70-130 )	
MSD	Hexavalent chromium (Cr VI)	0.50	0.48	MGL	96.0	( 70-130 )	

## QC Ref #350720 PH (H3=past HT, not compliant)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
DUP	PH (H3=past HT, not compliant)	4.89	4.90	UNIT		( 0-20 )	0.2

## QC Ref #350722 PH (H3=past HT, not compliant)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
DUP	PH (H3=past HT, not compliant)	8.14	8.17	UNIT		( 0-20 )	0.4

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#194600

Tronox LLC - Henderson  
(continued)

## QC Ref #350723 PH (H3=past HT, not compliant)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
DUP	PH (H3=past HT, not compliant)	7.76	7.77	UNIT		( 0-20 )	0.1

## QC Ref #350724 Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 27	01300030	UGL		( 0-0 )	
LCS1	Perchlorate	25.0	23.6	UGL	94.4	( 85-115 )	
LCS2	Perchlorate	25.0	24.0	UGL	96.0	( 85-115 )	
LCS3	Perchlorate	4	3.27	UGL	81.8	( 75-125 )	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	23.0	UGL	92.0	( 70-130 )	
MSD	Perchlorate	25.0	21.6	UGL	86.4	( 70-130 )	
RPD_LCS	Perchlorate	94.400	96.000	UGL	1.7	( 0-20 )	
RPD_MS	Perchlorate	92.000	86.400	UGL	6.3	( 0-20 )	

## QC Ref #350795 Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 27	01300715	UGL		( 0-0 )	
LCS1	Perchlorate	25.0	23.7	UGL	94.8	( 85-115 )	
LCS2	Perchlorate	25.0	24.6	UGL	98.4	( 85-115 )	
LCS3	Perchlorate	4	3.48	UGL	87.0	( 75-125 )	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	26.9	UGL	107.6	( 70-130 )	
MSD	Perchlorate	25.0	22.6	UGL	90.4	( 70-130 )	
RPD_LCS	Perchlorate	94.800	98.400	UGL	3.7	( 0-20 )	
RPD_MS	Perchlorate	107.600	90.400	UGL	17.4	( 0-20 )	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates  
are advisory only, unless otherwise specified in the method.



**MWH Laboratories**  
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#194600

Tronox LLC - Henderson  
(continued)

QC Ref #351057

Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 27	02010351	UGL		( 0-0 )	
LCS1	Perchlorate	25.0	24.3	UGL	97.2	( 85-115 )	
LCS2	Perchlorate	25.0	24.9	UGL	99.6	( 85-115 )	
LCS3	Perchlorate	4	3.67	UGL	91.8	( 75-125 )	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	20.9	UGL	83.6	( 70-130 )	
MSD	Perchlorate	25.0	21.2	UGL	84.8	( 70-130 )	
RPD_LCS	Perchlorate	97.200	99.600	UGL	2.4	( 0-20 )	
RPD_MS	Perchlorate	83.600	84.800	UGL	1.4	( 0-20 )	

QC Ref #351090

PH (H3=past HT, not compliant)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
DUP	PH (H3=past HT, not compliant)	7.61	7.67	UNIT		( 0-20 )	0.8

QC Ref #351211

Total Dissolved Solid (TDS)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01260076	MGL		( 0-0 )	
DUP	Total Dissolved Solid (TDS)	146	140	MGL		( 0-10 )	4.2
LCS1	Total Dissolved Solid (TDS)	175	172	MGL	98.3	( 85-115 )	
LCS2	Total Dissolved Solid (TDS)	700	696	MGL	99.4	( 85-115 )	
MBLK	Total Dissolved Solid (TDS)	ND	<10	MGL			
MRL_CHK	Total Dissolved Solid (TDS)	10.0	9	MGL	90.0	( 50-150 )	
RPD_LCS	Total Dissolved Solid (TDS)	98.286	99.429	MGL	1.2	( 0-20 )	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates  
are advisory only, unless otherwise specified in the method.



**MWH Laboratories**  
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#194600

Tronox LLC - Henderson  
(continued)

**QC Ref #351213                      Total Dissolved Solid (TDS)**

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300342	MGL		( 0-0 )	
DUP	Total Dissolved Solid (TDS)	4500	4480	MGL		( 0-10 )	0.4
LCS1	Total Dissolved Solid (TDS)	175	168	MGL	96.0	( 85-115 )	
LCS2	Total Dissolved Solid (TDS)	700	678	MGL	96.9	( 85-115 )	
MBLK	Total Dissolved Solid (TDS)	ND	<10	MGL			
MRL_CHK	Total Dissolved Solid (TDS)	10.0	9	MGL	90.0	( 50-150 )	
RPD_LCS	Total Dissolved Solid (TDS)	96.000	96.857	MGL	0.9	( 0-20 )	

**QC Ref #351230                      Chromium, Total, ICAP**

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300124	MGL		( 0-0 )	
LCS1	Chromium, Total, ICAP	1.00	0.954	MGL	95.4	( 85-115 )	
LCS2	Chromium, Total, ICAP	1.00	0.971	MGL	97.1	( 85-115 )	
MBLK	Chromium, Total, ICAP	ND	<0.010	MGL			
MRL_CHK	Chromium, Total, ICAP	0.010	0.0097	MGL	97.0	( 50-150 )	
MS	Chromium, Total, ICAP	1.00	0.949	MGL	94.9	( 70-130 )	
MSD	Chromium, Total, ICAP	1.00	0.953	MGL	95.3	( 70-130 )	

**QC Ref #351232                      Chromium, Total, ICAP**

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01310048	MGL		( 0-0 )	
LCS1	Chromium, Total, ICAP	1.00	0.956	MGL	95.6	( 85-115 )	
LCS2	Chromium, Total, ICAP	1.00	0.900	MGL	90.0	( 85-115 )	
MBLK	Chromium, Total, ICAP	ND	<0.010	MGL			
MRL_CHK	Chromium, Total, ICAP	0.010	0.0097	MGL	97.0	( 50-150 )	
MS	Chromium, Total, ICAP	1.00	0.942	MGL	94.2	( 70-130 )	
MSD	Chromium, Total, ICAP	1.00	0.996	MGL	99.6	( 70-130 )	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates  
are advisory only, unless otherwise specified in the method.



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#194600

Tronox LLC - Henderson  
(continued)

## QC Ref #351498 Total Dissolved Solid (TDS)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300028	MGL		( 0-0 )	
DUP	Total Dissolved Solid (TDS)	ND	ND	MGL		( 0-10 )	
LCS1	Total Dissolved Solid (TDS)	175	158	MGL	90.3	( 85-115 )	
LCS2	Total Dissolved Solid (TDS)	700	692	MGL	98.9	( 85-115 )	
MBLK	Total Dissolved Solid (TDS)	ND	<10	MGL			
MRL_CHK	Total Dissolved Solid (TDS)	10.0	12	MGL	120.0	( 50-150 )	
RPD_LCS	Total Dissolved Solid (TDS)	90.286	98.857	MGL	9.1	( 0-20 )	

## QC Ref #353634 Total Dissolved Solid (TDS)

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
AASPKSMP	Spiked sample	Lab # 27	01300127	MGL		( 0-0 )	
DUP	Total Dissolved Solid (TDS)	7310	7370	MGL		( 0-10 )	0.8
LCS1	Total Dissolved Solid (TDS)	175	162	MGL	92.6	( 85-115 )	
LCS2	Total Dissolved Solid (TDS)	700	668	MGL	95.4	( 85-115 )	
MBLK	Total Dissolved Solid (TDS)	ND	<10	MGL			
MRL_CHK	Total Dissolved Solid (TDS)	10.0	8	MGL	80.0	( 50-150 )	
RPD_LCS	Total Dissolved Solid (TDS)	92.571	95.429	MGL	3.0	( 0-20 )	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates  
are advisory only, unless otherwise specified in the method.