



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel. 626 386 1100  
Fax. 626 386 1101  
1 800 566 LABS (1 800 566 5227)


## Laboratory Report

for

Tronox LLC - Henderson  
PO Box 55

Henderson , NV 89009

Attention: Susan Crowley  
Fax: (405) 302-4607



ADE Andy Eaton  
Project Manager



Report#: 178141  
CLO4

Laboratory certifies that the test results meet all **NELAC** requirements unless noted in the Comments section or the Case Narrative. Following the cover page are Comments, QC Report, QC Summary, Data Report, Hits Report, totaling 5 page[s].





MWH Laboratories, a Division of MWH Americas, Inc.  
750 Royal Oaks Avenue Suite 100  
Monrovia CA 91016 (626) 386-1100 FAX (626) 386-1124

**Bottle Order for Kerr McGee Chemical Company - Henderson Standing**

Andrew Eaton ..... Your MWL Project Manager  
(626).386-1125 ..... Direct Phone/Voice Mail

Client Code KERRMCGEE-MP W Weekly  
Project Code CLO4 Week 1  
PO# / Job#  
Blanket PO

SO# 30716 18882 RS

**Sampler: Please Return this Paper with your samples**

0

Ship Sample Kits to  
Kerr McGee-Veolia Water  
Gate 1  
8000 West Lake Mead Drive  
Henderson, NV 89015

Billing Address

Send Report to  
Kerr McGee Henderson Plant  
PO Box 55  
Henderson, NV 89009

KMCLLC  
attn: Contract 304280  
PO Box 3049  
Livonia, MI 48150

Created by

Order Date

06/19/06

Date Needed by Client

Date Samples to Arrive at MWL

SHIP LOCATION

ATTN: Susan Crowley  
PHONE: 702-651-2234

ATTN: Susan Crowley  
PHONE: 702-651-2234  
FAX: 702-651-2310

Quote#

# of Samples Tests

Bottles-Qty for each sample, type & preservative if any

UN#

Important Comments

# of Samples	Tests	Bottles-Qty for each sample, type & preservative if any	UN#	Important Comments
2	CLO4	1 125ml poly /no preservative AMBER IF WITH CLO3!		Weekly Composites at INFLUENT-COMP EFFLUENT-COMP
-	-	-	-	COC should be labeled with IDs as above.
-	-	-	-	These are in lieu of the former samples that were arriving once a month as a series of weekly composites.
-	-	-	-	adjusted to 2 samples 5/05



THIS MEMORANDUM is an acknowledgement that a Bill of Lading has been issued and is not the Original Bill of Lading, nor a copy or duplicate, covering the property named herein, and is intended solely for filing or record.

SHIPPER'S NUMBER: 142054

RECEIVED subject to the classifications and lawfully filed tariffs in effect on the date of the issue of this Bill of Lading.

**From: TRONOX LLC**

the property described below, in apparent good order, except as noted (contents and condition of contents of packages unknown), marked, consigned, and destined as indicated below, which said Carrier (the word Carrier being understood throughout this contract as meaning any person or corporation in possession of the property under the contract) agrees to carry to its usual place of delivery at said destination, if on its route, otherwise to deliver to another Carrier on the route to said destination. It is mutually agreed, as to each Carrier of all or any said property over all or any portion of said route to destination, and as to each party at any time interested in all or any of said property, that every service to be performed hereunder shall be subject to all the terms and conditions of the Uniform Domestic Straight Bill of Lading set forth (1) in Uniform Freight Classification in effect on the date hereof, if this is a rail or a rail-water shipment, or (2) in the applicable motor carrier classification or tariff if this is a motor carrier shipment.

Shipper hereby certifies that he is familiar with all the terms and conditions of the said Bill of Lading, including those on the back thereof, set forth in the classification or tariff which governs the transportation of this shipment, and the said terms and conditions are hereby agreed to by the Shipper and accepted for himself and his assigns.

SCANNED

CARRIER <b>Federal Express</b>		Date <b>07/05/06</b>	FROM NO. STATION: STATE <b>Henderson, NV 89015</b>
MWH LABORATORIES 750 Royal Oaks Avenue, Suite #100 Monrovia, CA 91016-3629 Phone: 626-568-6400		Authorization <b>S.CROWLEY</b>	
		FULL NAME OF SHIPPER TRONOX LLC  <b>CODE NO. WCN IF 1321.10400</b> If it moves between two ports by water, the law requires that the Bill of Lading shall state whether it is Carrier's or Shippers weight.	
FREIGHT CHARGES <input checked="" type="checkbox"/> Prepaid <input type="checkbox"/> Collect		CUSTOMER PO OR REQ'N NO.	
N/AR		SHIPPED FROM <b>Henderson, NV</b>	
LINE NO.	DESCRIPTION AND CLASSIFICATION	STOCK NO.	TOTAL QUANTITY
	<b>Ice Chest with water samples 2 bottles -- Influent, Effluent 1 Ice Chest @ 10 lbs  Not Regulated</b>		<b>1 COOLER</b>
			Subject to Section 7 of Conditions of applicable Bill of Lading, if this shipment is to be delivered to the Consignee without recourse on the Consignor, the Consignor shall sign the following statement: The Carrier shall not make delivery of this shipment without payment of freight and all other lawful charges. <b>TRONOX LLC</b>
			The description and weight indicated on this Bill of Lading are correct. Subject to verification by the Governing Weighing and Inspection Bureau according to Agreement.
TRUCK SHIPMENTS			
PLACARDS OFFERED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		PLACARDS ACCEPTED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO	
NUMBER OF PACKAGES	GROSS WEIGHT	TARE WEIGHT	NET WEIGHT
		0	
1	TOTAL GROSS WEIGHT 10	TOTAL TARE WEIGHT 0	10
NOTE: Where the rate is dependent on value, shippers are required to state specifically in writing the agreed or declared value of the property. The agreed or declared value of the property is hereby specifically stated by the shipper to be not exceeding _____ per			FOR CHEMICAL EMERGENCY-SPILL, LEAK, FIRE, EXPOSURE OR ACCIDENT CALL CHEMTREC - DAY OR NIGHT 800-424-9300 483-7616 IN DISTRICT OF COLUMBIA 202-483-7616 FROM OUTSIDE THE CONTINENTAL US.  "Shippers imprint in lieu of stamp; not a part of Bill of Lading approved by the Interstate Commerce Commission"

THIS IS TO CERTIFY THAT THE ABOVE-NAMED MATERIALS ARE PROPERLY CLASSIFIED, DESCRIBED, PACKAGED, MARKED AND LABELED AND ARE IN PROPER CONDITION FOR TRANSPORTATION ACCORDING TO THE APPLICABLE REGULATIONS OF THE DEPARTMENT OF TRANSPORTATION.

TRONOX LLC  
Shipper permanent post office address of shipper, PO Box 268857, Oklahoma City, OK 73126-8857

PER

**Judi Durkin**

AGENT

PER

From: Origin ID: (702)651-2220  
KERR-MCGEE CHEMICAL  
KERR-MCGEE CHEMICAL  
8000 LAKE MEAD DRIVE  
SHIPPING DEPARTMENT  
HENDERSON, NV 89015




Ship Date: 05JUL06  
ActWgt: 10 LB  
System#: 2274147/INET2500  
Account#: S \*\*\*\*\*

REF: MSO #142054 - 10 LBS



Delivery Address Bar Code

SHIP TO: (626)568-6400 **BILL SENDER**  
**ATTN: MONTGOMERY LABS RECEIVING**  
**MONTGOMERY WATSON LABS**  
**750 ROYAL OAKS AVENUE**  
**#100**  
**MONROVIA, CA 91016**



**PRIORITY OVERNIGHT**

**THU**

TRK# **7919 8756 4450**


FORM 0201

Deliver By:  
06JUL06

**BUR A2**

**91016 -CA-US**

**QZ WHPA**



SCANNED

Shipping Label: Your shipment is complete

1. Use the 'Print' feature from your browser to send this page to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

**Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.**

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$500, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

MWH Laboratories  
750 Royal Oaks Drive, Monrovia, CA 91016  
PHONE: 626-386-1100/FAX: 626-386-1101

ACKNOWLEDGMENT OF SAMPLES RECEIVED

Tronox LLC - Henderson  
PO Box 55  
Henderson, NV 89009  
Attn: Susan Crowley  
Phone: 702-651-2234

Customer Code: KERRMCGEE-MP  
PO#: Susan Crowley PO  
Group#: 178141  
Project#: CLO4  
Proj Mgr: Andrew Eaton  
Phone: (626) 386-1125

The following samples were received from you on 07/06/06. They have been scheduled for the tests listed beside each sample. If this information is incorrect, please contact your service representative. Thank you for using MWH Laboratories.

Sample#	Sample Id	Tests Scheduled	Matrix	Sample Date
2607060360	EFFLUENT-COMP	CLO4	Water	01-jul-2006 10:00:00
2607060361	INFLUENT-COMP	CLO4	Water	01-jul-2006 10:00:00

Test Acronym Description

Test Acronym	Description
CLO4	Perchlorate



**MWH Laboratories**  
*A Division of MWH Americas, Inc.*

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

---

Report  
Comments  
#178141

**Client Specific Comments**

I hereby certify that all laboratory analytical data was generated by a laboratory certified by the NDEP for each constituent and media presented herein.

Client Signature: \_\_\_\_\_



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Hits Report  
#178141

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
06-jul-2006 16:02:06

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	2607060360	EFFLUENT-COMP				
	2607060361	INFLUENT-COMP				
07/19/06	Perchlorate		210000		ug/l	20000

SUMMARY OF POSITIVE DATA ONLY.





# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
Data Report  
#178141

Tronox LLC - Henderson  
Susan Crowley  
PO Box 55  
Henderson , NV 89009

Samples Received  
07/06/06

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
<b>EFFLUENT-COMP (2607060360)</b>				<b>Sampled on 07/01/06 10:00</b>				
	07/16/06 00:00	326085	( EPA 314	) Perchlorate	ND	ug/l	10	5
<b>INFLUENT-COMP (2607060361)</b>				<b>Sampled on 07/01/06 10:00</b>				
	07/19/06 00:00	326398	( EPA 314	) Perchlorate	210000	ug/l	20000	5000



**MWH Laboratories**  
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Summary  
#178141

Tronox LLC - Henderson

---

QC Ref #326085 - Perchlorate

2607060360

EFFLUENT-COMP

Analysis Date: 07/16/2006

Analyzed by: raja

QC Ref #326398 - Perchlorate

2607060361

INFLUENT-COMP

Analysis Date: 07/19/2006

Analyzed by: raja



# MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: 626 386 1100  
Fax: 626 386 1101  
1 800 566 LABS (1 800 566 5227)

Laboratory  
QC Report  
#178141

Tronox LLC - Henderson

## QC Ref #326085

## Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 26	07120574	UGL		( 0-0 )	
LCS1	Perchlorate	25.0	21.7	UGL	86.8	( 85-115 )	
LCS2	Perchlorate	25.0	26.1	UGL	104.4	( 85-115 )	
LCS3	Perchlorate	4	3.69	UGL	92.2	( 75-125 )	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	27.4	UGL	109.6	( 70-130 )	
MSD	Perchlorate	25.0	26.8	UGL	107.2	( 70-130 )	
RPD_LCS	Perchlorate	86.800	104.400	UGL	18.4	( 0-20 )	
RPD_MS	Perchlorate	109.600	107.200	UGL	2.2	( 0-20 )	

## QC Ref #326398

## Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 26	07050507	UGL		( 0-0 )	
LCS1	Perchlorate	25.0	24.7	UGL	98.8	( 85-115 )	
LCS2	Perchlorate	25.0	25.3	UGL	101.2	( 85-115 )	
LCS3	Perchlorate	4	4.05	UGL	101.2	( 75-125 )	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	24.9	UGL	99.6	( 70-130 )	
MSD	Perchlorate	25.0	24.7	UGL	98.8	( 70-130 )	
RPD_LCS	Perchlorate	98.800	101.200	UGL	2.4	( 0-20 )	
RPD_MS	Perchlorate	99.600	98.800	UGL	0.8	( 0-20 )	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.  
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates  
are advisory only, unless otherwise specified in the method.