



MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)


Laboratory Report

for

Tronox LLC - Henderson
PO Box 55

Henderson , NV 89009

Attention: Susan Crowley
Fax: (405) 302-4607



for
ADE Andy Eaton
Project Manager



Report#: 182538
Project: CL04
PO#: Susan Crowle

Laboratory certifies that the test results meet all **NELAC** requirements unless noted in the Comments section or the Case Narrative. Following the cover page are Comments, QC Report, QC Summary, Data Report, Hits Report, totaling 5 page[s].

182538



MONTGOMERY WATSON LABORATORIES

CHAIN OF CUSTODY RECORD

MWLARS USE ONLY:

750 Royal Oaks Drive, Suite 100, Monrovia, CA > 91016-3629
(626) 568-6400 (800) 566-5227

LOGIN COMMENTS:

SAMPLES CHECKED/LOGGED IN BY: [Signature]

SAMPLE TEMP, RECEIPT AT LAB: [Signature]

BLUE ICE: FROZEN PARTIALLY FROZEN THAWED

TO BE COMPLETED BY SAMPLER:

COMPANY / PROJECT NAME

KERRMCGEE-MP

PROJECT JOB # / P.O.#

Weekly Permit ClO₂ Compliance Samples

Wendy Prescott

[Signature: Wendy Prescott]
Susan Crowley
(702) 651-2234

Tronox LLC Henderson Plant
PO Box 55
Henderson, NV 89009

REFER TO ATTACHED BOTTLE ORDER FOR ANALYSES

ANALYSES REQUIRED (mark an 'X' in all tests required for each sample line)

TIME	DATE	LOCATION	IDENTIFIER, STATE ID#	MATRIX *	GRAB	COMP	ANALYSES REQUIRED (mark an 'X' in all tests required for each sample line)										SAMPLER COMMENTS										
							1	2	3	4	5	6	7	8	9	10		11	12								
10:00am	8-20-06 to 8-26-06		Effluent-Comp	RSW		X																				SBO- 18882	
10:00am	8-20-06 to 8-26-06		Influent-Comp	RSW		X																					

* MATRIX TYPES:

Reported by Volume:

CFW = Chlor(amin)ated Finished Water
FW = Other Finished Water

Reported by Weight:

SO = Soil
SL = Sludge

RGW = Raw Ground Water
RSW = Raw Surface Water

CWW = Chlorinated Waste Water
WW = Other Waste Water
SW = Storm Water

SIGNATURE

[Signature: Wendy Prescott]

PRINT NAME

Wendy Prescott

COMPANY/TITLE

Velox Water NA for Tronox LLC - Henderson

DATE

8/28/2006

TIME

12:00

RECEIVED BY:

C-O-C#

PAGE OF



MWH Laboratories, a Division of MWH Americas, Inc.
 750 Royal Oaks Avenue Suite 100
 Monrovia CA 91016 (626) 386-1100 FAX (626) 386-1124

Bottle Order for Ironox, LLC- Henderson Standing

Andrew Eaton..... Your MWL Project Manager
 (926).386.1125..... Direct Phone/Voice Mail

Client Code KERRMCGEE-MP W Weekly Period
 Project Code CLO4 Week 1
 PO# / Job#
 Blanket PO

SO# 31508 18882 **RS** **Sampler: Please Return this Paper with your samples**

Created by 0 Ship Sample Kits to
 Order Date 08/07/06 Veolia Water-Tronox LLC
 Gate 1
 8000 West Lake Mead Drive
 Henderson, NV 89015

Send Report to
 Ironox, LLC Henderson Plant
 PO Box 55
 Henderson, NV 89009

Billing Address
 Ironox, LLC
 Attn: Accounts Payable
 PO BOX 268859
 Oklahoma City, OK 73126-8859

ATTN: Susan Crowley
 PHONE: 702-651-2234

ATTN: Susan Crowley
 PHONE: 702-651-2234
 FAX: 702-651-2310

Quote#

SHIP LOCATION

of Samples Tests Bottles-Qty for each sample, type & preservative if any UN# Important Comments

# of Samples	Tests	Bottles-Qty for each sample, type & preservative if any	UN#	Important Comments
2	CLO4	1 125ml poly /no preservative AMBER IF WITH CLO3I		Weekly Composites at
-	-	-	-	INFLUENT-COMP
-	-	-	-	EFFLUENT-COMP
-	-	-	-	COC should be labeled with IDs as above.
-	-	-	-	These are in lieu of the former samples that were arriving once a month as a series of weekly composites.
-	-	-	-	adjusted to 2 samples 5/05

From: Origin ID: (702)651-2230
TRONOX LLC
TRONOX LLC
8000 LAKE MEAD PARKWAY
SHIPPING DEPARTMENT
HENDERSON, NV 89015



CLS65266/17/22


Ship Date: 28AUG06
ActWgt: 10 LB
System#: 2274147/INET2500
Account#: S *****

REF: MSO #142157 - 10 LBS



Delivery Address Bar Code

SHIP TO: (626)568-6400 **BILL SENDER**
ATTN: MONTGOMERY LABS RECEIVING
MONTGOMERY WATSON LABS
750 ROYAL OAKS AVENUE
#100
MONROVIA, CA 91016



PRIORITY OVERNIGHT

TUE

TRK# 7921 9205 5272


FORM 0201

Deliver By:
29AUG06

BUR A2

91016 -CA-US

QZ WHPA



SCANNED

Shipping Label: Your shipment is complete

1. Use the 'Print' feature from your browser to send this page to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$500, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

THIS MEMORANDUM is an acknowledgement that a Bill of Lading has been issued and is not the Original Bill of Lading, nor a copy or duplicate, covering the property named herein, and is intended solely for filing or record.

SHIPPER'S NUMBER: 142157

RECEIVED subject to the classifications and lawfully filed tariffs in effect on the date of the issue of this Bill of Lading.

From: TRONOX LLC

the property described below, in apparent good order, except as noted (contents and condition of contents of packages unknown), marked, consigned, and destined as indicated below, which said Carrier (the word Carrier being understood throughout this contract as meaning any person or corporation in possession of the property under the contract) agrees to carry to its usual place of delivery at said destination, if on its route, otherwise to deliver to another Carrier on the route to said destination. It is mutually agreed, as to each Carrier of all or any said property over all or any portion of said route to destination, and as to each party at any time interested in all or any of said property, that every service to be performed hereunder shall be subject to all the terms and conditions of the Uniform Domestic Straight Bill of Lading set forth (1) in Uniform Freight Classification in effect on the date hereof, if this is a rail or a rail-water shipment, or (2) in the applicable motor carrier classification or tariff if this is a motor carrier shipment.

Shipper hereby certifies that he is familiar with all the terms and conditions of the said Bill of Lading, including those on the back thereof, set forth in the classification or tariff which governs the transportation of this shipment, and the said terms and conditions are hereby agreed to by the Shipper and accepted for himself and his assigns.

CARRIER Federal Express		Date 08/28/06	FROM NO. STATION: STATE Henderson, NV 89015		
MWH LABORATORIES 750 Royal Oaks Avenue, Suite #100 Monrovia, CA 91016-3629 Phone: 626-568-6400			Authorization S. CROWLEY		
FREIGHT CHARGES <input checked="" type="checkbox"/> Prepaid <input type="checkbox"/> Collect			FULL NAME OF SHIPPER TRONOX LLC		
N/AR			CUSTOMER PO OR REQ'N NO.		
SHIPPED FROM Henderson, NV			CODE NO. WCN IF 1321.10400 If it moves between two ports by water, the law requires that the Bill of Lading shall state whether it is Carrier's or Shippers weight.		
LINE NO.	DESCRIPTION AND CLASSIFICATION	STOCK NO.	TOTAL QUANTITY		
	Ice Chest with water samples 2 bottles -- Influent, Effluent 1 Ice Chest @ 10 lbs Not Regulated		1 COOLER		
TRUCK SHIPMENTS PLACARDS OFFERED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO PLACARDS ACCEPTED <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO			FOR CHEMICAL EMERGENCY-SPILL, LEAK, FIRE, EXPOSURE OR ACCIDENT CALL CHEMTREC - DAY OR NIGHT 800-424-9300 483-7616 IN DISTRICT OF COLUMBIA 202-483-7616 FROM OUTSIDE THE CONTINENTAL US.		
NUMBER OF PACKAGES	GROSS WEIGHT	TARE WEIGHT	NET WEIGHT		
		0			
1	TOTAL GROSS WEIGHT 10	TOTAL TARE WEIGHT 0	10		
NOTE: Where the rate is dependent on value, shippers are required to state specifically in writing the agreed or declared value of the property. The agreed or declared value of the property is hereby specifically stated by the shipper to be not exceeding _____ per _____			"Shippers imprint in lieu of stamp; not a part of Bill of Lading approved by the Interstate Commerce Commission"		
THIS IS TO CERTIFY THAT THE ABOVE-NAMED MATERIALS ARE PROPERLY CLASSIFIED, DESCRIBED, PACKAGED, MARKED AND LABELED AND ARE IN PROPER CONDITION FOR TRANSPORTATION ACCORDING TO THE APPLICABLE REGULATIONS OF THE DEPARTMENT OF TRANSPORTATION.					
TRONOX LLC Shipper permanent post office address of shipper, PO Box 268857, Oklahoma City, OK 73126-8857		PER Judi Durkin	AGENT	PER	

MWH Laboratories
 750 Royal Oaks Drive, Monrovia, CA 91016
 PHONE: 626-386-1100/FAX: 626-386-1101

ACKNOWLEDGMENT OF SAMPLES RECEIVED

Tronox LLC - Henderson
 PO Box 55
 Henderson, NV 89009
 Attn: Susan Crowley
 Phone: 702-651-2234

Customer Code: KERRMCGEE-MP
 PO#: Susan Crowley PO
 Group#: 182538
 Project#: CLO4
 Proj Mgr: Andrew Eaton
 Phone: (626) 386-1125

The following samples were received from you on **08/29/06**. They have been scheduled for the tests listed beside each sample. If this information is incorrect, please contact your service representative. Thank you for using MWH Laboratories.

Sample#	Sample Id	Tests Scheduled	Matrix	Sample Date
2608290031	EFFLUENT-COMP	CLO4	Water	26-aug-2006 10:00:00
2608290032	INFLUENT-COMP	CLO4	Water	26-aug-2006 10:00:00

Test Acronym Description

Test Acronym	Description
CLO4	Perchlorate



MWH Laboratories
A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Report
Comments
#182538

Client Specific Comments

I hereby certify that all laboratory analytical data was generated by a laboratory certified by the NDEP for each constituent and media presented herein.

Client Signature: _____



MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
Hits Report
#182538

Tronox LLC - Henderson
Susan Crowley
PO Box 55
Henderson , NV 89009

Samples Received
29-aug-2006 10:09:29

Analyzed	Sample#	Sample ID	Result	Federal MCL	UNITS	MRL
	2608290031	EFFLUENT-COMP				
	2608290032	INFLUENT-COMP				
09/07/06	Perchlorate		217000		ug/l	20000

SUMMARY OF POSITIVE DATA ONLY.



MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
Data Report
#182538

Tronox LLC - Henderson
Susan Crowley
PO Box 55
Henderson , NV 89009

Samples Received
08/29/06

Prepared	Analyzed	QC Ref#	Method	Analyte	Result	Units	MRL	Dilution
EFFLUENT-COMP (2608290031)				Sampled on 08/26/06 10:00				
	09/07/06 00:00	332658	(EPA 314) Perchlorate	ND	ug/l	10	5
INFLUENT-COMP (2608290032)				Sampled on 08/26/06 10:00				
	09/07/06 00:00	332658	(EPA 314) Perchlorate	217000	ug/l	20000	5000



MWH Laboratories

A Division of MWH Americas, Inc.

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3829
Tel: 626 386 1100
Fax: 626 386 1101
1 800 566 LABS (1 800 566 5227)

Laboratory
QC Summary
#182538

Tronox LLC - Henderson

QC Ref #332658 - Perchlorate

Analysis Date: 09/07/2006

2608290031
2608290032

EFFLUENT-COMP
INFLUENT-COMP

Analyzed by: raja
Analyzed by: raja

Tronox LLC - Henderson

QC Ref #332658

Perchlorate

QC	Analyte	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPD (%)
MS	Spiked sample	Lab # 26	08240219	UGL		(0-0)	
LCS1	Perchlorate	25.0	25.8	UGL	103.2	(85-115)	
LCS2	Perchlorate	25.0	23.0	UGL	92.0	(85-115)	
LCS3	Perchlorate	4	4.68	UGL	117.0	(75-125)	
MBLK	Perchlorate	ND	<4.0	UGL			
MS	Perchlorate	25.0	24.3	UGL	97.2	(70-130)	
MSD	Perchlorate	25.0	21.6	UGL	86.4	(70-130)	
RPD_LCS	Perchlorate	103.200	92.000	UGL	11.5	(0-20)	
RPD_MS	Perchlorate	97.200	86.400	UGL	11.8	(0-20)	

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
Criteria for MS and DUP are advisory only, batch control is based on LCS. Criteria for duplicates
are advisory only, unless otherwise specified in the method.