



December 03, 2009

www.gel.com

Mr. Frank Hagar
Northgate Environmental Management, Inc.
1100 Quail St., Suite 102
Newport Beach, California 92660

Re: Tronox Henderson
Work Order: 239753

Dear Mr. Hagar:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on November 03, 2009, October 26, 2009, October 27, 2009, October 28, 2009, October 29, 2009, October 30, 2009 and October 31, 2009. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4453.

Sincerely,

for Edith Kent
Project Manager

Chain of Custody: 2027.001.01120, 2027.001.01121, 2027.001.01123, 2027.001.01124, 2027.001.01125,
2027.001.01126, 2027.001.01127, 2027.001.01128, 2027.001.01129, 2027.001.01131 and 2027.001.01134
Enclosures

Tronox LLC
Tronox Henderson
SDG:239753

Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation	4
Laboratory Certifications	28
Radiological Analysis.....	30
Sample Data Summary	43
Quality Control Data	63
Raw Data	67
Thorium	68
923093.....	69
Uranium.....	93
923094.....	94
Radium 228.....	118
922859.....	119
Radium 226.....	126
920697.....	127
Method Calibration Data	132
Alpha Spectroscopy	133
Gas Flow Proportional Counters	189
Lucas Cell Counters	273
Lucas 1	274
Lucas 2	295
Lucas 3	315
Lucas 4	335
Lucas 5	350
Lucas 7	371
Background and Efficiency Data	413
Runlogs	596

Case Narrative

**CASE NARRATIVE
for
Tronox LLC
Tronox Henderson
SDG:239753**

December 03, 2009

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The samples arrived at GEL Laboratories LLC, Charleston, South Carolina on November 03, 2009, October 26, 2009, October 27, 2009, October 28, 2009, October 29, 2009, October 30, 2009 and October 31, 2009 for analysis. Shipping container temperatures were checked, documented, and within specifications. The samples were delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There were no times of collection listed on the chain of custody for 2027.001.01121 and 2027.001.01128. The client was notified that the lab used the times listed on the sample containers. The client also verified that the samples on 2027.001.01125 were field filtered. Please refer to attached e-mails.

Items of Note

All samples under this SDG were logged as an open SDG until a sufficient amount of samples were received by the lab. The client was notified that the SDG was closed on November 9, 2009 and the turnaround time would start from then. Please see the attached e-mails for further details on all issues.

QC Issues

The following samples did not meet the Tronox QA program sample result uncertainty limit of <30% for Ra-226 with the results between 2 and 5 times the MDA and were counted for the maximum time: 239753006, 239753014 and 239753016. The following samples did not meet the Tronox QA program sample result uncertainty limit of <30% for Alpha Spec Uranium with the results between 2 and 5 times the MDA and were counted for the maximum time: and 239753018. The following samples did not meet the Tronox QA program sample result uncertainty limit of <30% for Alpha Spec Thorium with the results between 2 and 5 times the MDA and were counted for the maximum time: 239753006, 239753012 and 239753016. The following samples did not meet the Tronox QA program sample tracer yield requirements of 70-120% for Alpha Spec Uranium due to matrix issues: 239753001, 239753002, 239753009, 239753011 and 239753017. The following samples did not meet the Tronox QA program sample tracer yield requirements of 70-120% for Alpha Spec Thorium due to matrix issues: and 239753002. The following samples did not meet the Tronox QA program required detection limits for Alpha Spec Thorium due to limited sample volume and were counted for the maximum time: 239753001, 239753002, 239753005, 239753006, 239753011, and the method blank. The following samples did not meet the Tronox QA program required detection limits for Alpha Spec Uranium due to limited sample volume and were counted for the maximum time: 239753003.

Sample Identification

The laboratory received the following samples:

<u>Laboratory ID</u>	<u>Client ID</u>
239753001	M-141B
239753002	M-141009B
239753003	PB102309-A3

239753004	M-145B
239753005	M-139B
239753006	M-146B
239753007	M-144B
239753008	M-138B
239753009	M-138009B
239753010	M-138BDISS
239753011	M-138009BDISS
239753012	M-137B
239753013	M-137BDISS
239753014	M-148B
239753015	EB103009-GWA4
239753016	M-147B
239753017	M-147009B
239753018	EB110209-GWA3

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: Case Narrative, Chain of Custody, Cooler Receipt Checklist, Data Package Qualifier Definitions and data from the following fractions: Radiochemistry.

This data package, to the best of my knowledge, is in compliance with technical and administrative requirements.

Deatter Shaffer

for Edith Kent

Project Manager

Chain of Custody and Supporting Documentation



1100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9293

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027.001.01125

Page 1 of 1

Cooler # _____

Collection Area: III

Required Ship to Lab:		Required Project Information:		Required Invoice Information:		TAT: Standard 30 day		X	Rush	Mark One
Lab Name:	GEL Laboratories, LLC	Site ID#:	TRONOX LLC, HENDERSON	Send Invoice to:	Susan Crowley Tronox LLC	If Rush, Date due				
Address:	2040 Savage Road	Project #:	2027.001	Address:	PO Box 66	QC level Required:	Standard	Special	EPA Stage	Mark one
Charleston, SC 29407		Site Address:	560 W. Lake Mead Drive	City/State:	Henderson, NV 89008	Phone #:	(849)280-9293	NJ Reduced Deliverable Package?		
Lab P#: Edith M. Kent	City: Henderson	State: NV	Reimbursement project?	<input checked="" type="checkbox"/>	Non-reimbursable project?	<input checked="" type="checkbox"/>	Mark one	MA MCP Cert?	CT RCP Cert?	Mark One
Phone/Fax: (843)666-8171	Site PM Name: Derrick Willis	Phone/Fax: 849-375-7004	Send EDD to:	Frank Hagar Northgate Environmental Management, Inc	CC Hardcopy report to:	PDF Electronic Version Only	Lab Project ID (lab use)			
Lab P#: emk@gel.com	Site PM Email: derrick.willis@ngem.com		CC Hardcopy report to:	see additional comments below						
Applicable Lab Quote #: _____										
Requested Analyses										
EFPL903,17804,0, EML HASL 300.										
#	SAMPLE ID Character per box. (A-Z, 0-9, -, Samples IDs MUST BE UNIQUE	One	Valid Matrix Codes	MATRIX WATER GROUNDWATER WHITE WATER WHITE CC SLUDGE OIL FREE PRODUCT SOIL WATER WATER WATER ANIMAL TISSUE AIR SVE AIR BED LOAD	WATER WW WL SL OIL FP SO WT WT AT AE SVE BLD	SAMPLE CODE W WW WL SL OIL FP SO WT WT AT AE SVE BLD	SAMPLE DATE	SAMPLE TIME	# OF CONTAINERS	Preservatives
1	M-141B							1	N	X
2	M-141B							1	N	X
3	M-141009B							1	Y	X
4	M-141009B							1	Y	X
5	PB102309-A3							1	Y	X
6	PB102309-A3							1	Y	X
7										
8										
9										
10										
11										
12										
Relinquished by Affiliation:										
						DATE:	TIME:	ACQUITTED BY / AFFILIATION		
						10/23/2009	14:25	GES 10/23/2009		
								14:25		
Additional Comments/Special Instructions:										
FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (isotopic) and Uranium (Isotopic)										
All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: clindy.arnold@ngem.com frank.hagar@ngem.com										
Temp in DC		Samples on Ice?		In tact?		Y/N		Sample Receipt Conditions		
TDP Blank		US MAIL		FEDEX		Y/N		Initial		
UPS COURIER		PRINT Name of SAMPLER:		SIGNATURE of SAMPLER:		Signature of SAMPLER:		Signature of SAMPLER:		

Shipping Method (mark as appropriate): *FEDEX* Sampler Name and Signature: *Josh W. Ols*

Temp in DC: *Y/N* Samples on Ice: *Y/N* In tact?: *Y/N*

US MAIL: *Y/N* FEDEX: *Y/N* Print Name of Sampler: *Josh W. Ols*

Date Shipped: *10/23* Time: *14:25*

SAMPLE RECEIPT & REVIEW FORM

Client:	TRONOX			SDG/ARCO/C Work Order:	239753-1.
Received By:	C. Duffy			Date Received:	10/26/09
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.		
COC/Samples marked as radioactive?	X		Maximum Counts Observed*: 60		
Classified Radioactive II or III by RSO?	X				
COC/Samples marked containing PCBs?	X				
Shipped as a DOT Hazardous?	X		Hazard Class Shipped:	UN#:	
Samples identified as Foreign Soil?	X				

Sample Receipt Criteria			Yes	No	Comments/Qualifiers (Required for Non-Conforming Items)
1 Shipping containers received intact and sealed?	X				Circle Applicable: seals broken damaged container leaking container other (describe)
2 Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		X			Preservation Method: ice bags blue ice dry ice none other (describe) 220
3 Chain of custody documents included with shipment?	X				
4 Sample containers intact and sealed?	X				Circle Applicable: seals broken damaged container leaking container other (describe)
5 Samples requiring chemical preservation at proper pH?	X				Sample ID's, containers affected and observed pH: If Preservation added, Lot#:
6 VOA vials free of headspace (defined as < 6mm bubble)?		X			Sample ID's and containers affected:
7 Are Encore containers present?		X			(If yes, immediately deliver to Volatiles laboratory)
8 Samples received within holding time?	X				Id's and tests affected:
9 Sample ID's on COC match ID's on bottles?	X				Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	X				Sample ID's affected:
11 Number of containers received match number indicated on COC?	X				Sample ID's affected:
12 COC form is properly signed in relinquished/received sections?	X				

Comments:

7960 0137 0171

2397537

northgate
environmental management, inc.

1100 Quail Street, Suite 102, Newport Beach, CA 92660
(844) 260-9283

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

Collection Area: III

COC No. 2027.001.01124
Page: 1 of 1
Cooler # _____

2009/05/03

Required Ship to Lab:		Required Project Information:		Required Invoice Information:		TAT: Standard 30 day		X	Rush		
Lab Name: GEL Laboratories, LLC	Site ID #: TRONOX LLC, HENDERSON	Send invoice to: Susan Crowley		Address: PO Box 55	Tronox LLC	If Rush: Date due				Mark One	
Address: 2040 Savage Road	Project # 2027.001					QC level Required: Standard		Special	EPA Stage 4	Mark one	
Charleston, SC 29407	Site Address 560 W. Lake Mead Drive	City/State Henderson	State NV	Phone #: (949)260-9283		Reduced Deliverable Package?				Mark One	
Lab P.M.: Edith M. Kent						MA MCP Cert?		C/T RCP Cert?		Mark One	
Phone/Fax: (843)666-8171	Site PM Name Derrick Willis	Reimbursement project? X	Non-reimbursement project? X	Frank Hagar Northgate Environmental Management, Inc						Mark One	
Lab P.M. email: emk@gel.com	Phone/Fax: 949-376-7004	Send EDD to CC Hardcopy report to CC Hardcopy report to PDF Electronic Version Only				Lab Project ID (lab use)				Mark One	
Applicable Lab Quote #:	Site PM Email: derrick.williams@ngem.com	see additional comments below									
Requester's Name: EPA 903/180403 EML HASL 300*											
Comments/Lab Sample I.D.											
#	SAMPLE ID	Character per box. (A-Z, 0-9 / , -) Samples IDs MUST BE UNIQUE	Valid Matrix Codes	MATRIX	MATRIX CODE	SAMPLE TYPE	SAMPLE DATE	SAMPLE TIME	# OF CONTAINERS	Preservatives	
										NH3O	H2SO4
1	M-145B	One	WATER	WATER	WATER	WATER	WATER	WATER	1	X	X
2	M-145B		WATER	WATER	WATER	WATER	WATER	WATER	1	X	X
3	M-145BDISS		FREE PRODUCT	FREE PRODUCT	FREE PRODUCT	FREE PRODUCT	FREE PRODUCT	FREE PRODUCT	1	X	X
4	M-145BDISS		COL.	COL.	COL.	COL.	COL.	COL.	1	X	X
5			DO	DO	DO	DO	DO	DO	1	X	X
6			WIPE	WIPE	WIPE	WIPE	WIPE	WIPE	1	X	X
7			AMBIENT AIR	AMBIENT AIR	AMBIENT AIR	AMBIENT AIR	AMBIENT AIR	AMBIENT AIR	1	X	X
8			BIG AIR	BIG AIR	BIG AIR	BIG AIR	BIG AIR	BIG AIR	1	X	X
9			SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	1	X	X
10			SLUDGE	SLUDGE	SLUDGE	SLUDGE	SLUDGE	SLUDGE	1	X	X
11			LIQUID	LIQUID	LIQUID	LIQUID	LIQUID	LIQUID	1	X	X
12			ANIMAL TISSUE	ANIMAL TISSUE	ANIMAL TISSUE	ANIMAL TISSUE	ANIMAL TISSUE	ANIMAL TISSUE	1	X	X
Requisition Number / Verification: DRAFT DATE: 10/26/2023 BY: GES											
Additional Comments/Special Instructions: FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectrometry) Thorium (isotopic) and Uranium (isotopic)											
All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: Cindy.arnold@ngem.com frank.hagar@ngem.com											
SHIPPING: Method (mark as appropriate) SAMPLER NAME AND SIGNATURE: UPS COURIER FEDEX PRINT NAME OF SAMPLER: <i>Daren Brown</i> <i>John W. OTIS</i> US MAIL SIGNATURE OF SAMPLER: <i>John W. OTIS</i>											
Temp in DO Temp Blank? Samples in tact? Temp Blank? Samples in tact?											

SAMPLE RECEIPT & REVIEW FORM

Client: <i>Kerr/Northgate</i>	SDG/ARCO/C Work Order: <i>2397531</i> .		
Received By: <i>MK</i>	Date Received: <i>10-27-09</i>		
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group of further investigation.
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>		Maximum Counts Observed*: <i>Cpm 20</i>
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>		
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>		
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>		Hazard Class Shipped: UN#:
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>		

Sample Receipt Criteria		Yes	NA	N	Comments/Qualifiers (Required for Non-Conforming Items)			
1	Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)			
2	Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		<input checked="" type="checkbox"/>		Preservation Method: <i>23</i> ice bags blue ice dry ice <i>none</i> other (describe)			
3	Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>						
4	Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)			
5	Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>			Sample ID's, containers affected and observed pH: If Preservation added, Lot#:			
6	VOA vials free of headspace (defined as < 6mm bubble)?		<input checked="" type="checkbox"/>		Sample ID's and containers affected:			
7	Are Encore containers present?			<input checked="" type="checkbox"/>	(If yes, immediately deliver to Volatiles laboratory)			
8	Samples received within holding time?	<input checked="" type="checkbox"/>			ID's and tests affected:			
9	Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:			
10	Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>			Sample ID's affected:			
11	Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:			
12	COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>						

Comments:

FX 7970 5312 9534



environmental management, inc.

100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9293

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a FEGAI DOCUMENT. All relevant fields must be completed and accurate.

100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9293

COC No. 2027.001.01129

Page: 1 of

Collection Area: II

Collection Area: II

G northgate

environmental management, inc.

1100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9293

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027.001.01120
Page: 1 of 1
Cooler #: _____
Collection Area: IV

Required Ship to Lab:											
Required Project Information:			Required Invoice Information:			Analyses					
Lab Name: GEL Laboratories, LLC	Site ID#:	TRONOX LLC, HENDERSON	Sold invoice to:	Susan Crowley	Address:	PO Box 55	TAT: Standard 30 day	X	Rush	Mark One	
Address: 2040 Savage Road	Project #:	2027.001	City/State:	Henderson, NV 89009	Phone #:	(843)260-9293	If Rush, Date due				
Charleston, SC 29407	Site Address:	660 W Lake Mead Drive	Reimbursement project?	<input checked="" type="checkbox"/>	QC level Required:	Standard	QC Reduced Deliverable Package?	<input type="checkbox"/>	Special	EPA Stage	Mark one
Lab PM: Edith M. Kent	City:	Henderson	State:	NV	Non-Reimbursement project?	<input type="checkbox"/>	Mark one			4	
Phone/Fax: (843)556-8171	Site P/M Name:	Derrick Willis	Send EDD to:	Frank Hagar Northgate Environmental Management, Inc	MA MCP Cert?	<input type="checkbox"/>	CT RCP Cert?	<input type="checkbox"/>			Mark One
Lab PM email: emk@gel.com	Phone/Fax:	949-375-7004	CC Hardcopy report to:	frank.hagar@ngem.com	Lab Project ID (Lab use)						
Applicable Lab Quote #:	Site P/M Email:	derrick.willis@ngem.com	CC Hardcopy report to:	see additional comments below							
Requested Analyses											
#	SAMPLE ID	One	Valid Matrix Codes	MATRIX	SAMPLE TYPE=COMP	SAMPLE DATE	SAMPLE TIME	#OF CONTAINERS	Preservatives		
									WATER	WS	2 L Poly Clear
1	M-144B	Character per box. (A-Z, 0-9, /, -) Samples IDs MUST BE UNIQUE	DRINKING WATER	WW	SURFACE WATER	W		1	N	X	X
2	M-144B		GROUND WATER	WG	WATER OC	WQ		1	N	X	2 L Poly Clear
3	M-144BDS		WASTE WATER	WW	INDUSTRIAL	WI		1	Y	X	2 L Poly Clear - JWL
4	M-144BDS		FREE PRODUCT	WP	INDUSTRIAL	WI		1	Y	X	2 L Poly Clear - JWL
5			OL	ON	INDUSTRIAL	WI		1	Y	X	
6			WIPE	AA	OTHER	WA		1	Y	X	
7			AMBIENT AIR	AS	ANIMAL TISSUE	AT		1	Y	X	
8			SOIL	OS	SOIL GAS	SG		1	Y	X	
9											
10											
11											
12											
Additional Comments/Special Instructions:											
FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (isotopic) and Uranium (isotopic)											
All PDF Reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: cindy.arnold@ngem.com frank.hagar@ngem.com											
Temp in DC	Temp Black3	Sample Static3	Samples on Ice?	Date:	Time:	ACTUAL DATE/FAIR DATE	Date:	Time:	Sample Receipt Conditions		
Y/N	Y/N	Y/N	Y/N	10/27/1410	10:00 AM	10/27/1410	Y/N	Y/N	Y/N		
UPS COURIER	FEDEX	US MAIL	SHIPPING METHOD (mark as appropriate)	Print Name of SAMPLER:	Signature of SAMPLER:	JWL	Y/N	Y/N	Y/N		

SAMPLE RECEIPT & REVIEW FORM

Client:	<u>Kerr/Westgate</u>			SDG/ARCO/C Work Order: <u>2397537</u> .
Received By:	<u>MIC</u>			Date Received: <u>10-23-09</u>
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group of further investigation.	
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>	Maximum Counts Observed*: <u>over 30</u>		
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>			
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>			
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>	Hazard Class Shipped: UN#:		
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>			

Sample Receipt Criteria		Yes	No	Comments/Qualifiers (Required for Non-Conforming Items)
1	Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>		Circle Applicable: seals broken damaged container leaking container other (describe)
2	Samples requiring cold preservation within 0 ≤ 6 deg. C?	<input checked="" type="checkbox"/>		Preservation Method: ice bags dry ice <u>none</u> other (describe) <u>20°</u>
3	Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>		
4	Sample containers intact and sealed?	<input checked="" type="checkbox"/>		Circle Applicable: seals broken damaged container leaking container other (describe)
5	Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>		Sample ID's, containers affected and observed pH: If Preservation added, Lot#:
6	VOA vials free of headspace (defined as < 6mm bubble)?	<input checked="" type="checkbox"/>		Sample ID's and containers affected:
7	Are Encore containers present?		<input checked="" type="checkbox"/>	If yes, immediately deliver to Volatiles laboratory
8	Samples received within holding time?	<input checked="" type="checkbox"/>		Id's and tests affected:
9	Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>		Sample ID's and containers affected:
10	Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>		Sample ID's affected:
11	Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>		Sample ID's affected:
12	COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>		

Comments:

Fx 7970 57d6 8507

Northgate

environmental management, inc.

1100 Quail Street, Suite 102, Newport Beach, CA 92660
(844) 260-9283

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027.001.01123

Page: 1 of 1
Cooler #: _____

Collection Area IV

Required Project Information:												
Lab Name:	GEL Laboratories, LLC	Site ID #:	TRONOX LLC. HENDERSON			Send invoice to:	Susan Crowley			TAT: Standard 30 day	<input checked="" type="checkbox"/> Rush	
Address:	2040 Savage Road	Project #:	2027.001			Address:	PO Box 56			If Rush, Date due	Mark One	
Charlottesville, SC 29407		Site Address:	560 W. Lake Mead Drive			City/State:	Henderson, NV 89009			Phone #:	(844)260-9293	
Lab P.M.:	Edith M. Kent	City:	Henderson	State:	NV	Reimbursement project?	<input checked="" type="checkbox"/>			NJ Reduced Deliverable Package?	QC level Required: Standard	
Phone/Fax:	(843)655-9171	Site P.M. Name:	Derrick Willis			Send EDD to:	Frank Hagar Northgate Environmental Management, Inc			MA MCIP Cert?	Special EPA Stage 4	
Lab PM email:	emk@gel.com	Phone/Fax:	949-375-7004			CC Hardcopy report to:	PDF Electronic Version Only			CT RCP Cert?	Mark One	
Applicable Lab Quote #:		Site P.M Email:	derrick.willis@ngem.com			CC Hardcopy report to:	see additional comments below			Lab Project ID (lab use)		
Additional Comments/Special Instructions:												
FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (Isotopic) and Uranium (Isotopic)												
All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: cindy.arnold@ngem.com frank.hagar@ngem.com												
Required Analytes												
#	SAMPLE ID	Character per box.	One			# OF CONTAINERS			Preservatives			
							FIELD FILTERED? (Y/N)					
							N2S2O3					
							NaOH					
							HCl					
							HNO3					
							HSO4					
							Uptreated					
							Other					
							Metathiol					
							Na2S2O3					
							EPA 803, TSP420, EMIL HASL 300,					
							Analyses					
							Requester					
							Comments/Lab Sample I.D.					
1	M-138B	MATRIX	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
2	M-138B	CHRONO WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
3	M-138009B	WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
4	M-138009B	WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
5	M-138BDI	WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
6	M-138BDI	WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
7	M-138009BDI	WATER	WP	WATER	WATER	WATER	WATER	WATER	WATER	WATER	2 L Poly Clear	
8	M-138009BDI	AIR	WP	AIR	AIR	AIR	AIR	AIR	AIR	AIR	2 L Poly Clear	
9												
10												
11												
12												

Temp in QC	Sample in QC										
Tip Blank	Sample in QC										
Y/N	Y/N										
UPS COURIER FEDEX	PRINT Name of SAMPLER:	UPS COURIER FEDEX	PRINT Name of SAMPLER:	UPS COURIER FEDEX	PRINT Name of SAMPLER:	UPS COURIER FEDEX	PRINT Name of SAMPLER:	UPS COURIER FEDEX	PRINT Name of SAMPLER:	UPS COURIER FEDEX	PRINT Name of SAMPLER:
US MAIL	SIGNATURE of SAMPLER:										

SHIPPING METHOD (mark as applicable)	DATE	TIME	Sample Receipt Conditions
<i>John W. Hagar</i>	10/28/1330	12:00	Y/N Y/N Y/N
<i>John W. Hagar</i>	10/28/1330	12:00	Y/N Y/N Y/N
<i>John W. Hagar</i>	10/28/1330	12:00	Y/N Y/N Y/N
<i>John W. Hagar</i>	10/28/1330	12:00	Y/N Y/N Y/N

SAMPLE RECEIPT & REVIEW FORM

Client:	<i>Kerr/Worleygate</i>			SDG/ARCO/C Work Order:	<i>239753-1.</i>
Received By:	<i>MK</i>			Date Received:	<i>10-29-09</i>
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.		
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>		Maximum Counts Observed*: <i>Open 20</i>		
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>				
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>				
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>		Hazard Class Shipped:	UN#:	
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>				
Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)	
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)	
2 Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		<input checked="" type="checkbox"/>		Preservation Method: <i>do c</i> ice bags blue ice dry ice <input checked="" type="checkbox"/> none other (describe)	
3 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>				
4 Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)	
5 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>			Sample ID's, containers affected and observed pH: If Preservation added, Lot#:	
6 VOA vials free of headspace (defined as < 6mm bubble)?		<input checked="" type="checkbox"/>		Sample ID's and containers affected:	
7 Are Encore containers present?		<input checked="" type="checkbox"/>		(If yes, immediately deliver to Volatiles laboratory)	
8 Samples received within holding time?	<input checked="" type="checkbox"/>			Id's and tests affected:	
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:	
10 Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>			Sample ID's affected:	
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:	
12 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>				
Comments: <i>* M 138BDISS m 138009BDISS > LABELS = FILTERED: N</i>					
<i>Fx 7960 7379 1612</i>					

PM (or PMA) review: Initials *(RS)* Date *10/29/09*

239753%



environmental management, inc.
1100 Quail Street, Suite 102, Newport Beach, CA 92660
(844) 260-9293

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027.001.01121
Page: 1 of 1
Cooler #: _____

Collection Area: IV

2009/10/5 5:02:32

Required Ship to Lab:		Required Project Information:			Required Invoice Information:			TAT: Standard 30 day			X		Rush		Mark One		
Lab Name:	GEL Laboratories, LLC	Site ID #:	TRONOX LLC, HENDERSON		Send invoice to:	Susan Crowley Tronox LLC		If Rush, Date due									
Address:	2040 Savage Road	Project #:	2027.001		Address:	PO Box 66		QC level Required:			Standard		Special		EPA Stage	Mark one	
City:	Henderson	Site Address:	660 W. Lake Mead Drive		City/State:	Henderson, NV 89008		Phone #:	(849)260-9293		4						
Lab PW:	Edith M. Kent	Site PM:	Name:	Derrick Willis	Reimbursement project?	X	Non-reimbursable project?	X	Mark one	NJ Reduced Deliverable Package?							
Phone/Fax:	(843)686-8171	Site PM Email:	derrick.willis@ngem.com		Send EDD to:	Frank Hagar, Northgate Environmental Management, Inc		MA MCP Cert?	CT RCP Cert?		CT RCP Cert?					Mark One	
Lab PW email:	ermk@gel.com	Phone/Fax:	949-371-7004		CC Hardcopy report to:	PDF Electronic Version Only		Lab Project ID (lab use)									
Applicable Lab Quote #:		Site PM Email:	derrick.willis@ngem.com		CC Hardcopy report to:	see additional comments below		Comments/Lab Sample I.D.									
SAMPLE ID		One		Valid Matrix Codes		MATRIX		SAMPLE DATE		SAMPLE TIME		# OF CONTAINERS		Preservatives			
Character per box.		(A-Z, 0-9, -,)		One		WATER: NO SURFACE WATER NO WATER WW WATER/CO. SL SLUDGE SL OR ORGANIC SOIL SOIL WATER WATER ANIMAL TISSUE AN AE AE SO		G=GRAB C=COMP		G=GRAB C=COMP		N2S2O3 NaOH HCl HNO3 H2SO4 Unpreserved		FIELD FILTERED? (Y/N)			
#		Samples IDs MUST BE UNIQUE															
1	M-137B					WG G				X		1 N		X		2 L Poly Clear	
2	M-137B					WG G				X		1 N		X		2 L Poly Clear	
3	M-137BDISS					WG G				X		1 Y		X		2 L Poly Clear	
4	M-137BDISS					WG G				X		1 Y		X		2 L Poly Clear	
5																	
6																	
7																	
8																	
9																	
10																	
11																	
12																	
Additional Comments/Special Instructions: Full Digestion Specification EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (Isotopic) and Uranium (Isotopic) All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: cindy.arnold@ngem.com frank.hagar@ngem.com																	
Shutting (all three) (mark as appropriate)		Sample ID		Print Name of Sampler:		Signature of Sampler:		Date Sampled:		Time Sampled:		Date Rec'd:		Time Rec'd:		Sample Receipt Conditions:	
Temp in DC		in tact?		Sample ID		Signature		Date		Time		Date		Time		Sample Receipt Conditions	
Temp in DC		in tact?		Sample ID		Signature		Date		Time		Date		Time		Sample Receipt Conditions	
Temp in DC		in tact?		Sample ID		Signature		Date		Time		Date		Time		Sample Receipt Conditions	

Collection Area: III

Required Ship to Lab:		Required Project Information:			Required Invoice Information:			TAT: Standard 30 day			X	Rush	Mark One		
Lab Name:	GEL Laboratories, LLC	Site ID #:	TRONOX LLC, HENDERSON	Send Invoice to:	Susan Crowley	Address:	PO Box 66	If Rush, Date due	QC level Required:			Standard	Special	EPA Stage	Mark One
Address:	2040 Savage Road	Project #:	2027.001	City/State:	Henderson, NV 89009	Phone #:	(949) 260-9293	NJ Reduced Deliverable Package?						4	
Lab P.M.:	Edith M. Kent	Site Address:	660 W. Lake Mead Drive	Reimbursement project?	X	Non-Reimbursement project?		Mark one							
Phone/Fax:	(843)568-8171	Site PM Name:	Derrick Willis	Send EDD to:	Frank Hagar Northgate Environmental Management, Inc	Phone #:		MA/MCP Cert?				CT RCP Cert?			Mark One
Lab P.M. email:	emk@gel.com	Phone/Fax:	949-378-7004	CC Hardcopy report to:	PDF Electronic Version Only	CC Hardcopy report to:		Lab Project ID (lab use)							
Applicable Lab Quote #:		Site PM Email:	derrick.willis@ngem.com	see additional comments below											
Requestested Analyses															
EML HASL 300*, DML HASL 300, EPA 8031/SB20, DMSO3, HNO3, HCl, NaOH, Na2S2O3, Methanol, Other															
Comments/Lab Sample I.D.															
ITEM #	SAMPLE ID	Matrix	Valid Matrix Codes	Matrix Code	SAMPLE TYPE-COMP	#OF CONTAINERS	SAMPLE TIME	SAMPLE DATE	Preservatives						
									One	W/G	G	10/29/09	10/29/09	1	N
1	M-148B	WATER	WW	WW	WATER/SLUDGE	1	1	N	X	X	2 L Poly Clear				
2	M-148B	WATER	WW	WW	WATER/SLUDGE	1	1	N	X	X	2 L Poly Clear				
3	M-148B	WATER	WW	WW	WATER/SLUDGE	1	1	N	X	X	2 L Poly Clear				
4	M-148B	WATER	WW	WW	WATER/SLUDGE	1	1	N	X	X	2 L Poly Clear				
5															
6															
7															
8															
9															
10															
11															
12															
Additional Comments/Special Instructions:															
FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (isotopic) and Uranium (isotopic)															
All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: cindy.arnold@ngem.com frank.hagar@ngem.com															
SHIPPING METHOD (check as appropriate) UPS COURIER FEDEX US MAIL															
Sample Receipt Conditions Date: 10/29/09 Time: 14:45 Signature: John C. Willis															
Print Name of Sampler: John C. Willis Signature of Sampler: John C. Willis															
Term in OC Samples Impact? Tmp Blank? Samples in OC															

SAMPLE RECEIPT & REVIEW FORM

Client:	<u>Kerr / Northgate</u>		SDG/ARCO/C Work Order:	<u>239753 1-</u>
Received By:	<u>MK</u>		Date Received:	<u>10-30-09</u>
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group of further investigation.	
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>		Maximum Counts Observed*: <u>Open 30</u>	
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>			
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>			
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>		Hazard Class Shipped:	UN#:
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>			

Sample Receipt Criteria			Comments/Qualifiers (Required for Non-Conforming Items)		
	Yes	NA	No		
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)	
2 Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		<input checked="" type="checkbox"/>		Preservation Method: <u>ddc</u> ice bags blue ice dry ice <u>none</u> other (describe)	
3 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>				
4 Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: seals broken damaged container leaking container other (describe)	
5 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>			Sample ID's, containers affected and observed pH: If Preservation added, Lot#:	
6 VOA vials free of headspace (defined as < 6mm bubble)?		<input checked="" type="checkbox"/>		Sample ID's and containers affected:	
7 Are Encore containers present?			<input checked="" type="checkbox"/>	(If yes, immediately deliver to Volatiles laboratory)	
8 Samples received within holding time?	<input checked="" type="checkbox"/>			Id's and tests affected:	
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:	
10 Date & time on COC match date & time on bottles?			<input checked="" type="checkbox"/>	Sample ID's affected: <u>* see below</u>	
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:	
12 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>				

Comments:

* NO DATE OR TIME ON CHAIN FOR M-137B + M-137BDISS
 NO TIME ON CHAIN FOR M-148B

M-148B - 10-29-09 0910
 m-137B - 10-29-09 1330
 m-137BDISS - 10-29-09 1330

FX 7970 6563 0421

PM (or PMA) review: Initials gk Date 10/30/09

SAMPLE RECEIPT & REVIEW FORM

Client:	<u>Kerr / Northeast</u>			SDG/ARCO/C/Work Order: <u>2397531-</u>
Received By:	<u>MK</u>			Date Received: <u>10-31-09</u>
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.	
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>		Maximum Counts Observed*: <u>open 20</u>	
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>			
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>			
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>		Hazard Class Shipped:	UN#:
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>			

Sample Receipt Criteria			Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>					Circle Applicable: seals broken damaged container leaking container other (describe)
2 Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		<input checked="" type="checkbox"/>				Preservation Method: <u>19c</u> ice bags blue ice dry ice <u>none</u> other (describe)
3 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>					
4 Sample containers intact and sealed?	<input checked="" type="checkbox"/>					Circle Applicable: seals broken damaged container leaking container other (describe)
5 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>					Sample ID's, containers affected and observed pH: If Preservation added, Lot#:
6 VOA vials free of headspace (defined as < 6mm bubble)?		<input checked="" type="checkbox"/>				Sample ID's and containers affected:
7 Are Encore containers present?		<input checked="" type="checkbox"/>				(If yes, immediately deliver to Volatiles laboratory)
8 Samples received within holding time?	<input checked="" type="checkbox"/>					Id's and tests affected:
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>					Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>					Sample ID's affected:
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>					Sample ID's affected:
12 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>					

Comments:

FX 7960 8122 5281

2397537

northgate

environmental management, inc.

1100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9283

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027-001-01127
Page 1 of 1
Cooler # _____
Collection Area: III

Required Ship to Lab:		Required Project Information:		Required Invoice Information:		TAT: Standard 30 day		X	Rush																																																																																																																																																																						
Lab Name:	GEL Laboratories, LLC	Site ID #:	TRONOX LLC, HENDERSON	Send invoice to:	Susan Crowley	If Rush:	Date due			Mark One																																																																																																																																																																					
Address:	2040 Savage Road	Project #:	2027-001	Address:	PO Box 55																																																																																																																																																																										
Charleston, SC 29407		Site Address:	660 W. Lake Mead Drive	City/State:	Henderson, NV 89008	Phone #:	(849)250-9293																																																																																																																																																																								
Lab P.M.: Edith M. Kent		City:	Henderson	State:	NV	Reimbursement project?	<input checked="" type="checkbox"/>	X	Non-reimbursment project?	Mark one																																																																																																																																																																					
Phone/Fax:	(843)568-4171	Site PM Name:	Derrick Willis	Send EDD to:	Frank.Hagar@ngem.com	QC level Required:	Standard	Special	EPA Stage	Mark one																																																																																																																																																																					
Lab PM email:	emk@gel.com	Phone/Fax:	949-376-7004	CC Hardcopy report to:	PDF Electronic Version Only	MA MCP Cert?	CT RCP Cert?			Mark One																																																																																																																																																																					
Applicable Lab Quote #:		Site PM Email:	derrick.willis@ngem.com	CC Hardcopy report to:	see additional comments below	Lab Project ID (lab use)																																																																																																																																																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2" rowspan="2">SAMPLE ID</th> <th colspan="2" rowspan="2">Character per box. (A-Z, 0-9, /,)</th> <th colspan="2" rowspan="2"># OF CONTAINERS</th> <th colspan="2" rowspan="2">FIELD FILTERED (Y/N)</th> <th colspan="2" rowspan="2">Preservatives</th> <th colspan="2" rowspan="2">Comments/Lab Sample I.D.</th> </tr> <tr> <th>One</th> <th>M</th> <th>NaOH</th> <th>HCl</th> <th>HNO3</th> <th>H2SO4</th> <th>Uptreated</th> <th>Metallic</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>M-147B</td> <td>WG</td> <td>G</td> <td>11/2/2009</td> <td>1000</td> <td>1</td> <td>N</td> <td>X</td> <td>X</td> <td colspan="2">2 L Poly Clear</td> </tr> <tr> <td>2</td> <td>M-147B</td> <td>WG</td> <td>G</td> <td>11/2/2009</td> <td>1000</td> <td>1</td> <td>N</td> <td>X</td> <td>X</td> <td colspan="2">2 L Poly Clear</td> </tr> <tr> <td>3</td> <td>M-147BDISS</td> <td>WG</td> <td>G</td> <td>11/2/2009</td> <td>1</td> <td>Y</td> <td>X</td> <td>*</td> <td>*</td> <td colspan="2">2 L Poly Clear - JWS</td> </tr> <tr> <td>4</td> <td>M-147BDISS</td> <td>WG</td> <td>S</td> <td>4420000</td> <td>1</td> <td>Y</td> <td>X</td> <td>*</td> <td>*</td> <td colspan="2">2 L Poly Clear - JWS</td> </tr> <tr> <td>5</td> <td>M-147009B</td> <td>WG</td> <td>G</td> <td>11/2/2009</td> <td>1000</td> <td>1</td> <td>N</td> <td>X</td> <td>X</td> <td colspan="2">2 L Poly Clear</td> </tr> <tr> <td>6</td> <td>M-147009B</td> <td>WG</td> <td>G</td> <td>11/2/2009</td> <td>1000</td> <td>1</td> <td>N</td> <td>X</td> <td>X</td> <td colspan="2">2 L Poly Clear</td> </tr> <tr> <td>7</td> <td>M-147009BTS</td> <td>WG</td> <td>C</td> <td>4420000</td> <td>1</td> <td>Y</td> <td>X</td> <td>*</td> <td>*</td> <td colspan="2">2 L Poly Clear - JWS</td> </tr> <tr> <td>8</td> <td>M-147009BTS</td> <td>WG</td> <td>S</td> <td>11/2/2009</td> <td>1</td> <td>Y</td> <td>X</td> <td>*</td> <td>*</td> <td colspan="2">2 L Poly Clear - JWS</td> </tr> <tr> <td>9</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td colspan="2"></td> </tr> <tr> <td>10</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td colspan="2"></td> </tr> <tr> <td>11</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td colspan="2"></td> </tr> <tr> <td>12</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td colspan="2"></td> </tr> </tbody> </table>												SAMPLE ID		Character per box. (A-Z, 0-9, /,)		# OF CONTAINERS		FIELD FILTERED (Y/N)		Preservatives		Comments/Lab Sample I.D.		One	M	NaOH	HCl	HNO3	H2SO4	Uptreated	Metallic	1	M-147B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear		2	M-147B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear		3	M-147BDISS	WG	G	11/2/2009	1	Y	X	*	*	2 L Poly Clear - JWS		4	M-147BDISS	WG	S	4420000	1	Y	X	*	*	2 L Poly Clear - JWS		5	M-147009B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear		6	M-147009B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear		7	M-147009BTS	WG	C	4420000	1	Y	X	*	*	2 L Poly Clear - JWS		8	M-147009BTS	WG	S	11/2/2009	1	Y	X	*	*	2 L Poly Clear - JWS		9												10												11												12											
SAMPLE ID		Character per box. (A-Z, 0-9, /,)		# OF CONTAINERS		FIELD FILTERED (Y/N)		Preservatives		Comments/Lab Sample I.D.																																																																																																																																																																					
												One	M	NaOH	HCl	HNO3	H2SO4	Uptreated	Metallic																																																																																																																																																												
1	M-147B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear																																																																																																																																																																					
2	M-147B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear																																																																																																																																																																					
3	M-147BDISS	WG	G	11/2/2009	1	Y	X	*	*	2 L Poly Clear - JWS																																																																																																																																																																					
4	M-147BDISS	WG	S	4420000	1	Y	X	*	*	2 L Poly Clear - JWS																																																																																																																																																																					
5	M-147009B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear																																																																																																																																																																					
6	M-147009B	WG	G	11/2/2009	1000	1	N	X	X	2 L Poly Clear																																																																																																																																																																					
7	M-147009BTS	WG	C	4420000	1	Y	X	*	*	2 L Poly Clear - JWS																																																																																																																																																																					
8	M-147009BTS	WG	S	11/2/2009	1	Y	X	*	*	2 L Poly Clear - JWS																																																																																																																																																																					
9																																																																																																																																																																															
10																																																																																																																																																																															
11																																																																																																																																																																															
12																																																																																																																																																																															
Additional Comments/Special Instructions: FULL DIGESTION SPECIFICATION EMSL HASL 300* - DOE EMSL HASL 300 modified (alpha spectroscopy) Thorium (isotopic) and Uranium (isotopic) All PDF reports and EDDs will be uploaded to: Northgate Environmental Management, Inc. FTP site address provided to labs Notifications provided to: cindy.arnold@ngem.com frank.hagar@ngem.com																																																																																																																																																																															
Temp in QC	Temp Blank?	Sample in QC	Sample Blank?	On Ice?	Accepted by Lab/Att/Arch	Date:	Time:	Sample Receipt Conditions																																																																																																																																																																							
UPS COURIER	FEDEX	US MAIL																																																																																																																																																																													
<p>SHIPPING METHOD (mark as appropriate)</p> <p>Signature of CARRIER: <u>John W. Otis</u></p> <p>Print Name of CARRIER: FEDEX</p> <p>Signature of SAMPLER: <u>John W. Otis</u></p> <p>Print Name of SAMPLER: FEDEX</p>																																																																																																																																																																															

northgate

environmental management, inc.

1100 Quail Street, Suite 102, Newport Beach, CA 92660
(949) 260-9283

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed and accurate.

COC No. 2027.001.01134
Page: 1 of 1
Cooler #: _____

Collection Area: III

Required Ship to Lab:		Required Project Information:		Required Invoice Information:		TAT: Standard 30 day		X	Rush	Mark One
Lab Name:	GEL Laboratories, LLC	Site ID #:	TRONOX LLC, HENDERSON	Send Invoice to:	Susan Crowley Tronox LLC	If Rush, Date due:				
Address:	2040 Savage Road	Project #:	2027.001	Address:	PO Box 56	QC level Required:	Standard	Special	EPA Stage	Mark one
Charleston, SC 29407		Site Address:	560 W. Lake Mead Drive	City/State:	Henderson, NV 89009	Phone #:	(949)280-9283	NJ Reduced Deliverable Package?		
Lab P.M.: Edith M. Kent	City: Henderson	State: NV	Reimbursement project?	X	Non-reimbursement project?		Mark one			
Phone/Fax: (843)666-8171	Site PM Name:	Derrick Willis	Send EDD to:	Frank.Hagar@ngem.com	CC Hardcopy report to:	PDF Electronic Version Only	MA MCP Cert?	CTR CP Cert?		Mark One
Lab PM email: emk@gel.com	Phone/Fax:	848-375-7004	Site PM Email:	derrick.willis@ngem.com	CC Hardcopy report to:	see additional comments below	Lab Project ID (lab use)			
Applicable Lab Quote #:										
Additional Comments:										
<p>Sample ID: EB110209-GWA3 Character per box: One # Samples: MUST BE UNIQUE</p> <p>Valid Matrix Codes: MATRIX Matrix Code: G=GRAINE-C=COMP</p> <p>Waste Matrix Codes: MATRIX Matrix Code: G=GRAINE-C=COMP</p> <p>WATER WO SURFACE WATER WW</p> <p>GROUND WATER WO FREE SURFACE WATER WW</p> <p>WASTE WATER LF SLUDGE L</p> <p>FREE PRODUCT SO INHIBITANT S</p> <p>BIO ANIMAL TISSUE A</p> <p>SOIL EYE LASH AE</p> <p>WIRE AIR SVE LASH S</p> <p>SOIL GELS O</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Requesteted Analyses: EMA 903, ICP-MS, EML HASL 300, ICP-MS</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2 L Poly Clear</p>										
<p>Preservatives: FIELD FILTERED (Y/N)</p> <p>None</p> <p>NaOH</p> <p>HCl</p> <p>HNO3</p> <p>H2SO4</p> <p>Other</p> <p>Na2S2O3</p> <p>Methanol</p> <p>Umapreserved</p>										
<p>Comments/Lab Sample I.D.: 2</p>										

SAMPLE RECEIPT & REVIEW FORM

Client: <u>Kell/Northente</u>	SDG/ARCO/C Work Order: <u>239753-1</u>		
Received By: <u>mk</u>	Date Received: <u>11-3-09</u>		
Suspected Hazard Information	Yes	No	*If Counts > x2 area background on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.
COC/Samples marked as radioactive?	<input checked="" type="checkbox"/>		Maximum Counts Observed*: <u>yes 20</u>
Classified Radioactive II or III by RSO?	<input checked="" type="checkbox"/>		
COC/Samples marked containing PCBs?	<input checked="" type="checkbox"/>		
Shipped as a DOT Hazardous?	<input checked="" type="checkbox"/>		Hazard Class Shipped: UN#:
Samples identified as Foreign Soil?	<input checked="" type="checkbox"/>		

Sample Receipt Criteria			Comments/Qualifiers (Required for Non-Conforming Items)
	Yes	NA	No
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>		Circle Applicable: seals broken damaged container leaking container other (describe)
2 Samples requiring cold preservation within ($0 \leq 6$ deg. C)?		<input checked="" type="checkbox"/>	Preservation Method: ice bags blue ice dry ice <u>none</u> other (describe) <u>2 Kc</u>
3 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>		
4 Sample containers intact and sealed?	<input checked="" type="checkbox"/>		Circle Applicable: seals broken damaged container leaking container other (describe)
5 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>		Sample ID's, containers affected and observed pH: If Preservation added, Lot#:
6 VOA vials free of headspace (defined as < 6mm bubble)?		<input checked="" type="checkbox"/>	Sample ID's and containers affected:
7 Are Encore containers present?		<input checked="" type="checkbox"/>	(If yes, immediately deliver to Volatiles laboratory)
8 Samples received within holding time?	<input checked="" type="checkbox"/>		Id's and tests affected:
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>		Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>		Sample ID's affected:
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>		Sample ID's affected:
12 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>		

Comments:

EX 7970 7371 6153

Subject: RE: COC 2027.001.01123, Please verify
From: <frank.hagar@ngem.com>
Date: Thu, 29 Oct 2009 17:52:46 -0700
To: "Heather Shaffer" <Heather.Shaffer@gel.com>

The DISS samples are field filtered

From: Heather Shaffer [mailto:Heather.Shaffer@gel.com]
Sent: Thursday, October 29, 2009 12:50 PM
To: Cindy Arnold; Frank Hagar; Derrick Willis; Vivian Willis
Cc: Edie Kent
Subject: COC 2027.001.01123, Please verify

The containers for M-138BDIIS and M-138009BDIIS are marked on the label as "N" filtered. The chain of custody lists these samples as "Y" Filtered. Please verify if these two samples are or are not field filtered.

Thank you,
Heather

--
Heather Shaffer
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Road
Charleston, SC (USA) 29407
Main: 843.556.8171 x 4505
Fax: 843.766.1178
E-mail: heather.shaffer@gel.com
Web: www.gel.com

CONFIDENTIALITY NOTICE: This e-mail and any files transmitted with it are the property of The GEL Group, Inc. and its affiliates. All rights, including without limitation copyright, are reserved. The proprietary information contained in this e-mail message, and any files transmitted with it, is intended for the use of the recipient(s) named above. If the reader of this e-mail is not the intended recipient, you are hereby notified that you have received this e-mail in error and that any review, distribution or copying of this e-mail or any files transmitted with it is strictly prohibited. If you have received this e-mail in error, please notify the sender immediately and delete the original message and any files transmitted. The unauthorized use of this e-mail or any files transmitted with it is prohibited and disclaimed by The GEL Group, Inc. and its affiliates.

Subject: RE: Henderson Samples Received Today, 10/30/09 - Condition on Receipt
From: "Cindy Arnold" <carnold@ngem.com>
Date: Fri, 30 Oct 2009 18:18:58 +0000
To: emk@gel.com, Cindy.Arnold@ngem.com, Frank.Hagar@ngem.com, Derrick.Willis@ngem.com, Team.Kent@gel.com, vivian.willis@verdant-solutions.com

Thats fine Edie. Thank you, Cindy

----- Original Message ----- On 10/30/2009 3:34 PM Edie Kent wrote:
For samples M-137B and M-137BDIIS, COC# 2027.001.01121, there is no date or time of collection listed on the chain of custody. For sample M-148B, COC# 2027.001.01128, the time of collection is listed in both the sample date and sample time column on the chain of custody. GEL will use the following dates and times taken from the sample containers:

M-137B: 10/29/09 @ 13:30
M-137BDIIS: 10/29/09 @ 13:30
M-148B: 10/29/09 @ 0910

Edie

--
Edith M. Kent
Project Manager
GEL Laboratories, LLC
2040 Savage Road
Charleston, SC (USA) 29407
Direct: 843.769.7385 x4453
Main: 843.556.8171
Fax: 843.766.1178
E-mail: emk@gel.com
Web: www.gel.com

CONFIDENTIALITY NOTICE: This e-mail and any files transmitted with it are the property of The GEL Group, Inc. and its affiliates. All rights, including without limitation copyright, are reserved. The proprietary information contained in this e-mail message, and any files transmitted with it, is intended for the use of the recipient(s) named above. If the reader of this e-mail is not the intended recipient, you are hereby notified that you have received this e-mail in error and that any review, distribution or copying of this e-mail or any files transmitted with it is strictly prohibited. If you have received this e-mail in error, please notify the sender immediately and delete the original message and any files transmitted. The unauthorized use of this e-mail or any files transmitted with it is prohibited and disclaimed by The GEL Group, Inc. and its affiliates.

Subject: GEL Closed SDGs 239753

From: Heather Shaffer <Heather.Shaffer@gel.com>

Date: Mon, 09 Nov 2009 12:53:46 -0500

To: Cindy Arnold <Cindy.Arnold@ngem.com>, Frank Hagar <Frank.Hagar@ngem.com>, Edie Kent <emk@gel.com>, Derrick Willis <Derrick.Willis@ngem.com>

CC: Heather Shaffer <hea01394@gel.com>, Mercedes Simmons <mer01583@gel.com>

As of today we closed water SDG 239753.

Attached is a list of the samples in the SDG. As soon as we have completed the login review, you will receive the full receipt package for these SDG.

Thank you,
Heather

--

Heather Shaffer
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Road
Charleston, SC (USA) 29407
Main: 843.556.8171 x 4505
Fax: 843.766.1178
E-mail: heather.shaffer@gel.com
Web: www.gel.com

239753.xls

Content-Type: application/msexcel
Content-Encoding: base64

Subject: Re: SDG 239753 QC Issues - Alpha Spec Th, Alpha Spec U, Ra-226
From: Heather Shaffer <Heather.Shaffer@gel.com>
Date: Thu, 03 Dec 2009 12:53:23 -0500
To: Cindy Arnold <Cindy.Arnold@ngem.com>, Derrick Willis <Derrick.Willis@ngem.com>
CC: Edie Kent <emk@gel.com>

CORRECTION: THIS EMAIL IS FOR WORK ORDER 239753

Heather Shaffer wrote:

The following are the QC issues regarding this SDG for Ra-226, Alpha Spec Th and Alpha Spec U:

Ra-226 Issues

The following samples do not meet the Tronox QA program sample result uncertainty limit of <30% with activity between 2 and 5 times the MDA and were counted the maximum possible count time: 239753006, 239753014 and 239753016.

Thorium Issues:

Samples 239753001, 239753002, 239753005, 239753006 and 239753011 do not meet the Th-228 detection limits. The samples were analyzed with an appropriate aliquot for the method and matrix. Additionally, the samples were counted the maximum count time to achieve the lowest possible MDAs.

The method blank does not meet the Th-232 detection limits due to keeping the blank aliquot consistent with the other sample aliquots.

The following samples did not meet the Tronox QA program tracer yield requirement of 70-120%: 239753002. The samples met GEL's standard tracer yield requirements. The blank, LCS, and all other samples met the contract tracer yield requirements.

The following samples do not meet the Tronox QA program sample result uncertainty limit of <30% with activity between 2 and 5 times the MDA and were counted the maximum possible count time: 239753006, 239753012, and 239753016.

Uranium Issues:

The following samples did not meet the Tronox QA program tracer yield requirement of 70-120%: 239753001, 239753002, 239753009, 239753011, and 239753017. The samples met GEL's standard tracer yield requirements. The blank, LCS, and all other samples met the contract tracer yield requirements.

Sample 239753003 did not meet the Tronox QA program detection limit for U-233/234, U-235/236, and U-238. The sample size is restricted in the attempt to assure achieved yield recoveries meet the program yield requirements and to reduce the chance of tailing from U-233/234 activity into the U-235/236 region of interest. The samples were counted the maximum possible count time in order

to achieve the lowest possible MDA.

The following samples do not meet the Tronox QA program sample result uncertainty limit of <30% with activity between 2 and 5 times the MDA and were counted the maximum possible count time: 239753018.

Please note: The lab noticed that the levels of Uranium in these water samples were high compared to what they were used to seeing for this project. Attached is a Results greater than MDA report from the lab.

- This will be noted in the case narrative.

--
Heather Shaffer
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Road
Charleston, SC (USA) 29407
Main: 843.556.8171 x 4505
Fax: 843.766.1178
E-mail: heather.shaffer@gel.com
Web: www.gel.com

--
Heather Shaffer
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Road
Charleston, SC (USA) 29407
Main: 843.556.8171 x 4505
Fax: 843.766.1178
E-mail: heather.shaffer@gel.com
Web: www.gel.com

Laboratory Certifications

List of current GEL Certifications as of 03 December 2009

State	Certification
Arizona	AZ0668
Arkansas	88-0651
CLIA	42D0904046
California – NELAP	01151CA
Colorado	GEL
Connecticut	PH-0169
Dept. of Navy	NFESC 413
EPA Region 5	WG-15J
Florida – NELAP	E87156
Georgia	E87156 (FL/NELAP)
Georgia DW	967
Hawaii	N/A
ISO 17025	2567.01
Idaho	SC00012
Illinois – NELAP	200029
Indiana	C-SC-01
Kansas – NELAP	E-10332
Kentucky	90129
Louisiana – NELAP	03046
Maryland	270
Massachusetts	M-SC012
Nevada	SC00012
New Jersey – NELAP	SC002
New Mexico	FL NELAP E87156
New York – NELAP	11501
North Carolina	233
North Carolina DW	45709
Oklahoma	9904
Pennsylvania – NELAP	68-00485
South Carolina	10120001/10120002
Tennessee	TN 02934
Texas – NELAP	T104704235-07B-TX
U.S. Dept. of Agriculture	S-52597
Utah – NELAP	GEL
Vermont	VT87156
Virginia	00151
Washington	C1641

RADIOLOGICAL ANALYSIS

**Radiochemistry Case Narrative
Tronox LLC (ENSR)
SDG 239753**

Method/Analysis Information

Product: **Alphaspec Th, Liquid**

Analytical Method: DOE EML HASL-300, Th-01-RC Modified

Analytical Batch Number: 923093

Sample ID	Client ID
239753001	M-141B
239753002	M-141009B
239753003	PB102309-A3
239753004	M-145B
239753005	M-139B
239753006	M-146B
239753007	M-144B
239753008	M-138B
239753009	M-138009B
239753010	M-138BDISS
239753011	M-138009BDISS
239753012	M-137B
239753013	M-137BDISS
239753014	M-148B
239753015	EB103009-GWA4
239753016	M-147B
239753017	M-147009B
239753018	EB110209-GWA3
1201973223	Method Blank (MB)
1201973224	239753003(PB102309-A3) Sample Duplicate (DUP)
1201973225	239753003(PB102309-A3) Matrix Spike (MS)
1201973226	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-038 REV# 12.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met.

Standards Information

Standard solution(s) for these analyses are NIST traceable and used before the expiration date(s).

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

The blank volume is representative of the sample volume in this batch.

Designated QC

The following sample was used for QC: 239753003 (PB102309-A3).

QC Information

Refer to Non-Conformance Report.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

NCR Documentation

Nonconformance reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. The following NCR was generated for this SDG:
NCR 762500 was generated due to RDL less than MDA, Failed Recovery for Surrogate or Tracer and Other. 1. Samples 239753001, 239753002, 239753005, 239753006 and 239753011 do not meet the required detection limit for Th228. The blank, 1201973223, does not meet the required detection limit for Th232. 2. Sample 239753002 does not meet the client's tracer yield requirement of 70 - 120%. 3. Sample 239753006 has Thorium-230 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Samples 239753006 and 239753012 have Thorium-232 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Sample 239753016 has Thorium-228 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. 1. The blank, 1201973223, did not meet the detection limit due to keeping the blank volume consistent with the other sample aliquots. The samples were analyzed with an appropriate aliquot for the method and matrix. Additionally, the samples were counted 1000 minutes. Per GELs accredited methods and SOPs, further corrective action is not required. PM notified, reporting results. 2. The sample does meet the GEL standard tracer yield requirement. The Method Blank and the Laboratory Control sample meet the client's tracer yield requirement. PM notified, reporting results. 3. Samples were all counted the maximum count time of 1000 minutes to achieve the best possible uncertainties. PM notified, reporting results.

Manual Integration

No manual integrations were performed on data in this batch.

Additional Comments

Additional comments were not required for this sample set.

Qualifier information

Manual qualifiers were not required.

Method/Analysis Information

Product: Alphaspec U, Liquid

Analytical Method: DOE EML HASL-300, U-02-RC Modified

Analytical Batch Number: 923094

Sample ID	Client ID
239753001	M-141B
239753002	M-141009B
239753003	PB102309-A3
239753004	M-145B
239753005	M-139B
239753006	M-146B
239753007	M-144B
239753008	M-138B
239753009	M-138009B
239753010	M-138BDISS
239753011	M-138009BDISS
239753012	M-137B
239753013	M-137BDISS
239753014	M-148B
239753015	EB103009-GWA4
239753016	M-147B
239753017	M-147009B
239753018	EB110209-GWA3
1201973227	Method Blank (MB)
1201973228	239753004(M-145B) Sample Duplicate (DUP)
1201973229	239753004(M-145B) Matrix Spike (MS)
1201973230	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-011 REV# 18.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met.

Standards Information

Standard solution(s) for these analyses are NIST traceable and used before the expiration date(s).

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

The blank volume is representative of the sample volume in this batch.

Designated QC

The following sample was used for QC: 239753004 (M-145B).

QC Information

All of the QC samples met the required acceptance limits.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

NCR Documentation

Nonconformance reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. The following NCR was generated for this SDG:

NCR 762566 was generated due to RDL less than MDA, Failed Recovery for Surrogate or Tracer and Other. 1. Samples 239753001, 239753002, 239753009, 239753011, and 239753017 do not meet the client tracer yield requirements of 70 to 120 percent due to the matrix of the samples. 2. Sample 239753003 did not meet the detection limits for U-233/234, U-235/6, and U-233/234 as a result of the restricted size of the aliquot used. 3. Sample 239753018 has Uranium-233/234 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. 1. The GEL standard tracer yield requirements of 15 to 125 percent were met and the Method blank and the LCS for the batch did meet the client tracer yield recovery requirements. PM notified, reporting results. 2. The aliquot size used was appropriate for the analysis method used and a larger aliquot could lead to resolution and yield recovery issues. The sample was counted 1000 minutes to achieve the lowest possible MDA's. PM notified, reporting results. 3. Sample was counted the maximum count time of 1000 minutes to achieve the best possible uncertainties. PM notified, reporting results.

Manual Integration

No manual integrations were performed on data in this batch.

Additional Comments

The U-235/236 blank result is equal to the MDA but less than the detection limit.

Qualifier information

Manual qualifiers were not required.

Method/Analysis Information

Product: GFPC, Ra228, Liquid

Analytical Method: EPA 904.0/SW846 9320 Modified

Analytical Batch Number: 922859

Sample ID	Client ID
239753001	M-141B
239753002	M-141009B
239753003	PB102309-A3
239753004	M-145B
239753005	M-139B
239753006	M-146B
239753007	M-144B
239753008	M-138B
239753009	M-138009B
239753010	M-138BDISS
239753011	M-138009BDISS
239753012	M-137B
239753013	M-137BDISS
239753014	M-148B
239753015	EB103009-GWA4
239753016	M-147B
239753017	M-147009B
239753018	EB110209-GWA3
1201972468	Method Blank (MB)
1201972469	239753001(M-141B) Sample Duplicate (DUP)
1201972470	239753001(M-141B) Matrix Spike (MS)
1201972471	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-009 REV# 15.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met.

Standards Information

Standard solution(s) for these analyses are NIST traceable and used before the expiration date(s).

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

The blank volume is representative of the sample volume in this batch.

Designated QC

The following sample was used for QC: 239753001 (M-141B).

QC Information

All of the QC samples met the required acceptance limits.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Chemical Recoveries

All chemical recoveries meet the required acceptance limits for this sample set.

Miscellaneous Information:

NCR Documentation

Nonconformance reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A nonconformance report (NCR) was not generated for this SDG.

Additional Comments

Additional comments were not required for this sample set.

Qualifier information

Manual qualifiers were not required.

Method/Analysis Information

Product: Lucas Cell, Ra226, liquid

Analytical Method: EPA 903.1 Modified

Analytical Batch Number: 920697

Sample ID	Client ID
239753001	M-141B
239753002	M-141009B
239753003	PB102309-A3
239753004	M-145B
239753005	M-139B
239753006	M-146B
239753007	M-144B
239753008	M-138B
239753009	M-138009B
239753010	M-138BDISS
239753011	M-138009BDISS
239753012	M-137B
239753013	M-137BDISS
239753014	M-148B
239753015	EB103009-GWA4
239753016	M-147B
239753017	M-147009B
239753018	EB110209-GWA3
1201967363	Method Blank (MB)
1201967364	239753012(M-137B) Sample Duplicate (DUP)
1201967365	239753012(M-137B) Matrix Spike (MS)
1201967366	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-008 REV# 13.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met.

Standards Information

Standard solution(s) for these analyses are NIST traceable and used before the expiration date(s).

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

The blank volume is representative of the sample volume in this batch.

Designated QC

The following sample was used for QC: 239753012 (M-137B).

QC Information

All of the QC samples met the required acceptance limits.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

NCR Documentation

Nonconformance reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. The following NCR was generated for this SDG:
NCR 764411 was generated due to Other. 1. Samples 239753006, 239753014 and 239753016 have Radium-226 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Samples were all counted the maximum count time of 30 minutes to achieve the best possible uncertainties. 1. PM notified, reporting results.

Additional Comments

Additional comments were not required for this sample set.

Qualifier information

Manual qualifiers were not required.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Review Validation:

GEL requires all analytical data to be verified by a qualified data validator. In addition, all data designated for CLP or CLP-like packaging will receive a third level validation upon completion of the data package.

The following data validator verified the information presented in this case narrative:

Reviewer/Date: _____

 / 12/2/09

COMPANY - WIDE NONCONFORMANCE REPORT

Mo.Day Yr. 25-NOV-09	Division: Radiochemistry	Quality Criteria: Specifications	Type: Process			
Instrument Type: ALPHA SPECTROMETER	Test / Method: DOE EML HASL-300, Th-01-RC Modified	Matrix Type: Liquid	Client Code: KERR			
Batch ID: 923093	Sample Numbers: See Below					
Potentially affected work order(s)(SDG): 239753						
Application Issues: RDL less than MDA Failed Recovery for Surrogate or Tracer Other						
Specification and Requirements Nonconformance Description:	NRG Disposition:					
1. Samples 239753001, 239753002, 239753005, 239753006 and 239753011 do not meet the required detection limit for Th228. The blank, 1201973223, does not meet the required detection limit for Th232. 2. Sample 239753002 does not meet the client's tracer yield requirement of 70 - 120%. 3. Sample 239753006 has Thorium-230 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Samples 239753006 and 239753012 have Thorium-232 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Sample 239753016 has Thorium-228 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity.	1. The blank, 1201973223, did not meet the detection limit due to keeping the blank volume consistent with the other sample aliquots. The samples were analyzed with an appropriate aliquot for the method and matrix. Additionally, the samples were counted 1000 minutes. Per GELs accredited methods and SOPs, further corrective action is not required. PM notified, reporting results. 2. The sample does meet the GEL standard tracer yield requirement. The Method Blank and the Laboratory Control sample meet the client's tracer yield requirement. PM notified, reporting results. 3. Samples were all counted the maximum count time of 1000 minutes to achieve the best possible uncertainties. PM notified, reporting results.					

Originator's Name:

Joseph Moulden 25-NOV-09

Data Validator/Group Leader:

Eric Brimstain 25-NOV-09

COMPANY - WIDE NONCONFORMANCE REPORT			
Mo.Day Yr. 25-NOV-09	Division: Radiochemistry	Quality Criteria: Specifications	Type: Process
Instrument Type: ALPHA SPECTROMETER	Test / Method: DOE EML HASL-300, U-02-RC Modified	Matrix Type: Liquid	Client Code: KERR
Batch ID: 923094	Sample Numbers: See Below		
Potentially affected work order(s)(SDG): 239753			
Application Issues: RDL less than MDA Failed Recovery for Surrogate or Tracer Other			
Specification and Requirements Nonconformance Description:		NRG Disposition:	
1. Samples 239753001, 239753002, 239753009, 239753011, and 239753017 do not meet the client tracer yield requirements of 70 to 120 percent due to the matrix of the samples. 2. Sample 239753003 did not meet the detection limits for U-233/234, U-235/6, and U-233/234 as a result of the restricted size of the aliquot used. 3. Sample 239753018 has Uranium-233/234 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity.		1. The GEL standard tracer yield requirements of 15 to 125 percent were met and the Method blank and the LCS for the batch did meet the client tracer yield recovery requirements. PM notified, reporting results. 2. The aliquot size used was appropriate for the analysis method used and a larger aliquot could lead to resolution and yield recovery issues. The sample was counted 1000 minutes to achieve the lowest possible MDA's. PM notified, reporting results. 3. Sample was counted the maximum count time of 1000 minutes to achieve the best possible uncertainties. PM notified, reporting results.	

Originator's Name:

Jessica Downey 25-NOV-09

Data Validator/Group Leader:

Scott Moreland 30-NOV-09

COMPANY - WIDE NONCONFORMANCE REPORT			
Mo.Day Yr. 02-DEC-09	Division: Radiochemistry	Quality Criteria: Specifications	Type: Process
Instrument Type: LUCAS CELL DETECTOR	Test / Method: EPA 903.1 Modified	Matrix Type: Liquid	Client Code: KERR
Batch ID: 920697	Sample Numbers: See below		
Potentially affected work order(s)(SDG): 239753			
Application Issues: Other			
Specification and Requirements Nonconformance Description:		NRG Disposition:	
1. Samples 239753006, 239753014 and 239753016 have Radium-226 activity between two and five times the MDA and uncertainty greater than 30% of that respective activity. Samples were all counted the maximum count time of 30 minutes to achieve the best possible uncertainties.		1. PM notified, reporting results.	

Originator's Name:

Lyndsey Pace 02-DEC-09

Data Validator/Group Leader:

Lesley Anderson 02-DEC-09

SAMPLE DATA SUMMARY

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis Report for

KERR003 Tronox LLC

Client SDG: 239753 GEL Work Order: 239753

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the Certificate of Analysis.

The designation ND, if present, appears in the result column when the analyte concentration is not detected above the detection limit.

This data report has been prepared and reviewed in accordance with GEL Laboratories LLC standard operating procedures. Please direct any questions to your Project Manager, Edith Kent.

Reviewed by



12/2/09

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-141B	Project:	KERRHenderson
Sample ID:	239753001	Client ID:	KERR003
Matrix:	WG		
Collect Date:	23-OCT-09 10:00		
Receive Date:	26-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0254	+/-0.037	0.0638	0.030	pCi/L		KXM	11/20/09	1604	923093	1
							4					
Thorium-230	U	0.00978	+/-0.0136	0.0234	0.030	pCi/L						
Thorium-232	U	0.00734	+/-0.0107	0.0187	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		76.5	+/-1.05	0.0887	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		2.85	+/-0.226	0.0567	0.030	pCi/L						
Uranium-238		49.2	+/-0.840	0.0571	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	2.40	+/-1.63	2.49	3.00	pCi/L		JXC5	11/23/09	1726	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.228	+/-0.177	0.256	1.00	pCi/L		KSD1	12/01/09	1805	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			70.1	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			59.4	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			75.9	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID:	M-141009B	Project:	KERRHenderson
Sample ID:	239753002	Client ID:	KERR003
Matrix:	WG		
Collect Date:	23-OCT-09 10:00		
Receive Date:	26-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00331	+/-0.0219	0.0436	0.030	pCi/L		KXM	11/20/09	1604	923093	1
							4					
Thorium-230	U	-0.00786	+/-0.0115	0.0322	0.030	pCi/L						
Thorium-232	U	-0.00524	+/-0.0126	0.0322	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		75.1	+/-1.32	0.140	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		2.95	+/-0.291	0.0569	0.030	pCi/L						
Uranium-238		50.7	+/-1.08	0.0973	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228		2.39	+/-1.51	2.29	3.00	pCi/L		JXC5	11/23/09	1723	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.458	+/-0.222	0.241	1.00	pCi/L		KSD1	12/01/09	1840	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			67.5	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			37.6	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			90.8	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID: PB102309-A3
 Sample ID: 239753003
 Matrix: WG
 Collect Date: 23-OCT-09 12:15
 Receive Date: 26-OCT-09
 Collector: Client

Project: KERRHenderson
 Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00702	+/-0.0175	0.0322	0.030	pCi/L		KXM	11/20/09	1604	923093	
							4					
Thorium-230	U	0.00819	+/-0.0106	0.0181	0.030	pCi/L						
Thorium-232	U	0.00164	+/-0.00718	0.0157	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234	U	-0.00298	+/-0.0295	0.0599	0.030	pCi/L		KXM	11/21/09	1552	923094	
							4					
Uranium-235/236	U	0.00	+/-0.0152	0.0371	0.030	pCi/L						
Uranium-238	U	-0.00627	+/-0.0174	0.042	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	1.54	+/-1.72	2.89	3.00	pCi/L		JXC5	11/23/09	1726	922859	
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.275	+/-0.211	0.317	1.00	pCi/L		KSD1	12/01/09	1840	920697	
							4					

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			101	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			70.4	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			87.4	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-145B	Project:	KERRHenderson
Sample ID:	239753004	Client ID:	KERR003
Matrix:	WG		
Collect Date:	26-OCT-09 10:15		
Receive Date:	27-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00963	+/-0.0155	0.0276	0.030	pCi/L		KXM	11/20/09	1604	923093	1
							4					
Thorium-230	U	0.0123	+/-0.0103	0.0134	0.030	pCi/L						
Thorium-232	U	-0.00175	+/-0.00769	0.0194	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		18.2	+/-0.433	0.0384	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		0.702	+/-0.0956	0.0365	0.030	pCi/L						
Uranium-238		12.2	+/-0.355	0.0408	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228		3.05	+/-1.81	2.85	3.00	pCi/L		JXC5	11/23/09	1726	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.334	+/-0.243	0.360	1.00	pCi/L		KSD1	12/01/09	1840	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			105	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			82.6	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			86.6	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID: M-139B
 Sample ID: 239753005
 Matrix: WG
 Collect Date: 26-OCT-09 12:55
 Receive Date: 27-OCT-09
 Collector: Client

Project: KERRHenderson
 Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0191	+/-0.0282	0.0485	0.030	pCi/L		KXM	11/20/09	1604	923093	1
							4					
Thorium-230		0.0232	+/-0.0145	0.0183	0.030	pCi/L						
Thorium-232	U	-0.00166	+/-0.00975	0.0222	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		5.77	+/-0.222	0.0533	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		0.209	+/-0.0472	0.0207	0.030	pCi/L						
Uranium-238		3.81	+/-0.179	0.0242	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	0.00759	+/-1.31	2.48	3.00	pCi/L		JXC5	11/23/09	1726	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.490	+/-0.269	0.360	1.00	pCi/L		KSD1	12/01/09	1840	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			109	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			99.4	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			88.4	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-146B	Project:	KERRHenderson
Sample ID:	239753006	Client ID:	KERR003
Matrix:	WG		
Collect Date:	27-OCT-09 09:30		
Receive Date:	28-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0426	+/-0.0284	0.0436	0.030	pCi/L		KXM	11/20/09	1604	923093	1
							4					
Thorium-230		0.0599	+/-0.0225	0.0229	0.030	pCi/L						
Thorium-232		0.0428	+/-0.0187	0.0189	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		11.2	+/-0.341	0.0512	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		0.444	+/-0.0768	0.0366	0.030	pCi/L						
Uranium-238		7.62	+/-0.281	0.0359	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	0.391	+/-1.38	2.56	3.00	pCi/L		JXC5	11/23/09	1726	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.880	+/-0.318	0.347	1.00	pCi/L		KSD1	12/01/09	1840	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			102	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			86.3	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			72.0	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID: M-144B
Sample ID: 239753007
Matrix: WG
Collect Date: 27-OCT-09 12:25
Receive Date: 28-OCT-09
Collector: Client

Project: KERRHenderson
Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.013	+/-0.019	0.0329	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230		0.0157	+/-0.0115	0.0151	0.030	pCi/L						
Thorium-232	U	-1.87E-10	+/-0.00436	0.012	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		20.9	+/-0.466	0.0414	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		0.929	+/-0.112	0.0481	0.030	pCi/L						
Uranium-238		14.5	+/-0.388	0.0299	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	-0.654	+/-1.48	2.83	3.00	pCi/L		JXC5	11/23/09	1726	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.414	+/-0.215	0.271	1.00	pCi/L		KSD1	12/01/09	1840	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			104	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			83.0	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			88.1	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-138B	Project:	KERRHenderson
Sample ID:	239753008	Client ID:	KERR003
Matrix:	WG		
Collect Date:	28-OCT-09 11:15		
Receive Date:	29-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0253	+/-0.0184	0.027	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230		0.0327	+/-0.0162	0.0165	0.030	pCi/L						
Thorium-232		0.019	+/-0.0131	0.0165	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		40.9	+/-0.688	0.0434	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.20	+/-0.131	0.0285	0.030	pCi/L						
Uranium-238		20.7	+/-0.490	0.0231	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	1.16	+/-1.62	2.76	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.671	+/-0.304	0.374	1.00	pCi/L		KSD1	12/01/09	1910	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			106	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			74.7	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			86.6	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID:	M-138009B	Project:	KERRHenderson
Sample ID:	239753009	Client ID:	KERR003
Matrix:	WG		
Collect Date:	28-OCT-09 11:15		
Receive Date:	29-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00697	+/-0.0161	0.0297	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	0.00817	+/-0.0147	0.0264	0.030	pCi/L						
Thorium-232	U	0.0131	+/-0.012	0.0181	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		41.6	+/-0.794	0.0436	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.53	+/-0.170	0.0467	0.030	pCi/L						
Uranium-238		20.9	+/-0.562	0.0377	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	-0.616	+/-1.53	2.83	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.540	+/-0.257	0.323	1.00	pCi/L		KSD1	12/01/09	1910	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			114	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			56.7	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			94.9	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-138BDIIS	Project:	KERRHenderson
Sample ID:	239753010	Client ID:	KERR003
Matrix:	WG		
Collect Date:	28-OCT-09 11:15		
Receive Date:	29-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0194	+/-0.0163	0.0248	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	0.00336	+/-0.00932	0.0186	0.030	pCi/L						
Thorium-232	U	0.00336	+/-0.00466	0.00504	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		39.1	+/-0.687	0.0507	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.03	+/-0.124	0.0116	0.030	pCi/L						
Uranium-238		19.5	+/-0.485	0.0507	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	0.394	+/-1.42	2.60	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.252	+/-0.195	0.282	1.00	pCi/L		KSD1	12/01/09	1910	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			103	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			73.1	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			77.3	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID: M-138009BDIIS
Sample ID: 239753011
Matrix: WG
Collect Date: 28-OCT-09 11:15
Receive Date: 29-OCT-09
Collector: Client

Project: KERRHenderson
Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0346	+/-0.0251	0.039	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	-0.0032	+/-0.00887	0.0214	0.030	pCi/L						
Thorium-232	U	-0.0032	+/-0.0117	0.0259	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		38.5	+/-0.765	0.0728	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.10	+/-0.145	0.0468	0.030	pCi/L						
Uranium-238		19.2	+/-0.539	0.0302	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	2.19	+/-1.65	2.60	3.00	pCi/L		JXC5	11/23/09	1723	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.276	+/-0.173	0.227	1.00	pCi/L		KSD1	12/01/09	1910	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			98.3	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			57.3	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			81.4	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID: M-137B
 Sample ID: 239753012
 Matrix: WG
 Collect Date: 29-OCT-09 13:30
 Receive Date: 30-OCT-09
 Collector: Client

Project: KERRHenderson
 Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228		0.0733	+/-0.0301	0.0394	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230		0.103	+/-0.0276	0.0208	0.030	pCi/L						
Thorium-232		0.0711	+/-0.023	0.0187	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		34.0	+/-0.618	0.0359	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.01	+/-0.118	0.0108	0.030	pCi/L						
Uranium-238		19.0	+/-0.461	0.0323	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	0.624	+/-1.61	2.85	3.00	pCi/L		JXC5	11/23/09	1723	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.177	+/-0.159	0.242	1.00	pCi/L		KSD1	12/01/09	1910	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			97.2	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			77.7	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			94.7	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID: M-137BDIIS
 Sample ID: 239753013
 Matrix: WG
 Collect Date: 29-OCT-09 13:30
 Receive Date: 30-OCT-09
 Collector: Client

Project: KERRHenderson
 Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.0314	+/-0.0218	0.033	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	0.00461	+/-0.00998	0.0189	0.030	pCi/L						
Thorium-232	U	-0.00768	+/-0.00903	0.0235	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		37.1	+/-0.676	0.0246	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.16	+/-0.134	0.038	0.030	pCi/L						
Uranium-238		19.9	+/-0.496	0.0462	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	2.47	+/-1.81	2.85	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.580	+/-0.283	0.360	1.00	pCi/L		KSD1	12/01/09	2000	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			101	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			71.0	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			81.1	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID:	M-148B	Project:	KERRHenderson
Sample ID:	239753014	Client ID:	KERR003
Matrix:	WG		
Collect Date:	29-OCT-09 09:10		
Receive Date:	30-OCT-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228		0.0268	+/-0.0173	0.0229	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	0.00545	+/-0.00796	0.0139	0.030	pCi/L						
Thorium-232	U	0.00182	+/-0.00941	0.0201	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		32.3	+/-0.593	0.00852	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.19	+/-0.127	0.0269	0.030	pCi/L						
Uranium-238		21.9	+/-0.489	0.0217	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	1.55	+/-1.53	2.51	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.592	+/-0.234	0.235	1.00	pCi/L		KSD1	12/01/09	2000	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			96.0	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			80.8	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			84.2	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID: EB103009-GWA4
Sample ID: 239753015
Matrix: WG
Collect Date: 30-OCT-09 11:10
Receive Date: 31-OCT-09
Collector: Client

Project: KERRHenderson
Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00853	+/-0.0157	0.0282	0.030	pCi/L		KXM	11/20/09	1605	923093	1
							4					
Thorium-230	U	-0.00162	+/-0.00842	0.020	0.030	pCi/L						
Thorium-232	U	-0.00162	+/-0.00842	0.020	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		0.0272	+/-0.0176	0.0192	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236	U	0.00	+/-0.0122	0.0298	0.030	pCi/L						
Uranium-238	U	0.0176	+/-0.0148	0.0192	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	-1.49	+/-1.09	2.50	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.244	+/-0.196	0.293	1.00	pCi/L		KSD1	12/01/09	2000	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			104	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			91.7	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			85.9	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID: M-147B
Sample ID: 239753016
Matrix: WG
Collect Date: 02-NOV-09 10:00
Receive Date: 03-NOV-09
Collector: Client

Project: KERRHenderson
Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228		0.017	+/-0.0119	0.00647	0.030	pCi/L		KXM	11/20/09	1416	923093	
							4					
Thorium-230	U	0.00846	+/-0.0102	0.0162	0.030	pCi/L						
Thorium-232	U	-0.00212	+/-0.00587	0.0162	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		26.1	+/-0.527	0.0398	0.030	pCi/L		KXM	11/21/09	1552	923094	
							4					
Uranium-235/236		1.09	+/-0.120	0.0261	0.030	pCi/L						
Uranium-238		19.6	+/-0.456	0.0265	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	1.35	+/-1.69	2.87	3.00	pCi/L		JXC5	11/23/09	1727	922859	
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.784	+/-0.297	0.333	1.00	pCi/L		KSD1	12/01/09	2000	920697	
							4					

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			103	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			80.7	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			75.3	(15%-125%)

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
 Address : 1100 Quail St., Suite 102
 Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
 Project: **Tronox Henderson**

Client Sample ID:	M-147009B	Project:	KERRHenderson
Sample ID:	239753017	Client ID:	KERR003
Matrix:	WG		
Collect Date:	02-NOV-09 10:00		
Receive Date:	03-NOV-09		
Collector:	Client		

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00564	+/-0.0124	0.0244	0.030	pCi/L		KXM	11/20/09	1416	923093	1
							4					
Thorium-230	U	0.00499	+/-0.00978	0.0191	0.030	pCi/L						
Thorium-232	U	-0.00249	+/-0.00691	0.0191	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		26.2	+/-0.575	0.044	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236		1.07	+/-0.130	0.0311	0.030	pCi/L						
Uranium-238		19.6	+/-0.497	0.0251	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	0.949	+/-1.33	2.29	3.00	pCi/L		JXC5	11/23/09	1727	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226		0.303	+/-0.182	0.233	1.00	pCi/L		KSD1	12/01/09	2000	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			84.4	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			67.3	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			79.6	(15%-125%)

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Northgate Environmental Management, Inc.
Address : 1100 Quail St., Suite 102
Newport Beach, California 92660

Report Date: December 2, 2009

Contact: Mr. Frank Hagar
Project: **Tronox Henderson**

Client Sample ID: EB110209-GWA3
Sample ID: 239753018
Matrix: WG
Collect Date: 02-NOV-09 12:40
Receive Date: 03-NOV-09
Collector: Client

Project: KERRHenderson
Client ID: KERR003

Parameter	Qualifier	Result	Uncertainty	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Rad Alpha Spec Analysis												
<i>Alphaspec Th, Liquid "As Received"</i>												
Thorium-228	U	0.00543	+/-0.00998	0.0191	0.030	pCi/L		KXM	11/20/09	1416	923093	1
							4					
Thorium-230	U	0.00244	+/-0.0107	0.0234	0.030	pCi/L						
Thorium-232	U	-0.00244	+/-0.00677	0.0187	0.030	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>												
Uranium-233/234		0.067	+/-0.028	0.0276	0.030	pCi/L		KXM	11/21/09	1552	923094	2
							4					
Uranium-235/236	U	0.00309	+/-0.00605	0.00926	0.030	pCi/L						
Uranium-238		0.035	+/-0.023	0.0307	0.030	pCi/L						
Rad Gas Flow Proportional Counting												
<i>GFPC, Ra228, Liquid "As Received"</i>												
Radium-228	U	1.14	+/-1.54	2.63	3.00	pCi/L		JXC5	11/23/09	1723	922859	3
Rad Radium-226												
<i>Lucas Cell, Ra226, liquid "As Received"</i>												
Radium-226	U	0.108	+/-0.194	0.349	1.00	pCi/L		KSD1	12/01/09	2035	920697	4

The following Analytical Methods were performed

Method	Description	Analyst Comments
1	DOE EML HASL-300, Th-01-RC Modified	
2	DOE EML HASL-300, U-02-RC Modified	
3	EPA 904.0/SW846 9320 Modified	
4	EPA 903.1 Modified	

Surrogate/Tracer recovery	Test	Result	Nominal	Recovery%	Acceptable Limits
Actinium-227 Tracer	Alphaspec Th, Liquid "As Received"			90.2	(15%-125%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"			87.2	(15%-125%)
Barium-133 Tracer	GFPC, Ra228, Liquid "As Received"			79.8	(15%-125%)

QUALITY CONTROL DATA

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: December 2, 2009
Page 1 of 3

Northgate Environmental Management, Inc.
1100 Quail St., Suite 102
Newport Beach, California

Contact: Mr. Frank Hagar

Workorder: 239753

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Rad Alpha Spec											
Batch	923093										
QC1201973224	239753003	DUP									
Thorium-228			U	0.00702 +/-0.0175	U 0.00499 +/-0.00948	pCi/L	0.00		N/A	KXM4	11/20/09 14:16
Thorium-230			U	0.00819 +/-0.0106	U 0.00232 +/-0.00455	pCi/L	0.00		N/A		
Thorium-232			U	0.00164 +/-0.00718	U 0.00232 +/-0.00455	pCi/L	0.00		N/A		
QC1201973226	LCS				U 0.0016 +/-0.0134	pCi/L					11/20/09 16:05
Thorium-228		2.68			2.11 +/-0.112	pCi/L	78.7	(75%-125%)			
Thorium-230					U 0.00936 +/-0.00967	pCi/L		(75%-125%)			
QC1201973223	MB				U 0.00468 +/-0.0168	pCi/L					11/20/09 14:16
Thorium-228					U -0.00684 +/-0.00999	pCi/L					
Thorium-230					U -0.0205 +/-0.0141	pCi/L					
QC1201973225	239753003	MS									
Thorium-228			U	0.00702 +/-0.0175	U 0.00191 +/-0.0152	pCi/L					11/20/09 14:16
Thorium-230		2.68	U	0.00819 +/-0.0106	2.22 +/-0.134	pCi/L	82.9	(75%-125%)			
Thorium-232			U	0.00164 +/-0.00718	U 0.00 +/-0.00582	pCi/L		(75%-125%)			
Batch	923094										
QC1201973228	239753004	DUP									
Uranium-233/234				18.2 +/-0.433	18.6 +/-0.422	pCi/L	2.10	(0% - 20%)	KXM4	11/20/09 14:25	
Uranium-235/236				0.702 +/-0.0956	0.635 +/-0.0878	pCi/L	9.99	(0% - 20%)			
Uranium-238				12.2 +/-0.355	12.6 +/-0.346	pCi/L	2.52	(0% - 20%)			
QC1201973230	LCS				3.12 +/-0.167	pCi/L					11/20/09 14:25
Uranium-233/234					0.178 +/-0.0443	pCi/L					
Uranium-235/236											
Uranium-238		3.15			3.25	pCi/L	103	(75%-125%)			

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 239753

Page 2 of 3

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Rad Alpha Spec											
Batch	923094										
QC1201973227	MB				+/-0.171						
Uranium-233/234				U	0.0139	pCi/L				KXM4	11/20/09 14:25
Uranium-235/236					+/-0.0132						
Uranium-238					0.00895	pCi/L					
					+/-0.0101						
QC1201973229	239753004 MS			U	0.00	pCi/L					
Uranium-233/234					+/-0.0134						
Uranium-235/236											
Uranium-238											
QC1201972469	239753001 DUP										
Radium-228				U	2.40	U	0.755	pCi/L	0.00	N/A	JXC5 11/23/09 17:26
					+/-1.63		+/-1.26				
QC1201972471	LCS										
Radium-228		39.2					41.8	pCi/L		106	(75%-125%) 11/23/09 17:26
							+/-4.30				
QC1201972468	MB										
Radium-228				U			0.876	pCi/L			11/23/09 17:27
							+/-1.63				
QC1201972470	239753001 MS										
Radium-228		79.1	U	2.40			93.4	pCi/L		118	(75%-125%) 11/23/09 17:26
				+/-1.63			+/-9.04				
Rad Ra-226											
Batch	920697										
QC1201967364	239753012 DUP										
Radium-226				U	0.177	U	0.187	pCi/L	5.52	N/A	KSD1 12/01/09 20:35
					+/-0.159		+/-0.216				
QC1201967366	LCS										
Radium-226		24.2					19.4	pCi/L		80.4	(75%-125%) 12/01/09 21:25
							+/-1.24				
QC1201967363	MB										
Radium-226				U			0.320	pCi/L			12/01/09 20:35
							+/-0.260				
QC1201967365	239753012 MS										
Radium-226		121	U	0.177			102	pCi/L		84.4	(75%-125%) 12/01/09 20:35
				+/-0.159			+/-6.18				

Notes:

The Qualifiers in this report are defined as follows:

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 239753

Page 3 of 3

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
**	Analyte is a surrogate compound										
<	Result is less than value reported										
>	Result is greater than value reported										
A	The TIC is a suspected aldol-condensation product										
B	For General Chemistry and Organic analysis the target analyte was detected in the associated blank.										
BD	Results are either below the MDC or tracer recovery is low										
C	Analyte has been confirmed by GC/MS analysis										
D	Results are reported from a diluted aliquot of the sample										
F	Estimated Value										
H	Analytical holding time was exceeded										
J	Value is estimated										
M	M if above MDC and less than LLD										
M	Matrix Related Failure										
N/A	RPD or %Recovery limits do not apply.										
ND	Analyte concentration is not detected above the detection limit										
NJ	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier										
R	Sample results are rejected										
U	Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.										
UI	Gamma Spectroscopy--Uncertain identification										
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier										
Y	QC Samples were not spiked with this compound										
^	RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL. Qualifier Not Applicable for Radiochemistry.										
h	Preparation or preservation holding time was exceeded										

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more.

[^] The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

RAW DATA

THORIUM

Radiochemistry Batch Checklist, Rev 9

Batch# 923093 Product: Th Date: 11/25/09

Criteria:	Yes	No	Comments
Sample Solids are less than or equal to 100 mg for GAB.			N/A
Samples have been blank corrected (if required)			N/A
If activity less 10% MDA/ MDC, error is 150% or less of sample activity. If greater 10% MDA/ MDC, error is 40% or less. If below the MDA/ MDC, error is okay.	✓		
Instrument source check is within limits.	✓		
Instrument bkg check is within limits.			
Method RDL/ LLD has been met.		✓	NCR 762500
If duplicate activities are less 5% MDA/ MDC, then RPD is 100% or less. If greater 5% MDA/ MDC, then RPD 20% or less. If below the MDA/ MDC, the RPD is 0%.	✓		
Or meets the client's required RER acceptance criteria.			
Tracer yield is 15-125%. Carrier yield 25-125%.		✓	NCR 762500
Or meets the client's contract acceptance criteria.			
Method blank is less than the RDL/ LLD. (If rad samples, < 5% of lowest activity)	✓		
Sample was run within hold time.	✓		
Sample was correctly preserved if required.	✓		
Smears Taken for Radioactive batches.			N/A
Method Spike and LCS are within 75-125% or meets the client's contract acceptance criteria.	✓		
No blank spaces on data forms.	✓		
All line outs initialed and dated.			
No transcription errors are apparent.			
Aux data is correct.			N/A
Client Special requirements page has been checked.	✓		
Raw Data and/or spectrum are included and properly statused.	✓		
QC data entered into QC database and batch is in REVW	✓		
Hit notification complete (if necessary)			N/A
Batch entered into Case Narrative.	✓		
Batch non-conformances completed, if applicable.	✓		NCR 762500
Batch non-conformances second reviewed and disposition verified to be completed.	✓		NCR 762500
Aliquot Correction completed if required.			N/A
Review sample historical results if available (If REMP, results above MDC have been verified by historical results, recount or re-analysis.)	✓		

GEL Laboratories, LLC

revised 8/1/08

Primary Review Performed By:

Jopl M/L 11/25/09

Secondary Review Performed By:

2 - ~~est~~ 11/25/09

Thorium (Ac-227 Tracer) Que Sheet

16-NOV-09

Batch #: 923093 Analyst: KXM4 First Client Due Date:07-DEC-09
600 Tracer Isotope: Ac-227 Tracer Code: 0387-B-102 Expiration Date: 7-13-10 Internal Due Date:06-NOV-09
LCS Isotope: Th-230 LCS Code: A2796-5 Vol: 0.1mL Ac-227 Separation Date/Time: 11-19-09 15:55
Spike Isotope: Th-230 Spike Code: A2796-5 Expiration Date: 7-13-10
Prep Date: 11-18-09 Initials: MJM Pipet ID: 7971058 Balance ID: 16758257 Witness: MDA 11/18/09

Sample ID	Client Description	Type	Hazard Code	Min CRDL	Matrix	Client	Collection Date	Aliquot Pos.	Label #	Wet/Dry Aliquot (g/g/f)	Th Det #
239753001-1	M-141B	SAMPLE	.03	PCi/L	WATER	KERR003	23-OCT-09	0.800	1	<u>8000.0</u>	25
239753002-1	M-141009B	SAMPLE	.03	PCi/L	WATER	KERR003	23-OCT-09	0.800	2	<u>8000.0</u>	26
239753003-1	PB102309-A3	SAMPLE	.03	PCi/L	WATER	KERR003	23-OCT-09	0.800	3	<u>8000.0</u>	27
239753004-1	M-145B	SAMPLE	.03	PCi/L	WATER	KERR003	26-OCT-09	0.800	4	<u>8000.0</u>	28
239753005-1	M-139B	SAMPLE	.03	PCi/L	WATER	KERR003	26-OCT-09	0.800	5	<u>8000.0</u>	29
239753006-1	M-146B	SAMPLE	.03	PCi/L	WATER	KERR003	27-OCT-09	0.800	6	<u>8000.0</u>	30
239753007-1	M-144B	SAMPLE	.03	PCi/L	WATER	KERR003	27-OCT-09	0.800	7	<u>8000.0</u>	31
239753008-1	M-138B	SAMPLE	.03	PCi/L	WATER	KERR003	28-OCT-09	0.800	8	<u>8000.0</u>	32
239753009-1	M-138009B	SAMPLE	.03	PCi/L	WATER	KERR003	28-OCT-09	0.800	9	<u>8000.0</u>	33
239753010-1	M-138BDISS	SAMPLE	.03	PCi/L	WATER	KERR003	28-OCT-09	0.800	10	<u>8000.0</u>	34
239753011-1	M-138009BDISS	SAMPLE	.03	PCi/L	WATER	KERR003	28-OCT-09	0.800	11	<u>8000.0</u>	35
239753012-1	M-137B	SAMPLE	.03	PCi/L	WATER	KERR003	29-OCT-09	0.800	12	<u>8000.0</u>	36
239753013-1	M-137BDISS	SAMPLE	.03	PCi/L	WATER	KERR003	29-OCT-09	0.800	13	<u>8000.0</u>	37
239753014-1	M-148B	SAMPLE	.03	PCi/L	WATER	KERR003	29-OCT-09	0.800	14	<u>8000.0</u>	38
239753015-1	EB103009-GWA4	SAMPLE	.03	PCi/L	WATER	KERR003	30-OCT-09	0.800	15	<u>8000.0</u>	39
239753016-1	M-147B	SAMPLE	.03	PCi/L	WATER	KERR003	02-NOV-09	0.800	16	<u>8000.0</u>	40
239753017-1	M-147009B	SAMPLE	.03	PCi/L	WATER	KERR003	02-NOV-09	0.800	17	<u>8000.0</u>	41
239753018-1	EB110209-GWA3	SAMPLE	.03	PCi/L	WATER	KERR003	02-NOV-09	0.800	18	<u>8000.0</u>	42
1201973223-1	MB for batch 923093	MB	.03	PCi/L	WATER	QC ACCOUNT					
1201973224-1	PB102309-A3(239753003DUP)	DUP	.03	PCi/L	WATER	QC ACCOUNT	23-OCT-09	0.800	19	<u>8000.0</u>	43
1201973225-1	PB102309-A3(239753003MS)	MS	.03	PCi/L	WATER	QC ACCOUNT	23-OCT-09	0.800	20	<u>8000.0</u>	44
1201973226-1	LCS for batch 923093	LCS	.03	PCi/L	WATER	QC ACCOUNT					

Choose SOP Used: GL-RAD-A-038
 GL-RAD-A-045
 GL-RAD-A-043
 GL-RAD-A-032
 GL Laboratories LLC, Radiochemistry Division
Solid Sample Dissolution by: LEACH or DIGESTION

Circle One

Data Reviewed By: J. L. P. S. 11/15/09Date Reviewed: 11/15/09


GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 23-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753001_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :45-149AA5 AVERAGE %EFFICIENCY :32.8418 % YIELD : 70.102	COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 2.72263 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B025.CNF;1082 BKG DATE : 15-NOV-2009 EFF FILE : W025.CNF;322 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	500.000	484.000	16.000	4.0000	57.44000	2.19E+00	2.33E-01	9.76E-02	4.20E-02	2.01E-01
TH-228	5363.000	35.000	10.100	23.000	4.7958	99.94000	2.54E-02	3.70E-02	6.38E-02	2.81E-02	3.70E-02
TH229	4900.000	15.000	12.000	3.000	1.7321	99.52000	2.95E-02	2.05E-02	2.72E-02	9.90E-03	2.04E-02
TH-230	4625.000	6.000	4.000	2.000	1.4142	100.0000	9.78E-03	1.36E-02	2.34E-02	8.05E-03	1.36E-02
TH-232	3972.000	4.000	3.000	1.000	1.0000	100.0000	7.34E-03	1.07E-02	1.87E-02	5.69E-03	1.07E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 23-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753002_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78204 AVERAGE %EFFICIENCY :31.8739 % YIELD : 67.455	COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 2.61982 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B026.CNF;1083 BKG DATE : 15-NOV-2009 EFF FILE : W026.CNF;296 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	465.000	452.000	13.000	3.6056	57.44000	2.19E+00	2.42E-01	9.57E-02	4.06E-02	2.07E-01
TH-228	5363.000	11.000	1.225	8.000	2.8284	99.94000	3.31E-03	2.19E-02	4.36E-02	1.77E-02	2.19E-02
TH229	4900.000	9.000	0.000	9.000	3.0000	99.52000	2.51E-09	2.19E-02	4.46E-02	1.84E-02	2.19E-02
TH-230	4625.000	1.000	-3.000	4.000	2.0000	100.0000	-7.86E-03	1.15E-02	3.22E-02	1.22E-02	1.15E-02
TH-232	3972.000	2.000	-2.000	4.000	2.0000	100.0000	-5.24E-03	1.26E-02	3.22E-02	1.22E-02	1.26E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 23-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00			SAMPLE ID : S0239753003_TH SAMPLE QTY: 0.800 L
DETECTOR NUMBER :42484 AVERAGE %EFFICIENCY :34.1663 % YIELD : 100.659			COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.90940 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B027.CNF;1089 BKG DATE : 15-NOV-2009 EFF FILE : W027.CNF;323 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	738.000	723.000	15.000	3.8730	57.44000	2.19E+00	2.06E-01	6.36E-02	2.73E-02	1.63E-01
TH-228	5363.000	19.000	4.162	12.000	3.4641	99.94000	7.02E-03	1.75E-02	3.22E-02	1.36E-02	1.75E-02
TH229	4900.000	7.000	2.000	5.000	2.2361	99.52000	3.29E-03	1.12E-02	2.21E-02	8.56E-03	1.12E-02
TH-230	4625.000	8.000	5.000	3.000	1.7321	100.0000	8.19E-03	1.07E-02	1.81E-02	6.60E-03	1.06E-02
TH-232	3972.000	3.000	1.000	2.000	1.4142	100.0000	1.64E-03	7.18E-03	1.57E-02	5.39E-03	7.18E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 26-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753004_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78792 AVERAGE %EFFICIENCY :30.5052 % YIELD : 105.255	COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.08789 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B028.CNF;1093 BKG DATE : 15-NOV-2009 EFF FILE : W028.CNF;315 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	685.000	675.000	10.000	3.1623	57.44000	2.19E+00	2.09E-01	5.74E-02	2.38E-02	1.67E-01
TH-228	5363.000	15.000	5.350	7.000	2.6458	99.94000	9.63E-03	1.55E-02	2.76E-02	1.11E-02	1.55E-02
TH229	4900.000	9.000	5.000	4.000	2.0000	99.52000	8.81E-03	1.25E-02	2.17E-02	8.20E-03	1.25E-02
TH-230	4625.000	8.000	7.000	1.000	1.0000	100.0000	1.23E-02	1.03E-02	1.34E-02	4.08E-03	1.03E-02
TH-232	3972.000	2.000	-1.000	3.000	1.7321	100.0000	-1.75E-03	7.69E-03	1.94E-02	7.07E-03	7.69E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 26-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753005_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :33454 AVERAGE %EFFICIENCY :31.2344 % YIELD : 108.737	COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.22313 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B029.CNF;1084 BKG DATE : 15-NOV-2009 EFF FILE : W029.CNF;314 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	729.000	714.000	15.000	3.8730	57.44000	2.19E+00	2.06E-01	6.44E-02	2.76E-02	1.64E-01
TH-228	5363.000	44.000	11.197	30.000	5.4772	99.94000	1.91E-02	2.82E-02	4.85E-02	2.17E-02	2.82E-02
TH229	4900.000	11.000	2.000	9.000	3.0000	99.52000	3.33E-03	1.46E-02	2.83E-02	1.16E-02	1.46E-02
TH-230	4625.000	17.000	14.000	3.000	1.7321	100.0000	2.32E-02	1.46E-02	1.83E-02	6.68E-03	1.45E-02
TH-232	3972.000	4.000	-1.000	5.000	2.2361	100.0000	-1.66E-03	9.75E-03	2.22E-02	8.62E-03	9.75E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 27-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00			SAMPLE ID : S0239753006_TH SAMPLE QTY: 0.800 L
DETECTOR NUMBER :33447 AVERAGE %EFFICIENCY :32.3788 % YIELD : 101.662			COUNT DATE:20-NOV-2009 16:04:59 ELAPSED LIVE TIME(SEC): 59999.99 ANALYST :KXM4
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.94835 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B030.CNF;1081 BKG DATE : 15-NOV-2009 EFF FILE : W030.CNF;299 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	713.000	692.000	21.000	4.5826	57.44000	2.19E+00	2.10E-01	7.69E-02	3.37E-02	1.68E-01
TH-228	5363.000	49.000	24.283	22.000	4.6904	99.94000	4.26E-02	2.85E-02	4.36E-02	1.91E-02	2.84E-02
TH229	4900.000	9.000	-3.000	12.000	3.4641	99.52000	-5.16E-03	1.54E-02	3.29E-02	1.39E-02	1.54E-02
TH-230	4625.000	40.000	35.000	5.000	2.2361	100.0000	5.99E-02	2.28E-02	2.29E-02	8.90E-03	2.25E-02
TH-232	3972.000	28.000	25.000	3.000	1.7321	100.0000	4.28E-02	1.88E-02	1.89E-02	6.89E-03	1.87E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 27-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753007_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :79988 AVERAGE %EFFICIENCY :34.3928 % YIELD : 104.145	COUNT DATE:20-NOV-2009 16:05:00 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.04479 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B031.CNF;1079 BKG DATE : 15-NOV-2009 EFF FILE : W031.CNF;338 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	762.000	753.000	9.000	3.0000	57.44000	2.19E+00	1.97E-01	4.92E-02	2.03E-02	1.58E-01
TH-228	5363.000	25.000	8.044	14.000	3.7417	99.94000	1.30E-02	1.90E-02	3.29E-02	1.40E-02	1.90E-02
TH229	4900.000	14.000	-3.000	17.000	4.1231	99.52000	-4.74E-03	1.72E-02	3.50E-02	1.52E-02	1.72E-02
TH-230	4625.000	12.000	10.000	2.000	1.4142	100.0000	1.57E-02	1.16E-02	1.51E-02	5.17E-03	1.15E-02
TH-232	3972.000	1.000	0.000	1.000	1.0000	100.0000	-1.87E-10	4.36E-03	1.20E-02	3.66E-03	4.36E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 28-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753008_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78785 AVERAGE %EFFICIENCY :30.9759 % YIELD : 105.498	COUNT DATE:20-NOV-2009 16:05:00 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.09734 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B033.CNF;1078 BKG DATE : 15-NOV-2009 EFF FILE : W033.CNF;324 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	701.000	687.000	14.000	3.7417	57.44000	2.19E+00	2.05E-01	6.50E-02	2.77E-02	1.67E-01
TH-228	5363.000	24.000	14.303	7.000	2.6458	99.94000	2.53E-02	1.85E-02	2.70E-02	1.09E-02	1.84E-02
TH229	4900.000	29.000	24.000	5.000	2.2361	99.52000	4.16E-02	1.99E-02	2.32E-02	9.01E-03	1.98E-02
TH-230	4625.000	21.000	19.000	2.000	1.4142	100.0000	3.27E-02	1.63E-02	1.65E-02	5.67E-03	1.62E-02
TH-232	3972.000	13.000	11.000	2.000	1.4142	100.0000	1.90E-02	1.31E-02	1.65E-02	5.67E-03	1.31E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 28-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753009_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78202 AVERAGE %EFFICIENCY :30.1182 % YIELD : 114.346	COUNT DATE:20-NOV-2009 16:05:00 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.44098 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B035.CNF;1076 BKG DATE : 15-NOV-2009 EFF FILE : W035.CNF;313 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	743.000	724.000	19.000	4.3589	57.44000	2.19E+00	2.02E-01	7.03E-02	3.06E-02	1.63E-01
TH-228	5363.000	17.000	4.158	10.000	3.1623	99.94000	6.97E-03	1.61E-02	2.97E-02	1.23E-02	1.61E-02
TH229	4900.000	22.000	2.000	20.000	4.4721	99.52000	3.29E-03	2.09E-02	3.91E-02	1.71E-02	2.09E-02
TH-230	4625.000	13.000	5.000	8.000	2.8284	100.0000	8.17E-03	1.47E-02	2.64E-02	1.08E-02	1.47E-02
TH-232	3972.000	11.000	8.000	3.000	1.7321	100.0000	1.31E-02	1.20E-02	1.81E-02	6.59E-03	1.20E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 28-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753010_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78203 AVERAGE %EFFICIENCY :32.5831 % YIELD : 102.776	COUNT DATE:20-NOV-2009 16:05:00 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.99163 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B036.CNF;1074 BKG DATE : 15-NOV-2009 EFF FILE : W036.CNF;325 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	723.000	704.000	19.000	4.3589	57.44000	2.19E+00	2.04E-01	7.23E-02	3.15E-02	1.66E-01
TH-228	5363.000	20.000	11.236	6.000	2.4495	99.94000	1.94E-02	1.63E-02	2.48E-02	9.82E-03	1.63E-02
TH-229	4900.000	21.000	16.000	5.000	2.2361	99.52000	2.70E-02	1.69E-02	2.26E-02	8.79E-03	1.69E-02
TH-230	4625.000	5.000	2.000	3.000	1.7321	100.0000	3.36E-03	9.32E-03	1.86E-02	6.78E-03	9.32E-03
TH-232	3972.000	2.000	2.000	0.000	0.0000	100.0000	3.36E-03	4.66E-03	5.04E-03	0.00E+00	4.66E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 28-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753011_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :45-149BB5 AVERAGE %EFFICIENCY :35.8039 % YIELD : 98.313	COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.81831 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B037.CNF;1086 BKG DATE : 15-NOV-2009 EFF FILE : W037.CNF;301 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	767.000	740.000	27.000	5.1962	57.44000	2.19E+00	2.01E-01	8.03E-02	3.57E-02	1.63E-01
TH-228	5363.000	44.000	21.095	20.000	4.4721	99.94000	3.46E-02	2.52E-02	3.90E-02	1.71E-02	2.51E-02
TH229	4900.000	20.000	14.000	6.000	2.4495	99.52000	2.25E-02	1.61E-02	2.31E-02	9.16E-03	1.61E-02
TH-230	4625.000	3.000	-2.000	5.000	2.2361	100.0000	-3.20E-03	8.87E-03	2.14E-02	8.32E-03	8.87E-03
TH-232	3972.000	6.000	-2.000	8.000	2.8284	100.0000	-3.20E-03	1.17E-02	2.58E-02	1.05E-02	1.17E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 29-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753012_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :72532 AVERAGE %EFFICIENCY :34.2033 % YIELD : 97.212	COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.77554 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B038.CNF;1083 BKG DATE : 15-NOV-2009 EFF FILE : W038.CNF;315 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	707.000	699.000	8.000	2.8284	57.44000	2.19E+00	2.02E-01	5.06E-02	2.06E-02	1.64E-01
TH-228	5363.000	63.000	42.256	18.000	4.2426	99.94000	7.33E-02	3.03E-02	3.94E-02	1.71E-02	3.01E-02
TH229	4900.000	16.000	8.000	8.000	2.8284	99.52000	1.36E-02	1.64E-02	2.75E-02	1.12E-02	1.63E-02
TH-230	4625.000	65.000	61.000	4.000	2.0000	100.0000	1.03E-01	2.81E-02	2.08E-02	7.88E-03	2.76E-02
TH-232	3972.000	45.000	42.000	3.000	1.7321	100.0000	7.11E-02	2.33E-02	1.87E-02	6.82E-03	2.30E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 29-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753013_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :45-149BB2 AVERAGE %EFFICIENCY :36.4893 % YIELD : 100.508	COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.90354 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B039.CNF;1083 BKG DATE : 15-NOV-2009 EFF FILE : W039.CNF;292 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	789.000	771.000	18.000	4.2426	57.44000	2.19E+00	1.97E-01	6.45E-02	2.80E-02	1.58E-01
TH-228	5363.000	38.000	19.973	15.000	3.8730	99.94000	3.14E-02	2.18E-02	3.30E-02	1.42E-02	2.18E-02
TH229	4900.000	10.000	2.000	8.000	2.8284	99.52000	3.09E-03	1.28E-02	2.49E-02	1.02E-02	1.28E-02
TH-230	4625.000	7.000	3.000	4.000	2.0000	100.0000	4.61E-03	9.98E-03	1.89E-02	7.14E-03	9.98E-03
TH-232	3972.000	2.000	-5.000	7.000	2.6458	100.0000	-7.68E-03	9.03E-03	2.35E-02	9.45E-03	9.03E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 29-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753014_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78773 AVERAGE %EFFICIENCY :32.3093 % YIELD : 95.991	COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.72812 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B040.CNF;1086 BKG DATE : 15-NOV-2009 EFF FILE : W040.CNF;311 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	661.000	652.000	9.000	3.0000	57.44000	2.19E+00	2.07E-01	5.69E-02	2.34E-02	1.70E-01
TH-228	5363.000	21.000	14.440	4.000	2.0000	99.94000	2.68E-02	1.73E-02	2.29E-02	8.65E-03	1.73E-02
TH229	4900.000	14.000	9.000	5.000	2.2361	99.52000	1.64E-02	1.56E-02	2.45E-02	9.49E-03	1.56E-02
TH-230	4625.000	4.000	3.000	1.000	1.0000	100.0000	5.45E-03	7.96E-03	1.39E-02	4.22E-03	7.96E-03
TH-232	3972.000	4.000	1.000	3.000	1.7321	100.0000	1.82E-03	9.42E-03	2.01E-02	7.32E-03	9.41E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 30-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753015_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78205 AVERAGE %EFFICIENCY :33.2779 % YIELD : 104.204	COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.04708 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B041.CNF;1079 BKG DATE : 15-NOV-2009 EFF FILE : W041.CNF;315 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	745.000	729.000	16.000	4.0000	57.44000	2.19E+00	2.01E-01	6.48E-02	2.79E-02	1.62E-01
TH-228	5363.000	17.000	5.138	9.000	3.0000	99.94000	8.53E-03	1.57E-02	2.82E-02	1.16E-02	1.57E-02
TH229	4900.000	4.000	-8.000	12.000	3.4641	99.52000	-1.31E-02	1.28E-02	3.12E-02	1.31E-02	1.28E-02
TH-230	4625.000	3.000	-1.000	4.000	2.0000	100.0000	-1.62E-03	8.42E-03	2.00E-02	7.55E-03	8.42E-03
TH-232	3972.000	3.000	-1.000	4.000	2.0000	100.0000	-1.62E-03	8.42E-03	2.00E-02	7.55E-03	8.42E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 2-NOV-2009 00:00:00. AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753016_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78902 AVERAGE %EFFICIENCY :25.8533 % YIELD : 102.931	COUNT DATE:20-NOV-2009 14:16:26 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.99764 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B201.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W201.CNF;41 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	563.000	561.000	2.000	1.4142	57.44000	2.19E+00	2.17E-01	3.73E-02	1.28E-02	1.82E-01
TH-228	5363.000	10.000	7.899	0.000	0.0000	99.94000	1.70E-02	1.19E-02	6.47E-03	0.00E+00	1.19E-02
TH229	4900.000	12.000	9.000	3.000	1.7321	99.52000	1.91E-02	1.62E-02	2.35E-02	8.57E-03	1.61E-02
TH-230	4625.000	5.000	4.000	1.000	1.0000	100.0000	8.46E-03	1.02E-02	1.62E-02	4.92E-03	1.02E-02
TH-232	3972.000	0.000	-1.000	1.000	1.0000	100.0000	-2.12E-03	5.87E-03	1.62E-02	4.92E-03	5.86E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 2-NOV-2009 00:00:00. AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753017_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78903 AVERAGE %EFFICIENCY :26.7432 % YIELD : 84.429	COUNT DATE:20-NOV-2009 14:16:28 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.27907 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B202.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W202.CNF;41 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	479.000	476.000	3.000	1.7321	57.44000	2.19E+00	2.30E-01	5.08E-02	1.85E-02	1.98E-01
TH-228	5363.000	6.000	2.218	2.000	1.4142	99.94000	5.64E-03	1.24E-02	2.44E-02	8.37E-03	1.24E-02
TH229	4900.000	6.000	6.000	0.000	0.0000	99.52000	1.50E-02	1.21E-02	7.52E-03	0.00E+00	1.20E-02
TH-230	4625.000	3.000	2.000	1.000	1.0000	100.0000	4.99E-03	9.78E-03	1.91E-02	5.80E-03	9.78E-03
TH-232	3972.000	0.000	-1.000	1.000	1.0000	100.0000	-2.49E-03	6.91E-03	1.91E-02	5.80E-03	6.91E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 2-NOV-2009 00:00:00. AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S0239753018_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78905 AVERAGE %EFFICIENCY :25.5537 % YIELD : 90.216	COUNT DATE:20-NOV-2009 14:16:31 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.50380 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B203.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W203.CNF;42 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	487.000	486.000	1.000	1.0000	57.44000	2.19E+00	2.28E-01	3.44E-02	1.05E-02	1.95E-01
TH-228	5363.000	5.000	2.180	1.000	1.0000	99.94000	5.43E-03	9.98E-03	1.91E-02	5.79E-03	9.98E-03
TH229	4900.000	6.000	-2.000	8.000	2.8284	99.52000	-4.91E-03	1.80E-02	3.97E-02	1.61E-02	1.80E-02
TH-230	4625.000	3.000	1.000	2.000	1.4142	100.0000	2.44E-03	1.07E-02	2.34E-02	8.04E-03	1.07E-02
TH-232	3972.000	0.000	-1.000	1.000	1.0000	100.0000	-2.44E-03	6.77E-03	1.87E-02	5.68E-03	6.77E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 18-NOV-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S1201973223_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78907 AVERAGE %EFFICIENCY :25.2254 % YIELD : 97.972	COUNT DATE:20-NOV-2009 14:16:33 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.80503 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B204.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W204.CNF;41 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	524.000	521.000	3.000	1.7321	57.44000	2.19E+00	2.23E-01	4.64E-02	1.69E-02	1.89E-01
TH-228	5363.000	10.000	2.049	6.000	2.4495	99.94000	4.68E-03	1.68E-02	3.29E-02	1.30E-02	1.68E-02
TH229	4900.000	8.000	-2.000	10.000	3.1623	99.52000	-4.58E-03	1.90E-02	4.06E-02	1.68E-02	1.90E-02
TH-230	4625.000	1.000	-3.000	4.000	2.0000	100.0000	-6.84E-03	9.99E-03	2.80E-02	1.06E-02	9.99E-03
TH-232	3972.000	0.000	-9.000	9.000	3.0000	100.0000	-2.05E-02	1.41E-02	3.86E-02	1.59E-02	1.41E-02

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 23-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S1201973224_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78908 AVERAGE %EFFICIENCY :25.3700 % YIELD : 95.543	COUNT DATE:20-NOV-2009 14:16:36 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 3.71072 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B205.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W205.CNF;41 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	513.000	511.000	2.000	1.4142	57.44000	2.19E+00	2.24E-01	4.10E-02	1.41E-02	1.90E-01
TH-228	5363.000	5.000	2.086	1.000	1.0000	99.94000	4.99E-03	9.48E-03	1.83E-02	5.56E-03	9.48E-03
TH229	4900.000	4.000	4.000	0.000	0.0000	99.52000	9.34E-03	9.16E-03	7.00E-03	0.00E+00	9.15E-03
TH-230	4625.000	1.000	1.000	0.000	0.0000	100.0000	2.32E-03	4.55E-03	6.97E-03	0.00E+00	4.55E-03
TH-232	3972.000	1.000	1.000	0.000	0.0000	100.0000	2.32E-03	4.55E-03	6.97E-03	0.00E+00	4.55E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 23-OCT-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00	SAMPLE ID : S1201973225_TH SAMPLE QTY: 0.800 L		
DETECTOR NUMBER :78909 AVERAGE %EFFICIENCY :25.5419 % YIELD : 104.929	COUNT DATE:20-NOV-2009 14:16:37 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4		
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.07524 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B206.CNF;72 BKG DATE : 15-NOV-2009 EFF FILE : W206.CNF;41 CAL DATE : 23-OCT-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	570.000	565.000	5.000	2.2361	57.44000	2.19E+00	2.17E-01	5.19E-02	2.01E-02	1.82E-01
TH-228	5363.000	9.000	0.884	6.000	2.4495	99.94000	1.91E-03	1.52E-02	3.11E-02	1.23E-02	1.52E-02
TH229	4900.000	3.000	0.000	3.000	1.7321	99.52000	0.00E+00	1.01E-02	2.33E-02	8.51E-03	1.01E-02
TH-230	4625.000	1060.000	1059.000	1.000	1.0000	100.0000	2.22E+00	1.80E-01	1.61E-02	4.89E-03	1.34E-01
TH-232	3972.000	1.000	0.000	1.000	1.0000	100.0000	0.00E+00	5.82E-03	1.61E-02	4.89E-03	5.82E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.




GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923093 SAMPLE DATE : 18-NOV-2009 00:00:00 AC-227 SEPARATION : 19-NOV-2009 09:55:00			SAMPLE ID : S1201973226_TH SAMPLE QTY: 0.800 L
DETECTOR NUMBER :78793 AVERAGE %EFFICIENCY :33.4222 % YIELD : 108.024			COUNT DATE:20-NOV-2009 16:05:01 ELAPSED LIVE TIME(SEC): 60000.00 ANALYST :KXM4
MS/MSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	LCS/LCSD ID : A2796-J ISOTOPE : TH-230 PCI/L : 2.675E+00	TRACER ID : 0387-B-102 ISOTOPE : AC-227 NOMINAL : 3.88381 dpm RESULTS : 4.19543 dpm	LIB FILE : ENV_ALPHA_TH.N BKG FILE : B042.CNF;1078 BKG DATE : 15-NOV-2009 EFF FILE : W042.CNF;288 CAL DATE : 4-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
AC-227	5994.040	772.000	759.000	13.000	3.6056	57.44000	2.19E+00	1.97E-01	5.70E-02	2.42E-02	1.58E-01
TH-228	5363.000	13.000	1.020	9.000	3.0000	99.94000	1.60E-03	1.34E-02	2.65E-02	1.09E-02	1.34E-02
TH229	4900.000	10.000	2.000	8.000	2.8284	99.52000	3.13E-03	1.30E-02	2.53E-02	1.03E-02	1.30E-02
TH-230	4625.000	1350.000	1350.000	0.000	0.0000	100.0000	2.11E+00	1.60E-01	4.68E-03	0.00E+00	1.12E-01
TH-232	3972.000	8.000	6.000	2.000	1.4142	100.0000	9.36E-03	9.68E-03	1.49E-02	5.13E-03	9.67E-03

NOTE: Ac-227 results decay corrected to separation date/time.
NOTE: Corrections made to Th-228 net area due to Ra-223 ingrowth from tracer.



URANIUM

Radiochemistry Batch Checklist, Rev 9

Batch# 923094 Product: J Date: 11/28/09

Criteria:	Yes	No	Comments
Sample Solids are less than or equal to 100 mg for GAB.			NA
Samples have been blank corrected (if required)			NA
If activity less 10% MDA/ MDC, error is 150% or less of sample activity. If greater 10% MDA/ MDC, error is 40% or less. If below the MDA/ MDC, error is okay.	/		
Instrument source check is within limits.	/		
Instrument bkg check is within limits.			
Method RDL/ LLD has been met.	/		NCN# 762S66
If duplicate activities are less 5% MDA/ MDC, then RPD is 100% or less. If greater 5% MDA/ MDC, then RPD 20% or less. If below the MDA/ MDC, the RPD is 0%.	/		
Or meets the client's required RER acceptance criteria.			
Tracer yield is 15-125%. Carrier yield 25-125%.	/		
Or meets the client's contract acceptance criteria.			
Method blank is less than the RDL/ LLD. (If rad samples, < 5% of lowest activity)	/		CASE NARRATIVE
Sample was run within hold time.	/		
Sample was correctly preserved if required.	/		
Smears Taken for Radioactive batches.			NA
Method Spike and LCS are within 75-125% or meets the client's contract acceptance criteria.	/		
No blank spaces on data forms.	/		
All line outs initialed and dated.			
No transcription errors are apparent.			
Aux data is correct.			NA
Client Special requirements page has been checked.			
Raw Data and/or spectrum are included and properly statused.	/		
QC data entered into QC database and batch is in REVW	/		
Hit notification complete (if necessary)			NA
Batch entered into Case Narrative.	/		
Batch non-conformances completed, if applicable.	/		NCN# 762S66
Batch non-conformances second reviewed and disposition verified to be completed.	/		NCN# 762S66
Aliquot Correction completed if required.			NA
Review sample historical results if available (If REMP, results above MDC have been verified by historical results, recount or re-analysis.)	/		

GEL Laboratories, LLC

revised 8/1/08

Primary Review Performed By: David J. 11/28/09

Secondary Review Performed By: Lett - dated 11/28/09

11/28-12/1

95 Uranium Que Sheet

16-NOV-09

Batch #:	923094	Analyst:	KXM4	First Client Due Date:	07-DEC-09	Internal Due Date:	26-NOV-09
Tracer Isotope:	U-232	Tracer Code:	1283-E	Expiration Date:	1-15-10	Vol:	0.1mL
LCS Isotope:	U-238	LCS Code:	1163-C	Expiration Date:	4-16-10	Vol:	0.1mL
Spike Isotope:	U-238	Spike Code:	1163-G	Expiration Date:	4-16-10	Vol:	0.1mL
Prep Date:	11-17-09	Pipet ID:	29753058	Balance ID:	16750207		

Sample ID	Client Description	Type	Hazard Code	Min CRDL	Matrix	Client	Collection Date	Pos.	Label #	Aliquot # (e.g. A1)	U Det #
							Yel/Dry				
239753001-1	M-141B	SAMPLE		.03 pCi/L	WATER	KERR003	23-OCT-09	1	1	300.0	134
239753002-1	M-141009B	SAMPLE		.03 pCi/L	WATER	KERR003	23-OCT-09	2	2	300.0	129
239753003-1	PB102309-A3	SAMPLE		.03 pCi/L	WATER	KERR003	23-OCT-09	3	3	300.0	140
239753004-1	M-145B	SAMPLE		.03 pCi/L	WATER	KERR003	26-OCT-09	4	4	800.0	141
239753005-1	M-139B	SAMPLE		.03 pCi/L	WATER	KERR003	26-OCT-09	5	5	800.0	142
239753006-1	M-146B	SAMPLE		.03 pCi/L	WATER	KERR003	27-OCT-09	6	6	300.0	133
239753007-1	M-144B	SAMPLE		.03 pCi/L	WATER	KERR003	27-OCT-09	7	7	800.0	144
239753008-1	M-138B	SAMPLE		.03 pCi/L	WATER	KERR003	28-OCT-09	8	8	800.0	145
239753009-1	M-138009B	SAMPLE		.03 pCi/L	WATER	KERR003	28-OCT-09	9	9	300.0	146
239753010-1	M-138BDISS	SAMPLE		.03 pCi/L	WATER	KERR003	28-OCT-09	10	10	300.0	147
239753011-1	M-138009BDISS	SAMPLE		.03 pCi/L	WATER	KERR003	28-OCT-09	11	11	300.0	148
239753012-1	M-137B	SAMPLE		.03 pCi/L	WATER	KERR003	29-OCT-09	12	12	300.0	149
239753013-1	M-137BDISS	SAMPLE		.03 pCi/L	WATER	KERR003	29-OCT-09	13	13	300.0	150
239753014-1	M-148B	SAMPLE		.03 pCi/L	WATER	KERR003	29-OCT-09	14	14	800.0	151
239753015-1	EB10309-GWA4	SAMPLE		.03 pCi/L	WATER	KERR003	30-OCT-09	15	15	800.0	152
239753016-1	M-147B	SAMPLE		.03 pCi/L	WATER	KERR003	02-NOV-09	16	16	800.0	153
239753017-1	M-147009B	SAMPLE		.03 pCi/L	WATER	KERR003	02-NOV-09	17	17	800.0	154
239753018-1	EB110209-GWA3	SAMPLE		.03 pCi/L	WATER	KERR003	02-NOV-09	18	18	300.0	155
1201973227-1	MB for batch 923094	MB		.03 pCi/L	WATER	QC ACCOUNT		19	19	300.0	121
1201973228-1	M-145B(239753004DUP)	DUP		.03 pCi/L	WATER	QC ACCOUNT	26-OCT-09	20	20	800.0	122
1201973229-1	M-145B(239753004MS)	MS		.03 pCi/L	WATER	QC ACCOUNT	26-OCT-09	21	21	800.0	123
1201973230-1	LCS for batch 923094	LCS		.03 pCi/L	WATER	QC ACCOUNT		22	22	800.0	124

Choose SOP used: GL-RAD-A-011
GL-RAD-A-038
GL-RAD-A-045
GL-RAD-A-043

Solid Sample Dissolution by: TEACH or DIGESTION
KM
11-17-09

N/A

Data Reviewed By: John D. 11/17/09

Circle One

Initials: J.D. 11/17/09

Page 1 of 1

GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 23-OCT-2009 00:00:00

SAMPLE ID : S0239753001_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :65877
AVERAGE %EFFICIENCY :25.4206
% YIELD : 59.440

COUNT DATE:21-NOV-2009 15:52:09
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25236 dpm
RESULTS : 3.12198 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B138.CNF;370
BKG DATE : 15-NOV-2009
EFF FILE : W138.CNF;98
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	20548.000	20527.199	20.000	4.4721	100.0000	7.65E+01	1.10E+01	8.87E-02	3.88E-02	1.05E+00
U232	5302.100	803.000	793.000	10.000	3.1623	100.0000	2.96E+00	4.71E-01	6.61E-02	2.74E-02	2.08E-01
U-235	4391.000	623.000	619.000	4.000	2.0000	80.90000	2.85E+00	4.66E-01	5.67E-02	2.14E-02	2.26E-01
U-238	4184.730	13219.000	13212.000	7.000	2.6458	100.0000	4.92E+01	7.09E+00	5.71E-02	2.29E-02	8.40E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 23-OCT-2009 00:00:00

SAMPLE ID : S0239753002_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :76231
AVERAGE %EFFICIENCY :24.9134
% YIELD : 37.552

COUNT DATE:21-NOV-2009 15:52:12
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25236 dpm
RESULTS : 1.97238 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B139.CNF;367
BKG DATE : 15-NOV-2009
EFF FILE : W139.CNF;98
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	12500.000	12480.505	19.000	4.3589	100.0000	7.51E+01	1.18E+01	1.40E-01	6.10E-02	1.32E+00
U232	5302.100	505.000	491.000	14.000	3.7417	100.0000	2.96E+00	5.35E-01	1.23E-01	5.24E-02	2.69E-01
U-235	4391.000	397.000	396.000	1.000	1.0000	80.90000	2.95E+00	5.45E-01	5.69E-02	1.73E-02	2.91E-01
U-238	4184.730	8431.000	8423.000	8.000	2.8284	100.0000	5.07E+01	8.01E+00	9.73E-02	3.96E-02	1.08E+00

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 23-OCT-2009 00:00:00

SAMPLE ID : S0239753003_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :78771
AVERAGE %EFFICIENCY :25.5118
% YIELD : 70.430

COUNT DATE:21-NOV-2009 15:52:14
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25236 dpm
RESULTS : 3.69925 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B140.CNF;367
BKG DATE : 15-NOV-2009
EFF FILE : W140.CNF;103
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	12.000	-0.952	12.000	3.4641	100.0000	-2.98E-03	2.95E-02	5.99E-02	2.53E-02	2.95E-02
U232	5302.100	950.000	943.000	7.000	2.6458	100.0000	2.96E+00	4.56E-01	4.80E-02	1.93E-02	1.90E-01
U-235	4391.000	2.000	0.000	2.000	1.4142	80.90000	0.00E+00	1.52E-02	3.71E-02	1.27E-02	1.52E-02
U-238	4184.730	3.000	-2.000	5.000	2.2361	100.0000	-6.27E-03	1.74E-02	4.20E-02	1.63E-02	1.74E-02

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 26-OCT-2009 00:00:00

SAMPLE ID : S0239753004_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :76232
AVERAGE %EFFICIENCY :25.5443
% YIELD : 82.649

COUNT DATE:21-NOV-2009 15:52:17
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25193 dpm
RESULTS : 4.34065 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B141.CNF;370
BKG DATE : 15-NOV-2009
EFF FILE : W141.CNF;101
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	6843.000	6835.882	6.000	2.4495	100.0000	1.82E+01	2.60E+00	3.84E-02	1.52E-02	4.33E-01
U232	5302.100	1114.000	1108.000	6.000	2.4495	100.0000	2.96E+00	4.51E-01	3.84E-02	1.52E-02	1.75E-01
U-235	4391.000	216.000	213.000	3.000	1.7321	80.90000	7.02E-01	1.37E-01	3.65E-02	1.33E-02	9.56E-02
U-238	4184.730	4598.000	4591.000	7.000	2.6458	100.0000	1.22E+01	1.76E+00	4.08E-02	1.64E-02	3.55E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 26-OCT-2009 00:00:00

SAMPLE ID : S0239753005_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :64261
AVERAGE %EFFICIENCY :25.8306
% YIELD : 99.436

COUNT DATE:21-NOV-2009 15:52:19
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25193 dpm
RESULTS : 5.22232 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B142.CNF;364
BKG DATE : 15-NOV-2009
EFF FILE : W142.CNF;105
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	2656.000	2633.640	21.000	4.5826	100.0000	5.77E+00	8.14E-01	5.33E-02	2.34E-02	2.22E-01
U232	5302.100	1368.000	1348.000	20.000	4.4721	100.0000	2.96E+00	4.32E-01	5.22E-02	2.28E-02	1.60E-01
U-235	4391.000	78.000	77.000	1.000	1.0000	80.90000	2.09E-01	5.50E-02	2.07E-02	6.30E-03	4.72E-02
U-238	4184.730	1742.000	1739.000	3.000	1.7321	100.0000	3.81E+00	5.47E-01	2.42E-02	8.83E-03	1.79E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 27-OCT-2009 00:00:00

SAMPLE ID : S0239753006_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :65882
AVERAGE %EFFICIENCY :24.3577
% YIELD : 86.284

COUNT DATE:21-NOV-2009 15:52:22
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25179 dpm
RESULTS : 4.53145 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B143.CNF;367
BKG DATE : 15-NOV-2009
EFF FILE : W143.CNF;108
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	4202.000	4188.887	12.000	3.4641	100.0000	1.12E+01	1.61E+00	5.12E-02	2.16E-02	3.41E-01
U232	5302.100	1113.000	1103.000	10.000	3.1623	100.0000	2.96E+00	4.52E-01	4.75E-02	1.97E-02	1.76E-01
U-235	4391.000	137.000	134.000	3.000	1.7321	80.90000	4.44E-01	9.90E-02	3.66E-02	1.33E-02	7.68E-02
U-238	4184.730	2850.000	2845.000	5.000	2.2361	100.0000	7.62E+00	1.11E+00	3.59E-02	1.39E-02	2.81E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 27-OCT-2009 00:00:00

SAMPLE ID : S0239753007_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75551
AVERAGE %EFFICIENCY :25.0940
% YIELD : 82.993

COUNT DATE:21-NOV-2009 15:52:24
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25180 dpm
RESULTS : 4.35860 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B144.CNF;366
BKG DATE : 15-NOV-2009
EFF FILE : W144.CNF;102
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	7729.000	7720.897	7.000	2.6458	100.0000	2.09E+01	2.92E+00	4.14E-02	1.66E-02	4.66E-01
U232	5302.100	1103.000	1093.000	10.000	3.1623	100.0000	2.96E+00	4.45E-01	4.79E-02	1.99E-02	1.77E-01
U-235	4391.000	284.000	278.000	6.000	2.4495	80.90000	9.29E-01	1.70E-01	4.81E-02	1.90E-02	1.12E-01
U-238	4184.730	5363.000	5360.000	3.000	1.7321	100.0000	1.45E+01	2.04E+00	2.99E-02	1.09E-02	3.88E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 28-OCT-2009 00:00:00

SAMPLE ID : S0239753008_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :72526
AVERAGE %EFFICIENCY :25.0221
% YIELD : 74.702

COUNT DATE:21-NOV-2009 15:52:27
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25166 dpm
RESULTS : 3.92312 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B145.CNF;364
BKG DATE : 15-NOV-2009
EFF FILE : W145.CNF;107
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	13587.000	13580.010	6.000	2.4495	100.0000	4.09E+01	5.85E+00	4.34E-02	1.72E-02	6.88E-01
U232	5302.100	986.000	981.000	5.000	2.2361	100.0000	2.96E+00	4.59E-01	4.04E-02	1.57E-02	1.86E-01
U-235	4391.000	323.000	322.000	1.000	1.0000	80.90000	1.20E+00	2.15E-01	2.85E-02	8.66E-03	1.31E-01
U-238	4184.730	6888.000	6887.000	1.000	1.0000	100.0000	2.07E+01	2.99E+00	2.31E-02	7.01E-03	4.90E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 28-OCT-2009 00:00:00

SAMPLE ID : S0239753009_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :72527
AVERAGE %EFFICIENCY :25.2174
% YIELD : 56.669

COUNT DATE:21-NOV-2009 15:52:29
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25167 dpm
RESULTS : 2.97609 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B146.CNF;369
BKG DATE : 15-NOV-2009
EFF FILE : W146.CNF;109
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	10571.000	10567.243	3.000	1.7321	100.0000	4.16E+01	6.04E+00	4.36E-02	1.59E-02	7.94E-01
U232	5302.100	759.000	750.000	9.000	3.0000	100.0000	2.96E+00	4.76E-01	6.69E-02	2.75E-02	2.14E-01
U-235	4391.000	316.000	314.000	2.000	1.4142	80.90000	1.53E+00	2.78E-01	4.67E-02	1.60E-02	1.70E-01
U-238	4184.730	5301.000	5299.000	2.000	1.4142	100.0000	2.09E+01	3.06E+00	3.77E-02	1.30E-02	5.62E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 28-OCT-2009 00:00:00

SAMPLE ID : S0239753010_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75550
AVERAGE %EFFICIENCY :24.5119
% YIELD : 73.148

COUNT DATE:21-NOV-2009 15:52:32
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25166 dpm
RESULTS : 3.84149 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B147.CNF;369
BKG DATE : 15-NOV-2009
EFF FILE : W147.CNF;108
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	12452.000	12443.051	8.000	2.8284	100.0000	3.91E+01	5.62E+00	5.07E-02	2.07E-02	6.87E-01
U232	5302.100	951.000	941.000	10.000	3.1623	100.0000	2.96E+00	4.63E-01	5.57E-02	2.31E-02	1.91E-01
U-235	4391.000	265.000	265.000	0.000	0.0000	80.90000	1.03E+00	1.92E-01	1.16E-02	0.00E+00	1.24E-01
U-238	4184.730	6205.000	6197.000	8.000	2.8284	100.0000	1.95E+01	2.82E+00	5.07E-02	2.07E-02	4.85E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 28-OCT-2009 00:00:00

SAMPLE ID : S0239753011_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :74429
AVERAGE %EFFICIENCY :24.8842
% YIELD : 57.275

COUNT DATE:21-NOV-2009 15:52:34
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25166 dpm
RESULTS : 3.00790 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B148.CNF;368
BKG DATE : 15-NOV-2009
EFF FILE : W148.CNF;123
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	9760.000	9748.245	11.000	3.3166	100.0000	3.85E+01	5.59E+00	7.28E-02	3.05E-02	7.65E-01
U232	5302.100	753.000	748.000	5.000	2.2361	100.0000	2.96E+00	4.76E-01	5.30E-02	2.06E-02	2.13E-01
U-235	4391.000	227.000	225.000	2.000	1.4142	80.90000	1.10E+00	2.14E-01	4.68E-02	1.61E-02	1.45E-01
U-238	4184.730	4853.000	4852.000	1.000	1.0000	100.0000	1.92E+01	2.81E+00	3.02E-02	9.19E-03	5.39E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 29-OCT-2009 00:00:00

SAMPLE ID : S0239753012_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :33449
AVERAGE %EFFICIENCY :24.8326
% YIELD : 77.728

COUNT DATE:21-NOV-2009 15:52:37
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25152 dpm
RESULTS : 4.08191 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B149.CNF;372
BKG DATE : 15-NOV-2009
EFF FILE : W149.CNF;108
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	11662.000	11656.978	4.000	2.0000	100.0000	3.40E+01	4.77E+00	3.59E-02	1.36E-02	6.18E-01
U232	5302.100	1024.000	1013.000	11.000	3.3166	100.0000	2.96E+00	4.51E-01	5.38E-02	2.25E-02	1.84E-01
U-235	4391.000	280.000	280.000	0.000	0.0000	80.90000	1.01E+00	1.84E-01	1.08E-02	0.00E+00	1.18E-01
U-238	4184.730	6505.000	6502.000	3.000	1.7321	100.0000	1.90E+01	2.68E+00	3.23E-02	1.18E-02	4.61E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 29-OCT-2009 00:00:00

SAMPLE ID : S0239753013_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75552
AVERAGE %EFFICIENCY :24.7216
% YIELD : 70.986

COUNT DATE:21-NOV-2009 15:52:38
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25152 dpm
RESULTS : 3.72784 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B150.CNF;373
BKG DATE : 15-NOV-2009
EFF FILE : W150.CNF;116
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	11562.000	11560.070	1.000	1.0000	100.0000	3.71E+01	5.26E+00	2.46E-02	7.46E-03	6.76E-01
U232	5302.100	934.000	921.000	13.000	3.6056	100.0000	2.96E+00	4.59E-01	6.35E-02	2.69E-02	1.94E-01
U-235	4391.000	295.000	293.000	2.000	1.4142	80.90000	1.16E+00	2.11E-01	3.80E-02	1.30E-02	1.34E-01
U-238	4184.730	6210.000	6204.000	6.000	2.4495	100.0000	1.99E+01	2.84E+00	4.62E-02	1.83E-02	4.96E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 29-OCT-2009 00:00:00

SAMPLE ID : S0239753014_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75556
AVERAGE %EFFICIENCY :24.5337
% YIELD : 80.849

COUNT DATE:21-NOV-2009 15:52:40
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25152 dpm
RESULTS : 4.24582 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B151.CNF;368
BKG DATE : 15-NOV-2009
EFF FILE : W151.CNF;114
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	11370.000	11368.949	0.000	0.0000	100.0000	3.23E+01	4.51E+00	8.52E-03	0.00E+00	5.93E-01
U232	5302.100	1047.000	1041.000	6.000	2.4495	100.0000	2.96E+00	4.48E-01	4.09E-02	1.62E-02	1.81E-01
U-235	4391.000	339.000	338.000	1.000	1.0000	80.90000	1.19E+00	2.08E-01	2.69E-02	8.16E-03	1.27E-01
U-238	4184.730	7732.000	7731.000	1.000	1.0000	100.0000	2.19E+01	3.08E+00	2.17E-02	6.60E-03	4.89E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 30-OCT-2009 00:00:00

SAMPLE ID : S0239753015_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :76222
AVERAGE %EFFICIENCY :24.4246
% YIELD : 91.742

COUNT DATE:21-NOV-2009 15:52:43
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25138 dpm
RESULTS : 4.81774 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B152.CNF;365
BKG DATE : 15-NOV-2009
EFF FILE : W152.CNF;101
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	13.000	10.813	1.000	1.0000	100.0000	2.72E-02	1.80E-02	1.92E-02	5.85E-03	1.76E-02
U232	5302.100	1179.000	1176.000	3.000	1.7321	100.0000	2.96E+00	4.39E-01	2.78E-02	1.01E-02	1.69E-01
U-235	4391.000	2.000	0.000	2.000	1.4142	80.90000	0.00E+00	1.22E-02	2.98E-02	1.02E-02	1.22E-02
U-238	4184.730	8.000	7.000	1.000	1.0000	100.0000	1.76E-02	1.50E-02	1.92E-02	5.85E-03	1.48E-02

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 2-NOV-2009 00:00:00.

SAMPLE ID : S0239753016_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :76223
AVERAGE %EFFICIENCY :25.2320
% YIELD : 80.726

COUNT DATE:21-NOV-2009 15:52:46
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25097 dpm
RESULTS : 4.23892 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B153.CNF;360
BKG DATE : 15-NOV-2009
EFF FILE : W153.CNF;104
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	9460.000	9452.921	6.000	2.4495	100.0000	2.61E+01	3.66E+00	3.98E-02	1.58E-02	5.27E-01
U232	5302.100	1088.000	1069.000	19.000	4.3589	100.0000	2.96E+00	4.48E-01	6.44E-02	2.80E-02	1.80E-01
U-235	4391.000	319.000	318.000	1.000	1.0000	80.90000	1.09E+00	1.92E-01	2.61E-02	7.95E-03	1.20E-01
U-238	4184.730	7090.000	7088.000	2.000	1.4142	100.0000	1.96E+01	2.75E+00	2.65E-02	9.09E-03	4.56E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 2-NOV-2009 00:00:00.

SAMPLE ID : S0239753017_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :76224
AVERAGE %EFFICIENCY :25.4981
% YIELD : 67.255

COUNT DATE:21-NOV-2009 15:52:49
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25096 dpm
RESULTS : 3.53154 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B154.CNF;362
BKG DATE : 15-NOV-2009
EFF FILE : W154.CNF;102
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	7989.000	7983.092	5.000	2.2361	100.0000	2.62E+01	3.73E+00	4.40E-02	1.71E-02	5.75E-01
U232	5302.100	908.000	900.000	8.000	2.8284	100.0000	2.96E+00	4.60E-01	5.31E-02	2.16E-02	1.95E-01
U-235	4391.000	264.000	263.000	1.000	1.0000	80.90000	1.07E+00	1.98E-01	3.11E-02	9.44E-03	1.29E-01
U-238	4184.730	5971.000	5970.000	1.000	1.0000	100.0000	1.96E+01	2.80E+00	2.51E-02	7.64E-03	4.97E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 2-NOV-2009 00:00:00.

SAMPLE ID : S0239753018_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75553
AVERAGE %EFFICIENCY :25.8503
% YIELD : 87.199

COUNT DATE:21-NOV-2009 15:52:52
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25096 dpm
RESULTS : 4.57877 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B155.CNF;369
BKG DATE : 15-NOV-2009
EFF FILE : W155.CNF;111
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	31.000	26.806	3.000	1.7321	100.0000	6.70E-02	2.95E-02	2.76E-02	1.01E-02	2.80E-02
U232	5302.100	1192.000	1183.000	9.000	3.0000	100.0000	2.96E+00	4.39E-01	4.24E-02	1.74E-02	1.70E-01
U-235	4391.000	1.000	1.000	0.000	0.0000	80.90000	3.09E-03	6.07E-03	9.26E-03	0.00E+00	6.05E-03
U-238	4184.730	18.000	14.000	4.000	2.0000	100.0000	3.50E-02	2.35E-02	3.07E-02	1.16E-02	2.30E-02

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 17-NOV-2009 00:00:00

SAMPLE ID : S1201973227_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75545
AVERAGE %EFFICIENCY :24.7357
% YIELD : 94.360

COUNT DATE:20-NOV-2009 14:25:28
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.24889 dpm
RESULTS : 4.95286 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B121.CNF;412
BKG DATE : 15-NOV-2009
EFF FILE : W121.CNF;113
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	8.000	5.764	1.000	1.0000	100.0000	1.39E-02	1.33E-02	1.85E-02	5.61E-03	1.32E-02
U232	5302.100	1233.000	1225.000	8.000	2.8284	100.0000	2.96E+00	4.36E-01	3.90E-02	1.59E-02	1.67E-01
U-235	4391.000	3.000	3.000	0.000	0.0000	80.90000	8.95E-03	1.02E-02	8.95E-03	0.00E+00	1.01E-02
U-238	4184.730	4.000	0.000	4.000	2.0000	100.0000	0.00E+00	1.34E-02	2.97E-02	1.12E-02	1.34E-02

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 26-OCT-2009 00:00:00

SAMPLE ID : S1201973228_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :75546
AVERAGE %EFFICIENCY :25.2661
% YIELD : 89.740

COUNT DATE:20-NOV-2009 14:25:29
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25193 dpm
RESULTS : 4.71309 dpm

LIB FILE : ENV_ALPHA_UU.N
BKG FILE : B122.CNF;414
BKG DATE : 15-NOV-2009
EFF FILE : W122.CNF;116
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	7510.000	7496.799	12.000	3.4641	100.0000	1.86E+01	2.58E+00	4.75E-02	2.00E-02	4.22E-01
U232	5302.100	1200.000	1190.000	10.000	3.1623	100.0000	2.96E+00	4.39E-01	4.40E-02	1.83E-02	1.69E-01
U-235	4391.000	210.000	207.000	3.000	1.7321	80.90000	6.35E-01	1.24E-01	3.39E-02	1.24E-02	8.78E-02
U-238	4184.730	5060.000	5056.000	4.000	2.0000	100.0000	1.26E+01	1.75E+00	3.06E-02	1.16E-02	3.46E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 26-OCT-2009 00:00:00

SAMPLE ID : S1201973229_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :45-142V3
AVERAGE %EFFICIENCY :26.0328
% YIELD : 79.924

COUNT DATE:20-NOV-2009 14:25:33
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.25193 dpm
RESULTS : 4.19758 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B123.CNF;412
BKG DATE : 15-NOV-2009
EFF FILE : W123.CNF;112
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	7913.000	7910.898	1.000	1.0000	100.0000	2.14E+01	2.99E+00	2.07E-02	6.30E-03	4.72E-01
U232	5302.100	1105.000	1092.000	13.000	3.6056	100.0000	2.96E+00	4.45E-01	5.36E-02	2.27E-02	1.77E-01
U-235	4391.000	212.000	210.000	2.000	1.4142	80.90000	7.02E-01	1.36E-01	3.20E-02	1.10E-02	9.59E-02
U-238	4184.730	5670.000	5669.000	1.000	1.0000	100.0000	1.53E+01	2.15E+00	2.07E-02	6.30E-03	3.99E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



GEL Laboratories LLC
ALPHA SPECTROSCOPY REPORT

BATCH NUMBER: 923094
SAMPLE DATE : 17-NOV-2009 00:00:00

SAMPLE ID : S1201973230_UU
SAMPLE QTY: 0.800 L

DETECTOR NUMBER :45-142V2
AVERAGE %EFFICIENCY :25.9543
% YIELD : 93.527

COUNT DATE:20-NOV-2009 14:25:35
ELAPSED LIVE TIME(SEC): 60000.00
ANALYST :KXM4

MS/MSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00

LCS/LCSD
ID : 1163-G
ISOTOPE : U-238
PCI/L : 3.149E+00


TRACER
ID : 1283-E
ISOTOPE : U232
NOMINAL : 5.24889 dpm
RESULTS : 4.90913 dpm

LIB FILE : ENV_ALPHA_UUN
BKG FILE : B124.CNF;408
BKG DATE : 15-NOV-2009
EFF FILE : W124.CNF;108
CAL DATE : 18-NOV-2009

NUCLIDE ACTIVITY SUMMARY

NUCLIDE	ENERGY	GROSS AREA	NET AREA	BKG AREA	BKG Sg	%ABUN	ACTIVITY pCi/L	TPU 1.96-SIGMA	MDA pCi/L	Lc pCi/L	UNC pCi/L
U-3/4	4763.020	1350.000	1343.714	5.000	2.2361	100.0000	3.12E+00	4.56E-01	3.11E-02	1.21E-02	1.67E-01
U232	5302.100	1282.000	1274.000	8.000	2.8284	100.0000	2.96E+00	4.34E-01	3.75E-02	1.53E-02	1.63E-01
U-235	4391.000	62.000	62.000	0.000	0.0000	80.90000	1.78E-01	5.04E-02	8.60E-03	0.00E+00	4.43E-02
U-238	4184.730	1404.000	1401.000	3.000	1.7321	100.0000	3.25E+00	4.74E-01	2.57E-02	9.35E-03	1.71E-01

NOTE: Corrections made to U-3/4 net area due to tracer impurity



RADIUM 228

Radiochemistry Batch Checklist, Rev 9

Batch# 922859 Product: RA-228 Date: 11.25

Criteria:	Yes	No	Comments
Sample Solids are less than or equal to 100 mg for GAB.			N/A
Samples have been blank corrected (if required)			N/A
If activity less 10* MDA/ MDC, error is 150% or less of sample activity. If greater 10* MDA/ MDC, error is 40% or less. If below the MDA/ MDC, error is okay.	✓		
Instrument source check is within limits.	✓		
Instrument bkg check is within limits.			
Method RDL/ LLD has been met.	✓		
If duplicate activities are less 5* MDA/ MDC, then RPD is 100% or less. If greater 5* MDA/ MDC, then RPD 20% or less. If below the MDA/ MDC, the RPD is 0%. Or meets the client's required RER acceptance criteria.	✓		
Tracer yield is 15-125%. Carrier yield 25-125%. Or meets the client's contract acceptance criteria.	✓		
Method blank is less than the RDL/ LLD. (If rad samples, < 5% of lowest activity)	✓		
Sample was run within hold time.	✓		
Sample was correctly preserved if required.	✓		
Smears Taken for Radioactive batches.			N/A
Method Spike and LCS are within 75-125% or meets the client's contract acceptance criteria.	✓		
No blank spaces on data forms.			
All line outs initialed and dated.	✓		
No transcription errors are apparent.			
Aux data is correct.			N/A
Client Special requirements page has been checked.	✓		
Raw Data and/ or spectrum are included and properly statused.	✓		
QC data entered into QC database and batch is in REVW	✓		
Hit notification complete (if necessary)			N/A
Batch entered into Case Narrative.	✓		
Batch non-conformances completed, if applicable.			N/A
Batch non-conformances second reviewed and disposition verified to be completed.			N/A
Aliquot Correction completed if required.			N/A
Review sample historical results if available (If REMP, results above MDC have been verified by historical results, recount or re-analysis.)	✓		

GEL Laboratories, LLC

revised 8/1/08

Primary Review Performed By: Rhyean Brantley

Secondary Review Performed By: Melissa H. Hill

KERR

Radium-228 Que Sheet

Batch #: 922859
of 600
Tracer Isotope: Barium-133
Prep Date: 11-18-09
Initials: HS

Analyst: JXC5
Spike Code: 0503-6
LCS Code: 0503-9
Tracer Code: 0112-2
Pipet ID: 1795419
First Client Due Date: 12/07/2009
Expiration Date: 9-11-10
Expiration Date: 9-11-10
Expiration Date: 2-17-10
Balance ID: 2120-4803
Internal Due Date 11/26/2009
Ac-228 Ingrow: 11-19-09 / 1320
Vol: 0.1 mL
Vol: 0.1 mL
Vol: 0.1 mL
Witness: SFH-12-09

Sample ID	Client Description	Type	Hazard Code	Min CRDL	Matrix	Client	Collect Date & Time
239753001-1	M-141B	SAMPLE	3 pCi/L	WATER	KERR003		23-OCT-09 10:00 AM
- 239753002-1	M-141009B	SAMPLE	3 pCi/L	WATER	KERR003		23-OCT-09 10:00 AM
239753003-1	PB102309-A3	SAMPLE	3 pCi/L	WATER	KERR003		23-OCT-09 12:15 PM
239753004-1	M-145B	SAMPLE	3 pCi/L	WATER	KERR003		26-OCT-09 10:15 AM
239753005-1	M-139B	SAMPLE	3 pCi/L	WATER	KERR003		26-OCT-09 12:55 PM
239753006-1	M-146B	SAMPLE	3 pCi/L	WATER	KERR003		27-OCT-09 09:30 AM
239753007-1	M-144B	SAMPLE	3 pCi/L	WATER	KERR003		27-OCT-09 12:25 PM
239753008-1	M-138B	SAMPLE	3 pCi/L	WATER	KERR003		28-OCT-09 11:15 AM
239753009-1	M-138009B	SAMPLE	3 pCi/L	WATER	KERR003		28-OCT-09 11:15 AM
239753010-1	M-138BDISS	SAMPLE	3 pCi/L	WATER	KERR003		28-OCT-09 11:15 AM
239753011-1	M-138009BDISS	SAMPLE	3 pCi/L	WATER	KERR003		28-OCT-09 11:15 AM
239753012-1	M-137B	SAMPLE	3 pCi/L	WATER	KERR003		29-OCT-09 01:30 PM
239753013-1	M-137BDISS	SAMPLE	3 pCi/L	WATER	KERR003		29-OCT-09 01:30 PM
239753014-1	M-148B	SAMPLE	3 pCi/L	WATER	KERR003		29-OCT-09 09:10 AM
239753015-1	EB103009-GWA4	SAMPLE	3 pCi/L	WATER	KERR003		30-OCT-09 11:10 AM
239753016-1	M-147B	SAMPLE	3 pCi/L	WATER	KERR003	02-NOV-09 10:00 AM	16
239753017-1	M-147009B	SAMPLE	3 pCi/L	WATER	KERR003	02-NOV-09 10:00 AM	17
239753018-1	EB110209-GWA3	SAMPLE	3 pCi/L	WATER	KERR003	02-NOV-09 12:40 PM	18
1201972468-1	MB for batch 922859	MB	3 pCi/L	WATER	QC ACCOUNT	23-OCT-09 10:00 AM	19
1201972469-1	M-141B(239753001DUP)	DUP	3 pCi/L	WATER	QC ACCOUNT	23-OCT-09 10:00 AM	20
1201972470-1	M-141B(239753001MS)	MS	3 pCi/L	WATER	QC ACCOUNT	23-OCT-09 10:00 AM	21
1201972471-1	LCS for batch 922859	LCS	3 pCi/L	WATER	QC ACCOUNT		22

Pos. #	Vol (mL)	Ba Yield (%)	Gamma Det. #
1	200	14	75.86
2	200	13A	90.77
3	200	8A	87.36
4	200	1C	86.63
5	200	2A	86.41
6	200	7C	71.99
7	200	6B	88.11
8	200	5C	86.60
9	200	9D	94.88
10	200	7D	77.31
11	200	1A	81.38
12	200	9C	94.71
13	200	10D	81.12
14	200	1D	84.24
15	200	8C	85.91
16	200	7A	75.30
17	200	10C	79.64
18	200	12A	79.80
19	200	6D	88.88
20	200	9A	87.82
21	100	9B	88.74
22	200	3A	95.24

11-18-09
V.V.S.

Data Reviewed By: Ryan Brantly
Page 1 of 1

Radium-228 Liquid

Filename : RA228.xls
 File type : Excel
 Version # : 1.2.6
Batch : 922859
Analyst : JXC5
Prep Date : 11/16/2009
Ra-228 Abundance : 1
Ra-228 Method Uncertainty : 0.1268
Geometry: CaF₂ on 25mm Filter

Spike S/N : 0503-B
Spike Exp Date : 9/11/2010
Spike Activity (dpm/ml): 174.18
Spike Volume Added: 0.10
LCS S/N : 0503-B
LCS Exp Date : 9/11/2010
LCS Activity (dpm/ml): 174.18
LCS Volume Added: 0.10
Tracer S/N : 0112-J
Tracer Exp Date : 2/17/2010
Tracer Volume Added: 0.10

Pipet, 0.1 ml StdDev : +/-
 Pipet, 0.5 ml StdDev : +/-
 Pipet, 1 ml StdDev : +/-
Procedure Code : GFC28RAL
Paramname : Radium-228
Required MDA : 3 pCi/L
Halflife of Ra-228 : 5.75 years
Halflife of Ac-228 : 6.15 hours

				Tracer Calculations			Tracer Samp.		
				Tracer Concentration (cpm) (Ba-133 Ref.)	Tracer Ref. Count	Concentration (cpm) (Ba-133 Samp.)	Tracer Count	Tracer Aliquot (mL)	Tracer Aliquot StdDev. (mL)
Pos.	Sample ID	Sample Aliquot L	Sample Aliquot StdDev. L	Sample Date/Time	Uncertainty (cpm)	Concentration (cpm)	Uncertainty (cpm)	Concentration (cpm)	Uncertainty (cpm)
1	239753001.1	0.2000	1.6007E-05	10/23/2009 10:00	304.5	3.52%	231.0	4.11%	0.1
2	239753002.1	0.2000	1.6007E-05	10/23/2009 10:00	304.5	3.52%	276.4	3.71%	0.1
3	239753003.1	0.2000	1.6007E-05	10/23/2009 12:15	304.5	3.52%	266.0	3.79%	0.1
4	239753004.1	0.2000	1.6007E-05	10/26/2009 10:15	304.5	3.52%	263.8	3.81%	0.1
5	239753005.1	0.2000	1.6007E-05	10/26/2009 12:55	304.5	3.52%	269.2	3.77%	0.1
6	239753006.1	0.2000	1.6007E-05	10/27/2009 9:30	304.5	3.52%	219.2	4.24%	0.1
7	239753007.1	0.2000	1.6007E-05	10/27/2009 12:25	304.5	3.52%	288.3	3.78%	0.1
8	239753008.1	0.2000	1.6007E-05	10/28/2009 11:15	304.5	3.52%	263.7	3.81%	0.1
9	239753009.1	0.2000	1.6007E-05	10/28/2009 11:15	304.5	3.52%	288.9	3.62%	0.1
10	239753010.1	0.2000	1.6007E-05	10/28/2009 11:15	304.5	3.52%	235.4	4.07%	0.1
11	239753011.1	0.2000	1.6007E-05	10/28/2009 11:15	304.5	3.52%	247.8	3.95%	0.1
12	239753012.1	0.2000	1.6007E-05	10/29/2009 13:30	304.5	3.52%	288.4	3.63%	0.1
13	239753013.1	0.2000	1.6007E-05	10/29/2009 13:30	304.5	3.52%	247.0	3.96%	0.1
14	239753014.1	0.2000	1.6007E-05	10/29/2009 9:10	304.5	3.52%	266.5	3.87%	0.1
15	239753015.1	0.2000	1.6007E-05	10/30/2009 11:10	304.5	3.52%	261.6	3.83%	0.1
16	239753016.1	0.2000	1.6007E-05	11/2/2009 10:00	304.5	3.52%	229.3	4.13%	0.1
17	239753017.1	0.2000	1.6007E-05	11/2/2009 10:00	304.5	3.52%	242.5	4.00%	0.1
18	239753018.1	0.2000	1.6007E-05	11/2/2009 12:40	304.5	3.52%	243.0	3.98%	0.1
19	1201972468.1	0.2000	1.6007E-05	11/18/2009 09:00	304.5	3.52%	288.2	3.78%	0.1
20	1201972469.1	0.2000	1.6007E-05	10/23/2009 10:00	304.5	3.52%	267.4	3.78%	0.1
21	1201972470.1	0.1000	1.1370E-05	10/23/2009 10:00	304.5	3.52%	270.2	3.76%	0.1
22	1201972471.1	0.2000	1.6007E-05	11/18/2009 09:00	304.5	3.52%	260.0	3.61%	0.1

Pos.	Detector ID	Counting Time (min.)	Gross Counts	Alpha Beta	Beta cpm	Count Start Date/Time	Separation Date/Time	Ra-228 Decay	Ac-228 Decay	Ac-228 Count Correction	Calculated Sample Recovery %	Sample Recovery Error %	Calibration Data			Detector Efficiency (cpm/dpm)	Detector Efficiency Error (%)	Weekly Bkg Count (min.)
													Counted on	Calibration Date	Calibration Due Date			
1	1A	60	6	55	0.917	11/23/2009 17:26	11/23/2009 15:10	0.990	0.774	1.057	75.88%	2.88%	PIC	7/2/2009	7/31/2010	0.6303	0.00600	0.5448
2	13A	60	12	70	1.167	11/23/2009 17:23	11/23/2009 15:10	0.990	0.778	1.057	90.77%	2.74%	PIC	7/2/2009	7/31/2010	0.6410	0.00816	0.7148
3	8A	75	8	116	1.547	11/23/2009 17:26	11/23/2009 15:10	0.990	0.774	1.072	87.38%	2.77%	PIC	7/2/2009	7/31/2010	0.6247	0.00816	1.2820
4	1C	85	9	156	1.885	11/23/2009 17:26	11/23/2009 15:10	0.991	0.774	1.082	86.63%	2.78%	PIC	7/2/2009	7/31/2010	0.6176	0.00344	1.322
5	2A	60	10	44	0.735	11/23/2009 17:26	11/23/2009 15:10	0.991	0.773	1.057	88.41%	2.76%	PIC	7/2/2009	7/31/2010	0.6172	0.00349	0.732
6	7C	60	3	33	0.560	11/23/2009 17:26	11/23/2009 15:10	0.991	0.773	1.057	71.99%	2.98%	PIC	7/2/2009	7/31/2010	0.6178	0.00816	0.494
7	6B	75	22	82	1.089	11/23/2009 17:26	11/23/2009 15:10	0.991	0.773	1.072	88.11%	2.77%	PIC	7/2/2009	7/31/2010	0.6163	0.00816	1.2026
8	5C	75	11	104	1.387	11/23/2009 17:27	11/23/2009 15:10	0.991	0.773	1.072	86.60%	2.78%	PIC	7/2/2009	7/31/2010	0.6268	0.00816	1.184
9	9D	85	3	137	1.612	11/23/2009 17:27	11/23/2009 15:10	0.991	0.773	1.082	94.88%	2.71%	PIC	7/2/2009	7/31/2010	0.6433	0.00816	1.730
10	7D	60	4	41	0.688	11/23/2009 17:27	11/23/2009 15:10	0.991	0.773	1.057	77.31%	2.87%	PIC	7/2/2009	7/31/2010	0.6257	0.00816	0.622
11	14A	60	9	67	1.117	11/23/2009 17:23	11/23/2009 15:10	0.991	0.778	1.057	81.38%	2.83%	PIC	7/2/2009	7/31/2010	0.6393	0.00816	0.748
12	12C	60	6	82	1.367	11/23/2009 17:23	11/23/2009 15:10	0.992	0.778	1.057	94.71%	2.72%	PIC	7/2/2009	7/31/2010	0.6304	0.00816	1.246
13	10D	60	20	77	1.283	11/23/2009 17:27	11/23/2009 15:10	0.992	0.773	1.057	81.12%	2.83%	PIC	7/2/2009	7/31/2010	0.6320	0.00816	0.876
14	1D	60	4	54	0.900	11/23/2009 17:27	11/23/2009 15:10	0.992	0.772	1.057	84.24%	2.80%	PIC	7/2/2009	7/31/2010	0.6043	0.00511	0.646
15	8C	60	4	29	0.483	11/23/2009 17:27	11/23/2009 15:10	0.992	0.772	1.057	85.91%	2.78%	PIC	7/2/2009	7/31/2010	0.6339	0.00816	0.744
16	7A	60	4	55	0.917	11/23/2009 17:27	11/23/2009 15:10	0.993	0.772	1.057	75.30%	2.88%	PIC	7/2/2009	7/31/2010	0.6180	0.00816	0.714
17	10C	60	4	39	0.650	11/23/2009 17:27	11/23/2009 15:10	0.993	0.772	1.057	79.64%	2.84%	PIC	7/2/2009	7/31/2010	0.6250	0.00816	0.498
18	12D	60	5	54	0.900	11/23/2009 17:23	11/23/2009 15:10	0.993	0.778	1.057	79.80%	2.84%	PIC	7/2/2009	7/31/2010	0.6320	0.00816	0.714
19	6D	75	7	101	1.347	11/23/2009 17:27	11/23/2009 15:10	0.998	0.772	1.072	88.08%	2.77%	PIC	7/2/2009	7/31/2010	0.6120	0.00816	1.196
20	9A	60	8	46	0.767	11/23/2009 17:26	11/23/2009 15:10	0.990	0.774	1.057	87.82%	2.77%	PIC	7/2/2009	7/31/2010	0.6496	0.00816	0.628
21	9B	60	5	617	10.283	11/23/2009 17:26	11/23/2009 15:10	0.990	0.774	1.057	88.74%	2.76%	PIC	7/2/2009	7/31/2010	0.6356	0.00816	1.806
22	3A	60	83	8.717	8.717	11/23/2009 17:26	11/23/2009 15:10	0.998	0.774	1.057	95.24%	2.71%	PIC	7/2/2009	7/31/2010	0.5682	0.00943	1.386

Notes:

- 1 - Results are decay corrected to Sample Date/Time
 - 2 - Reference date for Spike Activity (dpm/m) is the batch Prep Date
 - 3 - Spike Nominals are decay corrected to Sample Date/Time
- * - RPD changed to 0% due to activity below MDA for 1201972469.1

Results Pos.	Decision Level pCi/L	Critical Required MDA pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate CPM	Net Count Rate CPM	2 SIGMA		Sample QC	Sample Type	RPD FER	Nominal pCi/L	Recovery
								Total Propri. pCi/L	Counting Uncertainty pCi/L					
1	1.5325	1.0820	3	2.4891	2.3975	0.3483	0.3687	0.1445	1.6310	1.7419	SAMPLE			
2	1.4340	1.0124	3	2.2907	2.3852	0.3233	0.4487	0.1523	1.6237	1.6237	SAMPLE			
3	1.8860	1.3315	3	2.8841	1.5407	0.5717	0.2967	1.7242	1.7684	1.7684	SAMPLE			
4	1.8679	1.3187	3	2.8472	3.0504	0.3046	0.5133	0.1557	1.8133	1.9725	SAMPLE			
5	1.5513	1.0953	3	2.4753	7.595E-03	87.7411	0.0013	0.1170	1.3063	1.3063	SAMPLE			
6	1.5635	1.1038	3	2.5571	0.3913	1.7997	0.0560	0.1008	1.3838	1.3838	SAMPLE			
7	1.8387	1.2981	3	2.8284	-0.6538	1.1573	-0.1127	0.1303	1.4826	1.4827	SAMPLE			
8	1.7938	1.2864	3	2.7614	1.1580	0.7132	0.2027	0.1444	1.6174	1.6441	SAMPLE			
9	1.8737	1.3229	3	2.8296	-0.6181	1.2688	-0.1182	0.1497	1.5294	1.5295	SAMPLE			
10	1.6133	1.1390	3	2.5993	0.3941	1.8328	0.0613	0.1124	1.4156	1.4192	SAMPLE			
11	1.6345	1.1540	3	2.6048	2.1886	0.3858	0.3687	0.1418	1.6500	1.7419	SAMPLE			
12	1.8377	1.2975	3	2.8535	0.6241	1.3177	0.1207	0.1590	1.6114	1.6192	SAMPLE			
13	1.8067	1.2755	3	2.8542	2.4699	0.3746	0.4073	0.1521	1.8079	1.9146	SAMPLE			
14	1.5629	1.1034	3	2.5122	1.5515	0.5033	0.2540	0.1276	1.5281	1.5784	SAMPLE			
15	1.5674	1.1066	3	2.4986	-1.4880	0.3759	-0.2607	0.0977	1.0930	1.0933	SAMPLE			
16	1.7953	1.2675	3	2.8688	1.3527	0.6385	0.2027	0.1283	1.6908	1.7258	SAMPLE			
17	1.4021	0.9899	3	2.2918	0.9487	0.7162	0.1520	0.1088	1.3805	1.3523	SAMPLE			
18	1.6451	1.1614	3	2.6287	1.1375	0.6897	0.1860	0.1282	1.5364	1.5635	SAMPLE			
19	1.8349	1.2855	3	2.8235	0.8762	0.9472	0.1507	0.1426	1.6259	1.6411	MB			
20	1.3742	0.9702	3	2.2128	0.7554	0.8548	0.1387	0.1185	1.2648	1.2794	239753001.1	DUP	0.0%	
21	4.7147	3.3286	3	7.2082	93.4250	0.0571	8.4773	0.4183	25.4861	239753001.1	WS			
22	2.1348	1.5072	3	3.2892	41.7577	0.0598	7.3307	0.3848	11.4747	11.4747	LCS		79.1254	
											39.2300	106.4%		

SampleID	Instr	Time (min.)	Alpha Counts	Beta Counts	Count Start Time	Count End Time	Machine
239753001	1A	60	6	55	11/23/2009 17:26	11/23/2009 18:26	PI/C
239753002	13A	60	12	70	11/23/2009 17:23	11/23/2009 18:23	PI/C
239753003	8A	75	8	116	11/23/2009 17:26	11/23/2009 18:41	PI/C
239753004	1C	85	9	156	11/23/2009 17:26	11/23/2009 18:51	PI/C
239753005	2A	60	10	44	11/23/2009 17:26	11/23/2009 18:26	PI/C
239753006	7C	60	3	33	11/23/2009 17:26	11/23/2009 18:26	PI/C
239753007	6B	75	22	82	11/23/2009 17:26	11/23/2009 18:41	PI/C
239753008	5C	75	11	104	11/23/2009 17:27	11/23/2009 18:42	PI/C
239753009	9D	85	3	137	11/23/2009 17:27	11/23/2009 18:52	PI/C
239753010	7D	60	4	41	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753011	14A	60	9	67	11/23/2009 17:23	11/23/2009 18:23	PI/C
239753012	12C	60	6	82	11/23/2009 17:23	11/23/2009 18:23	PI/C
239753013	10D	60	20	77	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753014	1D	60	4	54	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753015	8C	60	4	29	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753016	7A	60	4	55	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753017	10C	60	4	39	11/23/2009 17:27	11/23/2009 18:27	PI/C
239753018	12D	60	5	54	11/23/2009 17:23	11/23/2009 18:23	PI/C
1201972468	6D	75	7	101	11/23/2009 17:27	11/23/2009 18:42	PI/C
1201972469	9A	60	8	46	11/23/2009 17:26	11/23/2009 18:26	PI/C
1201972470	9B	60	5	617	11/23/2009 17:26	11/23/2009 18:26	PI/C
1201972471	3A	60	83	523	11/23/2009 17:26	11/23/2009 18:26	PI/C

ASSAY 20-Nov-09 4:25:11

Protocol id 8 228_REC
Time limit 180
Count limit 50000
Isotope Ba-133
Protocol date 9-Apr-07 10:03:07
Run id. 41

POS	RACK	BATCH	TIME	COUNTS	CPM	ERROR	% RECOVERY	COUNT TIME
1	98	1	180	1005	304.5	3.52		04:25:19
2	98	2	180	784	231	4.11	75.86	04:28:30
3	98	3	180	920	276.4	3.71	90.77	04:31:41
4	98	4	180	889	266	3.79	87.36	04:34:53
5	98	5	180	883	263.8	3.81	86.63	04:38:04
6	77	6	180	899	269.2	3.77	88.41	04:41:29
7	77	7	180	749	219.2	4.24	71.99	04:44:40
8	77	8	180	896	268.3	3.78	88.11	04:47:52
9	77	9	180	882	263.7	3.81	86.60	04:51:03
10	77	10	180	958	288.9	3.62	94.88	04:54:14
11	66	11	180	797	235.4	4.07	77.31	04:57:44
12	66	12	180	835	247.8	3.95	81.38	05:00:56
13	66	13	180	956	288.4	3.63	94.71	05:04:07
14	66	14	180	832	247	3.96	81.12	05:07:18
15	66	15	180	861	256.5	3.87	84.24	05:10:30
16	91	16	180	876	261.6	3.83	85.91	05:13:48
17	91	17	180	779	229.3	4.13	75.30	05:17:00
18	91	18	180	819	242.5	4	79.64	05:20:11
19	91	19	180	820	243	3.99	79.80	05:23:23
20	91	20	180	896	268.2	3.78	88.08	05:26:34
21	88	21	180	893	267.4	3.78	87.82	05:29:59
22	88	22	180	902	270.2	3.76	88.74	05:33:10
23	88	23	180	961	290	3.61	95.24	05:36:21

END OF ASSAY

RADIUM 226

Radiochemistry Batch Checklist, Rev 9

Batch# 920697 Product: 7a-226 Date: 12/2/09

Criteria:	Yes	No	Comments
Sample Solids are less than or equal to 100 mg for GAB.			NA
Samples have been blank corrected (if required)			NA
If activity less 10* MDA/ MDC, error is 150% or less of sample activity. If greater 10* MDA/ MDC, error is 40% or less. If below the MDA/ MDC, error is okay.	✓		
Instrument source check is within limits.	✓		
Instrument bkg check is within limits.	✓		
Method RDL/ LLD has been met.	✓		
If duplicate activities are less 5* MDA/ MDC, then RPD is 100% or less. If greater 5* MDA/ MDC, then RPD 20% or less. If below the MDA/ MDC, the RPD is 0%.			
Or meets the client's required RER acceptance criteria.	✓		
Tracer yield is 15-125%. Carrier yield 25-125%.			NA
Or meets the client's contract acceptance criteria.			
Method blank is less than the RDL/ LLD. (If rad samples, < 5% of lowest activity)	✓		
Sample was run within hold time.	✓		
Sample was correctly preserved if required.			NA
Smears Taken for Radioactive batches.			NA
Method Spike and LCS are within 75-125% or meets the client's contract acceptance criteria.	✓		
No blank spaces on data forms.			
All line outs initialed and dated.	✓		
No transcription errors are apparent.			
Aux data is correct.			NA
Client Special requirements page has been checked.	✓		
Raw Data and/ or spectrum are included and properly statused.	✓		
QC data entered into QC database and batch is in REVW	✓		
Hit notification complete (if necessary)			NA
Batch entered into Case Narrative.	✓		
Batch non-conformances completed, if applicable.	✓		NCR 764411
Batch non-conformances second reviewed and disposition verified to be completed.	✓		NCR 764411
Aliquot Correction completed if required.			NA
Review sample historical results if available (If REMP, results above MDC have been verified by historical results, recount or re-analysis.)	✓		

GEL Laboratories, LLC

revised 8/1/08

Primary Review Performed By:

Dymphney Pace

NETR 12/7/09

Secondary Review Performed By:

Z. Linch 12/2/09

Radium-226 Que Sheet

09-NOV-09

GEL Laboratories, Radiochemistry Division

Batch #: 920697

Spike Isotope: Radium-226 Spike Code: 01/26/10 Expiration Date: 01/26/10 Internal Due Date: 11/26/2009

LCS Isotope: Radium-226 LCS Code: 01/26/10 Expiration Date: 01/26/10

Bkg Count Time: 30 (Min) Sample Count Time: 30 (Min)

Prep Date: 11/24/09 Start Count Date: 11/24/09

Pipet ID: M-141B Balance ID: 01/26/09 Initials: LD Witness: DL 11-24-09

Comments: _____

Analyst: KSD1 First Client/Due Date: 12/07/2009

Expiration Date: 01/26/10 Vol: 0.1

End Initial/Degas Date/Time: 11/24/09

End LN De-em Date: 11/24/09

Bkg Count Time: 30 (Min) Start Count Date: 11/24/09

Initials: LD Witness: DL 11-24-09

Sample I	Client Description	Type	Hazard Code	Matrix	Min CRDL	Client	Position (Label)	Aliquot (ml or g)	End LN De-em Time	Start Count Time	Cell #	Det #	Bkg counts	Total Counts
239753001-1	M-141B	SAMPLE		WATER	1 pCML	KERR003	1	500	144S	1805	712	7	4	15
239753002-1	M-141009B	SAMPLE		WATER	1 pCML	KERR003	2	500	144S	1840	104	1	3	24
239753003-1	PB102309-A3	SAMPLE		WATER	1 pCML	KERR003	3	500	144S	1840	104	2	8	22
239753004-1	M-145B	SAMPLE		WATER	1 pCML	KERR003	4	500	144S	1840	201	3	8	23
239753005-1	M-139B	SAMPLE		WATER	1 pCML	KERR003	5	500	144S	1840	404	4	8	30
239753006-1	M-146B	SAMPLE		WATER	1 pCML	KERR003	6	500	144S	1840	204	5	8	49
239753007-1	M-144B	SAMPLE		WATER	1 pCML	KERR003	7	500	144S	1840	104	7	6	28
239753008-1	M-138B	SAMPLE		WATER	1 pCML	KERR003	8	500	151S	1910	101	1	8	37
239753009-1	M-138009B	SAMPLE		WATER	1 pCML	KERR003	9	500	151S	1910	101	2	8	35
239753010-1	M-138BDISS	SAMPLE		WATER	1 pCML	KERR003	10	500	151S	1910	201	3	4	15
239753011-1	M-138009BDISS	SAMPLE		WATER	1 pCML	KERR003	11	500	151S	1910	505	5	4	19
239753012-1	M-137B	SAMPLE		WATER	1 pCML	KERR003	12	500	151S	1910	100	7	4	13
239753013-1	M-137BDISS	SAMPLE		WATER	1 pCML	KERR003	13	500	153S	2000	101	1	8	34
239753014-1	M-148B	SAMPLE		WATER	1 pCML	KERR003	14	500	153S	2000	103	2	4	35
239753015-1	EB103009-GWA4	SAMPLE		WATER	1 pCML	KERR003	15	500	153S	2000	301	3	5	18
239753016-1	M-147B	SAMPLE		WATER	1 pCML	KERR003	16	500	153S	2000	501	5	8	46
239753017-1	M-147009B	SAMPLE		WATER	1 pCML	KERR003	17	500	153S	2000	101	7	4	20
239753018-1	EB110209-GWA3	SAMPLE		WATER	1 pCML	KERR003	18	500	1600	2035	101	1	8	13
1201967363-1	MB for batch 920697	MB		WATER	1 pCML	QC ACCOUNT	19	500	1600	2035	504	3	8	21
1201967364-1	M-137B(239753012DUP)	DUP		WATER	1 pCML	QC ACCOUNT	20	500	1600	2035	504	5	5	12
1201967365-1	M-137B(239753012MS)	MS		WATER	1 pCML	QC ACCOUNT	21	100	1600	2035	711	7	8	107
1201967366-1	LCS for batch 920697	LCS		WATER	1 pCML	QC ACCOUNT	22	500	1600	2125	101e	7	8	971

Data Reviewed By: Alphrey, Dale Date: 10/20/09
Page 1 of 1

Radium-226 Liquid

Filename : RA226.xls
 File type : Excel
 Version # : 1.2.5

Batch : 920697

Analyst : KSD1
 Prep Date : 11/24/2009

Ra-226 Abundance : 1

Ra-226 Method Uncertainty : 0.0918

Spike S/N : 0638-I
Spike Exp Date : 9/28/2010
Spike Activity (dpm/ml): 268.43
Spike Volume Added: 0.10

LCS S/N : 0638-I
LCS Exp Date : 9/28/2010
LCS Activity (dpm/ml): 268.43
LCS Volume Added: 0.10

Procedure Code : LUC26RAL
ParamName : Radium-226
Required MDA : 1 pCi/L
Halflife of Ra-226 : 1600 years
Halflife of Rn-222: 3.8235 days
Batch counted on : LUCAS CELL DETECTOR
BKG Count time : 30 min

Sample Characteristics			Count Raw Data			Weekly Background						
Pos.	Sample ID	Sample Aliquot L	Sample Aliquot SDev. L	Sample Date/Time	Cell Number	Gross Time (min.)	Gross Counts	Gross CPM	Counts	CPM	Count Time (min.)	Detector Efficiency (cpm/dpm)
1	239753001.1	0.5000	2.0256E-05	10/23/2009 10:00	712	30	15	0.500	4	0.133	30	2.0690
2	239753002.1	0.5000	2.0256E-05	10/23/2009 10:00	104	30	24	0.800	3	0.100	30	1.9720
3	239753003.1	0.5000	2.0256E-05	10/23/2009 12:15	204	30	22	0.733	8	0.267	30	2.1930
4	239753004.1	0.5000	2.0256E-05	10/26/2009 10:15	307	30	23	0.767	8	0.267	30	1.9310
5	239753005.1	0.5000	2.0256E-05	10/26/2009 12:55	404	30	30	1.000	8	0.267	30	1.9310
6	239753006.1	0.5000	2.0256E-05	10/27/2009 9:30	506	30	49	1.633	8	0.267	30	2.0040
7	239753007.1	0.5000	2.0256E-05	10/27/2009 12:25	709	30	28	0.933	6	0.200	30	2.2850
8	239753008.1	0.5000	2.0256E-05	10/28/2009 11:15	102	30	37	1.233	8	0.267	30	1.8550
9	239753009.1	0.5000	2.0256E-05	10/28/2009 11:15	207	30	35	1.167	8	0.267	30	2.1460
10	239753010.1	0.5000	2.0256E-05	10/28/2009 11:15	309	30	15	0.500	4	0.133	30	1.8770
11	239753011.1	0.5000	2.0256E-05	10/28/2009 11:15	505	30	19	0.633	4	0.133	30	2.3310
12	239753012.1	0.5000	2.0256E-05	10/29/2009 13:30	708	30	13	0.433	4	0.133	30	2.1880
13	239753013.1	0.5000	2.0256E-05	10/29/2009 13:30	112	30	34	1.133	8	0.267	30	1.9310
14	239753014.1	0.5000	2.0256E-05	10/29/2009 9:10	203	30	35	1.167	4	0.133	30	2.2540
15	239753015.1	0.5000	2.0256E-05	10/30/2009 11:10	311	30	18	0.600	6	0.200	30	2.1140
16	239753016.1	0.5000	2.0256E-05	11/2/2009 10:00	501	30	46	1.533	8	0.267	30	2.0870
17	239753017.1	0.5000	2.0256E-05	11/2/2009 10:00	707	30	20	0.667	4	0.133	30	2.2750
18	239753018.1	0.5000	2.0256E-05	11/2/2009 12:40	201	30	13	0.433	8	0.267	30	1.9930
19	1201967363.1	0.5000	2.0256E-05	11/24/2009 0:00	306	30	21	0.700	8	0.267	30	1.7470
20	1201967364.1	0.5000	2.0256E-05	10/29/2009 13:30	504	30	12	0.400	5	0.167	30	1.6150
21	1201967365.1	0.1000	1.1370E-05	10/29/2009 13:30	711	30	1071	35.700	8	0.267	30	2.2420
22	1201967366.1	0.5000	2.0256E-05	11/24/2009 0:00	706	30	971	32.367	8	0.267	30	2.1420

Detector Efficiency Error (cpm/dpm)	Cell Calibration Date	Cell Calibration Due Date	Rn-222 Ingrow			Count Start Date/Time	De-Gas to Ingrow	Rn-222 Corrections Ingrowth to Count	During Count	Ra-226 Decay
			End Date/Time	De-Gas Date/Time	12/1/2009 18:05					
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 14:25	12/1/2009 18:40	0.717	0.973	1.002	1.000	
0.05303	8/31/2009	8/31/2010	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.07722	12/19/2008	12/19/2009	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.06082	2/4/2009	2/4/2010	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.12371	3/2/2009	3/2/2010	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.14377	3/25/2009	3/25/2010	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 14:45	12/1/2009 18:40	0.718	0.971	1.002	1.000	
0.05303	8/31/2009	8/31/2010	11/24/2009 15:15	12/1/2009 15:15	12/1/2009 19:10	0.719	0.971	1.002	1.000	
0.07722	12/19/2008	12/19/2009	11/24/2009 15:15	12/1/2009 15:15	12/1/2009 19:10	0.719	0.971	1.002	1.000	
0.06082	2/4/2009	2/4/2010	11/24/2009 15:15	12/1/2009 15:15	12/1/2009 19:10	0.719	0.971	1.002	1.000	
0.14377	3/25/2009	3/25/2010	11/24/2009 15:15	12/1/2009 15:15	12/1/2009 19:10	0.719	0.971	1.002	1.000	
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 15:15	12/1/2009 19:10	0.719	0.971	1.002	1.000	
0.05303	8/31/2009	8/31/2010	11/24/2009 15:15	12/1/2009 15:35	12/1/2009 20:00	0.720	0.967	1.002	1.000	
0.07722	12/19/2008	12/19/2009	11/24/2009 15:15	12/1/2009 15:35	12/1/2009 20:00	0.720	0.967	1.002	1.000	
0.06082	2/4/2009	2/4/2010	11/24/2009 15:15	12/1/2009 15:35	12/1/2009 20:00	0.720	0.967	1.002	1.000	
0.14377	3/25/2009	3/25/2010	11/24/2009 15:15	12/1/2009 15:35	12/1/2009 20:00	0.720	0.967	1.002	1.000	
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 15:35	12/1/2009 20:00	0.720	0.967	1.002	1.000	
0.07722	12/19/2008	12/19/2009	11/24/2009 15:15	12/1/2009 16:00	12/1/2009 20:35	0.720	0.966	1.002	1.000	
0.06082	2/4/2009	2/4/2010	11/24/2009 15:15	12/1/2009 16:00	12/1/2009 20:35	0.720	0.966	1.002	1.000	
0.14377	3/25/2009	3/25/2010	11/24/2009 15:15	12/1/2009 16:00	12/1/2009 20:35	0.720	0.966	1.002	1.000	
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 16:00	12/1/2009 20:35	0.720	0.966	1.002	1.000	
0.06519	9/30/2009	9/30/2010	11/24/2009 15:15	12/1/2009 16:15	12/1/2009 21:25	0.721	0.962	1.002	1.000	

Notes:

1 Results are decay corrected to Sample Date/Time
2- Reference date for Spike Activity (dpm/ml) is the batch Prep Date
3- Spike Nominals are decay corrected to Sample Date/Time

2600

Results Pos.	Decision Level pCi/L	Critical Level pCi/L	Required MDA pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate CPM	Net Count Rate Error CPM	2 SIGMA		Nominal pCi/L	Recovery	
									Total Prop. Uncertainty pCi/L	Counting Uncertainty pCi/L			
1	0.1369	0.0966	1	0.2556	0.285	0.4016	0.3667	0.1453	0.1774	0.1845	SAMPLE		
2	0.1245	0.0879	1	0.2412	0.4880	0.2531	0.7000	0.1732	0.2221	0.2417	SAMPLE		
3	0.1828	0.1291	1	0.3169	0.2746	0.3988	0.4667	0.1826	0.2106	0.2202	SAMPLE		
4	0.2076	0.1466	1	0.3599	0.3341	0.3761	0.5000	0.1856	0.2431	0.2535	SAMPLE		
5	0.2076	0.1466	1	0.3599	0.4900	0.3063	0.7333	0.2055	0.2691	0.3071	SAMPLE		
6	0.2000	0.1412	1	0.3468	0.8800	0.2336	1.3667	0.2517	0.3176	0.4329	SAMPLE		
7	0.1519	0.1073	1	0.2710	0.4141	0.2729	0.7333	0.1944	0.2151	0.2337	SAMPLE		
8	0.2158	0.1523	1	0.3741	0.6714	0.2373	0.9667	0.2236	0.3044	0.3349	SAMPLE		
9	0.1865	0.1317	1	0.3234	0.5403	0.2548	0.9000	0.2186	0.2572	0.2869	SAMPLE		
10	0.1508	0.1065	1	0.2816	0.2517	0.4009	0.3667	0.1453	0.1955	0.2029	SAMPLE		
11	0.1214	0.0857	1	0.2267	0.2764	0.3506	0.5000	0.1599	0.1732	0.1963	SAMPLE		
12	0.1294	0.1013	1	0.2415	0.1767	0.4627	0.3000	0.1374	0.1586	0.1633	SAMPLE		
13	0.2079	0.1468	1	0.3604	0.5799	0.2548	0.8667	0.2160	0.2833	0.3079	SAMPLE		
14	0.1259	0.0889	1	0.2351	0.5923	0.2157	1.0333	0.2082	0.2339	0.2722	SAMPLE		
15	0.1644	0.1161	1	0.2933	0.2445	0.4128	0.4000	0.1633	0.1956	0.2026	SAMPLE		
16	0.1923	0.1358	1	0.3335	0.7842	0.2410	1.2667	0.2449	0.2972	0.3963	SAMPLE		
17	0.1248	0.0881	1	0.2330	0.3029	0.3130	0.5333	0.1633	0.1818	0.1937	SAMPLE		
18	0.2014	0.1422	1	0.3492	0.1081	0.9198	0.1667	0.1528	0.1941	0.1958	SAMPLE		
19	0.2298	0.1622	1	0.3984	0.3205	0.4187	0.4333	0.1795	0.2602	0.2692	MB		
20	0.1965	0.1387	1	0.3575	0.1867	0.6063	0.2333	0.1374	0.2155	0.2244	DUP	5.5%	
21	0.8952	0.6320	1	1.5522	102.1017	0.0721	35.4333	1.0949	6.1840	23.3646	239753012.1	MS	120.9196
22	0.1881	0.1328	1	0.3261	19.4337	0.0728	32.1000	1.0430	1.2376	4.4635	239753012.1	LCS	24.1832

METHOD CALIBRATION DATA

ALPHA SPECTROSCOPY

Alpha Spectroscopy Calibration Sources

The following is a summary of the procedure performed for preparing mixed alpha calibration standards:

A calibration stock solution was prepared by combining the following in a volumetric flask and diluting to 50 ml (51.4561 grams). These individual standards were first verified by direct precipitation of small aliquots of each standard (as described in Attachment I).

Isotope	Serial #	amount used (g)	dpm (note 1)
Gd-148	64445-278	0.2471	212.159287
Np-237	4341	1.8075	204.438594
Cm-244	4320A	7.2704	240.144737


Note 1: Dpm values are decay corrected to 2/7/2003.

Forty one weighted aliquots were then directly precipitated using Neodymium Flouride /HF system. The sources were then mounted on 0.1Poly-propylene filters and taped securely to 1 inch stainless steel planchettes for counting in an Alpha Spectroscopy system. The liquid fraction that passes through the filter is collected, traced with Am-241 and prepared for counting using the identical procedure. These samples are counted to ensure there is no more than 1% loss in the filtering processes. All sources pass this requirement. The DPM information for each source is listed in attachment II.

Certificate files were then created on the Alpha system used for acquisition and processing of data. Each source is assigned a name (AESS-001 through AEES-041). The information for the source activities is entered into the certificate files appropriate for the detector being used.

For example: If source AEES-001 is used for calibrating detector 25, the source data is entered into the certificate file name [env_alpha.cer]U025.cer.

The computer software uses these certificate files to calculate an energy calibration and determine the efficiency of the detector after counting the source.


Steve Nell
4/1/03

2002 Alpha Eff Source Stock Verification

Curium-244

Isotope	Value pCi/g	Isotope	Value pCi/g	Isotope	Value pCi/g	
SSTOCK2002A2_AM	106.000	SSTOCK2002A2_AM	90.100	SSTOCK2002A2_AM	96.080	
SSTOCK2002B2_AM	106.000	SSTOCK2002B2_AM	87.200	SSTOCK2002B2_AM	93.750	
SSTOCK2002C2_AM	106.000	SSTOCK2002C2_AM	93.500	SSTOCK2002C2_AM	96.560	
Mean Value (Counting) = StdDev =	106.000 0	Mean Value (Counting) = StdDev =	90.267 3.153305144	Mean Value (Counting) = StdDev =	95.463 1.503074627	
Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail = Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	108.1230 106 106 Pass 0 10.6 Pass	pCi/g PASS ① Fair & 2σ Pass Pass 10 % of Mean = Rule 2 (Pass/Fail)	pCi/g Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Pass	pCi/g 92.0900 83.96005638 96.57327696 Pass 6.306610289 9.0266666667	Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	95.6460 92.45718408 98.46848259 Pass 3.006149253 \\$546333333 Pass

The analyst prepared three standard verification sources for the mixed alpha stock standard using 0.1030 g for source #1, 0.1036 g for source #2, and 0.1028 g for source #3. Each standard was combined with 1.0 mL of Am-243 standard 0454-A and 0.1 mL of Nd carrier in a disposable centrifuge tube. Four mL of 2 M HCl was added to each standard and then diluted with 4 mL of DI water. 5 mL of ascorbic acid was added to each sample then one mL of 48% HF was added to precipitate Nd(IV) and Curium(IV) fluoride. After 30 minutes, each sample was filtered following routine procedures for alpha spectroscopy source preparation. Each source was counted using routine alpha spec procedures. pCi/l values for the Mixed Alpha Stock were calculated and compared to Am-243 certified values.

Neptunium-237

Isotope	Value pCi/g	Isotope	Value pCi/g	Isotope	Value pCi/g	
SSTOCK2002A2_AM	106.000	SSTOCK2002B2_AM	90.100	SSTOCK2002A2_AM	96.080	
SSTOCK2002B2_AM	106.000	SSTOCK2002B2_AM	87.200	SSTOCK2002B2_AM	93.750	
SSTOCK2002C2_AM	106.000	SSTOCK2002C2_AM	93.500	SSTOCK2002C2_AM	96.560	
Mean Value (Counting) = StdDev =	106.000 0	Mean Value (Counting) = StdDev =	90.267 3.153305144	Mean Value (Counting) = StdDev =	95.463 1.503074627	
Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail = Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	108.1230 106 106 Pass 0 10.6 Pass	pCi/g PASS ① Fair & 2σ Pass Pass 10 % of Mean = Pass	pCi/g Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Pass	pCi/g 92.0900 83.96005638 96.57327696 Pass 6.306610289 9.0266666667	Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	95.6460 92.45718408 98.46848259 Pass 3.006149253 \\$546333333 Pass

Gadolinium-148

Isotope	Value pCi/g	Isotope	Value pCi/g	Isotope	Value pCi/g
SSTOCK2002A2_AM	90.100	SSTOCK2002B2_AM	87.200	SSTOCK2002A2_AM	96.080
SSTOCK2002B2_AM	87.200	SSTOCK2002B2_AM	93.750	SSTOCK2002B2_AM	93.750
SSTOCK2002C2_AM	93.500	SSTOCK2002C2_AM	96.560	SSTOCK2002C2_AM	96.560
Mean Value (Counting) = StdDev =	90.267 3.153305144	Mean Value (Counting) = StdDev =	98.02% 98.02%	Mean Value (Counting) = StdDev =	99.81% 1.503074627
Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail = Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	92.0900 83.96005638 96.57327696 Pass 6.306610289 9.0266666667	pCi/g Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Pass	pCi/g 92.0900 83.96005638 96.57327696 Pass 6.306610289 9.0266666667	Target = Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	95.6460 92.45718408 98.46848259 Pass 3.006149253 \\$546333333 Pass

- ① The rule failed because the 3 results from 3 sources were the same.

Therefore, the std dev was zero. The intent of this is to ensure an appropriate amount of counts are achieved for proper determinations. Since for each standard the # of counts achieved was just under 10000 which has a counting error of nearly 1%. Because the standard's bias is < 2% from the known value the standard is acceptable.

Robert J. Lin 021208

Mixed alpha Isotope	Reference date =	Stock Dpm/g	Reference date	Half-life (years)	amount used for mixed	Dpm/g mixed	Decay corr dpm/g
Gd-148	64445-278 (0502)	44354.59289	9/5/2002	74.60	0.2471	212.9974853	212.159287
Np-237	Srm 4341 (0493)	5820	3/1/1992	2.14E+06	1.8075	204.4393182	204.438594
Cm-244	SRM 4320a (0490)	2223.6	2/1/1996	18.1	7.2704	314.1796879	240.144737
Source	Amount of standard used	dpm Gd-148	dpm Np-237	dpm Cm-244	dpm Gd-148	dpm Np-237	dpm Cm-244
AESS-001	1.0362	219.839	211.839	248.838	3.664	3.531	4.147 -
AESS-002	1.0344	219.458	211.471	248.406	3.658	3.525	4.140 -
AESS-003	1.034	219.373	211.390	248.310	3.656	3.523	4.138 -
AESS-004	1.0331	219.182	211.206	248.094	3.653	3.520	4.135 -
AESS-005	1.0353	219.649	211.655	248.622	3.661	3.528	4.144 -
AESS-006	1.0331	219.182	211.206	248.094	3.653	3.520	4.135 -
AESS-007	1.0348	219.542	211.553	248.502	3.659	3.526	4.142 -
AESS-008	1.0363	219.861	211.860	248.862	3.664	3.531	4.148 -
AESS-009	1.0352	219.627	211.635	248.598	3.660	3.527	4.143 -
AESS-010	1.0346	219.500	211.512	248.454	3.658	3.525	4.141 -
AESS-011	1.0353	219.649	211.655	248.622	3.661	3.528	4.144 -
AESS-012	1.0367	219.946	211.941	248.958	3.666	3.532	4.149 -
AESS-013	1.0396	220.561	212.534	249.654	3.676	3.542	4.161
AESS-014	1.0368	219.967	211.962	248.982	3.666	3.533	4.150
AESS-015	1.0363	219.861	211.860	248.862	3.664	3.531	4.148
AESS-016	1.0353	219.649	211.655	248.622	3.661	3.528	4.144
AESS-017	1.0356	219.712	211.717	248.694	3.662	3.529	4.145
AESS-018	1.0359	219.776	211.778	248.766	3.663	3.530	4.146
AESS-019	1.0349	219.564	211.574	248.526	3.659	3.526	4.142
AESS-020	1.0361	219.818	211.819	248.814	3.664	3.530	4.147
AESS-021	1.0348	219.542	211.553	248.502	3.659	3.526	4.142
AESS-022	1.0353	219.649	211.655	248.622	3.661	3.528	4.144
AESS-023	1.0353	219.649	211.655	248.622	3.661	3.528	4.144
AESS-024	1.0343	219.436	211.451	248.382	3.657	3.524	4.140
AESS-025	1.0364	219.882	211.880	248.886	3.665	3.531	4.148
AESS-026	1.0336	219.288	211.308	248.214	3.655	3.522	4.137
AESS-027	1.0353	219.649	211.655	248.622	3.661	3.528	4.144
AESS-028	1.0366	219.924	211.921	248.934	3.665	3.532	4.149

Attachment II

AESS-029	1.0355	219.691	211.696	248.670	3.662	3.528	4.144
AESS-030	1.0349	219.564	211.574	248.526	3.659	3.526	4.142
AESS-031	1.0343	219.436	211.451	248.382	3.657	3.524	4.140
AESS-032	1.0326	219.076	211.103	247.973	3.651	3.518	4.133
AESS-033	1.0308	218.694	210.735	247.541	3.645	3.512	4.126
AESS-034	1.0314	218.821	210.858	247.685	3.647	3.514	4.128
AESS-035	1.0303	218.588	210.633	247.421	3.643	3.511	4.124
AESS-036	1.0343	219.436	211.451	248.382	3.657	3.524	4.140
AESS-037	1.0353	219.649	211.655	248.622	3.661	3.528	4.144
AESS-038	1.0373	220.073	212.064	249.102	3.668	3.534	4.152
AESS-039	1.0334	219.245	211.267	248.166	3.654	3.521	4.136
AESS-040	1.0346	219.500	211.512	248.454	3.658	3.525	4.141
AESS-041	1.0352	219.627	211.635	248.598	3.660	3.527	4.143



National Institute of Standards & Technology

①490
0491

Certificate

Standard Reference Material 4320A Curium-244 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive curium-244 nitrate and nitric acid dissolved in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of alpha-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains curium-244 with a total activity of approximately 200 Bq. Curium-244 decays by alpha-particle emission to plutonium-240, which also decays by alpha-particle emission. None of the alpha particles escape from the SRM ampoule. During the decay process X-rays and gamma rays with energies from 40 keV to 1100 keV are also emitted. Most of these photons escape from the SRM ampoule but their intensities are so small that they do not represent a radiation hazard. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains nitric acid (HNO_3) with a concentration of 1 mole per liter of water. The solution is corrosive and represents a health hazard if it comes in contact with eyes or skin. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2. The ampoule should be opened only by persons qualified to handle both radioactive material and strong acid solution.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least February 2006.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) both because of the radioactivity and because of the strong acid.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, J.M.R. Hutchinson, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by N.M. Trahey.

Gaithersburg, Maryland 20899

February 1996 (Text only revised November 1997)

Thomas E. Gills, Chief
Standard Reference Materials Program

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood. In addition to the radioactive material, the solution contains strong acid and is corrosive.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle.
NEVER PIPETTE BY MOUTH
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4320A
 (Certified values are shown in bold type)

Source identification number	NIST SRM 4320A		
Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter (16.5 ± 0.5) mm Wall Thickness (0.60 ± 0.04) mm Barium content Less than 2.5% Lead-oxide content Less than 0.02% Other heavy elements Trace quantities		
Solution density	(1.030 ± 0.002) g·mL⁻¹ at 22.8 °C [b]*		
Solution mass	Approximately 5.15 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol·L ⁻¹)	Mass Fraction (g·g ⁻¹)
	H ₂ O	54	0.94
	HNO ₃	1.0	0.06
	HCl	<0.001	<4 × 10 ⁻⁵
	²⁴⁴ Cm ⁺³	5 × 10 ⁻¹¹	1 × 10 ⁻¹¹
Radiological Properties:			
Radionuclide	Curium-244		
Reference time	12:00 EST, 1 February 1996 [c]		
Massic activity of the solution [d]	37.06 Bq·g⁻¹ <i>74.12 Bq·g⁻¹</i>		
Relative expanded uncertainty (<i>k</i> =2)	0.68% [e] [f]		
Alpha-particle-emitting daughters	Plutonium-240: (0.22 ± 0.11) Bq·g⁻¹ [b] [c]		
Alpha-particle-emitting impurities	Curium-243: (0.005 ± 0.004) Bq·g⁻¹ [b] [g]		
Photon-emitting impurities	None detected [h]		
Half lives used in the decay corrections	Curium-244: (18.10 ± 0.02) a [i] Plutonium-240: (6563 ± 7) a [i]		
Calibration method	Two 4πα liquid-scintillation counting systems		

- [i] The stated uncertainty is the standard uncertainty. See reference [5].
- [j] Relative standard uncertainty of the input quantity x_i .
- [k] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [m] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_i(y)/y \equiv |\partial y/\partial x_i| \cdot u(x_i)/x_i = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u_i(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [n] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [p] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [q] The live time is determined by counting the pulses from a gated oscillator.
- [r] The standard uncertainty given is for the detected Cm-243 impurity. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(response\ per\ Bq\ of\ impurity)/(response\ per\ Bq\ of\ Cm-244)\} \cdot \{(Bq\ of\ impurity)/(Bq\ of\ Cm-244)\}$.
- [s] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(response\ per\ Bq\ of\ impurity)/(response\ per\ Bq\ of\ Cm-244)\} \cdot \{(Bq\ of\ impurity)/(Bq\ of\ Cm-244)\}$. Thus $u_i(y)/y$ is the relative change in y if the impurity were present with a mass activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] Evaluated Nuclear Structure Data File (ENSDF), February 1996.



ANALYTICS

0502

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 U.S.A.Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

64445-278

Gd-148 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master liquid radionuclide solution source. The master source was calibrated by liquid scintillation counting.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

ISOTOPE:	Gd-148
ACTIVITY (dps):	3.759 E3
HALF-LIFE:	74.6 years
CALIBRATION DATE:	September 5, 2002 12:00 EST
TOTAL UNCERTAINTY*:	2.7%
SYSTEMATIC:	1.9%
RANDOM:	0.8%

99% confidence level.

25
31
30
31
31
71
155

5.08493 grams 0.1M HCl solution.


30

P O NUMBER 3207RD, Item 1

31

SOURCE PREPARED BY:

31


M.D. Currie, Radiochemist

Q A APPROVED:

 9-6-02

0493



National Institute of Standards & Technology

Certificate

Standard Reference Material 4341 Radioactivity Standard

Radionuclide	Neptunium-237
Source identification	SRM 4341
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule ^{(1)*}
Solution mass	Approximately 5 grams
Solution composition	Neptunium-237 in 2 mol·L ⁻¹ nitric acid
Reference time	March 1992
Radioactivity concentration	97.0 Bq·g ⁻¹
Overall uncertainty	1.28 percent ⁽²⁾
Photon-emitting impurities	None detected ⁽³⁾
Alpha-particle-emitting impurities	None detected ⁽⁴⁾
Half life	$(2.14 \pm 0.11) \times 10^6$ years ⁽⁵⁾
Measuring instrument	NIST "0.8π"α defined-solid-angle counter with scintillation detector

This standard reference material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, J.M. Robin Hutchinson, Acting Group Leader.

Gaithersburg, MD
January 1993

William P. Reed, Chief
Standard Reference Materials Program

*Notes on back

NOTES

- (1) Approximately five milliliters of solution. Ampoule specifications:

body diameter	16.5 ± 0.5 mm
wall thickness	0.60 ± 0.04 mm
barium content	less than 2.5 percent
lead oxide content	less than 0.02 percent
other heavy elements	trace quantities

- (2) The overall uncertainty was formed by taking three times the quadratic combination of the standard deviations of the mean, or approximations thereof, for the following:

a) alpha-particle-emission-rate measurements	0.34 percent
b) background	0.01 percent
c) livetime	0.10 percent
d) detection efficiency	0.16 percent
e) count-rate-vs-energy extrapolation to zero energy	0.10 percent
f) half life	0.00 percent
g) gravimetric measurements	0.10 percent
h) alpha-emitting impurities	0.10 percent

- (3) The protactinium-233 daughter of neptunium-237 is approximately in equilibrium. The limit of detection for photon-emitting impurities is

$0.19 \gamma \cdot s^{-1} \cdot g^{-1}$ for energies between 30 and 307 keV and
 $0.01 \gamma \cdot s^{-1} \cdot g^{-1}$ for energies between 317 and 1750 keV,

provided that the impurity photons are separated in energy by 5 keV or more from photons emitted in the decay of neptunium-237 and progeny.

- (4) The limit of detection for alpha-particle-emitting impurities is

$0.10 \alpha \cdot s^{-1} \cdot g^{-1}$ for energies between 1.0 and 4.3 MeV and
 $0.05 \alpha \cdot s^{-1} \cdot g^{-1}$ for energies between 4.9 and 10 MeV.

- (5) Evaluated Nuclear Structure Data File (ENSDF), February 1990.

For further information please contact Dr. J.M. Robin Hutchinson at NIST.
Telephone: (301) 975-5532
FAX: (301) 926-7416

Instrument : CHAMBER 025

Detector : 45-149AA5
 Standard ID : AE55-025
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:11:20

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.535
NP-237	0807-B	28-FEB-2010	4768.800	4768.330
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.308

Energy Calibration Zero : 2328.778
 Energy Calibration Slope : 4.877499
 Energy Calibration Quadratic : 3.1643920E-04
 Energy Calibration Range : 2334 to 7655 KeV

The Energy Calibration : energy = 2328.778+(4.877499*Channel)+(3.1643920E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.044	3301.926	2.000000	0.4799997	70.71068	95.00000
NP-237	4433.147	4903.809	8.000000	1.919999	35.35534	95.00000
CM-244	5532.473	5887.442	103.0000	24.71998	9.853293	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:11:20
 Average Efficiency : 0.3284177
 Average Efficiency Error : 9.0522086E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	195.5670	28-FEB-2010	2988.044	3301.926	15199.00	0.3290007	1.4139536E-02	59.97539
NP-237	167.9916	28-FEB-2010	4433.147	4903.809	13268.00	0.3290368	1.6698070E-02	66.06133
CM-244	157.2432	28-FEB-2010	5532.473	5887.442	11579.00	0.3269968	1.6631199E-02	60.85799

Instrument : CHAMBER 026

Detector : 78204
 Standard ID : AE55-026
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:11:30

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.801
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2354.369
 Energy Calibration Slope : 4.953624
 Energy Calibration Quadratic : 3.1949743E-04
 Energy Calibration Range : 2359 to 7762 KeV

The Energy Calibration : energy = 2354.369+(4.953624*Channel)+(3.1949743E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.632	3302.017	1.000000	0.2399998	100.0000	95.00000
NP-237	4433.847	4905.782	9.000000	2.159998	33.33334	95.00000
CM-244	5532.384	5886.736	72.000000	17.27999	11.78511	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:11:30
 Average Efficiency : 0.3187393
 Average Efficiency Error : 9.3433373E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	199.5072	28-FEB-2010	2988.632	3302.017	15068.00	0.3197283	1.6198499E-02	48.77509
NP-237	168.0294	28-FEB-2010	4433.847	4905.782	13060.00	0.3237990	1.6436102E-02	68.84023
CM-244	160.5822	28-FEB-2010	5532.384	5886.736	11313.00	0.3130322	1.5926857E-02	52.34050

Instrument : CHAMBER 027

Detector : 42484
 Standard ID : AE55-027
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:11:40

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.724
NP-237	0807-B	28-FEB-2010	4768.800	4768.644
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.730

Energy Calibration Zero : 2368.047
 Energy Calibration Slope : 4.986534
 Energy Calibration Quadratic : 2.9072489E-04
 Energy Calibration Range : 2373 to 7779 KeV

The Energy Calibration : energy = 2368.047+(4.986534*Channel)+(2.9072489E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.848	3300.506	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4435.280	4902.342	7.000000	1.679999	37.79645	95.00000
CM-244	5534.038	5882.559	88.000000	21.11999	10.66004	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:11:40
 Average Efficiency : 0.3416629
 Average Efficiency Error : 1.0012131E-02
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	193.4238	28-FEB-2010	2990.848	3300.506	15371.00	0.3364211	1.7039858E-02	49.61097
NP-237	161.6154	28-FEB-2010	4435.280	4902.342	13349.00	0.3441128	1.7461589E-02	67.07381
CM-244	148.1754	28-FEB-2010	5534.038	5882.559	11500.00	0.3447433	1.7535616E-02	54.61600

Instrument : CHAMBER 028

Detector : 78792
 Standard ID : AE55-028
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:11:50

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.824
NP-237	0807-B	28-FEB-2010	4768.800	4768.798
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.948

Energy Calibration Zero : 2314.256
 Energy Calibration Slope : 4.950538
 Energy Calibration Quadratic : 3.2644827E-04
 Energy Calibration Range : 2319 to 7726 KeV

The Energy Calibration : energy = 2314.256+(4.950538*Channel)+(3.2644827E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.528	3302.181	2.000000	0.4799997	70.71068	95.00000
NP-237	4435.396	4902.811	8.000000	1.919999	35.35534	95.00000
CM-244	5530.503	5885.549	76.000000	18.23999	11.47079	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:11:50
 Average Efficiency : 0.3050518
 Average Efficiency Error : 8.9486167E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	199.6542	28-FEB-2010	2988.528	3302.181	14354.00	0.3043484	1.5429220E-02	44.19452
NP-237	168.1992	28-FEB-2010	4435.396	4902.811	12198.00	0.3021244	1.5351990E-02	59.07529
CM-244	156.7614	28-FEB-2010	5530.503	5885.549	10898.00	0.3088536	1.5724592E-02	43.16341

Instrument : CHAMBER 029

Detector : 33454
 Standard ID : AE55-029
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:00

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.660
NP-237	0807-B	28-FEB-2010	4768.800	4768.540
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.916

Energy Calibration Zero : 2352.222
 Energy Calibration Slope : 4.897179
 Energy Calibration Quadratic : 2.9826825E-04
 Energy Calibration Range : 2357 to 7680 KeV

The Energy Calibration : energy = 2352.222+(4.897179*Channel)+(2.9826825E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.921	3298.464	2.000000	0.4799997	70.71068	95.00000
NP-237	4435.920	4906.562	8.000000	1.919999	35.35534	95.00000
CM-244	5534.740	5883.900	86.000000	20.63999	10.78328	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:00
 Average Efficiency : 0.3123440
 Average Efficiency Error : 9.1590546E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	201.5742	28-FEB-2010	2988.921	3298.464	14665.00	0.3079810	1.5608890E-02	58.07315
NP-237	169.7700	28-FEB-2010	4435.920	4906.562	12802.00	0.3141530	1.5951235E-02	72.72359
CM-244	154.8234	28-FEB-2010	5534.740	5883.900	10984.00	0.3151226	1.6041664E-02	56.41142

Instrument : CHAMBER 030

Detector : 33447
 Standard ID : AE55-030
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:11
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:11

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.992
NP-237	0807-B	28-FEB-2010	4768.800	4768.798
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2380.611
 Energy Calibration Slope : 4.957869
 Energy Calibration Quadratic : 3.0952605E-04
 Energy Calibration Range : 2386 to 7782 KeV

The Energy Calibration : energy = 2380.611+(4.957869*Channel)+(3.0952605E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:30
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.078	3298.267	1.000000	0.2399998	100.0000	95.00000
NP-237	4434.110	4905.334	11.000000	2.639998	30.15113	95.00000
CM-244	5530.758	5884.331	76.000000	18.23999	11.47079	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:11
 Average Efficiency : 0.3237875
 Average Efficiency Error : 9.4913049E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	198.9792	28-FEB-2010	2990.078	3298.267	14918.00	0.3173854	1.6081888E-02	56.46090
NP-237	166.3758	28-FEB-2010	4434.110	4905.334	13266.00	0.3321642	1.6856849E-02	66.08606
CM-244	157.1856	28-FEB-2010	5530.758	5884.331	11410.00	0.3225155	1.6407013E-02	58.34272

Instrument : CHAMBER 031

Detector : 79988
 Standard ID : AE55-031
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:12
 Calibration Count Time : 240.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:20

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.800
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.862

Energy Calibration Zero : 2341.273
 Energy Calibration Slope : 4.909827
 Energy Calibration Quadratic : 3.3634368E-04
 Energy Calibration Range : 2346 to 7722 KeV

The Energy Calibration : energy = 2341.273+(4.909827*Channel)+(3.3634368E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.232	3301.398	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4436.778	4906.511	11.00000	2.640001	30.15113	95.00000
CM-244	5530.640	5883.813	45.00000	10.80000	14.90712	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:20
 Average Efficiency : 0.3439283
 Average Efficiency Error : 9.4762286E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	193.6650	28-FEB-2010	2990.232	3301.398	15614.00	0.3412869	1.4660587E-02	55.13291
NP-237	162.9186	28-FEB-2010	4436.778	4906.511	13684.00	0.3499030	1.7749120E-02	69.00159
CM-244	153.1968	28-FEB-2010	5530.640	5883.813	11785.00	0.3419108	1.7383847E-02	51.83997

Instrument : CHAMBER 033

Detector : 78785
 Standard ID : AE55-033
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:12
 Calibration Count Time : 240.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:35

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.801
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2378.839
 Energy Calibration Slope : 4.956201
 Energy Calibration Quadratic : 3.2238252E-04
 Energy Calibration Range : 2384 to 7792 KeV

The Energy Calibration : energy = 2378.839+(4.956201*Channel)+(3.2238252E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.294	3301.695	2.000000	0.4800001	70.71068	95.00000
NP-237	4433.763	4905.875	9.000000	2.160001	33.33334	95.00000
CM-244	5532.781	5887.338	79.000000	18.96000	11.25088	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:35
 Average Efficiency : 0.3097593
 Average Efficiency Error : 8.5533001E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	192.4158	28-FEB-2010	2988.294	3301.695	13801.00	0.3036067	1.3071683E-02	47.79169
NP-237	161.7816	28-FEB-2010	4433.763	4905.875	12172.00	0.3134331	1.5927175E-02	58.02074
CM-244	147.2670	28-FEB-2010	5532.781	5887.338	10457.00	0.3153150	1.6065678E-02	52.63453

Instrument : CHAMBER 035

Detector : 78202
 Standard ID : AE55-035
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:12
 Calibration Count Time : 240.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:49

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.811
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.071

Energy Calibration Zero : 2339.498
 Energy Calibration Slope : 4.957463
 Energy Calibration Quadratic : 3.2899898E-04
 Energy Calibration Range : 2344 to 7761 KeV

The Energy Calibration : energy = 2339.498+(4.957463*Channel)+(3.2899898E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.528	3298.375	4.000000	0.9600002	50.00000	95.00000
NP-237	4432.590	4905.766	17.000000	4.080001	24.25356	95.00000
CM-244	5534.227	5884.327	82.000000	19.68000	11.04315	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:49
 Average Efficiency : 0.3011819
 Average Efficiency Error : 8.3130784E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	198.6666	28-FEB-2010	2989.528	3298.375	14249.00	0.3035896	1.3062931E-02	46.51835
NP-237	168.2934	28-FEB-2010	4432.590	4905.766	12293.00	0.3042536	1.5458381E-02	64.58060
CM-244	158.8128	28-FEB-2010	5534.227	5884.327	10554.00	0.2950884	1.5032584E-02	50.29810

Instrument : CHAMBER 036

Detector : 78203
 Standard ID : AE55-036
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:12
 Calibration Count Time : 240.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:12:59

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.275
NP-237	0807-B	28-FEB-2010	4768.800	4768.954
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.157

Energy Calibration Zero : 2354.929
 Energy Calibration Slope : 4.916327
 Energy Calibration Quadratic : 3.5188830E-04
 Energy Calibration Range : 2360 to 7758 KeV

The Energy Calibration : energy = 2354.929+(4.916327*Channel)+(3.5188830E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.984	3301.734	3.000000	0.7200001	57.73503	95.00000
NP-237	4434.980	4906.333	10.000000	2.400001	31.62278	95.00000
CM-244	5532.965	5882.330	79.000000	18.96000	11.25088	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:12:59
 Average Efficiency : 0.3258305
 Average Efficiency Error : 8.9820670E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	201.3204	28-FEB-2010	2989.984	3301.734	15169.00	0.3189372	1.3707514E-02	56.87218
NP-237	167.4312	28-FEB-2010	4434.980	4906.333	13321.00	0.3314448	1.6819313E-02	67.32282
CM-244	156.4188	28-FEB-2010	5532.965	5882.330	11643.00	0.3305905	1.6812116E-02	56.23889

Instrument : CHAMBER 037

Detector : 45-149BB5
 Standard ID : AE55-037
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:13:10

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.661
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2384.221
 Energy Calibration Slope : 4.937421
 Energy Calibration Quadratic : 2.6759913E-04
 Energy Calibration Range : 2389 to 7721 KeV

The Energy Calibration : energy = 2384.221+(4.937421*Channel)+(2.6759913E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.569	3301.766	6.000000	1.439999	40.82483	95.00000
NP-237	4432.923	4904.219	17.000000	4.079998	24.25356	95.00000
CM-244	5532.481	5886.614	107.00000	25.67998	9.667364	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:13:10
 Average Efficiency : 0.3580391
 Average Efficiency Error : 9.8568676E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	197.7372	28-FEB-2010	2990.569	3301.766	16393.00	0.3509030	1.5061274E-02	69.74401
NP-237	167.1294	28-FEB-2010	4432.923	4904.219	14632.00	0.3646860	1.8482015E-02	78.18130
CM-244	154.7664	28-FEB-2010	5532.481	5886.614	12623.00	0.3620973	1.8390791E-02	72.24805

Instrument : CHAMBER 038

Detector : 72532
 Standard ID : AE55-038
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:13:20

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.800
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.935

Energy Calibration Zero : 2378.027
 Energy Calibration Slope : 4.946940
 Energy Calibration Quadratic : 3.2543507E-04
 Energy Calibration Range : 2383 to 7785 KeV

The Energy Calibration : energy = 2378.027+(4.946940*Channel)+(3.2543507E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.424	3299.282	2.000000	0.4799997	70.71068	95.00000
NP-237	4434.917	4906.502	10.000000	2.399998	31.62278	95.00000
CM-244	5532.790	5887.039	104.00000	24.95998	9.805807	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:13:20
 Average Efficiency : 0.3420325
 Average Efficiency Error : 9.4205979E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	200.1408	28-FEB-2010	2991.424	3299.282	15846.00	0.3351398	1.4392883E-02	55.53449
NP-237	170.0886	28-FEB-2010	4434.917	4906.502	14205.00	0.3479221	1.7639427E-02	64.95515
CM-244	157.7460	28-FEB-2010	5532.790	5887.039	12310.00	0.3464782	1.7604353E-02	59.85936

Instrument : CHAMBER 039

Detector : 45-149BB2
 Standard ID : AE55-039
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:13:31

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.799
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2388.746
 Energy Calibration Slope : 4.914010
 Energy Calibration Quadratic : 3.2089511E-04
 Energy Calibration Range : 2394 to 7757 KeV

The Energy Calibration : energy = 2388.746+(4.914010*Channel)+(3.2089511E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.039	3298.756	6.000000	1.439999	40.82483	95.00000
NP-237	4436.729	4905.040	12.000000	2.879998	28.86751	95.00000
CM-244	5532.232	5884.005	110.0000	26.39998	9.534626	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:13:31
 Average Efficiency : 0.3648929
 Average Efficiency Error : 1.0049243E-02
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	192.2418	28-FEB-2010	2988.039	3298.756	16118.00	0.3548785	1.5236211E-02	63.08267
NP-237	159.1506	28-FEB-2010	4436.729	4905.040	14402.00	0.3769790	1.9109026E-02	85.58052
CM-244	151.7142	28-FEB-2010	5532.232	5884.005	12591.00	0.3684085	1.8712135E-02	67.12554

Instrument : CHAMBER 040

Detector : 78773
 Standard ID : AE55-040
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:13:54

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3183.000
NP-237	0807-B	28-FEB-2010	4768.800	4768.799
CM-244	0910/0911-A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2355.549
 Energy Calibration Slope : 4.906727
 Energy Calibration Quadratic : 3.1722613E-04
 Energy Calibration Range : 2360 to 7713 KeV

The Energy Calibration : energy = 2355.549+(4.906727*Channel)+(3.1722613E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.808	3299.279	3.000000	0.7199996	57.73503	95.00000
NP-237	4436.136	4903.894	6.000000	1.439999	40.82483	95.00000
CM-244	5535.565	5886.878	96.000000	23.03999	10.20621	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:13:54
 Average Efficiency : 0.3230931
 Average Efficiency Error : 8.9092702E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	194.4828	28-FEB-2010	2988.808	3299.279	14753.00	0.3210947	1.3807181E-02	48.34544
NP-237	166.8174	28-FEB-2010	4436.136	4903.894	12953.00	0.3234969	1.6422754E-02	64.00413
CM-244	155.0100	28-FEB-2010	5535.565	5886.878	11367.00	0.3255583	1.6563144E-02	53.88326

Instrument : CHAMBER 041

Detector : 78205
 Standard ID : AE55-041
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:14:05

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.715
NP-237	0807-B	28-FEB-2010	4768.800	4768.680
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.847

Energy Calibration Zero : 2364.283
 Energy Calibration Slope : 4.953596
 Energy Calibration Quadratic : 3.3976510E-04
 Energy Calibration Range : 2369 to 7793 KeV

The Energy Calibration : energy = 2364.283+(4.953596*Channel)+(3.3976510E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.791	3302.487	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4436.678	4904.854	10.00000	2.399998	31.62278	95.00000
CM-244	5533.881	5884.420	95.00000	22.79999	10.25978	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:14:05
 Average Efficiency : 0.3327790
 Average Efficiency Error : 9.1675343E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	203.9034	28-FEB-2010	2988.791	3302.487	15730.00	0.3265573	1.4026052E-02	49.55199
NP-237	171.2268	28-FEB-2010	4436.678	4904.854	13823.00	0.3363137	1.7057342E-02	66.69939
CM-244	159.5796	28-FEB-2010	5533.881	5884.420	12167.00	0.3385442	1.7204376E-02	53.56832

Instrument : CHAMBER 042

Detector : 78793
 Standard ID : AE55-042
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 4-NOV-2009 08:33:13
 Calibration Count Time : 239.9998 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 4-NOV-2009 14:14:15

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	0810-B	28-FEB-2010	3183.000	3182.598
NP-237	0807-B	28-FEB-2010	4768.800	4768.622
CM-244	0910/0911-A	28-FEB-2010	5795.020	5794.973

Energy Calibration Zero : 2382.684
 Energy Calibration Slope : 4.893946
 Energy Calibration Quadratic : 3.5197593E-04
 Energy Calibration Range : 2388 to 7763 KeV

The Energy Calibration : energy = 2382.684+(4.893946*Channel)+(3.5197593E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 1-NOV-2009 19:30:31
 Background Count Time : 59999.99 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.964	3300.110	4.000000	0.9599994	50.00000	95.00000
NP-237	4432.825	4901.917	9.000000	2.159999	33.33334	95.00000
CM-244	5530.914	5884.067	116.0000	27.83998	9.284767	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 4-NOV-2009 14:14:15
 Average Efficiency : 0.3342218
 Average Efficiency Error : 9.2160990E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	188.7090	28-FEB-2010	2989.964	3300.110	14924.00	0.3347494	1.4391325E-02	47.46222
NP-237	159.6558	28-FEB-2010	4432.825	4901.917	12885.00	0.3362144	1.7069759E-02	62.70876
CM-244	150.5208	28-FEB-2010	5530.914	5884.067	11246.00	0.3315499	1.6871363E-02	52.78824

Instrument : CHAMBER 121

Detector : 75545
 Standard ID : AE55-005
 Standard Reference Date : 14-FEB-2008 09:35:18
 Calibration Analysis Date/Time : 18-NOV-2009 08:14:13
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:33:22

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.765
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2337.936
 Energy Calibration Slope : 4.942307
 Energy Calibration Quadratic : 2.8753869E-04
 Energy Calibration Range : 2343 to 7700 KeV

The Energy Calibration : energy = 2337.936+(4.942307*Channel)+(2.8753869E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:14:17
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.313	3302.511	1.000000	0.3000000	100.0000	95.00000
NP-237	4433.334	4901.898	2.000000	0.6000000	70.71068	95.00000
CM-244	5533.896	5885.024	6.000000	1.800000	40.82483	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:33:22
 Average Efficiency : 0.2473573
 Average Efficiency Error : 6.8073976E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	210.7452	28-FEB-2010	2990.313	3302.511	15217.00	0.2446291	1.0513312E-02	50.02479
NP-237	209.5938	28-FEB-2010	4433.334	4901.898	15770.00	0.2507932	1.2697713E-02	58.78762
CM-244	202.7478	28-FEB-2010	5533.896	5885.024	14097.00	0.2478902	1.2569192E-02	51.04265

Instrument : CHAMBER 122

Detector : 75546
 Standard ID : AE55-011
 Standard Reference Date : 14-FEB-2008 13:39:25
 Calibration Analysis Date/Time : 18-NOV-2009 08:14:17
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:33:39

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2333.121
 Energy Calibration Slope : 4.951138
 Energy Calibration Quadratic : 2.8171830E-04
 Energy Calibration Range : 2338 to 7698 KeV

The Energy Calibration : energy = 2333.121+(4.951138*Channel)+(2.8171830E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:14:23
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.580	3299.185	2.000000	0.6000000	70.71068	95.00000
NP-237	4436.363	4905.292	18.000000	5.4000000	23.57022	95.00000
CM-244	5532.326	5883.533	5.000000	1.5000000	44.72136	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:33:39
 Average Efficiency : 0.2526607
 Average Efficiency Error : 6.9489521E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	212.8284	28-FEB-2010	2991.580	3299.185	15866.00	0.2525603	1.0846268E-02	50.15437
NP-237	214.4868	28-FEB-2010	4436.363	4905.292	16139.00	0.2507320	1.2691127E-02	56.76629
CM-244	208.4184	28-FEB-2010	5532.326	5883.533	14895.00	0.2547987	1.2909920E-02	50.74705

Instrument : CHAMBER 123

Detector : 45-142V3
 Standard ID : AE55-006
 Standard Reference Date : 14-FEB-2008 09:35:18
 Calibration Analysis Date/Time : 18-NOV-2009 08:14:22
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:33:53

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.129

Energy Calibration Zero : 2373.252
 Energy Calibration Slope : 4.970928
 Energy Calibration Quadratic : 2.5967354E-04
 Energy Calibration Range : 2378 to 7736 KeV

The Energy Calibration : energy = 2373.252+(4.970928*Channel)+(2.5967354E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:14:28
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.604	3301.760	1.000000	0.3000000	100.0000	95.00000
NP-237	4434.252	4902.716	3.000000	0.9000000	57.73503	95.00000
CM-244	5533.878	5884.205	17.000000	5.1000000	24.25356	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:33:53
 Average Efficiency : 0.2603278
 Average Efficiency Error : 7.1630413E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	203.6952	28-FEB-2010	2988.604	3301.760	15566.00	0.2589007	1.1122215E-02	66.44209
NP-237	204.7038	28-FEB-2010	4434.252	4902.716	15697.00	0.2555905	1.2941348E-02	72.51353
CM-244	195.0060	28-FEB-2010	5533.878	5884.205	14643.00	0.2676548	1.3564456E-02	65.26345

Instrument : CHAMBER 124

Detector : 45-142V2
 Standard ID : AE55-012
 Standard Reference Date : 14-FEB-2008 13:39:25
 Calibration Analysis Date/Time : 18-NOV-2009 08:14:27
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:34:11

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2385.700
 Energy Calibration Slope : 5.030823
 Energy Calibration Quadratic : 2.4897119E-04
 Energy Calibration Range : 2391 to 7798 KeV

The Energy Calibration : energy = 2385.700+(5.030823*Channel)+(2.4897119E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:14:33
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2987.893	3299.314	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4432.634	4905.307	6.000000	1.800000	40.82483	95.00000
CM-244	5531.143	5884.210	7.000000	2.100000	37.79645	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:34:11
 Average Efficiency : 0.2595429
 Average Efficiency Error : 7.1389978E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	206.2200	28-FEB-2010	2987.893	3299.314	15727.00	0.2583798	1.1097851E-02	66.33055
NP-237	205.8930	28-FEB-2010	4432.634	4905.307	16250.00	0.2630527	1.3313570E-02	74.78283
CM-244	203.1954	28-FEB-2010	5531.143	5884.210	14692.00	0.2577755	1.3063119E-02	69.52264

Instrument : CHAMBER 138

Detector : 65877
 Standard ID : AE55-031
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:30
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:40:59

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.299
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.037

Energy Calibration Zero : 2378.122
 Energy Calibration Slope : 4.999195
 Energy Calibration Quadratic : 2.9128915E-04
 Energy Calibration Range : 2383 to 7803 KeV

The Energy Calibration : energy = 2378.122+(4.999195*Channel)+(2.9128915E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:15:45
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2992.359	3297.624	4.000000	1.200000	50.00000	95.00000
NP-237	4434.872	4902.951	26.000000	7.800000	19.61161	95.00000
CM-244	5530.666	5885.297	9.000000	2.700000	33.33334	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:40:59
 Average Efficiency : 0.2542060
 Average Efficiency Error : 7.0132604E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	193.6650	28-FEB-2010	2992.359	3297.624	14409.00	0.2520270	1.0842061E-02	57.19797
NP-237	162.9186	28-FEB-2010	4434.872	4902.951	12538.00	0.2563694	1.3021626E-02	60.08981
CM-244	153.1968	28-FEB-2010	5530.666	5885.297	10971.00	0.2551810	1.2989706E-02	57.60730

Instrument : CHAMBER 139

Detector : 76231
 Standard ID : AE55-026
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:35
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:41:12

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.019

Energy Calibration Zero : 2350.269
 Energy Calibration Slope : 4.923247
 Energy Calibration Quadratic : 3.1972880E-04
 Energy Calibration Range : 2355 to 7727 KeV

The Energy Calibration : energy = 2350.269+(4.923247*Channel)+(3.1972880E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:15:50
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.688	3302.273	2.000000	0.6000000	70.71068	95.00000
NP-237	4432.919	4902.312	25.000000	7.500000	20.00000	95.00000
CM-244	5530.906	5883.442	8.000000	2.400000	35.35534	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:41:12
 Average Efficiency : 0.2491344
 Average Efficiency Error : 7.3053949E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	199.5072	28-FEB-2010	2990.688	3302.273	14592.00	0.2477839	1.2558901E-02	54.20703
NP-237	168.0294	28-FEB-2010	4432.919	4902.312	12793.00	0.2536362	1.2878811E-02	52.46771
CM-244	160.5822	28-FEB-2010	5530.906	5883.442	11092.00	0.2462171	1.2530987E-02	49.40826

Instrument : CHAMBER 140

Detector : 78771
 Standard ID : AE55-032
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:39
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:41:26

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2341.379
 Energy Calibration Slope : 4.954690
 Energy Calibration Quadratic : 3.0003360E-04
 Energy Calibration Range : 2346 to 7730 KeV

The Energy Calibration : energy = 2341.379+(4.954690*Channel)+(3.0003360E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:15:55
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.559	3298.670	1.000000	0.3000000	100.0000	95.00000
NP-237	4433.640	4904.243	16.000000	4.800000	25.00000	95.00000
CM-244	5533.946	5886.857	2.000000	0.6000000	70.71068	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:41:26
 Average Efficiency : 0.2551176
 Average Efficiency Error : 7.0365276E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	195.2364	28-FEB-2010	2990.559	3298.670	14738.00	0.2557228	1.0996417E-02	45.42550
NP-237	165.9822	28-FEB-2010	4433.640	4904.243	12678.00	0.2545095	1.2924836E-02	53.64622
CM-244	153.7938	28-FEB-2010	5533.946	5886.857	10999.00	0.2548880	1.2974110E-02	52.32060

Instrument : CHAMBER 141

Detector : 76232
 Standard ID : AE55-027
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:43
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:41:39

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3182.673
NP-237	4341	28-FEB-2010	4768.800	4768.653
CM-244	4320A	28-FEB-2010	5795.020	5794.897

Energy Calibration Zero : 2360.567
 Energy Calibration Slope : 4.948405
 Energy Calibration Quadratic : 2.9607967E-04
 Energy Calibration Range : 2366 to 7738 KeV

The Energy Calibration : energy = 2360.567+(4.948405*Channel)+(2.9607967E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:15:59
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.766	3301.332	2.000000	0.6000000	70.71068	95.00000
NP-237	4433.993	4903.545	10.000000	3.000000	31.62278	95.00000
CM-244	5531.761	5883.798	6.000000	1.800000	40.82483	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:41:39
 Average Efficiency : 0.2554431
 Average Efficiency Error : 7.4935821E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	193.4238	28-FEB-2010	2988.766	3301.332	14292.00	0.2503223	1.2691112E-02	57.52324
NP-237	161.6154	28-FEB-2010	4433.993	4903.545	12568.00	0.2591544	1.3162418E-02	56.32706
CM-244	148.1754	28-FEB-2010	5531.761	5883.798	10692.00	0.2572228	1.3099612E-02	54.26664

Instrument : CHAMBER 142

Detector : 64261
 Standard ID : AE55-033
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:47
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:41:56

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2376.230
 Energy Calibration Slope : 4.966580
 Energy Calibration Quadratic : 3.0060238E-04
 Energy Calibration Range : 2381 to 7777 KeV

The Energy Calibration : energy = 2376.230+(4.966580*Channel)+(3.0060238E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:04
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.667	3300.258	1.000000	0.3000000	100.0000	95.00000
NP-237	4437.001	4903.079	26.000000	7.800000	19.61161	95.00000
CM-244	5533.701	5887.128	22.000000	6.600000	21.32007	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:41:56
 Average Efficiency : 0.2583063
 Average Efficiency Error : 7.1268436E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	192.4158	28-FEB-2010	2991.667	3300.258	14459.00	0.2545594	1.0950263E-02	56.98912
NP-237	161.7816	28-FEB-2010	4437.001	4903.079	12611.00	0.2596752	1.3188355E-02	65.52209
CM-244	147.2670	28-FEB-2010	5533.701	5887.128	10852.00	0.2624816	1.3364112E-02	55.63273

Instrument : CHAMBER 143

Detector : 65882
 Standard ID : AE55-028
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:51
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:42:11

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.019

Energy Calibration Zero : 2349.458
 Energy Calibration Slope : 4.977223
 Energy Calibration Quadratic : 2.6478578E-04
 Energy Calibration Range : 2354 to 7724 KeV

The Energy Calibration : energy = 2349.458+(4.977223*Channel)+(2.6478578E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:10
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.881	3299.612	10.00000	3.000000	31.62278	95.00000
NP-237	4434.630	4904.266	18.00000	5.400000	23.57022	95.00000
CM-244	5531.815	5883.100	11.00000	3.300000	30.15113	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:42:11
 Average Efficiency : 0.2435768
 Average Efficiency Error : 7.1454709E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	199.6542	28-FEB-2010	2990.881	3299.612	14261.00	0.2419440	1.2266767E-02	49.88798
NP-237	168.1992	28-FEB-2010	4434.630	4904.266	12424.00	0.2461090	1.2502169E-02	57.93009
CM-244	156.7614	28-FEB-2010	5531.815	5883.100	10677.00	0.2427591	1.2363405E-02	51.17299

Instrument : CHAMBER 144

Detector : 75551
 Standard ID : AE55-034
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:55
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:42:32

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2347.599
 Energy Calibration Slope : 4.937171
 Energy Calibration Quadratic : 3.0647003E-04
 Energy Calibration Range : 2353 to 7725 KeV

The Energy Calibration : energy = 2347.599+(4.937171*Channel)+(3.0647003E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:14
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.594	3301.779	4.000000	1.200000	50.00000	95.00000
NP-237	4433.735	4903.290	11.000000	3.300000	30.15113	95.00000
CM-244	5531.770	5884.080	9.000000	2.700000	33.33334	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:42:32
 Average Efficiency : 0.2509402
 Average Efficiency Error : 6.9225174E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	200.5488	28-FEB-2010	2989.594	3301.779	14579.00	0.2462479	1.0591129E-02	46.81347
NP-237	167.2962	28-FEB-2010	4433.735	4903.290	12833.00	0.2556286	1.2979213E-02	58.06931
CM-244	154.4388	28-FEB-2010	5531.770	5884.080	10977.00	0.2532673	1.2892167E-02	48.64608

Instrument : CHAMBER 145

Detector : 72526
 Standard ID : AE55-029
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 18-NOV-2009 08:15:59
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:42:45

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2352.373
 Energy Calibration Slope : 4.969179
 Energy Calibration Quadratic : 2.8356104E-04
 Energy Calibration Range : 2357 to 7738 KeV

The Energy Calibration : energy = 2352.373+(4.969179*Channel)+(2.8356104E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:19
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.032	3301.677	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4437.403	4902.601	7.000000	2.100000	37.79645	95.00000
CM-244	5531.625	5883.957	4.000000	1.200000	50.00000	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:42:45
 Average Efficiency : 0.2502206
 Average Efficiency Error : 7.3377313E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	201.5742	28-FEB-2010	2988.032	3301.677	14603.00	0.2454380	1.2439859E-02	54.61472
NP-237	169.7700	28-FEB-2010	4437.403	4902.601	12733.00	0.2499637	1.2693054E-02	54.42122
CM-244	154.8234	28-FEB-2010	5531.625	5883.957	11106.00	0.2557254	1.3014561E-02	53.58852

Instrument : CHAMBER 146

Detector : 72527
 Standard ID : AE55-035
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:04
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:43:01

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3182.984
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2350.006
 Energy Calibration Slope : 4.931883
 Energy Calibration Quadratic : 2.8951556E-04
 Energy Calibration Range : 2355 to 7704 KeV

The Energy Calibration : energy = 2350.006+(4.931883*Channel)+(2.8951556E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:23
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.037	3297.515	2.000000	0.6000000	70.71068	95.00000
NP-237	4436.256	4903.993	3.000000	0.9000000	57.73503	95.00000
CM-244	5534.939	5885.513	10.000000	3.0000000	31.62278	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:43:01
 Average Efficiency : 0.2521740
 Average Efficiency Error : 6.9548828E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	198.6666	28-FEB-2010	2991.037	3297.515	15016.00	0.2560428	1.1006447E-02	52.29327
NP-237	168.2934	28-FEB-2010	4436.256	4903.993	12722.00	0.2519628	1.2794692E-02	59.50724
CM-244	158.8128	28-FEB-2010	5534.939	5885.513	11023.00	0.2473175	1.2588392E-02	55.02975

Instrument : CHAMBER 147

Detector : 75550
 Standard ID : AE55-030
 Standard Reference Date : 15-FEB-2008 09:06:52
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:08
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:43:19

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.143

Energy Calibration Zero : 2342.758
 Energy Calibration Slope : 4.974234
 Energy Calibration Quadratic : 2.5187916E-04
 Energy Calibration Range : 2348 to 7700 KeV

The Energy Calibration : energy = 2342.758+(4.974234*Channel)+(2.5187916E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:28
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.625	3302.025	2.000000	0.6000000	70.71068	95.00000
NP-237	4434.897	4903.298	14.000000	4.200000	26.72612	95.00000
CM-244	5534.180	5884.256	5.000000	1.500000	44.72136	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:43:19
 Average Efficiency : 0.2451187
 Average Efficiency Error : 7.1902573E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	198.9792	28-FEB-2010	2988.625	3302.025	14282.00	0.2431632	1.2328269E-02	48.34250
NP-237	166.3758	28-FEB-2010	4434.897	4903.298	12254.00	0.2454244	1.2470047E-02	57.47663
CM-244	157.1856	28-FEB-2010	5534.180	5884.256	10884.00	0.2468401	1.2566833E-02	47.78167

Instrument : CHAMBER 148

Detector : 74429
 Standard ID : AE55-036
 Standard Reference Date : 18-FEB-2008 11:28:15
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:12
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:43:37

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.134
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2342.267
 Energy Calibration Slope : 4.959772
 Energy Calibration Quadratic : 2.7652038E-04
 Energy Calibration Range : 2347 to 7711 KeV

The Energy Calibration : energy = 2342.267+(4.959772*Channel)+(2.7652038E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:33
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.710	3299.671	3.000000	0.9000000	57.73503	95.00000
NP-237	4432.630	4901.757	17.000000	5.100000	24.25356	95.00000
CM-244	5534.227	5885.479	11.000000	3.300000	30.15113	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:43:37
 Average Efficiency : 0.2488421
 Average Efficiency Error : 6.8639922E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	201.3204	28-FEB-2010	2991.710	3299.671	14736.00	0.2479510	1.0662264E-02	49.91240
NP-237	167.4312	28-FEB-2010	4432.630	4901.757	12530.00	0.2493544	1.2665348E-02	49.84171
CM-244	156.4188	28-FEB-2010	5534.227	5885.479	10957.00	0.2495920	1.2705517E-02	53.03390

Instrument : CHAMBER 149

Detector : 33449
 Standard ID : AE55-037
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:16
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:43:50

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.207
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2394.373
 Energy Calibration Slope : 4.928394
 Energy Calibration Quadratic : 3.2330301E-04
 Energy Calibration Range : 2399 to 7780 KeV

The Energy Calibration : energy = 2394.373+(4.928394*Channel)+(3.2330301E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:38
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.436	3302.050	4.000000	1.200000	50.00000	95.00000
NP-237	4433.023	4902.650	6.000000	1.800000	40.82483	95.00000
CM-244	5531.642	5884.441	6.000000	1.800000	40.82483	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:43:50
 Average Efficiency : 0.2483255
 Average Efficiency Error : 6.8516782E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	197.7372	28-FEB-2010	2990.436	3302.050	14366.00	0.2460991	1.0587637E-02	68.92946
NP-237	167.1294	28-FEB-2010	4433.023	4902.650	12626.00	0.2517851	1.2787181E-02	66.29462
CM-244	154.7664	28-FEB-2010	5531.642	5884.441	10776.00	0.2481186	1.2634184E-02	65.00041

Instrument : CHAMBER 150

Detector : 75552
 Standard ID : AE55-043
 Standard Reference Date : 19-FEB-2008 00:32:27
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:20
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:44:03

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.870
CM-244	4320A	28-FEB-2010	5795.020	5795.095

Energy Calibration Zero : 2353.210
 Energy Calibration Slope : 4.970113
 Energy Calibration Quadratic : 2.7265481E-04
 Energy Calibration Range : 2358 to 7729 KeV

The Energy Calibration : energy = 2353.210+(4.970113*Channel)+(2.7265481E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:43
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.812	3302.301	1.000000	0.3000000	100.0000	95.00000
NP-237	4436.790	4906.431	6.000000	1.800000	40.82483	95.00000
CM-244	5534.181	5885.674	5.000000	1.500000	44.72136	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:44:03
 Average Efficiency : 0.2472161
 Average Efficiency Error : 6.8205683E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	197.7708	28-FEB-2010	2988.812	3302.301	14459.00	0.2476633	1.0653616E-02	51.90544
NP-237	168.7422	28-FEB-2010	4436.790	4906.431	12530.00	0.2474881	1.2570423E-02	58.37593
CM-244	156.3252	28-FEB-2010	5534.181	5885.674	10806.00	0.2463256	1.2542248E-02	51.11081

Instrument : CHAMBER 151

Detector : 75556
 Standard ID : AE55-038
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:24
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:44:18

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2346.500
 Energy Calibration Slope : 4.910573
 Energy Calibration Quadratic : 2.9834322E-04
 Energy Calibration Range : 2351 to 7688 KeV

The Energy Calibration : energy = 2346.500+(4.910573*Channel)+(2.9834322E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:48
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.916	3300.328	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4435.770	4902.424	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
CM-244	5532.149	5887.482	4.000000	1.200000	50.00000	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:44:18
 Average Efficiency : 0.2453373
 Average Efficiency Error : 6.7687407E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	200.1408	28-FEB-2010	2989.916	3300.328	14345.00	0.2428084	1.0446316E-02	54.48222
NP-237	170.0886	28-FEB-2010	4435.770	4902.424	12621.00	0.2473418	1.2561537E-02	57.49018
CM-244	157.7460	28-FEB-2010	5532.149	5887.482	10933.00	0.2469928	1.2573596E-02	57.42320

Instrument : CHAMBER 152

Detector : 76222
 Standard ID : AE55-044
 Standard Reference Date : 19-FEB-2008 00:32:27
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:28
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:44:34

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2341.313
 Energy Calibration Slope : 4.950351
 Energy Calibration Quadratic : 2.6349045E-04
 Energy Calibration Range : 2346 to 7687 KeV

The Energy Calibration : energy = 2341.313+(4.950351*Channel)+(2.6349045E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:53
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.311	3301.494	1.000000	0.3000000	100.0000	95.00000
NP-237	4435.919	4903.221	1.000000	0.3000000	100.0000	95.00000
CM-244	5532.929	5882.503	8.000000	2.400000	35.35534	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:44:34
 Average Efficiency : 0.2442460
 Average Efficiency Error : 6.7417030E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	194.4510	28-FEB-2010	2989.311	3301.494	14186.00	0.2471355	1.0634735E-02	48.77798
NP-237	166.6248	28-FEB-2010	4435.919	4903.221	12052.00	0.2410947	1.2253158E-02	55.60851
CM-244	155.8290	28-FEB-2010	5532.929	5882.503	10651.00	0.2435444	1.2403901E-02	54.29796

Instrument : CHAMBER 153

Detector : 76223
 Standard ID : AE55-039
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:32
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:44:49

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2333.060
 Energy Calibration Slope : 4.940231
 Energy Calibration Quadratic : 2.9184026E-04
 Energy Calibration Range : 2338 to 7698 KeV

The Energy Calibration : energy = 2333.060+(4.940231*Channel)+(2.9184026E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:16:57
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2990.256	3302.449	4.000000	1.200000	50.00000	95.00000
NP-237	4433.519	4902.304	10.000000	3.000000	31.62278	95.00000
CM-244	5534.705	5886.108	7.000000	2.100000	37.79645	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:44:49
 Average Efficiency : 0.2523201
 Average Efficiency Error : 6.9638924E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	192.2418	28-FEB-2010	2990.256	3302.449	14256.00	0.2511956	1.0808469E-02	45.93060
NP-237	159.1506	28-FEB-2010	4433.519	4902.304	12260.00	0.2567171	1.3043687E-02	53.80658
CM-244	151.7142	28-FEB-2010	5534.705	5886.108	10631.00	0.2496969	1.2717671E-02	49.20813

Instrument : CHAMBER 154

Detector : 76224
 Standard ID : AE55-045
 Standard Reference Date : 19-FEB-2008 00:32:27
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:36
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:45:05

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.070

Energy Calibration Zero : 2339.159
 Energy Calibration Slope : 4.959938
 Energy Calibration Quadratic : 2.7331465E-04
 Energy Calibration Range : 2344 to 7705 KeV

The Energy Calibration : energy = 2339.159+(4.959938*Channel)+(2.7331465E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:17:02
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.570	3301.543	4.000000	1.200000	50.00000	95.00000
NP-237	4434.233	4903.159	8.000000	2.400000	35.35534	95.00000
CM-244	5535.283	5886.305	8.000000	2.400000	35.35534	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:45:05
 Average Efficiency : 0.2549808
 Average Efficiency Error : 7.0385467E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	186.9936	28-FEB-2010	2988.570	3301.543	14056.00	0.2546200	1.0958767E-02	48.42456
NP-237	160.8066	28-FEB-2010	4434.233	4903.159	12136.00	0.2515155	1.2781434E-02	55.84683
CM-244	145.8384	28-FEB-2010	5535.283	5886.305	10609.00	0.2592019	1.3202298E-02	48.53014

Instrument : CHAMBER 155

Detector : 75553
 Standard ID : AE55-040
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 18-NOV-2009 08:16:40
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 18-NOV-2009 13:45:22

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.199
NP-237	4341	28-FEB-2010	4768.800	4768.862
CM-244	4320A	28-FEB-2010	5795.020	5795.045

Energy Calibration Zero : 2368.783
 Energy Calibration Slope : 4.941553
 Energy Calibration Quadratic : 3.2590213E-04
 Energy Calibration Range : 2374 to 7771 KeV

The Energy Calibration : energy = 2368.783+(4.941553*Channel)+(3.2590213E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 15-NOV-2009 17:17:06
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2991.569	3299.187	5.000000	1.500000	44.72136	95.00000
NP-237	4433.980	4905.235	4.000000	1.200000	50.00000	95.00000
CM-244	5531.102	5885.121	11.000000	3.300000	30.15113	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 18-NOV-2009 13:45:22
 Average Efficiency : 0.2585027
 Average Efficiency Error : 7.1280575E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	194.4828	28-FEB-2010	2991.569	3299.187	14883.00	0.2592175	1.1144735E-02	52.61243
NP-237	166.8174	28-FEB-2010	4433.980	4905.235	12881.00	0.2573634	1.3066485E-02	59.47372
CM-244	155.0100	28-FEB-2010	5531.102	5885.121	11253.00	0.2586618	1.3161110E-02	51.99247

Instrument : CHAMBER 201

Detector : 78902
 Standard ID : AE55-039
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:38
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:32:21

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3182.826
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2358.063
 Energy Calibration Slope : 4.962597
 Energy Calibration Quadratic : 2.9971759E-04
 Energy Calibration Range : 2363 to 7754 KeV

The Energy Calibration : energy = 2358.063+(4.962597*Channel)+(2.9971759E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:20
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.109	3301.625	1.000000	0.3000000	100.0000	95.00000
NP-237	4432.695	4903.768	2.000000	0.6000000	70.71068	95.00000
CM-244	5534.085	5887.332	18.000000	5.4000000	23.57022	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:32:21
 Average Efficiency : 0.2585327
 Average Efficiency Error : 7.1318159E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	192.2418	28-FEB-2010	2988.109	3301.625	14633.00	0.2576852	1.1082157E-02	44.32004
NP-237	159.1506	28-FEB-2010	4432.695	4903.768	12467.00	0.2610968	1.3262650E-02	56.19632
CM-244	151.7142	28-FEB-2010	5534.085	5887.332	10984.00	0.2572168	1.3093179E-02	45.13528

Instrument : CHAMBER 202

Detector : 78903
 Standard ID : AE55-045
 Standard Reference Date : 19-FEB-2008 00:32:27
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:42
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:23:05

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2348.852
 Energy Calibration Slope : 4.968003
 Energy Calibration Quadratic : 2.7484546E-04
 Energy Calibration Range : 2354 to 7724 KeV

The Energy Calibration : energy = 2348.852+(4.968003*Channel)+(2.7484546E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:24
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2989.259	3297.622	0.0000000E+00	0.0000000E+00	0.0000000E+00	95.00000
NP-237	4437.128	4901.565	1.000000	0.3000000	100.0000	95.00000
CM-244	5534.664	5886.243	21.00000	6.300000	21.82179	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:23:05
 Average Efficiency : 0.2674317
 Average Efficiency Error : 7.3762531E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	186.9936	28-FEB-2010	2989.259	3297.622	14760.00	0.2672196	1.1490363E-02	44.90806
NP-237	160.8066	28-FEB-2010	4437.128	4901.565	12659.00	0.2624004	1.3325701E-02	51.42311
CM-244	145.8384	28-FEB-2010	5534.664	5886.243	11217.00	0.2732188	1.3902702E-02	48.56720

Instrument : CHAMBER 203

Detector : 78905
 Standard ID : AE55-040
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:47
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:23:13

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.801
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2360.050
 Energy Calibration Slope : 4.948787
 Energy Calibration Quadratic : 3.0506149E-04
 Energy Calibration Range : 2365 to 7747 KeV

The Energy Calibration : energy = 2360.050+(4.948787*Channel)+(3.0506149E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:28
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.440	3301.203	8.000000	2.400000	35.35534	95.00000
NP-237	4435.134	4905.455	8.000000	2.400000	35.35534	95.00000
CM-244	5534.910	5882.378	26.000000	7.800000	19.61161	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:23:13
 Average Efficiency : 0.2555374
 Average Efficiency Error : 7.0476462E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	194.4828	28-FEB-2010	2988.440	3301.203	14732.00	0.2564026	1.1025699E-02	49.14829
NP-237	166.8174	28-FEB-2010	4435.134	4905.455	12659.00	0.2529034	1.2843485E-02	57.22518
CM-244	155.0100	28-FEB-2010	5534.910	5882.378	11218.00	0.2570517	1.3080092E-02	55.29720

Instrument : CHAMBER 204

Detector : 78907
 Standard ID : AE55-046
 Standard Reference Date : 19-FEB-2008 19:35:48
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:51
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:23:21

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2357.577
 Energy Calibration Slope : 4.971226
 Energy Calibration Quadratic : 2.7621095E-04
 Energy Calibration Range : 2363 to 7738 KeV

The Energy Calibration : energy = 2357.577+(4.971226*Channel)+(2.7621095E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:33
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.337	3301.930	11.00000	3.300000	30.15113	95.00000
NP-237	4437.014	4901.750	13.00000	3.900000	27.73501	95.00000
CM-244	5535.287	5887.123	32.00000	9.600000	17.67767	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:23:21
 Average Efficiency : 0.2522539
 Average Efficiency Error : 6.9599389E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	194.7474	28-FEB-2010	2988.337	3301.930	14610.00	0.2539103	1.0920225E-02	51.81709
NP-237	164.6658	28-FEB-2010	4437.014	4901.750	12297.00	0.2488496	1.2643373E-02	55.31627
CM-244	151.3824	28-FEB-2010	5535.287	5887.123	10807.00	0.2534880	1.2907310E-02	54.27145

Instrument : CHAMBER 205

Detector : 78908
 Standard ID : AE55-041
 Standard Reference Date : 18-FEB-2008 15:31:47
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:55
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:23:29

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.799
CM-244	4320A	28-FEB-2010	5795.020	5795.021

Energy Calibration Zero : 2363.378
 Energy Calibration Slope : 4.959655
 Energy Calibration Quadratic : 2.9583176E-04
 Energy Calibration Range : 2368 to 7752 KeV

The Energy Calibration : energy = 2363.378+(4.959655*Channel)+(2.9583176E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:36
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2987.957	3301.178	2.000000	0.6000000	70.71068	95.00000
NP-237	4436.162	4901.400	3.000000	0.9000000	57.73503	95.00000
CM-244	5530.781	5883.461	11.000000	3.3000000	30.15113	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:23:29
 Average Efficiency : 0.2537003
 Average Efficiency Error : 6.9936696E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	203.9034	28-FEB-2010	2987.957	3301.178	15153.00	0.2515763	1.0812583E-02	48.70498
NP-237	171.2268	28-FEB-2010	4436.162	4901.400	13162.00	0.2562121	1.3003838E-02	57.91314
CM-244	159.5796	28-FEB-2010	5530.781	5883.461	11419.00	0.2542548	1.2933665E-02	55.11108

Instrument : CHAMBER 206

Detector : 78909
 Standard ID : AE55-047
 Standard Reference Date : 19-FEB-2008 00:32:27
 Calibration Analysis Date/Time : 23-OCT-2009 07:44:59
 Calibration Count Time : 300.0000 min.

Subsection 1: Energy Calibration

Energy/Channel Equation

$$\text{energy} = \text{Cal_Zero} + (\text{e1} * \text{C}) + (\text{e2} * \text{C}^2)$$

where : Cal_Zero = Energy Calibration Zero
 e1 = Energy Calibration Slope
 e2 = Energy Calibration Quadratic
 C = Channel

Energy Calibration Date/Time : 23-OCT-2009 13:23:40

Cal. Isotopes	Source Id	Expiration Date	Standard Energy (KeV)	Actual Energy (KeV)
GD-148	6445-278	28-FEB-2010	3183.000	3183.000
NP-237	4341	28-FEB-2010	4768.800	4768.800
CM-244	4320A	28-FEB-2010	5795.020	5795.020

Energy Calibration Zero : 2359.196
 Energy Calibration Slope : 4.958711
 Energy Calibration Quadratic : 2.8597465E-04
 Energy Calibration Range : 2364 to 7737 KeV

The Energy Calibration : energy = 2359.196+(4.958711*Channel)+(2.8597465E-04*Channel^2)

Subsection 2: Background Calibration

Background Analysis Date/Time : 18-OCT-2009 17:41:40
 Background Count Time : 60000.00 sec.

Cal. Isotopes	Start Energy (KeV)	End Energy (KeV)	Counts in 1000 min	Counts during Cal	% Error	% Confidence
GD-148	2988.534	3301.541	2.000000	0.6000000	70.71068	95.00000
NP-237	4435.146	4904.801	2.000000	0.6000000	70.71068	95.00000
CM-244	5532.900	5884.749	16.000000	4.8000000	25.00000	95.00000

Subsection 3: Efficiency Calibration

Efficiency Calibration Date/Time : 23-OCT-2009 13:23:40
 Average Efficiency : 0.2554192
 Average Efficiency Error : 7.0435088E-03
 % Confidence : 95.00000

Cal. Isotopes	DPM	Exp. Date	Start Engy (KeV)	End Engy (KeV)	Counts	EFF.	EFF Err	Resolution
GD-148	197.4804	28-FEB-2010	2988.534	3301.541	15042.00	0.2578534	1.1083830E-02	50.51255
NP-237	168.3948	28-FEB-2010	4435.146	4904.801	12737.00	0.2521141	1.2802137E-02	55.90126
CM-244	154.6032	28-FEB-2010	5532.900	5884.749	11117.00	0.2554787	1.3001921E-02	55.72073

GAS FLOW PROPORTIONAL COUNTERS

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414
(843)556-8171

Gas Flow Proportional Counter Calibration Package

Method: Re-228 (PL)

- 1) Is all calibration standard information enclosed for:
primary standard certificate?
secondary standard(s) documentation?
standard preparation information?
standard < 1 Year old or verified?
- 2) Are the detector graphs included?
beta absorption curves?
beta plateau?
- 3) Is the raw count data included for:
the plateau generation?
the absorption curve generation?
the calibration verification?
the crosstalk calculations?
- 4) Are the calibration verification calculations included?
are verification recoveries 100% +/- 25%
- 5) Is the method Carrier Standardization included?

YES	NO	Comments
<input checked="" type="checkbox"/>		
		Average Efficiency
<input checked="" type="checkbox"/>		
		N/A

Prepared By: D.L.

Date: 7/2/09

Reviewed By: Angela G.

Date: 7/2/09

Effective Date: 7/2/09

Ra-228 Calibration PROTEAN Detectors

Detector	Source	Separation date	Count date	Ac-228 decay	Spike Vol. Ra-228 (mL)	Std. Act. Ra-228 dpm/mL	Standard Nominal	Separation time Decay Corrected Volume corrected				Ra-228 eff (cpm/dpm)
								raw beta	ct. time (min)	Beta cpm	corrected* cpm	
1A	1	7/1/09 10:45	7/1/2009 13:36	0.7249	1.5	6363.2	9544.8	13584	3	4521.3	6237.434348	0.6535
1A	2	7/1/09 10:45	7/1/2009 13:52	0.7032	1.5	6363.2	9544.8	12775	3	4258.3	6055.521583	0.6344
1A	3	7/1/09 10:45	7/1/2009 13:48	0.7083	1.5	6363.2	9544.8	12750	3	4250.0	6000.085083	0.6286
1A	4	7/1/09 10:45	7/1/2009 13:41	0.7170	1.5	6363.2	9544.8	12410	3	4136.7	5769.683602	0.6045
1B	1	7/1/09 10:45	7/1/2009 13:41	0.7174	1.5	6363.2	9544.8	13292	3	4430.7	6176.07771	0.6471
1B	2	7/1/09 10:45	7/1/2009 13:36	0.7246	1.5	6363.2	9544.8	13274	3	4424.7	6106.181463	0.6397
1B	3	7/1/09 10:45	7/1/2009 13:52	0.7031	1.5	6363.2	9544.8	12869	3	4233.0	6020.43969	0.6308
1B	4	7/1/09 10:45	7/1/2009 13:48	0.7082	1.5	6363.2	9544.8	12072	3	4024.0	5682.267909	0.5953
1C	1	7/1/09 10:45	7/1/2009 13:48	0.7085	1.5	6363.2	9544.8	12813	3	4271.0	6028.410186	0.6316
1C	2	7/1/09 10:45	7/1/2009 13:41	0.7172	1.5	6363.2	9544.8	12979	3	4326.3	6032.15531	0.6320
1C	3	7/1/09 10:45	7/1/2009 13:36	0.7245	1.5	6363.2	9544.8	12755	3	4251.7	5668.722998	0.6149
1C	4	7/1/09 10:45	7/1/2009 13:52	0.7030	1.5	6363.2	9544.8	11917	3	3972.3	5650.765354	0.5920
1D	1	7/1/09 10:45	7/1/2009 13:52	0.7033	1.5	6363.2	9544.8	12473	3	4157.7	5911.258105	0.6193
1D	2	7/1/09 10:45	7/1/2009 13:48	0.7084	1.5	6363.2	9544.8	12484	3	4161.3	5874.170562	0.6154
1D	3	7/1/09 10:45	7/1/2009 13:41	0.7171	1.5	6363.2	9544.8	12289	3	4096.3	5712.363902	0.5985
1D	4	7/1/09 10:45	7/1/2009 13:36	0.7243	1.5	6363.2	9544.8	12115	3	4038.3	5575.474435	0.5841
2A	1	7/1/09 10:45	7/1/2009 13:57	0.6960	1.5	6363.2	9544.8	12499	3	4166.3	5988.085459	0.6272
2A	2	7/1/09 10:45	7/1/2009 14:15	0.6728	1.5	6363.2	9544.8	12103	3	4034.3	5996.8905	0.6283
2A	3	7/1/09 10:45	7/1/2009 14:09	0.6815	1.5	6363.2	9544.8	11988	3	3989.3	5854.110901	0.6133
2A	4	7/1/09 10:45	7/1/2009 14:02	0.6899	1.5	6363.2	9544.8	11855	3	3951.7	5728.227222	0.6001
2B	1	7/1/09 10:45	7/1/2009 14:02	0.6903	1.5	6363.2	9544.8	12471	3	4157.0	6022.286434	0.6309
2B	2	7/1/09 10:45	7/1/2009 13:57	0.6958	1.5	6363.2	9544.8	12492	3	4104.0	5984.232843	0.6270
2B	3	7/1/09 10:45	7/1/2009 14:15	0.6727	1.5	6363.2	9544.8	11892	3	3984.0	5892.884561	0.6174
2B	4	7/1/09 10:45	7/1/2009 14:09	0.6814	1.5	6363.2	9544.8	11539	3	3846.3	5644.974311	0.5914
2C	1	7/1/09 10:45	7/1/2009 14:08	0.6817	1.5	6363.2	9544.8	12050	3	4016.7	5892.005142	0.6173
2C	2	7/1/09 10:45	7/1/2009 14:02	0.6901	1.5	6363.2	9544.8	11914	3	3971.3	5754.571355	0.6029
2C	3	7/1/09 10:45	7/1/2009 13:58	0.6957	1.5	6363.2	9544.8	11994	3	3998.0	5746.92868	0.6021
2C	4	7/1/09 10:45	7/1/2009 14:15	0.6726	1.5	6363.2	9544.8	10889	3	3629.7	5396.37168	0.5854
2D	1	7/1/09 10:45	7/1/2009 14:15	0.6729	1.5	6363.2	9544.8	12010	3	4003.3	5949.493049	0.6233
2D	2	7/1/09 10:45	7/1/2009 14:08	0.6816	1.5	6363.2	9544.8	12124	3	4041.3	5929.303014	0.6212
2D	3	7/1/09 10:45	7/1/2009 14:02	0.6900	1.5	6363.2	9544.8	12168	3	4056.0	5878.360714	0.6159
2D	4	7/1/09 10:45	7/1/2009 13:58	0.6954	1.5	6363.2	9544.8	11692	3	3897.3	5604.156523	0.5871
3A	1	7/1/09 10:45	7/1/2009 14:19	0.6675	1.5	6363.2	9544.8	11194	3	3731.3	5589.748519	0.5856
3A	2	7/1/09 10:45	7/1/2009 14:35	0.6482	1.5	6363.2	9544.8	14227	4	3556.8	5486.792678	0.5748
3A	3	7/1/09 10:45	7/1/2009 14:30	0.6548	1.5	6363.2	9544.8	14180	4	3545.0	5414.108112	0.5672
3A	4	7/1/09 10:45	7/1/2009 14:25	0.6608	1.5	6363.2	9544.8	13754	4	3438.5	5203.464549	0.5452
3B	1	7/1/09 10:45	7/1/2009 14:25	0.6612	1.5	6363.2	9544.8	15370	4	3842.5	5811.010789	0.6088
3B	2	7/1/09 10:45	7/1/2009 14:20	0.6673	1.5	6363.2	9544.8	11695	3	3898.3	5842.303251	0.6121
3B	3	7/1/09 10:45	7/1/2009 14:35	0.6481	1.5	6363.2	9544.8	14905	4	3726.3	5749.171166	0.6023
3B	4	7/1/09 10:45	7/1/2009 14:30	0.6547	1.5	6363.2	9544.8	14220	4	3555.0	5430.231301	0.5689
3C	1	7/1/09 10:45	7/1/2009 14:29	0.6552	1.5	6363.2	9544.8	15644	4	3911.0	5969.527404	0.6254
3C	2	7/1/09 10:45	7/1/2009 14:25	0.6611	1.5	6363.2	9544.8	15984	4	3991.0	6036.911214	0.6325
3C	3	7/1/09 10:45	7/1/2009 14:20	0.6672	1.5	6363.2	9544.8	11701	3	3900.3	5846.033242	0.6125
3C	4	7/1/09 10:45	7/1/2009 14:35	0.6480	1.5	6363.2	9544.8	14729	4	3682.3	5682.352456	0.5953
3D	1	7/1/09 10:45	7/1/2009 14:35	0.6484	1.5	6363.2	9544.8	15152	4	3788.0	5842.430209	0.6121
3D	2	7/1/09 10:45	7/1/2009 14:30	0.6550	1.5	6363.2	9544.8	15168	4	3792.0	5789.343603	0.6065
3D	3	7/1/09 10:45	7/1/2009 14:25	0.6610	1.5	6363.2	9544.8	15295	4	3823.8	5785.011122	0.6061
3D	4	7/1/09 10:45	7/1/2009 14:20	0.6670	1.5	6363.2	9544.8	10942	3	3647.3	5468.022172	0.5729
4A	1	7/1/09 10:45	7/1/2009 14:40	0.6418	1.5	6363.2	9544.8	15298	4	3824.5	5959.288371	0.6243
4A	2	7/1/09 10:45	7/1/2009 15:00	0.6187	1.5	6363.2	9544.8	14897	4	3724.3	6019.957238	0.6307
4A	3	7/1/09 10:45	7/1/2009 14:53	0.6266	1.5	6363.2	9544.8	15050	4	3762.5	6005.095127	0.6291
4A	4	7/1/09 10:45	7/1/2009 14:48	0.6325	1.5	6363.2	9544.8	14462	4	3615.5	5715.951787	0.5989
4B	1	7/1/09 10:45	7/1/2009 14:48	0.6329	1.5	6363.2	9544.8	15335	4	3833.8	6057.768128	0.6347
4B	2	7/1/09 10:45	7/1/2009 14:41	0.6416	1.5	6363.2	9544.8	15513	4	3878.3	6044.745331	0.6333
4B	3	7/1/09 10:45	7/1/2009 15:00	0.6186	1.5	6363.2	9544.8	14521	4	3630.3	5888.56525	0.6148
4B	4	7/1/09 10:45	7/1/2009 14:53	0.6265	1.5	6363.2	9544.8	14328	4	3582.0	5715.747589	0.5990
4C	1	7/1/09 10:45	7/1/2009 14:53	0.6268	1.5	6363.2	9544.8	14733	4	3683.3	5876.583259	0.6157
4C	2	7/1/09 10:45	7/1/2009 14:48	0.6327	1.5	6363.2	9544.8	14902	4	3725.5	5888.011911	0.6169
4C	3	7/1/09 10:45	7/1/2009 14:41	0.6414	1.5	6363.2	9544.8	14856	4	3714.0	5790.010642	0.6066
4C	4	7/1/09 10:45	7/1/2009 15:00	0.6185	1.5	6363.2	9544.8	13733	4	3433.3	5550.795964	0.5816
4D	1	7/1/09 10:45	7/1/2009 15:00	0.6188	1.5	6363.2	9544.8	14167	4	3541.8	5723.884149	0.5997
4D	2	7/1/09 10:45	7/1/2009 14:53	0.6267	1.5	6363.2	9544.8	14204	4	3551.0	5866.467573	0.5937
4D	3	7/1/09 10:45	7/1/2009 14:48	0.6326	1.5	6363.2	9544.8	14131	4	3532.8	5584.07765	0.5850
4D	4	7/1/09 10:45	7/1/2009 14:41	0.6413	1.5	6363.2	9544.8	13978	4	3494.5	5449.182717	0.5709
5A	1	7/1/09 10:45	7/1/2009 15:06	0.6112	1.5	6363.2	9544.8	14870	4	3717.5	6082.165089	0.6372
5A	2	7/1/09 10:45	7/1/2009 15:21	0.5943	1.5	6363.2	9544.8	14487	4	3821.8	6094.223373	0.6385
5A	3	7/1/09 10:45	7/1/2009 15:17	0.5996	1.5	6363.2	9544.8	14259	4	3564.8	5945.170793	0.6229
5A	4	7/1/09 10:45	7/1/2009 15:12	0.6047	1.5	6363.2	9544.8	13957	4	3489.3	5770.592799	0.6046
5B	1	7/1/09 10:45	7/1/2009 15:12	0.6050	1.5	6363.2	9544.8	14869	4	3717.3	6144.005028	0.6437
5B	2	7/1/09 10:45	7/1/2009 15:06	0.6111	1.5	6363.2	9544.8	14821	4	3705.3	6063.072791	0.6352
5B	3	7/1/09 10:45	7/1/2009 15:21	0.5942	1.5	6363.2	9544.8	14289	4	3572.3	6011.872812	0.6299
5B	4	7/1/09 10:45	7/1/2009 15:17									

5C	4	7/1/09 10:45	7/1/2009 15:21	0.5941	1.5	6363.2	9544.8	13831	4	3457.8	5819.905873	0.6097	0.6368
5D	1	7/1/09 10:45	7/1/2009 15:21	0.5943	1.5	6363.2	9544.8	14321	4	3580.3	6024.014899	0.6311	
5D	2	7/1/09 10:45	7/1/2009 15:17	0.5993	1.5	6363.2	9544.8	14642	4	3680.5	6107.538025	0.6399	
5D	3	7/1/09 10:45	7/1/2009 15:12	0.6048	1.5	6363.2	9544.8	14443	4	3610.8	5970.409434	0.6255	Average EFF 0.6237
5D	4	7/1/09 10:45	7/1/2009 15:07	0.6107	1.5	6363.2	9544.8	13954	4	3488.5	5711.973074	0.5984	
6A	1	7/1/09 10:45	7/1/2009 15:27	0.5885	1.5	6363.2	9544.8	14018	4	3504.5	5955.42076	0.6239	
6A	2	7/1/09 10:45	7/1/2009 15:40	0.5735	1.5	6363.2	9544.8	12283	3.5	3509.4	6118.819734	0.6411	
6A	3	7/1/09 10:45	7/1/2009 15:36	0.5779	1.5	6363.2	9544.8	12111	3.5	3480.3	5987.187856	0.6273	Average EFF 0.6221
6A	4	7/1/09 10:45	7/1/2009 15:32	0.5826	1.5	6363.2	9544.8	11598	3.5	3313.7	5687.952648	0.5959	
6B	1	7/1/09 10:45	7/1/2009 15:32	0.5824	1.5	6363.2	9544.8	12151	3.5	3471.7	5961.398905	0.6246	
6B	2	7/1/09 10:45	7/1/2009 15:27	0.5885	1.5	6363.2	9544.8	14371	4	3592.8	6105.389624	0.6397	
6B	3	7/1/09 10:45	7/1/2009 15:40	0.5734	1.5	6363.2	9544.8	11705	3.5	3344.3	5831.983307	0.6110	Average EFF
6B	4	7/1/09 10:45	7/1/2009 15:36	0.5779	1.5	6363.2	9544.8	11388	3.5	3253.7	5630.295163	0.5899	0.6163
6C	1	7/1/09 10:45	7/1/2009 15:36	0.5778	1.5	6363.2	9544.8	12161	3.5	3474.6	6013.224586	0.6300	
6C	2	7/1/09 10:45	7/1/2009 15:32	0.5821	1.5	6363.2	9544.8	12063	3.5	3452.3	5930.638446	0.6213	
6C	3	7/1/09 10:45	7/1/2009 15:27	0.5883	1.5	6363.2	9544.8	13638	4	3409.5	5795.433731	0.6072	Average EFF 0.6111
6C	4	7/1/09 10:45	7/1/2009 15:40	0.5733	1.5	6363.2	9544.8	11218	3.5	3205.1	5590.212659	0.5857	
6D	1	7/1/09 10:45	7/1/2009 15:40	0.5732	1.5	6363.2	9544.8	11987	3.5	3424.9	5974.547886	0.6259	
6D	2	7/1/09 10:45	7/1/2009 15:36	0.5777	1.5	6363.2	9544.8	12183	3.5	3480.9	6025.235519	0.6313	
6D	3	7/1/09 10:45	7/1/2009 15:32	0.5819	1.5	6363.2	9544.8	11882	3.5	3394.9	5833.810262	0.6112	Average EFF 0.6120
6D	4	7/1/09 10:45	7/1/2009 15:27	0.5881	1.5	6363.2	9544.8	13018	4	3254.5	5533.699914	0.5798	
7A	1	7/1/09 10:45	7/1/2009 15:46	0.5673	1.5	6363.2	9544.8	12007	3.5	3430.6	6047.285806	0.6336	
7A	2	7/1/09 10:45	7/1/2009 16:00	0.5525	1.5	6363.2	9544.8	11655	3.5	3330.0	6027.308696	0.6315	
7A	3	7/1/09 10:45	7/1/2009 15:56	0.5569	1.5	6363.2	9544.8	11445	3.5	3270.0	5871.972756	0.6152	Average EFF 0.6180
7A	4	7/1/09 10:45	7/1/2009 15:50	0.5627	1.5	6363.2	9544.8	11121	3.5	3174.7	5646.894018	0.5916	
7B	1	7/1/09 10:45	7/1/2009 15:51	0.5622	1.5	6363.2	9544.8	11968	3.5	3419.4	6082.664171	0.6373	
7B	2	7/1/09 10:45	7/1/2009 15:46	0.5673	1.5	6363.2	9544.8	12050	3.5	3442.9	6069.322745	0.6359	
7B	3	7/1/09 10:45	7/1/2009 16:00	0.5524	1.5	6363.2	9544.8	11675	3.5	3335.7	6038.785014	0.6327	Average EFF
7B	4	7/1/09 10:45	7/1/2009 15:56	0.5587	1.5	6363.2	9544.8	11271	3.5	3220.3	5784.331251	0.6060	0.6280
7C	1	7/1/09 10:45	7/1/2009 15:56	0.5566	1.5	6363.2	9544.8	11781	3.5	3366.0	6047.202464	0.6336	
7C	2	7/1/09 10:45	7/1/2009 15:51	0.5621	1.5	6363.2	9544.8	11760	3.5	3380.0	5978.073192	0.6263	
7C	3	7/1/09 10:45	7/1/2009 15:46	0.5670	1.5	6363.2	9544.8	11766	3.5	3361.7	5928.878357	0.6212	Average EFF 0.6178
7C	4	7/1/09 10:45	7/1/2009 16:00	0.5523	1.5	6363.2	9544.8	10888	3.5	3110.9	5632.598965	0.5901	
7D	1	7/1/09 10:45	7/1/2009 16:00	0.5522	1.5	6363.2	9544.8	11605	3.5	3315.7	6004.271132	0.6291	
7D	2	7/1/09 10:45	7/1/2009 15:56	0.5565	1.5	6363.2	9544.8	11920	3.5	3405.7	6119.509991	0.6411	
7D	3	7/1/09 10:45	7/1/2009 15:51	0.5619	1.5	6363.2	9544.8	11933	3.5	3409.4	6067.346561	0.6357	Average EFF 0.6257
7D	4	7/1/09 10:45	7/1/2009 15:46	0.5668	1.5	6363.2	9544.8	11305	3.5	3230.0	5698.36602	0.5970	
8A	1	7/1/09 10:45	7/1/2009 16:06	0.5466	1.5	6363.2	9544.8	11673	3.5	3335.1	6101.651756	0.6393	
8A	2	7/1/09 10:45	7/1/2009 16:19	0.5333	1.5	6363.2	9544.8	11172	3.5	3192.0	5985.379105	0.6271	
8A	3	7/1/09 10:45	7/1/2009 16:15	0.5377	1.5	6363.2	9544.8	11258	3.5	3216.6	5982.329368	0.6268	Average EFF 0.6247
8A	4	7/1/09 10:45	7/1/2009 16:10	0.5424	1.5	6363.2	9544.8	10977	3.5	3136.3	5782.059146	0.6058	
8B	1	7/1/09 10:45	7/1/2009 16:10	0.5423	1.5	6363.2	9544.8	11583	3.5	3309.4	6102.412618	0.6393	
8B	2	7/1/09 10:45	7/1/2009 16:06	0.5466	1.5	6363.2	9544.8	11758	3.5	3359.4	6146.082528	0.6439	
8B	3	7/1/09 10:45	7/1/2009 16:19	0.5332	1.5	6363.2	9544.8	11499	3.5	3285.4	6161.727069	0.6456	Average EFF 0.6332
8B	4	7/1/09 10:45	7/1/2009 16:15	0.5376	1.5	6363.2	9544.8	10844	3.5	3098.3	5763.600098	0.6038	
8C	1	7/1/09 10:45	7/1/2009 16:15	0.5375	1.5	6363.2	9544.8	11539	3.5	3296.9	6133.762218	0.6426	
8C	2	7/1/09 10:45	7/1/2009 16:10	0.5422	1.5	6363.2	9544.8	11774	3.5	3364.0	6204.011354	0.6500	
8C	3	7/1/09 10:45	7/1/2009 16:06	0.5465	1.5	6363.2	9544.8	11611	3.5	3317.4	6070.574762	0.6360	Average EFF 0.6339
8C	4	7/1/09 10:45	7/1/2009 16:19	0.5331	1.5	6363.2	9544.8	10809	3.5	3088.3	5793.080291	0.6069	
8D	1	7/1/09 10:45	7/1/2009 16:19	0.5330	1.5	6363.2	9544.8	11301	3.5	3228.9	6057.336905	0.6346	
8D	2	7/1/09 10:45	7/1/2009 16:15	0.5374	1.5	6363.2	9544.8	11412	3.5	3260.6	6087.58377	0.6357	
8D	3	7/1/09 10:45	7/1/2009 16:10	0.5421	1.5	6363.2	9544.8	11680	3.5	3314.1	6145.674775	0.6439	Average EFF 0.6281
8D	4	7/1/09 10:45	7/1/2009 16:06	0.5464	1.5	6363.2	9544.8	10918	3.5	3119.4	5709.327085	0.5982	
9A	1	7/1/09 10:45	7/1/2009 16:24	0.5280	1.5	6363.2	9544.8	11605	3.5	3315.7	6280.207813	0.6580	
9A	2	7/1/09 10:45	7/1/2009 16:42	0.5106	1.5	6363.2	9544.8	11281	3.5	3223.1	6313.016372	0.6614	
9A	3	7/1/09 10:45	7/1/2009 16:33	0.5196	1.5	6363.2	9544.8	11301	3.5	3228.9	6214.402502	0.6511	Average EFF
9A	4	7/1/09 10:45	7/1/2009 16:29	0.5236	1.5	6363.2	9544.8	10987	3.5	3139.1	5995.155865	0.6281	0.6496
9B	1	7/1/09 10:45	7/1/2009 16:29	0.5235	1.5	6363.2	9544.8	11151	3.5	3186.0	6085.406603	0.6376	
9B	2	7/1/09 10:45	7/1/2009 16:24	0.5280	1.5	6363.2	9544.8	11462	3.5	3274.9	6202.821366	0.6499	
9B	3	7/1/09 10:45	7/1/2009 16:42	0.5104	1.5	6363.2	9544.8	11004	3.5	3144.0	6160.125852	0.6454	Average EFF
9B	4	7/1/09 10:45	7/1/2009 16:33	0.5195	1.5	6363.2	9544.8	10581	3.5	3023.1	5819.569586	0.6097	0.6356
9C	1	7/1/09 10:45	7/1/2009 16:33	0.5194	1.5	6363.2	9544.8	11026	3.5	3150.3	6064.880483	0.6354	
9C	2	7/1/09 10:45	7/1/2009 16:29	0.5235	1.5	6363.2	9544.8	11281	3.5	3223.1	6157.122814	0.6451	
9C	3	7/1/09 10:45	7/1/2009 16:24	0.5279	1.5	6363.2	9544.8	11016	3.5	3147.4	5962.583098	0.6247	Average EFF
9C	4	7/1/09 10:45	7/1/2009 16:42	0.5103	1.5	6363.2	9544.8	10297	3.5	2942.0	5765.244836	0.6040	0.6273
9D	1	7/1/09 10:45	7/1/2009 16:38	0.5146	1.5	6363.2	9544.8	11135	3.5	3181.4	6182.4976	0.6477	
9D	2	7/1/09 10:45	7/1/2009 16:33	0.5193	1.5	6363.2	9544.8	11412	3.5	3260.6	6278.391381	0.6578	
9D	3	7/1/09 10:45	7/1/2009 16:29	0.5234	1.5	6363.2	9544.8	11340	3.5	3240.0	6190.682442	0.6488	Average EFF
9D	4	7/1/09 10:45	7/1/2009 16:24	0.5278	1.5	6363.2	9544.8	10912	3.5	3117.7	5907.401951	0.6189	0.6433
10A	1	7/1/09 10:45	7/1/2009 16:47	0.5057	1.5	6363.2	9544.8	10991	3.5	3140.3	6209.984837	0.6506	
10A	2	7/1/09 10:45	7/1/2										

10D	3	7/1/09 10:45	7/1/2009 16:53	0.5000	1.5	6363.2	9544.8	10643	3.5	3040.9	6081.577364	0.6372	Average EFF 0.6320
10D	4	7/1/09 10:45	7/1/2009 16:48	0.5053	1.5	6363.2	9544.8	10064	3.5	2875.4	5690.501596	0.5982	
11A	1	7/1/09 10:45	7/1/2009 11:56	0.8745	1.5	6363.2	9544.8	14773	3	4924.3	5631.22443	0.5900	
11A	2	7/1/09 10:45	7/1/2009 12:08	0.8547	1.5	6363.2	9544.8	14429	3	4809.7	5627.17636	0.5896	
11A	3	7/1/09 10:45	7/1/2009 12:04	0.8607	1.5	6363.2	9544.8	14454	3	4818.0	5597.851728	0.5865	Average EFF 0.5825
11A	4	7/1/09 10:45	7/1/2009 12:00	0.8677	1.5	6363.2	9544.8	14013	3	4671.0	5383.193838	0.5640	
11B	1	7/1/09 10:45	7/1/2009 12:00	0.8681	1.5	6363.2	9544.8	16203	3	5401.0	6221.768068	0.6518	
11B	2	7/1/09 10:45	7/1/2009 11:56	0.8742	1.5	6363.2	9544.8	16106	3	5388.7	6141.073627	0.6434	
11B	3	7/1/09 10:45	7/1/2009 12:08	0.8545	1.5	6363.2	9544.8	15843	3	5214.3	6102.154531	0.6393	Average EFF 0.6372
11B	4	7/1/09 10:45	7/1/2009 12:04	0.8606	1.5	6363.2	9544.8	15133	3	5044.3	5861.738123	0.6141	
11C	1	7/1/09 10:45	7/1/2009 12:04	0.8609	1.5	6363.2	9544.8	15637	3	5212.3	6054.305139	0.6343	
11C	2	7/1/09 10:45	7/1/2009 12:00	0.8680	1.5	6363.2	9544.8	15919	3	5306.3	6113.481467	0.6405	
11C	3	7/1/09 10:45	7/1/2009 11:56	0.8740	1.5	6363.2	9544.8	16452	3	5484.0	6274.376359	0.6574	Average EFF
11C	4	7/1/09 10:45	7/1/2009 12:08	0.8544	1.5	6363.2	9544.8	14887	3	4982.3	5808.157492	0.6085	0.6352
11D	1	7/1/09 10:45	7/1/2009 12:08	0.8548	1.5	6363.2	9544.8	15807	3	5202.3	6085.822645	0.6376	
11D	2	7/1/09 10:45	7/1/2009 12:04	0.8608	1.5	6363.2	9544.8	15944	3	5314.7	6174.136045	0.6469	
11D	3	7/1/09 10:45	7/1/2009 12:00	0.8679	1.5	6363.2	9544.8	16098	3	5386.0	6182.998937	0.6478	Average EFF
11D	4	7/1/09 10:45	7/1/2009 11:56	0.8738	1.5	6363.2	9544.8	15191	3	5083.7	5794.733717	0.6071	0.6348
12A	1	7/1/09 10:45	7/1/2009 12:15	0.8437	1.5	6363.2	9544.8	15450	3	5150.0	6104.026984	0.6395	
12A	2	7/1/09 10:45	7/1/2009 12:28	0.8234	1.5	6363.2	9544.8	15016	3	5005.3	6078.958289	0.6369	
12A	3	7/1/09 10:45	7/1/2009 12:24	0.8296	1.5	6363.2	9544.8	14984	3	4994.7	6020.558384	0.6308	Average EFF
12A	4	7/1/09 10:45	7/1/2009 12:20	0.8358	1.5	6363.2	9544.8	14530	3	4843.3	5794.58497	0.6071	0.6286
12B	1	7/1/09 10:45	7/1/2009 12:20	0.8362	1.5	6363.2	9544.8	15404	3	5134.7	6140.635636	0.6433	
12B	2	7/1/09 10:45	7/1/2009 12:15	0.8437	1.5	6363.2	9544.8	15807	3	5202.3	6166.05496	0.6460	
12B	3	7/1/09 10:45	7/1/2009 12:28	0.8232	1.5	6363.2	9544.8	15060	3	5020.0	6097.91718	0.6389	Average EFF
12B	4	7/1/09 10:45	7/1/2009 12:24	0.8295	1.5	6363.2	9544.8	14553	3	4851.0	5848.11587	0.6127	0.6352
12C	1	7/1/09 10:45	7/1/2009 12:24	0.8300	1.5	6363.2	9544.8	15183	3	5061.0	6097.649845	0.6388	
12C	2	7/1/09 10:45	7/1/2009 12:20	0.8361	1.5	6363.2	9544.8	15651	3	5217.0	6239.881493	0.6537	
12C	3	7/1/09 10:45	7/1/2009 12:15	0.8436	1.5	6363.2	9544.8	15216	3	5072.0	6012.519531	0.6299	Average EFF
12C	4	7/1/09 10:45	7/1/2009 12:28	0.8231	1.5	6363.2	9544.8	14117	3	4705.7	5716.805229	0.5989	0.6304
12D	1	7/1/09 10:45	7/1/2009 12:28	0.8235	1.5	6363.2	9544.8	15174	3	5058.0	6141.959419	0.6435	
12D	2	7/1/09 10:45	7/1/2009 12:24	0.8298	1.5	6363.2	9544.8	15137	3	5045.7	6080.699807	0.6371	
12D	3	7/1/09 10:45	7/1/2009 12:20	0.8359	1.5	6363.2	9544.8	15418	3	5139.3	6148.142699	0.6441	Average EFF
12D	4	7/1/09 10:45	7/1/2009 12:15	0.8434	1.5	6363.2	9544.8	14566	3	4855.3	5756.75774	0.6031	0.6320
13A	1	7/1/09 10:45	7/1/2009 12:33	0.8153	1.5	6363.2	9544.8	15230	3	5076.7	6226.552932	0.6524	
13A	2	7/1/09 10:45	7/1/2009 12:50	0.7902	1.5	6363.2	9544.8	14784	3	4928.0	6236.596242	0.6534	
13A	3	7/1/09 10:45	7/1/2009 12:41	0.8031	1.5	6363.2	9544.8	14851	3	4950.3	6164.384216	0.6458	Average EFF 0.6410
13A	4	7/1/09 10:45	7/1/2009 12:37	0.8090	1.5	6363.2	9544.8	14183	3	4727.7	5843.553624	0.6122	
13B	1	7/1/09 10:45	7/1/2009 12:37	0.8094	1.5	6363.2	9544.8	15625	3	5208.3	6434.850276	0.6742	
13B	2	7/1/09 10:45	7/1/2009 12:33	0.8153	1.5	6363.2	9544.8	15450	3	5150.0	6316.496573	0.6618	
13B	3	7/1/09 10:45	7/1/2009 12:50	0.7901	1.5	6363.2	9544.8	14689	3	4896.3	6197.297391	0.6493	Average EFF
13B	4	7/1/09 10:45	7/1/2009 12:41	0.8029	1.5	6363.2	9544.8	14377	3	4792.3	5968.757323	0.6253	0.6526
13C	1	7/1/09 10:45	7/1/2009 12:41	0.8033	1.5	6363.2	9544.8	15426	3	5142.0	6401.251014	0.6707	
13C	2	7/1/09 10:45	7/1/2009 12:37	0.8093	1.5	6363.2	9544.8	15315	3	5105.0	6307.973396	0.6609	
13C	3	7/1/09 10:45	7/1/2009 12:33	0.8152	1.5	6363.2	9544.8	15288	3	5096.0	6251.046762	0.6549	Average EFF
13C	4	7/1/09 10:45	7/1/2009 12:50	0.7900	1.5	6363.2	9544.8	14222	3	4740.7	6001.208943	0.6287	0.6538
13D	1	7/1/09 10:45	7/1/2009 12:50	0.7903	1.5	6363.2	9544.8	14492	3	4830.7	6112.65055	0.6404	
13D	2	7/1/09 10:45	7/1/2009 12:46	0.7958	1.5	6363.2	9544.8	14858	3	4952.7	6223.19528	0.6520	
13D	3	7/1/09 10:45	7/1/2009 12:37	0.8092	1.5	6363.2	9544.8	14873	3	4957.7	6126.881339	0.6419	Average EFF
13D	4	7/1/09 10:45	7/1/2009 12:33	0.8151	1.5	6363.2	9544.8	14389	3	4796.3	5884.197712	0.6165	0.6377
14A	1	7/1/09 10:45	7/1/2009 12:54	0.7834	1.5	6363.2	9544.8	14463	3	4821.0	6153.598507	0.6447	
14A	2	7/1/09 10:45	7/1/2009 13:17	0.7507	1.5	6363.2	9544.8	14137	3	4712.3	6277.53373	0.6577	
14A	3	7/1/09 10:45	7/1/2009 13:13	0.7571	1.5	6363.2	9544.8	14022	3	4674.0	6173.627369	0.6468	Average EFF 0.6393
14A	4	7/1/09 10:45	7/1/2009 13:02	0.7727	1.5	6363.2	9544.8	13451	3	4483.7	5802.830587	0.6080	
14B	1	7/1/09 10:45	7/1/2009 13:01	0.7730	1.5	6363.2	9544.8	14039	3	4679.7	6054.030301	0.6343	
14B	2	7/1/09 10:45	7/1/2009 12:54	0.7834	1.5	6363.2	9544.8	14398	3	4799.3	6126.324754	0.6418	
14B	3	7/1/09 10:45	7/1/2009 13:17	0.7505	1.5	6363.2	9544.8	13475	3	4491.7	5984.510182	0.6270	Average EFF 0.6266
14B	4	7/1/09 10:45	7/1/2009 13:13	0.7569	1.5	6363.2	9544.8	13077	3	4359.0	5758.643863	0.6033	
14C	1	7/1/09 10:45	7/1/2009 13:12	0.7573	1.5	6363.2	9544.8	14116	3	4705.3	6213.261445	0.6510	
14C	2	7/1/09 10:45	7/1/2009 13:02	0.7729	1.5	6363.2	9544.8	14187	3	4729.0	6118.427365	0.6410	
14C	3	7/1/09 10:45	7/1/2009 12:55	0.7832	1.5	6363.2	9544.8	14409	3	4803.0	6132.734423	0.6425	Average EFF 0.6375
14C	4	7/1/09 10:45	7/1/2009 13:17	0.7505	1.5	6363.2	9544.8	13229	3	4409.7	5875.993199	0.6156	
14D	1	7/1/09 10:45	7/1/2009 13:17	0.7508	1.5	6363.2	9544.8	13927	3	4642.3	6183.314452	0.6478	
14D	2	7/1/09 10:45	7/1/2009 13:12	0.7572	1.5	6363.2	9544.8	14089	3	4696.3	6202.348821	0.6498	
14D	3	7/1/09 10:45	7/1/2009 13:02	0.7728	1.5	6363.2	9544.8	13912	3	4637.3	6000.768164	0.6287	Average EFF 0.6326
14D	4	7/1/09 10:45	7/1/2009 12:55	0.7830	1.5	6363.2	9544.8	13545	3	4515.0	5766.084113	0.6041	

*Background is considered negligible

SampleID	Instr	Time (min.)	Alpha Counts	Beta Counts	Count Start Time	Count End Time
1 1A		3	126	13564	7/1/2009 13:36	7/1/2009 13:39
2 1A		3	136	12775	7/1/2009 13:52	7/1/2009 13:55
3 1A		3	135	12750	7/1/2009 13:48	7/1/2009 13:51
4 1A		3	142	12410	7/1/2009 13:41	7/1/2009 13:44
1 1B		3	115	13292	7/1/2009 13:41	7/1/2009 13:44
2 1B		3	136	13274	7/1/2009 13:36	7/1/2009 13:39
3 1B		3	131	12699	7/1/2009 13:52	7/1/2009 13:55
4 1B		3	129	12072	7/1/2009 13:48	7/1/2009 13:51
1 1C		3	207	12813	7/1/2009 13:48	7/1/2009 13:51
2 1C		3	221	12979	7/1/2009 13:41	7/1/2009 13:44
3 1C		3	189	12755	7/1/2009 13:36	7/1/2009 13:39
4 1C		3	179	11917	7/1/2009 13:52	7/1/2009 13:55
1 1D		3	558	12473	7/1/2009 13:52	7/1/2009 13:55
2 1D		3	582	12484	7/1/2009 13:48	7/1/2009 13:51
3 1D		3	632	12289	7/1/2009 13:41	7/1/2009 13:44
4 1D		3	568	12115	7/1/2009 13:36	7/1/2009 13:39
1 2A		3	424	12499	7/1/2009 13:57	7/1/2009 14:00
2 2A		3	449	12103	7/1/2009 14:15	7/1/2009 14:18
3 2A		3	419	11968	7/1/2009 14:09	7/1/2009 14:12
4 2A		3	417	11855	7/1/2009 14:02	7/1/2009 14:05
1 2B		3	42	12471	7/1/2009 14:02	7/1/2009 14:05
2 2B		3	39	12492	7/1/2009 13:57	7/1/2009 14:00
3 2B		3	54	11892	7/1/2009 14:15	7/1/2009 14:18
4 2B		3	69	11539	7/1/2009 14:09	7/1/2009 14:12
1 2C		3	504	12050	7/1/2009 14:08	7/1/2009 14:11
2 2C		3	527	11914	7/1/2009 14:02	7/1/2009 14:05
3 2C		3	496	11994	7/1/2009 13:58	7/1/2009 14:01
4 2C		3	499	10889	7/1/2009 14:15	7/1/2009 14:18
1 2D		3	543	12010	7/1/2009 14:15	7/1/2009 14:18
2 2D		3	508	12124	7/1/2009 14:08	7/1/2009 14:11
3 2D		3	542	12168	7/1/2009 14:02	7/1/2009 14:05
4 2D		3	544	11692	7/1/2009 13:58	7/1/2009 14:01
1 3A		3	1397	11194	7/1/2009 14:19	7/1/2009 14:22
2 3A		4	1809	14227	7/1/2009 14:35	7/1/2009 14:39
3 3A		4	1757	14180	7/1/2009 14:30	7/1/2009 14:34
4 3A		4	1725	13754	7/1/2009 14:25	7/1/2009 14:29
1 3B		4	914	15370	7/1/2009 14:25	7/1/2009 14:29
2 3B		3	731	11695	7/1/2009 14:20	7/1/2009 14:23
3 3B		4	960	14905	7/1/2009 14:35	7/1/2009 14:39
4 3B		4	922	14220	7/1/2009 14:30	7/1/2009 14:34
1 3C		4	671	15644	7/1/2009 14:29	7/1/2009 14:33
2 3C		4	722	15964	7/1/2009 14:25	7/1/2009 14:29
3 3C		3	558	11701	7/1/2009 14:20	7/1/2009 14:23
4 3C		4	647	14729	7/1/2009 14:35	7/1/2009 14:39
1 3D		4	651	15152	7/1/2009 14:35	7/1/2009 14:39
2 3D		4	722	15168	7/1/2009 14:30	7/1/2009 14:34
3 3D		4	684	15295	7/1/2009 14:25	7/1/2009 14:29
4 3D		3	466	10942	7/1/2009 14:20	7/1/2009 14:23
1 4A		4	412	15298	7/1/2009 14:40	7/1/2009 14:44
2 4A		4	407	14897	7/1/2009 15:00	7/1/2009 15:04
3 4A		4	389	15050	7/1/2009 14:53	7/1/2009 14:57

4
7/12/09

4 4A	4	417	14462	7/1/2009 14:48	7/1/2009 14:52
1 4B	4	58	15335	7/1/2009 14:48	7/1/2009 14:52
2 4B	4	61	15513	7/1/2009 14:41	7/1/2009 14:45
3 4B	4	53	14521	7/1/2009 15:00	7/1/2009 15:04
4 4B	4	72	14328	7/1/2009 14:53	7/1/2009 14:57
1 4C	4	532	14733	7/1/2009 14:53	7/1/2009 14:57
2 4C	4	545	14902	7/1/2009 14:48	7/1/2009 14:52
3 4C	4	486	14856	7/1/2009 14:41	7/1/2009 14:45
4 4C	4	540	13733	7/1/2009 15:00	7/1/2009 15:04
1 4D	4	1158	14167	7/1/2009 15:00	7/1/2009 15:04
2 4D	4	1192	14204	7/1/2009 14:53	7/1/2009 14:57
3 4D	4	1136	14131	7/1/2009 14:48	7/1/2009 14:52
4 4D	4	1149	13978	7/1/2009 14:41	7/1/2009 14:45
1 5A	4	424	14870	7/1/2009 15:06	7/1/2009 15:10
2 5A	4	395	14487	7/1/2009 15:21	7/1/2009 15:25
3 5A	4	403	14259	7/1/2009 15:17	7/1/2009 15:21
4 5A	4	389	13957	7/1/2009 15:12	7/1/2009 15:16
1 5B	4	428	14869	7/1/2009 15:12	7/1/2009 15:16
2 5B	4	440	14821	7/1/2009 15:06	7/1/2009 15:10
3 5B	4	420	14289	7/1/2009 15:21	7/1/2009 15:25
4 5B	4	414	13809	7/1/2009 15:17	7/1/2009 15:21
1 5C	4	436	14676	7/1/2009 15:17	7/1/2009 15:21
2 5C	4	443	15122	7/1/2009 15:12	7/1/2009 15:16
3 5C	4	433	14958	7/1/2009 15:07	7/1/2009 15:11
4 5C	4	416	13831	7/1/2009 15:21	7/1/2009 15:25
1 5D	4	451	14321	7/1/2009 15:21	7/1/2009 15:25
2 5D	4	452	14642	7/1/2009 15:17	7/1/2009 15:21
3 5D	4	444	14443	7/1/2009 15:12	7/1/2009 15:16
4 5D	4	414	13954	7/1/2009 15:07	7/1/2009 15:11
1 6A	4	272	14018	7/1/2009 15:27	7/1/2009 15:31
2 6A	3.5	246	12283	7/1/2009 15:40	7/1/2009 15:44
3 6A	3.5	231	12111	7/1/2009 15:36	7/1/2009 15:40
4 6A	3.5	229	11598	7/1/2009 15:32	7/1/2009 15:35
1 6B	3.5	540	12151	7/1/2009 15:32	7/1/2009 15:36
2 6B	4	592	14371	7/1/2009 15:27	7/1/2009 15:31
3 6B	3.5	498	11705	7/1/2009 15:40	7/1/2009 15:44
4 6B	3.5	498	11388	7/1/2009 15:36	7/1/2009 15:40
1 6C	3.5	462	12161	7/1/2009 15:36	7/1/2009 15:40
2 6C	3.5	468	12083	7/1/2009 15:32	7/1/2009 15:36
3 6C	4	534	13638	7/1/2009 15:27	7/1/2009 15:31
4 6C	3.5	455	11218	7/1/2009 15:40	7/1/2009 15:44
1 6D	3.5	456	11987	7/1/2009 15:40	7/1/2009 15:44
2 6D	3.5	468	12183	7/1/2009 15:36	7/1/2009 15:40
3 6D	3.5	496	11882	7/1/2009 15:32	7/1/2009 15:36
4 6D	4	525	13018	7/1/2009 15:27	7/1/2009 15:31
1 7A	3.5	466	12007	7/1/2009 15:46	7/1/2009 15:50
2 7A	3.5	491	11655	7/1/2009 16:00	7/1/2009 16:04
3 7A	3.5	444	11445	7/1/2009 15:56	7/1/2009 15:59
4 7A	3.5	477	11121	7/1/2009 15:50	7/1/2009 15:54
1 7B	3.5	418	11968	7/1/2009 15:51	7/1/2009 15:54
2 7B	3.5	448	12050	7/1/2009 15:46	7/1/2009 15:50
3 7B	3.5	460	11675	7/1/2009 16:00	7/1/2009 16:04

4 7B	3.5	413	11271	7/1/2009 15:56	7/1/2009 16:00
1 7C	3.5	471	11781	7/1/2009 15:56	7/1/2009 16:00
2 7C	3.5	457	11760	7/1/2009 15:51	7/1/2009 15:54
3 7C	3.5	454	11766	7/1/2009 15:46	7/1/2009 15:50
4 7C	3.5	406	10888	7/1/2009 16:00	7/1/2009 16:04
1 7D	3.5	359	11605	7/1/2009 16:00	7/1/2009 16:04
2 7D	3.5	391	11920	7/1/2009 15:56	7/1/2009 16:00
3 7D	3.5	386	11933	7/1/2009 15:51	7/1/2009 15:55
4 7D	3.5	400	11305	7/1/2009 15:46	7/1/2009 15:50
1 8A	3.5	348	11673	7/1/2009 16:06	7/1/2009 16:09
2 8A	3.5	340	11172	7/1/2009 16:19	7/1/2009 16:22
3 8A	3.5	298	11258	7/1/2009 16:15	7/1/2009 16:18
4 8A	3.5	327	10977	7/1/2009 16:10	7/1/2009 16:13
1 8B	3.5	124	11583	7/1/2009 16:10	7/1/2009 16:13
2 8B	3.5	112	11758	7/1/2009 16:06	7/1/2009 16:09
3 8B	3.5	110	11499	7/1/2009 16:19	7/1/2009 16:23
4 8B	3.5	102	10844	7/1/2009 16:15	7/1/2009 16:18
1 8C	3.5	202	11539	7/1/2009 16:15	7/1/2009 16:18
2 8C	3.5	196	11774	7/1/2009 16:10	7/1/2009 16:14
3 8C	3.5	203	11611	7/1/2009 16:06	7/1/2009 16:09
4 8C	3.5	207	10809	7/1/2009 16:19	7/1/2009 16:23
1 8D	3.5	240	11301	7/1/2009 16:19	7/1/2009 16:23
2 8D	3.5	248	11412	7/1/2009 16:15	7/1/2009 16:18
3 8D	3.5	233	11660	7/1/2009 16:10	7/1/2009 16:14
4 8D	3.5	235	10918	7/1/2009 16:06	7/1/2009 16:10
1 9A	3.5	39	11605	7/1/2009 16:24	7/1/2009 16:28
2 9A	3.5	49	11281	7/1/2009 16:42	7/1/2009 16:46
3 9A	3.5	47	11301	7/1/2009 16:33	7/1/2009 16:36
4 9A	3.5	64	10987	7/1/2009 16:29	7/1/2009 16:32
1 9B	3.5	53	11151	7/1/2009 16:29	7/1/2009 16:32
2 9B	3.5	39	11462	7/1/2009 16:24	7/1/2009 16:28
3 9B	3.5	45	11004	7/1/2009 16:42	7/1/2009 16:46
4 9B	3.5	51	10581	7/1/2009 16:33	7/1/2009 16:36
1 9C	3.5	49	11026	7/1/2009 16:33	7/1/2009 16:36
2 9C	3.5	49	11281	7/1/2009 16:29	7/1/2009 16:32
3 9C	3.5	40	11016	7/1/2009 16:24	7/1/2009 16:28
4 9C	3.5	60	10297	7/1/2009 16:42	7/1/2009 16:46
1 9D	3.5	65	11135	7/1/2009 16:38	7/1/2009 16:41
2 9D	3.5	53	11412	7/1/2009 16:33	7/1/2009 16:37
3 9D	3.5	54	11340	7/1/2009 16:29	7/1/2009 16:32
4 9D	3.5	77	10912	7/1/2009 16:24	7/1/2009 16:28
1 10A	3.5	71	10991	7/1/2009 16:47	7/1/2009 16:51
2 10A	4	106	11959	7/1/2009 17:12	7/1/2009 17:16
3 10A	3.5	70	10553	7/1/2009 16:58	7/1/2009 17:01
4 10A	3.5	95	10338	7/1/2009 16:53	7/1/2009 16:56
1 10B	4	139	11110	7/1/2009 17:03	7/1/2009 17:07
2 10B	3.5	102	10812	7/1/2009 16:47	7/1/2009 16:51
3 10B	4	103	11422	7/1/2009 17:12	7/1/2009 17:16
4 10B	3.5	110	9967	7/1/2009 16:58	7/1/2009 17:01
1 10C	3.5	74	10482	7/1/2009 16:58	7/1/2009 17:01
2 10C	3.5	79	10535	7/1/2009 16:53	7/1/2009 16:57
3 10C	3.5	87	10723	7/1/2009 16:47	7/1/2009 16:51

4 10C	4	95	11066	7/1/2009 17:13	7/1/2009 17:17
1 10D	4	102	12021	7/1/2009 17:13	7/1/2009 17:17
2 10D	3.5	75	10614	7/1/2009 16:58	7/1/2009 17:01
3 10D	3.5	78	10643	7/1/2009 16:53	7/1/2009 16:57
4 10D	3.5	81	10064	7/1/2009 16:48	7/1/2009 16:51
1 11A	3	31	14773	7/1/2009 11:56	7/1/2009 11:59
2 11A	3	23	14429	7/1/2009 12:08	7/1/2009 12:11
3 11A	3	33	14454	7/1/2009 12:04	7/1/2009 12:07
4 11A	3	49	14013	7/1/2009 12:00	7/1/2009 12:03
1 11B	3	43	16203	7/1/2009 12:00	7/1/2009 12:03
2 11B	3	53	16106	7/1/2009 11:56	7/1/2009 11:59
3 11B	3	46	15643	7/1/2009 12:08	7/1/2009 12:11
4 11B	3	42	15133	7/1/2009 12:04	7/1/2009 12:07
1 11C	3	27	15637	7/1/2009 12:04	7/1/2009 12:07
2 11C	3	38	15919	7/1/2009 12:00	7/1/2009 12:03
3 11C	3	33	16452	7/1/2009 11:56	7/1/2009 11:59
4 11C	3	46	14887	7/1/2009 12:08	7/1/2009 12:11
1 11D	3	43	15607	7/1/2009 12:08	7/1/2009 12:11
2 11D	3	42	15944	7/1/2009 12:04	7/1/2009 12:07
3 11D	3	32	16098	7/1/2009 12:00	7/1/2009 12:03
4 11D	3	39	15191	7/1/2009 11:56	7/1/2009 11:59
1 12A	3	29	15450	7/1/2009 12:15	7/1/2009 12:18
2 12A	3	28	15016	7/1/2009 12:28	7/1/2009 12:31
3 12A	3	31	14984	7/1/2009 12:24	7/1/2009 12:27
4 12A	3	46	14530	7/1/2009 12:20	7/1/2009 12:23
1 12B	3	26	15404	7/1/2009 12:20	7/1/2009 12:23
2 12B	3	31	15607	7/1/2009 12:15	7/1/2009 12:18
3 12B	3	34	15060	7/1/2009 12:28	7/1/2009 12:31
4 12B	3	49	14553	7/1/2009 12:24	7/1/2009 12:27
1 12C	3	24	15183	7/1/2009 12:24	7/1/2009 12:27
2 12C	3	44	15651	7/1/2009 12:20	7/1/2009 12:23
3 12C	3	46	15216	7/1/2009 12:15	7/1/2009 12:18
4 12C	3	60	14117	7/1/2009 12:28	7/1/2009 12:31
1 12D	3	48	15174	7/1/2009 12:28	7/1/2009 12:31
2 12D	3	37	15137	7/1/2009 12:24	7/1/2009 12:27
3 12D	3	25	15418	7/1/2009 12:20	7/1/2009 12:23
4 12D	3	59	14566	7/1/2009 12:15	7/1/2009 12:18
1 13A	3	50	15230	7/1/2009 12:33	7/1/2009 12:36
2 13A	3	36	14784	7/1/2009 12:50	7/1/2009 12:53
3 13A	3	41	14851	7/1/2009 12:41	7/1/2009 12:44
4 13A	3	49	14183	7/1/2009 12:37	7/1/2009 12:40
1 13B	3	39	15625	7/1/2009 12:37	7/1/2009 12:40
2 13B	3	41	15450	7/1/2009 12:33	7/1/2009 12:36
3 13B	3	37	14689	7/1/2009 12:50	7/1/2009 12:53
4 13B	3	47	14377	7/1/2009 12:41	7/1/2009 12:44
1 13C	3	54	15426	7/1/2009 12:41	7/1/2009 12:44
2 13C	3	41	15315	7/1/2009 12:37	7/1/2009 12:40
3 13C	3	36	15288	7/1/2009 12:33	7/1/2009 12:36
4 13C	3	34	14222	7/1/2009 12:50	7/1/2009 12:53
1 13D	3	47	14492	7/1/2009 12:50	7/1/2009 12:53
2 13D	3	50	14858	7/1/2009 12:46	7/1/2009 12:49
3 13D	3	43	14873	7/1/2009 12:37	7/1/2009 12:40

4 13D	3	47	14389	7/1/2009 12:33	7/1/2009 12:36
1 14A	3	44	14463	7/1/2009 12:54	7/1/2009 12:57
2 14A	3	41	14137	7/1/2009 13:17	7/1/2009 13:20
3 14A	3	45	14022	7/1/2009 13:13	7/1/2009 13:16
4 14A	3	51	13451	7/1/2009 13:02	7/1/2009 13:05
1 14B	3	42	14039	7/1/2009 13:01	7/1/2009 13:04
2 14B	3	36	14398	7/1/2009 12:54	7/1/2009 12:57
3 14B	3	47	13475	7/1/2009 13:17	7/1/2009 13:20
4 14B	3	47	13077	7/1/2009 13:13	7/1/2009 13:16
1 14C	3	26	14116	7/1/2009 13:12	7/1/2009 13:15
2 14C	3	35	14187	7/1/2009 13:02	7/1/2009 13:05
3 14C	3	37	14409	7/1/2009 12:55	7/1/2009 12:58
4 14C	3	38	13229	7/1/2009 13:17	7/1/2009 13:20
1 14D	3	16	13927	7/1/2009 13:17	7/1/2009 13:20
2 14D	3	32	14089	7/1/2009 13:12	7/1/2009 13:15
3 14D	3	16	13912	7/1/2009 13:02	7/1/2009 13:05
4 14D	3	47	13545	7/1/2009 12:55	7/1/2009 12:58

Notes:
 1 - Results are decay corrected to Sample Date/Time
 2 - Reference date for Spike Activity (dpm/ml) is the batch Prep Date
 3 - Spike Nominals are decay corrected to Sample Date/Time

* indicates results calculated at 100% recovery

Decision Level	Critical Required	MDA pcU/L	Sample Act. Conc. pcU/L	Act. Rate pcU/L	Net Count CPM	Net Rate Error CPM	2 SIGMA		2 SIGMA		Nominal pcU/L	Recovery
							Counting pCPM	Uncertainty pCPM	Total Prop.	Uncertainty Sample Type		
									pcU/L	pcU/L		
0.3471	0.2451	1	0.7182	133.0399	0.0254	131.6880	2.9686	5.9178	21.6468	LCS	164.3409	81.6%
0.3647	0.2575	1	0.9659	145.2821	0.0243	139.8173	2.9508	5.9071	21.4655	LCS	164.3409	81.0%
0.5369	0.3790	1	0.8753	159.8528	0.0239	150.0937	2.8583	5.8279	20.5368	LCS	164.3409	88.4%
0.4695	0.3314	1	0.8057	127.0000	0.0237	128.0933	2.8583	6.6057	25.6756	LCS	164.3409	97.3%
0.4261	0.3008	1	1.2813	141.0616	0.0247	135.4387	3.0211	6.1673	22.7300	LCS	164.3409	77.3%
0.7598	0.5365	1	0.7515	141.8559	0.0253	131.7983	2.9881	6.2613	22.9053	LCS	164.3409	85.8%
0.3798	0.2881	1	0.8072	145.8192	0.0251	131.8887	2.9686	6.4252	23.5374	LCS	164.3409	86.3%
0.6150	0.2830	1	1.1343	128.6854	0.0284	108.9047	2.7042	6.3116	21.1935	LCS	164.3409	88.7%
0.6347	0.4481	1	1.5022	135.4510	0.0268	119.8900	2.8455	6.3115	21.9803	LCS	164.3409	78.9%
0.9035	0.6379	1	1.0779	141.2554	0.0255	128.6447	2.9382	6.3235	22.8259	LCS	164.3409	82.4%
0.6078	0.4291	1	0.9887	155.5960	0.0247	137.7700	3.0378	6.7244	25.0836	LCS	164.3409	86.0%
0.5473	0.3864	1	1.1054	135.5336	0.0264	124.2433	2.8886	6.1761	21.9739	LCS	164.3409	94.7%
0.6283	0.4436	1	1.4942	138.9155	0.0254	125.4287	2.9134	6.2833	22.1127	LCS	164.3409	82.5%
0.9036	0.6379	1	1.3079	145.9256	0.0252	130.3400	2.9824	6.5032	23.5621	LCS	164.3409	83.3%
0.7676	0.5419	1	1.3000	147.9861	0.0268	124.2633	2.8910	6.7471	24.0105	LCS	164.3409	88.8%
0.7520	0.6309	1	0.9027	134.9811	0.0269	120.7040	2.8427	6.1231	21.8265	LCS	164.3409	90.0%
0.4899	0.3895	1	1.2076	131.4732	0.0271	117.9500	2.8170	6.1544	21.3797	LCS	164.3409	82.1%
0.6874	0.4924	1	1.1419	148.2299	0.0259	132.9873	2.9884	6.4406	23.6659	LCS	164.3409	89.0%
0.6530	0.4610	1	1.3064	156.3706	0.0255	139.2187	3.0605	6.7377	25.2968	LCS	164.3409	95.2%
0.7661	0.5409	1	1.3079	145.9256	0.0252	130.3400	2.9824	6.2523	21.8127	LCS	164.3409	81.7%
0.8889	0.4871	1	1.1987	141.1863	0.0270	118.9860	2.8288	6.2446	22.2843	LCS	164.3409	83.4%
0.6079	0.4292	1	1.0862	137.0396	0.0269	120.3027	2.8412	6.3436	23.6775	LCS	164.3409	88.8%
0.9609	0.6713	1	1.5725	146.0958	0.0264	127.0307	2.9317	6.6044	23.4785	LCS	164.3409	88.0%
0.4376	0.3980	1	0.8562	144.5849	0.0268	121.3713	2.8488	6.6518	21.8896	LCS	164.3409	88.0%
0.4227	0.2984	1	0.8830	134.2380	0.0275	113.7227	2.7577	6.3803	21.8577	LCS	164.3409	81.7%
0.4860	0.3079	1	0.8480	137.6373	0.0270	118.4987	2.8152	6.4094	22.3723	LCS	164.3409	83.8%
0.3862	0.2797	1	0.7956	151.8835	0.0262	128.6313	2.7839	6.7858	24.6098	LCS	164.3409	92.4%
0.4480	0.3163	1	0.8867	152.1131	0.0261	130.4707	2.9538	6.7449	24.6318	LCS	164.3409	92.6%
0.6832	0.4470	1	1.1278	127.4251	0.0279	109.4120	2.7108	6.2072	20.8518	LCS	164.3409	77.8%
0.8817	0.6831	1	1.6167	135.1471	0.0273	117.2540	2.8197	6.3869	21.8896	LCS	164.3409	82.2%
0.5779	0.4080	1	1.0463	146.5894	0.0263	127.3240	2.9214	6.5982	23.7610	LCS	164.3409	89.2%
0.8422	0.5946	1	1.4501	141.4935	0.0272	117.4880	2.8147	6.6441	23.0149	LCS	164.3409	86.1%
1.9822	1.3641	1	2.9747	135.0540	0.0285	109.6040	2.7857	6.7277	22.0820	LCS	164.3409	82.2%
0.4205	0.2969	1	0.8569	130.5505	0.0276	112.2200	2.7400	7.1247	21.2682	LCS	164.3409	89.4%
0.7972	0.5629	1	1.3835	137.7974	0.0261	112.5273	2.7540	6.4182	21.9266	LCS	164.3409	81.4%
0.4437	0.3132	1	0.8728	144.2824	0.0269	118.7633	2.8301	6.6832	23.4437	LCS	164.3409	87.6%
0.4475	0.3159	1	0.8728	135.4549	0.0253	141.3227	3.0173	6.7736	21.8707	LCS	164.3409	91.8%
0.8154	0.5757	1	1.3963	150.5313	0.0265	128.3747	2.9406	6.7718	24.4459	LCS	164.3409	81.8%
0.4063	0.2868	1	0.8104	146.5894	0.0263	113.5507	2.7553	6.3927	21.8871	LCS	164.3409	81.8%
0.8422	0.5946	1	1.4501	141.4935	0.0272	117.4880	2.8147	6.6441	23.0149	LCS	164.3409	82.2%
0.3879	0.2385	1	0.8569	146.9063	0.0268	121.4088	2.8489	7.0545	23.8548	LCS	164.3409	89.4%
0.4816	0.3400	1	0.8558	146.9063	0.0249	148.2120	3.1520	5.4891	21.2301	LCS	164.3409	80.1%
0.4437	0.3132	1	0.8728	144.8386	0.0271	117.5853	2.7540	6.7698	23.5500	LCS	164.3409	82.1%
0.3452	0.2423	1	0.6763	135.4549	0.0253	141.3227	3.0173	6.7736	21.8705	LCS	164.3409	90.6%
0.3289	0.2322	1	0.6397	131.6831	0.0247	150.2887	3.1984	5.4434	21.2188	LCS	164.3409	87.6%
0.2949	0.2082	1	0.5922	149.8038	0.0237	168.2880	3.0826	5.7928	23.8866	LCS	164.3409	80.1%
0.3879	0.2385	1	0.6397	151.8473	0.0225	172.6707	3.3968	5.8549	24.3615	LCS	164.3409	92.2%
0.4816	0.3400	1	0.8577	131.6839	0.0249	148.2120	3.1520	5.4891	21.2301	LCS	164.3409	80.1%
0.4437	0.3132	1	0.8728	144.8386	0.0246	153.3873	3.2186	5.5463	21.7215	LCS	164.3409	82.1%
0.7488	0.5287	1	1.2322	134.8566	0.0235	167.9807	3.3535	5.8232	23.8892	LCS	164.3409	97.2%
0.4447	0.3140	1	0.8052	147.8517	0.0238	147.3080	3.0735	5.4434	21.2188	LCS	164.3409	87.6%
0.6180	0.4365	1	1.0494	143.9479	0.0241	162.8890	3.3080	5.7315	23.1384	LCS	164.3409	80.1%
0.3427	0.2420	1	0.6860	135.0573	0.0248	148.3593	3.1490	5.6820	21.7752	LCS	164.3409	82.2%
0.5897	0.4234	1	1.0256	146.9063	0.0249	163.4987	3.3053	5.7852	23.4616	LCS	164.3409	88.8%
0.3116	0.2341	1	0.6469	146.0021	0.0240	163.4987	3.3053	5.7852	23.4616	LCS	164.3409	82.4%
0.6355	0.4487	1	1.0806	159.6717	0.0235	174.3747	3.4225	6.1425	25.6134	LCS	164.3409	90.6%
0.3136	0.2214	1	0.6265	132.0825	0.0251	144.5507	3.078	5.5850	21.3060	LCS	164.3409	80.4%
1.4618	1.0321	1	2.2506	135.6135	0.0254	145.4707	3.1981	5.8215	21.9070	LCS	164.3409	82.5%
0.3185	0.2249	1	0.6390	141.6238	0.0245	154.5427	3.2133	5.7718	22.7980	LCS	164.3409	86.2%
0.3327	0.2349	1	0.6546	146.7439	0.0242	158.9520	3.2579	5.8888	23.8017	LCS	164.3409	89.3%

149 712109

SampleID	Instr	Time (min.)	Alpha Counts	Beta Counts	Count Start Time	Count End Time	Machine
1	1A	15	36	1980	7/2/2009 8:39	7/2/2009 8:54	Protean
2	1B	15	27	1959	7/2/2009 8:40	7/2/2009 8:55	Protean
3	1C	15	44	2108	7/2/2009 8:40	7/2/2009 8:55	Protean
4	1D	15	108	2265	7/2/2009 8:40	7/2/2009 8:55	Protean
5	2A	15	69	1838	7/2/2009 8:40	7/2/2009 8:55	Protean
6	2B	15	8	2053	7/2/2009 8:40	7/2/2009 8:55	Protean
7	2C	15	96	1982	7/2/2009 8:40	7/2/2009 8:55	Protean
8	2D	15	93	1984	7/2/2009 9:08	7/2/2009 9:23	Protean
1	3A	15	233	1645	7/2/2009 9:08	7/2/2009 9:23	Protean
2	3B	15	99	1821	7/2/2009 9:08	7/2/2009 9:23	Protean
3	3C	15	96	1942	7/2/2009 9:08	7/2/2009 9:23	Protean
4	3D	15	90	2076	7/2/2009 9:08	7/2/2009 9:23	Protean
5	4A	15	79	1877	7/2/2009 9:08	7/2/2009 9:23	Protean
6	4B	15	13	1909	7/2/2009 9:08	7/2/2009 9:23	Protean
7	4C	15	97	1974	7/2/2009 9:09	7/2/2009 9:24	Protean
8	4D	15	181	1880	7/2/2009 9:25	7/2/2009 9:40	Protean
1	5A	15	53	1818	7/2/2009 9:26	7/2/2009 9:41	Protean
2	5B	15	59	1785	7/2/2009 9:26	7/2/2009 9:41	Protean
3	5C	15	43	2009	7/2/2009 9:26	7/2/2009 9:41	Protean
4	5D	15	59	2107	7/2/2009 9:26	7/2/2009 9:41	Protean
5	6A	15	35	1800	7/2/2009 9:27	7/2/2009 9:42	Protean
6	6B	15	71	1816	7/2/2009 9:27	7/2/2009 9:42	Protean
7	6C	15	81	1933	7/2/2009 9:27	7/2/2009 9:42	Protean
8	6D	15	81	1826	7/2/2009 9:47	7/2/2009 10:02	Protean
1	7A	15	75	1711	7/2/2009 9:48	7/2/2009 10:03	Protean
2	7B	15	59	1783	7/2/2009 9:48	7/2/2009 10:03	Protean
3	7C	15	74	1934	7/2/2009 9:48	7/2/2009 10:03	Protean
4	7D	15	83	1963	7/2/2009 9:48	7/2/2009 10:03	Protean
5	8A	15	49	1653	7/2/2009 9:48	7/2/2009 10:03	Protean
6	8B	15	20	1788	7/2/2009 9:48	7/2/2009 10:03	Protean
7	8C	15	34	1920	7/2/2009 9:48	7/2/2009 10:03	Protean
8	8D	15	45	1782	7/2/2009 10:07	7/2/2009 10:22	Protean
1	9A	15	17	1689	7/2/2009 10:06	7/2/2009 10:21	Protean
2	9B	15	13	1706	7/2/2009 10:06	7/2/2009 10:21	Protean
3	9C	15	13	1802	7/2/2009 10:06	7/2/2009 10:21	Protean
4	9D	15	15	1945	7/2/2009 10:06	7/2/2009 10:21	Protean
5	10A	15	10	1708	7/2/2009 10:07	7/2/2009 10:22	Protean
6	10B	15	19	1743	7/2/2009 10:07	7/2/2009 10:22	Protean
7	10C	15	15	1826	7/2/2009 10:07	7/2/2009 10:22	Protean
8	10D	15	14	1769	7/2/2009 10:22	7/2/2009 10:37	Protean
1	11A	15	19	2125	7/2/2009 7:26	7/2/2009 7:41	Protean
2	11B	15	22	2260	7/2/2009 7:26	7/2/2009 7:41	Protean
3	11C	15	13	2544	7/2/2009 7:26	7/2/2009 7:41	Protean
4	11D	15	14	2596	7/2/2009 7:26	7/2/2009 7:41	Protean
5	12A	15	17	2235	7/2/2009 7:26	7/2/2009 7:41	Protean
6	12B	15	10	2330	7/2/2009 7:26	7/2/2009 7:41	Protean
7	12C	15	16	2530	7/2/2009 7:26	7/2/2009 7:41	Protean
8	12D	15	10	2463	7/2/2009 7:26	7/2/2009 7:41	Protean
1	13A	15	11	2231	7/2/2009 7:49	7/2/2009 8:04	Protean
2	13B	15	13	2190	7/2/2009 7:49	7/2/2009 8:04	Protean
3	13C	15	11	2458	7/2/2009 7:49	7/2/2009 8:04	Protean

449
7/2/09

ver_pic_09.xls

4	13D	15	12	2635	7/2/2009 7:50	7/2/2009 8:05	Protean
5	14A	15	11	2173	7/2/2009 7:50	7/2/2009 8:05	Protean
6	14B	15	11	2281	7/2/2009 7:50	7/2/2009 8:05	Protean
7	14C	15	14	2323	7/2/2009 7:50	7/2/2009 8:05	Protean
8	14D	15	14	2388	7/2/2009 7:50	7/2/2009 8:05	Protean

Ra-228	Cal Date	7/2/2009	Exp Date	7/31/2009	
Protean	A0	A1	A2	A3	A4
1A	6.30258E-01				
1B	6.28221E-01				
1C	6.17615E-01				
1D	6.04341E-01				
2A	6.17224E-01				
2B	6.16681E-01				
2C	5.96919E-01				
2D	6.11886E-01				
3A	5.68218E-01				
3B	5.98041E-01				
3C	6.16431E-01				
3D	5.99405E-01				
4A	6.20765E-01				
4B	6.20459E-01				
4C	6.05183E-01				
4D	5.87325E-01				
5A	6.25790E-01				
5B	6.28027E-01				
5C	6.36802E-01				
5D	6.23741E-01				
6A	6.22050E-01				
6B	6.16280E-01				
6C	6.11053E-01				
6D	6.12043E-01				
7A	6.17961E-01				
7B	6.27962E-01				
7C	6.17791E-01				
7D	6.25720E-01				
8A	6.24723E-01				
8B	6.33167E-01				
8C	6.33890E-01				
8D	6.28089E-01				
9A	6.496412E-01				
9B	6.356321E-01				
9C	6.273008E-01				
9D	6.432553E-01				
10A	6.389066E-01				
10B	6.137441E-01				
10C	6.249999E-01				
10D	6.319781E-01				
11A	5.82502E-01				
11B	6.37172E-01				
11C	6.35171E-01				
11D	6.34840E-01				
12A	6.28566E-01				
12B	6.35234E-01				
12C	6.30366E-01				
12D	6.31956E-01				
13A	6.40953E-01				

13B	6.52643E-01
13C	6.53798E-01
13D	6.37701E-01
14A	6.39290E-01
14B	6.26611E-01
14C	6.37531E-01
14D	6.32609E-01


MPC 9600 Plateau

Instrument 1 MPC 9604 Detector A

7/1/2009

Alpha Volts: 1575

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	1	
765	0	
795	0 +0.00	
825	0 >100	
855	1 >100	
885	0 +55.56	
915	2 +66.67	
945	0 >100	
975	2 >100	
1005	42 >100	
1035	145 >100	
1065	544 >100	
1095	1136 >100	
1125	1967 >100	
1155	2845 >100	
1185	4078 >100	
1215	5483 +93.18	
1245	7400 +83.35	
1275	9328 +75.40	

1305	11640 +69.78
1335	14241 +62.88
1365	17534 +55.91
1395	20127 +45.04
1425	23254 +31.29
1455	24902 +20.41
1485	25605 +10.49
1515	26310 +6.44
1545	26535 +5.31
1575	26953 +2.79
1605	27399 +1.83
1635	27000 +1.71
1665	27255 +1.62
1695	27723 +3.14
1725	27705 +1.56
1755	28072 +1.15
1785	27729 +1.43
1815	28194 +3.24
1845	28243
1875	29191


MPC 9600 Plateau

Instrument 1 MPC 9604 Detector B

7/1/2009

Alpha Volts: 1575

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	+55.56
795	1	+83.33
825	1	+55.56
855	0	>100
885	1	+0.00
915	0	+0.00
945	1	>100
975	0	>100
1005	4	>100
1035	56	>100
1065	292	>100
1095	890	>100
1125	1841	>100
1155	2936	>100
1185	4179	>100
1215	5837	>100
1245	7821	+91.28
1275	10638	+83.88

1305	13188	+75.92
1335	16818	+67.60
1365	20420	+59.86
1395	24341	+47.85
1425	27854	+35.51
1455	30288	+23.26
1485	31798	+14.54
1515	32622	+8.32
1545	33496	+5.11
1575	33475	+4.43
1605	33903	+3.09
1635	34654	+2.46
1665	34485	+1.74
1695	34445	+1.84
1725	34908	+3.91
1755	35401	+6.80
1785	36062	+10.27
1815	37505	+14.30
1845	39508	
1875	41843	


MPC 9600 Plateau

Instrument 1 MPC 9604 Detector C

7/1/2009

Alpha Volts: 1575

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1		1305	14817	+71.06
735	0		1335	17823	+63.34
765	1 +0.00		1365	21704	+53.63
795	0 >100		1395	25422	+42.55
825	1 -55.56		1425	28424	+29.21
855	1 +55.56		1455	30244	+18.11
885	0 >100		1485	31305	+10.10
915	1 >100		1515	31989	+6.07
945	0 >100		1545	32223	+3.43
975	4 >100		1575	32671	+2.15
1005	32 >100		1605	32621	+1.68
1035	206 >100		1635	32837	+1.52
1065	639 >100		1665	32961	+2.01
1095	1416 >100		1695	33249	+2.64
1125	2551 >100		1725	33409	+3.21
1155	3619 >100		1755	33931	+4.07
1185	5037 +98.68		1785	34234	+7.20
1215	6875 +91.19		1815	34909	+10.28
1245	8915 +85.53		1845	36660	
1275	11519 +77.28		1875	38205	


MPC 9600 Plateau

Instrument 1 MPC 9604 Detector D

7/1/2009

Alpha Volts: 1575

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	15202	+66.36
735	1		1335	18216	+57.86
765	0 +0.00		1365	21597	+45.58
795	1 +0.00		1395	24648	+32.96
825	0 +0.00		1425	26505	+19.92
855	1 >100		1455	27475	+11.42
885	0 >100		1485	27836	+7.08
915	0 >100		1515	28609	+4.51
945	0 >100		1545	28896	+2.93
975	8 >100		1575	28862	+1.66
1005	75 >100		1605	28969	+0.36
1035	303 >100		1635	29292	+0.80
1065	872 >100		1665	28836	+1.06
1095	1656 >100		1695	29279	+1.48
1125	2729 >100		1725	29439	+3.59
1155	3862 >100		1755	29642	+4.07
1185	5425 +98.19		1785	30243	+6.51
1215	7256 +88.82		1815	30699	+7.79
1245	9510 +81.89		1845	31876	
1275	11944 +74.07		1875	32444	


MPC 9600 Plateau

Instrument 2 MPC 9604 Detector A

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts


705	0	
735	1	
765	0	+83.33
795	0	-83.33
825	1	>100
855	0	>100
885	1	+100.00
915	1	>100
945	2	>100
975	12	>100
1005	91	>100
1035	421	>100
1065	1239	>100
1095	2155	>100
1125	3527	>100
1155	4974	>100
1185	6647	+97.44
1215	9250	+89.00
1245	12041	+82.15
1275	15094	+73.81

1305	19017	+67.45
1335	23157	+59.23
1365	27625	+45.78
1395	31465	+32.72
1425	33352	+20.41
1455	35084	+11.74
1485	35819	+7.11
1515	36292	+3.35
1545	36527	+1.63
1575	36540	+0.87
1605	36585	+0.48
1635	36742	+1.76
1665	36691	+1.53
1695	37461	+1.89
1725	37073	+3.07
1755	37603	+4.02
1785	38346	+6.58
1815	39111	+7.95
1845	40115	
1875	41409	

MPC 9600 Plateau
Alpha Volts: 705

Instrument 2 MPC 9604 Detector B
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	12541	+83.18
735	1		1335	16192	+74.48
765	0		1365	20083	+67.17
795	0 >100		1395	24273	+58.43
825	0 >100		1425	29090	+46.86
855	0 >100		1455	33223	+34.56
885	0 >100		1485	35608	+22.67
915	0 >100		1515	37581	+13.63
945	1 >100		1545	38762	+8.18
975	2 >100		1575	39185	+4.42
1005	3 >100		1605	39484	+3.06
1035	14 >100		1635	39806	+2.61
1065	127 >100		1665	40264	+2.03
1095	500 >100		1695	40353	+2.32
1125	1332 >100		1725	40431	+3.28
1155	2373 >100		1755	41127	+7.09
1185	3614 >100		1785	41882	+12.40
1215	5227 >100		1815	44049	+18.52
1245	7060 +97.33		1845	46950	
1275	9574 +90.30		1875	51097	


MPC 9600 Plateau

Instrument 2 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	1 >100	
945	0 >100	
975	17 >100	
1005	87 >100	
1035	438 >100	
1065	1055 >100	
1095	2114 >100	
1125	3282 >100	
1155	4625 >100	
1185	6554 +97.66	
1215	8743 +88.09	
1245	11345 +81.31	
1275	14261 +74.60	

1305	18216 +67.74
1335	21995 +58.11
1365	26173 +46.11
1395	29479 +32.75
1425	32186 +20.62
1455	33022 +12.13
1485	33981 +7.22
1515	34520 +4.95
1545	35095 +2.07
1575	35014 +0.38
1605	34812 +0.55
1635	34859 +1.11
1665	35460 +1.94
1695	35273 +1.95
1725	35629 +2.73
1755	35811 +5.77
1785	36656 +6.44
1815	37896 +9.21
1845	38145
1875	40283


MPC 9600 Plateau

Instrument 2 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1		1305	18675	+65.94
735	0		1335	22620	+55.69
765	0	+83.33	1365	26869	+44.63
795	2	+55.56	1395	29957	+32.08
825	1	>100	1425	32494	+20.49
855	0	>100	1455	33836	+11.98
885	0	>100	1485	34627	+6.45
915	0	>100	1515	34849	+3.22
945	2	>100	1545	35298	+1.98
975	9	>100	1575	35180	+2.37
1005	89	>100	1605	35503	+1.57
1035	439	>100	1635	36006	+0.99
1065	1198	>100	1665	35722	+0.89
1095	2164	>100	1695	35597	+0.93
1125	3436	>100	1725	36188	+1.86
1155	4917	>100	1755	36272	+1.90
1185	6762	+96.59	1785	36389	+2.55
1215	9006	+89.14	1815	36529	+4.39
1245	11800	+81.34	1845	37459	
1275	15132	+73.59	1875	38170	


MPC 9600 Plateau

Instrument 3 MPC 9604 Detector A

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	+55.56
795	1	>100
825	1	+0.00
855	1	>100
885	0	>100
915	0	>100
945	0	>100
975	9	>100
1005	53	>100
1035	302	>100
1065	878	>100
1095	1805	>100
1125	2887	>100
1155	4163	>100
1185	5842	+99.81
1215	7959	+90.90
1245	10323	+83.03
1275	13250	+75.91

1305	16654	+68.57
1335	20416	+59.26
1365	24191	+47.28
1395	27643	+34.04
1425	29891	+21.08
1455	31183	+12.30
1485	31558	+6.67
1515	32444	+4.05
1545	32413	+2.90
1575	32704	+0.81
1605	32837	+0.71
1635	32629	+0.49
1665	32797	+0.16
1695	32964	+1.32
1725	32746	+1.40
1755	33308	+1.56
1785	33318	+3.21
1815	33456	+3.92
1845	34283	
1875	34815	


MPC 9600 Plateau

Instrument 3 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	1	
765	0	-55.56
795	0	>100
825	1	>100
855	3	+33.33
885	0	+0.00
915	1	>100
945	2	>100
975	29	>100
1005	165	>100
1035	613	>100
1065	1394	>100
1095	2558	>100
1125	3702	>100
1155	5222	>100
1185	7161	+96.06
1215	9507	+89.18
1245	12552	+81.52
1275	16030	+73.64

1305	19810	+64.73
1335	23962	+52.62
1365	28091	+39.27
1395	30594	+25.61
1425	32381	+14.86
1455	33206	+8.91
1485	33832	+4.41
1515	34260	+3.01
1545	34071	+2.33
1575	34623	+1.34
1605	34848	+1.22
1635	34564	+0.89
1665	34733	+1.01
1695	35144	+2.76
1725	35084	+3.66
1755	35839	+3.97
1785	36332	+5.39
1815	36654	+5.35
1845	37609	
1875	38164	


MPC 9600 Plateau

Instrument 3 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts


705	1	
735	1	
765	1	
795	0 >100	
825	0 +0.00	
855	0 >100	
885	1 >100	
915	1 >100	
945	1 >100	
975	17 >100	
1005	122 >100	
1035	533 >100	
1065	1287 >100	
1095	2493 >100	
1125	3753 >100	
1155	5482 >100	
1185	7538 +99.39	
1215	10305 +90.31	
1245	13415 +82.57	
1275	17141 +75.13	

1305	21412 +66.80
1335	26262 +56.32
1365	30679 +43.71
1395	34466 +31.61
1425	36949 +20.14
1455	38998 +11.16
1485	39313 +5.34
1515	39625 +2.44
1545	39751 +2.04
1575	40227 +1.45
1605	40228 +0.56
1635	40255 +0.13
1665	40075 +1.22
1695	40384 +1.95
1725	40900 +3.50
1755	41028 +3.05
1785	41899 +3.71
1815	41767 +5.64
1845	42852
1875	44132

MPC 9600 Plateau
Alpha Volts: 705

Instrument 3 MPC 9604 Detector D
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

705	0	
735	1	
765	0	+0.00
795	1	>100
825	0	+83.33
855	0	-83.33
885	1	>100
915	0	>100
945	1	>100
975	12	>100
1005	51	>100
1035	298	>100
1065	848	>100
1095	1649	>100
1125	2535	>100
1155	3602	>100
1185	5036	+98.31
1215	6880	+91.37
1245	8822	+82.29
1275	11546	+74.61


VOLTS COUNTS %/100 Volts

1305	14171	+66.45
1335	17362	+54.90
1365	20310	+43.83
1395	22647	+30.82
1425	24551	+20.19
1455	25440	+11.69
1485	26124	+5.90
1515	26245	+2.21
1545	26428	+1.39
1575	26151	+2.69
1605	26721	+2.72
1635	27168	+2.80
1665	27007	+0.87
1695	27135	+0.70
1725	27089	+1.24
1755	27414	+1.43
1785	27373	+3.21
1815	27581	+4.34
1845	28332	
1875	28750	

MPC 9600 Plateau
Alpha Volts: 705

Instrument 4 MPC 9604 Detector A
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	16442	+66.24
735	0		1335	20146	+57.40
765	0		1365	23769	+46.40
795	0 >100		1395	26926	+34.68
825	2 +55.56		1425	29276	+24.40
855	1 >100		1455	31037	+15.28
885	0 -55.56		1485	32197	+7.91
915	3 >100		1515	32425	+4.33
945	0 >100		1545	32314	+2.14
975	16 >100		1575	33071	+2.66
1005	114 >100		1605	32918	+2.52
1035	451 >100		1635	33435	+1.02
1065	1100 >100		1665	33382	+0.73
1095	2068 >100		1695	33349	+1.07
1125	3189 >100		1725	33324	+1.28
1155	4386 >100		1755	34001	+2.26
1185	6094 +94.81		1785	33701	+3.08
1215	8184 +87.09		1815	34304	+2.97
1245	10489 +78.88		1845	34744	
1275	13273 +72.66		1875	35012	


MPC 9600 Plateau

Instrument 4 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	15747	+62.38
735	1		1335	19230	+54.19
765	0 +0.00		1365	22255	+44.46
795	1 >100		1395	25299	+32.45
825	0 >100		1425	27370	+22.24
855	0 >100		1455	28625	+14.10
885	0 >100		1485	29467	+8.56
915	0 >100		1515	30213	+5.29
945	2 >100		1545	30326	+2.77
975	31 >100		1575	30564	+1.57
1005	176 >100		1605	30548	+1.52
1035	550 >100		1635	30820	+0.85
1065	1218 >100		1665	30898	+0.79
1095	2114 >100		1695	30779	+0.44
1125	3212 >100		1725	30934	+0.45
1155	4416 >100		1755	31008	+0.96
1185	6066 +92.28		1785	30991	+2.01
1215	7936 +85.60		1815	31196	+3.80
1245	10288 +76.79		1845	31781	
1275	13020 +70.59		1875	32406	


MPC 9600 Plateau

Instrument 4 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts


705	0	
735	1	
765	0	+55.56
795	2	+0.00
825	0	-55.56
855	1	>100
885	0	>100
915	0	>100
945	2	>100
975	24	>100
1005	134	>100
1035	558	>100
1065	1361	>100
1095	2511	>100
1125	3762	>100
1155	5246	>100
1185	7268	+96.29
1215	9733	+88.98
1245	12701	+79.94
1275	16176	+73.13

1305	19796	+65.77
1335	24338	+57.55
1365	28686	+45.86
1395	32750	+32.27
1425	34919	+20.83
1455	36434	+11.45
1485	37487	+5.80
1515	37623	+3.32
1545	37528	+2.07
1575	38277	+2.12
1605	38338	+2.70
1635	38426	+1.12
1665	39007	+1.06
1695	38592	+0.64
1725	38870	+0.63
1755	38868	+1.30
1785	39238	+1.45
1815	39169	+2.34
1845	39570	
1875	40086	

MPC 9600 Plateau
Alpha Volts: 705

Instrument 4 MPC 9604 Detector D
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	+0.00
795	0	>100
825	1	+83.33
855	1	+55.56
885	0	+0.00
915	1	>100
945	1	>100
975	60	>100
1005	297	>100
1035	855	>100
1065	1647	>100
1095	2700	>100
1125	3921	>100
1155	5471	+96.54
1185	7042	+90.21
1215	9405	+82.23
1245	12266	+76.33
1275	14989	+69.38

VOLTS COUNTS %/100 Volts

1305	18491	+61.09
1335	22444	+51.56
1365	25756	+37.44
1395	28379	+23.82
1425	29517	+14.00
1455	30309	+8.08
1485	30874	+6.03
1515	31345	+3.66
1545	31782	+2.17
1575	31567	+1.31
1605	31789	+0.78
1635	31963	+1.34
1665	31956	+0.29
1695	32123	+0.20
1725	31850	+1.46
1755	32114	+2.39
1785	32665	+3.95
1815	32876	+4.96
1845	33399	
1875	34206	


MPC 9600 Plateau

Instrument 5 MPC 9604 Detector A

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	1	
795	1 +83.33	
825	1 -83.33	
855	1 >100	
885	0 -55.56	
915	0 >100	
945	1 >100	
975	9 >100	
1005	76 >100	
1035	308 >100	
1065	814 >100	
1095	1600 >100	
1125	2598 >100	
1155	3596 >100	
1185	5065 +96.05	
1215	6773 +90.23	
1245	8717 +81.43	
1275	11391 +74.83	

VOLTS COUNTS %/100 Volts

1305	13974 +68.00
1335	17170 +58.62
1365	20456 +47.04
1395	23332 +33.83
1425	24996 +21.10
1455	26290 +12.40
1485	26683 +7.74
1515	27270 +4.43
1545	27590 +3.48
1575	27635 +1.71
1605	27932 +1.20
1635	27807 +0.88
1665	28006 +0.62
1695	27964 +0.63
1725	28112 +0.98
1755	28020 +2.84
1785	28392 +3.76
1815	29028 +5.17
1845	29220
1875	29849


MPC 9600 Plateau

Instrument 5 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	1 >100	
855	1 +41.67	
885	2 -33.33	
915	0 >100	
945	1 >100	
975	17 >100	
1005	87 >100	
1035	336 >100	
1065	1010 >100	
1095	1955 >100	
1125	3124 >100	
1155	4486 >100	
1185	6017 >100	
1215	8507 +91.20	
1245	11148 +82.59	
1275	14003 +74.21	

VOLTS COUNTS %/100 Volts

1305	17414 +68.46
1335	21540 +59.98
1365	25854 +46.75
1395	29222 +33.38
1425	31128 +21.52
1455	32995 +13.26
1485	33846 +8.09
1515	34289 +3.25
1545	34528 +2.00
1575	34311 +1.78
1605	34866 +1.78
1635	35046 +1.14
1665	35087 -0.26
1695	34795 +0.11
1725	34857 +0.93
1755	35220 +2.81
1785	35363 +3.98
1815	36028 +4.79
1845	36577
1875	37207


MPC 9600 Plateau

Instrument 5 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	1 >100	
885	0 >100	
915	1 >100	
945	2 >100	
975	7 >100	
1005	56 >100	
1035	305 >100	
1065	982 >100	
1095	1874 >100	
1125	2890 >100	
1155	4260 >100	
1185	6001 >100	
1215	8050 +91.54	
1245	10895 +82.98	
1275	13556 +76.26	

VOLTS COUNTS %/100 Volts


1305	17085 +68.24	
1335	21135 +59.99	
1365	25066 +47.39	
1395	28530 +33.93	
1425	30823 +22.30	
1455	32287 +12.93	
1485	33217 +6.71	
1515	33474 +3.57	
1545	33517 +1.17	
1575	33921 +1.13	
1605	33584 +1.27	
1635	34014 +1.12	
1665	34116 +0.98	
1695	34225 -0.22	
1725	33980 +0.58	
1755	33971 +1.96	
1785	34541 +3.64	
1815	34954 +5.38	
1845	35375	
1875	36384	

MPC 9600 Plateau

Instrument 5 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705 Beta Volts: 1575



VOLTS COUNTS %/100 Volts


705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	6 >100	
1005	81 >100	
1035	318 >100	
1065	897 >100	
1095	1710 >100	
1125	2714 >100	
1155	3925 >100	
1185	5395 +97.31	
1215	7282 +88.49	
1245	9426 +81.36	
1275	12007 +75.65	

VOLTS COUNTS %/100 Volts

1305	15025 +68.87
1335	18640 +58.97
1365	22048 +45.84
1395	24877 +32.08
1425	26653 +20.83
1455	27899 +13.08
1485	28670 +8.43
1515	29257 +5.13
1545	29568 +2.06
1575	29683 +0.52
1605	29362 +0.57
1635	29589 +0.80
1665	29870 +1.82
1695	29783 +0.90
1725	30077 +0.75
1755	29889 +2.02
1785	30152 +3.33
1815	30656 +6.54
1845	31211
1875	32389

MPC 9600 Plateau
Alpha Volts: 705

Instrument 6 MPC 9604 Detector A 7/1/2009
Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	7 >100	
1005	31 >100	
1035	238 >100	
1065	810 >100	
1095	1637 >100	
1125	2743 >100	
1155	3932 >100	
1185	5579 >100	
1215	7602 +94.41	
1245	10078 +84.86	
1275	13091 +77.67	

VOLTS COUNTS %/100 Volts

1305	16217 +71.57
1335	20184 +63.76
1365	24605 +53.98
1395	28528 +41.40
1425	31675 +28.02
1455	33899 +17.93
1485	34826 +10.65
1515	35815 +6.13
1545	36225 +4.15
1575	36456 +2.28
1605	36747 +1.47
1635	36801 +1.26
1665	36859 +0.85
1695	37095 +1.85
1725	37072 +4.01
1755	37724 +6.65
1785	38802 +10.33
1815	40036 +14.71
1845	41975
1875	45123


MPC 9600 Plateau

Instrument 6 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	1 +83.33	
855	1 -83.33	
885	0 >100	
915	0 >100	
945	5 >100	
975	18 >100	
1005	125 >100	
1035	482 >100	
1065	1255 >100	
1095	2318 >100	
1125	3540 >100	
1155	5288 >100	
1185	7168 +98.51	
1215	9760 +88.48	
1245	12656 +81.52	
1275	16065 +74.58	


VOLTS COUNTS %/100 Volts

1305	20094 +68.67	
1335	24665 +59.40	
1365	29591 +47.86	
1395	33376 +34.51	
1425	36440 +22.50	
1455	38024 +13.58	
1485	39187 +7.04	
1515	39608 +3.63	
1545	39722 +2.10	
1575	39894 +2.32	
1605	40298 +2.09	
1635	40711 +1.41	
1665	40574 +0.80	
1695	40608 +1.02	
1725	40839 +1.28	
1755	41201 +1.97	
1785	41065 +3.74	
1815	41711 +5.42	
1845	42917	
1875	43699	

MPC 9600 Plateau
Alpha Volts: 705

Instrument 6 MPC 9604 Detector C
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	1	+0.00
795	0	>100
825	0	+0.00
855	0	>100
885	1	>100
915	1	>100
945	2	>100
975	8	>100
1005	70	>100
1035	353	>100
1065	990	>100
1095	1956	>100
1125	3024	>100
1155	4400	>100
1185	6173	+99.75
1215	8230	+89.85
1245	10904	+82.36
1275	13747	+76.18


VOLTS COUNTS %/100 Volts

1305	17350	+67.80
1335	21371	+60.27
1365	25084	+49.32
1395	29177	+36.15
1425	31927	+24.86
1455	33217	+14.70
1485	34545	+7.74
1515	35097	+4.64
1545	34927	+2.96
1575	35439	+2.21
1605	35939	+2.41
1635	35763	+0.94
1665	36053	+0.35
1695	35886	+1.15
1725	36066	+1.77
1755	36379	+3.03
1785	36768	+4.80
1815	37193	+6.14
1845	38320	
1875	39061	

MPC 9600 Plateau
Alpha Volts: 705

Instrument 6 MPC 9604 Detector D
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	1 +0.00	
825	0 >100	
855	0 +0.00	
885	0 >100	
915	1 >100	
945	0 >100	
975	14 >100	
1005	109 >100	
1035	481 >100	
1065	1177 >100	
1095	2133 >100	
1125	3243 >100	
1155	4554 >100	
1185	6285 +98.38	
1215	8468 +89.75	
1245	11266 +83.13	
1275	14088 +74.43	


VOLTS COUNTS %/100 Volts

1305	17954 +65.82
1335	21482 +57.64
1365	25373 +45.78
1395	29042 +34.80
1425	31373 +23.29
1455	33143 +14.25
1485	34006 +8.49
1515	34662 +4.71
1545	34892 +3.14
1575	35129 +1.86
1605	35411 +1.49
1635	35380 +0.62
1665	35554 +0.65
1695	35385 +1.18
1725	35755 +1.89
1755	35907 +3.26
1785	36305 +4.62
1815	36870 +6.98
1845	37807
1875	39047

MPC 9600 Plateau
Alpha Volts: 705

Instrument 7 MPC 9604 Detector A
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	1 +83.33	
855	1 -83.33	
885	0 -55.56	
915	0 >100	
945	1 >100	
975	3 >100	
1005	42 >100	
1035	242 >100	
1065	613 >100	
1095	1353 >100	
1125	2213 >100	
1155	3256 >100	
1185	4474 >100	
1215	5932 +94.10	
1245	8072 +87.32	
1275	10579 +79.61	

VOLTS COUNTS %/100 Volts

1305	13228 +70.36
1335	16271 +60.12
1365	19506 +49.19
1395	22188 +36.46
1425	24373 +24.43
1455	25649 +15.99
1485	26433 +9.58
1515	27195 +5.74
1545	27367 +3.24
1575	27490 +1.86
1605	27608 +1.22
1635	27841 +1.33
1665	27695 +1.11
1695	27999 +1.42
1725	27992 +2.04
1755	28289 +2.52
1785	28408 +4.56
1815	28863 +5.70
1845	29664
1875	30148


MPC 9600 Plateau

Instrument 7 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	16978	+70.97
735	0		1335	20569	+61.39
765	0		1365	24989	+48.97
795	0 >100		1395	28389	+36.69
825	0 >100		1425	30977	+24.05
855	0 >100		1455	32727	+14.93
885	0 >100		1485	33697	+8.42
915	1 >100		1515	34195	+4.89
945	1 >100		1545	34437	+3.49
975	3 >100		1575	34850	+2.11
1005	34 >100		1605	35174	+1.62
1035	221 >100		1635	34923	+0.68
1065	825 >100		1665	35250	+0.35
1095	1709 >100		1695	35171	+1.24
1125	2873 >100		1725	35237	+1.02
1155	4078 >100		1755	35584	+2.79
1185	5858 >100		1785	35587	+4.59
1215	7809 +91.82		1815	36485	+6.74
1245	10336 +85.02		1845	37270	
1275	13215 +77.79		1875	38453	


MPC 9600 Plateau

Instrument 7 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	4 >100	
1005	57 >100	
1035	277 >100	
1065	817 >100	
1095	1666 >100	
1125	2766 >100	
1155	4077 >100	
1185	5667 >100	
1215	7694 +91.50	
1245	10209 +84.83	
1275	12950 +77.50	

VOLTS COUNTS %/100 Volts

1305	16543 +70.03
1335	20257 +60.71
1365	24245 +48.17
1395	27602 +35.50
1425	30019 +23.48
1455	31614 +14.53
1485	32522 +8.91
1515	33103 +5.28
1545	33572 +2.60
1575	33695 +0.70
1605	33525 +1.48
1635	33477 +0.99
1665	34432 +1.49
1695	33745 +1.43
1725	34149 +1.60
1755	34350 +3.69
1785	34955 +3.62
1815	35251 +4.44
1845	35592
1875	36382


MPC 9600 Plateau

Instrument 7 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	5 >100	
1005	29 >100	
1035	204 >100	
1065	609 >100	
1095	1354 >100	
1125	2316 >100	
1155	3418 >100	
1185	4654 >100	
1215	6455 +92.99	
1245	8669 +86.45	
1275	10931 +79.15	


VOLTS COUNTS %/100 Volts

1305	14016 +71.42
1335	17436 +62.21
1365	20814 +50.32
1395	23760 +36.91
1425	26302 +24.91
1455	27519 +15.17
1485	28410 +8.91
1515	28843 +5.41
1545	29396 +3.58
1575	29357 +1.54
1605	29719 +0.51
1635	29358 +0.23
1665	29623 +0.57
1695	29509 +2.12
1725	29896 +2.84
1755	30165 +4.42
1785	30570 +5.65
1815	31180 +6.95
1845	31995
1875	32717

MPC 9600 Plateau
Alpha Volts: 705

Instrument 8 MPC 9604 Detector A
Beta Volts: 1575

7/1/2009



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	1 >100	
975	9 >100	
1005	96 >100	
1035	468 >100	
1065	1084 >100	
1095	2286 >100	
1125	3479 >100	
1155	4912 >100	
1185	6819 +98.23	
1215	9153 +89.05	
1245	12105 +83.21	
1275	15122 +75.24	

1305	19482 +67.45
1335	23344 +59.35
1365	27793 +45.86
1395	31916 +34.29
1425	33979 +21.61
1455	35993 +11.71
1485	36530 +7.04
1515	36796 +3.11
1545	37393 +2.44
1575	37279 +1.41
1605	37650 +0.49
1635	37458 +0.91
1665	37579 +0.12
1695	37828 +1.10
1725	37535 +1.72
1755	38104 +2.18
1785	38416 +4.12
1815	38633 +4.92
1845	39649
1875	40366


MPC 9600 Plateau

Instrument 8 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	0 >100	
1005	20 >100	
1035	122 >100	
1065	511 >100	
1095	1263 >100	
1125	2390 >100	
1155	3641 >100	
1185	5246 >100	
1215	7212 +98.32	
1245	9897 +89.80	
1275	12742 +82.40	

VOLTS COUNTS %/100 Volts

1305	16337 +74.91
1335	20471 +68.07
1365	25012 +57.86
1395	29694 +47.48
1425	33409 +35.17
1455	37013 +23.27
1485	38629 +14.35
1515	39529 +7.69
1545	40284 +4.34
1575	40711 +2.52
1605	40642 +1.97
1635	40879 +1.11
1665	41405 +0.98
1695	41011 +0.30
1725	41182 +0.41
1755	41178 +3.28
1785	41573 +6.47
1815	42858 +10.82
1845	44440
1875	46780


MPC 9600 Plateau

Instrument 8 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	4 >100	
1005	46 >100	
1035	202 >100	
1065	697 >100	
1095	1532 >100	
1125	2614 >100	
1155	3953 >100	
1185	5474 >100	
1215	7466 +93.09	
1245	9842 +86.73	
1275	12814 +80.29	

VOLTS COUNTS %/100 Volts

1305	16303	+72.82
1335	20309	+64.32
1365	24364	+53.82
1395	28527	+40.95
1425	31774	+28.74
1455	33631	+16.87
1485	35030	+9.25
1515	35208	+5.21
1545	35741	+3.27
1575	36019	+2.95
1605	36373	+2.21
1635	36484	+2.27
1665	36713	+2.28
1695	37093	+2.46
1725	37325	+4.17
1755	37543	+7.52
1785	38833	+13.43
1815	40656	+19.49
1845	43753	
1875	47246	


MPC 9600 Plateau

Instrument 8 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705

Beta Volts: 1575



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	1	+0.00
795	0	>100
825	0	>100
855	0	>100
885	0	>100
915	0	>100
945	0	>100
975	5	>100
1005	47	>100
1035	243	>100
1065	792	>100
1095	1744	>100
1125	2933	>100
1155	4123	>100
1185	5780	>100
1215	7791	+91.58
1245	10478	+84.93
1275	13118	+77.50


VOLTS COUNTS %/100 Volts

1305	16889	+70.18
1335	20600	+61.29
1365	24824	+50.40
1395	28208	+38.85
1425	31539	+25.79
1455	33391	+16.06
1485	33991	+8.60
1515	34782	+5.01
1545	35201	+4.10
1575	35380	+2.50
1605	35849	+1.87
1635	35784	+1.79
1665	36000	+1.43
1695	36269	+2.10
1725	36381	+3.46
1755	36733	+6.86
1785	37669	+11.78
1815	39465	+16.64
1845	41803	
1875	44665	

MPC 9600 Plateau
Alpha Volts: 870

Instrument 9 MPC 9604 Detector A
Beta Volts: 1530

7/1/2009



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	1	+0.00
795	0	>100
825	0	>100
855	0	>100
885	0	>100
915	0	>100
945	0	>100
975	2	>100
1005	33	>100
1035	203	>100
1065	668	>100
1095	1403	>100
1125	2545	>100
1155	3800	>100
1185	5363	>100
1215	7355	+95.00
1245	9807	+87.69
1275	12700	+80.28

VOLTS COUNTS %/100 Volts

1305	16226	+71.71
1335	20083	+61.95
1365	23913	+49.99
1395	27526	+36.97
1425	30193	+24.54
1455	31747	+14.71
1485	32544	+7.71
1515	33198	+3.66
1545	33188	+1.51
1575	33227	+0.73
1605	33278	+1.04
1635	33518	+1.38
1665	33565	+1.95
1695	33774	+1.99
1725	34135	+3.30
1755	34244	+3.67
1785	35022	+4.84
1815	35229	+5.93
1845	36179	
1875	36821	


MPC 9600 Plateau

Instrument 9 MPC 9604 Detector B

7/1/2009

Alpha Volts: 870

Beta Volts: 1530



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	4 >100	
1005	45 >100	
1035	300 >100	
1065	836 >100	
1095	1742 >100	
1125	2896 >100	
1155	4198 >100	
1185	5849 >100	
1215	7887 +92.20	
1245	10561 +83.55	
1275	13442 +76.62	

VOLTS COUNTS %/100 Volts

1305	16723 +68.78
1335	20749 +60.55
1365	24686 +48.78
1395	28343 +35.24
1425	30657 +24.31
1455	32208 +15.22
1485	33662 +9.32
1515	34098 +4.47
1545	34326 +2.17
1575	34133 +1.60
1605	34758 +1.41
1635	34706 +1.35
1665	34769 +0.30
1695	34830 -0.10
1725	34850 +0.90
1755	34613 +2.41
1785	35351 +3.87
1815	35849 +4.97
1845	36285
1875	36814


MPC 9600 Plateau

Instrument 9 MPC 9604 Detector C

7/1/2009

Alpha Volts: 870

Beta Volts: 1530



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	1 >100	
915	1 >100	
945	2 >100	
975	3 >100	
1005	64 >100	
1035	349 >100	
1065	970 >100	
1095	1982 >100	
1125	3328 >100	
1155	5012 >100	
1185	6669 >100	
1215	9448 +92.67	
1245	12293 +86.58	
1275	15917 +76.99	

VOLTS COUNTS %/100 Volts

1305	20192 +70.39
1335	24524 +60.97
1365	29650 +48.44
1395	33904 +35.09
1425	36549 +22.73
1455	38217 +13.58
1485	39628 +7.51
1515	40035 +3.73
1545	40020 +1.92
1575	40236 +2.06
1605	40680 +1.62
1635	40953 +1.03
1665	40643 +0.43
1695	40882 +1.41
1725	40979 +2.18
1755	41654 +2.20
1785	41602 +2.27
1815	41935 +4.50
1845	42259
1875	44183


MPC 9600 Plateau

Instrument 9 MPC 9604 Detector D

7/1/2009

Alpha Volts: 870

Beta Volts: 1530



VOLTS	COUNTS	%/100 Volts
705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	1 >100	
945	0 >100	
975	5 >100	
1005	35 >100	
1035	186 >100	
1065	618 >100	
1095	1280 >100	
1125	2141 >100	
1155	3268 >100	
1185	4659 >100	
1215	6343 +90.68	
1245	8064 +83.46	
1275	10497 +77.03	

VOLTS	COUNTS	%/100 Volts
1305	13319	+70.94
1335	16319	+61.35
1365	19577	+50.27
1395	22498	+36.85
1425	24782	+23.90
1455	25761	+15.37
1485	26486	+8.38
1515	27503	+5.11
1545	27223	+2.67
1575	27453	+1.71
1605	27604	+2.70
1635	28021	+0.78
1665	28059	+1.05
1695	27548	+0.90
1725	28280	+2.16
1755	28290	+3.51
1785	28600	+4.46
1815	28879	+6.35
1845	29913	
1875	30417	


MPC 9600 Plateau

Instrument 10 MPC 9604 Detector A

7/1/2009

Alpha Volts: 870

Beta Volts: 1552



VOLTS COUNTS %/100 Volts

705	0	
735	1	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	1 >100	
1005	37 >100	
1035	198 >100	
1065	687 >100	
1095	1491 >100	
1125	2580 >100	
1155	3920 >100	
1185	5588 >100	
1215	7384 +91.32	
1245	9794 +84.81	
1275	12572 +79.73	

VOLTS COUNTS %/100 Volts

1305	16076 +72.76
1335	19985 +63.85
1365	24102 +50.95
1395	27819 +36.01
1425	30228 +23.86
1455	31343 +14.40
1485	32811 +8.77
1515	33243 +6.10
1545	33518 +3.25
1575	34010 +1.98
1605	34061 +1.59
1635	33973 +0.97
1665	34346 +0.93
1695	34366 +1.72
1725	34341 +1.54
1755	34860 +2.47
1785	34897 +4.50
1815	35377 +6.60
1845	36458
1875	37630


MPC 9600 Plateau

Instrument 10 MPC 9604 Detector B

7/1/2009

Alpha Volts: 870

Beta Volts: 1552



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	14469	+71.08
735	0		1335	17904	+63.07
765	0		1365	21677	+51.20
795	0 >100		1395	25027	+38.06
825	0 >100		1425	27237	+24.55
855	0 >100		1455	28914	+14.61
885	0 >100		1485	29480	+8.48
915	0 >100		1515	30075	+5.06
945	1 >100		1545	30374	+3.42
975	7 >100		1575	30738	+1.68
1005	28 >100		1605	30703	+1.08
1035	190 >100		1635	30679	+0.77
1065	597 >100		1665	30902	+1.46
1095	1474 >100		1695	30992	+1.89
1125	2383 >100		1725	31224	+2.40
1155	3680 >100		1755	31397	+3.27
1185	5131 >100		1785	31826	+4.13
1215	6808 +89.95		1815	32236	+5.59
1245	8990 +83.03		1845	32782	
1275	11493 +77.30		1875	33632	


MPC 9600 Plateau

Instrument 10 MPC 9604 Detector C

7/1/2009

Alpha Volts: 870

Beta Volts: 1552



VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	1 >100	
915	0 >100	
945	2 >100	
975	2 >100	
1005	36 >100	
1035	220 >100	
1065	780 >100	
1095	1712 >100	
1125	2926 >100	
1155	4297 >100	
1185	6097 >100	
1215	8397 +95.11	
1245	11155 +85.84	
1275	14430 +78.79	

VOLTS COUNTS %/100 Volts


1305	18051 +71.16
1335	22586 +62.34
1365	26973 +51.47
1395	31137 +38.24
1425	34321 +25.70
1455	36267 +15.37
1485	37197 +9.21
1515	37851 +5.38
1545	38622 +3.00
1575	38600 +1.55
1605	38538 +1.03
1635	38786 +0.91
1665	39129 +1.38
1695	38832 +1.20
1725	39323 +2.00
1755	39390 +3.35
1785	40031 +4.86
1815	40466 +6.64
1845	41713
1875	42620

MPC 9600 Plateau

Instrument 10 MPC 9604 Detector D 7/1/2009

Alpha Volts: 870

Beta Volts: 1552



VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	0 >100	
975	3 >100	
1005	49 >100	
1035	244 >100	
1065	764 >100	
1095	1584 >100	
1125	2677 >100	
1155	3763 >100	
1185	5395 >100	
1215	7350 +93.71	
1245	9655 +83.52	
1275	12504 +76.82	

VOLTS COUNTS %/100 Volts

1305	15430 +69.87	
1335	19258 +61.49	
1365	23018 +50.06	
1395	26562 +35.34	
1425	28750 +22.67	
1455	29911 +13.20	
1485	30798 +8.01	
1515	31375 +4.83	
1545	31684 +3.74	
1575	31721 +2.38	
1605	32398 +1.44	
1635	32154 +0.64	
1665	32157 -0.77	
1695	32152 +0.99	
1725	32029 +1.41	
1755	32699 +3.00	
1785	32566 +4.71	
1815	33351 +5.92	
1845	34031	
1875	34941	


Plateau 7/1/09

Instrument 11 MPC 9604 Detector A

7/1/2009

Alpha Volts: 1515

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

705	0	
735	1	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	0 >100	
945	1 +0.00	
975	0 >100	
1005	0 >100	
1035	2 >100	
1065	9 >100	
1095	61 >100	
1125	248 >100	
1155	528 >100	
1185	882 >100	
1215	1270 >100	
1245	1786 >100	
1275	2478 +93.67	

VOLTS COUNTS %/100 Volts

1305	3225	+87.64
1335	4189	+80.15
1365	5428	+75.12
1395	6662	+68.60
1425	8241	+58.14
1455	9857	+46.65
1485	11018	+33.24
1515	11953	+21.01
1545	12538	+13.57
1575	12760	+8.35
1605	13114	+5.84
1635	13258	+4.78
1665	13430	+3.99
1695	13551	+5.46
1725	13771	+8.65
1755	14204	+16.44
1785	14916	+30.03
1815	16579	+48.74
1845	19717	
1875	25029	


Plateau 7/1/09

Instrument 11 MPC 9604 Detector B

7/1/2009

Alpha Volts: 1515

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	0 >100	
915	1 >100	
945	0 >100	
975	11 >100	
1005	47 >100	
1035	280 >100	
1065	610 >100	
1095	1192 >100	
1125	1789 >100	
1155	2466 >100	
1185	3337 +94.91	
1215	4526 +88.85	
1245	5885 +78.40	
1275	7518 +72.09	

1305	8947 +65.63
1335	11238 +56.58
1365	13246 +46.66
1395	14838 +30.69
1425	16166 +20.11
1455	16396 +11.95
1485	17161 +5.61
1515	17274 +3.59
1545	17144 -0.00
1575	17323 +0.80
1605	17136 +2.21
1635	17484 +1.94
1665	17638 +2.16
1695	17580 +0.85
1725	17655 +1.05
1755	17700 +1.98
1785	17857 +2.38
1815	18006 +3.36
1845	18140
1875	18468


Plateau 7/1/09

Instrument 11 MPC 9604 Detector C

7/1/2009

Alpha Volts: 1515

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1	
735	0	
765	0	+0.00
795	0	>100
825	1	+0.00
855	0	>100
885	0	+0.00
915	0	>100
945	1	>100
975	7	>100
1005	46	>100
1035	191	>100
1065	540	>100
1095	957	>100
1125	1597	>100
1155	2217	>100
1185	3154	+98.74
1215	4239	+89.75
1245	5550	+79.98
1275	6980	+73.12

1305	8636	+66.44
1335	10593	+56.56
1365	12582	+46.23
1395	13957	+33.45
1425	15443	+21.49
1455	16048	+13.14
1485	16331	+6.45
1515	16603	+4.19
1545	16736	+2.73
1575	16884	+1.11
1605	16875	+1.91
1635	16813	+2.86
1665	17257	+2.60
1695	17425	+1.58
1725	17238	+0.49
1755	17230	+0.63
1785	17482	+3.27
1815	17468	+4.46
1845	17977	
1875	18163	


Plateau 7/1/09

Instrument 11 MPC 9604 Detector D

7/1/2009

Alpha Volts: 1515

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	7679	+65.97
735	0		1335	9737	+57.57
765	0		1365	11301	+45.87
795	0 >100		1395	12767	+31.71
825	0 >100		1425	13767	+19.90
855	1 +83.33		1455	14399	+10.72
885	1 +55.56		1485	14467	+4.38
915	0 >100		1515	14671	+2.12
945	1 >100		1545	14576	+2.61
975	9 >100		1575	14808	+1.80
1005	60 >100		1605	14974	+3.15
1035	173 >100		1635	14872	+1.76
1065	480 >100		1665	15248	-0.41
1095	911 >100		1695	15067	-0.27
1125	1508 >100		1725	14784	-0.43
1155	2024 >100		1755	15044	+2.01
1185	2872 +97.38		1785	15163	+2.82
1215	3858 +89.30		1815	15333	+3.61
1245	5070 +78.02		1845	15278	
1275	6322 +73.30		1875	15817	


Plateau 7/1/09

Instrument 12 MPC 9604 Detector A

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts


705	0		1305	6302	+80.03
735	1		1335	8191	+73.78
765	0		1365	10140	+66.18
795	0 >100		1395	12247	+55.83
825	0 >100		1425	14468	+43.92
855	0 >100		1455	16303	+31.28
885	0 >100		1485	17411	+18.64
915	0 >100		1515	18150	+9.87
945	0 >100		1545	18275	+5.30
975	1 >100		1575	18496	+3.16
1005	3 >100		1605	18685	+2.66
1035	17 >100		1635	18820	+2.63
1065	84 >100		1665	18855	+4.16
1095	267 >100		1695	19152	+7.70
1125	709 >100		1725	19706	+13.90
1155	1299 >100		1755	20640	+26.51
1185	1813 >100		1785	22308	+40.92
1215	2638 >100		1815	26460	+51.46
1245	3777 +96.47		1845	31616	
1275	4915 +87.98		1875	37348	

Plateau 7/1/09

Instrument 12 MPC 9604 Detector B 7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	1 +83.33	
885	1 -83.33	
915	0 -55.56	
945	0 >100	
975	1 >100	
1005	43 >100	
1035	165 >100	
1065	557 >100	
1095	1055 >100	
1125	1775 >100	
1155	2470 >100	
1185	3617 +98.46	
1215	4757 +90.95	
1245	6186 +83.59	
1275	8021 +77.85	

1305	10207 +70.42
1335	12473 +60.75
1365	14900 +48.87
1395	17101 +35.36
1425	18643 +22.53
1455	19350 +12.34
1485	19848 +6.68
1515	20014 +3.51
1545	20278 +2.03
1575	20186 +0.80
1605	20375 +0.32
1635	20209 +1.36
1665	20364 +0.83
1695	20607 +2.43
1725	20429 +2.51
1755	20924 +3.64
1785	20984 +5.11
1815	21470 +5.63
1845	21773
1875	22346


Plateau 7/1/09

Instrument 12 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1		1305	9543	+67.01
735	0		1335	11617	+56.47
765	0		1365	13791	+45.47
795	0 >100		1395	15387	+31.66
825	0 >100		1425	16819	+20.02
855	0 >100		1455	17210	+11.63
885	1 +0.00		1485	17742	+6.05
915	0 >100		1515	17892	+3.04
945	0 >100		1545	18070	+1.09
975	7 >100		1575	17856	+1.43
1005	52 >100		1605	18054	+0.42
1035	214 >100		1635	18287	+1.06
1065	590 >100		1665	17969	+0.78
1095	1201 >100		1695	18187	+1.48
1125	1759 >100		1725	18317	+4.89
1155	2569 >100		1755	18518	+4.76
1185	3440 +95.13		1785	19156	+5.18
1215	4583 +87.74		1815	19100	+5.18
1245	5985 +81.67		1845	19496	
1275	7682 +74.54		1875	19842	


Plateau 7/1/09

Instrument 12 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts


705	0		1305	9144	+69.92
735	0		1335	11120	+58.43
765	0		1365	13399	+45.40
795	0 >100		1395	14711	+32.57
825	0 >100		1425	16134	+20.69
855	0 >100		1455	16805	+13.46
885	0 >100		1485	17209	+7.90
915	0 >100		1515	17500	+4.31
945	0 >100		1545	17812	+3.48
975	4 >100		1575	17629	+2.80
1005	26 >100		1605	18066	+2.23
1035	169 >100		1635	18122	+1.44
1065	483 >100		1665	18166	+1.20
1095	955 >100		1695	17967	+1.60
1125	1639 >100		1725	18469	+3.41
1155	2233 >100		1755	18409	+6.35
1185	3262 +98.61		1785	18884	+9.47
1215	4306 +89.77		1815	19535	+11.98
1245	5662 +82.36		1845	20630	
1275	7113 +76.36		1875	21076	

Plateau 7/1/09

Instrument 13 MPC 9604 Detector A 7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0	
735	1	
765	0 +55.56	
795	2 >100	
825	0 +0.00	
855	0 >100	
885	1 >100	
915	0 >100	
945	1 >100	
975	14 >100	
1005	104 >100	
1035	281 >100	
1065	720 >100	
1095	1302 >100	
1125	1834 >100	
1155	2544 >100	
1185	3485 +92.28	
1215	4624 +85.50	
1245	5878 +77.82	
1275	7515 +71.49	

1305	9209 +64.55
1335	11200 +55.94
1365	13123 +43.27
1395	14957 +29.04
1425	15658 +17.41
1455	16123 +8.01
1485	16530 +4.92
1515	16437 +2.71
1545	16704 +0.83
1575	16707 +2.14
1605	16602 +0.55
1635	17024 -0.28
1665	16684 -0.42
1695	16597 -0.85
1725	16711 +1.27
1755	16796 +1.51
1785	16903 +1.57
1815	16880 +1.46
1845	17066
1875	17085


Plateau 7/1/09

Instrument 13 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts


VOLTS COUNTS %/100 Volts

705	0		1305	9666	+64.39
735	0		1335	11722	+55.91
765	0		1365	13680	+44.91
795	0 >100		1395	15677	+31.56
825	0 >100		1425	16786	+19.46
855	0 >100		1455	17283	+10.57
885	0 >100		1485	17608	+5.95
915	1 >100		1515	17972	+3.32
945	0 >100		1545	18006	+1.84
975	4 >100		1575	17970	+1.58
1005	70 >100		1605	18104	+0.74
1035	257 >100		1635	18351	+0.24
1065	648 >100		1665	18016	+0.16
1095	1116 >100		1695	18080	-0.63
1125	1784 >100		1725	18283	+0.29
1155	2560 >100		1755	18047	-0.47
1185	3531 +96.11		1785	18110	-0.32
1215	4568 +89.22		1815	18040	+1.17
1245	6137 +81.65		1845	18200	
1275	7855 +74.42		1875	18320	

Plateau 7/1/09

Instrument 13 MPC 9604 Detector C 7/1/2009

Alpha Volts: 705 Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	11573	+64.95
735	0		1335	13929	+56.47
765	0		1365	16726	+43.82
795	0 >100		1395	18834	+29.38
825	0 >100		1425	19743	+16.84
855	0 >100		1455	20314	+7.95
885	0 >100		1485	20860	+4.16
915	0 >100		1515	20670	+3.23
945	0 >100		1545	20844	+2.09
975	9 >100		1575	21330	+2.48
1005	93 >100		1605	21188	+1.16
1035	325 >100		1635	21280	-0.32
1065	834 >100		1665	21237	+0.08
1095	1525 >100		1695	21202	+0.42
1125	2318 >100		1725	21254	+0.60
1155	3233 >100		1755	21406	+1.41
1185	4357 +92.07		1785	21326	+0.42
1215	5755 +85.64		1815	21619	+0.16
1245	7438 +78.35		1845	21282	
1275	9463 +70.89		1875	21478	


Plateau 7/1/09

Instrument 13 MPC 9604 Detector D

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	1		1305	7524	+61.93
735	0		1335	9002	+55.36
765	0		1365	10542	+44.70
795	0 >100		1395	12064	+31.21
825	0 >100		1425	12981	+19.20
855	0 >100		1455	13192	+10.41
885	0 >100		1485	13570	+5.93
915	0 >100		1515	13820	+4.08
945	0 >100		1545	13866	+0.75
975	9 >100		1575	13880	+0.21
1005	58 >100		1605	13695	+0.59
1035	228 >100		1635	13950	+0.77
1065	544 >100		1665	13954	+1.92
1095	936 >100		1695	13911	+0.19
1125	1468 >100		1725	14116	+0.02
1155	2110 >100		1755	13908	-0.24
1185	2770 +94.71		1785	13960	-0.81
1215	3670 +85.91		1815	13939	+0.71
1245	4937 +79.46		1845	13931	
1275	6066 +70.79		1875	14071	


Plateau 7/1/09

Instrument 14 MPC 9604 Detector A

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0	
735	0	
765	0	
795	0 >100	
825	0 >100	
855	0 >100	
885	1 +0.00	
915	0 >100	
945	0 >100	
975	0 >100	
1005	18 >100	
1035	137 >100	
1065	430 >100	
1095	865 >100	
1125	1444 >100	
1155	2151 >100	
1185	2981 >100	
1215	4168 +92.14	
1245	5377 +84.73	
1275	6924 +74.92	

1305	8778 +67.49
1335	10502 +57.68
1365	12516 +46.36
1395	14215 +35.88
1425	15472 +22.01
1455	16469 +12.99
1485	16342 +6.70
1515	16874 +3.07
1545	16918 +2.53
1575	16950 +0.58
1605	16943 +0.95
1635	17008 +2.13
1665	17130 +2.45
1695	17403 +2.43
1725	17377 +2.43
1755	17515 +4.88
1785	17710 +7.54
1815	18533 +9.04
1845	18905
1875	19415


Plateau 7/1/09

Instrument 14 MPC 9604 Detector B

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	8797	+65.44
735	0		1335	10726	+54.47
765	0		1365	12570	+41.11
795	0 >100		1395	13917	+26.79
825	0 >100		1425	14687	+15.44
855	1 +0.00		1455	15048	+8.47
885	0 >100		1485	15318	+5.00
915	0 >100		1515	15494	+3.76
945	0 >100		1545	15606	+3.04
975	3 >100		1575	15776	+2.35
1005	40 >100		1605	15889	+1.44
1035	210 >100		1635	15907	-0.16
1065	590 >100		1665	15881	+0.64
1095	983 >100		1695	15741	+1.21
1125	1645 >100		1725	16124	+3.63
1155	2342 >100		1755	16076	+5.41
1185	3045 +96.43		1785	16588	+5.79
1215	4201 +90.42		1815	16830	+7.53
1245	5579 +83.64		1845	17185	
1275	7121 +74.44		1875	17682	


Plateau 7/1/09

Instrument 14 MPC 9604 Detector C

7/1/2009

Alpha Volts: 705

Beta Volts: 1515



VOLTS COUNTS %/100 Volts


VOLTS COUNTS %/100 Volts

705	0		1305	10118	+69.76
735	0		1335	12269	+59.65
765	0		1365	14810	+47.35
795	0 >100		1395	16773	+33.46
825	0 >100		1425	18104	+20.13
855	0 >100		1455	18720	+11.98
885	1 +0.00		1485	19122	+6.50
915	0 >100		1515	19580	+4.77
945	0 >100		1545	19527	+2.48
975	2 >100		1575	19902	+0.81
1005	21 >100		1605	19690	+0.53
1035	132 >100		1635	19739	+0.23
1065	491 >100		1665	19765	+1.29
1095	1036 >100		1695	19932	+1.40
1125	1698 >100		1725	19976	+2.72
1155	2517 >100		1755	20051	+2.92
1185	3468 >100		1785	20523	+4.26
1215	4721 +91.83		1815	20542	+5.57
1245	6175 +85.13		1845	21035	
1275	8025 +76.82		1875	21528	

Plateau 7/1/09

Instrument 14 MPC 9604 Detector D 7/1/2009

Alpha Volts: 705 Beta Volts: 1515



VOLTS COUNTS %/100 Volts

VOLTS COUNTS %/100 Volts

705	0		1305	8095	+71.16
735	0		1335	10052	+58.38
765	0		1365	11990	+47.92
795	0 >100		1395	13400	+35.01
825	0 >100		1425	14808	+23.58
855	0 >100		1455	15554	+13.45
885	0 >100		1485	15987	+6.39
915	0 >100		1515	15861	+3.45
945	0 >100		1545	16156	+2.18
975	1 >100		1575	16297	+1.72
1005	14 >100		1605	16297	+1.33
1035	130 >100		1635	16208	+1.62
1065	363 >100		1665	16526	+2.92
1095	785 >100		1695	16581	+3.94
1125	1357 >100		1725	16832	+5.91
1155	1996 >100		1755	17039	+8.68
1185	2735 +99.45		1785	17800	+11.53
1215	3785 +94.20		1815	18351	+11.46
1245	4857 +86.43		1845	19265	
1275	6571 +78.80		1875	19468	

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

66002-278

Ra-228 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated using a germanium gamma spectrometer system.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	Ra-228
ACTIVITY (dps):	2.367 E4
HALF-LIFE:	5.75 years
CALIBRATION DATE:	April 23, 2003 12:00 EST
TOTAL UNCERTAINTY*:	2.4%

*95% Confidence Level

Impurities: γ -impurities (other than decay products) <0.1%,
Ra-226 <0.1%

5.31628 grams 4M HCl solution with 100 μ g/g Ba carrier.

P O NUMBER 3219 RD, Item 1

SOURCE PREPARED BY: M. Taskaeva
M. Taskaeva, Radiochemist

Q A APPROVED: J.M. Monty 4-23-03

GEL Standard Traceability Log Rad

Source Material Info	
Parent Code:	0553-A
Prepared By:	Lonnie Morris
Carrier Conc:	0.5M HCl
Reference Date:	04/23/2003
Ampoule Mass (g):	5.0235 g
Uncertainty:	+/-
LogBook No:	RC-S-035-068

A Solution Material Info	
Isotope:	Radium-228 SPIKE
Prepared By:	Lonnie Morris
Prep Date:	04/25/2003
Verification Date:	04/27/2005
Expiration Date:	04/27/2006
Primary Code:	0553-B
Dilution(mL):	1000 mL
Mass of Parent(g):	30.535 g
Density(g/mL):	
Balance ID:	

Calculations Converting parent activity to dpm/mL/dpm/g

(Mass of parent(g)) * (Parm Activity (dpm/mL)) * (conversion dpm to dpm) / (Dilution Vol) = Parent Activity
 (dpm/mL)

(Mass of parent(g)) * (Parm Activity (dpm/mL)) * (conversion dpm to dpm) / Density (g/mL)/ (Dilution Vol) = Parent Activity (dpm/g)

(30.535 g) * (13419.8626 dpm/mL) * (1 dpm/dpm) / (1000 mL) = 409.7755 dpm/mL

(30.535 g) * (13419.8626 dpm/mL) * (1 dpm/dpm) / (g/mL) / (1000 mL) = dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date

GEL Laboratories LLC
 Version 1.0 9/18/2000

ANALYTICS

0503

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 · U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

64673-278

Ra-228 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated using a germanium gamma spectrometer system.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	Ra-228
ACTIVITY (dps):	1.939 E4
HALF-LIFE:	5.75 years
CALIBRATION DATE:	October 1, 2002 12:00 EST
TOTAL UNCERTAINTY*:	3.6%
SYSTEMATIC:	3.4%
RANDOM:	1.1%

*99% Confidence Level

Impurities: γ -impurities <0.1%

5.02617 grams 0.1M HCl solution with 110 μ g/g Ba carrier.

P O NUMBER 3208RD, Item 2

SOURCE PREPARED BY: M. Taskaeva
M. Taskaeva, Radiochemist

Q A APPROVED: M. Mty 10-202

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0503
Prepared By:	Angela Johnson
Carrier Conc:	0.1 M HCL
Reference Date:	10/01/2002
Ampoule Mass (g):	5.02617 g
Uncertainty:	+/- 3.6 %
LogBook No:	RC S 035 018

A Solution Material Info	
Isotope:	Radium-228
Prepared By:	Angela Johnson
Prep Date:	02/20/2003
Verification Date:	04/09/2004
Expiration Date:	04/09/2005
Primary Code:	0503-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.4737 g
Density(g/mL):	0.9992
Balance ID:	

Calculations Converting parent activity to dpm/mL/dpm/g

$$(\text{Mass of parent(g)}) * (\text{Parm Activity (dps)}) * (\text{conversion dpm to dps}) / (\text{Ampoule Mass(g)} * \text{(Dilution Vol)}) = \text{Parent Activity (dpm/mL)}$$

$$(\text{Mass of parent(g)}) * (\text{Parm Activity (dps)}) * (\text{conversion dpm to dps}) / \text{Density} / (\text{Ampoule Mass (g)} * \text{(Dilution Vol)}) = \text{Parent Activity (dpm/g)}$$

$$(4.4737 \text{ g}) * (19390 \text{ dps}) * (60 \text{ dpm/dps}) / (5.02617 \text{ g} * 100 \text{ mL}) = 10355.2060 \text{ dpm/mL}$$

$$(4.4737 \text{ g}) * (19390 \text{ dps}) * (60 \text{ dpm/dps}) / (0.9992 \text{ g/mL}) / (5.02617 \text{ g} * 100 \text{ mL}) = 10363.0820 \text{ dpm/g}$$

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
04/02/2003	Lonnie Morris	39.71	1000	0503-B	411.518 dpm/mL	09/13/2008	09/13/2009

GEL Laboratories LLC
Version 1.0 9/18/2000

Verification for Ra-228 Standard 0503-B

Standard					
	Isotope	Detector CPM	BKG CPM	NET CPM	Detector Eff Mass. Used (mL) Source DPM/mL
D. Roy	0503-B	1962.0000	45.6000	1916.4000	9.263763 1.0000 206.8705773
9/13/2008	0503-B	1983.2000	45.6000	1937.6000	9.263763 1.0000 209.1590642
	0503-B	1927.0000	45.6000	1881.4000	9.263763 1.0000 203.092415
					206.3740189

Mean Value (Counting) = 206.3740189 **StdDev =** 3.063655617 **dpm/mL**

Certificate Value = 200.596 **dpm/mL**

Lower Limit = 200.2467076 **dpm/mL**
Upper Limit = 212.5013301 **dpm/mL**
Pass

Rule 1 Pass/Fail

Two sigma = 6.127311233

10 % of Mean = 20.63740189

Pass

Rule 2 (Pass/Fail)

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements.

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-228 source 0503-B by transferring portions of the standard into glass liquid scintillation vials. Ten mL of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and background source were dark adapted for two hours and counted on LSC Gold for Ra-228 source standard verification. The Ra-228 efficiency calibration which was used for verification calculations was performed on 9/13/08 using source 0683-A (Ra-228). Calibration data is recorded in this logbook under Ra-228 0683-A. Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A - B)/(C)(D)$$

where:

A = Ver. source cpm,

B = BKG cpm,

C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

Daniell May 9/16/08

Onslow & Johnson 9/11/08

9/11/04

PAGE: 1

ID : TOTAL ACTIVITY

16 SEP 2008 16:24

USER:11 COMMENT:GOLD

PRESET TIME : 5.00
DATA CALC : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD
COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : EDIT
TWO PHASE : NO AQC : NO CYCLE REPEATS : 1 DISK : OFF
SCINTILLATOR: LIQUID LUMEX:YES LOW SAMPLE REJ: 0
LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

CHAN: 0.0 - 990.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 0.0 - 1000.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

SAM	POS	TIME	H#	WIND1	WIND2	LUMEX	ELAPSED
NO		MIN		CPM %ERROR	CPM %ERROR	%	TIME
1	11-1	5.00	98.2	50.40 12.60	54.00 12.17	0.41	5.55
2	11-2	1.30	99.3	7802.31 1.99	7803.08 1.99	0.00	7.81
3	11-3	1.30	100.4	7782.31 1.99	7786.15 1.99	0.00	10.14
4	11-4	1.35	99.2	7581.48 1.98	7585.19 1.98	0.01	12.51
5	11-5	5.00	97.9	45.60 13.25	47.20 13.02	0.43	18.61
6	11-6	5.00	110.7	1962.00 2.02	1964.80 2.02	0.01	24.65
7	11-7	5.00	110.8	1983.20 2.01	1984.80 2.01	0.01	30.75
8	11-8	5.00	110.7	1927.00 2.04	1927.80 2.04	0.02	36.85

8/16/08

Sample Count Start Time:

16 Sep 2008 16:46:59

Data Capture Date:

9/16/2008 16:52:01

User Filename:

S11091611-5A.WK1

U11091611-1A.WK1

Spectrum Type

Log Counts

User Number:

11

User Id:

TOTAL ACTIVITY

User Comment:

GOLD

Isotope Name:

14C

Scintillator:

LIQUID

Sample, Rack-Pos, Time:

5 11-5 5.00


H#, Total Counts:

97.9 69

Start, End, X-Axis:

0 990 Channel Number

SPECTRUM PLOT
USER 11 - TOTAL ACTIVITY




2008-09-16

Sample Count Start Time: 16 Sep 2008 16:53:01
Data Capture Date: 9/16/2008 16:58:06
User Filename: S11091611-6A.WK1
U11091611-1A.WK1

Spectrum Type Log Counts
User Number: 11
User Id: TOTAL ACTIVITY
User Comment: GOLD
Isotope Name: 14C
Scintillator: LIQUID
Sample, Rack-Pos, Time: 6 11-6 5.00
H#, Total Counts: 110.7 7666
Start, End, X-Axis: 0 990 Channel Number

SPECTRUM PLOT
USER 11 - TOTAL ACTIVITY




9/16/08
16:59:07

Sample Count Start Time: 16 Sep 2008 16:59:07
Data Capture Date: 9/16/2008 17:04:12
User Filename: S11091611-7A.WK1
U11091611-1A.WK1

Spectrum Type Log Counts
User Number: 11
User Id: TOTAL ACTIVITY
User Comment: GOLD
Isotope Name: 14C
Scintillator: LIQUID
Sample, Rack-Pos, Time: 7 11-7 5.00
H#, Total Counts: 110.8 7726
Start, End, X-Axis: 0 990 Channel Number

SPECTRUM PLOT
USER 11 - TOTAL ACTIVITY




2008
Sept 16

Sample Count Start Time: 16 Sep 2008 17:05:13
Data Capture Date: 9/16/2008 17:10:18
User Filename: S11091611-8A.WK1
U11091611-1A.WK1

Spectrum Type Log Counts
User Number: 11
User Id: TOTAL ACTIVITY
User Comment: GOLD
Isotope Name: 14C
Scintillator: LIQUID
Sample, Rack-Pos, Time: 8 11-8 5.00
H#, Total Counts: 110.7 7557
Start, End, X-Axis: 0 990 Channel Number

SPECTRUM PLOT
USER 11 - TOTAL ACTIVITY



Radium-228 Que Sheet

Sep 5 2009

Batch #:	881540	Analyst:DXM2	First Client Due Date:	NA	Internal Due Date@07/03/2009
Spike Isotope:	Radium-228	Spike Code: A NA	Expiration Date:	NA	Ac-228 Ingrw: 2025 6/30/09
LCS Isotope:	Radium-228	LCS Code: DS 03-B	Expiration Date:	9/13/09	Vol: N/A
Tracer Isotope:	Barium-133	Tracer Code: 0112-1	Expiration Date:	2/17/10	Vol: 2
Prep Date:	6/30/09	Initials: JRS	Pipet ID:	1734212	Vol: 0.1

Balance ID: NA

Collect Date & Time

Pos. #

Vol (mL)

Ba Yield (%)

Gamma Det. #

Hazard Code.	Min CRDL	Matrix	Client	Collect Date & Time	Pos. #	Vol (mL)	Ba Yield (%)	Gamma Det. #
				16-JUN-09 03:56 PM	1	20	100.83	1
				16-JUN-09 03:56 PM	2	20	108.20	2
				16-JUN-09 03:56 PM	3	20	114.22	3
				16-JUN-09 03:56 PM	4	20	120.58	4
				16-JUN-09 03:56 PM	5	20	105.84	5
				16-JUN-09 03:56 PM	6	20	102.70	6
				16-JUN-09 03:56 PM	7	20	112.83	7
				16-JUN-09 03:56 PM	8	20	111.91	8

7/26/09
JRS

Data Reviewed By:

Comments:

ASSAY 30-Jun-09 19:32:06

Protocol id 8 228_REC
Time limit 180
Count limit 50000
Isotope Ba-133
Protocol date 9-Apr-07 10:03:07
Run id. 54

POS	RACK	BATCH	TIME	COUNTS	CPM	ERROR	% RECOVERY	COUNT TIME
1	97	1	180	779	229.3	4.13		19:32:13
2	97	2	180	785	231.2	4.11	100.83	19:35:24
3	97	3	180	835	248.1	3.95	108.20	19:38:35
4	97	4	180	877	261.9	3.83	114.22	19:41:47
5	97	5	180	921	276.5	3.71	120.58	19:44:58
6	72	6	180	819	242.7	4	105.84	19:48:17
7	72	7	180	798	235.5	4.07	102.70	19:51:28
8	72	8	180	867	258.7	3.85	112.82	19:54:40
9	72	9	180	861	256.6	3.87	111.91	19:57:51

END OF ASSAY

2/2/09

LUCAS CELL COUNTERS

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414
(843)556-8171

Lucas Cell Calibration Package

	YES	NO	Comments
1) Is all calibration standard information enclosed for:			
the primary standard certificate?	✓		
the secondary standard(s) documentation?	✓		
standard preparation information?	✓		
standard < 1 Year old or verified?	✓		
2) Is the efficiency calibration report included ?	✓		
3) Is the raw count data included for:			
Cell constant determination?	✓		
Plateau generation?	✓		
4) Are the calibration verifications included?	✓		
5) Are the instrument settings included:			
HVPS settings?	✓		
6) Has the CELLEFF.xls file been updated ?	✓		
7) Have the calibration dates been updated in ALPHALIMS ?	✓		

Prepared By: Kelli B. Domke

Date: 8/31/09

Reviewed By: Angela J. G.

Date: 8/31/09

Effective Date: 8/31/09

Ra-226 Cell Constants

standard ID: 0299-H
Volume added (mL): 0.1
Standard Reference Activity (DPM/mL): 2483.21

<u>Lucas cell #</u>	<u>Cell constant</u>	<u>Standard Source</u>	<u>Date/Time of count</u>	<u>Date/time flushed to cell</u>	<u>Bkg Counts</u>	<u>bkg cpm</u>	<u>total counts</u>	<u>time min</u>	<u>cpm</u>	<u>Known activity</u>	<u>t1 (days) end-flush to count</u>	<u>t2 (days) end-degas to count</u>	<u>t3 (days) Std Ref Date to count</u>	<u>Decay from Std Ref Date to count</u>	
101	1.846	Average	1.956	cal 7	8/27/2009 16:35	8/27/2009 13:30	8/2/2009 11:30	4479	15	298.60	248.32	6.08333	0.12847	3544	0.9958
101	1.960	Sidev	0.107	cal 9	8/24/2009 14:20	8/24/2009 9:30	8/18/2009 13:40	4581	15	305.40	248.32	5.82639	0.20139	3541	0.9958
101	2.060			cal 1	8/21/2009 15:00	8/21/2009 9:30	8/18/2009 13:40	2945	15	196.33	248.32	2.82639	0.22917	3538	0.9958
102	1.862	Average	1.855	cal 5	8/27/2009 15:50	8/27/2009 12:40	8/2/2009 10:50	4510	15	300.67	248.32	6.07639	0.13194	3544	0.9958
102	1.850	Sidev	0.096	cal 10	8/24/2009 14:45	8/24/2009 9:55	8/18/2009 13:40	4330	15	288.67	248.32	5.84375	0.20139	3541	0.9958
102	1.853			cal 2	8/21/2009 15:20	8/21/2009 9:50	8/18/2009 13:40	2659	15	177.27	248.32	2.84028	0.22917	3538	0.9958

<u>Lucas cell #</u>	<u>Cell constant</u>	<u>Standard Source</u>	<u>Date/Time of count</u>	<u>Date/time flushed to cell</u>	<u>Bkg Counts</u>	<u>bkg cpm</u>	<u>total counts</u>	<u>time min</u>	<u>cpm</u>	<u>Known activity</u>	<u>t1 (days) end-flush to count</u>	<u>t2 (days) end-degas to count</u>	<u>t3 (days) Std Ref Date to count</u>	<u>Decay from Std Ref Date to count</u>	
104	2.073	Average	1.972	cal 1	8/27/2009 14:25	8/27/2009 9:35	8/2/2009 11:00	3070	15	204.67	248.32	2.94097	0.20139	3544	0.9958
104	1.855	Sidev	0.110	cal 11	8/24/2009 15:15	8/24/2009 10:15	8/18/2009 13:40	4343	15	289.53	248.32	5.85764	0.20833	3541	0.9958
104	1.987			cal 3	8/21/2009 15:50	8/21/2009 10:10	8/18/2009 13:40	2858	15	190.53	248.32	2.85417	0.23611	3538	0.9958

<u>Lucas cell #</u>	<u>Cell constant</u>	<u>Standard Source</u>	<u>Date/Time of count</u>	<u>Date/time flushed to cell</u>	<u>Bkg Counts</u>	<u>bkg cpm</u>	<u>total counts</u>	<u>time min</u>	<u>cpm</u>	<u>Known activity</u>	<u>t1 (days) end-flush to count</u>	<u>t2 (days) end-degas to count</u>	<u>t3 (days) Std Ref Date to count</u>	<u>Decay from Std Ref Date to count</u>	
106	1.985	Average	1.836	cal 2	8/27/2009 14:55	8/27/2009 10:00	8/2/2009 11:20	2940	15	196.00	248.32	2.94444	0.20486	3544	0.9958
106	1.738	Sidev	0.131	cal 12	8/24/2009 15:35	8/24/2009 10:40	8/18/2009 13:40	4078	15	271.87	248.32	5.87500	0.20486	3541	0.9958
106	1.786			cal 4	8/21/2009 16:30	8/21/2009 10:30	8/18/2009 13:40	2572	15	171.47	248.32	2.86606	0.25000	3538	0.9958
107	2.025	Average	1.981	cal 8	8/27/2009 16:55	8/27/2009 13:50	8/2/2009 11:55	4910	15	327.33	248.32	6.07986	0.12847	3544	0.9958
107	2.054	Sidev	0.102	cal 1	8/24/2009 15:55	8/24/2009 11:00	8/2/2009 10:50	3090	15	206.00	248.32	3.00694	0.20486	3541	0.9958
107	1.864			cal 5	8/21/2009 16:45	8/21/2009 10:50	8/18/2009 13:40	2696	15	179.73	248.32	2.88194	0.24653	3538	0.9958
108	1.906	Average	1.946	cal 6	8/27/2009 16:05	8/27/2009 13:05	8/2/2009 11:15	4623	15	308.20	248.32	6.07639	0.12500	3544	0.9958
108	1.975	Sidev	0.036	cal 2	8/24/2009 16:25	8/24/2009 11:20	8/21/2009 10:50	2978	15	198.53	248.32	3.02083	0.21181	3541	0.9958
108	1.957			cla 6	8/21/2009 17:00	8/21/2009 11:15	8/18/2009 13:40	2846	15	189.73	248.32	2.89931	0.23958	3538	0.9958

E053028 < Put in Machines.xls (Lucas Cell Tab)

8/13/109

$$\frac{1.10}{1.10} = .900$$

Voltage ~~1.10~~
8/13/09

8/13/09

Ra-226 Calibration Sheet

Standard ID: 0'V9-μ
 Volume Added (mL): .01
 Expiration Date: 8/11/10

* count time 15 min

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 1	500	8/18/09 13:40	8/19/09 09:30	8/20/09 14:10	101	1	34250
Cal 1	500	8/18/09 13:10	8/19/09 09:50	8/20/09 14:25	101	1	2778
Cal 3	500	8/18/09 13:40	8/19/09 10:10	8/21/09 14:45	104	1	2182
Cal 4	500	8/18/09 13:10	8/19/09 10:30	8/21/09 14:30	100	1	2572
Cal 5	500	8/18/09 13:40	8/19/09 10:50	8/21/09 14:15	101	1	2694
Cal 6	500	8/18/09 13:40	8/19/09 11:15	8/21/09 17:00	108	1	2844
Cal 7	500	8/18/09 13:40	8/19/09 11:30	8/21/09 17:15	111	1	2712
Cal 8	500	8/18/09 13:10	8/19/09 11:55	8/21/09 17:35	112	1	2731
Cal 9							
Cal 10							
Cal 11							
Cal 12							

W8/18/09

8/13/09

10/8/2009

$$\text{Voltage} = 0.9$$

Ra-226 Calibration Sheet

Standard ID: 0191-14
Volume Added (mL): 0.1
Expiration Date:

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count	Cell #	Det #	Total Counts
Cul 9	500	8/18/04 1340	8/19/04 0130	8/19/04 1735	101	1	8439
Cul 10	500	8/18/04 1340	8/19/04 0155	8/19/04 1745	101	1	8444
Cul 11	500	8/18/04 1340	8/19/04 0155	8/19/04 1755	104	1	1343
Cul 12	500	8/18/04 1340	8/19/04 1010	8/20/04 1009	1535	104	1
Cul 13	500	8/19/04 1050	8/19/04 1100	8/24/04 1555	101	1	3090
Cul 14	500	8/19/04 1050	8/19/04 1110	8/24/04 109	108	1	2978
Cul 15	500	8/19/04 1050	8/19/04 1115	8/24/04 1700	111	1	3139
Cul 16	500	8/19/04 1050	8/19/04 1140	8/24/04 1715	112	1	3019

1918.18.471

~~119~~ 3109

Voltage - D.9

Ra-226 Calibration Sheet

Standard ID: D001-4
Volume Added (mL): 1/10
Expiration Date:

* 15 min counts

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 1	500	8/27/09 1101	8/27/09 0435	8/27/09 1425	104	1	3070
Cal 2	500	8/27/09 1144	8/27/09 1000	8/27/09 1455	106	1	2940
Cal 3	500	8/27/09 1125	8/27/09 1010	8/27/09 1512	111	1	3177
Cal 4	500	8/27/09 1240	8/27/09 1050	8/27/09 1530	112	1	2895
Cal 5	500	8/27/09 1050	8/27/09 1040	8/27/09 1550	107	1	4510
Cal 6	500	8/27/09 1115	8/27/09 1305	8/27/09 1605	108	1	41623
Cal 7	500	8/27/09 1130	8/27/09 1330	8/27/09 1635	101	1	44179
Cal 8	500	8/27/09 1155	8/27/09 1350	8/27/09 1655	107	1	4910

1931.09

2019.8.4

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number GL-RMP-A-008 Isotope RA-226
 Date Standards Prepared 4/15/05 Cocktail Type Used NA
 Standard ID 0299-H Matrix of Vial/Planchett NA
 Amount Used (g or mL) 0.1 WA
 Standard Activity (DPM/g or mL) 2483.1133 NA
 Reference Date 12/15/99 Type of Scintillation Vial NA
 Expiration Date 8/1/10 Pipette ID Used 1429303
 Residue/Carrier Agent 0.1M HCl Balance ID Used 38080104
 Quenching Agent NA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	CG11				
2	CG12				
3	CG13				
4	CG14				
5	CG15				
6	CG16				
7	CG17				
8	CG18				
9	CG19				
10	CG110				
11	CG111				
12	CG112				
<u>WAS Q20105</u>					

Prepared By:

Lily S. De Leon

Date

8/31/09

Reviewed By:

Angie A. Gh

Date

8/31/09

Rev 1 RLM 9/10/97

ee'd

8-21-00

Nycomed Amersham plc
Amersham Laboratories

0299

CALIBRATION
No. 0146

ISSUED
TO:
Nycomed Amersham plc
Radiation & Radioactivity
Calibration Laboratory
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

ISSUED
FOR:
AEA Technology plc
Isotrak
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

Description Principal radionuclide: Radium-226

Product code: RAY44
Solution number: R4/131/89

Measurement Reference time: 1200 GMT on 15 December 1999

Nuclear data Nuclear data quoted on this certificate are taken from the Joint European File, Version 2.2.

Expression of uncertainty The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, which corresponds to a t -distribution with $v_{eff} = \infty$ effective degrees of freedom. This corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Unless indicated, all other uncertainties are expressed at the confidence level associated with one standard uncertainty.

The format used for the uncertainties in the values of radionuclidic purity is illustrated in the following examples:

6.5(21)	-	6.5 ± 2.1
6.54(21)	-	6.54 ± 0.21
6.543(21)	-	6.543 ± 0.021

Date of
issue

17th December 1999 V10 8133105

Nycomed

WSPN

Standard Traceability Log Rad

Source Material Info		A Solution Material Info	
Isotope:		Radium-226	
Prepared By:	Angela Johnson	Prepared By:	Angela Johnson
Prep Date:	09/15/2000	Prep Date:	09/15/2000
Verification Date:	01/23/2008	Expiration Date:	01/23/2009
Carrier Conc:	0.5 M HCL	Primary Code:	0299-A
Reference Date:	12/15/1999	Dilution(mL):	100 mL
Ampoule Mass (g):	5.0368 g	Mass of Parent(g):	4.6634 g
Uncertainty:	+/- 2.5 %	Density(g/mL):	1.0012
LogBook No:	RC S 027 128	Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

$$(\text{Mass of parent(g)} * (\text{Parm Activity (kBq/g)}) * (\text{conversion dpm to kBq}) / \text{Density (g/mL)} / (\text{Dilution Vol}) = \text{Parent Activity (dpm/mL)}$$

$$(\text{Mass of parent(g)} * (\text{Parm Activity (kBq/g)}) * (\text{conversion dpm to kBq}) / \text{Density (g/mL)} / (\text{Dilution Vol}) = \text{Parent Activity (dpm/g)}$$

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (100 \text{ mL}) = 122414.2500 \text{ dpm/mL}$$

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (1.0012 \text{ g/mL}) / (100 \text{ mL}) = 122273.3377 \text{ dpm/g}$$

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	01/26/2009	01/26/2010
08/07/2009	Mary Aders	5.0767	250	0299-H	2483.2133 dpm/mL	08/07/2009	08/07/2010

Voltage Curve Ludlum #1

Voltage (kV)	Count Time (min)	Counts	Date/Time
0.20	1.00	0	8/21/09 13:20
0.25	1.00	0	8/21/09 13:21
0.30	1.00	0	8/21/09 13:22
0.35	1.00	0	8/21/09 13:23
0.40	1.00	0	8/21/09 13:24
0.45	1.00	0	8/21/09 13:25
0.50	1.00	0	8/21/09 13:26
0.55	1.00	1534	8/21/09 13:27
0.60	1.00	19637	8/21/09 13:28
0.65	1.00	47206	8/21/09 13:29
0.70	1.00	80410	8/21/09 13:30
0.75	1.00	104945	8/21/09 13:31
0.80	1.00	122514	8/21/09 13:32
0.85	1.00	134160	8/21/09 13:33
0.90	1.00	144753	8/21/09 13:34
0.95	1.00	151057	8/21/09 13:35
1.00	1.00	157429	8/21/09 13:36
1.05	1.00	163110	8/21/09 13:37
1.10	1.00	166034	8/21/09 13:38
1.15	1.00	168121	8/21/09 13:39
1.20	1.00	171347	8/21/09 13:40
1.25	1.00	173388	8/21/09 13:41
1.30	1.00	175958	8/21/09 13:42
1.35	1.00	182719	8/21/09 13:43
1.40	1.00	195871	8/21/09 13:44
1.45	1.00	231584	8/21/09 13:45
1.50	1.00	303021	8/21/09 13:46
1.55	1.00	387838	8/21/09 13:47


Detector set to operate at 0.90 kV

8/21/09

Ludlum Detector Voltage Curve

—♦— Voltage Curve Ludlum #1

Operating Voltage = 0.9 kV



8/3/09

Control Limits for Lucas Cell Counter #1

Analyst: KSD1

Date: 8/31/2009

Count #	Detector #1
1	138383
2	138269
3	141307
4	140521
5	132825
6	135924
7	139231
8	138298
9	135342
10	138056
11	138123
12	139159
13	138410
14	138251
15	138438
16	138080
17	137814
18	137961
19	137248
20	137477

Average = 137955.9

Std. Dev. = 1775.5

+3 S. D. = 143282.4266

+2 S. D. = 141506.901

Mean = 137955.9

-2 S. D. = 134404.799

-3 S. D. = 132629.2734

Control Limits 8/31/2009 * Operating Voltage changed to 0.9 kV
Detector #1
Upper Limit 143282
Lower Limit 132629

49
8/31/09

	Eff	Cal Date
101	1.956	8/31/2009
102	1.855	8/31/2009
104	1.972	8/31/2009
106	1.836	8/31/2009
107	1.981	8/31/2009
108	1.946	8/31/2009
111	2.024	8/31/2009
112	1.931	8/31/2009

Lucas	Ra-226	
Oldest Cal	01/23/2008	
Detector	Eff Error	Cal Date
1	0.0530	8/31/2009
2	0.0772	12/19/2008
3	0.0608	1/23/2008
4	0.1237	3/2/2009
5	0.1438	3/25/2009
6	0.0661	8/4/2009
7	0.0855	11/21/2008

8/09

Ra-226 WATER

Batch : LCSVER
 Date : 8/20/2008
 Analyst : KSD1

Procedure Code : LUC26RAL
 Paramname : Radium-226
 MDA : 1 pCi/L
 Bkg Count Time: 30 min

Sample ID	Sample Vol L	Count Time min	Gross counts cts	Cell #	Cell Const. num	BKG cpm	BKG cpm	Ra-226 MDA pCi/L	Ra-226 RESULT pCi/L	Ra-226 ERROR pCi/L	COUNT DATE/TIME
Ver 2	0.500	30	689	101	1.956	8	0.267	0.5907	25.3156	1.9236	8/31/2009 14:35
Ver 6	0.500	30	697	102	1.855	4	0.133	0.4721	27.1986	2.0367	8/31/2009 15:05
Ver 2	0.500	30	656	104	1.972	8	0.267	0.6303	25.7021	2.0032	8/28/2009 14:00
Ver 4	0.500	30	638	106	1.836	8	0.267	0.6304	24.9919	1.9762	8/31/2009 15:40
Ver 7	0.500	30	629	107	1.981	8	0.267	0.6257	24.4533	1.9479	8/28/2009 17:50
Ver 5	0.500	30	693	108	1.946	8	0.267	0.5959	25.6861	1.9459	8/31/2009 16:15
Ver 3	0.500	30	672	111	2.024	8	0.267	0.6129	25.6096	1.9713	8/28/2009 14:35
Ver 4	0.500	30	631	112	1.931	8	0.267	0.6411	25.1365	1.9990	8/28/2009 15:10

✓ 8/31/09

Sample ID	Sample Dup	Det #	Run Date	Sample Type	Standard ID	NC	NC units	Recovery/RPD
Ver 2		1	8/31/2009 14:35	LCS	0638-H	24.17	pCi/L	105%
Ver 3		1	8/31/2009 15:05	LCS	0638-H	24.17	pCi/L	113%
Ver 2		1	8/28/2009 14:00	LCS	0638-H	24.17	pCi/L	106%
Ver 4		1	8/31/2009 15:40	LCS	0638-H	24.17	pCi/L	103%
Ver 7		1	8/28/2009 17:50	LCS	0638-H	24.17	pCi/L	101%
Ver 8		1	8/31/2009 16:15	LCS	0638-H	24.17	pCi/L	106%
Ver 3		1	8/28/2009 14:35	LCS	0638-H	24.17	pCi/L	106%
Ver 4		1	8/28/2009 15:10	LCS	0638-H	24.17	pCi/L	104%

DEGASSING DATE/TIME	DE-EMAN. DATE/TIME	DEGASS-DE-EM	dE-EM-COUNT	constant	constant	constant	Net CPM cpm	In growth constant
8/28/2009 10:20	8/31/2009 11:10	72.83	3.42	0.4230	0.9745	1.0019	22.7000	0.4130
8/28/2009 10:40	8/31/2009 11:30	72.83	3.58	0.4230	0.9733	1.0019	23.1000	0.4125
8/25/2009 16:00	8/28/2009 10:20	66.33	3.67	0.3940	0.9727	1.0019	21.6000	0.3839
8/28/2009 11:00	8/31/2009 11:55	72.92	3.75	0.4234	0.9721	1.0019	21.0000	0.4123
8/25/2009 16:00	8/28/2009 12:00	68.00	5.83	0.4015	0.9569	1.0019	20.7000	0.3850
8/28/2009 11:20	8/31/2009 12:15	72.92	4.00	0.4234	0.9703	1.0019	22.8333	0.4115
8/25/2009 16:00	8/28/2009 10:40	66.67	3.92	0.3955	0.9709	1.0019	22.1333	0.3847
8/25/2009 16:00	8/28/2009 11:00	67.00	4.17	0.3970	0.9690	1.0019	20.7667	0.3854

Ver 24
1 mL

Rn-226 Verification Sheet

W1 .9 voltage

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Ver 1	500	8/28/09 1600	8/28/09 0555	8/28/09 1210	101	1	8	525
Ver 1	500	8/28/09 1600	8/28/09 1010	8/28/09 1400	104	1	8	656
Ver 5	500	8/28/09 1600	8/28/09 1040	8/28/09 1435	111	1	8	670
Ver 4	500	8/28/09 1600	8/28/09 1110	8/28/09 1510	111	1	8	631
Ver 5	500	8/28/09 1600	8/28/09 1110	8/28/09 1510	106	1	8	678
Ver 6	500	8/28/09 1600	8/28/09 1140	8/28/09 1540	104	1	8	654
Ver 7	500	8/28/09 1600	8/28/09 1100	8/28/09 1500	107	1	8	629
Ver 4	500	8/28/09 1600	8/28/09 1100	8/28/09 1500	106	1	8	630
Ver 1	500	8/28/09 1600	8/28/09 1100	8/28/09 1500	105	1	8	629
Ver 5	500	8/28/09 1600	8/28/09 1100	8/28/09 1500	105	1	8	630
Ver 1	500	8/28/09 1600	8/28/09 1100	8/28/09 1500	101	1	8	689
Ver 3	500	8/28/09 1600	8/28/09 1130	8/28/09 1505	102	1	4	697
Ver 4	500	8/28/09 1600	8/28/09 1100	8/28/09 1540	104	1	8	635
Ver 5	500	8/28/09 1600	8/28/09 1100	8/28/09 1545	108	1	8	693

Ver 24
1 mL

4/21/09

4/21/09

General Engineering Laboratories Verification Source Preparation Sheet

Applicable SOP Number	JL-1410-A-006	Isotope	
Date Standards Prepared	11/23/08	Cocktail Type Used	NA
Standard ID	06138-H	Matrix of Vial/Planchett	NA
Amount Used (g or mL)	0.1		NA
Standard Activity (DPM/g or mL)	6.67 x 10 ¹⁸ 118.8845 K08181105	Type of Scintillation Vial	NA
Reference Date	1/23/04	Pipette ID Used	1429303
Expiration Date	1/17/10	Balance ID Used	360080204
Residue/Carrier Agent	NA	Quenching Agent	NA

Prepared By: Kelli D. Deere Date: 10/10/09
Reviewed By: Angela G. G. Date: 8/31/09

Rev 1 FEB 9/10/97

ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 - U.S.A.

0638

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

67519-278

Ra-226 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated using a germanium gamma spectrometer system.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

Analytics maintains traceability to the National Institute of Standards and Technology through participation in a Measurements Assurance Program as described in USNRC Reg. Guide 4.15, Revision 1, February 1979.


ISOTOPE:	Ra-226
ACTIVITY (dps):	2.353 E4
HALF-LIFE:	1.600 E3 years
CALIBRATION DATE:	January 23, 2004 12:00 EST
RELATIVE EXPANDED UNCERTAINTY (k=2):	3.3%

Impurities: γ -impurities (other than decay products) <0.1%

5.01065 grams 0.1M HCl solution with 50 μ g/g Ba carrier.

P O NUMBER 3231RD, Item 5

SOURCE PREPARED BY:


M. D. Currie, Radiochemist

Q A APPROVED:



WD 8/3/05

Standard Traceability Log Rad

WARNING! Training must be completed!!

Alphalims will be locked out if training is not completed within 1 week of assignment Contact
Quality if additional time is needed to complete training

A Solution Material Info

Source Material Info	Isotope:	
Prepared By:	Radium-226	Amanda Fehr
Prep Date:	01/16/2006	
Verification Date:	04/09/2009	
Expiration Date:	04/09/2010	
Primary Code:	0638-A	
Dilution(mL):	100 mL	
Mass of Parent(g):	4.8398 g	
Density(g/mL):	1.0266	
Balance ID:	38080204	

Calculations Converting parent activity to dpm/mL/dpm/g

(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / (Ampoule Mass(g) * (Dilution Vol)) = Parent Activity (dpm/mL)
(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / Density / (Ampoule Mass(g) * (Dilution Vol)) = Parent Activity (dpm/g)
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (5.01065 g * 100 mL) = 13636.6133 dpm/mL
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (5.01065 g/mL) / (1.0266 g/mL) = 13282.9676 dpm/g

WU QP8n109

Secondary Standards					
Prep Date	Preparer	Mass Primary Dilution (mL)	Code	Conc dpm/mL	Verification Date
01/17/2006	Amanda Fehr	2.1041	100	0638-B	279.0211 dpm/mL
07/17/2006	Mary Aders	2.1313	100	0638-C	282.6281 dpm/mL
03/28/2007	Daniel Roy	2.1025	100	0638-D	279.2744 dpm/mL
03/28/2007	Daniel Roy	45.468	250	0638-E	2415.7999 dpm/mL
12/18/2007	Daniel Roy	2.014	100	0638-F	267.519 dpm/mL
02/12/2008	Daniel Roy	.5004	100	0638-G	66.468 dpm/mL
07/23/2008	Daniel Roy	5.0607	250	0638-H	268.8845 dpm/mL

GEL Laboratories LLC
Version 1.0 9/18/2000

WY 13015

Verification for Ra-226 Standard 0638-H

D. Roy
7/23/2008

Isotope	Value	Uncertainty
0638-H	11.852	1.1079
0638-H	12.092	1.1141
0638-H	12.372	1.1216

Mean Value (Counting) = 12.106 100.13 Pass
Stdev = 0.260353631 Rule 3 (Pass/Fail)

Target = 12.09
Lower Limit = 11.5848594
Upper Limit = 12.62627393
Rule 1 Pass/Fail Pass
Two sigma = 0.520707263
10 % of Mean = 1.210556667
Rule 2 (Pass/Fail) Pass

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0638-H by transferring portions of the degassed standard into tared glass liquid scintillation vials. 10 mL of DI Water and 10 mL of mineral oil were added to each vial and the vials were shaken. A Blank vial was prepared in a similar fashion using 10 mL of DI Water and 10 mL of mineral oil. The standard verification vials and Background source were dark adapted for two hours and counted on LSC RED using source standard verification. Each verification source calculation was performed as follows:

Source dpm/g = (A - B)/(C)(D)
where:
A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

David Roy 8/14/08
Lisa L. Johnson 8/14/08

LAD 8/13/08

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414
(843)556-8171

Lucas Cell Calibration Package

- 1) Is all calibration standard information enclosed for:
the primary standard certificate?
the secondary standard(s) documentation?
standard preparation information?
standard < 1 Year old or verified?
- 2) Is the efficiency calibration report included ?
- 3) Is the raw count data included for:
Cell constant determination?
Plateau generation?
- 4) Are the calibration verifications included?
- 5) Are the instrument settings included:
HVPS settings?
- 6) Has the CELLEFF.xls file been updated ?
- 7) Have the calibration dates been updated in ALPHALIMS ?

YES	NO	Comments
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	
✓	✓	

Prepared By: Kelli Donnell

Date: 12/19/08

Reviewed By: Mary G. Adens

Date: 12/19/08

Effective Date: 12/19/08

Ra-226 Cell Constants

Standard Reference date: 12/15/1999
 Standard ID: 0299-G
 Volume added (mL): 0.1
 Standard Reference Activity (DPM/mL): 2466.35

Lucas cell #	Cell constant	Standard Source	Date/time of count	Date/time flushed to cell	Date/time end of degas	Date/time	count	Known activity	t1 (days) end-degas	t2 (days) end-flush	t3 (days) Std Ref Date	Decay from Std Ref Date				
201	2.021	Average	1.993	Cal 14	9/15/2008 15:45	9/15/2008 9:05	9/12/2008 13:20	0.267	5596	30	186.53	243.02	0.27778	3198	0.9962	
201	2.043	StdDev	0.068	Cal 14	9/18/2008 13:00	9/18/2008 8:10	9/15/2008 9:05	0.267	5949	30	198.30	243.02	2.96181	0.20139	3201	0.9962
201	1.915			Cal 14	9/25/2008 19:35	9/25/2008 9:15	9/22/2008 10:00	0.267	5361	30	178.70	243.02	2.96875	0.43066	3208	0.9962
202	2.436	Average	2.261	Cal 13	9/15/2008 16:20	9/15/2008 9:35	9/12/2008 13:20	0.267	6779	30	225.97	243.02	2.84975	0.28125	3198	0.9962
202	2.209	StdDev	0.156	Cal 13	9/18/2008 13:50	9/18/2008 8:45	9/15/2008 9:35	0.267	6425	30	214.17	243.02	2.96528	0.21181	3201	0.9962
202	2.137			Cal 14	10/21/2008 13:50	10/20/2008 13:45	10/13/2008 16:00	0.267	9248	30	308.27	243.02	6.90625	1.00347	3234	0.9962
203	2.255	Average	2.254	Cal 43	9/15/2008 16:50	9/15/2008 10:00	9/12/2008 13:20	0.267	6300	30	210.00	243.02	2.86111	0.28472	3198	0.9962
203	2.273	StdDev	0.019	Cal 43	9/18/2008 14:25	9/18/2008 9:15	9/15/2008 10:00	0.267	6613	30	220.49	243.02	2.96875	0.21528	3201	0.9962
203	2.234			Cal 43	9/25/2008 21:00	9/25/2008 10:15	9/22/2008 10:00	0.267	6298	30	209.93	243.02	3.01042	0.44732	3208	0.9962
204	2.194	Average	2.193	Cal 15	9/15/2008 17:25	9/15/2008 10:30	9/12/2008 13:20	0.267	6132	30	204.40	243.02	2.88194	0.28819	3198	0.9962
204	2.300	StdDev	0.102	Cal 15	9/18/2008 14:55	9/18/2008 9:35	9/15/2008 10:30	0.267	6571	30	222.37	243.02	2.96181	0.22222	3201	0.9962
204	2.096			Cal 15	9/30/2008 14:05	9/30/2008 9:10	9/26/2008 9:45	0.133	7535	30	251.17	243.02	3.97569	0.20486	3213	0.9962
205	1.677	Average	1.788	Cal 13	10/21/2008 8:30	10/20/2008 14:05	10/13/2008 16:00	0.267	7584	30	252.80	243.02	6.92014	0.76736	3233	0.9962
205	1.730	StdDev	0.167	Cal 44	9/18/2008 16:00	9/18/2008 10:05	9/15/2008 10:55	0.167	4998	30	166.63	243.02	2.96528	0.24653	3201	0.9962
205	1.990			Cal 44	9/30/2008 14:45	9/30/2008 9:40	9/26/2008 9:45	0.167	7170	30	239.00	243.02	3.98653	0.21181	3213	0.9962
206	2.240	Average	2.259	Cal 46	9/15/2008 21:10	9/15/2008 11:25	9/12/2008 13:20	0.233	6216	30	207.20	243.02	2.92014	0.40625	3198	0.9962
206	2.283	StdDev	0.030	Cal 46	9/18/2008 16:35	9/18/2008 10:25	9/15/2008 11:25	0.267	6604	30	220.13	243.02	2.95833	0.25634	3201	0.9962
206	2.245			Cal 46	9/30/2008 15:20	9/30/2008 10:15	9/26/2008 9:45	0.267	8125	30	270.83	243.02	4.02083	0.21181	3213	0.9962
207	2.187	Average	2.146	Cal 36	9/15/2008 21:40	9/15/2008 11:50	9/12/2008 13:20	0.267	6094	30	203.13	243.02	2.93750	0.40972	3198	0.9962
207	2.141	StdDev	0.038	Cal 36	9/18/2008 17:55	9/18/2008 10:40	9/15/2008 11:50	0.267	6105	30	203.50	243.02	2.95139	0.30208	3201	0.9962
207	2.110			Cal 36	9/30/2008 16:00	9/30/2008 10:45	9/26/2008 9:45	0.233	7656	30	255.20	243.02	4.04167	0.21875	3213	0.9962
208	2.239	Average	2.253	Cal 39	9/15/2008 22:15	9/15/2008 12:13	9/12/2008 13:20	0.267	6559	30	200.60	243.02	2.65466	0.44167	3198	0.9962
208	2.243	StdDev	0.135	Cal 39	9/18/2008 16:30	9/18/2008 11:40	9/15/2008 12:15	0.133	6374	30	212.47	243.02	2.94792	0.61426	3201	0.9962
208	2.146			Cal 39	9/30/2008 16:35	9/30/2008 11:40	9/15/2008 11:40	0.133	6594	30	230.03	243.02	4.98599	0.98599	3213	0.9962
209	2.471	Average	2.291	Cal 19	9/15/2008 22:45	9/15/2008 13:50	9/12/2008 13:20	0.033	7073	30	235.77	243.02	3.02083	0.37153	3198	0.9962
209	2.212	StdDev	0.137	Cal 19	9/18/2008 19:15	9/18/2008 11:15	9/15/2008 13:50	0.067	6170	30	205.67	243.02	2.89236	0.33338	3201	0.9962
209	2.420			Cal 19	9/30/2008 17:25	9/30/2008 11:40	9/26/2008 9:45	0.100	8795	30	283.17	243.02	4.07986	0.23958	3213	0.9962
210	2.320	Average	2.253	Cal 47	9/15/2008 23:15	9/15/2008 14:15	9/12/2008 13:20	0.033	6665	30	222.17	243.02	3.03819	0.37500	3198	0.9962
210	2.210	StdDev	0.059	Cal 47	9/18/2008 19:45	9/18/2008 11:30	9/15/2008 14:15	0.100	6142	30	204.73	243.02	2.88542	0.34375	3201	0.9962
210	2.230			Cal 47	9/30/2008 18:00	9/30/2008 12:05	9/26/2008 9:45	0.033	8116	30	270.53	243.02	4.09722	0.24653	3213	0.9962
211	2.140	Average	2.171	Cal 37	9/15/2008 23:50	9/15/2008 14:30	9/12/2008 13:20	0.033	6150	30	205.00	243.02	3.04861	0.38889	3198	0.9962
211	2.238	StdDev	0.057	Cal 37	9/18/2008 22:20	9/18/2008 12:35	9/15/2008 14:30	0.133	6207	30	206.90	243.02	2.92014	0.40625	3201	0.9962
211	2.136			Cal 37	9/30/2008 18:30	9/30/2008 13:35	9/26/2008 9:45	0.100	7917	30	263.90	243.02	4.15972	0.20486	3213	0.9962
212	2.405	Average	2.322	Cal 42	9/16/2008 0:20	9/15/2008 14:50	9/12/2008 13:20	0.033	6926	30	230.87	243.02	3.06250	0.38583	3198	0.9962
212	2.315	StdDev	0.081	Cal 42	9/18/2008 22:55	9/18/2008 12:50	9/15/2008 14:50	0.267	6405	30	213.50	243.02	2.91667	0.42014	3201	0.9962
212	2.244			Cal 42	9/30/2008 19:50	9/30/2008 14:00	9/26/2008 9:45	0.267	8287	30	276.23	243.02	4.17708	0.24306	3213	0.9962

12/14/14

12/14/14

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 14	500	9/12/08 1000	9/13/08 0915	9/13/08 1935	201	v	0	5361
Cal 13	500	9/12/08 1000	9/13/08 0950	9/13/08 2010	202	v	2	5845
Cal 43	500	9/12/08 1000	9/13/08 1015	9/13/08 2100	203	v	0	6298
Cal 15	500	9/12/08 1000						
Cal 44	500	9/12/08 1000						
Cal 46	500	9/12/08 1000						
Cal 36	500	9/12/08 1000						
Cal 30	500	9/12/08 1000						
Cal 19	500	9/12/08 1000						
Cal 47	500	9/12/08 1000						
Cal 37	500	9/12/08 1000						
Cal 42	500	9/12/08 1000						

W/M
12/18/08

V/V
12/18/08

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 14	500	9/11/2008 1320	9/11/2008 1325	9/11/2008 1545	201	2	8	5594
Cal 13	500	9/11/2008 1320	9/11/2008 1325	9/11/2008 1620	201	2	8	6719
Cal 43	500	9/11/2008 1320	9/11/2008 1325	9/11/2008 1650	203	2	8	6300
Cal 15	500	9/11/2008 1320	9/11/2008 1325	9/15.08 1725	204	2	8	6132
Cal 44	500	9/11/2008 1320	9/11/2008 1325	9/15.08 1805	205	2	5	6132
Cal 45	500	9/11/2008 1320	9/11/2008 1115	9/15.09 2110	204	2	7	6216
Cal 36	500	9/11/2008 1320	9/11/2008 1150	9/15.08 2140	207	2	8	6094
Cal 30	500	9/11/2008 1320	9/11/2008 1115	9/15.08 2215	208	4	8	6258
Cal 19	500	9/11/2008 1320	9/11/2008 1350	9/15.08 2245	209	2	1	7073
Cal 47	500	9/11/2008 1320	9/11/2008 1415	9/15.08 2315	210	2	1	6665
Cal 57	500	9/11/2008 1320	9/11/2008 1430	9/15.08 2350	211	2	1	6150
Cal 42	500	9/11/2008 1320	9/11/2008 1450	9/16.08 0020	212	2	1	6911
								644121808

Cal 14
9/11/2008
K9
V9

Ra-226 Verification Sheet

July 11th 1900
W.M. Ladd

Ra-226 Verification Sheet

Ra-226 Verification Sheet

101748
new
12/19/04

Verification for Ra-226 Standard 0299-G

		Isotope	Detector CPM	BKG CPM	NET CPM	Detector Eff	Mass. Used (G)	Standard	Source DPM/G
4/2/2008	D. Roy	0299-G N1	2536.9600	52.4000	2484.5600	1.97186	0.5057	2562.667649	
		0299-G N2	2520.2500	52.4000	2467.8500	1.97186	0.5056	2545.935781	
		0299-G N3	2532.5000	52.4000	2480.1000	1.97186	0.5042	2565.677715	
		Mean Value (Counting) =	2558.093715		104.944421	Pass	Average =	2558.093715	
		Stdev =	10.63610098		0.00415782	Rule 3 (Pass/Fail)			

Certificate Value = 2437.6 dpm/mL
 Lower Limit = 2536.821513 dpm/mL
 Upper Limit = 2579.365917 dpm/mL
 *exception taken due to full recovery of standard

Rule 1 Pass/Fail Fail
 Two sigma = 21.27220197 dpm/mL
 10 % of Mean = 255.8093715 dpm/mL
 Rule 2 (Pass/Fail) Pass

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-G by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and ten mLs of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Gold for Radium source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 4/02/08 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 0024. Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A \cdot B)/(C)(D)$$

where:

- A = Ver. source cpm,
- B = BKG cpm,
- C = System efficiency, (cpm/dpm), and
- D = mass used for standard verification.

RAD.SOP.M-001

MCAT 12119108

W.W.Maloy

4/9/08

David Gray 4/10/08

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0299
Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCL
Reference Date:	12/15/1999
Ampoule Mass (g):	5.0368 g
Uncertainty:	+/- 2.5 %
LogBook No:	RC S 027 128

A Solution Material Info	
Isotope:	Radium-226
Prepared By:	Angela Johnson
Prep Date:	09/15/2000
Verification Date:	01/23/2008
Expiration Date:	01/23/2009
Primary Code:	0299-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.6634 g
Density(g/mL):	1.0012
Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL)/ (Dilution Vol) = Parent Activity (dpm/g)

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (100 mL) = 122414.2500 dpm/mL

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (1.0012 g/mL) / (100 mL) = 122273.3377 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	04/02/2008	04/02/2009

GEL Laboratories LLC
Version 1.0 9/18/2000

Rev 1.01.01.00
12/18/2008

General Engineering Laboratories Verification Source Preparation Sheet

Applicable SOP Number GL-001-A-008

Isotope

Date Standards Prepared 4/15/05

Cocktail Type Used NA

Standard ID 2014.0

Matrix of Vial/Planchett N/A

Amount Used (g or ml)

180

Standard Activity (DPM/g or mL) 1146.54

Type of Scintillation Vial

Reference Date 3/13/99

Pipette ID Used 1429303

Expiration Date 4/2109

Balance ID Used 56040216

Residue/Carrier Agent

Quenching Agent NA

www.ijerpi.org

Kelli & Dennis

- 2 -

12 | 19108

Background Bar

Mary G. Blane

Date _____

1219 | 58

Rev 1 RLM 9/10/97

0299

UKAS ACCREDITED CALIBRATION LABORATORY No. 0146

ement Reference time for solution number R4/131/89:

1200 GMT on 15 December 1999

Radioactive concentration of radium-226:

43.75 kilobecquerels per gram of solution

which is equivalent to:

1.183 microcuries per gram of solution

Mass of solution:

5.0368 grams

Total activity of radium-226:

220.4 kilobecquerels

which is equivalent to:

5.956 microcuries

Recommended half life:

1600 years

Method of measurement:

The activity of the solution was measured using a high pressure re-entrant ionisation chamber calibrated with a large number of absolutely standardised solutions.

Calibration date: 15 December 1999

The calibration date is provided for added information only, and must not be confused with the reference date on pages 1 and 2 of the certificate. It is the reference date that must be used in all calculations relating to the values of activity.

Expanded uncertainty in the radioactive concentration quoted above: $\pm 2.5\%$

Combined Type A uncertainty: $\pm 0.2\%$

Combined Type B uncertainty: $\pm 1.3\%$

nonradioactive impurities found by high-resolution gamma ray spectrometry, or in any other examination of the solution, are listed below expressed as percentages of the activity of the principal radionuclide at the reference time.

Chemical composition: Carrier free in 0.5M HCl

Quality assurance requirements: This product meets the quality assurance requirements for achieving traceability to NIST as defined in ANSI N42.22-1995.

Storage time: 1 year = 365.25 days

Notes: At the reference date radium-226 was shown to be in radioactive equilibrium with its daughter nuclides down the decay chain to polonium-214 and thallium-210, the precursors of lead-210. The ionisation chamber was calibrated using a standard supplied by the National Institute of Standards and Technology, Washington DC, USA.

MSL 12/10/08
KOB 12/10/08

Ra-226 WATER

Batch : LCSVER
 Date : 10/31/2008
 Analyst : KSD1

Procedure Code : LUC26RAL
 Parmname : Radium-226
 MDA : 1 pCi/L
 Instrument Used : LUCAS CELL DETECTOR

Bkg Count Time: 30 min

Sample ID	Sample Vol L	Count Time min	Gross counts cts	Cell #	Cell Const.	BKG cpm	Ra-226 MDA pCi/L	Ra-226 RESULT pCi/L	Ra-226 ERROR pCi/L	COUNT DATE/TIME	
										num	count
VER 1	0.500	30	1014	201	1.983	0.267	0.3504	22.1841	1.3817	11/17/2008 15:10	V0
VER 2	0.500	30	1056	202	2.261	0.267	0.3089	20.3702	1.2427	11/17/2008 15:45	
VER 3	0.500	30	726	203	2.254	0.267	0.5419	24.4866	1.8110	10/30/2008 16:05	
VER 4	0.500	30	737	204	2.193	0.267	0.5519	25.3188	1.8580	10/30/2008 18:20	
VER 5	0.500	30	937	205	1.799	0.267	0.3882	22.6936	1.4718	11/17/2008 16:20	
VER 6	0.500	30	780	206	2.259	0.267	0.5373	26.1045	1.8604	10/30/2008 20:20	
VER 7	0.500	30	711	207	2.146	0.267	0.5705	25.2245	1.8858	10/30/2008 22:00	
VER 8	0.500	30	593	208	2.283	0.267	0.5132	16.9552	1.4723	11/20/2008 16:40	
VER 9	0.500	30	630	209	2.291	0.133	0.4042	21.0513	1.6596	10/30/2008 23:40	12/10/08
VER 10	0.500	30	691	210	2.253	0.033	0.2527	23.7356	1.7736	10/31/2008 1:15	
VER 11	0.500	30	1067	211	2.171	0.267	0.3314	22.0840	1.3401	11/17/2008 21:55	
VER 12	0.500	30	648	212	2.322	0.133	0.4223	22.6294	1.7586	10/31/2008 9:15	

Sample ID	Sample Dup	Det #	Run Date	Sample Type	Standard ID	NC	NC units	Recovery/RPD
201		2	11/17/2008 10:20	LCS	0638-F	24.10	pCi/L	92%
202		2	11/17/2008 10:45	LCS	0638-F	24.10	pCi/L	85%
203		2	10/30/2008 11:05	LCS	0638-F	24.10	pCi/L	102%
204		2	10/30/2008 12:30	LCS	0638-F	24.10	pCi/L	105%
205		2	11/17/2008 11:10	LCS	0638-F	24.10	pCi/L	94%
206		2	10/30/2008 13:10	LCS	0638-F	24.10	pCi/L	108%
207		2	10/30/2008 13:25	LCS	0638-F	24.10	pCi/L	105%
208		2	11/20/2008 11:45	LCS	0638-F	24.10	pCi/L	70%
209		2	10/30/2008 14:05	LCS	0638-F	24.10	pCi/L	87%
210		2	10/30/2008 14:25	LCS	0638-F	24.10	pCi/L	98%
211		2	11/17/2008 12:20	LCS	0638-F	24.10	pCi/L	92%
212		2	10/30/2008 14:55	LCS	0638-F	24.10	pCi/L	94%
				dE-EM-COUNT	constant	constant	Net CPM cpm	Ingrowth constant
11/10/2008 15:35	11/17/2008 10:20	162.75	4.83	0.7073	0.9642	1.0019	33.5333	0.6833
11/10/2008 15:35	11/17/2008 10:45	163.17	5.00	0.7083	0.9630	1.0019	34.9333	0.6833
10/27/2008 14:20	10/30/2008 11:05	68.75	5.00	0.4049	0.9630	1.0019	23.9333	0.3907
10/27/2008 14:20	10/30/2008 12:30	70.17	5.83	0.4113	0.9569	1.0019	24.3000	0.3943
11/10/2008 15:35	11/17/2008 11:10	163.58	5.17	0.7092	0.9617	1.0019	30.9667	0.6833
10/27/2008 14:20	10/30/2008 13:10	70.83	7.17	0.4142	0.9473	1.0019	25.7333	0.3931
10/27/2008 14:20	10/30/2008 13:25	71.08	8.58	0.4153	0.9373	1.0019	23.4330	0.3900
11/17/2008 11:10-11/20/2008 11:45		72.58	4.92	0.4219	0.9636	1.0019	17.5000	-0.4073
10/27/2008 14:20	10/30/2008 14:05	71.75	9.58	0.4182	0.9302	1.0019	20.8670	0.3998
10/27/2008 14:20	10/30/2008 14:25	72.08	10.83	0.4197	0.9215	1.0019	23.0003	0.3875
11/10/2008 15:35	11/17/2008 12:20	164.75	9.58	0.7117	0.9302	1.0019	35.3000	0.6633
10/27/2008 14:20	10/30/2008 14:55	72.58	18.33	0.4219	0.8707	1.0019	21.4670	0.3681

1811-1910

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
W W 1	500	11/11/08 1535	11/11/08 1620	11/11/08 1510	201	1	8	1014
2	500	11/11/08 1535	11/11/08 1615	11/11/08 1545	202	2	8	1054
3	500	11/11/08 1535	11/11/08 1610	11/11/08 1620	205	2	8	937
4	500	11/11/08 1535	11/11/08 1615	11/11/08 2050	208	2	8	786
5	500	11/11/08 1535	11/11/08 1610	11/11/08 2120	209	1	8	1200
6	500	11/11/08 1535	11/11/08 1610	11/11/08 2155	211	2	8	1067
7	500	11/11/08 1535	11/11/08 1645	11/11/08 1330	101	1	8	981
8	500	11/11/08 1535	11/11/08 0900	11/11/08 1405	103	1	8	1164
9	500	11/11/08 1535	11/11/08 0920	11/11/08 1445	105	1	8	871
10								
11								
12								
13	500							
14								
15								
16								
17								
18								
19								
20								
21								
22								
23								
24								
25								
26								
27								
28								
29								
30								
31								
32								
33								
34								
35								
36								
37								
38								
39								
40								
41								
42								
43								
44								
45								
46								
47								
48								
49								
50								
51								
52								
53								
54								
55								
56								
57								
58								
59								
60								
61								
62								
63								
64								
65								
66								
67								
68								
69								
70								
71								
72								
73								
74								
75								
76								
77								
78								
79								
80								
81								
82								
83								
84								
85								
86								
87								
88								
89								
90								
91								
92								
93								
94								
95								
96								
97								
98								
99								
100								
101								
102								
103								
104								
105								
106								
107								
108								
109								
110								
111								
112								
113								
114								
115								
116								
117								
118								
119								
120								
121								
122								
123								
124								
125								
126								
127								
128								
129								
130								
131								
132								
133								
134								
135								
136								
137								
138								
139								
140								
141								
142								
143								
144								
145								
146								
147								
148								
149								
150								
151								
152								
153								
154								
155								
156								
157								
158								
159								
160								
161								
162								
163								
164								
165								
166								
167								
168								
169								
170								
171								
172								
173								
174								
175								
176								
177								
178								
179								
180								
181								
182								
183								
184								
185								
186								
187								
188								
189								
190								
191								
192								
193								
194								
195								
196								
197								
198								
199								
200								
201								
202								
203								
204								
205								
206								
207								
208								
209								
210								
211								
212								
213								
214								
215								
216								
217								
218								
219								
220								
221								
222								
223								
224								
225								
226								
227								
228								
229								
230								
231								
232								
233								
234								
235								
236								
237								
238								
239								
240				</				

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
VWR 1	500	10/30/08 1420	10/30/08 1420	10/30/08 1500	101	A	4	452
VWR 2	500	10/30/08 1420	10/30/08 1420	10/30/08 1535	1012	C	4	180
VWR 3	500	10/30/08 1420	10/30/08 1420	10/30/08 1605	1013	B	2	8
VWR 4	500	10/30/08 1420	10/30/08 1420	10/30/08 1620	1014	B	8	726
VWR 5	500	10/30/08 1420	10/30/08 1420	10/30/08 1700	1015	C	2	737
VWR 6	500	10/30/08 1420	10/30/08 1420	10/30/08 1810	1016	D	6	663
VWR 7	500	10/30/08 1420	10/30/08 1420	10/30/08 2020	1017	E	8	780
VWR 8	500	10/30/08 1420	10/30/08 1420	10/30/08 2200	1018	F	2	711
VWR 9	500	10/30/08 1420	10/30/08 1420	10/30/08 2300	1019	G	4	497
VWR 10	500	10/30/08 1420	10/30/08 1420	10/30/08 2340	1020	H	4	630
VWR 11	500	10/30/08 1420	10/30/08 1420	10/31/08 0115	1021	I	1	601
VWR 12	500	10/30/08 1420	10/30/08 1420	10/31/08 0835	1022	J	3	123
VWR 13	500	10/30/08 1420	10/30/08 1420	10/31/08 0915	1023	K	4	648

Verification for Ra-226 Standard 0638-F

D Roy 12/27/2007	Isotope 0638-F N1	Detector CPM 1239.9000	BKG CPM 31.5000	NET CPM 1208.4000	Detector Eff Mass. Used (mL)	Source DPM/mL
	0638-F N2	1222.8000	31.5000	1191.3000	4.624018	1.0000
	0638-F N3	1219.4000	31.5000	1187.9000	4.624018	1.0000
					4.624018	1.0000
Mean Value (Counting) = Sdev =	258.6206772 2.375965421	96.8384646 0.000918707	Pass Rule 3 (Pass/Fail)		Average =	258.6206772
Certificate Value = Lower Limit = Upper Limit = Rule 1 Pass/Fail Two sigma = 10 % of Mean = Rule 2 (Pass/Fail)	267.1 253.8687464 263.3726081 Fail 4.751930843 25.866206772 Pass					
						*exception taken due to full recovery of standard

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0638-F by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and 10 mL Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC YELLOW using source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 12/27/07 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 (0024-A). Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A - B)/(C)(D)$$

where:

- A = Ver. source cpm,
- B = BKG cpm,
- C = System efficiency, (cpm/dpm), and
- D = mass used for standard verification.

Reference RAD SOP M-001

W 12/19/08

W 12/19/08
Amanda L. Lehr 14107

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number

GL-RAD-A-008

Isotope

RA-126

Date Standards Prepared

12/18/07

Cocktail Type Used

NA

Standard ID

0638-F

Matrix of Vial/Planchett

NA

Amount Used (g or mL)

0.1

NA

Standard Activity (DPM/g or mL)

167.519

Type of Scintillation Vial

NA

Reference Date

11/23/04

Pipette ID Used

1429303

Expiration Date

12/20/08

Balance ID Used

3604046

Residue/Carrier Agent

0.1M HCl

Quenching Agent

NA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	Ver 1				
2	Ver 2				
3	Ver 3				
4	Ver 4				
5	Ver 5				
6	Ver 6				
7	Ver 7				
8	Ver 8				
9	Ver 9				
10	Ver 10				
11	Ver 11				
12	Ver 12				

Prepared By:

Kelli Dierel

Date

12/19/08

Reviewed By:

Mary J. Hens

Date


12/19/08

Rev 1 RLM 9/10/97

VoltageCurve det2

Page 1

12/19/08
12/19/08



201	1.993	12/19/2008
202	2.261	12/19/2008
203	2.254	12/19/2008
204	2.193	12/19/2008
205	1.799	12/19/2008
206	2.259	12/19/2008
207	2.146	12/19/2008
209	2.291	12/19/2008
210	2.253	12/19/2008
211	2.171	12/19/2008
212	2.322	12/19/2008

Next
12/19/2008

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414

(843)556-8171

Lucas Cell Calibration Package

	YES	NO	Comments
1) Is all calibration standard information enclosed for: the primary standard certificate? the secondary standard(s) documentation? standard preparation information? standard < 1 Year old or verified?	✓		
2) Is the efficiency calibration report included ?	✓		
3) Is the raw count data included for: Cell constant determination? Plateau generation?	✓		
4) Are the calibration verifications included?	✓		
5) Are the instrument settings included: HVPS settings?	✓		
6) Has the CELLEFF.xls file been updated ?	✓		
7) Have the calibration dates been updated in ALPHALIMS ?	✓		

Prepared By: Kelli Penel

Date: 2/3/01

Reviewed By: Hugh Hens

Date: 2/4/59

Effective Date: 214139

Ra-226 Cell Constants

Standard Reference date: 12/15/1999
Standard ID: 0299-G
Volume added (mL): 0.1
Standard Reference Activity (DPM/ml): 2446.35

Lucas	Cell #	Cell constant	Standard Source	Date/Time of count	Date/time flushed to cell	Date/time end of degas	Bkg cpm	Total counts	Known activity	t1 (days) end-degas	t2 (days) end-flush	t3 (days) Std Ref Date	Decay from to count	
301	1.867	Average	2.021	43	1/20/2009 11:05	1/19/2009 10:10	1/9/2009 15:45	0.267	9355	30	311.83	243.67	9.76736 0.03819	3324 0.9961
301	2.184	StdDev	0.159	43	1/29/2009 11:50	1/28/2009 8:50	1/26/2009 13:00	0.267	6239	30	207.97	243.67	8.82539 0.12500	3333 0.9961
301	2.011			43	1/26/2009 14:35	1/22/2009 9:25	1/22/2009 9:10	0.267	7282	30	242.73	243.67	4.01042 0.21528	3331 0.9961
302	2.082	Average	2.131	13	1/30/2009 11:30	1/26/2009 8:30	1/26/2009 13:00	0.267	7401	30	246.70	243.67	3.815250 0.12500	3334 0.9961
302	2.225	StdDev	0.082	47	1/29/2009 13:30	1/29/2009 9:20	1/26/2009 13:00	0.233	6395	30	211.17	243.67	2.84722 0.17361	3334 0.9961
302	2.086			47	1/26/2009 15:30	1/26/2009 9:55	1/22/2009 9:10	0.267	7555	30	251.83	243.67	4.03125 0.23264	3331 0.9961
303	1.958	Average	2.136	19	1/20/2009 13:40	1/19/2009 11:00	1/6/2009 15:45	0.267	9695	30	323.17	243.67	9.80208 1.11111	3325 0.9961
303	2.218	StdDev	0.154	19	1/22/2009 20:35	1/22/2009 10:05	1/19/2009 15:00	0.267	5938	30	197.83	243.67	2.79514 0.43750	3327 0.9961
303	2.231			19	1/26/2009 17:20	1/26/2009 10:25	1/22/2009 9:10	0.267	8028	30	267.60	243.67	4.05208 0.28819	3331 0.9961
305	1.897	Average	2.057	42	1/20/2009 14:50	1/19/2009 11:35	1/6/2009 15:45	0.200	9357	30	311.90	243.67	9.82639 1.19542	3325 0.9961
305	2.191	StdDev	0.149	42	1/22/2009 21:50	1/22/2009 11:05	1/19/2009 15:00	0.267	5921	30	197.37	243.67	2.83681 0.44792	3327 0.9961
305	2.083			42	1/26/2009 23:00	1/26/2009 11:20	1/22/2009 9:10	0.267	7280	30	242.67	243.67	4.09028 0.48611	3331 0.9961
306	1.730	Average	1.747	44	1/20/2009 15:20	1/19/2009 11:50	1/5/2009 15:45	0.167	8521	30	284.03	243.67	9.83891 1.14583	3325 0.9961
306	1.691	StdDev	0.067	30	1/29/2009 14:30	1/29/2009 10:20	1/26/2009 13:00	0.233	4869	30	162.30	243.67	2.88889 0.17361	3334 0.9961
306	1.821			44	1/26/2009 23:30	1/26/2009 11:50	1/22/2009 9:10	0.267	6387	30	212.90	243.67	4.11111 0.48611	3331 0.9961
307	1.818	Average	1.931	15	1/20/2009 15:50	1/19/2009 12:05	1/6/2009 15:45	0.267	8944	30	298.13	243.67	9.84722 1.15625	3325 0.9961
307	2.085	StdDev	0.145	36	1/30/2009 12:55	1/30/2009 9:10	1/26/2009 13:00	0.267	7442	30	248.07	243.67	3.84028 0.15625	3335 0.9961
308	1.881			15	1/27/2009 0:05	1/26/2009 12:10	1/22/2009 9:10	0.267	6598	30	219.93	243.67	4.12500 0.49653	3331 0.9961
308	2.129	Average	1.950	44	1/29/2009 15:50	1/28/2009 11:05	1/26/2009 13:00	0.133	6149	30	204.97	243.67	2.92014 0.19792	3334 0.9961
308	1.858	StdDev	0.155	14	1/23/2009 9:35	1/22/2009 13:45	1/19/2009 15:00	0.267	4829	30	160.97	243.67	2.94792 0.82639	3327 0.9961
308	1.862			14	1/27/2009 8:30	1/26/2009 13:15	1/22/2009 9:10	0.267	6226	30	207.53	243.67	4.17014 0.80208	3331 0.9961
309	1.857	Average	1.877	13	1/20/2009 17:20	1/19/2009 13:35	1/9/2009 15:45	0.033	9149	30	304.97	243.67	9.90972 1.15625	3325 0.9961
309	1.964	StdDev	0.079	13	1/23/2009 10:30	1/22/2009 14:05	1/19/2009 15:00	0.267	5100	30	170.00	243.67	2.96181 0.85069	3327 0.9961
309	1.810			13	1/27/2009 9:05	1/26/2009 13:30	1/22/2009 9:10	0.267	6046	30	201.53	243.67	4.18056 0.81597	3331 0.9961
311	2.140	Average	2.114	15	1/29/2009 16:40	1/29/2009 11:20	1/26/2009 13:00	0.267	6176	30	205.87	243.67	2.93056 0.22222	3334 0.9961
311	2.212	StdDev	0.114	28	1/23/2009 12:20	1/22/2009 14:25	1/19/2009 15:00	0.267	5698	30	189.93	243.67	2.97569 0.971319	3328 0.9961
311	1.988			28	1/27/2009 10:15	1/26/2009 13:45	1/22/2009 9:10	0.267	6607	30	220.23	243.67	4.19097 0.885417	3331 0.9961
312	1.871	Average	1.944	36	1/20/2009 15:16	1/19/2009 14:10	1/9/2009 15:45	0.100	9135	30	304.50	243.67	9.93403 1.21250	3325 0.9961
312	2.014	StdDev	0.071	14	1/29/2009 17:10	1/29/2009 11:35	1/26/2009 13:00	0.167	5814	30	193.80	243.67	2.94997 0.23264	3334 0.9961
312	1.946			36	1/27/2009 11:10	1/26/2009 14:00	1/22/2009 9:10	0.267	6446	30	214.87	243.67	4.20339 0.88194	3331 0.9961

Ra-226 Verification Sheet

Page 31

3

Ra-226 Verification Sheet

#3

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
600 Cal 43	500	1126109 1300	1126109 0850	1126109 1150	301	3	8	6239
Cal 47	500	1126109 1300	1126109 0910	1126109 1330	302	3	7	6335
Cal 49	500	1126109 1300	1126109 0450	1126109 1400	304	3	2	6472
Cal 50	500	1126109 1300	1126109 1010	1126109 1430	304	3	7	4869
Cal 42	500	1126109 1300	1126109 1045	1126109 1515	307	3	3	5088
Cal 44	500	1126109 1300	1126109 1105	1126109 1550	308	3	3	5109
Cal 45	500	1126109 1300	1126109 1110	1126109 1640	311	3	8	6176
Cal 44	500	1126109 1300	1126109 1135	1126109 1710	312	3	5	5814
Cal 113	500	1126109 1300						
Cal 28	500	1126109 1300						
Cal 36	500		1126109 1300					
Cal 37	500		1126109 1300					

KD
213109
WIA
211151

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 43	500	11/10/09 1545	11/10/09 1010	11/10/09 1105	301	3	8	1355
Cal 47	500	11/10/09 1545	11/10/09 1540	11/10/09 1150	302	3	8	8433
Cal 19	500	11/10/09 1545	11/10/09 1100	11/10/09 1340	303	3	8	13109
Cal 30	500	11/10/09 1545	11/10/09 1440	11/10/09 1470	304	3	8	16213109
Cal 42	500	11/10/09 1545	11/10/09 1135	11/10/09 1450	305	3	5	1357
Cal 44	500	11/10/09 1545	11/10/09 1150	11/10/09 1520	306	3	7	8521
Cal 15	500	11/10/09 1545	11/10/09 1105	11/10/09 1550	307	3	8	8944
Cal 14	500	11/10/09 1545	11/10/09 1515	11/10/09 1645	308	3	3	6938
Cal 13	500	11/10/09 1545	11/10/09 1335	11/10/09 1720	309	3	1	9149
Cal 18	500	11/10/09 1545	11/10/09 1355	11/10/09 1840	311	3	8	8648
Cal 36	500	11/10/09 1545	11/10/09 1410	11/10/09 1916	312	3	1	9135
Cal 37	500	11/10/09 1545						

KD
213109
KD
213109
KD
213109
KD
213109

KD
213109
KD
213109

Ra-226 Verification Sheet

Page	320	Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cat 43	500	114104	1500	1122104	0410	1122104 1525	309	3	8	10408
Cat 44	500	114105	1500	1122105	0405	1122105 1405	302	3	8	6493
Cat 45	500	114106	1500	1122106	0405	1122106 1405	302	3	8	10409
Cat 46	500	114107	1500	1122107	0405	1122107 1405	303	3	8	5938
Cat 47	500	114108	1500	1122108	0405	1122108 1405	304	3	8	5240
Cat 48	500	114109	1500	1122109	0405	1122109 1405	305	3	8	5921
Cat 49	500	114110	1500	1122110	0405	1122110 1405	305	3	8	5593
Cat 50	500	114111	1500	1122111	0405	1122111 1405	306	3	8	10409
Cat 51	500	114112	1500	1122112	0405	1122112 1405	306	3	8	5870
Cat 52	500	114113	1500	1122113	0405	1122113 1405	307	3	8	10409
Cat 53	500	114114	1500	1122114	0405	1122114 1405	307	3	8	5870
Cat 54	500	114115	1500	1122115	0405	1122115 1405	308	3	8	5870
Cat 55	500	114116	1500	1122116	0405	1122116 1405	309	3	8	5100
Cat 56	500	114117	1500	1122117	0405	1122117 1405	311	3	8	5698
Cat 57	500	114118	1500	1122118	0405	1122118 1405	311	3	8	5881
Cat 58	500	114119	1500	1122119	0405	1122119 1405	311	3	8	10409
Cat 59	500	114120	1500	1122120	0405	1122120 1405	311	3	8	5881
Cat 60	500	114121	1500	1122121	0405	1122121 1405	311	3	8	10409
Cat 61	500	114122	1500	1122122	0405	1122122 1405	311	3	8	5881
Cat 62	500	114123	1500	1122123	0405	1122123 1405	311	3	8	10409
Cat 63	500	114124	1500	1122124	0405	1122124 1405	311	3	8	5881
Cat 64	500	114125	1500	1122125	0405	1122125 1405	311	3	8	10409
Cat 65	500	114126	1500	1122126	0405	1122126 1405	311	3	8	5881
Cat 66	500	114127	1500	1122127	0405	1122127 1405	311	3	8	10409
Cat 67	500	114128	1500	1122128	0405	1122128 1405	311	3	8	5881
Cat 68	500	114129	1500	1122129	0405	1122129 1405	311	3	8	10409
Cat 69	500	114130	1500	1122130	0405	1122130 1405	311	3	8	5881
Cat 70	500	114131	1500	1122131	0405	1122131 1405	311	3	8	10409
Cat 71	500	114132	1500	1122132	0405	1122132 1405	311	3	8	5881
Cat 72	500	114133	1500	1122133	0405	1122133 1405	311	3	8	10409
Cat 73	500	114134	1500	1122134	0405	1122134 1405	311	3	8	5881
Cat 74	500	114135	1500	1122135	0405	1122135 1405	311	3	8	10409
Cat 75	500	114136	1500	1122136	0405	1122136 1405	311	3	8	5881
Cat 76	500	114137	1500	1122137	0405	1122137 1405	311	3	8	10409
Cat 77	500	114138	1500	1122138	0405	1122138 1405	311	3	8	5881
Cat 78	500	114139	1500	1122139	0405	1122139 1405	311	3	8	10409
Cat 79	500	114140	1500	1122140	0405	1122140 1405	311	3	8	5881
Cat 80	500	114141	1500	1122141	0405	1122141 1405	311	3	8	10409
Cat 81	500	114142	1500	1122142	0405	1122142 1405	311	3	8	5881
Cat 82	500	114143	1500	1122143	0405	1122143 1405	311	3	8	10409
Cat 83	500	114144	1500	1122144	0405	1122144 1405	311	3	8	5881
Cat 84	500	114145	1500	1122145	0405	1122145 1405	311	3	8	10409
Cat 85	500	114146	1500	1122146	0405	1122146 1405	311	3	8	5881
Cat 86	500	114147	1500	1122147	0405	1122147 1405	311	3	8	10409
Cat 87	500	114148	1500	1122148	0405	1122148 1405	311	3	8	5881
Cat 88	500	114149	1500	1122149	0405	1122149 1405	311	3	8	10409
Cat 89	500	114150	1500	1122150	0405	1122150 1405	311	3	8	5881
Cat 90	500	114151	1500	1122151	0405	1122151 1405	311	3	8	10409
Cat 91	500	114152	1500	1122152	0405	1122152 1405	311	3	8	5881
Cat 92	500	114153	1500	1122153	0405	1122153 1405	311	3	8	10409
Cat 93	500	114154	1500	1122154	0405	1122154 1405	311	3	8	5881
Cat 94	500	114155	1500	1122155	0405	1122155 1405	311	3	8	10409
Cat 95	500	114156	1500	1122156	0405	1122156 1405	311	3	8	5881
Cat 96	500	114157	1500	1122157	0405	1122157 1405	311	3	8	10409
Cat 97	500	114158	1500	1122158	0405	1122158 1405	311	3	8	5881
Cat 98	500	114159	1500	1122159	0405	1122159 1405	311	3	8	10409
Cat 99	500	114160	1500	1122160	0405	1122160 1405	311	3	8	5881
Cat 100	500	114161	1500	1122161	0405	1122161 1405	311	3	8	10409
Cat 101	500	114162	1500	1122162	0405	1122162 1405	311	3	8	5881
Cat 102	500	114163	1500	1122163	0405	1122163 1405	311	3	8	10409
Cat 103	500	114164	1500	1122164	0405	1122164 1405	311	3	8	5881
Cat 104	500	114165	1500	1122165	0405	1122165 1405	311	3	8	10409
Cat 105	500	114166	1500	1122166	0405	1122166 1405	311	3	8	5881
Cat 106	500	114167	1500	1122167	0405	1122167 1405	311	3	8	10409
Cat 107	500	114168	1500	1122168	0405	1122168 1405	311	3	8	5881
Cat 108	500	114169	1500	1122169	0405	1122169 1405	311	3	8	10409
Cat 109	500	114170	1500	1122170	0405	1122170 1405	311	3	8	5881
Cat 110	500	114171	1500	1122171	0405	1122171 1405	311	3	8	10409
Cat 111	500	114172	1500	1122172	0405	1122172 1405	311	3	8	5881
Cat 112	500	114173	1500	1122173	0405	1122173 1405	311	3	8	10409
Cat 113	500	114174	1500	1122174	0405	1122174 1405	311	3	8	5881
Cat 114	500	114175	1500	1122175	0405	1122175 1405	311	3	8	10409
Cat 115	500	114176	1500	1122176	0405	1122176 1405	311	3	8	5881
Cat 116	500	114177	1500	1122177	0405	1122177 1405	311	3	8	10409
Cat 117	500	114178	1500	1122178	0405	1122178 1405	311	3	8	5881
Cat 118	500	114179	1500	1122179	0405	1122179 1405	311	3	8	10409
Cat 119	500	114180	1500	1122180	0405	1122180 1405	311	3	8	5881
Cat 120	500	114181	1500	1122181	0405	1122181 1405	311	3	8	10409
Cat 121	500	114182	1500	1122182	0405	1122182 1405	311	3	8	5881
Cat 122	500	114183	1500	1122183	0405	1122183 1405	311	3	8	10409
Cat 123	500	114184	1500	1122184	0405	1122184 1405	311	3	8	5881
Cat 124	500	114185	1500	1122185	0405	1122185 1405	311	3	8	10409
Cat 125	500	114186	1500	1122186	0405	1122186 1405	311	3	8	5881
Cat 126	500	114187	1500	1122187	0405	1122187 1405	311	3	8	10409
Cat 127	500	114188	1500	1122188	0405	1122188 1405	311	3	8	5881
Cat 128	500	114189	1500	1122189	0405	1122189 1405	311	3	8	10409
Cat 129	500	114190	1500	1122190	0405	1122190 1405	311	3	8	5881
Cat 130	500	114191	1500	1122191	0405	1122191 1405	311	3	8	10409
Cat 131	500	114192	1500	1122192	0405	1122192 1405	311	3	8	5881
Cat 132	500	114193	1500	1122193	0405	1122193 1405	311	3	8	10409
Cat 133	500	114194	1500	1122194	0405	1122194 1405	311	3	8	5881
Cat 134	500	114195	1500	1122195	0405	1122195 1405	311	3	8	10409
Cat 135	500	114196	1500	1122196	0405	1122196 1405	311	3	8	5881
Cat 136	500	114197	1500	1122197	0405	1122197 1405	311	3	8	10409
Cat 137	500	114198	1500	1122198	0405	1122198 1405	311	3	8	5881
Cat 138	500	114199	1500	1122199	0405	1122199 1405	311	3	8	10409
Cat 139	500	114200	1500	1122200	0405	1122200 1405	311	3	8	5881
Cat 140	500	114201	1500	1122201	0405	1122201 1405	311	3	8	10409
Cat 141	500	114202	1500	1122202	0405	1122202 1405	311	3	8	5881
Cat 142	500	114203	1500	1122203	0405	1122203 1405	311	3	8	10409
Cat 143	500	114204	1500	112220						

Ra-226 Verification Sheet

Cell for #3

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 43	500	11/21/09 0910	11/26/09 0115	11/26/09 0455	301	3	00	1182
Cal 47	500	11/21/09 0910	11/26/09 0155	11/21/09 1530	302	3	00	1555
Cal 55	500	11/21/09 0910	11/26/09 0115	11/26/09 0010	303	00	00	8028
Cal 30	500	11/21/09 0910	11/26/09 0150	11/26/09 1645	304	3	00	5162
Cal 42	500	11/21/09 0910	11/26/09 0100	11/26/09 2300	305	3	00	7280
Cal 44	500	11/21/09 0910	11/26/09 0150	11/26/09 2330	306	3	00	6387
Cal 15	500	11/21/09 0910	11/21/09 1710	11/27/09 0005	307	3	00	6098
Cal 14	500	11/21/09 0910	11/26/09 1315	11/27/09 0830	308	3	00	6726
Cal 13	500	11/21/09 0910	11/26/09 1330	11/27/09 0005	309	3	00	6046
Cal 28	500	11/21/09 0910	11/26/09 1345	11/27/09 1015	311	3	00	6007
Cal 34	500	11/21/09 0910	11/26/09 1400	11/27/09 1110	312	3	00	6446
					313			
					314			
					315			
					316			
					317			
					318			
					319			
					320			
					321			
					322			
					323			
					324			
					325			
					326			
					327			
					328			
					329			
					330			
					331			
					332			
					333			
					334			
					335			
					336			
					337			
					338			
					339			
					340			
					341			
					342			
					343			
					344			
					345			
					346			
					347			
					348			
					349			
					350			
					351			
					352			
					353			
					354			
					355			
					356			
					357			
					358			
					359			
					360			
					361			
					362			
					363			
					364			
					365			
					366			
					367			
					368			
					369			
					370			
					371			
					372			
					373			
					374			
					375			
					376			
					377			
					378			
					379			
					380			
					381			
					382			
					383			
					384			
					385			
					386			
					387			
					388			
					389			
					390			
					391			
					392			
					393			
					394			
					395			
					396			
					397			
					398			
					399			
					400			
					401			
					402			
					403			
					404			
					405			
					406			
					407			
					408			
					409			
					410			
					411			
					412			
					413			
					414			
					415			
					416			
					417			
					418			
					419			
					420			
					421			
					422			
					423			
					424			
					425			
					426			
					427			
					428			
					429			
					430			
					431			
					432			
					433			
					434			
					435			
					436			
					437			
					438			
					439			
					440			
					441			
					442			
					443			
					444			
					445			
					446			
					447			
					448			
					449			
					450			
					451			
					452			
					453			
					454			
					455			
					456			
					457			
					458			
					459			
					460			
					461			
					462			
					463			
					464			
					465			
					466			
					467			
					468			
					469			
					470			
					471			
					472			
					473			
					474			
					475			
					476			
					477			
					478			
					479			
					480			
					481			
					482			
					483			
					484			
					485			
					486			
					487			
					488			
					489			
					490			
					491			
					492			
					493			
					494			
					495			
					496			
					497			
					498			
					499			
					500			
					501			
					502			
					503			
					504			
					505			
					506			
					507			
					508			
					509			
					510			
					511			
					512			
					513			
					514			
					515			
					516			
					517			
					518			
					519			
					520			
					521			
					522			
					523			
					524			
					525			
					526			
					527			
					528			
					529			
					530			
					531			
					532			
					533			
					534			
				</td				

Verification for Ra-226 Standard 0299-G

		Isotope	Detector CPM	BKG CPM	NET CPM	Detector Eff	Standard	Source DPM/G
4/2/2008	D. Roy	0299-G N1	2536.9600	52.4000	2484.5600	1.917186	0.5057	2562.667649
		0299-G N2	2520.2500	52.4000	2467.8500	1.917186	0.5056	2545.935781
		0299-G N3	2532.5000	52.4000	2480.1000	1.917186	0.5042	2565.677715
Mean Value (Counting) =		2558.093715		104,944,421	Pass		Average =	2558.093715
StDev =		10.63610098		0.00415782	Rule 3 (Pass/Fail)			
Certificate Value =		2437.6	dpm/mL					
Lower Limit =		2536.821513	dpm/mL					
Upper Limit =		2579.365917	dpm/mL					
Rule 1 Pass/Fail		Fail	*exception taken due to full recovery of standard					
Two sigma =		21.27220197	dpm/mL					
10 % of Mean =		255.8093715	dpm/mL					
Rule 2 (Pass/Fail)		Pass						

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-G by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and ten mLs of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Gold for Radium source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 4/02/08 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 0024. Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A - B)/(C)(D)$$

where:

- A = Ver. source cpm,
- B = BKG cpm,
- C = System efficiency, (cpm/dpm), and
- D = mass used for standard verification.

RAD SOP M-001

10/24/08
S. J. X

GEL Standard Traceability Log Rad

Source Material Info		A Solution Material Info	
Parent Code:	0299	Isotope:	Radium-226
Prepared By:	Angela Johnson	Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCL	Prep Date:	09/15/2000
Reference Date:	12/15/1999	Verification Date:	01/23/2008
Ampoule Mass (g):	5.0368 g	Expiration Date:	01/23/2009
Uncertainty:	+/- 2.5 %	Primary Code:	0299-A
LogBook No:	RC S 027 128	Dilution(mL):	100 mL
		Mass of Parent(g):	4.6634 g
		Density(g/mL):	1.0012
		Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL)/ (Dilution Vol) = Parent Activity (dpm/g)

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (100 mL) = 122414.2500 dpm/mL

(4.6634 g) * (43.75 kBq/g) * (60000 dpm/kBq) / (1.0012 g/mL) / (100 mL) = 122273.3377 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	04/02/2008	04/02/2009

GEL Laboratories LLC
Version 1.0 9/18/2000

LR 2/3/09
LLT 2/4/09

General Engineering Laboratories Verification Source Preparation Sheet

Applicable SOP Number GL-RAD-A-008

Isotope DA-221

Date Standards Prepared 4/5/05

Cocktail Type Used NA

Standard ID 0299-G

Matrix of Vial/Planchett NA

Amount Used (g or ml)

NA

Standard Activity (DPM/g or mL) 2446.347

Type of Scintillation Vial N/A

Reference Date 12/15/99

ANSWER The answer is 1000. The first two digits of the number are 10, so the answer is 1000.

Expiration Date 4/21/09

Balance ID Used 30040216

Residue/Carrier Agent 0.5 M HCl

Quenching Agent NA

Prepared By: Kelli & Novel Date: 2/20/14

Date _____

213109

Reviewed By John H. Adams Date 2/4/99

Date

214159

Rev 1 RLM 9/10/97

0299

UKAS ACCREDITED CALIBRATION LABORATORY No. 0146

ement Reference time for solution number R4/131/89:

1200 GMT on 15 December 1999

Radioactive concentration of radium-226:

43.75 kilobecquerels per gram of solution

which is equivalent to:

1.183 microcuries per gram of solution

Mass of solution:

5.0368 grams

Total activity of radium-226:

220.4 kilobecquerels

which is equivalent to:

5.956 microcuries

Recommended half life:

1600 years

Method of measurement:

The activity of the solution was measured using a high pressure re-entrant ionisation chamber calibrated with a large number of absolutely standardised solutions.

Calibration date: 15 December 1999

The calibration date is provided for added information only, and must not be confused with the reference date on pages 1 and 2 of the certificate. It is the reference date that must be used in all calculations relating to the values of activity.

acy Expanded uncertainty in the radioactive concentration quoted above: $\pm 2.5\%$

Combined Type A uncertainty: $\pm 0.2\%$

Combined Type B uncertainty: $\pm 1.3\%$

nucleic The estimated activities of any radioactive impurities found by high-resolution gamma ray spectrometry, or in any other examination of the solution, are listed below expressed as percentages of the activity of the principal radionuclide at the reference time.

ical Carrier free in 0.5M HCl
osition

arks This product meets the quality assurance requirements for achieving traceability to NIST as defined in ANSI N42.22-1995.

1 year = 365.25 days

At the reference date radium-226 was shown to be in radioactive equilibrium with its daughter nuclides down the decay chain to polonium-214 and thallium-210, the precursors of lead-210. The ionisation chamber was calibrated using a standard supplied by the National Institute of Standards and Technology, Washington DC, USA.

KB 213109
NIST 2101-91

Ra-226 WATER

Batch : LCSVER
Date : 1/2/2009
Analyst : KSD1

Procedure Code : LUC26RAL
Parmname : Radium-226
MDA : 1 pCi/L
Instrument Used : LUCAS CELL DETECTOR

Bkg Count Time: 30 min

Sample ID	Sample Vol L	Count Time min	Gross counts cts	Cell #	Cell Const. num	Ra-226		Ra-226 COUNT DATE/TIME
						BKG cpm	MDA pCi/L	
1	0.500	30	656	301	2.021	0.267	0.4919	1/30/2009 15:05
1	0.500	30	655	302	2.131	0.267	0.5554	2/2/2009 13:40
2	0.500	30	914	303	2.136	0.267	0.4647	1/30/2009 15:40
3	0.500	30	791	305	2.057	0.267	0.4845	23.8718 1.6891 1/30/2009 17:05
4	0.500	30	768	306	1.747	0.267	0.5709	27.2885 1.9605 1/30/2009 17:37
2	0.500	30	720	307	1.931	0.267	0.6113	27.3779 2.0335 2/2/2009 14:15
5	0.500	30	730	308	1.950	0.267	0.5149	23.3957 1.7254 1/30/2009 19:05
6	0.500	30	764	309	1.877	0.267	0.5908	28.0944 2.0238 1/31/2009 10:20
7	0.500	30	594	311	2.114	0.267	0.5510	20.3087 1.6667 2/2/2009 8:25
8	0.500	30	542	312	1.944	0.267	0.8009	26.8983 2.3154

Wet sign: [Signature]

Handwritten notes: b01M2 on 1/30/2009

Sample ID	Cell #	Det #	Run Date	Sample Type	Standard ID	NC	NC units	Recovery/RPD
1	301	3	1/30/2009 10:40	LCS	0638-F	24.10	pCi/L	83%
2	302	3	2/2/2009 9:15	LCS	0638-F	24.10	pCi/L	94%
2	303	3	1/30/2009 11:05	LCS	0638-F	24.10	pCi/L	110%
3	305	3	1/30/2009 11:30	LCS	0638-F	24.10	pCi/L	99%
4	306	3	1/30/2009 11:45	LCS	0638-F	24.10	pCi/L	113%
2	307	3	2/2/2009 9:40	LCS	0638-F	24.10	pCi/L	114%
5	308	3	1/30/2009 12:00	LCS	0638-F	24.10	pCi/L	97%
3	309	3	1/30/2009 13:05	LCS	0638-F	24.10	pCi/L	117%
7	311	3	1/30/2009 13:20	LCS	0638-F	24.10	pCi/L	84%
8	312	3	1/30/2009 13:40	LCS	0638-F	24.10	pCi/L	112%
<hr/>								
DEGASSING DATE/TIME		DE-EMAN. DATE/TIME	DEGASS- DE-EM	dE-EM- COUNT	constant	constant	Net CPM cpm	Ingrowth constant
1/26/2009 16:05	1/30/2009 10:40	90.58	4.42	0.4954	0.9672	1.0019	21.6000	0.4800
1/30/2009 10:00	2/2/2009 9:15	71.25	4.42	0.4160	0.9672	1.0019	21.5667	0.4032
1/26/2009 16:05	1/30/2009 11:05	91.00	4.58	0.4969	0.9660	1.0019	30.1997	0.4809
1/26/2009 16:05	1/30/2009 11:30	91.42	5.58	0.4985	0.9587	1.0019	26.1000	0.4788
1/26/2009 16:05	1/30/2009 11:45	91.67	5.87	0.4995	0.9567	1.0019	25.3330	0.4787
1/30/2009 10:00	2/2/2009 9:40	71.67	4.58	0.4179	0.9660	1.0019	23.7330	0.4044
1/26/2009 16:05	1/30/2009 12:00	91.92	7.08	0.5004	0.9479	1.0019	24.0667	0.4753
1/26/2009 16:05	1/30/2009 13:05	93.00	21.25	0.5045	0.8518	1.0019	25.1997	0.4305
1/26/2009 16:05	1/30/2009 13:20	93.25	28.00	0.5054	0.8095	1.0019	19.5330	0.4099
1/26/2009 16:05	1/30/2009 13:40	93.58	66.75	0.5067	0.6041	1.0019	17.7997	0.3067

SOMK
2/2

#3

Ra-226 Verification Sheet

Sample of ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Vr 1	500	11/20/09 1605	11/20/09 1640	11/20/09 1505	301	3	8	1054
Vr 2	500	11/20/09 1605	11/20/09 1605	11/20/09 1540	303	3	8	914
Vr 3	500	11/20/09 1605	11/20/09 1630	11/20/09 1705	305	3	8	791
Vr 4	500	11/20/09 1605	11/20/09 1645	11/20/09 1737	306	3	8	768
Vr 5	500	11/20/09 1605	11/20/09 1600	11/20/09 1905	308	3	8	730
Vr 6	500	11/20/09 1605	11/20/09 1605	1.31.09 1020	309	3	8	764
Vr 7	500	11/20/09 1605	11/20/09 1620	1.31.09 1720	311	3	8	594
Vr 8	500	11/20/09 1605	11/20/09 1640	11/20/09 1815	312	3	8	542
Vr 9	500	11/20/09 1605						
Vr 10	500		11/20/09 1605					
Vr 11	500		11/20/09 1605					
Vr 12	500		11/20/09 1605					

11/20/09

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
UV 1	500	11/21/04 1000	11/21/04 0415	11/21/04 1340	3001	3	8	155
UV 1	500	11/21/04 1000	11/21/04 0440	11/21/04 1415	3007	3	8	120
Net 1	500	11/20/04 1000	11/20/04 0400	11/20/04 1450	3001	3	8	754

140 213109

140 213109

Verification for Ra-226 Standard 0638-F

D. Roy 2/2/2009	Isotope	Value	Uncertainty
	0638-F #1	24.629	1.7426
	0638-F #2	24.438	1.7557
	0638-F #3	22.791	1.6808
Mean Value (Counting) =	23.953	99.60	Pass
Stdev =	1.010781096		Rule 3 (Pass/Fail)

Target =	24.05
Lower Limit =	21.93100448
Upper Limit =	25.97412886
Rule 1 Pass/Fail	Pass
Two sigma =	2.021562191
10 % of Mean =	2.395256667
Rule 2 (Pass/Fail)	Pass

- Rule 1** = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements
Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.
Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for standard 0638-F using 0.1 mL for each source. Each source was counted using routine Lucas cell procedures. Calibration for 0299-G was used in this verification.

160 24109
H. J. Hart 212109
J. Mandelkern
214109

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number GL-RAD-008 Isotope Rn-226
 Date Standards Prepared 12/12/2007 Cocktail Type Used NIA
 Standard ID 0635-F Matrix of Vial/Planchett NIA
 Amount Used (g or mL) 0.1 mL NIA
 Standard Activity (DPM/g or mL) 267.51 mg dpm/mL NIA
 Reference Date 1/23/2004 Type of Scintillation Vial NIA
 Expiration Date 2/14/09 Pipette ID Used 1429303
 Residue/Carrier Agent 0.1 mL HCl Balance ID Used NIA
 Quenching Agent NIA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	Ver 1				
2	Ver 2				
3	Ver 3				
4	Ver 4				
5	Ver 5				
6	Ver 6				
7	Ver 7				
8	Ver 8				
9	Ver 9				
10	Ver 10				
11	Ver 11				
12	Ver 12				

Prepared By: Kelli b noelle

Date 2/3/09

Reviewed By: May Griswold

Date 2/4/09


Rev 1 RLM 9/10/97

Voltage Curve 1-09

Voltage Curve Ludlum # 3				
Volts	Counts	Date	Time	Detector
0.00	0	1/20/2009	13:45	3
0.05	0	1/20/2009	13:46	3
0.10	0	1/20/2009	13:47	3
0.15	0	1/20/2009	13:48	3
0.20	0	1/20/2009	13:49	3
0.25	0	1/20/2009	14:00	3
0.30	0	1/20/2009	14:01	3
0.35	0	1/20/2009	14:02	3
0.40	0	1/20/2009	14:03	3
0.45	0	1/20/2009	14:04	3
0.50	0	1/20/2009	14:05	3
0.55	3914	1/20/2009	14:06	3
0.60	34392	1/20/2009	14:07	3
0.65	96643	1/20/2009	14:08	3
0.70	128361	1/20/2009	14:09	3
0.75	142888	1/20/2009	14:10	3
0.80	154583	1/20/2009	14:11	3
0.85	163087	1/20/2009	14:12	3
0.90	167801	1/20/2009	14:13	3
0.95	172317	1/20/2009	14:14	3
1.00	188508	1/20/2009	14:15	3

KEA-214109
 160
 2/3/09

Ludlum 3 Voltage Curve



1.1×10^{-9}

301	2.021	2/4/2009
302	2.131	2/4/2009
303	2.136	2/4/2009
305	2.057	2/4/2009
306	1.747	2/4/2009
307	1.931	2/4/2009
308	1.950	2/4/2009
309	1.877	2/4/2009
311	2.114	2/4/2009
312	1.944	2/4/2009

New
2/4/09
160
2/4/09

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414

[843]556-8171

Lucas Cell Calibration Package

	YES	NO	Comments
1) Is all calibration standard information enclosed for: the primary standard certificate? the secondary standard(s) documentation? standard preparation information? standard < 1 Year old or verified?	✓		
2) Is the efficiency calibration report included ?	✓		
3) Is the raw count data included for: Cell constant determination? Plateau generation?	✓		
4) Are the calibration verifications included?	✓		
5) Are the instrument settings included: HVPS settings?	✓		
6) Has the CELLEFF.xls file been updated ?	✓		
7) Have the calibration dates been updated in ALPHALIMS ?	✓		

Prepared By: Kelli Donell

Date: 4/28/09

Reviewed By: Angela J. Ohs

Date: 3/2/09

Effective Date: 3/21/09

Ra-226 Cell Constants

Standard Reference date: 12/15/1999
standard ID: 0.299-G
Volume added (mL): 0.1
Standard Reference Activity (DPM/mL): 2446.35

Lucas cell #	Cell constant	Source	Standard of count	Date/Time of count	Date/Time flushed to cell	degas	bkg	total counts	count	time min	cpm	activity	known	t1 (days)	t2 (days)	t3 (days)	decay from Std Ref Date
401	1.689	Average	1.574	3	2/23/2009 16:15	2/20/2009 10:30	0.267	4580	30	152.67	243.66	2.71181	0.23958	3359	0.9960	0.9960	
401	1.585	StdDev	0.121	3	2/27/2009 13:15	2/27/2009 9:00	0.267	5474	30	182.47	243.66	3.70486	0.17708	3363	0.9960	0.9960	
401	1.448			38	2/25/2009 14:40	2/20/2009 7:55	0.267	5677	30	169.23	243.66	4.60417	0.28125	3361	0.9960	0.9960	
402	2.133	Average	2.118	43	2/23/2009 16:55	2/23/2009 11:05	0.267	5817	30	193.90	243.66	2.73611	0.24306	3359	0.9960	0.9960	
402	2.173	StdDev	0.064	43	2/27/2009 14:10	2/27/2009 9:30	0.267	7507	30	250.23	243.66	3.72569	0.19444	3363	0.9960	0.9960	
402	2.048			15	2/25/2009 15:25	2/20/2009 8:15	0.267	8017	30	267.23	243.66	4.61806	0.29861	3361	0.9960	0.9960	
403	1.475	Average	1.463	7	2/23/2009 18:30	2/23/2009 11:30	0.267	4011	30	133.70	243.66	2.75347	0.29167	3359	0.9960	0.9960	
403	1.495	StdDev	0.039	7	2/27/2009 14:50	2/27/2009 10:00	0.267	5182	30	172.73	243.66	3.74653	0.20139	3363	0.9960	0.9960	
403	1.419			14	2/25/2009 15:55	2/20/2009 8:35	0.267	5562	30	185.40	243.66	4.63194	0.30556	3361	0.9960	0.9960	
404	1.792	Average	1.931	42	2/23/2009 19:05	2/23/2009 13:10	0.267	5005	30	166.83	243.66	2.82292	0.24653	3359	0.9960	0.9960	
404	2.142	StdDev	0.186	42	2/27/2009 15:25	2/27/2009 10:30	0.267	7443	30	248.10	243.66	3.76736	0.20486	3363	0.9960	0.9960	
404	1.859			46	2/25/2009 20:20	2/20/2009 8:55	0.267	7075	30	235.83	243.66	4.64583	0.47569	3361	0.9960	0.9960	
405	2.066	Average	1.903	38	3/2/2009 13:40	3/2/2009 10:30	0.267	8602	30	286.73	243.66	4.85417	0.13194	3366	0.9960	0.9960	
405	1.899	StdDev	0.161	13	2/27/2009 16:00	2/27/2009 10:55	0.267	6612	30	220.40	243.66	3.78472	0.21181	3363	0.9960	0.9960	
405	1.745			47	2/25/2009 20:55	2/25/2009 10:10	0.267	6721	30	224.03	243.66	4.68792	0.44792	3361	0.9960	0.9960	
409	1.805	Average	2.036	30	2/24/2009 0:30	2/23/2009 15:20	0.267	5039	30	167.97	243.66	2.91319	0.36194	3359	0.9960	0.9960	
409	2.153	StdDev	0.200	44	2/3/2009 21:10	2/3/2009 15:00	0.267	7949	30	264.97	243.66	4.17361	0.25694	3363	0.9960	0.9960	
409	2.149			44	2/27/2009 16:35	2/27/2009 11:30	0.267	7516	30	250.53	243.66	3.80903	0.21181	3363	0.9960	0.9960	
410	1.869	Average	1.886	28	2/26/2009 8:50	2/25/2009 13:05	0.267	6838	30	227.93	243.66	4.81944	0.82292	3361	0.9960	0.9960	
410	1.965	StdDev	0.072	15	2/4/2009 8:30	2/3/2009 15:30	0.267	6708	30	223.60	243.67	4.19444	0.79833	3339	0.9960	0.9960	
410	1.824			48	2/24/2009 8:00	2/23/2009 15:40	0.267	4840	30	161.33	243.66	2.92708	0.68056	3359	0.9960	0.9960	
411	1.824	Average	1.824	36	2/24/2009 8:40	2/23/2009 15:55	0.267	4839	30	161.30	243.66	2.92750	0.68792	3359	0.9960	0.9960	
411	1.811	StdDev	0.013	30	2/27/2009 17:45	2/27/2009 12:20	0.267	6357	30	211.90	243.66	3.84375	0.22569	3363	0.9960	0.9960	
411	1.836			9	2/26/2009 9:30	2/25/2009 13:40	0.267	6734	30	224.47	243.66	4.84375	0.82639	3361	0.9960	0.9960	
412	1.947	Average	1.967	34	2/26/2009 10:15	2/25/2009 14:05	0.267	7137	30	237.90	243.66	4.86111	0.84028	3361	0.9960	0.9960	
412	2.131	StdDev	0.156	48	2/27/2009 18:20	2/23/2009 12:45	0.267	7495	30	249.83	243.66	3.86111	0.22264	3363	0.9960	0.9960	
412	1.822			35	2/24/2009 9:40	2/23/2009 16:10	0.267	4818	30	160.60	243.66	2.94792	0.72917	3359	0.9960	0.9960	

EffErr 0.123705 < Put in Machines.xls (Lucas Cell Tab)

Angela Ogle 3/2/09

Valerie Venello 3/2/09

401	1.574	3/2/2009
402	2.118	3/2/2009
403	1.463	3/2/2009
404	1.931	3/2/2009
405	1.903	3/2/2009
409	2.036	3/2/2009
410	1.886	3/2/2009
411	1.824	3/2/2009
412	1.967	3/2/2009

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number GL-RAD-A-008 Isotope Ru 226
 Date Standards Prepared 4/15/09 Cocktail Type Used NA
 Standard ID D294G Matrix of Vial/Planchett NA
 Amount Used (g or ml) 0.1 NA
 Standard Activity (DPM/g or mL) 2446.347 NA
 Reference Date 11/15/09 Type of Scintillation Vial NA
 Expiration Date 4/15/09 Pipette ID Used 1429303
 Residue/Carrier Agent 0.5M HCl Balance ID Used 36040216
 Quenching Agent NA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
3	Cai3				
43	Cai43				
7	Cai7				
41	Cai41				
13	Cai43				
44	Cai44				
30	Cai30				
48	Cai48				
36	Cai36				
35	Cai35				
38	Cai38				
15	Cai15				
14	Cai14				
46	Cai46				
47	Cai47				

Prepared By:

Juli Denee

Date

3/21/09

Reviewed By:

Angie L Ohr

Date

3/21/09

Rev 1 RLM 9/10/97

General Engineering Laboratories Verification Source Preparation Sheet

Applicable SOP Number	GL-PAB-AUDS	Isotope	22N-226
Date Standards Prepared	4/15/09	Cocktail Type Used	NA
Standard ID	B7A99G	Matrix of Vial/Planchett	NA
Amount Used (g or ml)	0.2 100 312/09 0.1		NA
Standard Activity (DPM/g or ml.)	2.446.347	Type of Scintillation Vial	NA
Reference Date	12/15/09	Pipette ID Used	1429305
Expiration Date	4/12/09	Balance ID Used	3604021C
Residue/Carrier Agent	b-SM-HCl	Quenching Agent	NA

Prepared By: Julie Dorree, Date: 12/20/09

Page

312109

Reviewed By: A. K. S. **Date:** _____

Date _____

312109

Reviewed By: _____ Rev. 1 RIM 9/10/97

Standard Traceability Log Rad

340 of 600

Source Material Info		A Solution Material Info	
Parent Code:	0299	Isotope:	Radium-226
Prepared By:	Angela Johnson	Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCL	Prep Date:	09/15/2000
Reference Date:	12/15/1999	Verification Date:	01/23/2008
Ampoule Mass (g)	5.0368 g	Expiration Date:	01/23/2009
Uncertainty:	+/- 2.5 %	Primary Code:	0299-A
LogBook No:	RC S 027 128	Dilution(mL):	100 mL
		Mass of Parent(g):	4.6634 g
		Density(g/mL):	1.0012
		Balance ID:	

Calculations Converting parent activity to dpm/mL/dpm/g

$$(Mass\ of\ parent(g)) * (Parm\ Activity\ (kBq/g)) * (conversion\ dpm\ to\ kBq) / (Dilution\ Vol) = Parent\ Activity\ (dpm/mL)$$

$$(Mass\ of\ parent(g)) * (Parm\ Activity\ (kBq/g)) * (conversion\ dpm\ to\ kBq) / Density\ (g/mL) * (Dilution\ Vol) = Parent\ Activity\ (dpm/g)$$

$$(4.6634\ g) * (43.75\ kBq/g) * (60000\ dpm/kBq) / (100\ mL) = 122414.2500\ dpm/mL$$

$$(4.6634\ g) * (43.75\ kBq/g) * (60000\ dpm/kBq) / (1.0012\ g/mL) / (100\ mL) = 122273.3377\ dpm/g$$

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	04/02/2008	04/02/2009

8-21-00

Nycomed Amersham plc
Amersham Laboratories

029



Nycomed Amersham plc
Radiation & Radioactivity
Calibration Laboratory
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

ISSUED
FOR:

AEA Technology plc
Isotrak
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

ion Principal radionuclide: Radium-226

Product code: RAY44
Solution number: R4/131/89

ment Reference time: 1200 GMT on 15 December 1999

data. Nuclear data quoted on this certificate are taken from the Joint European File, Version 2.2.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, which corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Unless indicated, all other uncertainties are expressed at the confidence level associated with one standard uncertainty.

The format used for the uncertainties in the values of radionuclidic purity is illustrated in the following examples:

6.5(21)	-	6.5 ± 2.1
6.54(21)	-	6.54 ± 0.21
6.543(21)	-	6.543 ± 0.021

Date of
issue

17th December 1999 No 312109

Nycomed
Amersham

Verification for Ra-226 Standard 0299-G

		Standard			
		Source DPM/G	Mass. Used (G)	Detector Eff	NET CPM
4/2/2008	Isotope	BKG CPM	1.917186	2484.5600	2536.9600
D. Roy	0299-G N1	52.4000	0.5057	1.917186	2520.2500
	0299-G N2	52.4000	0.5056	1.917186	2532.5000
	0299-G N3	52.4000	0.5042	1.917186	2480.1000
Mean Value (Counting) =	2558.093715	104.944421	Average =	2558.093715	0.00415782
Stdev =	10.63610098	Pass			Rule 3 (Pass/Fail)
Certificate Value =	2437.6	dpm/mL			
Lower Limit =	2536.821513	dpm/mL			
Upper Limit =	2579.365917	dpm/mL			
Rule 1 Pass/Fail	Fail	*exception taken due to full recovery of standard			
Two sigma =	21.27220197	dpm/mL			
10 % of Mean =	255.8093715	dpm/mL			
Rule 2 (Pass/Fail)	Pass				

Verification Rules

Rule 1 = The certificate value (NOT Including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-G by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and ten mLs of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Gold for Radium source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 4/02/08 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 0024. Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A - B)/(C)(D)$$

where:

- A = Ver. source cpm,
- B = BKG cpm,
- C = System efficiency, (cpm/dpm), and
- D = mass used for standard verification.

RAD SOP M-001

4/10/08
Dawn Dry
M/NVR

Ra-226 Verification Sheet

卷之三

Ra-226 Verification Sheet

Call #11

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cell 13	500	11/20/09 11:15	11/23/09 10:30	11/23/09 14:15	401	4	0	1480
43	500	11/20/09 17:25	11/23/09 11:05	11/23/09 11:55	402	4	0	1517
7	500	11/20/09 17:25	11/23/09 11:30	2-24-09 18:20	403	4	0	4011
41	500	11/20/09 17:25	11/23/09 13:10	2-23-09 19:05	404	4	0	5005
41	500	11/20/09 17:25	11/23/09 13:10	2-23-09 19:35	405	4	0	4224
41	500	11/20/09 17:25	11/23/09 13:40	2-23-09 19:55	406	4	0	2155
41	500	11/20/09 17:25	11/23/09 14:05	2-23-09 21:50	407	4	0	2598
41	500	11/20/09 17:25	11/23/09 14:05	2-23-09 23:30	408	4	0	1359
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 00:00	409	4	0	4829
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 00:30	410	4	0	4840
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 01:00	411	4	0	4829
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 01:30	412	4	0	4829
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 02:00	413	4	0	4829
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 02:30	414	4	0	4829
41	500	11/20/09 17:25	11/23/09 14:05	2-24-09 03:00	415	4	0	4829
35	500	11/20/09 17:25	11/23/09 14:10	11/23/09 04:10	416	4	0	4818

Ra-226 Verification Sheet

丁
七

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
SCU1 28	500	10/10/17 17:25	10/10/17 19:09	10/10/17 19:40	401	4	3	56,977
15	500	10/10/17 17:25	10/10/17 19:15	10/10/17 19:25	401	4	3	80,117
14	500	10/10/17 17:25	10/10/17 19:35	10/10/17 19:55	402	4	3	55,622
44	500	10/10/17 17:25	10/10/17 19:55	10/10/17 20:20	404	4	3	70,755
41	500	10/10/17 17:25	10/10/17 20:59	10/10/17 20:55	405	4	3	67,211
10	500	10/10/17 17:25	10/10/17 21:01	10/10/17 20:22	406	4	3	100,1
15	500	10/10/17 17:25	10/10/17 21:01	10/10/17 20:47	407	4	3	28,271
13	500	10/10/17 17:25	10/10/17 21:01	10/10/17 20:55	408	4	3	31,377
29	500	10/10/17 17:25	10/10/17 21:11	10/10/17 20:50	409	4	3	51,649
28	500	10/10/17 17:25	10/10/17 21:25	10/10/17 20:50	410	4	3	48,388
9	500	10/10/17 17:25	10/10/17 21:25	10/10/17 20:50	411	4	3	67,344
34	500	10/10/17 17:25	10/10/17 21:25	10/10/17 20:50	412	4	3	71,377

卷之三

Rg.226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cai 45 Cal 3	500	2/23/09 16:55	2/27/09 6:00	2/27/09 7:05	401	4	8	54114
Cai 43	500	2/23/09 16:05	2/27/09 06:30	2/27/09 14:16	402	4	8	1507
Cai 7	500	2/23/09 16:05	2/27/09 10:00	2/27/09 16:50	403	4	8	5182
Cai 41	500	2/23/09 16:05	2/27/09 10:30	2/27/09 15:25	404	4	8	7443
Cai 13	500	2/23/09 16:05	2/27/09 10:55	2/27/09 16:00	405	4	8	6612
Cai 44	500	2/23/09 16:05	2/27/09 11:30	2/27/09 16:35	409	4	8	7516
Cai 9	500	2/23/09 16:05	2/27/09 11:50	2/27/09 17:15	410	4	8	7850
Cai 50	500	2/23/09 16:05	2/27/09 12:20	2/27/09 17:45	411	4	8	2357
Cai 48	500	2/23/09 16:05	2/27/09 17:45	2/27/09 18:20	412	4	8	7495

3121051

Ra-226 Verification Sheet

Cell #4

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cell 34	500	11/15/09 14:00	3/12/09 10:00	3/12/09 13:40	405	4	8	8602

11/15/09
3/12/09


voltage curve -09

Voltage Curve Ludlum # 4				
Volts (K.V.)	Counts	Date	Time	Detector
0.20	0	2/2/2009	9:00	4
0.25	0	2/2/2009	9:00	4
0.30	0	2/2/2009	9:00	4
0.35	0	2/2/2009	9:00	4
0.40	0	2/2/2009	9:00	4
0.45	473	2/2/2009	9:00	4
0.50	25577	2/2/2009	9:00	4
0.55	77365	2/2/2009	9:00	4
0.60	105618	2/2/2009	9:00	4
0.65	122379	2/2/2009	9:00	4
0.70	140073	2/2/2009	9:00	4
0.75	149183	2/2/2009	9:00	4
0.80	140046	2/2/2009	9:00	4
0.85	149183	2/2/2009	9:00	4
0.90	155553	2/2/2009	9:00	4
0.95	161020	2/2/2009	9:00	4
1.00	165182	2/2/2009	9:00	4
1.05	182720	2/2/2009	9:00	4
1.10	215932	2/2/2009	9:00	4
1.15	292211	2/2/2009	9:00	4
1.20	449383	2/2/2009	9:00	4

~~3/2/09~~

W 3/2109

Ludlum 4 Voltage Curve



Ward

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414
(843)556-8171

Lucas Cell Calibration Package

(501-512)

	YES	NO	Comments
1) Is all calibration standard information enclosed for: the primary standard certificate? the secondary standard(s) documentation? standard preparation information? standard < 1 Year old or verified?	✓		
2) Is the efficiency calibration report included ?	✓		
3) Is the raw count data included for: Cell constant determination? Plateau generation?			
4) Are the calibration verifications included?	✓		
5) Are the instrument settings included: HVPS settings?	✓		
6) Has the CELLEFF.xls file been updated ?	✓		
7) Have the calibration dates been updated in ALPHALIMS ?	✓		

Prepared By: Will S. Donner

Date: 3/24/09

Reviewed By: Angel A. Olin

Date: 3/25/09

Effective Date: 3/25/09

Ra-226 Cell Constants

Standard ID: 0239-E
 Volume added (mL): 0.1
 Standard Reference Activity (DPM/mL): 2434.34

Lucas	Cell #	Cell constant	Standard Source	Date/Time of count	Date/time flushed to cell	Date/time end of degas	Total counts	Time min	Activity dpm	Known activity end-degas	t1 (days) end-flush	t2 (days) end-flush	t3 (days)	Std Ref Date	Decay from Std Ref Date to count	
501	1.927	Average	2.087	15	3/6/2009 7:50	3/3/2009 8:15	2/25/2009 14:00	5281	30	176.03	243.03	5.76042	2.98264	3369	0.9960	
501	2.086	Sidev	0.160	9	3/1/2009 10:40	3/1/2009 12:50	3/5/2009 14:00	7611	30	253.70	243.03	4.95139	0.90972	3374	0.9960	
501	2.247	Average	1.878	42	3/12/2009 13:30	3/12/2009 9:10	3/6/2009 15:25	10210	30	340.33	243.03	5.73958	0.18056	3376	0.9960	
502	1.772	Average	1.878	16	3/18/2009 8:25	3/17/2009 12:50	3/10/2009 14:00	7951	30	265.03	243.03	6.95139	0.81597	3381	0.9960	
502	2.045	Sidev	0.146	14	3/11/2009 11:15	3/10/2009 13:20	3/5/2009 14:00	7474	30	249.13	243.03	4.97222	0.91319	3374	0.9960	
502	1.816			19	3/12/2009 14:20	3/12/2009 9:35	3/6/2009 15:25	8243	30	274.77	243.03	5.75694	0.19782	3376	0.9960	
503	1.581	Average	1.601	46	3/6/2009 9:20	3/5/2009 9:20	2/25/2009 14:00	7250	30	241.67	243.03	7.80556	1.00000	3369	0.9960	
503	1.633	Sidev	0.028	42	3/19/2009 20:15	3/19/2009 15:15	3/12/2009 12:10	8282	30	276.07	243.03	7.12847	0.20833	3383	0.9960	
503	1.588			44	3/12/2009 14:50	3/12/2009 10:00	3/6/2009 15:25	7214	30	240.47	243.03	5.77431	0.20139	3376	0.9960	
504	1.592	Average	1.615	47	3/6/2009 10:30	3/5/2009 9:40	2/25/2009 14:00	7262	30	242.07	243.03	7.81944	1.03472	3369	0.9960	
504	1.611	Sidev	0.025	34	3/11/2009 12:30	3/10/2009 14:05	3/5/2009 14:00	5889	30	196.30	243.03	5.00347	0.93403	3375	0.9960	
504	1.641	Average	2.331	16	3/19/2008 20:50	3/19/2009 15:30	3/12/2009 12:10	8310	30	277.00	243.03	7.13889	0.22222	3383	0.9960	
505	2.364			19	3/11/2009 13:00	3/10/2009 14:30	3/5/2009 10:05	2/25/2009 14:00	10654	30	355.13	243.03	7.83681	1.10764	3370	0.9960
505	2.438	Sidev	0.127	23	3/12/2009 17:01	3/12/2009 10:50	3/5/2009 14:00	8924	30	287.47	243.03	5.02083	0.93750	3375	0.9960	
505	2.190	Average	2.004	25	3/6/2009 13:10	3/5/2009 10:30	2/25/2009 14:00	8576	30	285.87	243.03	7.85417	1.11111	3370	0.9960	
506	1.902	Sidev	0.112	47	3/11/2009 13:30	3/10/2009 15:05	3/5/2009 14:00	7804	30	260.13	243.03	5.04514	0.93403	3375	0.9960	
506	2.124	Average	1.985	13	3/12/2009 17:40	3/12/2009 11:15	3/6/2009 15:25	8954	30	288.47	243.03	5.82639	0.26736	3376	0.9960	
507	1.708	Average	1.701	23	3/6/2009 13:45	3/5/2009 10:55	2/25/2009 14:00	7695	30	256.50	243.03	7.87153	1.11806	3370	0.9960	
507	1.722	Sidev	0.024	25	3/11/2009 14:20	3/10/2009 15:27	3/5/2009 14:00	6315	30	210.50	243.03	5.06042	0.95347	3375	0.9960	
507	1.674			43	3/12/2009 18:30	3/12/2009 11:35	3/6/2009 15:25	7535	30	251.17	243.03	5.84028	0.28819	3376	0.9960	
508	1.605	Average	1.534	39	3/6/2009 14:20	3/5/2009 11:25	2/25/2009 14:00	7236	30	241.20	243.03	7.89236	1.12153	3370	0.9960	
508	1.497	Sidev	0.062	44	3/19/2009 21:30	3/19/2009 15:45	3/12/2009 12:10	7561	30	252.03	243.03	7.14931	0.23958	3383	0.9960	
508	1.499			3	3/12/2009 20:45	3/12/2009 12:10	3/6/2009 15:25	6680	30	222.67	243.03	5.86458	0.35764	3376	0.9960	
509	1.730	Average	1.798	28	3/6/2009 14:50	3/5/2009 11:45	2/25/2009 14:00	7795	30	259.83	243.03	7.90625	1.12847	3370	0.9960	
509	1.857	Sidev	0.064	39	3/11/2009 15:25	3/10/2009 16:05	3/5/2009 14:00	6810	30	227.00	243.03	5.08681	0.97222	3375	0.9960	
509	1.806	Average	1.458	36	3/12/2009 21:20	3/12/2009 12:35	3/6/2009 15:25	8049	30	268.30	243.03	5.88194	0.36458	3376	0.9960	
510	1.460	Sidev	0.024	28	3/11/2009 16:05	3/10/2009 16:20	3/5/2009 14:00	5246	30	174.87	243.03	5.09722	0.89858	3375	0.9960	
510	1.481			35	3/12/2009 21:55	3/12/2009 12:50	3/6/2009 15:25	6589	30	219.63	243.03	5.89236	0.37847	3376	0.9960	
511	1.839	Average	1.959	34	3/6/2009 16:30	3/5/2009 13:20	2/25/2009 14:00	8316	30	277.20	243.03	7.97222	1.13194	3370	0.9960	
511	1.995	Sidev	0.106	46	3/11/2009 16:50	3/10/2009 16:35	3/5/2009 14:00	7283	30	242.77	243.03	5.10764	1.01042	3375	0.9960	
511	2.041			37	3/12/2009 22:40	3/12/2009 13:10	3/6/2009 15:25	9068	30	302.27	243.03	5.90625	0.39583	3376	0.9960	
512	1.796	Average	1.956	48	3/11/2009 17:35	3/10/2009 16:50	3/5/2009 14:00	6542	30	218.07	243.03	5.11806	1.03125	3375	0.9960	
512	2.100	Sidev	0.152	38	3/12/2009 23:15	3/12/2009 13:30	3/6/2009 15:25	9322	30	310.73	243.03	5.92014	0.40625	3376	0.9960	
512	1.972			48	3/18/2009 13:00	3/17/2009 14:00	3/10/2009 14:00	8653	30	288.43	243.03	7.00000	0.95833	3382	0.9960	

*Backgrounds are not significant enough to be considered in calculations. ANSI N42.25-1997 (B.2).

**Calibration
Verification Sheet**
Ra-226 3114109
4-19

Cal #5

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 15	500	2/15/09 1450	3/3/09 0815	3/6/09 0150	501	5	8	5181
Cal 14	500	2/15/09 1400	3/3/09 0845	3/6/09 0840	502	5	1	42508
					503	5	10	313109
								68600
Cal 46	500	2/15/09 1400	3/5/09 0940	3/6/09 0900	503	5	3	7150
Cal 47	500	2/15/09 1400	3/5/09 0940	3/6/09 1020	504	5	1	7242
Cal 14	500	2/25/09 1400	3/15/09 1005	3/6/09 1240	505	5	3	10654
Cal 15	500	2/25/09 1450	3/15/09 1050	3/6/09 1316	506	5	8	8574
Cal 13	500	2/25/09 1450	3/15/09 1055	3/6/09 1345	507	5	4	7095
Cal 39	500	2/25/09 1400	3/15/09 1125	3/6/09 1470	508	5	1	1234
Cal 18	500	2/25/09 1400	3/15/09 1145	3/6/09 1450	509	5	8	1795
Cal 9	500	2/25/09 1400	3/15/09 1210	3/6/09 1525	510	5	2	6578
Cal 34	500	2/25/09 1400	3/15/09 1320	3/6/09 1630	511	5	4	8316

31/16
Mallard

Calibration Ra-226 Verification Sheet

1931-1940

1931-1949

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 14	9	3/15/04 14:00	3/11/04 12:50	3/11/04 10:40	501	5	8	7411
Cal 14	500	3/15/04 14:00	3/11/04 11:15	3/11/04 11:15	502	5	5	1414
Cal 15	500	3/15/04 14:00	3/11/04 15:55	3/11/04 15:55	503	5	8	7352
Cal 15	500	3/15/04 14:00	3/11/04 14:05	3/11/04 13:00	504	5	4	9889
Cal 16	500	3/15/04 14:00	3/11/04 14:05	3/11/04 13:00	505	5	2	8924
Cal 17	500	3/15/04 14:00	3/11/04 14:30	3/11/04 13:00	506	5	8	7804
Cal 18	500	3/15/04 14:00	3/11/04 15:05	3/11/04 13:20	507	5	4	6445
Cal 19	500	3/15/04 14:00	3/11/04 15:05	3/11/04 14:10	508	5	4	6810
Cal 20	500	3/15/04 14:00	3/11/04 15:05	3/11/04 14:55	509	5	8	7283
Cal 21	500	3/15/04 14:00	3/11/04 16:30	3/11/04 15:25	510	5	3	5242
Cal 22	500	3/15/04 14:00	3/11/04 16:30	3/11/04 16:30	511	5	8	6542
Cal 23	500	3/15/04 14:00	3/11/04 16:30	3/11/04 17:35	512	5	8	6542

四月九日

Calibration

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cal 42	500	3/16/09 1525	3/12/09 1010	3/12/09 1330	501	5	8	10210
Cal 19	500	3/16/09 1525	3/12/09 1010	3/12/09 1410	502	5	8	8243
Cal 44	500	3/16/09 1525	3/12/09 1006	3/12/09 1450	503	5	2	7114
Cal 40	500	3/16/09 1525	3/12/09 1005	3/12/09 1515	504	5	0	1620
Cal 1	500	3/16/09 1525	3/12/09 1050	3/12/09 1730	505	5	5	9884
Cal 13	500	3/16/09 1525	3/12/09 1115	3/12/09 17140	506	5	8	8254
Cal 43	500	3/16/09 1525	3/12/09 1135	3/12/09 18130	507	5	6	7535
Cal 3	500	3/16/09 1525	3/12/09 1110	3/12/09 20145	508	5	0	6480
Cal 14	500	3/16/09 1525	3/12/09 1235	3/12/09 21120	509	5	8	8049
Cal 35	500	3/16/09 1525	3/12/09 1250	3/12/09 21155	510	5	1	6589
Cal 37	500	3/16/09 1525	3/12/09 1310	3/12/09 22140	511	5	8	9068
Cal 38	500	3/16/09 1525	3/12/09 1330	3/12/09 23115	512	5	5	9322

3114109

**Calibration
Ra-226 Verification Sheet**

Call # S/S

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
Cu118	500	31/10/09 1400	31/11/09 1250	31/10/09 08125	502	5	5	1451
Cu139	500	31/10/09 1400	31/11/09 1325	31/10/09 08555	503	5	5	1855
Cu118	500	31/10/09 1400	31/11/09 1345	31/10/09 08555	504	5	5	1804
Cu140	500	31/10/09 1400	31/11/09 1400	31/10/09 1300	512	5	8	81053
Cu115	500	31/10/09 1400	31/11/09 1527	31/10/09 1420	507	5	4	6315

Ra-226 Calibration Sheet

Standard ID: 02aef-0

Volume Added (mL):

Volume Autenu (ML): 111
Expiration Date: 11/10/09

1909

8-21-00

Nycomed Amersham plc
Amersham Laboratories

0299



Nycomed Amersham plc
Radiation & Radioactivity
Calibration Laboratory
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

ISSUED
FOR:

AEA Technology plc
Isotrak
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

Product code: RAY44
Solution number: R4/131/89

ion Principal radionuclide: Radium-226

ment Reference time: 1200 GMT on 15 December 1999

data Nuclear data quoted on this certificate are taken from the Joint European File, Version 2.2.

ion of The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, which
inties for a t -distribution with $v_{eff} = \infty$ effective degrees of freedom corresponds to a coverage probability of approximately
95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Unless indicated, all other uncertainties are expressed at the confidence level associated with one standard
uncertainty.

The format used for the uncertainties in the values of radionuclidic purity is illustrated in the following examples:

6.5(21)	=	6.5 ± 2.1
6.54(21)	=	6.54 ± 0.21
6.543(21)	=	6.543 ± 0.021

Date of

17th December 1999

Standard Traceability Log Rad

Source Material Info	
Parent Code:	0299
Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCL
Reference Date:	12/15/1999
Ampoule Mass (g):	5.0368 g
Uncertainty:	+/- 2.5 %
LogBook No:	RC S 027 128

A Solution Material Info	
Isotope:	Radium-226
Prepared By:	Angela Johnson
Prep Date:	09/15/2000
Verification Date:	01/23/2008
Expiration Date:	01/23/2009
Primary Code:	0299-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.6634 g
Density(g/mL):	1.0012
Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL) / (Dilution Vol) = Parent Activity (dpm/g)

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (100 \text{ mL}) = 122414.2500 \text{ dpm/mL}$$

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (1.0012 \text{ g/mL}) / (100 \text{ mL}) = 122273.3377 \text{ dpm/g}$$

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	04/02/2008	04/02/2009

GEL Laboratories LLC
Version 1.0 9/18/2000



Verification for Ra-226 Standard 0299-G

Standard					
	Isotope	Detector CPM	BKG CPM	NET CPM	Detector Eff
4/2/2008 D. Roy	0299-G N1	2536.9600	52.4000	2484.5600	1.917186
	0299-G N2	2520.2500	52.4000	2467.8500	1.917186
	0299-G N3	2532.5000	52.4000	2480.1000	1.917186
Mean Value (Counting) =	2558.093715			104.94421	Pass
Stdev =	10.63610098			0.00415782	Rule 3 (Pass/Fail)
Certificate Value =	2437.6	dpm/mL			
Lower Limit =	2536.821513	dpm/mL			
Upper Limit =	2579.3635917	dpm/mL			
Rule 1 Pass/Fail	Fail	*exception taken due to full recovery of standard			
Two sigma =	21.27220197	dpm/mL			
10 % of Mean =	255.8093715	dpm/mL			
Rule 2 (Pass/Fail)	Pass				

Verification Rules

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 10% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-G by transferring portions of the standard into tared glass liquid scintillation vials. One mL of DI Water and ten mLs of Ready Gel liquid scintillation cocktail was added to each vial and the vials were shaken to mix. A Blank vial was prepared in a similar fashion using 1 mL of DI water and 10 mL of Ready Gel cocktail. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Gold for Radium source standard verification. The Ra-226 efficiency calibration which was used for verification calculations was performed on 4/02/08 using source 0024-A (Ra-226). Calibration data is recorded in this logbook under Ra-226 0024. Each verification source calculation was performed as follows:

$$\text{Source dpm/g} = (A - B)/(C)(D)$$

where:

A = Ver. source cpm,

B = BKG cpm,

C = System efficiency, (cpm/dpm), and

D = mass used for standard verification.

4/11/08
David Dry
Hand signed. See photo
4/11/08

General Engineering Laboratories
Verification Source Preparation Sheet
 Calibration

Applicable SOP Number GL-RM0-A-008 Isotope RA-226
 Date Standards Prepared 07/15/09 Cocktail Type Used NA
 Standard ID D999-6 Matrix of Vial/Planchett NA
 Amount Used (g or ml) 0.1 Type of Scintillation Vial NA
 Standard Activity (DPM/g or ml) 14446.347 Pipette ID Used 1429303
 Reference Date 12/15/09 Balance ID Used 316D40216
 Expiration Date 4/12/09 Quenching Agent NA
 Residue/Carrier Agent D5M HCl

	Standard Number	Quenching Vol (uL)/Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
15	Ca115				
46	Ca146				
47	Ca147				
16	Ca116				
25	Ca125				
23	Ca123				
39	Ca139				
28	Ca128				
9	Ca119				
34	Ca134				
42	Ca142				
19	Ca119				
44	Ca144				
7	Ca117				
13	Ca113				

Prepared By: Kelli Dorian Date 3124109

Reviewed By: _____ Date _____


Rev 1 RLM 9/10/97

Voltage

W 3/24/09

Voltage Curve Ludlum # 5				
Volts	Counts	Date	Time	Detector
0.00	0	2/25/2009	9:20	5
0.05	0	2/25/2009	9:20	5
0.10	0	2/25/2009	9:20	5
0.15	0	2/25/2009	9:20	5
0.20	0	2/25/2009	9:20	5
0.25	0	2/25/2009	9:20	5
0.30	0	2/25/2009	9:20	5
0.35	0	2/25/2009	9:20	5
0.40	0	2/25/2009	9:20	5
0.45	0	2/25/2009	9:20	5
0.50	3611	2/25/2009	9:20	5
0.55	47984	2/25/2009	9:20	5
0.60	94752	2/25/2009	9:20	5
0.65	135854	2/25/2009	9:20	5
0.70	155952	2/25/2009	9:20	5
0.75	155696	2/25/2009	9:20	5
0.80	161972	2/25/2009	9:20	5
0.85	168840	2/25/2009	9:20	5
0.90	175598	2/25/2009	9:20	5
0.95	239969	2/25/2009	9:20	5
1.00	397249	2/25/2009	9:20	5

Luddum 5 Voltage Curve



Luddum 5

Ra-226 WATER

Batch : LCSVER
 Date : 2/20/2008
 Analyst : DXM2

Procedure Code : LUC26RAL
 Parmname : Radium-226
 MDA : 1 pCi/L
 Instrument Used : LUCAS CELL DETECTOR

Bkg Count Time: 30 min

Sample ID	Sample Vol L	Count Time min	Gross counts cts	Cell #	Cell Const. num	BKG cpm	Ra-226 MDA pCi/L	Ra-226 RESULT pCi/L	COUNT DATE/TIME	
									Ra-226 ERROR pCi/L	3/16/2009 15:10
Ver 1	0.500	30	766	501	2.087	0.267	0.6041	28.8142	2.0728	3/16/2009 19:25
Ver 2	0.500	30	537	502	1.878	0.167	0.5682	23.0223	1.9747	3/16/2009 20:20
Ver 3	0.500	30	518	503	1.601	0.267	0.8071	25.9035	2.2832	3/20/2009 19:00
Ver 4	0.500	30	701	504	1.615	0.267	0.6021	26.2570	1.9774	3/16/2009 22:00
Ver 5	0.500	30	680	505	2.331	0.033	0.2559	23.5744	1.7758	3/20/2009 19:40
Ver 6	0.500	30	893	506	2.004	0.267	0.4859	27.0593	1.7988	3/16/2009 23:00
Ver 7	0.500	30	488	507	1.701	0.267	0.7287	22.0004	2.0008	3/16/2009 23:30
Ver 8	0.500	30	544	508	1.534	0.033	0.3760	27.7023	2.3344	3/20/2009 20:50
Ver 9	0.500	30	768	509	1.798	0.267	0.5430	25.9694	1.8657	3/17/2009 5:00
Ver 10	0.500	30	432	510	1.458	0.033	0.3700	21.6379	2.0476	3/17/2009 5:35
Ver 11	0.500	30	577	511	1.959	0.267	0.5934	21.2369	1.7694	3/17/2009 6:10
Ver 12	0.500	30	723	512	1.956	0.267	0.5945	26.7349	1.9815	

Sample ID	Sample Dup	Det #	Run Date	Sample Type	Standard ID	NC	NC units	Recovery/RPD
501	5	5	3/16/2009 15:10	LCS	0638-F	24.05	pCi/L	120%
502	5	5	3/16/2009 19:25	LCS	0638-F	24.05	pCi/L	96%
503	5	5	3/16/2009 20:20	LCS	0638-F	24.05	pCi/L	108%
504	5	5	3/20/2009 19:00	LCS	0638-F	24.05	pCi/L	109%
505	5	5	3/16/2009 22:00	LCS	0638-F	24.05	pCi/L	98%
506	5	5	3/20/2009 19:40	LCS	0638-F	24.05	pCi/L	113%
507	5	5	3/16/2009 23:00	LCS	0638-F	24.05	pCi/L	91%
508	5	5	3/16/2009 23:30	LCS	0638-F	24.05	pCi/L	115%
509	5	5	3/20/2009 20:50	LCS	0638-F	24.05	pCi/L	108%
510	5	5	3/17/2009 5:00	LCS	0638-F	24.05	pCi/L	90%
511	5	5	3/17/2009 5:35	LCS	0638-F	24.05	pCi/L	88%
512	5	5	3/17/2009 6:10	LCS	0638-F	24.05	pCi/L	111%
DEGASSING DATE/TIME		DE-EMAN. DATE/TIME	DEGASS-DE-EM	dE-EM-COUNT	constant	constant	Net CPM cpm	Ingrowth constant
3/13/2009 15:30	3/16/2009 9:45	66.25	5.42	0.3936	0.9599	1.0019	25.2667	0.3785
3/13/2009 15:30	3/16/2009 10:10	66.67	9.25	0.3955	0.9325	1.0019	17.7333	0.3695
3/13/2009 15:30	3/16/2009 10:30	67.00	9.83	0.3970	0.9284	1.0019	17.0000	0.3693
3/16/2009 14:00	3/20/2009 13:05	95.08	5.92	0.5122	0.9563	1.0019	23.1000	0.4908
3/13/2009 15:30	3/16/2009 11:25	67.92	10.58	0.4012	0.9232	1.0019	22.6333	0.3711
3/16/2009 14:00	3/20/2009 13:20	95.33	6.33	0.5131	0.9533	1.0019	29.5000	0.4901
3/13/2009 15:30	3/16/2009 13:50	70.33	9.17	0.4120	0.9331	1.0019	15.9997	0.3852
3/13/2009 15:30	3/16/2009 13:50	70.33	9.67	0.4120	0.9296	1.0019	18.1000	0.3837
3/16/2009 14:00	3/20/2009 13:45	95.75	7.08	0.5147	0.9479	1.0019	25.3333	0.4888
3/13/2009 5:30	3/16/2009 14:25	80.92	14.58	0.4571	0.8957	1.0019	14.3667	0.4103
3/13/2009 5:30	3/16/2009 14:45	81.25	14.83	0.4585	0.8941	1.0019	18.9663	0.4107
3/13/2009 5:30	3/16/2009 15:00	81.50	15.17	0.4595	0.8918	1.0019	23.8330	0.4106

Ra-226 Verification Sheet

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count	Cell #	Det #	Background CPM	Total Counts
VUR 1	500	3/13/09 1530	3/16/09 0045	3/16/09 1510	501	5	8	766
VUR 2	500	3/13/09 1530	3/16/09 0010	3/16/09 1510	502	5	85	537
VUR 3	500	3/13/09 1530	3/16/09 0030	3/16/09 2020	503	5	8	518
VUR 4	500	3/13/09 1530	3/16/09 1100	3/16/09 2145	504	5	8	577
VUR 5	500	3/13/09 1530	3/16/09 1125	3/16/09 2200	505	5	8	630
VUR 6	500	3/13/09 1530	3/16/09 1155	3/16/09 2230	506	5	8	707
VUR 7	500	3/13/09 1530	3/16/09 1320	3/16/09 2300	507	5	8	468
VUR 8	500	3/13/09 1530	3/16/09 1350	3/16/09 2330	508	5	8140	544
VUR 9	500	3/13/09 1530	3/16/09 1410	3/17/09 0415	509	5	8	640
VUR 10	500	3/13/09 1530	3/16/09 1425	3/17/09 0315	510	5	8140	432
VUR 11	500	3/13/09 1530	3/16/09 1445	3/17/09 0535	511	5	8	577
VUR 12	500	3/13/09 1530	3/16/09 1500	3/17/09 0610	512	5	8	713

Ra-226 Verification Sheet

Standard ID: 06e38f

Volume Added (ml):

Volume Added (mL): 0.1
Expiration Date: 12/10

W.B. Murray
January 19

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number GL-DMP-A-008 Isotope RA 226
 Date Standards Prepared 1/16/09 Cocktail Type Used NA
 Standard ID 0638-F Matrix of Vial/Planchett NA
 Amount Used (g or mL) 0.1 NA
 Standard Activity (DPM/g or mL) 2107.519 NA
 Reference Date 1/23/04 Pipette ID Used 1429303
 Expiration Date 2/2/10 Balance ID Used 36080204
 Residue/Carrier Agent NA Quenching Agent NA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	Ver 1				
2	Ver 2				
3	Ver 3				
4	Ver 4				
5	Ver 5				
6	Ver 6				
7	Ver 7				
8	Ver 8				
9	Ver 9				
10	Ver 10				
11	Ver 11				
12	Ver 12				

Prepared By:

Kelli Diesel

Date:

3/24/09

Reviewed By:

Angela J. Ogle

Date:

3/25/09

Rev 1 RLM 9/10/97

GEL Standard Traceability Log Rad

Source Material Info	
Parent Code:	0638
Prepared By:	Amanda Fehr
Carrier Conc:	0.1M HCl
Reference Date:	01/23/2004
Ampoule Mass (g):	5.01065 g
Uncertainty:	+/- 3.3 %
LogBook No:	RC-S-037-037

A Solution Material Info	
Isotope:	Radium-226
Prepared By:	Amanda Fehr
Prep Date:	01/16/2006
Verification Date:	03/04/2007
Expiration Date:	03/04/2008
Primary Code:	0638-A
Dilution(mL):	100 mL
Mass of Parent(g):	4.8398 g
Density(g/mL):	1.0266
Balance ID:	38080204

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / (Ampoule Mass(g) *(Dilution Vol)) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / Density / (Ampoule Mass (g) * (Dilution Vol)) = Parent Activity (dpm/g)

(4.8398 g) * (23530 dps) * (60 dpm/dps) / (5.01065 g * 100 mL) = 13636.6133 dpm/mL

(4.8398 g) * (23530 dps) * (60 dpm/dps) / (1.0266 g/mL)/ (5.01065 g * 100 mL) = 13282.9676 dpm/g

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
01/17/2006	Amanda Fehr	2.1041	100	0638-B	279.0211 dpm/mL	01/17/2007	01/17/2008
07/17/2006	Mary Aders	2.1313	100	0638-C	282.6281 dpm/mL	07/26/2006	07/26/2007
03/28/2007	Daniel Roy	2.1025	100	0638-D	279.2744 dpm/ml	04/08/2007	04/08/2008
03/28/2007	Daniel Roy	45.468	250	0638-E	2415.7999 dpm/ml	04/09/2008	04/08/2009
12/18/2007	Daniel Roy	2.014	100	0638-F	267.519 dpm/ml	02/02/2009	02/02/2010
02/12/2008	Daniel Roy	.5004	100	0638-G	66.468 dpm/ml	03/04/2008	03/04/2009
07/23/2008	Daniel Roy	5.0607	250	0638-H	268.8845 dpm/ml	07/23/2008	07/23/2009

Verification for Ra-226 Standard 0638-F

D. Roy 2/2/2009	Isotope	Value	Uncertainty
	0638-F #1	24.629	1.7426
	0638-F #2	24.438	1.7557
	0638-F #3	22.791	1.6808

Mean Value (Counting) =	23.953	99.60	Pass
Stdev =	1.010781096		Rule 3 (Pass/Fail)

Target =	24.05
Lower Limit =	21.93100448
Upper Limit =	25.97412886
Rule 1 Pass/Fail	Pass
Two sigma =	2.021562191
10 % of Mean =	2.395256667
Rule 2 (Pass/Fail)	Pass

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for standard 0638-F using 0.1 mL for each source. Each source was counted using routine Lucas cell procedures. Calibration for 0299-G was used in this verification.

160 3124109

General Engineering Laboratories

2040 Savage Road, Charleston, SC 29414
(843)556-8171

Lucas Cell Calibration Package

(701-712)

- 1) Is all calibration standard information enclosed for:
 - the primary standard certificate?
 - the secondary standard(s) documentation?
 - standard preparation information?
 - standard < 1 Year old or verified?

- 2) Is the efficiency calibration report included ?

- 3) Is the raw count data included for:
 - Cell constant determination?
 - Plateau generation?

- 4) Are the calibration verifications included?

- 5) Are the instrument settings included:
 - HVPS settings?

- 6) Has the CELLEFF.xls file been updated ?

- 7) Have the calibration dates been updated in ALPHALIMS ?

	YES	NO	Comments
1)	✓		
	✓		
	✓		
	✓		
2)	✓		
3)	✓		
	✓		
4)	✓		
5)	✓		
6)	✓		
7)	✓		

Prepared By: Kelli Spangler

Date: 9/30/09

Reviewed By: Augie J. G.

Date: 9/30/09

Effective Date: 9/30/09

Ra-226 Cell Constants

Standard Reference date: 12/15/1999
 Standard ID: Q299-H
 Volume added (mL): 0.1
 Standard Reference Activity (DPM/ml): 2483.21

Lucas cell #	Cell constant	Standard Source	Date/Time of count	Date/time flushed to cell	degas	cpm	counts	min	count	time	Known activity	t1 (days) end-degas	t2 (days) end-flush	t3 (days) to count	Std Ref Date	Decay from Std Ref Date to count
701	2.180	Average	2.107	Cal 12	9/21/2009 17:00	9/21/2009 12:55	9/18/2009 17:00	6158	30	205.27	243.02	2.82986	0.17014	3569	0.9958	
701	2.025	Stdev	0.078	Cal 1	9/15/2009 17:45	9/21/2009 13:45	9/1/2009 10:30	6595	15	439.67	243.02	14.13542	0.166667	3563	0.9958	
701	2.117			Cal 1	9/18/2009 18:15	9/18/2009 13:20	9/15/2009 10:00	3219	15	214.60	243.02	3.13889	0.20486	3566	0.9958	
702	2.101	Average	2.033	Cal 2	9/24/2009 18:05	9/24/2009 14:05	9/21/2009 17:00	3014	15	200.93	243.02	2.87847	0.16667	3572	0.9958	
702	2.020	Stdev	0.063	Cal 2	9/15/2009 18:10	9/15/2009 14:10	9/1/2009 10:30	6583	15	438.87	243.02	14.15278	0.166667	3563	0.9958	
702	1.977			Cal 11	9/21/2009 17:25	9/21/2009 13:20	9/18/2009 17:00	5611	30	187.03	243.02	2.84722	0.17014	3569	0.9958	
703	2.218	Average	2.221	Cal 10	9/21/2009 18:00	9/21/2009 13:45	9/18/2009 17:00	6317	30	210.57	243.02	2.86458	0.17708	3569	0.9958	
703	2.279	Stdev	0.057	Cal 3	9/24/2009 18:25	9/24/2009 14:35	9/21/2009 17:00	3292	15	219.47	243.02	2.89891	0.15972	3572	0.9958	
703	2.165			Cal 3	9/18/2009 19:00	9/18/2009 14:55	9/15/2009 10:00	3364	15	224.27	243.02	3.20466	0.17014	3566	0.9958	
704	2.302	Average	2.235	Cal 9	9/21/2009 18:35	9/21/2009 14:20	9/18/2009 17:00	6599	30	219.97	243.02	2.88889	0.17708	3569	0.9958	
704	2.255	Stdev	0.079	Cal 4	9/24/2009 18:45	9/24/2009 15:00	9/21/2009 17:00	3274	15	218.27	243.02	2.91667	0.15625	3572	0.9958	
704	2.148			Cal 4	9/18/2009 19:15	9/18/2009 15:20	9/15/2009 10:00	3356	15	223.73	243.02	3.22222	0.16319	3566	0.9958	
705	2.032	Average	2.107	Cal 5	9/18/2009 19:40	9/18/2009 15:45	9/15/2009 10:00	3187	15	212.47	243.02	2.323958	0.16319	3566	0.9958	
705	2.090	Stdev	0.084	Cal 5	9/24/2009 19:05	9/24/2009 15:25	9/21/2009 17:00	3050	15	203.33	243.02	2.93403	0.15278	3572	0.9958	
705	2.198			Cal 8	9/21/2009 19:10	9/21/2009 14:45	9/18/2009 17:00	6321	30	210.70	243.02	2.90625	0.18403	3569	0.9958	
706	2.093	Average	2.142	Cal 7	9/21/2009 20:07	9/21/2009 15:05	9/18/2009 17:00	6013	30	200.43	243.02	2.92014	0.20972	3569	0.9958	
706	2.109	Stdev	0.071	Cal 6	9/24/2009 19:25	9/24/2009 15:45	9/21/2009 17:00	3089	15	205.93	243.02	2.94752	0.15278	3572	0.9958	
706	2.223			Cal 6	9/18/2009 19:55	9/18/2009 16:10	9/15/2009 10:00	3505	15	233.67	243.02	3.25684	0.15625	3566	0.9958	
707	2.154	Average	2.275	Cal 7	9/18/2009 20:15	9/18/2009 16:30	9/15/2009 10:00	3406	15	227.07	243.02	3.27083	0.15625	3566	0.9958	
707	2.386	Stdev	0.116	Cal 7	9/24/2009 19:45	9/24/2009 16:05	9/21/2009 17:00	3506	15	233.73	243.02	2.96181	0.15278	3572	0.9958	
707	2.287			Cal 6	9/21/2009 20:35	9/21/2009 15:25	9/18/2009 17:00	6586	30	219.53	243.02	2.93403	0.152528	3569	0.9958	
708	2.253	Average	2.188	Cal 8	9/24/2009 20:00	9/24/2009 16:30	9/21/2009 17:00	3330	15	222.00	243.02	2.98958	0.17917	3572	0.9958	
708	2.110	Stdev	0.180	Cal 1	9/28/2009 18:35	9/28/2009 15:05	9/24/2009 17:00	7591	30	253.03	243.02	3.92014	0.14583	3576	0.9958	
708	1.923			Cal 8	9/18/2009 20:25	9/18/2009 16:50	9/15/2009 10:00	3055	15	203.67	243.02	3.28472	0.14931	3566	0.9958	
709	2.088	Average	2.285	Cal 9	9/18/2009 21:03	9/18/2009 17:15	9/15/2009 10:00	3324	15	221.60	243.02	3.30208	0.15833	3566	0.9958	
709	2.352	Stdev	0.168	Cal 4	9/21/2009 21:50	9/21/2009 16:20	9/18/2009 17:00	6623	30	227.43	243.02	2.97222	0.22917	3569	0.9958	
709	2.400			Cal 9	9/24/2009 20:20	9/24/2009 16:45	9/21/2009 17:00	3554	15	236.93	243.02	2.98958	0.14931	3572	0.9958	
710	2.512	Average	2.409	Cal 3	9/21/2009 22:21	9/21/2009 16:35	9/18/2009 17:00	7291	30	243.03	243.02	2.98624	0.24028	3569	0.9958	
710	2.436	Stdev	0.119	cal 10	9/24/2009 20:50	9/24/2009 17:00	9/21/2009 17:00	3611	15	240.73	243.02	3.00000	0.15972	3572	0.9958	
711	2.212	Average	2.242	Cal 11	9/18/2009 21:37	9/18/2009 17:45	9/15/2009 10:00	3536	15	235.73	243.02	3.32292	0.16111	3566	0.9958	
711	2.302	Stdev	0.052	Cal 11	9/24/2009 22:05	9/24/2009 17:15	9/21/2009 17:00	3395	15	226.33	243.02	3.01042	0.20139	3572	0.9958	
711	2.211			Cal 2	9/21/2009 22:52	9/21/2009 16:55	9/18/2009 17:00	6432	30	214.40	243.02	2.99653	0.247792	3569	0.9958	
712	2.292	Average	2.069	Cal 1	9/21/2009 23:40	9/21/2009 17:10	9/18/2009 17:00	6657	30	221.90	243.02	3.00694	0.27083	3569	0.9958	
712	1.928	Stdev	0.195	Cal 11	9/15/2009 22:15	9/15/2009 17:35	9/1/2009 10:30	6263	15	417.53	243.02	14.29514	0.19444	3563	0.9958	
712	1.988			Cal 12	9/24/2009 22:27	9/24/2009 17:30	9/21/2009 17:00	2938	15	195.87	243.02	3.02093	0.20625	3572	0.9958	

EFERR 0.065186 < Put in Machines.xls (Lucas Cell Tab)

EFERR 0.065186 < Put in Machines.xls (Lucas Cell Tab)

#7

Ra-226 Calibration Sheet

Standard ID: 0191-4

Volume Added (mL): 8.110

Expiration Date: 8/1/10 * 15 min

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 1	500	9/1/09 10:30	9/1/09 13:45	9/15/09 1745	7D1	7	6595
Cal 2	500	9/1/09 10:30	9/1/09 14:10	9/15/09 1810	7D2	7	6583
Cal 3	500	9/1/09 10:30	9/1/09 14:25	9/15/09 1895	103	7	51070
Cal 4	500	9/1/09 10:30	9/1/09 15:15	9/15/09 1900	104	7	6639
Cal 5	9/1/09 10:30	9/1/09 15:40	9/15/09 1915	105	7	5579	
Cal 6	9/1/09 10:30	9/1/09 16:05	9/15/09 1915	106	7	5347	
Cal 7	9/1/09 10:30	9/1/09 16:30	9/15/09 2000	101	1	5370	
Cal 8	9/1/09 10:30	9/1/09 16:45	9/15/09 2030	108	7	6203	
Cal 9	9/1/09 10:30	9/1/09 17:05	9/15/09 2110	110	7	6458	
Cal 10	9/1/09 10:30	9/1/09 17:20	9/15/09 2155	111	7	5935	
Cal 11	500	9/1/09 10:30	9/1/09 1735	9/15/09 2215	112	7	6263

10/1/09

10/1/09

9/1/09
9/1/09

10/1/09

Ra-226 Calibration Sheet

Standard ID: b6m-4

Volume Added (mL): 0.1

Expiration Date: 5/17/10

* 15min

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 1	500	9/18/09 13:20	9/18/09 13:20	9/18/09 18:15	101	1	3219
Cal 2	500	9/18/09 14:05	9/18/09 14:05	9/18/09 18:35	101	1	3420
Cal 3	500	9/18/09 14:55	9/18/09 14:55	9/18/09 19:00	103	1	3364
Cal 4	500	9/18/09 15:20	9/18/09 15:20	9/18/09 19:15	104	1	3356
Cal 5	500	9/18/09 15:45	9/18/09 15:45	9/18/09 19:40	105	1	3187
Cal 6	500	9/18/09 16:10	9/18/09 16:10	9/18/09 19:55	106	1	3505
Cal 7	500	9/18/09 16:55	9/18/09 16:55	9/18/09 20:15	107	1	3404
Cal 8	500	9/18/09 17:30	9/18/09 17:30	9/18/09 20:25	108	1	3055
Cal 9	500	9/18/09 18:05	9/18/09 17:15	9/18/09 21:03	109	1	3324
Cal 10	500	9/18/09 18:50	9/18/09 17:30	9/18/09 21:20	110	1	3135
Cal 11	500	9/18/09 19:25	9/18/09 17:45	9/18/09 21:37	111	1	3536
Cal 12	500	9/18/09 19:50	9/18/09 18:00	9/18/09 21:48	112	1	5663

1009130105

1009130107

1009130105

Ra-226 Calibration Sheet

Standard ID: M144
 Volume Added (mL): 0.1
 Expiration Date: 01/10

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 11	500	9/18/09 1700	9/21/09 1555	9/21/09 1000	701	7	6158
Cal 11	500	9/18/09 1700	9/21/09 1520	9/21/09 1725	702	7	5611
Cal 10	500	9/18/09 1700	9/21/09 1545	9/21/09 1800	703	7	6317
Cal 9	500	9/18/09 1700	9/21/09 1420	9/21/09 1835	704	7	6599
Cal 10	500	9/18/09 1700	9/21/09 1445	9/21/09 1910	705	7	6321
Cal 7	500	9/18/09 1700	9/21/09 1505	9/21/09 2007	706	7	6013
Cal 4	500	9/18/09 1700	9/21/09 1625	9/21/09 2035	107	7	6586
Cal 4	500	9/18/09 1700	9/21/09 1605	9/21/09 2112	708	7	7155
Cal 4	500	9/18/09 1700	9/21/09 1620	9/21/09 2150	109	1	6823
Cal 3	500	9/18/09 1700	9/21/09 1635	9/21/09 2221	110	7	7291
Cal 1	500	9/18/09 1700	9/21/09 1655	9/21/09 2352	111	7	6432
Cal 1	500	9/18/09 1700	9/21/09 1710	9/21/09 2340	112	7	6657

WIP

9/13/09
AT 9/13/09

Ra-226 Calibration Sheet

Standard ID: 0119-11

Volume Added (mL): 0.1

Expiration Date: 07/10

* 15 min counts

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
Cal 1	500	9/24/09 1700	9/24/09 17240	9/24/09 17475	7D1	7	3105
Cal 1	500	9/24/09 1700	9/24/09 1705	9/24/09 1805	102	7	3014
Cal 1	500	9/24/09 1700	9/24/09 1725	9/24/09 1825	103	1	3292
Cal 1	500	9/24/09 1700	9/24/09 1600	9/24/09 1845	104	7	3274
Cal 1	500	9/24/09 1700	9/24/09 1525	9/24/09 1905	105	7	3050
Cal 1	500	9/24/09 1700	9/24/09 1645	9/24/09 1925	106	7	3089
Cal 1	500	9/24/09 1700	9/24/09 1605	9/24/09 1945	107	7	3504
Cal 1	500	9/24/09 1700	9/24/09 1630	9/24/09 2000	108	7	3330
Cal 1	500	9/24/09 1700	9/24/09 1645	9/24/09 2020	109	7	3554
Cal 1	500	9/24/09 1700	9/24/09 1700	9/24/09 2050	110	7	3611
Cal 1	500	9/24/09 1700	9/24/09 1715	9/24/09 2205	111	7	3395
Cal 1	500	9/24/09 1700	9/24/09 1730	9/24/09 2227	112	7	2938

W W 1120109

24/30/09

Ra-226 Calibration Sheet

Standard ID: 09/01-4

Volume Added (mL): 0.1

Volume Added (mL):

Volume Added (mL): 50
Expiration Date:

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Total Counts
MM 1	500	01/18/19 1505	01/18/19 1509	01/18/19 1835	108	7	7591

242

Yours affecly

General Engineering Laboratories
Verification Source Preparation Sheet
 Calibration

Applicable SOP Number GL-RAD-A-008
 Isotope RA-226
 Date Standards Prepared 4/15/05
 Cocktail Type Used NA
 Standard ID 6229-H
 Matrix of Vial/Planchett NA
 Amount Used (g or mL) 0.1
 NA
 Standard Activity (DPM/g or mL) 24B3.2133
 Type of Scintillation Vial NA
 Reference Date 12/15/99
 Pipette ID Used 1429303
 Expiration Date 8/1/10
 Balance ID Used 38080204
 Residue/Carrier Agent 0.1 M HCl
 Quenching Agent NA

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	Cal 1				
2	Cal 2				
3	Cal 3				
4	Cal 4				
5	Cal 5				
6	Cal 6				
7	Cal 7				
8	Cal 8				
9	Cal 9				
10	Cal 10				
11	Cal 11				
12	Cal 12				

Prepared By:

Kelli S. Donnelly

Date

4/12/05

Reviewed By:

Angela Gibson

Date

9/30/09

Rev 1 RLM 9/10/97

ee'd 8-21-00
Nycomed Amersham plc
Amersham Laboratories

0299

CALIBRATION
No. 0146

ISSUED
BY:
Nycomed Amersham plc
Radiation & Radioactivity
Calibration Laboratory
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

ISSUED
FOR:
AEA Technology plc
Isotak
Amersham Laboratories
White Lion Road
Amersham
Buckinghamshire
HP7 9LL

Description Principal radionuclide: Radium-226

Product code: RAY44
Solution number: R4/131/89

Measurement Reference time: 1200 GMT on 15 December 1999

Nuclear data Nuclear data quoted on this certificate are taken from the Joint European File, Version 2.2.

Expression of uncertainty The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, which corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Unless indicated, all other uncertainties are expressed at the confidence level associated with one standard uncertainty.

The format used for the uncertainties in the values of radionuclidic purity is illustrated in the following examples:

6.5(21)	-	6.5 ± 2.1
6.54(21)	-	6.54 ± 0.21
6.543(21)	-	6.543 ± 0.021

Date of
issue

17th December 1999

WD9120109

Nycomed

Standard Traceability Log Rad

Source Material Info		A Solution Material Info	
Parent Code:	0299	Isotope:	Radium-226
Prepared By:	Angela Johnson	Prepared By:	Angela Johnson
Carrier Conc:	0.5 M HCl	Prep Date:	09/15/2000
Reference Date:	12/15/1999	Verification Date:	01/23/2008
Ampoule Mass (g):	5.0368 g	Expiration Date:	01/23/2009
Uncertainty:	+/- 2.5 %	Primary Code:	0299-A
LogBook No:	RC S 027 128	Dilution(mL):	100 mL
		Mass of Parent(g):	4.6634 g
		Density(g/mL):	1.0012
		Balance ID:	

Calculations Converting parent activity to dpm/mL|dpm/g

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / (Dilution Vol) = Parent Activity (dpm/mL)

(Mass of parent(g)) * (Parm Activity (kBq/g)) * (conversion dpm to kBq) / Density (g/mL) / (Dilution Vol) = Parent Activity (dpm/g)

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (100 \text{ mL}) = 122414.2500 \text{ dpm/mL}$$

$$(4.6634 \text{ g}) * (43.75 \text{ kBq/g}) * (60000 \text{ dpm/kBq}) / (100012 \text{ g/mL}) / (100 \text{ mL}) = 122273.3377 \text{ dpm/g}$$

Secondary Standards

Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL	Verification Date	Expiration Date
08/26/2003	Angela Johnson	1.9909	100	0299-E	2434.34 dpm/mL	11/04/2004	11/04/2005
08/26/2003	Angela Johnson	1.9872	100	0299-F	2429.82 dpm/mL	08/26/2004	08/26/2005
04/05/2005	Amanda Fehr	5.0018	250	0299-G	2446.3471 dpm/mL	01/26/2009	01/26/2010
08/07/2009	Mary Aders	5.0767	250	0299-H	2483.2133 dpm/mL	08/07/2009	08/07/2010

WV 9/20/05

Verification for Ra-226 Standard 0299-H

M. Aders 8/7/2009	Isotope	Value	Uncertainty
	0299-H	111.440	2.5408
	0299-H	115.924	2.5878
	0299-H	111.780	2.5407
Mean Value (Counting) =	113.048	101.49	Pass
Stdev =	2.496414563		Rule 3 (Pass/Fail)
Target =	111.39		
Lower Limit =	108.0550709		
Upper Limit =	118.0407291		
Rule 1 Pass/Fail		Pass	
Two sigma =	4.992829126		
10 % of Mean =	11.30479		
Rule 2 (Pass/Fail)		Pass	

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0299-H by transferring portions of the degassed standard into tared glass liquid scintillation vials. 10 mL of DI Water and 10 mL of mineral oil were added to each vial and the vials were shaken. A Blank vial was prepared in a similar fashion using 10 mL of DI Water and 10 mL of mineral oil. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Red using source standard verification. Each verification source calculation was performed as follows:

Source dpm/g = $(A - B)/(C)(D)$
where:
A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

Aug 7th 2009
Amanda Aders 81-3861

Radon-222 Liquid

Filename : RN2222.XLS
 File type : Excel
 Version # : 1.2.4

Batch : 891920
 Analyst : MLA
 Prep Date : 8/7/2009

Rn-222 Abundance : 1

Rn-222 Method Uncertainty : 0.1111

Geometry : 10ML MINERAL OIL/10ML
 SAMPLE

Spike S/N :	N/A	LCS S/N :	0299-H
Spike Exp Date :	N/A	LCS Exp Date :	8/7/2010
Spike Activity (dpm/ml) :	N/A	LCS Activity (dpm/ml) :	2472.85
Spike Volume Added:	N/A	LCS Volume Added:	0.10
Spike Date/Time: 8/7/2009 14:00		Procedure Code :	LSC22RNL
		Paramname :	Radon-222
		Required MDA :	200
		Halflife of Radon-222 :	3.823

Pipet, 0.1 ml Stdev : +/- 0.000701 ml
 Pipet, 0.5 ml Stdev : +/- 0.002564 ml

Sample Characteristics	Sample			Count raw Data			Background		
	Sample ID	Sample Aliquot L	Sample StDev. L	Sample Date/Time	Rack Position #	Counting Time (min.)	Quench#	Gross cpm	Background Count Start Date/Time
1	1201897288.1	1.0000	2.0399E-05	8/7/2009 0:00	8-2	15	43.3	517.53	8.47 15 8/12/2009 7:48
2	1201897289.1	1.0000	2.0399E-05	8/7/2009 0:00	8-3	15	44.6	538.8	8.47 15 8/12/2009 8:04
3	1201897270.1	1.0000	2.0399E-05	8/7/2009 0:00	8-4	15	45	520.6	8.47 15 8/12/2009 8:20

Pos.	Calibration Data		Detector		Backgrounds		Correction Factors		Net Sample		
	Counted on	Calibration Date	Calibration Due Date	Efficiency (cpm/dpm)	Error (cpm/dpm)	Rack Position #	Date/Time	Spike Date/Time	Rn-222 Ingrowth	Rn-222 Count Correction	Activity for MS pCi/l
1	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	8-1	8/12/2009 7:31	8/7/2009 14:00	0.577	0.577	
2	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	8-1	8/12/2009 7:31	8/7/2009 14:00	0.578	0.578	
3	LSCRED	7/28/2009	7/31/2010	3.5654	0.00792	8-1	8/12/2009 7:31	8/7/2009 14:00	0.579	0.579	

Notes:

- Results are decay corrected to Sample Date/Time

- Reference date for Spike Activity (dom/mn) is the batch

Spikes Nominals are decimal corrected to Sample Date/Time

ଓଡ଼ିଆ ଲେଖକ ପରିଚୟ - ୧

Results	1 SIGMA			2 SIGMA			3 SIGMA		
	Decision Level pCi/L	Critical Level pCi/L	Required MDA pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate CPM	Net Count Rate Error CPM	Total Prop. Uncertainty pCi/L
Pos.									Sample QC
1	0.5420	0.3827	200	0.8092	111.4397	0.0141	509.0600	5.9217	24.4606
2	0.5412	0.3821	200	0.8080	115.9238	0.0139	530.3300	6.0403	25.4391
3	0.5404	0.3816	200	0.8068	111.7802	0.0140	512.1300	5.9390	24.5345

ID #: R#4-12345

12 AUG 2009 07:43

USER #: 12 COMMENT: RED

PRESET TIME: 15.00
 DATA CALC: CPM HR YES SAMPLE REPEATS: 1 PRINTER: EDITT
 COUNT PLATE: NO TCA: NO REPLICATE: 1 RS232: EDIT
 TWO PHASE: NO ADU: NO CYCLE REPEATS: 1 DISY: OFF
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REL: 0
 LOW LEVEL: YES HALF LIFE CORRECTION DATE: none


CHAN1: 600.0 - 375.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
 CHAN2: 0.0 - 200.0 %ERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	HR	WIND1		WIND2		WIND1		WIND2		LUMEX %	ELAPSED TIME
				RAW	CPM	RAW	CPM	CPM	%ERROR	CPM	%ERROR		
1	3-1	15.00	37.1	32.47	27.73	8.47	17.75	27.73	9.81	0.47	11.1		
2	3-2	15.00	43.3	517.53	407.33	517.53	2.27	607.33	2.10	6.97	11.1		
3	3-3	15.00	48.6	539.80	428.67	538.80	2.22	628.67	2.06	6.97	11.1		
4	3-4	15.00	45.0	520.40	610.00	520.60	2.24	610.00	2.09	6.97	11.1		


Sample Count Start Time: 12 Aug 2009 07:31:52
Data Capture Date 12 Aug 2009 07:47:25
User Filename S12081208-1A.XLS
U12081208-1A.XLS
Spectrum Type Log Counts
User Number 12
User Id RN-222
User Comment RED
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 1 8-1 15.00
H#, Total Counts: 39.1 422
Start, End, X-Axis: 0 990 Channel Number

SPECTRUM PLOT
USER 12 - RN-222




Sample Count Start Time: 12 Aug 2009 07:48:04
Data Capture Date 12 Aug 2009 08:03:28
User Filename S12081208-2A.XLS
U12081208-1A.XLS
Spectrum Type Log Counts
User Number 12
User Id RN-222
User Comment RED
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 2 8-2 15.00
H#, Total Counts: 43.3 9166
Start, End, X-Axis: 0 990 Channel Number

S P E C T R U M P L O T
USER 12 - RN-222



Sample Count Start Time: 12 Aug 2009 08:04:11
Data Capture Date 12 Aug 2009 08:19:35
User Filename S12081208-3A.XLS
U12081208-1A.XLS


Spectrum Type Log Counts
User Number 12
User Id RN-222
User Comment RED
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 3 8-3 15.00
H#, Total Counts: 44.6 9492
Start, End, X-Axis: 0 990 Channel Number



Sample Count Start Time: 12 Aug 2009 08:20:17
Data Capture Date 12 Aug 2009 08:35:41
User Filename S12081208-4A.XLS
U12081208-1A.XLS

Spectrum Type Log Counts
User Number 12
User Id RN-222
User Comment RED
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 4 8-4 15.00
H#, Total Counts: 45.0 9197
Start, End, X-Axis: 0 990 Channel Number

S P E C T R U M P L O T
USER 12 - RN-222



08/07/2009

Batch #: 891920
 Spike Isotope: Radium-226 Spike Code: 0
 LCS Isotope: Radium-226 LCS Code: D2C1G-H
 Prep Date: 2011-07-01 Pipet ID: 21100008 Initials: AS Witness:

Analyst:MLA

First Client Due Date:

Expiration Date: 2011-07-01 Vol: Nom Conc: Expiration Date: 2011-07-01 Vol: Nom Conc:

Comments

Internal Due Date 08/17/2009

Sample ID	Client Description	Type	Hazard	Min	Code	CRDL	Matrix	Client	Collection Date	Label	Wet/Dry Sample Mass (g/mL)	LSC Rack #	Time Spike Added
1201897268-1	LCS for batch 891920	LCS		.2	pCi/mL		DRINKING WATQC ACCOUNT		20-JUL-09 12:00 PM				14°C
1201897269-1	LCS for batch 891920	LCS		.2	pCi/mL		DRINKING WATQC ACCOUNT		20-JUL-09 12:00 PM				14°C
1201897270-1	LCS for batch 891920	LCS		.2	pCi/mL		DRINKING WATQC ACCOUNT		20-JUL-09 12:00 PM				14°C

Bkg Rack #:

Comments: _____

Data Reviewed By: _____

Instrument Used: LS6500 (Red) 7065155, LS6500 (Black) 7069123, LS6500 (Blue) 7067083, LS6500 (Green) 7067404
 Wallac (Yellow) 4040127, Wallac (Pink) 22000082, Purple 7069123, Silver 70600656

Voltage Curve Ludlum #7


Voltage (kV)	Count Time (min)	Counts	Date/Time
0.20	1.00	0	9/15/09 12:13
0.25	1.00	0	9/15/09 12:14
0.30	1.00	0	9/15/09 12:15
0.35	1.00	0	9/15/09 12:16
0.40	1.00	3788	9/15/09 12:17
0.45	1.00	41827	9/15/09 12:18
0.50	1.00	87578	9/15/09 12:19
0.55	1.00	119153	9/15/09 12:20
0.60	1.00	136757	9/15/09 12:21
0.65	1.00	146242	9/15/09 12:22
0.70	1.00	154066	9/15/09 12:23
0.75	1.00	160997	9/15/09 12:24
0.80	1.00	164506	9/15/09 12:25
0.85	1.00	168023	9/15/09 12:26
0.90	1.00	171900	9/15/09 12:27
0.95	1.00	174082	9/15/09 12:28
1.00	1.00	181331	9/15/09 12:29
1.05	1.00	211928	9/15/09 12:30
1.10	1.00	293552	9/15/09 12:31
1.15	1.00	544079	9/15/09 12:32
1.20	1.00	827973	9/15/09 12:33
1.25	1.00	1214090	9/15/09 12:34

Detector set to operate at 0.70 kV

49
9/30/09

Ludlum Detector Voltage Curve

—◆— Voltage Curve Ludlum #7



DAILY CALIBRATION RANGE

Trial	Counts	Date	Time	Detector	STATISTICS
1	154335	9/15/2009	13:30	7	
2	153698	9/15/2009	13:31	7	
3	153933	9/15/2009	13:32	7	
4	154196	9/15/2009	13:33	7	
5	154114	9/15/2009	13:34	7	Average 150711.30
6	153766	9/15/2009	13:35	7	St. Dev. 3407.47
7	154409	9/15/2009	13:36	7	+ 3 S.D. 160933.72
8	154086	9/15/2009	13:37	7	+ 2 S.D. 157526.25
9	153833	9/15/2009	13:38	7	Average 150711.30
10	153689	9/15/2009	13:39	7	- 2 S.D. 143896.35
11	148183	9/16/2009	10:25	7	- 3 S.D. 140488.88
12	148142	9/16/2009	10:35	7	UPPER 160934
13	148193	9/16/2009	10:36	7	LOWER 140489
14	147463	9/16/2009	10:37	7	
15	147251	9/16/2009	10:39	7	
16	146697	9/17/2009	4:25	7	
17	146925	9/17/2009	5:45	7	
18	147238	9/17/2009	6:00	7	
19	147239	9/17/2009	6:15	7	
20	146836	9/17/2009	6:30	7	

X ^a₉ 13010⁹

701	2.107	9/30/2009
702	2.033	9/30/2009
703	2.221	9/30/2009
704	2.235	9/30/2009
705	2.107	9/30/2009
706	2.142	9/30/2009
707	2.275	9/30/2009
708	2.188	9/30/2009
709	2.285	9/30/2009
710	2.409	9/30/2009
711	2.242	9/30/2009
712	2.069	9/30/2009

X¹₉ 9/30/09

Ra-226 WATER

Batch : LCSVER
 Date : 9/22/2009
 Analyst : KSD1
 Bkg Count Time: 30 min

Procedure Code : LUC26RAL
 Parmname : Radium-226
 MDA : 1 pCi/L
 Instrument Used : LUCAS CELL DETECTOR

Sample ID	Sample Vol L	Count Time min	Gross counts cts	Cell #	Cell Const. num	BKG cpm	Ra-226 MDA pCi/L	Ra-226 RESULT pCi/L	Ra-226 ERROR pCi/L	COUNT DATE/TIME
VER 1	0.500	15	636	701	2.107	0.267	0.5512	24.6163	1.9283	9/30/2009 9:20
VER 2	0.500	15	625	702	2.033	0.267	0.5247	27.0835	2.1404	9/29/2009 16:10
VER 3	0.500	15	625	703	2.221	0.267	0.4811	24.8342	1.9627	9/29/2009 16:45
VER 4	0.500	15	587	704	2.235	0.267	0.4786	23.1944	1.8925	9/29/2009 17:15
VER 5	0.500	15	511	705	2.107	0.267	0.5081	21.4146	1.8751	9/29/2009 17:50
VER 6	0.500	15	580	706	2.142	0.267	0.4998	23.9310	1.9645	9/29/2009 18:25
VER 7	0.500	15	539	707	2.275	0.267	0.4643	20.6372	1.7586	9/29/2009 18:40
VER 8	0.500	15	525	708	2.188	0.267	0.4816	20.8572	1.8013	9/29/2009 19:00
VER 9	0.500	15	559	709	2.285	0.267	0.4615	21.2888	1.7807	9/29/2009 19:40
VER 10	0.500	15	694	710	2.409	0.267	0.4093	23.4767	1.7593	9/30/2009 9:50
VER 11	0.500	15	537	711	2.242	0.267	0.4690	20.7776	1.7739	9/29/2009 20:20
VER 12	0.500	15	552	712	2.069	0.267	0.5096	23.2132	1.9542	9/29/2009 21:10

AAQ
9/30/09

Sample ID	Det #	Run Date	Sample Type	Standard ID	NC	NC units	Recovery/RPD
			LCS	0638-F	24.05	pCi/L	102%
701	7	9/29/2009 15:35	LCS	0638-F	24.05	pCi/L	113%
702	7	9/29/2009 16:10	LCS	0638-F	24.05	pCi/L	103%
703	7	9/29/2009 16:45	LCS	0638-F	24.05	pCi/L	96%
704	7	9/29/2009 17:15	LCS	0638-F	24.05	pCi/L	89%
705	7	9/29/2009 17:50	LCS	0638-F	24.05	pCi/L	100%
706	7	9/29/2009 18:25	LCS	0638-F	24.05	pCi/L	86%
707	7	9/29/2009 18:40	LCS	0638-F	24.05	pCi/L	87%
708	7	9/29/2009 19:00	LCS	0638-F	24.05	pCi/L	89%
709	7	9/29/2009 19:40	LCS	0638-F	24.05	pCi/L	98%
710	7	9/29/2009 20:00	LCS	0638-F	24.05	pCi/L	86%
711	7	9/29/2009 20:20	LCS	0638-F	24.05	pCi/L	97%
712	7	9/29/2009 21:10	LCS	0638-F	24.05	pCi/L	
			DEGASS- DE-EM	dE-EM- COUNT	constant	constant	In growth constant
DEGASSING DATE/TIME	DE-EMAN. DATE/TIME						
9/22/2009 14:30	9/30/2009 6:00	183.50	3.33	0.7498	0.9751	1.0009	42.1333
9/22/2009 14:30	9/29/2009 10:00	163.50	6.17	0.7090	0.9545	1.0009	41.4000
9/22/2009 14:30	9/29/2009 10:15	163.75	6.50	0.7095	0.9521	1.0009	41.4000
9/22/2009 14:30	9/29/2009 10:30	164.00	6.75	0.7101	0.9503	1.0009	38.8667
9/22/2009 14:30	9/29/2009 10:50	164.33	7.00	0.7108	0.9485	1.0009	33.8000
9/22/2009 14:30	9/29/2009 11:15	164.75	7.17	0.7117	0.9473	1.0009	38.4000
9/22/2009 14:30	9/29/2009 12:45	166.25	5.92	0.7150	0.9563	1.0009	35.6663
9/22/2009 14:30	9/29/2009 13:10	166.67	5.83	0.7159	0.9569	1.0009	34.7333
9/22/2009 14:30	9/29/2009 13:35	167.08	6.08	0.7168	0.9551	1.0009	37.0000
9/22/2009 14:30	9/30/2009 6:30	184.00	3.33	0.7507	0.9751	1.0009	46.0000
9/22/2009 14:30	9/29/2009 14:20	167.83	6.00	0.7184	0.9557	1.0009	35.5333
9/22/2009 14:30	9/29/2009 14:40	168.17	6.50	0.7191	0.9521	1.0009	36.5333

A19
9/30/09

Ra-226 Verification Sheet

Count time: 15 min

L#
VWS

Ra-226 Verification Sheet

COUNT 15 min

Sample ID	Volume (mL)	End Degas Date/Time	End De-em Date/Time	Start Count Date/Time	Cell #	Det #	Background CPM	Total Counts
VRL16	600	9/22/09 1430	9/30/09 000	9/30/09 0911W	701	7	3	636
VRL17	600	9/22/09 1430	9/30/09 030	9/30/09 0950	710	7	8	694

JLQ
9/30/09

JLQ 9/30/09

General Engineering Laboratories
Verification Source Preparation Sheet

Applicable SOP Number	GL RAP-A 008	Isotope	Pu-239
Date Standards Prepared	1/23/09	Cocktail Type Used	NA
Standard ID	DC25814	Matrix of Vial/Planchett	NA NA NA
Amount Used (g or mL)	0.1	Type of Scintillation Vial	NA
Standard Activity (DPM/g or mL)	268,8845	Pipette ID Used	1429303
Reference Date	1/23/04	Balance ID Used	380580204
Expiration Date	1/17/10	Quenching Agent	NA
Residue/Carrier Agent	NR		

	Standard Number	Quenching Vol (uL)/ Residue Volume(mL)	Initial Wt. (g)	Final Wt. (g)	Net Wt. (mg)
1	Ver 1				
2	Ver 2				
3	Ver 3				
4	Ver 4				
5	Ver 5				
6	Ver 6				
7	Ver 7				
8	Ver 8				
9	Ver 9				
10	Ver 10				
11	Ver 11				
12	Ver 12				
13	Ver 13				
14	Ver 17				

Prepared By:

Kelli & Daniel

Date:

9/30/09

Reviewed By:

Ashe & Ogt

Date:

9/30/09

Rev 1 RLM 9/10/97

ANALYTICS

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 - U.S.A.

0638

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION Standard Radionuclide Source

67519-278

Ra-226 5 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master solution. The master solution was calibrated using a germanium gamma spectrometer system.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

Analytics maintains traceability to the National Institute of Standards and Technology through participation in a Measurements Assurance Program as described in USNRC Reg. Guide 4.15, Revision 1, February 1979.

ISOTOPE:	Ra-226
ACTIVITY (dps):	2.353 E4
HALF-LIFE:	1.600 E3 years
CALIBRATION DATE:	January 23, 2004 12:00 EST
RELATIVE EXPANDED UNCERTAINTY (k=2):	3.3%

Impurities: γ -impurities (other than decay products) <0.1%

5.01065 grams 0.1M HCl solution with 50 μ g/g Ba carrier.

P O NUMBER 3231RD, Item 5

SOURCE PREPARED BY:

M. D. Currie
M. D. Currie, Radiochemist

Q A APPROVED:

M. D. Currie 1/26/04

W/120105

Standard Traceability Log Rad

WARNING! Training must be completed!!
Alphalims will be locked out if training is not completed within 1 week of assignment Contact
Quality if additional time is needed to complete training

Source Material Info		A Solution Material Info	
Parent Code:	0638	Isotope:	Radium-226
Prepared By:	Amanda Fehr	Prepared By:	Amanda Fehr
Carrier Conc:	0.1M HCl	Prep Date:	01/16/2006
Reference Date:	01/23/2004	Verification Date:	04/09/2009
Ampoule Mass (g):	5.01065 g	Expiration Date:	04/09/2010
Uncertainty:	+/- 3.3 %	Primary Code:	0638-A
LogBook No:	RC-S-037-037	Dilution(mL):	100 mL
		Mass of Parent(g):	4.8398 g
		Density(g/mL):	1.0266
		Balance ID:	38080204

Calculations Converting parent activity to dpm/mL/dpm/g

(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / (Ampoule Mass(g) * (Dilution Vol)) = Parent Activity (dpm/mL)
(Mass of parent(g)) * (Parm Activity (dps)) * (conversion dpm to dps) / Density / (Ampoule Mass (g) * (Dilution Vol)) = Parent Activity (dpm/g)
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (5.01065 g * 100 mL) = 13636.6133 dpm/mL
(4.8398 g) * (23530 dps) * (60 dpm/dps) / (1.0266 g/mL) / (5.01065 g * 100 mL) = 13282.9676 dpm/g

WIP/2009

Secondary Standards					
Prep Date	Preparer	Mass Primary	Dilution (mL)	Code	Conc dpm/mL
01/17/2006	Amanda Fehr	2.1041	100	0638-B	279.0211 dpm/mL
07/17/2006	Mary Aders	2.1313	100	0638-C	282.6281 dpm/mL
03/28/2007	Daniel Roy	2.1025	100	0638-D	279.2744 dpm/mL
03/28/2007	Daniel Roy	45.468	250	0638-E	2415.7999 dpm/mL
12/18/2007	Daniel Roy	2.014	100	0638-F	267.519 dpm/ml
02/12/2008	Daniel Roy	.5004	100	0638-G	66.468 dpm/ml
07/23/2008	Daniel Roy	5.0607	250	0638-H	268.8845 dpm/ml

GEL Laboratories LLC
Version 1.0 9/18/2000

W M 1260105

Verification for Ra-226 Standard 0638-H

M. Aders 7/17/2009	Isotope	Value	Uncertainty
	0638-H	12.025	1.2237
	0638-H	10.739	1.1752
	0638-H	12.348	1.2298

Mean Value (Counting) = 11.704 96.86 Pass
Stdev = 0.85081728 Rule 3 (Pass/Fail)

Target = 12.08
Lower Limit = 10.00223211
Upper Limit = 13.40550123
Rule 1 Pass/Fail = Pass
Two sigma = 1.701634559
10 % of Mean = 1.170386667
Rule 2 (Pass/Fail) = Fail *Exception taken due to full reaccovery of standard

Rule 1 = The certificate value (NOT including any uncertainty) shall lie within the 95% confidence interval determined from the mean and two sigma standard deviation of the three measurements

Rule 2 = The two sigma value used for the 95% confidence interval shall not exceed 10% of the mean value of the three verification measurements.

Rule 3 = The determined mean value shall be within 5% of the certificate value.

The analyst prepared three standard verification sources for Ra-226 source 0638-H by transferring portions of the degassed standard into tared glass liquid scintillation vials. 10 mL of DI Water and 10 mL of mineral oil were added to each vial and the vials were shaken. A Blank vial was prepared in a similar fashion using 10 mL of DI Water and 10 mL of mineral oil. The standard verification vials and Background source were dark adapted for two hours and counted on LSC Green using source standard verification. Each verification source calculation was performed as follows:

Source dpm/g = $(A - B)/(C)(D)$
where:
A = Ver. source cpm,
B = BKG cpm,
C = System efficiency, (cpm/dpm), and
D = mass used for standard verification.

Reference RAD SOP M-001

Aug 2009 7/30/09
L. J. Schaefer 7/17/09
Nanamarett 7/30/09

Radon-222 Liquid

Filename : RN222.XLS
 File type : Excel
 Version # : 1.2.4

Batch : 886194
 Analyst : MLA
 Prep Date : 7/17/2009

Rn-222 Abundance : 1

Rn-222 Method Uncertainty : 0.0556
 Geometry : 10ML MINERAL OIL/10ML
 SAMPLE

Spike S/N :	N/A	LCS S/N :	0638-H
Spike Exp Date :	N/A	LCS Exp Date :	7/23/2009
Spike Activity (dpm/ml):	N/A	LCS Activity (dpm/ml):	268.25
Spike Volume Added:	N/A	LCS Volume Added:	0.10
Spike Date/Time:	7/17/2009 15:00	Procedure Code :	LSC99TCL
Paramname :	Radon-222	Required MDA :	50
Halflife of Radon-222 :	3.823	pCi/L days	

Pipet, 0.1 ml Sidev : +/- 0.0000701 ml
 Pipet, 0.5 ml Sidev : +/- 0.002564 ml

Pos.	Sample ID	Sample Aliquot L	Sample StDev. L	Sample Date/Time	Count raw Data			Background Count			Background Count		
					Rack Position #	Counting Time (min.)	Quench#	Gross cpm	cpm	Time (min.)	Start Date/Time	Count	Start Date/Time
1	1201883284.1	1.0000	2.0399E-05	7/17/2009 15:00	22-2	15	50.3	43.73	8.20	15	7/20/2009 11:53	0.594	
2	1201883285.1	1.0000	2.0399E-05	7/17/2009 15:00	22-3	15	50	38.2	8.20	15	7/20/2009 12:09	0.592	
3	1201883286.1	1.0000	2.0399E-05	7/17/2009 15:00	22-4	15	49.1	45.4	8.20	15	7/20/2009 12:26	0.591	

Pos.	Counted on	Calibration Data		Detector		Backgrounds		Correction Factors		Net Sample Activity for MS pCi/L
		Calibration Date	Calibration Due Date	Detector Efficiency (cpm/dpm)	Error (cpm/dpm)	Rack Position #	Count Start Date/Time	Spike Date/Time	Rn-222 Ingrowth	
1	LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	22-1	7/20/2009 11:36	7/17/2009 15:00	0.406	0.406
2	LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	22-1	7/20/2009 11:36	7/17/2009 15:00	0.408	0.408
3	LSCGREEN	3/25/2009	3/31/2010	3.4365	0.00792	22-1	7/20/2009 11:36	7/17/2009 15:00	0.409	0.409

Notes:
 1 - Results are decay corrected to Sample Date/Time
 2 - Reference date for Spike Activity (dpm/ml) is the batch Prep Date
 3 - Spike Nominals are decay corrected to Sample Date/Time

Results Pos.	Decision Level pCi/L	Critical Level pCi/L	Required MDA pCi/L	MDA pCi/L	Sample Act. Conc. pCi/L	Sample Act. Error pCi/L	Net Count Rate Error CPM	Net Count Rate Error CPM	2 SIGMA		2 SIGMA		
									Total Prop. pCi/L	Counting Uncertainty pCi/L	Total Prop. pCi/L	Counting Uncertainty pCi/L	
1	0.8104	0.5722	50	1.2114	12.0246	0.0525	35.8600	1.8619	1.2237	1.8026	LCS	12.0832	99.5%
2	0.8078	0.5703	50	1.2075	10.7393	0.0564	32.1300	1.7939	1.1752	1.6669	LCS	12.0832	88.9%
3	0.8053	0.5685	50	1.2037	12.3477	0.0514	37.0600	1.8833	1.2298	1.8330	LCS	12.0832	102.2%

Radon 222 Que Sheet

07/17/2009

Batch #:886194
Spike Isotope: Radium-226 Spike Code: 11010 Expiration Date: 7/23/09 Vol: 1.1 Nom Conc: _____
LCS Isotope: Radium-226 LCS Code: 11010 Expiration Date: 7/23/09 Vol: 1.1 Nom Conc: _____
Prep Date: 7/17/09 Initials: LSC Witness: AT

Analyst:MLA

Internal Due Date:07/22/2009

Expiration Date: 7/23/09 Vol: 1.1 Nom Conc: _____
Expiry Date: 7/23/09 Vol: 1.1 Nom Conc: _____

Comments

Sample ID	Client Description	Type	Hazard Code	Min CRDL	Matrix	Client	Collection Date	Label	Wet/Dry Sample Mass (g/mL)	LSC Rack #	Time Spike Added
1201883284-1	LCS for batch 886194	LCS		50 pCi/L	WATER	QC ACCOUNT	15-JUL-09 10:45 AM	1		22-2	
1201883285-1	LCS for batch 886194	LCS		50 pCi/L	WATER	QC ACCOUNT	15-JUL-09 10:45 AM	2		22-3	
1201883286-1	LCS for batch 886194	LCS		50 pCi/L	WATER	QC ACCOUNT	15-JUL-09 10:45 AM	3		22-4	

Bkg Rack #: 22-1

Comments: _____

Data Reviewed By: _____

Instrument Used: LS6000 (Red) 7065155, LS6500 (Black) 7069123, LS6500 (Blue) 7067083, LS6500 (Green) 7067404
Wallac (Yellow) 4040127, Wallac (Pink) 22000082, Purple 7069123, Silver 70600656

PAGE: 1

ID # FCBM-22222

20 JUL 2009 11:16

USER: CA COMMENT: GREEN

PRESENT TIME : 15.00
DATA CALC : CRM H# : YES SAMPLE REPEATS : 1 PRINTER : EDIT
COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : EDIT
TWO PHASE : NO AQC : NO CYCLE REPEATS : 1 DISK : OFF
SCINTILLATOR: LIQUID LUMEX: YES LOW SAMPLE REJ: 0 RWM LIST : OFF
LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 630.0 - 975.0 XERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0
CHAN: 300.0 - 900.0 XERROR: 2.00 FACTOR: 1.000000 BKG. SUB: 0


ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND 1 CPM	WIND 1 %ERROR	WIND 2 CPM	WIND 2 %ERROR	LUMEX %	ELAPSED TIME
1	221-1	15.00	47.9	8.20	18.03	19.13	11.81	0.38	15.92
2	221-2	15.00	50.3	43.73	7.81	60.67	6.63	0.16	32.28
3	221-3	15.00	50.0	38.20	8.36	52.27	7.14	0.17	48.66
4	221-4	15.00	49.1	45.40	7.66	62.93	6.51	0.15	65.03

Sample Count Start Time: 20 Jul 2009 11:36:58
Data Capture Date 20 Jul 2009 11:52:21
User Filename S16072022-1B.XLS
U16072022-1B.XLS

Spectrum Type Log Counts
User Number 16
User Id RN-222
User Comment GREEN
Isotope Name ¹⁴C
Scintillator LIQUID
Sample, Rack-Pos, Time: 1 22-1 15.00
H#, Total Counts: 47.9 412
Start, End, X-Axis: 0 990 Channel Number


SPECTRUM PLOT
USER 16 - RN-222



Sample Count Start Time: 20 Jul 2009 11:53:20
Data Capture Date 20 Jul 2009 12:08:43
User Filename S16072022-2B.XLS
U16072022-1B.XLS

Spectrum Type Log Counts
User Number 16
User Id RN-222
User Comment GREEN
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 2 22-2 15.00
H#, Total Counts: 50.3 1100
Start, End, X-Axis: 0 990 Channel Number


SPECTRUM PLOT
USER 16 - RN-222



Sample Count Start Time: 20 Jul 2009 12:09:43
Data Capture Date 20 Jul 2009 12:25:05
User Filename S16072022-3B.XLS
U16072022-1B.XLS

Spectrum Type Log Counts
User Number 16
User Id RN-222
User Comment GREEN
Isotope Name 14C
Scintillator LIQUID
Sample, Rack-Pos, Time: 3 22-3 15.00
H#, Total Counts: 50.0 956
Start, End, X-Axis: 0 990 Channel Number


S P E C T R U M P L O T
USER 16 - RN-222



Sample Count Start Time: 20 Jul 2009 12:26:05
Data Capture Date 20 Jul 2009 12:41:28
User Filename S16072022-4B.XLS
U16072022-1B.XLS

Spectrum Type Log Counts
User Number 16
User Id RN-222
User Comment GREEN
Isotope Name ¹⁴C
Scintillator LIQUID
Sample, Rack-Pos, Time: 4 22-4 15.00
H#, Total Counts: 49.1 1123
Start, End, X-Axis: 0 990 Channel Number

S P E C T R U M P L O T
USER 16 - RN-222




BACKGROUND AND EFFICIENCY DATA

LUCAS1 EFF

Generated 12/02/2009


+3 Sigma 143282



● Denotes Outlier

LUCAS2 EFF

Generated 12/01/2009




● Denotes Outlier

LUCAS3 EFF

Generated 12/01/2009

+3 Sigma 128619




● Denotes Outlier

LUCAS4 EFF


Generated 12/01/2009

+3 Sigma 145550

CPM



● Denotes Outlier




● Denotes Outlier

LUCAS7 EFF

Generated 12/01/2009

+3 Sigma 160934

CPM




● Denotes Outlier

PIC10C ALPHA BKG


Generated 11/23/2009

CPM




PIC10C ALPHA EFF

CPM




PIC10C ALPHA EFF Cross Talk

CPM



PIC10C BETA BKG

CPM




● Denotes Outlier

PIC10C BETA EFF

Generated 11/23/2009

+3 Sigma 26060


CPM



PIC10D ALPHA BKG


Generated 11/23/2009

CPM




PIC10D ALPHA EFF

CPM




PIC10D ALPHA EFF Cross Talk

CPM



PIC10D BETA BKG

CPM




● Denotes Outlier

PIC10D BETA EFF

Generated 11/23/2009

+3 Sigma 44250

CPM



PIC10D BETA EFF Cross Talk

+3 Sigma 0.000924


+2 Sigma 0.000778

Mean 0.000486

-2 Sigma 0.000193

-3 Sigma 0.0000473

DATE




● Denotes Outlier

PIC12C ALPHA BKG


Generated 11/23/2009

CPM




PIC12C ALPHA EFF

CPM




PIC12C ALPHA EFF Cross Talk

CPM



PIC12C BETA BKG

CPM




● Denotes Outlier

PIC12C BETA EFF

Generated 11/23/2009

+3 Sigma 26350

CPM



PIC12C BETA EFF Cross Talk

+3 Sigma 0.0002203


+2 Sigma 0.000178

Mean 0.0000945

-2 Sigma 0.00001064

-3 Sigma -0.0000313

DATE




● Denotes Outlier

PIC12D ALPHA BKG


Generated 11/23/2009

CPM




PIC12D ALPHA EFF

CPM



PIC12D ALPHA EFF Cross Talk


CPM



PIC12D BETA BKG

+3 Sigma 1.101

CPM




● Denotes Outlier

PIC12D BETA EFF

Generated 11/23/2009

+3 Sigma 45160

CPM



PIC12D BETA EFF Cross Talk

+3 Sigma 0.000249


+2 Sigma 0.0002099

Mean 0.000132

-2 Sigma 0.00005408

-3 Sigma 0.0000151

DATE




● Denotes Outlier

PIC13A ALPHA BKG


Generated 11/23/2009

CPM




PIC13A ALPHA EFF

CPM




PIC13A ALPHA EFF Cross Talk

CPM



PIC13A BETA BKG

CPM




● Denotes Outlier

PIC13A BETA EFF

Generated 11/23/2009

+3 Sigma 20740

CPM



PIC13A BETA EFF Cross Talk

+3 Sigma 0.00037

+2 Sigma 0.000314

Mean 0.0002032

-2 Sigma 0.000092

-3 Sigma 0.0000364


DATE

DATE


● Denotes Outlier

PIC14A ALPHA BKG


Generated 11/23/2009




PIC14A ALPHA EFF



PIC14A ALPHA EFF Cross Talk



PIC14A BETA BKG




● Denotes Outlier


PIC14A BETA EFF

Generated 11/23/2009

CPM



PIC14A BETA EFF Cross Talk




● Denotes Outlier

PIC1A ALPHA BKG

Generated 11/23/2009
+3 Sigma 0.242


CPM



PIC1A ALPHA EFF

+3 Sigma 11270


CPM



PIC1A ALPHA EFF Cross Talk

+3 Sigma 0.288


CPM



PIC1A BETA BKG

+3 Sigma 0.9406

CPM




● Denotes Outlier

PIC1A BETA EFF

Generated 11/23/2009

+3 Sigma 20440

CPM



PIC1A BETA EFF Cross Talk

+3 Sigma 0.00132


+2 Sigma 0.001181

Mean 0.0009009

-2 Sigma 0.000621

-3 Sigma 0.000481

DATE




● Denotes Outlier

PIC1C ALPHA BKG


Generated 11/23/2009

CPM




PIC1C ALPHA EFF

CPM




PIC1C ALPHA EFF Cross Talk

CPM



PIC1C BETA BKG

CPM




● Denotes Outlier


PIC1C BETA EFF

Generated 11/23/2009

CPM



PIC1C BETA EFF Cross Talk




● Denotes Outlier

PIC1D ALPHA BKG

Generated 11/23/2009


CPM



PIC1D ALPHA EFF

DATE


CPM



PIC1D ALPHA EFF Cross Talk

DATE


CPM



PIC1D BETA BKG

DATE

CPM




● Denotes Outlier


PIC1D BETA EFF

Generated 11/23/2009

CPM




PIC1D BETA EFF Cross Talk




● Denotes Outlier

PIC2A ALPHA BKG


Generated 11/23/2009




PIC2A ALPHA EFF



PIC2A ALPHA EFF Cross Talk



PIC2A BETA BKG




● Denotes Outlier


PIC2A BETA EFF

Generated 11/23/2009

CPM




PIC2A BETA EFF Cross Talk




● Denotes Outlier

PIC3A ALPHA BKG


Generated 11/23/2009




PIC3A ALPHA EFF



PIC3A ALPHA EFF Cross Talk



PIC3A BETA BKG




● Denotes Outlier

PIC3A BETA EFF

Generated 11/23/2009

+3 Sigma 20110

CPM



PIC3A BETA EFF Cross Talk


+3 Sigma 0.0364

+2 Sigma 0.0324

Mean 0.0243

-3 Sigma 0.0122

-2 Sigma 0.0163




● Denotes Outlier

PIC5C ALPHA BKG

Generated 11/23/2009

CPM



PIC5C ALPHA EFF

+3 Sigma 12250


+2 Sigma 12197

Mean 12090

-2 Sigma 11983

-3 Sigma 11930

CPM



PIC5C ALPHA EFF Cross Talk

+3 Sigma 0.2706


+2 Sigma 0.268

Mean 0.263

-2 Sigma 0.258

-3 Sigma 0.256

CPM



PIC5C BETA BKG

+3 Sigma 1.75


+2 Sigma 1.52

Mean 1.046

-2 Sigma 0.576

-3 Sigma 0.3403

CPM




● Denotes Outlier

PIC5C BETA EFF

Generated 11/23/2009


+3 Sigma 22270

CPM



DATE

PIC5C BETA EFF Cross Talk




DATE

● Denotes Outlier

PIC6B ALPHA BKG


Generated 11/23/2009

CPM




PIC6B ALPHA EFF

CPM




PIC6B ALPHA EFF Cross Talk

CPM



PIC6B BETA BKG

CPM




● Denotes Outlier

PIC6B BETA EFF

Generated 11/23/2009

+3 Sigma 23370

CPM



PIC6B BETA EFF Cross Talk

+3 Sigma 0.0138

+2 Sigma 0.01208

Mean 0.00861

-2 Sigma 0.00515

-3 Sigma 0.00342

DATE


● Denotes Outlier

PIC6D ALPHA BKG

Generated 11/23/2009


+3 Sigma 0.538

CPM




PIC6D ALPHA EFF

CPM




PIC6D ALPHA EFF Cross Talk

CPM



PIC6D BETA BKG

● Denotes Outlier




PIC6D BETA EFF

Generated 11/23/2009

+3 Sigma 43710

CPM



PIC6D BETA EFF Cross Talk


+3 Sigma 0.01031

+2 Sigma 0.009107

-2 Sigma 0.00428

-3 Sigma 0.003075

Mean 0.0067




● Denotes Outlier

PIC7A ALPHA BKG


Generated 11/23/2009

CPM




PIC7A ALPHA EFF

CPM




PIC7A ALPHA EFF Cross Talk

CPM



PIC7A BETA BKG

● Denotes Outlier




PIC7A BETA EFF

Generated 11/23/2009

+3 Sigma 20810

CPM



PIC7A BETA EFF Cross Talk


+3 Sigma 0.0117

+2 Sigma 0.01049

Mean 0.00799

-2 Sigma 0.00548


-3 Sigma 0.00423




● Denotes Outlier

PIC7C ALPHA BKG


Generated 11/23/2009




PIC7C ALPHA EFF



PIC7C ALPHA EFF Cross Talk



PIC7C BETA BKG




● Denotes Outlier


PIC7C BETA EFF

Generated 11/23/2009

CPM




PIC7C BETA EFF Cross Talk




● Denotes Outlier

PIC7D ALPHA BKG


Generated 11/23/2009




PIC7D ALPHA EFF



PIC7D ALPHA EFF Cross Talk



PIC7D BETA BKG




● Denotes Outlier

PIC7D BETA EFF

Generated 11/23/2009

+3 Sigma 25290

CPM



PIC7D BETA EFF Cross Talk

+3 Sigma 0.00925


+2 Sigma 0.00815

Mean 0.00595

-2 Sigma 0.00375

-3 Sigma 0.00265

DATE




● Denotes Outlier

PIC8A ALPHA BKG

Generated 11/23/2009

+3 Sigma 0.4007


CPM



PIC8A ALPHA EFF

+3 Sigma 11150


CPM



PIC8A ALPHA EFF Cross Talk

+3 Sigma 0.268


CPM



PIC8A BETA BKG

+3 Sigma 1.82

CPM




● Denotes Outlier


PIC8A BETA EFF

Generated 11/23/2009

CPM



PIC8A BETA EFF Cross Talk




● Denotes Outlier

PIC8C ALPHA BKG


Generated 11/23/2009

CPM




PIC8C ALPHA EFF

CPM




PIC8C ALPHA EFF Cross Talk

CPM



PIC8C BETA BKG

CPM




● Denotes Outlier


PIC8C BETA EFF

Generated 11/23/2009

CPM




PIC8C BETA EFF Cross Talk




● Denotes Outlier

PIC9A ALPHA BKG


Generated 11/23/2009




PIC9A ALPHA EFF



PIC9A ALPHA EFF Cross Talk



PIC9A BETA BKG




● Denotes Outlier

PIC9A BETA EFF

Generated 11/23/2009

CPM



PIC9A BETA EFF Cross Talk

+3 Sigma 0.000344


+2 Sigma 0.000293

Mean 0.00019

-2 Sigma 0.0000865

-3 Sigma 0.0000349

DATE




● Denotes Outlier

PIC9B ALPHA BKG


Generated 11/23/2009

CPM




PIC9B ALPHA EFF

CPM




PIC9B ALPHA EFF Cross Talk

CPM



PIC9B BETA BKG

CPM




● Denotes Outlier

PIC9B BETA EFF

Generated 11/23/2009

+3 Sigma 22590

CPM



PIC9B BETA EFF Cross Talk

+3 Sigma 0.000372


+2 Sigma 0.000322

Mean 0.000223

-2 Sigma 0.000124

-3 Sigma 0.0000743


DATE




● Denotes Outlier

PIC9D ALPHA BKG


Generated 11/23/2009




PIC9D ALPHA EFF



PIC9D ALPHA EFF Cross Talk



PIC9D BETA BKG




● Denotes Outlier


PIC9D BETA EFF

Generated 11/23/2009

CPM




PIC9D BETA EFF Cross Talk




● Denotes Outlier


QA filename : DKA100:[ENV_ALPHA.QA.W]W025.QAF; 4
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 250000 through 0, 4500000




QA filename : DKA100:[ENV_ALPHA.QA.W]W025.QAF; 4
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 60.0000 through 105.000




QA filename : DKA100:[ENV_ALPHA.QA,B]B025.QAF;2
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W026.QAF; 3
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 307728 through 0, 327728




QA filename : DKA100:[ENV_ALPHA.QA.W]W026.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 85.8763 through 94.9159




QA filename : DKA100:[ENV_ALPHA.QA,B]B026.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W027.QAF; 4
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 324980 through 0, 344980




QA filename : DKA100:[ENV_ALPHA.QA.W]W027.QAF; 4
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 81.2030 through 89.7506




QA filename : DKA100:[ENV_ALPHA.QA,B]B027.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W028.QAF; 4
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 295040 through 0, 315040




QA filename : DKA100:[ENV_ALPHA.QA.W]W028.QAF; 4
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 85.1965 through 94.1645




QA filename : DKA100:[ENV_ALPHA.QA,B]B028.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W029.QAF; 6
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 250000 through 0, 450000




QA filename : DKA100:[ENV_ALPHA.QA.W]W029.QAF;6
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 60.0000 through 105.000




QA filename : DKA100:[ENV_ALPHA.QA,B]B029.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W030.QAF; 3
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 250000 through 0, 450000




QA filename : DKA100:[ENV_ALPHA.QA.W]W030.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:08 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 60.0000 through 105.000




QA filename : DKA100:[ENV_ALPHA.QA,B]B030.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:51 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W031.QAF; 4
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 338790 through 0, 358790




QA filename : DKA100:[ENV_ALPHA.QA.W]W031.QAF; 4
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 81.9728 through 90.6016




QA filename : DKA100:[ENV_ALPHA.QA,B]B031.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:52 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W033.QAF; 3
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 304222 through 0, 327748




QA filename : DKA100:[ENV_ALPHA.QA.W]W033.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 83.9373 through 86.9661




QA filename : DKA100:[ENV_ALPHA.QA,B]B033.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:52 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W035.QAF; 3
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 294859 through 0, 314859



QA filename : DKA100:[ENV_ALPHA.QA.W]W035.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 85.3984 through 94.3878




QA filename : DKA100:[ENV_ALPHA.QA,B]B035.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:52 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




Background Rate


QA filename : DKA100:[ENV_ALPHA.QA.W]W036.QAF;2
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 318717 through 0, 338717




QA filename : DKA100:[ENV_ALPHA.QA.W]W036.QAF;2
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:09 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 84.6422 through 93.5518




QA filename : DKA100:[ENV_ALPHA.QA,B]B036.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:52 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W037.QAF; 4
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 250000 through 0, 450000




QA filename : DKA100:[ENV_ALPHA.QA.W]W037.QAF; 4
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 60.0000 through 105.000




QA filename : DKA100:[ENV_ALPHA.QA,B]B037.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W038.QAF; 3
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 327380 through 0, 362086




QA filename : DKA100:[ENV_ALPHA.QA.W]W038.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 87.5715 through 92.1899




QA filename : DKA100:[ENV_ALPHA.QA.B]B038.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W039.QAF; 3
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 250000 through 0, 450000




QA filename : DKA100:[ENV_ALPHA.QA.W]W039.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 60.0000 through 105.000




QA filename : DKA100:[ENV_ALPHA.QA,B]B039.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W040.QAF; 3
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 313016 through 0, 333016




QA filename : DKA100:[ENV_ALPHA.QA.W]W040.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 82.8065 through 91.5229




QA filename : DKA100:[ENV_ALPHA.QA,B]B040.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W041.QAF; 5
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 320943 through 0, 340943




QA filename : DKA100:[ENV_ALPHA.QA.W]W041.QA.F;5
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 86.6435 through 95.7639




QA filename : DKA100:[ENV_ALPHA.QA,B]B041.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02




QA filename : DKA100:[ENV_ALPHA.QA.W]W042.QAF; 3
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 322243 through 0, 342243




QA filename : DKA100:[ENV_ALPHA.QA.W]W042.QAF;3
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 4-MAY-2009 09:38:10 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 80 . 6389 through 89 . 1273




QA filename : DKA100:[ENV_ALPHA.QA,B]B042.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 12:03:53 through 5-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 2.000000E-02



QA filename : DKA100:[ENV_ALPHA.QA.W]W121.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:45:23 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 237686 through 0, 257686




QA filename : DKA100:[ENV_ALPHA.QA.W]W121.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:45:23 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 89.4263 through 98.8395




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B121.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:50:31 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W122.QAF;1
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:45:28 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.242659 through 0.262659




QA filename : DKA100:[ENV_ALPHA.QA.W]W122.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:45:28 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 90.5949 through 100.131




```

QA filename          : DKA100:[ENV_ALPHA,QA,B]B122,QAF;1
Parameter Name      : BACKRATE (Background Rate)
Start/End Dates    : 3-MAY-2009 13:50:35 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000


```



QA filename : DKA100:[ENV_ALPHA,QA,w]w123,QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:45:32 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 249752 through 0, 269752




QA filename : DKA100:[ENV_ALPHA.QA.W]W123.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:45:32 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 86.0496 through 95.1074




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA,QA,B]B123,QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:50:39 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : OKA100: [ENV_ALPHA.QA.W]W124.QAF;1
 Parameter Name : AVRGEFF (Average Efficiency)
 Start/End Dates : 18-MAY-2009 09:45:37 through 19-NOV-2009 12:00:00
 Lower/Upper Lmts: 0.251398 through 0.271398




QA filename : DKA100:[ENV_ALPHA.QA.W]W124.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:45:37 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 86.8862 through 96.0322




QA filename : DKA100:[ENV_ALPHA.QA,B]B124.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:50:44 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W138.QAF;1
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:46:47 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0,247085 through 0,267085




QA filename : DKA100:[ENV_ALPHA.QA.W]W138.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:46:47 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 82.8399 through 91.5599




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B138.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:51:44 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W139.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:46:51 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 240299 through 0, 260299




QA filename : DKA100:[ENV_ALPHA.QA.W]W139.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:46:51 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 85.8145 through 94.8477




QA filename : DKA100:[ENV_ALPHA.QA,B]B139.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:51:49 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W140.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:46:56 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.246178 through 0.266178




QA filename : DKA100:[ENV_ALPHA.QA.W]W140.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:46:56 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 83 . 8171 through 92 . 6399




QA filename : DKA100:[ENV_ALPHA.QA,B]B140.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:51:53 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W141.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:00 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.247845 through 0.267845




QA filename : DKA100:[ENV_ALPHA.QA.W]W141.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:00 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 80, 85.95 through 89, 37.11




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B141.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:51:57 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W142.QAF; 2
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:04 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 252182 through 0, 272182




QA filename : DKA100:[ENV_ALPHA.QA.W]W142.QAF;2
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:04 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 81.0245 through 89.5533




```

QA filename          : DKA100:[ENV_ALPHA,QA,B]B142.QAF;1
Parameter Name      : BACKRATE (Background Rate)
Start/End Dates    : 3-MAY-2009 13:52:01 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000

```




QA filename : DKA100:[ENV_ALPHA.QA.W]W143.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:09 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 233879 through 0, 253879



Average Efficiency


QA filename : DKA100:[ENV_ALPHA.QA.W]W143.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:09 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.9200 through 93.8590




```

QA filename          : DKA100:[ENV_ALPHA,QA,B]B143.QAF;1
Parameter Name      : BACKRATE (Background Rate)
Start/End Dates    : 3-MAY-2009 13:52:05 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000


```



QA filename : DKA100:[ENV_ALPHA.QA.W]W144.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:13 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 237963 through 0, 257963




QA filename : DKA100:[ENV_ALPHA.QA.W]W144.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:13 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.6507 through 93.5613




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B144.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:10 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA,QA,w]w145,QAf;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:17 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 239850 through 0, 259850




QA filename : DKA100:[ENV_ALPHA.QA.W]W145.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:17 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.9354 through 93.8760




QA filename : DKA100:[ENV_ALPHA,QA,B]B145.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:14 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W146.QAF;2
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:22 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.241831 through 0.261831




QA filename : DKA100:[ENV_ALPHA.QA.W]W146.QAF;2
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:22 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.8578 through 93.7902




QA filename : DKA100:[ENV_ALPHA.QA,B]B146.QAF;2
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:19 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : OKA100:[ENV_ALPHA.QA.W]W147.QAF;1
 Parameter Name : AVRGEFF (Average Efficiency)
 Start/End Dates : 18-MAY-2009 09:47:26 through 19-NOV-2009 12:00:00
 Lower/Upper Lmts: 0,237046 through 0,257046




QA filename : DKA100:[ENV_ALPHA.QA.W]W147.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:26 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.9777 through 93.9227




QA filename : DKA100:[ENV_ALPHA.QA,B]B147.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:23 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA,W]W148.QAF;1
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:30 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0,237934 through 0,257934




QA filename : DKA100:[ENV_ALPHA.QA.W]W148.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:30 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 85.0831 through 94.0393




QA filename : DKA100:[ENV_ALPHA,QA,B]B148.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:27 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W149.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:34 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.242495 through 0.262495




QA filename : DKA100:[ENV_ALPHA.QA.W]W149.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:34 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.8126 through 93.7402




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B149.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:31 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA,QA,w]w150,QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:39 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 238314 through 0, 258314




QA filename : DKA100:[ENV_ALPHA.QA.W]W150.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:39 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 84.4039 through 93.2885




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B150.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:35 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA,QA,w]w151.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:43 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 233693 through 0, 253693




QA filename : DKA100:[ENV_ALPHA.QA.W]W151.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:43 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 86.5749 through 95.6881




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B151.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:39 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : OKA100:[ENV_ALPHA.QA.W]W152.QAF;1
 Parameter Name : AVRGEFF (Average Efficiency)
 Start/End Dates : 18-MAY-2009 09:47:48 through 19-NOV-2009 12:00:00
 Lower/Upper Lmts: 0.238479 through 0.258479




QA filename : DKA100:[ENV_ALPHA.QA.W]W152.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:48 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 83,7180 through 92,5304




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA,QA,B]B152.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:43 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA,QA,w]w153.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:52 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 244738 through 0, 264738




QA filename : DKA100:[ENV_ALPHA.QA.W]W153.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:52 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 81.3634 through 89.9280




QA filename : DKA100:[ENV_ALPHA,QA,B]B153.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:47 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W154.QAF;1
Parameter Name : AVERAGEEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:47:57 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.251386 through 0.271386




QA filename : DKA100:[ENV_ALPHA.QA.W]W154.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:47:57 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 79.9003 through 88.3109




QA filename : DKA100:[ENV_ALPHA.QA,B]B154.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:51 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W155.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 18-MAY-2009 09:48:01 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 247241 through 0, 267241




QA filename : DKA100:[ENV_ALPHA.QA.W]W155.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 18-MAY-2009 09:48:01 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 83.6873 through 92.4965




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B155.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:52:54 through 19-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W201.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:32 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 249568 through 0, 269568




QA filename : DKA100:[ENV_ALPHA.QA.W]W201.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 27-MAY-2009 07:45:32 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 81.1299 through 89.6699




QA filename : DKA100:[ENV_ALPHA.QA,B]B201.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:55:50 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W202.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:36 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 255511 through 0, 275511




QA filename : DKA100:[ENV_ALPHA.QA.W]W202.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 27-MAY-2009 07:45:36 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 80, 8649 through 89, 3769




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B202.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:55:54 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W203.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:40 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 252203 through 0, 272203




QA filename : DKA100:[ENV_ALPHA.QA.W]W203.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 27-MAY-2009 07:45:40 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 83,7993 through 92,6203



QA filename : DKA100:[ENV_ALPHA.QA,B]B203.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:55:58 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W204.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:44 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 242368 through 0, 262368




Average Efficiency


QA filename : DKA100:[ENV_ALPHA.QA.W]W204.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 27-MAY-2009 07:45:44 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 82.7661 through 91.4783




QA filename : DKA100:[ENV_ALPHA.QA,B]B204.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:56:02 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



QA filename : DKA100:[ENV_ALPHA.QA.W]W205.QAF;1
Parameter Name : AVRGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:48 through 24-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 245702 through 0, 265702




QA filename : DKA100:[ENV_ALPHA.QA.W]W205.QAF;1
Parameter Name : NLACTVITY-GD148 (NUCLIDE ACTIVITY GD-148)
Start/End Dates : 27-MAY-2009 07:45:48 through 24-NOV-2009 12:00:00
Lower/Upper Lmts: 86.7285 through 95.8579




NUCLIDE ACTIVITY GD-


QA filename : DKA100:[ENV_ALPHA.QA,B]B205.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:56:06 through 24-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000




QA filename : DKA100:[ENV_ALPHA.QA.W]W206.QAF;1
Parameter Name : AVERAGEFF (Average Efficiency)
Start/End Dates : 27-MAY-2009 07:45:52 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0, 246228 through 0, 266228



QA filename : OKA100:[ENV_ALPHA,QA,W]W206.QAF;1
 Parameter Name : NLACTVITY-60148 (NUCLIDE ACTIVITY 60-148)
 Start/End Dates : 27-MAY-2009 07:45:52 through 25-NOV-2009 12:00:00
 Lower/Upper Lmts: 85,1104 through 94,0694



QA filename : DKA100:[ENV_ALPHA.QA,B]B206.QAF;1
Parameter Name : BACKRATE (Background Rate)
Start/End Dates : 3-MAY-2009 13:56:09 through 25-NOV-2009 12:00:00
Lower/Upper Lmts: 0.000000E+00 through 0.100000



RUNLOGS

Instrument Run Log

Instrument Type: LUCAS CELL DETECTOR

Batch ID: 920697

Sample ID	Sample Type	Analyst	Instrument	Run Date	Status	Geometry	Calibration Date
239753001	SAMPLE	KSD1	LUCAS7	01-DEC-09 18:05	DONE	Lucas Cell	30-SEP-09 00:00
239753002	SAMPLE	KSD1	LUCAS1	01-DEC-09 18:40	DONE	Lucas Cell	31-AUG-09 00:00
239753003	SAMPLE	KSD1	LUCAS2	01-DEC-09 18:40	DONE	Lucas Cell	19-DEC-08 00:00
239753004	SAMPLE	KSD1	LUCAS3	01-DEC-09 18:40	DONE	Lucas Cell	04-FEB-09 00:00
239753005	SAMPLE	KSD1	LUCAS4	01-DEC-09 18:40	DONE	Lucas Cell	02-MAR-09 00:00
239753006	SAMPLE	KSD1	LUCAS5	01-DEC-09 18:40	DONE	Lucas Cell	25-MAR-09 00:00
239753007	SAMPLE	KSD1	LUCAS7	01-DEC-09 18:40	DONE	Lucas Cell	30-SEP-09 00:00
239753008	SAMPLE	KSD1	LUCAS1	01-DEC-09 19:10	DONE	Lucas Cell	31-AUG-09 00:00
239753009	SAMPLE	KSD1	LUCAS2	01-DEC-09 19:10	DONE	Lucas Cell	19-DEC-08 00:00
239753010	SAMPLE	KSD1	LUCAS3	01-DEC-09 19:10	DONE	Lucas Cell	04-FEB-09 00:00
239753011	SAMPLE	KSD1	LUCAS5	01-DEC-09 19:10	DONE	Lucas Cell	25-MAR-09 00:00
239753012	SAMPLE	KSD1	LUCAS7	01-DEC-09 19:10	DONE	Lucas Cell	30-SEP-09 00:00
239753013	SAMPLE	KSD1	LUCAS1	01-DEC-09 20:00	DONE	Lucas Cell	31-AUG-09 00:00
239753014	SAMPLE	KSD1	LUCAS2	01-DEC-09 20:00	DONE	Lucas Cell	19-DEC-08 00:00
239753015	SAMPLE	KSD1	LUCAS3	01-DEC-09 20:00	DONE	Lucas Cell	04-FEB-09 00:00
239753016	SAMPLE	KSD1	LUCAS5	01-DEC-09 20:00	DONE	Lucas Cell	25-MAR-09 00:00
239753017	SAMPLE	KSD1	LUCAS7	01-DEC-09 20:00	DONE	Lucas Cell	30-SEP-09 00:00
239753018	SAMPLE	KSD1	LUCAS2	01-DEC-09 20:35	DONE	Lucas Cell	19-DEC-08 00:00
1201967363 MB		KSD1	LUCAS3	01-DEC-09 20:35	DONE	Lucas Cell	04-FEB-09 00:00
1201967364 DUP		KSD1	LUCAS5	01-DEC-09 20:35	DONE	Lucas Cell	25-MAR-09 00:00
1201967365 MS		KSD1	LUCAS7	01-DEC-09 20:35	DONE	Lucas Cell	30-SEP-09 00:00
1201967366 LCS		KSD1	LUCAS7	01-DEC-09 21:25	DONE	Lucas Cell	30-SEP-09 00:00

Instrument Run Log

Instrument Type: GFPC

Batch ID: 922859

Sample ID	Sample Type	Analyst	Instrument	Run Date	Status	Geometry	Calibration Date
239753002	SAMPLE	JXC5	PIC13A	23-NOV-09 17:23	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753011	SAMPLE	JXC5	PIC14A	23-NOV-09 17:23	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753012	SAMPLE	JXC5	PIC12C	23-NOV-09 17:23	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753018	SAMPLE	JXC5	PIC12D	23-NOV-09 17:23	DONE	CeF on 25mm Filter	02-JUL-09 00:00
1201972469	DUP	JXC5	PIC9A	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
1201972470	MS	JXC5	PIC9B	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
1201972471	LCS	JXC5	PIC3A	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753001	SAMPLE	JXC5	PIC1A	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753003	SAMPLE	JXC5	PIC8A	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753004	SAMPLE	JXC5	PIC1C	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753005	SAMPLE	JXC5	PIC2A	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753006	SAMPLE	JXC5	PIC7C	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753007	SAMPLE	JXC5	PIC6B	23-NOV-09 17:26	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753008	SAMPLE	JXC5	PIC5C	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753009	SAMPLE	JXC5	PIC9D	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753010	SAMPLE	JXC5	PIC7D	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753013	SAMPLE	JXC5	PIC10D	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753014	SAMPLE	JXC5	PIC1D	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753015	SAMPLE	JXC5	PIC8C	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753016	SAMPLE	JXC5	PIC7A	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
239753017	SAMPLE	JXC5	PIC10C	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00
1201972468	MB	JXC5	PIC6D	23-NOV-09 17:27	DONE	CeF on 25mm Filter	02-JUL-09 00:00

Instrument Run Log

Instrument Type: ALPHA SPECTROMETER

Batch ID: 923093

Sample ID	Sample Type	Analyst	Instrument	Run Date	Status	Geometry	Calibration Date
239753016	SAMPLE	KXM4	1201	20-NOV-09 14:16	DONE		
239753017	SAMPLE	KXM4	1202	20-NOV-09 14:16	DONE		
239753018	SAMPLE	KXM4	1203	20-NOV-09 14:16	DONE		
1201973223	MB	KXM4	1204	20-NOV-09 14:16	DONE		
1201973224	DUP	KXM4	1205	20-NOV-09 14:16	DONE		
1201973225	MS	KXM4	1206	20-NOV-09 14:16	DONE		
239753001	SAMPLE	KXM4	1025	20-NOV-09 16:04	DONE		
239753002	SAMPLE	KXM4	1026	20-NOV-09 16:04	DONE		
239753003	SAMPLE	KXM4	1027	20-NOV-09 16:04	DONE		
239753004	SAMPLE	KXM4	1028	20-NOV-09 16:04	DONE		
239753005	SAMPLE	KXM4	1029	20-NOV-09 16:04	DONE		
239753006	SAMPLE	KXM4	1030	20-NOV-09 16:04	DONE		
239753007	SAMPLE	KXM4	1031	20-NOV-09 16:05	DONE		
239753008	SAMPLE	KXM4	1033	20-NOV-09 16:05	DONE		
239753009	SAMPLE	KXM4	1035	20-NOV-09 16:05	DONE		
239753010	SAMPLE	KXM4	1036	20-NOV-09 16:05	DONE		
239753011	SAMPLE	KXM4	1037	20-NOV-09 16:05	DONE		
239753012	SAMPLE	KXM4	1038	20-NOV-09 16:05	DONE		
239753013	SAMPLE	KXM4	1039	20-NOV-09 16:05	DONE		
239753014	SAMPLE	KXM4	1040	20-NOV-09 16:05	DONE		
239753015	SAMPLE	KXM4	1041	20-NOV-09 16:05	DONE		
1201973226	LCS	KXM4	1042	20-NOV-09 16:05	DONE		

Instrument Run Log

Instrument Type: ALPHA SPECTROMETER

Batch ID: 923094

Sample ID	Sample Type	Analyst	Instrument	Run Date	Status	Geometry	Calibration Date
1201973227	MB	KXM4	1121	20-NOV-09 14:25	DONE		
1201973228	DUP	KXM4	1122	20-NOV-09 14:25	DONE		
1201973229	MS	KXM4	1123	20-NOV-09 14:25	DONE		
1201973230	LCS	KXM4	1124	20-NOV-09 14:25	DONE		
239753001	SAMPLE	KXM4	1138	21-NOV-09 15:52	DONE		
239753002	SAMPLE	KXM4	1139	21-NOV-09 15:52	DONE		
239753003	SAMPLE	KXM4	1140	21-NOV-09 15:52	DONE		
239753004	SAMPLE	KXM4	1141	21-NOV-09 15:52	DONE		
239753005	SAMPLE	KXM4	1142	21-NOV-09 15:52	DONE		
239753006	SAMPLE	KXM4	1143	21-NOV-09 15:52	DONE		
239753007	SAMPLE	KXM4	1144	21-NOV-09 15:52	DONE		
239753008	SAMPLE	KXM4	1145	21-NOV-09 15:52	DONE		
239753009	SAMPLE	KXM4	1146	21-NOV-09 15:52	DONE		
239753010	SAMPLE	KXM4	1147	21-NOV-09 15:52	DONE		
239753011	SAMPLE	KXM4	1148	21-NOV-09 15:52	DONE		
239753012	SAMPLE	KXM4	1149	21-NOV-09 15:52	DONE		
239753013	SAMPLE	KXM4	1150	21-NOV-09 15:52	DONE		
239753014	SAMPLE	KXM4	1151	21-NOV-09 15:52	DONE		
239753015	SAMPLE	KXM4	1152	21-NOV-09 15:52	DONE		
239753016	SAMPLE	KXM4	1153	21-NOV-09 15:52	DONE		
239753017	SAMPLE	KXM4	1154	21-NOV-09 15:52	DONE		
239753018	SAMPLE	KXM4	1155	21-NOV-09 15:52	DONE		