

Comments Regarding Appendix C:

Perchlorate results for 33 samples were reanalyzed for result confirmation and both results were reported by the laboratory. The least technically acceptable result for each pair was marked rejected in the data validation process. These rejected results are listed in Attachment B, Section IX of this Data Validation Summary Report. Both sets of results are included in the Electronic Data Deliverable (Appendix D). The rejected results have a final_validation_qualifier of "R".