

Table III. Reason Codes and Definitions

| Reason Code | Explanation |
|-------------|---|
| a | qualified due to low abundance (radiochemical activity) |
| be | qualified due to equipment blank contamination |
| bf | qualified due to field blank contamination |
| bl | qualified due to lab blank contamination |
| bt | qualified due to trip blank contamination |
| bp | qualified due to pump blank contamination (wells w/o dedicated pumps, when contamination is detected in the Pump Blk) |
| br | qualified due to filter blank contamination (aqueous Hexavalent Chromium and Dissolved sample fractions) |
| c | qualified due to calibration problems |
| cp | qualified due to insufficient ingrowth (radiochemical only) |
| dc | duel column confirmation %D exceeded |
| e | concentration exceeded the calibration range |
| fd | qualified due to field duplicate imprecision |
| h | qualified due to holding time exceedance |
| i | qualified due to internal standard areas |
| k | qualified as Estimated Maximum Possible Concentrations (dioxins and PCB congeners) |
| l | qualified due to LCS recoveries |
| ld | qualified due to lab duplicate imprecision (matrix duplicate, MSD, LCSD) |
| m | qualified due to matrix spike recoveries |
| nb | qualified due to negative lab blank contamination (nondetect results only) |
| nd | qualified due to non-detected target analyte |
| o | other |
| p | qualified as a false positive due to contamination during shipping |
| pH | sample preservation not within acceptance range |
| q | qualified due to quantitation problem |
| s | qualified due to surrogate recoveries |
| sd | serial dilution did not meet control criteria |
| sp | SQL < detect < PQL |
| st | sample receipt temperature exceeded |
| t | qualified due to elevated helium tracer concentrations |
| vh | volatile headspace detected in aqueous sample containers submitted for VOC analysis |
| x | qualified due to low % solids |
| z | qualified due to ICS results |