

TABLE OF CONTENTS

Cover Page	2
QC Checklist	3
Quality Investigation Report.....	4
Sequence Log	8
Summary Sheet.....	11
Initial Calibration.....	16
QC: (Method Blank; MRL Check; LCS/LCSD; Matrix Spike 2708020271 MS/MSD)	46
Periodic QC	53
Samples	58
QC: (Method Blank; MRL Check; LCS/LCSD; Matrix Spike 2708020327 MS/MSD)	61
Samples	70
Periodic QC	75
Samples	79
Periodic QC	89
Samples	94
Standards Preparation Worksheet and Certificates of Analysis	96
Digestion Logs.....	152

Level IV Data Package

MWH Group 212147

Method: 6010 CR

Sample No.:

2708020317
2708020318
2708020319
2708020320
2708020321
2708020322
2708020323
2708020324
2708020325
2708020326
2708020327
2708020328
2708020329
2708020330
2708020331
2708020342
2708020343
2708020344

EPA 200.7/6010B QC Check List

Analyst W3 Analysis Date 8/13/07 Reviewer/Date WZ 8/16/07

Instrument PerKin Elmer Optima 4300DV

All sample analyzed within 6 month holding time

All sample raw concentration below the high standard or linear
anged samples marked for dilution and rerun

Initial and closing QC

- ICV within +/- 5%
- Linearity check +/- 10%
- ICSAB +/- 20%
- 1 PPM check +/- 10%
- MRL +/- 50%

Middle, closing and batch QC

- FilterCheck < 1/2 MRL
- MBLANK < 1/2 MRL
- LCS +/- 15%
- MS/MSD +/- 30% (200.7) +/- 25% (6010B)
- CCV/MCV/ECV +/- 10%
- ICB/CCB/ECB < 1/2 MRL
- CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/- 20%
- RPD between LCS/LCSD is within +/- 20%
- Internal standards +/- 20%
- All pH of the samples are < 2

No more than 20 samples per batch

MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

QIR needed for failed QC for 2n MS/MSD "M3"

Special Det Code noted on the cover sheet Cr6010.

R value for multi point calibration is > 0.995

Proper MRL check ran for special low MRL samples

Reagent and Standards used for
Optima 4300 DV
Updated 08/08/07

Int: W3
Date: 8/13/07

Method 200.7/6010

Received by Supervisor on 14-aug-2007

QIR initiated by: wbh

QUALITY INVESTIGATION REPORT

QIR No.: METL_213187

Analysis date: 081307
Analyst: wbh
Method reference: ML-EPA 601
Analytical instrument: OPTIMA
Extraction Date: 080207
Prepared By: jrf

Group	Sample#	Sample ID	Customer	QC Ref	Test	PM
212147	2708020323	M-19	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020321	M-34	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020322	M-35	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020320	M-50	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020327	I-K	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020324	M-39	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020326	M-61	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020325	M-68	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020331	I-I	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020328	I-J	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020329	I-Z	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020330	M-67	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020343	EB-2	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020332	I-V	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020342	M-12A	KERRMCGEE-MP	380839	CR6010	ADE
212147	2708020344	MD-2	KERRMCGEE-MP	380839	CR6010	ADE
212263	2708030226	M-73	KERRMCGEE-MP	380839	CR6010	ADE
212263	2708030225	M-74	KERRMCGEE-MP	380839	CR6010	ADE
212263	2708030224	M-87	KERRMCGEE-MP	380839	CR6010	ADE
212263	2708030227	M-88	KERRMCGEE-MP	380839	CR6010	ADE

Brief Description: (include reason for non-compliance-Root Cause)

Sample 2708020332 was used as 2nd set MS/D pair for this batch of samples. A flag (M3) was added because spike recovery for the MS failed acceptance limits at 20%. All other QCs passed. LCS/D were recovering 93/91%. And the MSD recovered 86 percent. Another spike set 2708020327 recovered 91 and 90%.

Corrective Action Taken/Prevention:

Sample was qualified with an M3 flag and all samples with this batch were reported. Matrix interference suspected.

Impact on Data Quality:

Sample concentration was also high and at the same level as the spike. None as MS/D recoveries do not affect quality.

LIMS user:mce Date/time stamp:24-aug-2007 15:37:37

Data Disposition/Acceptable/Method/Regulations:

Flag sample -0332 with M3 since analyte conc. in sample was 20x of spiked amount. Batch data acceptable based on passing LCSs and 1st MS/MSD.

LIMS user:ycy Date/time stamp:24-aug-2007 17:51:08

Annotation:

M3 - The accuracy of the spike recovery value is reduced since the analyte concentration in the sample is disproportionate to the spike level. The method control sample recovery was acceptable.

LIMS user:ycy Date/time stamp:24-aug-2007 17:51:08

Client Contact:

ok to report with M3 flag for impacted sample

LIMS user:ade Date/time stamp:25-aug-2007 08:46:43

Detail Report for QIR group#

213187

Group	Sample#	Sample ID	Customer	QC Ref	Test	Analyst	Analysis Date	Prep	Prep Date	Inst
212147	2708020320	M-50	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020321	M-34	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020322	M-35	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020323	M-19	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020324	M-39	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020325	M-68	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020326	M-61	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020327	I-K	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020328	I-J	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020329	I-2	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020330	M-67	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020331	I-I	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020332	I-V	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020342	M-12A	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020343	EB-2	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212147	2708020344	MD-2	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212263	2708030224	M-87	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212263	2708030225	M-74	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212263	2708030226	M-73	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA
212263	2708030227	M-88	KERRMCGEE-MP	380839	CR6010	wbh	08/13/07 00:00	jrf	08/12/07 00:00	OPTIMA

Batch# 380839 CR6010

Analyte	QC	Actual	Found	Lower	Yield	Upper	Statu
Chromium, Total, ICAP	LCS1	1.00	0.930	85.0	93.0	115.0	OK
Chromium, Total, ICAP	LCS2	1.00	0.911	85.0	91.1	115.0	OK
Chromium, Total, ICAP	MBLK	ND	ND	0.0		0.0	OK
Chromium, Total, ICAP	MRL_CHK	0.010	0.009	50.0	99.4	150.0	OK
Chromium, Total, ICAP	MS	1.00	0.914	70.0	91.4	130.0	OK
Chromium, Total, ICAP	MSD	1.00	0.895	70.0	89.5	130.0	OK

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	8/13/07	19:11	1	9.9636	9.96 ✓	95-105	99.6%
LINEARITY	8/13/07	19:14	1	0.0016	.0016		
ICSA	8/13/07	19:18	1	-0.0011	ND ✓	80-120	
ICSAB	8/13/07	19:22	1	.23743	.237 ✓	80-120	94.9%
Wash	8/13/07	19:26	1	-0.0001	ND		
QC-25 ppm	8/13/07	19:32	1	.95734	.960 ✓		
CCV	8/13/07	19:37	1	4.9148	4.91 ✓	90-110	98.2%
ICB	8/13/07	19:48	1	0.0000	0.0000		
MRL	8/13/07	19:54	1	0.0098	.0098 ✓	50-150	97.5%
FILTER CHECK	8/13/07	19:57	1	-0.0002	ND		
MRL6010	8/13/07	20:01	1	-0.0027	ND <i>No Sp. Se 0. Perm 0.1</i>		
MBLANK	8/13/07	20:05	1	-0.0001	ND <i>Wrong</i>		
LCS	8/13/07	20:08	1	.95525	.955 ✓ <i>fx locatn</i>	85-115	95.5%
LCSD	8/13/07	20:12	1	.94145	.941 ✓	85-115	94.1%
2708020271_2X	8/13/07	20:16	2	.21518	.220 ✓		
2708020271_2XMS	8/13/07	20:20	2	1.1761	1.18 ✓	0.961	48.0 Q <i>96.1</i>
2708020271_2XMSD	8/13/07	20:24	2	1.1556	1.16 ✓	0.941	47.0 Q <i>94.1</i>
2708020271_2XT	8/13/07	20:24	2		2.00	70 - 130	54.1 <i>54.1</i>
2708020272_2X	8/13/07	20:28	2	.48722	.490 ✓		
2708020272_2XMS	8/13/07	20:32	2	1.4206	1.42	0.933	46.6 Q <i>93.3</i>
CCV	8/13/07	20:36	1	4.8635	4.86	90-110	97.2% <i>97.2%</i>
CCB	8/13/07	20:39	1	0.0000	0.0000		
2708020272_2XMSD	8/13/07	20:43	2	1.4252	1.43	0.938	46.9 Q <i>93.8</i>
2708020272_2XT	8/13/07	20:43	2		2.00	70 - 130	
2708020256_20X	8/13/07	20:46	20	29.366	29 ✓		
2708020257_20X	8/13/07	20:51	20	28.933	29 ✓		
2708020258_20X	8/13/07	20:55	20	16.424	16 ✓		
2708020259_20X	8/13/07	20:59	20	43.453	43 ✓		
2708020260_20X	8/13/07	21:04	20	30.262	30 ✓		
2708020261_20X	8/13/07	21:08	20	30.987	31 ✓		
2708020262_20X	8/13/07	21:12	20	25.971	26 ✓		
2708020263_20X	8/13/07	21:17	20	14.973	15 ✓		
2708020264_20X	8/13/07	21:21	20	12.622	13 ✓		
CCV	8/13/07	21:25	1	4.7959	4.8 ✓	90-110	95.9%
CCB	8/13/07	21:29	1	0.0022	.0022		
MCV	8/13/07	21:32	1	2.3859	2.39 ✓	90-110	95.4%
2708020265_10X	8/13/07	21:35	10	11.131	11 ✓		
2708020266_10X	8/13/07	21:40	10	9.4001	9.4 ✓		
2708020267_5X	8/13/07	21:44	5	4.3750	4.4 ✓		
2708020268_2X	8/13/07	21:48	2	1.8711	1.9 ✓		
2708020269_2X	8/13/07	21:53	2	.79782	.8 ✓		
2708020270_2X	8/13/07	21:57	2	.44294	.440 ✓		
2708020317_2X	8/13/07	22:02	2	0.0161	.016 ✓		
2708020319_5X	8/13/07	22:06	5	11.356	11 ✓		
2708020318_2X	8/13/07	22:11	2	.06087	.061 ✓		
MBLANK	8/13/07	22:15	1	0.0007	0.0007 ✓		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
CCV	8/13/07	22:19	1	4.7548	4.75 ✓	90-110	95.0%
CCB	8/13/07	22:24	1	0.0009	0.0009 ✓		
LCS	8/13/07	22:28	1	.93005	.93 ✓	85-115	93.0%
LCSD	8/13/07	22:31	1	.91146	.911 ✓	85-115	91.1%
2708020327_2X	8/13/07	22:34	2	.96783	.970 ✓		
2708020327_2XMS	8/13/07	22:38	2	1.8816	1.88 ✓	0.9141	91.4% Q
2708020327_2XMSD	8/13/07	22:42	2	1.8625	1.86 ✓	0.8951	89.5% Q
2708020327_2XT	8/13/07	22:42	2		2.00	70 - 138	87.5%
2708020332_20X	8/13/07	22:46	20	21.102	21 ✓		
2708020332_20XMS	8/13/07	22:50	20	21.300	21.3 ✓	0.1981	19.8% Q
2708020332_20XMSD	8/13/07	22:55	20	21.963	22 ✓	0.8611	86.1% Q
2708020332_20XT	8/13/07	22:55	20		20.00	70 - 130	86.1%
2708020320_20X	8/13/07	22:59	20	29.441	29 ✓		
2708020321_20X	8/13/07	23:03	20	15.929	16 ✓		
CCV	8/13/07	23:08	1	4.7324	4.73 ✓	90-110	94.6%
CCB	8/13/07	23:11	1	0.0022	.0022 ✓		
MCV	8/13/07	23:16	1	2.3620	2.36 ✓	90-110	94.4%
2708020322_10X	8/13/07	23:19	10	9.4485	9.4 ✓		
2708020323_2X	8/13/07	23:24	2	.38131	.380 ✓		
2708020324_5X	8/13/07	23:28	5	4.5793	4.6 ✓		
2708020325_2X	8/13/07	23:32	2	.91385	.910 ✓		
2708020326_2X	8/13/07	23:37	2	1.0224	1.0 ✓		
2708020328_2X	8/13/07	23:41	2	2.3597	2.4 ✓		
2708020329_5X	8/13/07	23:46	5	9.4749	9.5 ✓		
2708020330_5X	8/13/07	23:50	5	5.5403	5.5 ✓		
2708020331_20X	8/13/07	23:54	20	19.785	20 ✓		
2708020342_10X	8/13/07	23:59	10	12.578	13 ✓		
CCV	8/14/07	0:05	1	4.7427	4.74 ✓	90-110	94.8%
CCB	8/14/07	0:13	1	0.0010	.001 ✓		
2708020343_2X	8/14/07	0:17	2	0.0033	.0033 ✓		
2708020344_5X	8/14/07	0:20	5	11.969	12 ✓		
2708030224_2X	8/14/07	0:25	2	1.8657	1.9 ✓		
2708030225_2X	8/14/07	0:29	2	.76172	.760 ✓		
2708030226_2X	8/14/07	0:34	2	2.8850	2.9 ✓		
2708030227_2X	8/14/07	0:38	2	.87133	.870 ✓		
MBLANK	8/14/07	0:43	1	0.0006	0.0005		
LCS	8/14/07	0:46	1	.91435	.914 ✓	85-115	91.4%
LCSD	8/14/07	0:50	1	.93482	.935 ✓	85-115	93.4%
2708040004_5X	8/14/07	0:54	5	5.0792	5.1 ✓		
CCV	8/14/07	1:01	1	4.7245	4.72 ✓	90-110	94.4%
CCB	8/14/07	1:10	1	0.0003	0.0002		
MCV	8/14/07	1:15	1	2.3292	2.33 ✓	90-110	93.4%
2708040004_5XMS	8/14/07	1:18	5	5.8624	5.86 ✓	0.7831	78.3% Q
2708040004_5XMSD	8/14/07	1:22	5	5.7952	5.8 ✓	0.7161	71.6% Q
2708040004_5XT	8/14/07	1:22	5		5.00	70 - 130	71.6%
2708040006_20X	8/14/07	1:27	20	29.675	30 ✓		
2708040006_20XMS	8/14/07	1:31	20	29.922	29.9 ✓	0.2471	24.7% Q
2708040006_20XMSD	8/14/07	1:35	20	29.989	30 ✓	0.3141	31.4% Q
2708040006_20XT	8/14/07	1:35	20		20.00	70 - 130	
2708030228_2X	8/14/07	1:40	2	1.3358	1.3 ✓		

GA 212/147

GA 212/147
0.380859

GA 212/263

GA 212/263
09-360840

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2708030229_2X	8/14/07	1:44	2	.46897	.470 ✓		
2708030230_2X	8/14/07	1:49	2	1.7594	1.8 ✓		
2708030231_2X	8/14/07	1:53	2	0.0797	.080 ✓		
2708030232_2X	8/14/07	1:57	2	.19337	.190 ✓		
CCV	8/14/07	2:05	1	4.7200	4.72 ✓	90-110	94.4%
CCB	8/14/07	2:11	1	0.0003	0.0003		
2708030233_20X	8/14/07	2:14	20	32.887	33 ✓		
2708030234_2X	8/14/07	2:19	2	2.5954	2.6 ✓		
2708030235_2X	8/14/07	2:23	2	.96589	.970 ✓		
2708030239_5X	8/14/07	2:27	5	0.0419	.042 ✓		
2708040001_2X	8/14/07	2:31	2	0.0746	.075 ✓		
2708040002_2X	8/14/07	2:36	2	0.0933	.093 ✓		
2708040003_5X	8/14/07	2:40	5	7.4532	7.5 ✓		
2708040005_5X	8/14/07	2:45	5	4.7907	4.8 ✓		
2708040007_20X	8/14/07	2:49	20	25.929	26 ✓		
2708040008_20X	8/14/07	2:53	20	22.921	23 ✓		
CCV	8/14/07	3:01	1	4.7569	4.76 ✓	90-110	95.1%
CCB	8/14/07	3:10	1	0.0008	0.0007 ✓		
MCV	8/14/07	3:15	1	2.3193	2.32 ✓	90-110	92.7%
2708040009_20X	8/14/07	3:19	20	27.091	27 ✓		
2708040010_2X	8/14/07	3:24	2	0.0357	.036 ✓		
2708040011_2X -	8/14/07	3:28	2	0.0534	.053 ✓		
ICSA	8/14/07	3:33	1	-0.0009	ND ✓	80-120	
ICSAB	8/14/07	3:37	1	.22613	.226 ✓	80-120	90.4%
Wash	8/14/07	3:41	1	0.0000	0		
QC-25 1ppm	8/14/07	3:47	1	.19770	.2 ✓		
ECV	8/14/07	3:52	1	4.6691	4.67 ✓	90-110	93.3%
ECB	8/14/07	3:59	1	-0.0003	ND ✓		
MRL	8/14/07	4:05	1	0.0090	.009 ✓	50-150	90.4%

ICP SUMMARY SHEET

File ID: 070813A
Date Started: 8/13/07
Analyst ID: WBH

SAMPLE ID

LINEARITY	(19:14)	Wash	(19:26)	FILTER CHECK	(19:57)
2708020271_2	(20:16)	2708020272_2	(20:28)	2708020256_2	(20:46)
2708020257_2	(20:51)	2708020258_2	(20:55)	2708020259_2	(20:59)
2708020260_2	(21:04)	2708020261_2	(21:08)	2708020262_2	(21:12)
2708020263_2	(21:17)	2708020264_2	(21:21)	2708020265_1	(21:35)
2708020266_1	(21:40)	2708020267_5	(21:44)	2708020268_2	(21:48)
2708020269_2	(21:53)	2708020270_2	(21:57)	2708020317_2	(22:02)
2708020319_5	(22:06)	2708020318_2	(22:11)	2708020327_2	(22:34)
2708020332_2	(22:46)	2708020320_2	(22:59)	2708020321_2	(23:03)
2708020322_1	(23:19)	2708020323_2	(23:24)	2708020324_5	(23:28)
2708020325_2	(23:32)	2708020326_2	(23:37)	2708020328_2	(23:41)
2708020329_5	(23:46)	2708020330_5	(23:50)	2708020331_2	(23:54)
2708020342_1	(23:59)	2708020343_2	(0:17)	2708020344_5	(0:20)
2708030224_2	(0:25)	2708030225_2	(0:29)	2708030226_2	(0:34)
2708030227_2	(0:38)	2708040004_5	(0:54)	2708040006_2	(1:27)
2708030228_2	(1:40)	2708030229_2	(1:44)	2708030230_2	(1:49)
2708030231_2	(1:53)	2708030232_2	(1:57)	2708030233_2	(2:14)
2708030234_2	(2:19)	2708030235_2	(2:23)	2708030239_5	(2:27)
2708040001_2	(2:31)	2708040002_2	(2:36)	2708040003_5	(2:40)
2708040005_5	(2:45)	2708040007_2	(2:49)	2708040008_2	(2:53)
2708040009_2	(3:19)	2708040010_2	(3:24)	2708040011_2	(3:28)
Wash	(3:41)				

Cr6010

COMMENT:

MRL for Cr6010 \Rightarrow 0.00994 μ SL.

Analyst: WBJ

Approved By: WR 8/16/07

BATCH NUMBER for 070813A

Amulya
8/14/07

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2708020271_2X

2708020271_2X	2708020272_2X	2708020256_20X
2708020257_20X	2708020258_20X	2708020259_20X
2708020260_20X	2708020261_20X	2708020262_20X
2708020263_20X	2708020264_20X	2708020265_10X
2708020266_10X	2708020267_5X	2708020268_2X
2708020269_2X	2708020270_2X	2708020317_2X
2708020319_5X	2708020318_2X	

Batch ID: 2708020327_2X

2708020327_2X	2708020332_20X	2708020320_20X
2708020321_20X	2708020322_10X	2708020323_2X
2708020324_5X	2708020325_2X	2708020326_2X
2708020328_2X	2708020329_5X	2708020330_5X
2708020331_20X	2708020342_10X	2708020343_2X
2708020344_5X	2708030224_2X	2708030225_2X
2708030226_2X	2708030227_2X	

Batch ID: 2708040004_5X

2708040004_5X	2708040006_20X	2708030228_2X
2708030229_2X	2708030230_2X	2708030231_2X
2708030232_2X	2708030233_20X	2708030234_2X
2708030235_2X	2708030239_5X	2708040001_2X
2708040002_2X	2708040003_5X	2708040005_5X
2708040007_20X	2708040008_20X	2708040009_20X
2708040010_2X	2708040011_2X	

Landscap Summary

File ID: 070813A

Date: 8/13/07

Analyst: WBH

Page: 4

Sample ID	Time	CR	CU	FE	K	MG	MN	MO	NA	NI	PB	SB
ICV	19:11	9.96/10	10.1/10	10.0/10	101/100	102/100	10.0/10	103/100	103/100	9.98/10	10.3/10	10.1/10
LINEARITY	19:14	0.002	-0.015	101	316	200	0.004	0.002	319/300	0.001	-0.007	0.013
ICSA	19:18	-0.001	-0.016	100/100	0.390	254/250	0.003	0.003	0.186	0.000	-0.046	0.010
ICSA B	19:22	0.237/.25	0.233/.25	97.0/100	0.168	244/250	0.250/.25	0.002	0.066	0.449/.5	0.440/.5	0.011
Wash	19:26	-0.001	0.0000	-0.0021	0.0007	0.0009	-0.0004	0.002	0.0192	0.001	0.0010	-0.0058
QC-25 ppm	19:32	0.9573	0.9309	1.014	9.552	1.078	0.9991	0.9150	1.052	1.005	1.032	0.9333
CCV	19:37	4.91/5	4.77/5	5.16/5	49.7/50	52.4/50	5.02/5	4.89/5	51.7/50	5.06/5	5.12/5	4.92/5
ICB	19:48	0.0000	0.0002	0.0001	-0.0034	0.0008	-0.0004	0.0001	0.0111	0.0000	0.0015	-0.0025
MRL	19:54	0.010/.01	0.010/.01	0.021/.02	0.902/1	0.107/.1	N/A	0.019/.02	1.04/1	0.020/.02	0.022/.02	N/A
FILTER CHECK	19:57	-0.002	-0.001	-0.0011	-0.0646	0.0001	N/A	-0.003	0.0038	-0.000	0.0018	N/A
MRL6010	20:01	-0.003	-0.007	-0.003	-0.029	-0.000	N/A	-0.001	-0.000	-0.000	0.007(.02)	N/A
MBLANK	20:05	-0.001	-0.001	0.0016	-0.0656	0.0045	N/A	0.000	0.0247	-0.000	0.0024	N/A
LCS	20:08	0.955/1	0.961/1	5.03/5	19.3/20	20.6/20	N/A	0.953/1	51.0/50	0.498/.5	1.04/1	N/A
LCSD	20:12	0.941/1	0.956/1	5.07/5	19.6/20	20.7/20	N/A	0.933/1	51.1/50	0.487/.5	1.02/1	N/A
2708020271_2X	20:16	0.2152	0.0046	0.0014	35.25	181.5	N/A	0.0275	880.6	-0.015	-0.0882	N/A
2708020271_2XMSD	20:20	1.176	1.007	5.318	57.15	201.2	N/A	0.9954	924.9	0.4808	1.019	N/A
2708020271_2XMSD	20:24	1.156	0.9911	5.153	56.01	198.6	N/A	0.9852	913.7	0.4742	0.9896	N/A
2708020272_2X	20:28	0.4872	0.0042	0.3238	46.57	214.6	N/A	0.0301	856.1	0.0009	-0.129	N/A
2708020272_2XMS	20:32	1.421	0.9819	5.469	66.09	228.0	N/A	0.9981	882.7	0.4698	0.9949	N/A
CCV	20:36	4.86/5	4.72/5	5.02/5	49.1/50	52.4/50	N/A	4.77/5	51.8/50	4.98/5	5.04/5	N/A
CCB	20:39	0.0000	0.0001	0.0012	0.0133	-0.0001	N/A	0.0006	0.1645	-0.001	0.0026	N/A
2708020272_2XMSD	20:43	1.425	0.9781	5.421	66.67	232.0	N/A	0.9777	898.7	0.4606	0.9758	N/A
2708020256_20X	20:46	29.37	0.0149	0.0959	55.77	529.6	N/A	0.0362	2374.2	-0.032	-0.0329	N/A
2708020257_20X	20:51	28.93	0.0132	0.0334	46.83	564.4	N/A	0.0456	2336.3	-0.043	-0.0493	N/A
2708020258_20X	20:55	16.42	0.0091	0.1123	41.79	462.1	N/A	0.0456	1655.3	-0.034	-0.0980	N/A
2708020259_20X	20:59	43.45	0.0523	23.27	53.91	664.0	N/A	0.0253	2330.0	0.1392	-0.0547	N/A
2708020260_20X	21:04	30.26	0.0364	0.9707	54.77	651.7	N/A	0.0373	2318.1	-0.038	-0.0576	N/A
2708020261_20X	21:08	30.99	0.0110	0.0027	45.47	624.0	N/A	0.0392	2290.3	-0.068	-0.0876	N/A
2708020262_20X	21:12	25.97	0.0091	-0.0300	36.73	492.3	N/A	0.0311	2334.0	0.0039	-0.0390	N/A
2708020263_20X	21:17	14.97	0.0013	-0.0169	47.11	433.5	N/A	0.0260	1901.8	-0.028	-0.0308	N/A
2708020264_20X	21:21	12.62	0.0060	0.3496	35.36	388.5	N/A	0.0269	1710.5	0.0005	-0.0399	N/A
CCV	21:25	4.80/5	4.63/5	5.16/5	50.2/50	53.0/50	N/A	4.75/5	52.0/50	5.00/5	5.03/5	N/A
CCB	21:29	0.0022	-0.001	-0.0012	-0.0200	-0.0000	N/A	0.0008	0.1064	-0.001	0.0019	N/A
MCV	21:32	2.39/2.5	2.28/2.5	2.58/2.5	24.6/25	26.8/25	N/A	2.32/2.5	26.0/25	2.50/2.5	2.51/2.5	N/A
2708020265_10X	21:35	11.13	0.0058	0.0157	43.19	330.3	N/A	0.0289	1733.6	-0.017	-0.0306	N/A
2708020266_10X	21:40	9.400	0.0053	1.122	40.64	295.8	N/A	0.0303	1853.6	0.0003	0.0038	N/A
2708020267_5X	21:44	4.375	0.0051	0.0080	38.68	224.6	N/A	0.0321	1597.4	-0.033	-0.0220	N/A
2708020268_2X	21:48	1.871	0.0028	0.0024	32.71	164.5	N/A	0.0341	1292.4	-0.006	-0.0099	N/A
2708020269_2X	21:53	0.7978	0.0029	0.0069	39.80	221.9	N/A	0.0272	1196.9	0.0002	-0.0104	N/A
2708020270_2X	21:57	0.4429	0.0032	0.0300	45.51	265.4	N/A	0.0206	1115.3	-0.004	-0.0178	N/A
2708020317_2X	22:02	0.0161	0.0024	0.0040	84.40	0.0168	N/A	0.0168	359.6	-0.006	-0.0043	N/A
2708020319_5X	22:06	11.36	0.0090	0.1397	23.07	278.3	N/A	0.0162	1642.2	-0.007	-0.0139	N/A
2708020318_2X	22:11	0.0609	0.0059	0.0744	18.58	199.2	N/A	0.0175	636.5	0.0004	-0.0028	N/A
MBLANK	22:15	0.0007	-0.0003	0.0015	-0.0280	0.0019	N/A	-0.0004	0.4622	-0.006	0.0020	N/A
CCV	22:19	4.75/5	4.55/5	4.99/5	48.9/50	53.2/50	N/A	4.61/5	51.9/50	4.86/5	4.90/5	N/A
CCB	22:24	0.0009	-0.0002	-0.0004	-0.0450	0.0009	N/A	0.0002	0.1693	-0.003	0.0019	N/A
LCS	22:28	0.930/1	0.917/1	5.08/5	19.4/20	20.8/20	N/A	0.902/1	49.8/50	0.474/.5	0.975/1	N/A
LCSD	22:31	0.911/1	0.904/1	5.04/5	19.4/20	20.8/20	N/A	0.902/1	50.0/50	0.474/.5	0.984/1	N/A
2708020327_2X	22:34	0.9678	0.0022	0.0086	21.47	268.7	N/A	0.0615	912.8	0.0001	-0.0147	N/A

Landscape Summary

File ID: 070813A

Date: 8/13/07

Analyst: WBH

Page: 5

Sample_ID	Time	CR	CU	FE	K	MG	MN	MO	NA	NI	PB	SB
2708020327_2XMS	22:38	1.882	0.9429	5.014	42.70	292.7	N/A	0.9687	983.4	0.4596	0.9289	N/A
2708020327_2XMSD	22:42	1.863	0.9330	4.987	42.32	288.6	N/A	0.9594	963.1	0.4556	0.9215	N/A
2708020332_20X	22:46	21.10	-0.003	0.0486	64.93	523.8	N/A	0.0359	\$1908.6	-0.008	-0.471	N/A
2708020332_20XMS	22:50	21.30	0.8765	5.050	81.93	519.2	N/A	0.9212	\$1874.3	0.4537	0.9266	N/A
2708020332_20XMSD	22:55	21.96	0.9208	5.140	82.61	532.0	N/A	0.9569	\$1916.9	0.4689	0.9693	N/A
2708020320_20X	22:59	29.44	-0.017	1.021	35.17	292.8	N/A	0.0423	\$2968.5	-0.009	-0.043	N/A
2708020321_20X	23:03	15.93	0.0044	-0.146	43.52	392.1	N/A	0.0385	\$1711.6	-0.0063	-0.0264	N/A
CCV	23:08	4.73/5	4.52/5	5.08/5	49.8/50	53.4/50	N/A	4.62/5	51.5/50	4.87/5	4.90/5	N/A
CCB	23:11	0.0022	-0.001	-0.012	-0.423	-0.000	N/A	0.0008	0.1601	0.0002	0.0011	N/A
MCV	23:16	2.36/2.5	2.24(2.5)	2.64/2.5	25.5/25	27.2/25	N/A	2.29/2.5	25.8/25	2.47/2.5	2.47/2.5	N/A
2708020322_10X	23:19	9.449	0.0041	-0.0099	82.14	254.6	N/A	0.0599	\$1309.1	-0.010	-0.134	N/A
2708020323_2X	23:24	0.3813	0.0029	0.0302	10.88	157.8	N/A	0.1619	\$1123.3	0.0047	-0.124	N/A
2708020324_5X	23:28	4.579	0.0031	0.0612	26.96	440.5	N/A	0.0220	895.4	-0.030	-0.211	N/A
2708020325_2X	23:32	0.9139	0.0037	0.2015	26.30	313.5	N/A	0.0951	\$1244.8	0.0016	-0.134	N/A
2708020326_2X	23:37	1.022	0.0019	-0.0336	19.26	274.5	N/A	0.0419	843.2	0.0002	-0.154	N/A
2708020328_2X	23:41	2.360	0.0026	-0.126	20.40	333.7	N/A	0.0300	900.9	0.0015	-0.145	N/A
2708020329_5X	23:46	9.475	0.0020	0.0023	38.00	464.0	N/A	0.0264	\$1241.8	0.0003	-0.175	N/A
2708020330_5X	23:50	5.540	0.0023	0.0405	26.77	454.3	N/A	0.0234	947.7	-0.012	-0.226	N/A
2708020331_20X	23:54	19.79	-0.021	0.0282	71.66	477.9	N/A	0.0364	\$1974.2	0.0026	-0.291	N/A
2708020342_10X	23:59	12.58	0.0063	0.4790	45.70	21.55	N/A	0.0286	\$2325.5	-0.024	0.0005	N/A
CCV	0:05	4.74/5	4.53/5	5.09/5	49.5/50	53.9/50	N/A	4.64/5	51.4/50	4.90/5	4.93/5	N/A
CCB	0:13	0.0010	-0.006	-0.003	-0.776	-0.012	N/A	0.0001	0.1676	-0.006	0.0012	N/A
2708020343_2X	0:17	0.0033	-0.001	0.0678	-0.0686	0.2255	N/A	0.0004	0.6041	-0.004	0.0057	N/A
2708020344_5X	0:20	11.97	0.0037	0.3767	46.00	20.45	N/A	0.0327	\$2254.1	-0.005	-0.026	N/A
2708020324_2X	0:25	1.866	0.0018	0.1374	12.70	101.0	N/A	0.0148	426.0	0.0028	-0.0028	N/A
2708030225_2X	0:29	0.7617	0.0024	-0.041	24.21	258.0	N/A	0.1031	\$1334.7	0.0009	-0.133	N/A
2708030226_2X	0:34	2.885	0.0019	0.1220	15.67	193.3	N/A	0.0165	471.5	-0.008	-0.101	N/A
2708030227_2X	0:38	0.8713	0.0031	0.0487	25.41	232.5	N/A	0.1180	\$1544.2	0.0021	-0.140	N/A
MELANK	0:43	0.0006	-0.007	0.0021	-0.721	0.0049	N/A	-0.0004	0.5792	-0.005	0.0030	N/A
LCS	0:46	0.914/1	0.895/1	5.02/5	19.9/20	21.6/20	N/A	0.885/1	52.0/50	0.468/5	0.965/1	N/A
LCS	0:50	0.935/1	0.913/1	5.16/5	19.8/20	21.5/20	N/A	0.911/1	51.4/50	0.482/5	0.985/1	N/A
2708040004_5X	0:54	5.079	0.0021	0.0148	25.81	271.9	N/A	0.0219	\$1151.5	-0.036	-0.199	N/A
CCV	1:01	4.72/5	4.47(5)	5.13/5	50.4/50	54.0/50	N/A	4.59/5	51.8/50	4.86/5	4.87/5	N/A
CCB	1:10	0.0003	-0.0006	-0.024	-0.700	0.0014	N/A	0.0002	0.1596	-0.007	0.0019	N/A
MCV	1:15	2.33/2.5	2.18(2.5)	2.58/2.5	25.0/25	27.3/25	N/A	2.24(2.5)	25.9/25	2.44/2.5	2.43/2.5	N/A
2708040004_5XMS	1:18	5.862	0.9216	5.191	46.41	288.0	N/A	0.9297	\$1184.2	0.4617	0.9530	N/A
2708040004_5XMSD	1:22	5.795	0.9069	4.997	45.31	283.3	N/A	0.8934	\$1156.9	0.4460	0.9158	N/A
2708040006_20X	1:27	29.68	-0.029	-0.0096	41.55	508.1	N/A	0.0314	\$2459.3	-0.009	-0.232	N/A
2708040006_20XMS	1:31	29.92	0.8867	5.092	58.08	515.6	N/A	0.9242	\$2439.7	0.4550	0.9315	N/A
2708040006_20XMSD	1:35	29.99	0.8866	5.158	59.40	522.2	N/A	0.9194	\$2475.4	0.4532	0.9137	N/A
2708030228_2X	1:40	1.336	0.0011	0.1122	9.807	85.01	N/A	0.0152	416.8	-0.002	-0.024	N/A
2708030229_2X	1:44	0.4690	0.0033	0.0796	9.126	92.07	N/A	0.0103	536.2	-0.012	-0.062	N/A
2708030230_2X	1:49	1.759	0.0010	0.0315	14.75	157.4	N/A	0.0129	521.1	-0.017	-0.104	N/A
2708030231_2X	1:53	0.0797	0.0004	0.0261	4.837	30.06	N/A	0.0103	158.7	-0.014	-0.019	N/A
2708030232_2X	1:57	0.1934	0.0005	0.0892	6.452	54.89	N/A	0.0107	289.6	-0.003	-0.023	N/A
CCV	2:05	4.72/5	4.46(5)	5.21/5	51.3/50	54.4/50	N/A	4.59/5	52.2/50	4.88/5	4.86/5	N/A
CCB	2:11	0.0003	-0.006	-0.017	-0.0095	0.0007	N/A	0.0004	0.1268	-0.005	0.0030	N/A
2708030233_20X	2:14	32.89	-0.003	0.2271	37.08	515.3	N/A	0.0404	\$2669.2	-0.009	-0.297	N/A
2708030234_2X	2:19	2.595	0.0007	0.8022	21.15	43.32	N/A	0.0238	986.5	-0.007	-0.009	N/A
2708030235_2X	2:23	0.9659	0.0000	4.864	14.61	128.5	N/A	0.0173	608.3	-0.007	-0.005	N/A

Landscape Summary

File ID: 070813A

Date: 8/13/07

Analyst: WBH

Page: 6

Sample ID	Time	CR	CU	FE	K	MG	MN	MO	NA	NI	PB	SB
2708030239 5X	2:27	0.0419	0.0197	0.1604	159.9	\$1147.9	N/A	0.5230	\$7911.1	0.0550	-0.0366	N/A
2708040001 2X	2:31	0.0746	0.0000	0.0367	5.634	38.81	N/A	0.0083	176.6	-0.0014	-0.0028	N/A
2708040002 2X	2:36	0.0933	0.0022	0.0685	4.179	28.90	N/A	0.0094	147.8	-0.0015	-0.0006	N/A
2708040003 5X	2:40	7.453	0.0019	0.0151	29.49	267.3	N/A	0.0188	\$1686.5	-0.0041	-0.0132	N/A
2708040005 5X	2:45	4.791	0.0027	0.1978	36.20	336.1	N/A	0.0281	\$1290.5	-0.0003	-0.0230	N/A
2708040007 20X	2:49	25.93	-0.0061	0.0999	41.88	454.4	N/A	0.0228	\$2384.1	-0.0173	0.0042	N/A
2708040008 20X	2:53	22.92	-0.0063	-0.0077	38.59	447.0	N/A	0.0278	\$2165.5	-0.0174	-0.0185	N/A
CCV	3:01	4.76/5	4.49(5)	5.18/5	51.2/50	54.9/50	N/A	4.61/5	52.7/50	4.89/5	4.89/5	N/A
CCB	3:10	0.0008	-0.0007	-0.0007	-0.0821	0.0017	N/A	0.0002	0.1332	-0.0008	0.0028	N/A
MCV	3:15	2.32/2.5	2.16(2.5)	2.59/2.5	25.5/25	27.6(25)	N/A	2.19(2.5)	26.3/25	2.39/2.5	2.37/2.5	N/A
2708040009 20X	3:19	27.09	-0.0024	0.0632	37.67	424.5	N/A	0.0581	\$2465.2	-0.0154	-0.0251	N/A
2708040010 2X	3:24	0.0357	0.0007	0.0552	14.12	138.2	N/A	0.0270	518.1	-0.0024	-0.0057	N/A
2708040011 2X	3:28	0.0534	0.0016	0.3418	9.994	162.1	N/A	0.0131	637.2	-0.0005	-0.0063	N/A
ICSAE	3:33	-0.01	-0.015	99.2/100	0.063	257/250	N/A	0.002	0.776	-0.001	-0.038	N/A
ICSAE	3:37	0.225/.25	0.212/.25	97.2/100	-0.011	250/250	N/A	0.002	0.465	0.424/.5	0.408/.5	N/A
Wash	3:41	0.0000	-0.0009	-0.0005	-0.0796	0.0008	N/A	-0.0003	0.1691	-0.0006	0.0034	N/A
QC-25 ppm	3:47	0.1977	0.1790	1.022	9.823	1.128	N/A	0.0093	1.284	0.0093	0.0199	N/A
ECV	3:52	4.67/5	4.41(5)	5.24/5	52.0/50	54.4/50	N/A	4.50/5	52.5/50	4.78/5	4.76/5	N/A
ECB	3:59	-0.003	-0.0008	0.0006	-0.0818	0.0001	N/A	0.0001	0.0989	-0.0007	0.0026	N/A
MRL	4:05	0.009/.01	0.008/.01	0.022/.02	0.918/1	0.112/.1	N/A	0.017/.02	1.11/1	0.019/.02	0.022/.02	N/A

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
-2.0	15.0	332357.3
-1.6	15.0	451683.8
-1.2	15.0	590002.7
-0.8	15.0	728483.9
-0.4	15.0	819616.4
0.0	15.0	824559.1
0.4	15.0	798705.8
0.8	15.0	673091.0
1.2	15.0	547644.5
1.6	15.0	418050.5
2.0	15.0	292380.9
0.0	10.0	6732.8
0.0	10.5	18151.5
0.0	11.0	29878.4
0.0	11.5	47226.7
0.0	12.0	74779.1
0.0	12.5	178059.6
0.0	13.0	261061.7
0.0	13.5	361162.6
0.0	14.0	500029.9
0.0	14.5	751190.4
0.0	15.0	835082.7
0.0	15.5	722693.9
0.0	16.0	673038.3
0.0	16.5	485847.3
0.0	17.0	392956.9
0.0	17.5	286593.9
0.0	18.0	218974.3
0.0	18.5	142388.0
0.0	19.0	39918.9
0.0	19.5	18136.3
0.0	20.0	7888.2
-0.8	15.0	653111.3
-0.4	15.0	716739.4
0.0	15.0	752703.9
0.4	15.0	714987.1
0.8	15.0	576721.1
0.0	13.0	234818.2
0.0	13.5	307689.7
0.0	14.0	421459.9
0.0	14.5	634824.2
0.0	15.0	728686.1
0.0	15.5	765318.3
0.0	16.0	736690.8
0.0	16.5	540067.9
0.0	17.0	383443.8

8/13/2007 18:40:13 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 765318.3 for Axial viewing

Y viewing position set to 15.5 mm having Peak intensity 765318.3 for Axial viewing

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
-2.0	15.0	293532.1
-1.6	15.0	417556.5
-1.2	15.0	529811.3
-0.8	15.0	650837.5
-0.4	15.0	705012.0
0.0	15.0	707703.6
0.4	15.0	682449.0
0.8	15.0	625134.4
1.2	15.0	517619.3
1.6	15.0	364622.4
2.0	15.0	274474.2
0.0	10.0	10554.2
0.0	10.5	26957.0
0.0	11.0	42349.4

0.0	11.5	67942.2
0.0	12.0	104486.9
0.0	12.5	226689.9
0.0	13.0	326313.0
0.0	13.5	418632.3
0.0	14.0	528396.5
0.0	14.5	774576.1
0.0	15.0	801894.0
0.0	15.5	706087.9
0.0	16.0	598171.6
0.0	16.5	386445.2
0.0	17.0	274874.4
0.0	17.5	203521.8
0.0	18.0	140139.2
0.0	18.5	86589.8
0.0	19.0	20005.3
0.0	19.5	8099.9
0.0	20.0	4760.8
-0.8	15.0	651506.6
-0.4	15.0	706677.0
0.0	15.0	711301.0
0.4	15.0	672293.4
0.8	15.0	593429.9
0.0	13.0	334592.3
0.0	13.5	446893.8
0.0	14.0	559667.7
0.0	14.5	715242.3
0.0	15.0	712820.0
0.0	15.5	667149.5
0.0	16.0	574737.2
0.0	16.5	375408.5
0.0	17.0	274975.0

8/13/2007 18:44:41 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 715242.3 for Axial viewing

Y viewing position set to 14.5 mm having Peak intensity 715242.3 for Axial viewing

Align View X Radial for analyte Mn 257.610

X-position	Y-position	Intensity
-7.0	15.0	1539.0
-6.5	15.0	2213.8
-6.0	15.0	2410.0
-5.5	15.0	2340.5
-5.0	15.0	2582.2
-4.5	15.0	3238.3
-4.0	15.0	4382.8
-3.5	15.0	6740.8
-3.0	15.0	10708.7
-2.5	15.0	17121.8
-2.0	15.0	25927.1
-1.5	15.0	40334.1
-1.0	15.0	52578.2
-0.5	15.0	57293.1
0.0	15.0	59108.2
0.5	15.0	51597.8
1.0	15.0	43508.2
1.5	15.0	35699.5
2.0	15.0	27052.6
2.5	15.0	17211.3
3.0	15.0	7497.2
3.5	15.0	6693.7
4.0	15.0	6535.8
4.5	15.0	7315.8
5.0	15.0	8057.8
5.5	15.0	8587.8
6.0	15.0	8530.0
6.5	15.0	8238.8
7.0	15.0	8443.3

8/13/2007 18:51:29 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 59108.2 for Radial viewing

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
-2.0	15.0	3197782.7
-1.6	15.0	4317417.3
-1.2	15.0	5586326.8
-0.8	15.0	6879567.3
-0.4	15.0	7624401.8
0.0	15.0	7918817.4
0.4	15.0	7493374.5
0.8	15.0	6419580.0
1.2	15.0	5138303.2
1.6	15.0	3876028.4
2.0	15.0	2816586.5
0.0	10.0	66505.8
0.0	10.5	181297.8
0.0	11.0	301322.5
0.0	11.5	471895.6
0.0	12.0	730635.0
0.0	12.5	1724967.8
0.0	13.0	2496290.9
0.0	13.5	3505596.3
0.0	14.0	4668120.5
0.0	14.5	7089557.8
0.0	15.0	7823237.9
0.0	15.5	7946207.4
0.0	16.0	7353615.8
0.0	16.5	5387275.0
0.0	17.0	4147348.1
0.0	17.5	3072424.6
0.0	18.0	2266946.0
0.0	18.5	1570025.6
0.0	19.0	447044.5
0.0	19.5	206279.5
0.0	20.0	89779.0
-0.8	15.5	7005183.8
-0.4	15.5	7822659.9
0.0	15.5	7976004.6
0.4	15.5	7440038.1
0.8	15.5	6423200.3
0.0	13.5	3584564.4
0.0	14.0	4624549.7
0.0	14.5	7052902.8
0.0	15.0	7851864.8
0.0	15.5	8029735.5
0.0	16.0	7415696.2
0.0	16.5	5395373.1
0.0	17.0	4136708.2
0.0	17.5	3082816.3

8/13/2007 18:56:18 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 8029735.5 for Axial viewing

Y viewing position set to 15.5 mm having Peak intensity 8029735.5 for Axial viewing

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
-2.0	15.0	3193376.7
-1.6	15.0	4428470.5
-1.2	15.0	5788529.9
-0.8	15.0	6953547.5
-0.4	15.0	7813627.4
0.0	15.0	8046489.0
0.4	15.0	7520803.7
0.8	15.0	6548788.2
1.2	15.0	5215173.2
1.6	15.0	3954236.4
2.0	15.0	2815685.7
0.0	10.0	116003.2

0.0	10.5	301304.7
0.0	11.0	479423.0
0.0	11.5	734465.9
0.0	12.0	1141443.6
0.0	12.5	2418607.2
0.0	13.0	3470996.7
0.0	13.5	4658538.2
0.0	14.0	5947844.0
0.0	14.5	7925697.2
0.0	15.0	8061167.4
0.0	15.5	7441636.8
0.0	16.0	6515235.5
0.0	16.5	4262231.9
0.0	17.0	3097559.4
0.0	17.5	2259432.1
0.0	18.0	1543982.2
0.0	18.5	948160.4
0.0	19.0	206470.5
0.0	19.5	89587.9
0.0	20.0	50978.6
-0.8	15.0	6994896.6
-0.4	15.0	7785201.3
0.0	15.0	7967603.9
0.4	15.0	7316751.6
0.8	15.0	6480574.0
0.0	13.0	3605385.9
0.0	13.5	4639525.4
0.0	14.0	5989381.2
0.0	14.5	7825247.4
0.0	15.0	8027273.0
0.0	15.5	7505222.5
0.0	16.0	6619169.0
0.0	16.5	4220384.9
0.0	17.0	3097707.7

8/13/2007 18:59:41 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 8027273.0 for Axial viewing

Y viewing position set to 15.0 mm having Peak intensity 8027273.0 for Axial viewing

=====
Analysis Begun

Start Time: 8/13/2007 19:04:27 Plasma On Time: 8/13/2007 17:30:51
 Logged In Analyst: Owner Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 069N-na Autosampler Model: AS-93plus

Sample Information File: C:\pe\Owner\Sample Information\070810A.sif
 Batch ID: 070810a
 Results Data Set: 070813A
 Results Library: C:\pe\Owner\Results\Results.mdb

=====
Method Loaded

Method Name: 200.7&6010_070703 Method Last Saved: 7/16/2007 14:45:32
 IEC File: 070703.iec MSF File:
 Method Description: 200.7/6010_070703

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Mo	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a
Alx	Lin, Calc Int	Peak Area	Axial	Sca	No
Bex	Lin, Calc Int	Peak Area	Axial	Sca	No

=====
 Sequence No.: 1 Autosampler Location: 0
 Sample ID: Calib Blank 1 Date Collected: 8/13/2007 19:04:28
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

=====
Nebulizer Parameters: Calib Blank 1
 Analyte Back Pressure Flow
 All 222.0 kPa 0.65 L/min

=====
Mean Data: Calib Blank 1

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc. Units
Sca	257893.2	12337.30	4.78%	100 %
Yr	260710.7	5575.44	2.14%	100 %
Ag†	1744.5	38.49	2.21%	[0.00] mg/L
Al†	40.8	48.60	119.20%	[0.00] mg/L
As†	7.8	1.13	14.57%	[0.00] mg/L
B_†	278.6	9.67	3.47%	[0.00] mg/L

Bat	-16.1	14.44	89.76%	[0.00]	mg/L
Bet	-4031.5	125.52	3.11%	[0.00]	mg/L
Cat	984.5	13.28	1.35%	[0.00]	mg/L
Cdt	77.1	0.24	0.31%	[0.00]	mg/L
Cot	-16.7	3.39	20.30%	[0.00]	mg/L
Cr	309.9	18.43	5.95%	[0.00]	mg/L
Cu	4495.3	175.79	3.91%	[0.00]	mg/L
Fet	-14.0	0.31	2.24%	[0.00]	mg/L
Kf	394.4	44.26	11.22%	[0.00]	mg/L
Mgt	-85.7	7.57	8.84%	[0.00]	mg/L
Mnt	516.7	33.44	6.47%	[0.00]	mg/L
Mot	37.2	3.57	9.58%	[0.00]	mg/L
Nat	-646.0	37.88	5.86%	[0.00]	mg/L
Nit	35.0	2.80	7.99%	[0.00]	mg/L
Pbt	-69.0	3.84	5.57%	[0.00]	mg/L
Sbt	-7.1	2.04	28.90%	[0.00]	mg/L
Set	-0.5	5.77	>999.9%	[0.00]	mg/L
Tlt	-17.5	1.58	8.98%	[0.00]	mg/L
Vt	181.5	27.81	15.32%	[0.00]	mg/L
Znt	305.0	10.59	3.47%	[0.00]	mg/L
Alxt	250.5	89.69	35.81%	[0.00]	ug/L
Bext	-4031.5	125.52	3.11%	[0.00]	ug/L

Sequence No.: 2
 Sample ID: Standard 2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 8/13/2007 19:08:14
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
 All 222.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	231067.5	99.99	0.04%	89.6 %
Yr	243482.8	3761.52	1.54%	93.4 %
Ag†	409447.5	6284.42	1.53%	[2] mg/L
Al†	49105.5	196.34	0.40%	[10] mg/L
As†	16923.6	44.14	0.26%	[10] mg/L
B_†	132246.2	524.62	0.40%	[5.02] mg/L
Ba†	817449.6	1875.07	0.23%	[10] mg/L
Be†	13149539.2	95187.34	0.72%	[4.01] mg/L
Ca†	984805.2	29780.44	3.02%	[100] mg/L
Cd†	117119.9	202.10	0.17%	[5.01] mg/L
Co†	226700.2	8.37	0.00%	[10] mg/L
Cr†	699939.4	1445.81	0.21%	[9.97] mg/L
Cu†	4273001.8	8126.61	0.19%	[10] mg/L
Fe†	18232.5	58.14	0.32%	[9.98] mg/L
K†	132506.9	1064.88	0.80%	[100] mg/L
Mg†	406259.8	11731.95	2.89%	[100] mg/L
Mn†	5440923.1	6864.14	0.13%	[10] mg/L
Mo†	148872.2	299.46	0.20%	[9.98] mg/L
Na†	336530.3	10049.57	2.99%	[100] mg/L
Ni†	217835.0	151.39	0.07%	[10] mg/L
Pb†	53504.6	72.43	0.14%	[10] mg/L
Sb†	15895.2	18.67	0.12%	[10] mg/L
Se†	15046.7	68.70	0.46%	[10] mg/L
Tl†	29791.4	0.56	0.00%	[10] mg/L
V†	1624476.5	4771.19	0.29%	[10] mg/L
Zn†	507510.8	157.30	0.03%	[10] mg/L
Alx†	1013089.6	2787.77	0.28%	[10000] ug/L
Bex†	13149539.2	95187.34	0.72%	[4010] ug/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	204700	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	4911	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	1692	0.00000	1.000000	
B_	1	Lin, Calc Int	0.0	26340	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	81740	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	3279000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	9848	0.00000	1.000000	
Cd	1	Lin, Calc Int	0.0	23380	0.00000	1.000000	
Co	1	Lin, Calc Int	-0.0	22670	0.00000	1.000000	
Cr	1	Lin, Calc Int	0.0	70200	0.00000	1.000000	
Cu	1	Lin, Calc Int	0.0	427300	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	1827	0.00000	1.000000	
K	1	Lin, Calc Int	0.0	1325	0.00000	1.000000	
Mg	1	Lin, Calc Int	-0.0	4063	0.00000	1.000000	
Mn	1	Lin, Calc Int	0.0	544100	0.00000	1.000000	
Mo	1	Lin, Calc Int	0.0	14920	0.00000	1.000000	
Na	1	Lin, Calc Int	0.0	3365	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	21780	0.00000	1.000000	
Pb	1	Lin, Calc Int	0.0	5350	0.00000	1.000000	
Sb	1	Lin, Calc Int	0.0	1590	0.00000	1.000000	
Se	1	Lin, Calc Int	0.0	1505	0.00000	1.000000	

Tl	1	Lin, Calc Int	0.0	2979	0.00000	1.000000
V	1	Lin, Calc Int	0.0	162400	0.00000	1.000000
Zn	1	Lin, Calc Int	0.0	50750	0.00000	1.000000
Alx	1	Lin, Calc Int	0.0	101.3	0.00000	1.000000
Bex	1	Lin, Calc Int	0.0	3279	0.00000	1.000000

Sequence No.: 3
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 8/13/2007 19:11:37
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
 All 222.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	228374.9	88.6 %		0.28			0.32%
Yr	241092.5	92.5 %		0.07			0.08%
Ag†	407698.1	1.99 mg/L		0.001	1.99 mg/L	0.001	0.06%
	QC value within limits for Ag	Recovery = 99.57%					
Al†	49285.9	10.0 mg/L		0.01	10.0 mg/L	0.01	0.05%
	QC value within limits for Al	Recovery = 100.37%					
As†	17365.2	10.3 mg/L		0.13	10.3 mg/L	0.13	1.23%
	QC value within limits for As	Recovery = 102.60%					
B_†	132572.1	5.01 mg/L		0.015	5.01 mg/L	0.015	0.30%
	QC value within limits for B_	Recovery = 100.24%					
Ba†	820711.8	10.0 mg/L		0.01	10.0 mg/L	0.01	0.11%
	QC value within limits for Ba	Recovery = 100.40%					
Be†	13240658.7	4.04 mg/L		0.022	4.04 mg/L	0.022	0.54%
	QC value within limits for Be	Recovery = 100.94%					
Ca†	1009464.8	103 mg/L		1.4	103 mg/L	1.4	1.34%
	QC value within limits for Ca	Recovery = 102.50%					
Cd†	117297.7	4.88 mg/L		0.004	4.88 mg/L	0.004	0.08%
	QC value within limits for Cd	Recovery = 97.56%					
Co†	227186.4	10.0 mg/L		0.02	10.0 mg/L	0.02	0.16%
	QC value within limits for Co	Recovery = 100.21%					
Cr†	699490.9	9.96 mg/L		0.000	9.96 mg/L	0.000	0.00%
	QC value within limits for Cr	Recovery = 99.64%					
Cu†	4297256.6	10.1 mg/L		0.01	10.1 mg/L	0.01	0.12%
	QC value within limits for Cu	Recovery = 100.66%					
Fe†	18349.9	10.0 mg/L		0.02	10.0 mg/L	0.02	0.16%
	QC value within limits for Fe	Recovery = 100.44%					
K†	133349.9	101 mg/L		0.1	101 mg/L	0.1	0.11%
	QC value within limits for K	Recovery = 100.64%					
Mg†	415285.6	102 mg/L		1.5	102 mg/L	1.5	1.46%
	QC value within limits for Mg	Recovery = 102.22%					
Mn†	5456247.2	10.0 mg/L		0.01	10.0 mg/L	0.01	0.13%
	QC value within limits for Mn	Recovery = 100.28%					
Mo†	149132.8	10.00 mg/L		0.003	10.00 mg/L	0.003	0.03%
	QC value within limits for Mo	Recovery = 99.97%					
Na†	345105.0	103 mg/L		0.9	103 mg/L	0.9	0.88%
	QC value within limits for Na	Recovery = 102.55%					
Ni†	217503.1	9.98 mg/L		0.009	9.98 mg/L	0.009	0.10%
	QC value within limits for Ni	Recovery = 99.85%					
Pb†	54921.7	10.3 mg/L		0.14	10.3 mg/L	0.14	1.37%
	QC value within limits for Pb	Recovery = 102.65%					
Sb†	16286.6	10.1 mg/L		0.18	10.1 mg/L	0.18	1.81%
	QC value within limits for Sb	Recovery = 100.88%					
Se†	15467.7	10.3 mg/L		0.23	10.3 mg/L	0.23	2.25%
	QC value within limits for Se	Recovery = 102.99%					
Tl†	30641.3	10.3 mg/L		0.17	10.3 mg/L	0.17	1.68%
	QC value within limits for Tl	Recovery = 103.14%					
V†	1626031.4	10.1 mg/L		0.00	10.1 mg/L	0.00	0.03%
	QC value within limits for V	Recovery = 100.65%					
Zn†	507148.6	9.92 mg/L		0.001	9.92 mg/L	0.001	0.01%
	QC value within limits for Zn	Recovery = 99.25%					
Alx†	1019710.5	10100 ug/L		5.0	10.1 mg/L	0.00	0.05%
	QC value within limits for Alx	Recovery = 100.65%					
Bex†	13240658.7	4040 ug/L		22.0	4.04 mg/L	0.022	0.54%
	QC value within limits for Bex	Recovery = 100.94%					

All analyte(s) passed QC.

Sequence No.: 4
 Sample ID: LINEARITY
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 9
 Date Collected: 8/13/2007 19:14:59
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte	Back Pressure	Flow
All	222.0 kPa	0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	227167.5	88.1	%	0.97			1.10%
Yr	230492.4	88.4	%	1.03			1.16%
Ag†	-3410.4	-0.0167	mg/L	0.00096	-0.0167 mg/L	0.00096	5.74%
	QC value within limits for Ag	Recovery = Not calculated					
Al†	-16.5	-0.00336	mg/L	0.011174	-0.00336 mg/L	0.011174	332.80%
	QC value within limits for Al	Recovery = Not calculated					
As†	-73.6	0.0175	mg/L	0.00174	0.0175 mg/L	0.00174	9.93%
	QC value within limits for As	Recovery = Not calculated					
B_†	583.3	0.0221	mg/L	0.00313	0.0221 mg/L	0.00313	14.13%
	QC value within limits for B_	Recovery = Not calculated					
Ba†	91.1	0.00111	mg/L	0.000073	0.00111 mg/L	0.000073	6.54%
	QC value within limits for Ba	Recovery = Not calculated					
Be†	-612.3	-0.00019	mg/L	0.000010	-0.00019 mg/L	0.000010	5.24%
	QC value within limits for Be	Recovery = Not calculated					
Ca†	3036801.5	308	mg/L	2.6	308 mg/L	2.6	0.83%
	QC value within limits for Ca	Recovery = 102.79%					
Cd†	-30.7	-0.00054	mg/L	0.000191	-0.00054 mg/L	0.000191	35.11%
	QC value within limits for Cd	Recovery = Not calculated					
Co†	64.3	0.00283	mg/L	0.000149	0.00283 mg/L	0.000149	5.26%
	QC value within limits for Co	Recovery = Not calculated					
Cr†	115.5	0.00165	mg/L	0.000024	0.00165 mg/L	0.000024	1.45%
	QC value within limits for Cr	Recovery = Not calculated					
Cu†	-6613.5	-0.0155	mg/L	0.00061	-0.0155 mg/L	0.00061	3.91%
	QC value within limits for Cu	Recovery = Not calculated					
Fe†	184559.1	101	mg/L	0.2	101 mg/L	0.2	0.20%
	QC value within limits for Fe	Recovery = 101.02%					
K†	419175.5	316	mg/L	3.1	316 mg/L	3.1	0.98%
	QC value within limits for K	Recovery = 105.45%					
Mg†	813310.4	200	mg/L	1.9	200 mg/L	1.9	0.96%
	QC value within limits for Mg	Recovery = Not calculated					
Mn†	2099.8	0.00386	mg/L	0.000039	0.00386 mg/L	0.000039	1.01%
	QC value within limits for Mn	Recovery = Not calculated					
Mo†	27.2	0.00182	mg/L	0.000531	0.00182 mg/L	0.000531	29.15%
	QC value within limits for Mo	Recovery = Not calculated					
Na†	1072522.5	319	mg/L	3.8	319 mg/L	3.8	1.20%
	QC value within limits for Na	Recovery = 106.23%					
Ni†	22.3	0.00102	mg/L	0.000371	0.00102 mg/L	0.000371	36.24%
	QC value within limits for Ni	Recovery = Not calculated					
Pb†	-36.6	-0.00684	mg/L	0.002190	-0.00684 mg/L	0.002190	32.04%
	QC value within limits for Pb	Recovery = Not calculated					
Sb†	20.8	0.0130	mg/L	0.00088	0.0130 mg/L	0.00088	6.76%
	QC value within limits for Sb	Recovery = Not calculated					
Se†	-497.9	-0.134	mg/L	0.0045	-0.134 mg/L	0.0045	3.35%
	QC value within limits for Se	Recovery = Not calculated					
Tl†	84.6	0.0284	mg/L	0.00096	0.0284 mg/L	0.00096	3.40%
	QC value within limits for Tl	Recovery = Not calculated					
V†	-2517.4	-0.0155	mg/L	0.00022	-0.0155 mg/L	0.00022	1.42%
	QC value within limits for V	Recovery = Not calculated					
Zn†	1470.7	0.0290	mg/L	0.00007	0.0290 mg/L	0.00007	0.25%
	QC value within limits for Zn	Recovery = Not calculated					
Alx†	618.4	6.10	ug/L	0.386	0.00610 mg/L	0.000386	6.32%
	QC value within limits for Alx	Recovery = Not calculated					
Bex†	-612.3	-0.187	ug/L	0.0098	-0.00019 mg/L	0.000010	5.24%
	QC value within limits for Bex	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 5
 Sample ID: ICSA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 8/13/2007 19:18:46
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte	Back Pressure	Flow
All	221.0 kPa	0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	236329.4	91.6 %		0.31			0.34%
Yr	236675.6	90.8 %		0.08			0.09%
Agt	-3410.9	-0.0167 mg/L		0.00011	-0.0167 mg/L	0.00011	0.67%
	QC value within limits for Ag	Recovery = Not calculated					
Alf	1237178.3	252 mg/L		3.8	252 mg/L	3.8	1.51%
	QC value within limits for Al	Recovery = 100.78%					
Ast	-328.9	-0.134 mg/L		0.0005	-0.134 mg/L	0.0005	0.36%
	QC value within limits for As	Recovery = Not calculated					
B_f	1787.7	0.0679 mg/L		0.00061	0.0679 mg/L	0.00061	0.90%
	QC value within limits for B	Recovery = Not calculated					
Baf	115.7	0.00142 mg/L		0.000098	0.00142 mg/L	0.000098	6.95%
	QC value within limits for Ba	Recovery = Not calculated					
Bef	-827.2	-0.00025 mg/L		0.000018	-0.00025 mg/L	0.000018	7.28%
	QC value within limits for Be	Recovery = Not calculated					
Car	2574452.5	261 mg/L		4.7	261 mg/L	4.7	1.82%
	QC value within limits for Ca	Recovery = 104.57%					
Cdf	-47.3	0.00137 mg/L		0.000003	0.00137 mg/L	0.000003	0.22%
	QC value within limits for Cd	Recovery = Not calculated					
Cof	38.2	0.00168 mg/L		0.000215	0.00168 mg/L	0.000215	12.74%
	QC value within limits for Co	Recovery = Not calculated					
Crt	-74.5	-0.00106 mg/L		0.000174	-0.00106 mg/L	0.000174	16.34%
	QC value within limits for Cr	Recovery = Not calculated					
Cuf	-6980.4	-0.0163 mg/L		0.00020	-0.0163 mg/L	0.00020	1.24%
	QC value within limits for Cu	Recovery = Not calculated					
Fef	183184.4	100 mg/L		0.4	100 mg/L	0.4	0.36%
	QC value within limits for Fe	Recovery = 100.27%					
Kf	516.4	0.390 mg/L		0.0422	0.390 mg/L	0.0422	10.84%
	QC value within limits for K	Recovery = Not calculated					
Mgf	1032069.2	254 mg/L		3.9	254 mg/L	3.9	1.54%
	QC value within limits for Mg	Recovery = 101.62%					
Mnf	1695.4	0.00312 mg/L		0.000007	0.00312 mg/L	0.000007	0.23%
	QC value within limits for Mn	Recovery = Not calculated					
Mof	40.4	0.00271 mg/L		0.000382	0.00271 mg/L	0.000382	14.10%
	QC value within limits for Mo	Recovery = Not calculated					
Naf	624.6	0.186 mg/L		0.0102	0.186 mg/L	0.0102	5.48%
	QC value within limits for Na	Recovery = Not calculated					
Nif	-3.1	-0.00014 mg/L		0.000068	-0.00014 mg/L	0.000068	48.28%
	QC value within limits for Ni	Recovery = Not calculated					
Pbf	-248.7	-0.0465 mg/L		0.00144	-0.0465 mg/L	0.00144	3.09%
	QC value within limits for Pb	Recovery = Not calculated					
Sbf	16.2	0.0101 mg/L		0.00368	0.0101 mg/L	0.00368	36.24%
	QC value within limits for Sb	Recovery = Not calculated					
Sef	-511.9	-0.145 mg/L		0.0008	-0.145 mg/L	0.0008	0.56%
	QC value within limits for Se	Recovery = Not calculated					
Tlf	54.5	0.0182 mg/L		0.00220	0.0182 mg/L	0.00220	12.05%
	QC value within limits for Tl	Recovery = Not calculated					
Vf	-2564.3	-0.0158 mg/L		0.00007	-0.0158 mg/L	0.00007	0.42%
	QC value within limits for V	Recovery = Not calculated					
Znf	1228.6	0.0242 mg/L		0.00017	0.0242 mg/L	0.00017	0.72%
	QC value within limits for Zn	Recovery = Not calculated					
Alxt	Saturated2						
	Unable to evaluate QC.						
Bexf	-827.2	-0.252 ug/L		0.0184	-0.00025 mg/L	0.000018	7.28%
	QC value within limits for Bex	Recovery = Not calculated					

All analyte(s) passed QC. One or more analytes were not evaluated.

Sequence No.: 6
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 11
 Date Collected: 8/13/2007 19:22:31
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte Back Pressure Flow
 All 221.0 kPa 0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	235764.7	91.4 %		2.10			2.29%
Yr	237098.5	90.9 %		0.18			0.20%
Ag†	49978.7	0.244 mg/L		0.0012	0.244 mg/L	0.0012	0.49%
	QC value less than the lower limit for Ag Recovery = 48.83%						
Al†	1187694.1	242 mg/L		1.6	242 mg/L	1.6	0.68%
	QC value within limits for Al Recovery = 96.75%						
As†	-329.1	-0.136 mg/L		0.0003	-0.136 mg/L	0.0003	0.21%
	QC value less than the lower limit for As Recovery = Not calculated						
B_†	1716.5	0.0647 mg/L		0.00018	0.0647 mg/L	0.00018	0.27%
	QC value greater than the upper limit for B Recovery = Not calculated						
Ba†	20201.5	0.247 mg/L		0.0056	0.247 mg/L	0.0056	2.28%
	QC value within limits for Ba Recovery = 98.85%						
Be†	794977.3	0.242 mg/L		0.0020	0.242 mg/L	0.0020	0.82%
	QC value within limits for Be Recovery = 96.97%						
Ca†	2463750.4	250 mg/L		1.4	250 mg/L	1.4	0.56%
	QC value within limits for Ca Recovery = 100.07%						
Cd†	10681.1	0.461 mg/L		0.0119	0.461 mg/L	0.0119	2.58%
	QC value within limits for Cd Recovery = 92.24%						
Co†	5258.1	0.232 mg/L		0.0059	0.232 mg/L	0.0059	2.54%
	QC value within limits for Co Recovery = 92.78%						
Cr†	16669.0	0.237 mg/L		0.0063	0.237 mg/L	0.0063	2.66%
	QC value within limits for Cr Recovery = 94.97%						
Cu†	99324.4	0.233 mg/L		0.0021	0.233 mg/L	0.0021	0.91%
	QC value within limits for Cu Recovery = 93.07%						
Fe†	177217.5	97.0 mg/L		0.27	97.0 mg/L	0.27	0.27%
	QC value within limits for Fe Recovery = 97.00%						
K†	223.2	0.168 mg/L		0.0340	0.168 mg/L	0.0340	20.20%
	QC value within limits for K Recovery = Not calculated						
Mg†	991205.4	244 mg/L		2.0	244 mg/L	2.0	0.84%
	QC value within limits for Mg Recovery = 97.59%						
Mn†	135800.4	0.250 mg/L		0.0015	0.250 mg/L	0.0015	0.59%
	QC value within limits for Mn Recovery = 99.84%						
Mo†	35.3	0.00237 mg/L		0.000744	0.00237 mg/L	0.000744	31.44%
	QC value within limits for Mo Recovery = Not calculated						
Na†	223.6	0.0664 mg/L		0.01899	0.0664 mg/L	0.01899	28.59%
	QC value within limits for Na Recovery = Not calculated						
Ni†	9789.0	0.449 mg/L		0.0116	0.449 mg/L	0.0116	2.59%
	QC value within limits for Ni Recovery = 89.88%						
Pb†	2354.6	0.440 mg/L		0.0114	0.440 mg/L	0.0114	2.59%
	QC value within limits for Pb Recovery = 88.01%						
Sb†	23.6	0.0110 mg/L		0.00329	0.0110 mg/L	0.00329	29.96%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-483.0	-0.132 mg/L		0.0054	-0.132 mg/L	0.0054	4.06%
	QC value less than the lower limit for Se Recovery = Not calculated						
Tl†	48.5	0.0169 mg/L		0.00014	0.0169 mg/L	0.00014	0.86%
	QC value within limits for Tl Recovery = Not calculated						
V†	36429.1	0.226 mg/L		0.0014	0.226 mg/L	0.0014	0.61%
	QC value within limits for V Recovery = 90.22%						
Zn†	26890.9	0.527 mg/L		0.0127	0.527 mg/L	0.0127	2.42%
	QC value within limits for Zn Recovery = 105.36%						
Alx†	Saturated2 Unable to evaluate QC.						
Bex†	794977.3	242 ug/L		2.0	0.242 mg/L	0.0020	0.82%
	QC value within limits for Bex Recovery = 96.97%						

QC Failed. Continue with analysis.

Sequence No.: 7
 Sample ID: Wash
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 19:26:19
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Wash

Analyte Back Pressure Flow
 All 221.0 kPa 0.65 L/min

Mean Data: Wash

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	262878.5	102	%	1.3			1.30%
Yr	251579.1	96.5	%	0.05			0.05%
Ag†	330.7	0.00162	mg/L	0.000251	0.00162 mg/L	0.000251	15.52%
	QC value within limits for Ag	Recovery = Not calculated					
Al†	-55.8	-0.0114	mg/L	0.00507	-0.0114 mg/L	0.00507	44.58%
	QC value within limits for Al	Recovery = Not calculated					
As†	-0.0	-0.00001	mg/L	0.003058	-0.00001 mg/L	0.003058	>999.9%
	QC value within limits for As	Recovery = Not calculated					
B_†	76.0	0.00288	mg/L	0.000073	0.00288 mg/L	0.000073	2.54%
	QC value within limits for B_	Recovery = Not calculated					
Ba†	-1.9	-0.00002	mg/L	0.000026	-0.00002 mg/L	0.000026	113.33%
	QC value within limits for Ba	Recovery = Not calculated					
Be†	159.9	0.00005	mg/L	0.000008	0.00005 mg/L	0.000008	15.90%
	QC value within limits for Be	Recovery = Not calculated					
Ca†	34.9	0.00355	mg/L	0.000294	0.00355 mg/L	0.000294	8.29%
	QC value within limits for Ca	Recovery = Not calculated					
Cd†	2.8	0.00012	mg/L	0.000062	0.00012 mg/L	0.000062	51.29%
	QC value within limits for Cd	Recovery = Not calculated					
Co†	3.5	0.00015	mg/L	0.000258	0.00015 mg/L	0.000258	167.49%
	QC value within limits for Co	Recovery = Not calculated					
Cr†	-5.7	-0.00008	mg/L	0.000041	-0.00008 mg/L	0.000041	50.13%
	QC value within limits for Cr	Recovery = Not calculated					
Cu†	9.7	0.00002	mg/L	0.000000	0.00002 mg/L	0.000000	0.79%
	QC value within limits for Cu	Recovery = Not calculated					
Fe†	-3.9	-0.00215	mg/L	0.000862	-0.00215 mg/L	0.000862	40.18%
	QC value within limits for Fe	Recovery = Not calculated					
K†	0.9	0.00065	mg/L	0.029362	0.00065 mg/L	0.029362	>999.9%
	QC value within limits for K	Recovery = Not calculated					
Mg†	3.5	0.00086	mg/L	0.000511	0.00086 mg/L	0.000511	59.10%
	QC value within limits for Mg	Recovery = Not calculated					
Mn†	-219.4	-0.00040	mg/L	0.000001	-0.00040 mg/L	0.000001	0.35%
	QC value within limits for Mn	Recovery = Not calculated					
Mo†	-2.6	-0.00018	mg/L	0.000072	-0.00018 mg/L	0.000072	40.60%
	QC value within limits for Mo	Recovery = Not calculated					
Na†	64.6	0.0192	mg/L	0.01090	0.0192 mg/L	0.01090	56.75%
	QC value within limits for Na	Recovery = Not calculated					
Ni†	-1.4	-0.00006	mg/L	0.000033	-0.00006 mg/L	0.000033	52.44%
	QC value within limits for Ni	Recovery = Not calculated					
Pb†	5.3	0.00099	mg/L	0.000607	0.00099 mg/L	0.000607	61.12%
	QC value within limits for Pb	Recovery = Not calculated					
Sb†	-9.3	-0.00584	mg/L	0.001869	-0.00584 mg/L	0.001869	32.01%
	QC value within limits for Sb	Recovery = Not calculated					
Se†	-0.4	-0.00025	mg/L	0.003919	-0.00025 mg/L	0.003919	>999.9%
	QC value within limits for Se	Recovery = Not calculated					
Tl†	5.7	0.00193	mg/L	0.001258	0.00193 mg/L	0.001258	65.27%
	QC value within limits for Tl	Recovery = Not calculated					
V†	-22.9	-0.00014	mg/L	0.000120	-0.00014 mg/L	0.000120	84.96%
	QC value within limits for V	Recovery = Not calculated					
Zn†	13.6	0.00027	mg/L	0.000044	0.00027 mg/L	0.000044	16.26%
	QC value within limits for Zn	Recovery = Not calculated					
Alx†	62.5	0.617	ug/L	0.1913	0.00062 mg/L	0.000191	31.03%
	QC value within limits for Alx	Recovery = Not calculated					
Bex†	159.9	0.0488	ug/L	0.00775	0.00005 mg/L	0.000008	15.90%
	QC value within limits for Bex	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 8
 Sample ID: QC-25 lppm
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 12
 Date Collected: 8/13/2007 19:29:46
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QC-25 lppm

Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: QC-25 lppm

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	268220.7	104 %	2.0			1.93%
Yr	261327.0	100 %	1.1			1.10%
Ag†	172908.2	0.845 mg/L	0.0170	0.845 mg/L	0.0170	2.01%
	QC value less than the lower limit for Ag Recovery = 84.46%					
Al†	4880.3	0.994 mg/L	0.0089	0.994 mg/L	0.0089	0.90%
	QC value within limits for Al Recovery = 99.38%					
As†	1585.4	0.937 mg/L	0.0216	0.937 mg/L	0.0216	2.31%
	QC value within limits for As Recovery = 93.68%					
B_†	23460.0	0.889 mg/L	0.0181	0.889 mg/L	0.0181	2.04%
	QC value less than the lower limit for B_ Recovery = 88.85%					
Ba†	81857.9	1.00 mg/L	0.022	1.00 mg/L	0.022	2.24%
	QC value within limits for Ba Recovery = 100.14%					
Be†	3054853.8	0.932 mg/L	0.0001	0.932 mg/L	0.0001	0.01%
	QC value within limits for Be Recovery = 93.16%					
Ca†	10069.9	1.02 mg/L	0.014	1.02 mg/L	0.014	1.37%
	QC value within limits for Ca Recovery = 102.25%					
Cd†	21431.1	0.904 mg/L	0.0141	0.904 mg/L	0.0141	1.56%
	QC value within limits for Cd Recovery = 90.43%					
Co†	22642.7	0.999 mg/L	0.0207	0.999 mg/L	0.0207	2.08%
	QC value within limits for Co Recovery = 99.88%					
Cr†	67063.3	0.955 mg/L	0.0218	0.955 mg/L	0.0218	2.28%
	QC value within limits for Cr Recovery = 95.53%					
Cu†	400073.9	0.937 mg/L	0.0173	0.937 mg/L	0.0173	1.85%
	QC value within limits for Cu Recovery = 93.72%					
Fe†	1832.9	1.00 mg/L	0.012	1.00 mg/L	0.012	1.20%
	QC value within limits for Fe Recovery = 100.33%					
K†	12503.3	9.44 mg/L	0.035	9.44 mg/L	0.035	0.37%
	QC value within limits for K Recovery = 94.36%					
Mg†	4332.7	1.07 mg/L	0.011	1.07 mg/L	0.011	1.02%
	QC value within limits for Mg Recovery = 106.65%					
Mn†	539886.6	0.992 mg/L	0.0001	0.992 mg/L	0.0001	0.01%
	QC value within limits for Mn Recovery = 99.23%					
Mo†	13624.7	0.913 mg/L	0.0178	0.913 mg/L	0.0178	1.94%
	QC value within limits for Mo Recovery = 91.34%					
Na†	3550.7	1.06 mg/L	0.010	1.06 mg/L	0.010	0.98%
	QC value within limits for Na Recovery = 105.51%					
Ni†	21844.8	1.00 mg/L	0.020	1.00 mg/L	0.020	2.04%
	QC value within limits for Ni Recovery = 100.28%					
Pb†	5435.4	1.02 mg/L	0.023	1.02 mg/L	0.023	2.22%
	QC value within limits for Pb Recovery = 101.59%					
Sb†	1496.6	0.926 mg/L	0.0192	0.926 mg/L	0.0192	2.07%
	QC value within limits for Sb Recovery = 92.62%					
Se†	1363.4	0.908 mg/L	0.0216	0.908 mg/L	0.0216	2.37%
	QC value within limits for Se Recovery = 90.80%					
Tl†	3034.3	1.02 mg/L	0.019	1.02 mg/L	0.019	1.87%
	QC value within limits for Tl Recovery = 102.11%					
V†	149337.9	0.925 mg/L	0.0213	0.925 mg/L	0.0213	2.30%
	QC value within limits for V Recovery = 92.46%					
Zn†	49349.7	0.966 mg/L	0.0180	0.966 mg/L	0.0180	1.86%
	QC value within limits for Zn Recovery = 96.56%					
Alx†	94390.7	932 ug/L	9.1	0.932 mg/L	0.0091	0.98%
	QC value within limits for Alx Recovery = 93.17%					
Bex†	3054853.8	932 ug/L	0.1	0.932 mg/L	0.0001	0.01%
	QC value within limits for Bex Recovery = 93.16%					

QC Failed. Retry.

Sequence No.: 9
 Sample ID: QC-25 ppm
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 12
 Date Collected: 8/13/2007 19:32:37
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QC-25 ppm
 Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: QC-25 ppm

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	263786.8	102 %		0.3			0.31%
Yr	258515.4	99.2 %		0.99			1.00%
Ag†	172841.0	0.844 mg/L		0.0014	0.844 mg/L	0.0014	0.17%
	QC value less than the lower limit for Ag Recovery = 84.43%						
Al†	4946.9	1.01 mg/L		0.014	1.01 mg/L	0.014	1.40%
	QC value within limits for Al Recovery = 100.74%						
As†	1606.9	0.949 mg/L		0.0025	0.949 mg/L	0.0025	0.27%
	QC value within limits for As Recovery = 94.95%						
B_†	23569.8	0.893 mg/L		0.0019	0.893 mg/L	0.0019	0.21%
	QC value less than the lower limit for B_ Recovery = 89.27%						
Ba†	82053.7	1.00 mg/L		0.007	1.00 mg/L	0.007	0.69%
	QC value within limits for Ba Recovery = 100.38%						
Be†	3084809.1	0.941 mg/L		0.0101	0.941 mg/L	0.0101	1.07%
	QC value within limits for Be Recovery = 94.07%						
Ca†	10202.0	1.04 mg/L		0.015	1.04 mg/L	0.015	1.44%
	QC value within limits for Ca Recovery = 103.59%						
Cd†	21331.7	0.900 mg/L		0.0010	0.900 mg/L	0.0010	0.11%
	QC value less than the lower limit for Cd Recovery = 89.98%						
Co†	22598.1	0.997 mg/L		0.0007	0.997 mg/L	0.0007	0.07%
	QC value within limits for Co Recovery = 99.68%						
Cr†	67209.7	0.957 mg/L		0.0052	0.957 mg/L	0.0052	0.54%
	QC value within limits for Cr Recovery = 95.73%						
Cu†	397370.5	0.931 mg/L		0.0061	0.931 mg/L	0.0061	0.66%
	QC value within limits for Cu Recovery = 93.09%						
Fe†	1851.9	1.01 mg/L		0.014	1.01 mg/L	0.014	1.36%
	QC value within limits for Fe Recovery = 101.37%						
K†	12656.8	9.55 mg/L		0.036	9.55 mg/L	0.036	0.37%
	QC value within limits for K Recovery = 95.52%						
Mg†	4378.4	1.08 mg/L		0.013	1.08 mg/L	0.013	1.19%
	QC value within limits for Mg Recovery = 107.77%						
Mn†	543579.0	0.999 mg/L		0.0126	0.999 mg/L	0.0126	1.26%
	QC value within limits for Mn Recovery = 99.91%						
Mo†	13648.6	0.915 mg/L		0.0045	0.915 mg/L	0.0045	0.49%
	QC value within limits for Mo Recovery = 91.50%						
Na†	3539.4	1.05 mg/L		0.001	1.05 mg/L	0.001	0.09%
	QC value within limits for Na Recovery = 105.17%						
Ni†	21897.1	1.01 mg/L		0.008	1.01 mg/L	0.008	0.76%
	QC value within limits for Ni Recovery = 100.52%						
Pb†	5521.6	1.03 mg/L		0.002	1.03 mg/L	0.002	0.23%
	QC value within limits for Pb Recovery = 103.20%						
Sb†	1508.0	0.933 mg/L		0.0008	0.933 mg/L	0.0008	0.09%
	QC value within limits for Sb Recovery = 93.33%						
Se†	1374.0	0.915 mg/L		0.0040	0.915 mg/L	0.0040	0.43%
	QC value within limits for Se Recovery = 91.51%						
Tl†	3074.9	1.03 mg/L		0.004	1.03 mg/L	0.004	0.40%
	QC value within limits for Tl Recovery = 103.47%						
V†	149546.4	0.926 mg/L		0.0073	0.926 mg/L	0.0073	0.79%
	QC value within limits for V Recovery = 92.59%						
Zn†	49530.6	0.969 mg/L		0.0056	0.969 mg/L	0.0056	0.58%
	QC value within limits for Zn Recovery = 96.91%						
Alx†	93592.3	924 ug/L		8.9	0.924 mg/L	0.0089	0.96%
	QC value within limits for Alx Recovery = 92.38%						
Sex†	3084809.1	941 ug/L		10.1	0.941 mg/L	0.0101	1.07%

QC value within limits for Bex Recovery = 94.07%
QC Failed. Continue with analysis.

Sequence No.: 10
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/13/2007 19:37:00
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	247239.1	95.9 %		0.77			0.80%
Yr	247567.2	95.0 %		0.70			0.73%
Ag†	145530.1	0.711 mg/L		0.1118	0.711 mg/L	0.1118	15.73%
	QC value within limits for Ag Recovery = 71.09%						
Al†	25003.1	5.09 mg/L		0.068	5.09 mg/L	0.068	1.33%
	QC value within limits for Al Recovery = 101.83%						
As†	8342.1	4.93 mg/L		0.006	4.93 mg/L	0.006	0.11%
	QC value within limits for As Recovery = 98.58%						
B_†	63742.3	2.41 mg/L		0.005	2.41 mg/L	0.005	0.21%
	QC value within limits for B_ Recovery = 96.38%						
Bar	406329.2	4.97 mg/L		0.001	4.97 mg/L	0.001	0.03%
	QC value within limits for Ba Recovery = 99.41%						
Be†	6525437.7	1.99 mg/L		0.030	1.99 mg/L	0.030	1.53%
	QC value within limits for Be Recovery = 99.50%						
Ca†	512692.2	52.1 mg/L		0.08	52.1 mg/L	0.08	0.15%
	QC value within limits for Ca Recovery = 104.12%						
Cd†	59095.1	2.46 mg/L		0.003	2.46 mg/L	0.003	0.11%
	QC value within limits for Cd Recovery = 98.46%						
Co†	113684.8	5.01 mg/L		0.002	5.01 mg/L	0.002	0.04%
	QC value within limits for Co Recovery = 100.30%						
Cr†	345042.4	4.91 mg/L		0.017	4.91 mg/L	0.017	0.35%
	QC value within limits for Cr Recovery = 98.30%						
Cu†	2035223.5	4.77 mg/L		0.013	4.77 mg/L	0.013	0.27%
	QC value within limits for Cu Recovery = 95.35%						
Fe†	9424.1	5.16 mg/L		0.107	5.16 mg/L	0.107	2.07%
	QC value within limits for Fe Recovery = 103.17%						
K†	65880.4	49.7 mg/L		0.62	49.7 mg/L	0.62	1.24%
	QC value within limits for K Recovery = 99.44%						
Mg†	212996.4	52.4 mg/L		0.02	52.4 mg/L	0.02	0.04%
	QC value within limits for Mg Recovery = 104.86%						
Mn†	2728672.7	5.02 mg/L		0.004	5.02 mg/L	0.004	0.09%
	QC value within limits for Mn Recovery = 100.30%						
Mo†	72772.6	4.88 mg/L		0.015	4.88 mg/L	0.015	0.31%
	QC value within limits for Mo Recovery = 97.57%						
Na†	173915.9	51.7 mg/L		0.10	51.7 mg/L	0.10	0.19%
	QC value within limits for Na Recovery = 103.36%						
Ni†	110142.9	5.06 mg/L		0.021	5.06 mg/L	0.021	0.42%
	QC value within limits for Ni Recovery = 101.13%						
Pb†	27391.7	5.12 mg/L		0.004	5.12 mg/L	0.004	0.07%
	QC value within limits for Pb Recovery = 102.39%						
Sb†	7950.1	4.92 mg/L		0.010	4.92 mg/L	0.010	0.20%
	QC value within limits for Sb Recovery = 98.47%						
Se†	7602.8	5.06 mg/L		0.025	5.06 mg/L	0.025	0.49%
	QC value within limits for Se Recovery = 101.26%						
Tl†	15539.0	5.23 mg/L		0.001	5.23 mg/L	0.001	0.02%
	QC value within limits for Tl Recovery = 104.59%						
V†	795964.4	4.93 mg/L		0.011	4.93 mg/L	0.011	0.22%
	QC value within limits for V Recovery = 98.54%						
Zn†	263142.8	5.15 mg/L		0.018	5.15 mg/L	0.018	0.35%
	QC value within limits for Zn Recovery = 103.01%						
Alx†	491360.1	4850 ug/L		29.9	4.85 mg/L	0.030	0.62%
	QC value within limits for Alx Recovery = 97.00%						
Bex†	6525437.7	1990 ug/L		30.4	1.99 mg/L	0.030	1.53%
	QC value within limits for Bex Recovery = 99.50%						

All analyte(s) passed QC.

Sequence No.: 11
 Sample ID: ICB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 19:40:20
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	261637.5	101	%	0.3			0.32%
Yr	255025.6	97.8	%	0.31			0.31%
Ag†	3844.9	0.0188	mg/L	0.00031	0.0188 mg/L	0.00031	1.64%
	QC value greater than the upper limit for Ag Recovery = Not calculated						
Al†	-1.3	-0.00026	mg/L	0.007468	-0.00026 mg/L	0.007468	>999.9%
	QC value within limits for Al Recovery = Not calculated						
As†	13.5	0.00799	mg/L	0.002443	0.00799 mg/L	0.002443	30.59%
	QC value within limits for As Recovery = Not calculated						
B_†	427.6	0.0162	mg/L	0.00149	0.0162 mg/L	0.00149	9.20%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-10.2	-0.00013	mg/L	0.000104	-0.00013 mg/L	0.000104	83.42%
	QC value within limits for Ba Recovery = Not calculated						
Be†	110.0	0.00003	mg/L	0.000039	0.00003 mg/L	0.000039	115.33%
	QC value within limits for Be Recovery = Not calculated						
Ca†	12.7	0.00129	mg/L	0.000868	0.00129 mg/L	0.000868	67.27%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	15.8	0.00054	mg/L	0.000031	0.00054 mg/L	0.000031	5.81%
	QC value within limits for Cd Recovery = Not calculated						
Co†	-2.1	-0.00009	mg/L	0.000368	-0.00009 mg/L	0.000368	400.35%
	QC value within limits for Co Recovery = Not calculated						
Cr†	0.7	0.00001	mg/L	0.000024	0.00001 mg/L	0.000024	242.33%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	172.8	0.00040	mg/L	0.000218	0.00040 mg/L	0.000218	53.86%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-0.6	-0.00035	mg/L	0.000325	-0.00035 mg/L	0.000325	93.85%
	QC value within limits for Fe Recovery = Not calculated						
K†	-16.1	-0.0122	mg/L	0.01473	-0.0122 mg/L	0.01473	121.20%
	QC value within limits for K Recovery = Not calculated						
Mg†	6.3	0.00156	mg/L	0.000058	0.00156 mg/L	0.000058	3.74%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-206.7	-0.00038	mg/L	0.000007	-0.00038 mg/L	0.000007	1.81%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	14.0	0.00094	mg/L	0.000277	0.00094 mg/L	0.000277	29.64%
	QC value within limits for Mo Recovery = Not calculated						
Na†	135.6	0.0403	mg/L	0.02568	0.0403 mg/L	0.02568	63.72%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-6.0	-0.00027	mg/L	0.000123	-0.00027 mg/L	0.000123	45.01%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	11.2	0.00209	mg/L	0.000530	0.00209 mg/L	0.000530	25.33%
	QC value within limits for Pb Recovery = Not calculated						
Sb†	0.9	0.00056	mg/L	0.002258	0.00056 mg/L	0.002258	405.01%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-1.1	-0.00076	mg/L	0.001811	-0.00076 mg/L	0.001811	239.29%
	QC value within limits for Se Recovery = Not calculated						
Tl†	5.2	0.00176	mg/L	0.000094	0.00176 mg/L	0.000094	5.34%
	QC value within limits for Tl Recovery = Not calculated						
V†	-18.1	-0.00011	mg/L	0.000110	-0.00011 mg/L	0.000110	99.09%
	QC value within limits for V Recovery = Not calculated						
Zn†	43.0	0.00085	mg/L	0.000092	0.00085 mg/L	0.000092	10.80%
	QC value within limits for Zn Recovery = Not calculated						
Alx†	109.8	1.08	ug/L	0.991	0.00108 mg/L	0.000991	91.42%
	QC value within limits for Alx Recovery = Not calculated						
Bex†	110.0	0.0336	ug/L	0.03870	0.00003 mg/L	0.000039	115.33%
	QC value within limits for Bex Recovery = Not calculated						

QC Failed. Retry.

Sequence No.: 12
 Sample ID: ICB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 19:42:55
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
 All 221.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	264330.7	102 %		1.4			1.37%
Yr	254016.5	97.4 %		0.03			0.03%
Ag†	3111.8	0.0152 mg/L		0.00032	0.0152 mg/L	0.00032	2.09%
	QC value greater than the upper limit for Ag Recovery = Not calculated						
Al†	22.9	0.00467 mg/L		0.003426	0.00467 mg/L	0.003426	73.37%
	QC value within limits for Al Recovery = Not calculated						
As†	3.9	0.00229 mg/L		0.001814	0.00229 mg/L	0.001814	79.28%
	QC value within limits for As Recovery = Not calculated						
B_†	243.7	0.00925 mg/L		0.000818	0.00925 mg/L	0.000818	8.84%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-9.8	-0.00012 mg/L		0.000026	-0.00012 mg/L	0.000026	21.82%
	QC value within limits for Ba Recovery = Not calculated						
Be†	86.9	0.00003 mg/L		0.000018	0.00003 mg/L	0.000018	66.42%
	QC value within limits for Be Recovery = Not calculated						
Ca†	30.5	0.00310 mg/L		0.001607	0.00310 mg/L	0.001607	51.90%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	6.1	0.00022 mg/L		0.000199	0.00022 mg/L	0.000199	89.24%
	QC value within limits for Cd Recovery = Not calculated						
Co†	2.1	0.00009 mg/L		0.000001	0.00009 mg/L	0.000001	0.60%
	QC value within limits for Co Recovery = Not calculated						
Cr†	-8.5	-0.00012 mg/L		0.000113	-0.00012 mg/L	0.000113	93.53%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	61.7	0.00014 mg/L		0.000355	0.00014 mg/L	0.000355	245.65%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-1.0	-0.00055 mg/L		0.001017	-0.00055 mg/L	0.001017	184.80%
	QC value within limits for Fe Recovery = Not calculated						
K†	-68.3	-0.0516 mg/L		0.00375	-0.0516 mg/L	0.00375	7.28%
	QC value within limits for K Recovery = Not calculated						
Mg†	2.5	0.00060 mg/L		0.000565	0.00060 mg/L	0.000565	93.49%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-236.6	-0.00043 mg/L		0.000013	-0.00043 mg/L	0.000013	2.94%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	5.7	0.00038 mg/L		0.000179	0.00038 mg/L	0.000179	46.73%
	QC value within limits for Mo Recovery = Not calculated						
Na†	39.5	0.0117 mg/L		0.01933	0.0117 mg/L	0.01933	164.53%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-2.5	-0.00011 mg/L		0.000088	-0.00011 mg/L	0.000088	77.83%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	14.5	0.00271 mg/L		0.000425	0.00271 mg/L	0.000425	15.71%
	QC value within limits for Pb Recovery = Not calculated						
Sb†	-5.7	-0.00358 mg/L		0.000335	-0.00358 mg/L	0.000335	9.36%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-6.1	-0.00404 mg/L		0.001636	-0.00404 mg/L	0.001636	40.53%
	QC value within limits for Se Recovery = Not calculated						
Tl†	7.8	0.00262 mg/L		0.000248	0.00262 mg/L	0.000248	9.47%
	QC value within limits for Tl Recovery = Not calculated						
V†	-28.4	-0.00018 mg/L		0.000190	-0.00018 mg/L	0.000190	108.38%
	QC value within limits for V Recovery = Not calculated						
Zn†	14.9	0.00029 mg/L		0.000332	0.00029 mg/L	0.000332	112.83%
	QC value within limits for Zn Recovery = Not calculated						
Alx†	25.8	0.255 ug/L		0.0912	0.00025 mg/L	0.000091	35.79%
	QC value within limits for Alx Recovery = Not calculated						
Bex†	86.9	0.0265 ug/L		0.01761	0.00003 mg/L	0.000018	66.42%

QC value within limits for Bex Recovery = Not calculated
QC Failed. Retry.

Sequence No.: 13

Sample ID: ICB

Analyst:

Initial Sample Wt:

Dilution:

Autosampler Location: 0

Date Collected: 8/13/2007 19:45:33

Data Type: Original

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
All 220.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	267577.8	104 %		0.4			0.41%
Yr	253537.2	97.2 %		0.51			0.52%
Ag†	2423.5	0.0118 mg/L		0.00087	0.0118 mg/L	0.00087	7.34%
	QC value greater than the upper limit for Ag Recovery = Not calculated						
Al†	28.5	0.00581 mg/L		0.003644	0.00581 mg/L	0.003644	62.73%
	QC value within limits for Al Recovery = Not calculated						
As†	0.4	0.00023 mg/L		0.004127	0.00023 mg/L	0.004127	>999.9%
	QC value within limits for As Recovery = Not calculated						
B_†	166.2	0.00631 mg/L		0.000262	0.00631 mg/L	0.000262	4.15%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-6.8	-0.00008 mg/L		0.000025	-0.00008 mg/L	0.000025	30.37%
	QC value within limits for Ba Recovery = Not calculated						
Be†	210.7	0.00006 mg/L		0.000012	0.00006 mg/L	0.000012	19.06%
	QC value within limits for Be Recovery = Not calculated						
Ca†	24.1	0.00244 mg/L		0.000027	0.00244 mg/L	0.000027	1.09%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	1.7	0.00007 mg/L		0.000076	0.00007 mg/L	0.000076	109.36%
	QC value within limits for Cd Recovery = Not calculated						
Co†	-3.8	-0.00017 mg/L		0.000216	-0.00017 mg/L	0.000216	130.12%
	QC value within limits for Co Recovery = Not calculated						
Cr†	-4.4	-0.00006 mg/L		0.000040	-0.00006 mg/L	0.000040	62.53%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	8.7	0.00002 mg/L		0.000087	0.00002 mg/L	0.000087	429.06%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-0.5	-0.00028 mg/L		0.001366	-0.00028 mg/L	0.001366	482.81%
	QC value within limits for Fe Recovery = Not calculated						
K†	-57.0	-0.0430 mg/L		0.04269	-0.0430 mg/L	0.04269	99.25%
	QC value within limits for K Recovery = Not calculated						
Mg†	1.7	0.00043 mg/L		0.000068	0.00043 mg/L	0.000068	15.85%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-250.3	-0.00046 mg/L		0.000002	-0.00046 mg/L	0.000002	0.35%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	3.5	0.00024 mg/L		0.000021	0.00024 mg/L	0.000021	8.82%
	QC value within limits for Mo Recovery = Not calculated						
Na†	9.6	0.00285 mg/L		0.006064	0.00285 mg/L	0.006064	212.43%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-5.9	-0.00027 mg/L		0.000215	-0.00027 mg/L	0.000215	78.77%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	7.8	0.00147 mg/L		0.000617	0.00147 mg/L	0.000617	42.06%
	QC value within limits for Pb Recovery = Not calculated						
Sb†	1.8	0.00115 mg/L		0.002330	0.00115 mg/L	0.002330	202.13%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-1.1	-0.00071 mg/L		0.004101	-0.00071 mg/L	0.004101	573.66%
	QC value within limits for Se Recovery = Not calculated						
Tl†	6.5	0.00218 mg/L		0.000825	0.00218 mg/L	0.000825	37.88%
	QC value within limits for Tl Recovery = Not calculated						
V†	3.6	0.00002 mg/L		0.000246	0.00002 mg/L	0.000246	>999.9%
	QC value within limits for V Recovery = Not calculated						
Zn†	3.6	0.00007 mg/L		0.000098	0.00007 mg/L	0.000098	136.31%
	QC value within limits for Zn Recovery = Not calculated						
Alx†	29.4	0.290 ug/L		0.5995	0.00029 mg/L	0.000600	206.73%
	QC value within limits for Alx Recovery = Not calculated						

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	263475.8	102 %		0.4			0.41%
Yr	254576.6	97.6 %		0.43			0.44%
B ₋ t	1294.0	0.0490 mg/L		0.00018	0.0490 mg/L	0.00018	0.38%
	QC value within limits for B ₋ Recovery = 98.03%						
Ba†	1617.0	0.0198 mg/L		0.00008	0.0198 mg/L	0.00008	0.38%
	QC value within limits for Ba Recovery = 98.91%						
Be†	3362.9	0.00103 mg/L		0.000017	0.00103 mg/L	0.000017	1.70%
	QC value within limits for Be Recovery = 102.55%						
Ca†	10207.1	1.04 mg/L		0.001	1.04 mg/L	0.001	0.05%
	QC value within limits for Ca Recovery = 103.65%						
Cd†	190.9	0.00836 mg/L		0.000090	0.00836 mg/L	0.000090	1.07%
	QC value greater than the upper limit for Cd Recovery = 167.20%						
Co†	1133.0	0.0500 mg/L		0.00012	0.0500 mg/L	0.00012	0.24%
	QC value within limits for Co Recovery = 99.95%						
Cr†	695.0	0.00976 mg/L		0.000169	0.00976 mg/L	0.000169	1.73%
	QC value within limits for Cr Recovery = 97.57%						
Cu†	4167.3	0.00980 mg/L		0.000149	0.00980 mg/L	0.000149	1.52%
	QC value within limits for Cu Recovery = 98.00%						
Fe†	38.8	0.0213 mg/L		0.00114	0.0213 mg/L	0.00114	5.35%
	QC value within limits for Fe Recovery = 106.27%						
K†	1195.2	0.902 mg/L		0.0532	0.902 mg/L	0.0532	5.90%
	QC value within limits for K Recovery = 90.20%						
Mg†	436.7	0.107 mg/L		0.0008	0.107 mg/L	0.0008	0.76%
	QC value within limits for Mg Recovery = 107.50%						
Mo†	281.9	0.0189 mg/L		0.00020	0.0189 mg/L	0.00020	1.04%
	QC value within limits for Mo Recovery = 94.49%						
Na†	3486.0	1.04 mg/L		0.012	1.04 mg/L	0.012	1.12%
	QC value within limits for Na Recovery = 103.59%						
Ni†	443.5	0.0204 mg/L		0.00008	0.0204 mg/L	0.00008	0.38%
	QC value within limits for Ni Recovery = 101.79%						
Pb†	118.4	0.0221 mg/L		0.00128	0.0221 mg/L	0.00128	5.77%
	QC value within limits for Pb Recovery = 110.65%						
Tl†	324.8	0.109 mg/L		0.0007	0.109 mg/L	0.0007	0.60%
	QC value within limits for Tl Recovery = 109.04%						
V†	291.2	0.00185 mg/L		0.000096	0.00185 mg/L	0.000096	5.21%
	QC value within limits for V Recovery = 92.31%						
Zn†	895.7	0.0175 mg/L		0.00010	0.0175 mg/L	0.00010	0.58%
	QC value within limits for Zn Recovery = 87.55%						
QC Failed. Continue with analysis.							

Sequence No.: 17
 Sample ID: FILTER_CHECK
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 16
 Date Collected: 8/13/2007 19:57:52
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: FILTER_CHECK

Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: FILTER_CHECK

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	264917.0	103	%	1.0			0.98%
Yr	254463.6	97.6	%	1.18			1.21%
B_f	37.5	0.00142	mg/L	0.000264	0.00142	mg/L	0.000264 18.54%
Ba†	-10.2	-0.00012	mg/L	0.000002	-0.00012	mg/L	0.000002 1.66%
Be†	8.3	0.00000	mg/L	0.000016	0.00000	mg/L	0.000016 649.81%
Ca†	-94.8	-0.00962	mg/L	0.001093	-0.00962	mg/L	0.001093 11.35%
Cd†	-4.3	-0.00018	mg/L	0.000030	-0.00018	mg/L	0.000030 16.01%
Co†	-0.6	-0.00002	mg/L	0.000266	-0.00002	mg/L	0.000266 >999.9%
Cr†	-13.3	-0.00019	mg/L	0.000209	-0.00019	mg/L	0.000209 110.50%
Cu†	-43.0	-0.00010	mg/L	0.000069	-0.00010	mg/L	0.000069 68.62%
Fe†	-1.9	-0.00107	mg/L	0.001705	-0.00107	mg/L	0.001705 159.79%
K†	-85.6	-0.0646	mg/L	0.00869	-0.0646	mg/L	0.00869 13.44%
Mg†	0.2	0.00005	mg/L	0.001443	0.00005	mg/L	0.001443 >999.9%
Mn†	-4.6	-0.00031	mg/L	0.000299	-0.00031	mg/L	0.000299 96.14%
Nat	12.8	0.00380	mg/L	0.012744	0.00380	mg/L	0.012744 335.32%
Ni†	-0.9	-0.00004	mg/L	0.000056	-0.00004	mg/L	0.000056 135.80%
Pb†	9.7	0.00181	mg/L	0.000826	0.00181	mg/L	0.000826 45.52%
Tl†	2.8	0.00094	mg/L	0.002544	0.00094	mg/L	0.002544 270.58%
V†	12.8	0.00008	mg/L	0.000019	0.00008	mg/L	0.000019 24.25%
Zn†	-145.5	-0.00287	mg/L	0.000062	-0.00287	mg/L	0.000062 2.18%

see run

Sequence No.: 18
Sample ID: MRL6010
Analyst: Walter Hsieh
Initial Sample Wt:
Dilution: 1X

Autosampler Location: 23
Date Collected: 8/13/2007 20:01:32
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: MRL6010

Analyte Back Pressure Flow
All 220.0 kPa 0.65 L/min

Mean Data: MRL6010

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	395035.5	153 %	%	60.1			39.25%
Yr	276954.9	106 %	%	92.2			86.78%
B_t	-86.2	-0.00327 mg/L	mg/L	0.002881	-0.00327 mg/L	0.002881	88.05%
Bat	3.2	0.00004 mg/L	mg/L	0.000021	0.00004 mg/L	0.000021	53.92%
Bef	3019.4	0.00092 mg/L	mg/L	0.000129	0.00092 mg/L	0.000129	13.97%
Ca†	-92.3	-0.00938 mg/L	mg/L	0.012005	-0.00938 mg/L	0.012005	128.05%
Cdt	-26.3	-0.00113 mg/L	mg/L	0.000805	-0.00113 mg/L	0.000805	71.46%
Cof	6.1	0.00027 mg/L	mg/L	0.000221	0.00027 mg/L	0.000221	82.12%
Crf	-186.8	-0.00266 mg/L	mg/L	0.000566	-0.00266 mg/L	0.000566	21.28%
Cut	-2982.4	-0.00698 mg/L	mg/L	0.001434	-0.00698 mg/L	0.001434	20.55%
Fet	-5.9	-0.00322 mg/L	mg/L	0.011821	-0.00322 mg/L	0.011821	367.08%
K†	-38.2	-0.0288 mg/L	mg/L	0.23262	-0.0288 mg/L	0.23262	807.52%
Mgt	-1.7	-0.00043 mg/L	mg/L	0.018422	-0.00043 mg/L	0.018422	>999.9%
Mot	-16.7	-0.00112 mg/L	mg/L	0.000498	-0.00112 mg/L	0.000498	44.60%
Nat	-1.4	-0.00042 mg/L	mg/L	0.128660	-0.00042 mg/L	0.128660	>999.9%
Nit	-6.0	-0.00027 mg/L	mg/L	0.000363	-0.00027 mg/L	0.000363	132.35%
Pbt	38.9	0.00727 mg/L	mg/L	0.002414	0.00727 mg/L	0.002414	33.21%
Tlt	6.2	0.00209 mg/L	mg/L	0.002305	0.00209 mg/L	0.002305	110.13%
V†	-83.7	-0.00053 mg/L	mg/L	0.000446	-0.00053 mg/L	0.000446	84.27%
Znt	-198.7	-0.00391 mg/L	mg/L	0.000724	-0.00391 mg/L	0.000724	18.51%

Sequence No.: 15
 Sample ID: MRL6010
 Analyst: Walter Esiet
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 24
 Date Collected: 8/14/2007 09:10:38
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL6010

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: MRL6010

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	631584.6	105 %		1.9			1.84%
Yr	370573.9	104 %		1.7			1.62%
Agf	2817.4	0.0102 mg/L		0.00004	0.0102 mg/L	0.00004	0.42%
Alf	295.9	0.0630 mg/L		0.00334	0.0630 mg/L	0.00334	5.30%
Asf	219.1	0.0968 mg/L		0.00138	0.0968 mg/L	0.00138	1.42%
B_f	1695.1	0.0528 mg/L		0.00050	0.0528 mg/L	0.00050	0.94%
Bat	1517.5	0.0207 mg/L		0.00024	0.0207 mg/L	0.00024	1.16%
Bef	3286.1	0.00112 mg/L		0.00013	0.00112 mg/L	0.00013	1.17%
Caf	10009.6	1.08 mg/L		0.001	1.08 mg/L	0.001	0.05%
Cof	170.8	0.00463 mg/L		0.000023	0.00463 mg/L	0.000023	0.49%
Cct	1446.6	0.0533 mg/L		0.00073	0.0533 mg/L	0.00073	1.37%
Crt	757.8	0.00994 mg/L		0.000101	0.00994 mg/L	0.000101	1.01%
Cuf	3947.0	0.0107 mg/L		0.00006	0.0107 mg/L	0.00006	0.55%
Fef	37.5	0.0231 mg/L		0.00074	0.0231 mg/L	0.00074	3.21%
Kf	1321.2	1.01 mg/L		0.021	1.01 mg/L	0.021	2.07%
Mgf	481.9	0.108 mg/L		0.0020	0.108 mg/L	0.0020	1.87%
Mnf	1238.9	0.00225 mg/L		0.000037	0.00225 mg/L	0.000037	1.63%
Mcf	263.9	0.0199 mg/L		0.00056	0.0199 mg/L	0.00056	2.81%
Naf	3160.7	1.06 mg/L		0.011	1.06 mg/L	0.011	1.04%
Nif	479.6	0.0213 mg/L		0.00029	0.0213 mg/L	0.00029	1.34%
Pcf	111.7	0.0228 mg/L		0.00164	0.0228 mg/L	0.00164	7.20%
Scf	92.6	0.0445 mg/L		0.00018	0.0445 mg/L	0.00018	0.39%
Scf	145.3	0.103 mg/L		0.0036	0.103 mg/L	0.0036	3.52%
Tlf	325.5	0.112 mg/L		0.0011	0.112 mg/L	0.0011	0.98%
Vf	310.5	0.00194 mg/L		0.000295	0.00194 mg/L	0.000295	5.24%
Zrf	1198.5	0.0240 mg/L		0.00061	0.0240 mg/L	0.00061	2.53%
Alxf	7148.3	74.4 ug/L		0.88	0.0744 mg/L	0.00088	1.19%
Bexf	3286.1	1.12 ug/L		0.013	0.00112 mg/L	0.000013	1.17%

MRL bala.
 Return to 070814
 8/14/07

Sequence No.: 19
Sample ID: MBLANK
Analyst: Walter Hsieh
Initial Sample Wt:
Dilution: 1X

Autosampler Location: 38
Date Collected: 8/13/2007 20:05:18
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: MBLANK

Analyte Back Pressure Flow
All 219.0 kPa 0.65 L/min

Mean Data: MBLANK

Analyte	Mean Corrected			Std.Dev.	Sample		RSD
	Intensity	Conc.	Units		Conc.	Units	
Sca	272687.9	106	%	0.3			0.24%
Yr	256801.0	98.5	%	0.82			0.83%
B_†	38.0	0.00144	mg/L	0.000089	0.00144	mg/L	0.000089 6.19%
Ba†	-4.8	-0.00006	mg/L	0.000038	-0.00006	mg/L	0.000038 64.02%
Be†	76.4	0.00002	mg/L	0.000006	0.00002	mg/L	0.000006 24.21%
Ca†	492.4	0.0500	mg/L	0.000069	0.0500	mg/L	0.000069 1.39%
Cd†	-0.8	-0.00003	mg/L	0.000175	-0.00003	mg/L	0.000175 520.85%
Co†	-0.2	-0.00001	mg/L	0.000071	-0.00001	mg/L	0.000071 810.65%
Cr†	-6.5	-0.00009	mg/L	0.000057	-0.00009	mg/L	0.000057 62.35%
Cu†	-37.2	-0.00009	mg/L	0.000158	-0.00009	mg/L	0.000158 181.83%
Fe†	2.9	0.00157	mg/L	0.000015	0.00157	mg/L	0.000015 0.97%
K†	-87.0	-0.0656	mg/L	0.02648	-0.0656	mg/L	0.02648 40.35%
Mg†	18.3	0.00450	mg/L	0.000530	0.00450	mg/L	0.000530 11.78%
Mo†	-0.2	-0.00001	mg/L	0.000287	-0.00001	mg/L	0.000287 >999.9%
Na†	83.1	0.0247	mg/L	0.02126	0.0247	mg/L	0.02126 86.12%
Ni†	-0.3	-0.00002	mg/L	0.000028	-0.00002	mg/L	0.000028 177.48%
Pb†	13.0	0.00244	mg/L	0.000906	0.00244	mg/L	0.000906 37.15%
Tl†	2.7	0.00090	mg/L	0.002447	0.00090	mg/L	0.002447 271.08%
V†	-3.6	-0.00002	mg/L	0.000156	-0.00002	mg/L	0.000156 684.79%
Zn†	89.1	0.00176	mg/L	0.000042	0.00176	mg/L	0.000042 2.39%

Sequence No.: 20
 Sample ID: LCS
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 39
 Date Collected: 8/13/2007 20:08:59
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: LCS

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	257550.0	99.9 %	1.79			1.79%
Yr	254796.7	97.7 %	1.07			1.10%
B_t	12529.7	0.474 mg/L	0.0010	0.474 mg/L	0.0010	0.21%
Bat	77869.2	0.953 mg/L	0.0095	0.953 mg/L	0.0095	0.99%
Bef	163821.6	0.0500 mg/L	0.00050	0.0500 mg/L	0.00050	0.99%
Ca†	501425.8	50.9 mg/L	0.07	50.9 mg/L	0.07	0.14%
Cd†	5133.2	0.223 mg/L	0.0028	0.223 mg/L	0.0028	1.26%
Co†	22768.6	1.00 mg/L	0.013	1.00 mg/L	0.013	1.31%
Cr†	67063.1	0.955 mg/L	0.0066	0.955 mg/L	0.0066	0.69%
Cu†	410209.3	0.961 mg/L	0.0054	0.961 mg/L	0.0054	0.56%
Fet	9188.7	5.03 mg/L	0.026	5.03 mg/L	0.026	0.53%
K†	25635.5	19.3 mg/L	0.15	19.3 mg/L	0.15	0.80%
Mg†	83719.3	20.6 mg/L	0.04	20.6 mg/L	0.04	0.19%
Mor	14210.9	0.953 mg/L	0.0127	0.953 mg/L	0.0127	1.33%
Nat	171567.3	51.0 mg/L	0.17	51.0 mg/L	0.17	0.33%
Nit	10852.6	0.498 mg/L	0.0065	0.498 mg/L	0.0065	1.31%
Pb†	5542.4	1.04 mg/L	0.017	1.04 mg/L	0.017	1.61%
Tlt	3179.1	1.07 mg/L	0.014	1.07 mg/L	0.014	1.33%
V†	154367.8	0.956 mg/L	0.0077	0.956 mg/L	0.0077	0.81%
Znt	51744.5	1.02 mg/L	0.007	1.02 mg/L	0.007	0.69%

Sequence No.: 21
 Sample ID: LCSD
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 40
 Date Collected: 8/13/2007 20:12:44
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min.

Mean Data: LCSD

Analyte	Mean Corrected			Sample			RSD
	Intensity	Conc. Units	Calib	Conc. Units	Std.Dev.	Std.Dev.	
Sca	263250.2	102 %			0.5		0.49%
Yr	251007.8	96.3 %			1.14		1.18%
B_t	12437.3	0.470 mg/L		0.470 mg/L	0.0022	0.0022	0.46%
Bat	76752.4	0.939 mg/L		0.939 mg/L	0.0017	0.0017	0.19%
Set	160781.3	0.0490 mg/L		0.0490 mg/L	0.00022	0.00022	0.45%
Cat	501523.4	50.9 mg/L		50.9 mg/L	0.04	0.04	0.07%
Cdt	5037.6	0.219 mg/L		0.219 mg/L	0.0016	0.0016	0.72%
Cot	22339.8	0.985 mg/L		0.985 mg/L	0.0038	0.0038	0.38%
Crt	66094.3	0.941 mg/L		0.941 mg/L	0.0043	0.0043	0.46%
Cuf	408251.2	0.956 mg/L		0.956 mg/L	0.0034	0.0034	0.36%
Fet	9270.3	5.07 mg/L		5.07 mg/L	0.046	0.046	0.91%
Kt	25958.9	19.6 mg/L		19.6 mg/L	0.05	0.05	0.23%
Mgt	84288.9	20.7 mg/L		20.7 mg/L	0.16	0.16	0.76%
Mot	13920.3	0.933 mg/L		0.933 mg/L	0.0024	0.0024	0.26%
Nat	172105.1	51.1 mg/L		51.1 mg/L	0.02	0.02	0.03%
Nit	10601.8	0.487 mg/L		0.487 mg/L	0.0019	0.0019	0.40%
Pbt	5430.9	1.02 mg/L		1.02 mg/L	0.006	0.006	0.57%
Tlt	3119.2	1.05 mg/L		1.05 mg/L	0.007	0.007	0.67%
Vt	151866.7	0.940 mg/L		0.940 mg/L	0.0040	0.0040	0.43%
Znt	50968.5	1.00 mg/L		1.00 mg/L	0.003	0.003	0.27%

Sequence No.: 22
 Sample ID: 2708020271_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 41
 Date Collected: 8/13/2007 20:16:28
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020271_2X

Analyte Back Pressure Flow
 All 220.0 kPa 0.65 L/min

Mean Data: 2708020271_2X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	240251.2	93.2	%	0.08			0.08%
Yr	239608.3	91.9	%	1.32			1.44%
B_t	38887.9	1.48	mg/L	0.004	2.95	mg/L	0.24%
Baf	1063.2	0.0130	mg/L	0.00003	0.0260	mg/L	0.21%
Bef	-959.6	-0.00029	mg/L	0.000005	-0.00059	mg/L	1.73%
Ca	1895014.4	192	mg/L	1.8	385	mg/L	0.91%
Cdf	42.3	0.00181	mg/L	0.000008	0.00361	mg/L	0.45%
Cof	-14.2	-0.00063	mg/L	0.000091	-0.00125	mg/L	14.49%
Crt	7553.5	0.108	mg/L	0.0008	0.215	mg/L	0.77%
Cut	984.4	0.00230	mg/L	0.000271	0.00461	mg/L	11.78%
Fet	1.3	0.00072	mg/L	0.000759	0.00145	mg/L	104.71%
Kt	23356.9	17.6	mg/L	0.07	35.3	mg/L	0.41%
Mgt	368732.5	90.8	mg/L	0.90	182	mg/L	0.99%
Mof	205.0	0.0137	mg/L	0.00017	0.0275	mg/L	1.22%
Naf	1481669.5	440	mg/L	4.2	881	mg/L	0.95%
Nif	-16.2	-0.00074	mg/L	0.000353	-0.00148	mg/L	47.64%
Pbf	-21.9	-0.00410	mg/L	0.000538	-0.00820	mg/L	13.11%
Tlt	55.8	0.0188	mg/L	0.00307	0.0376	mg/L	16.35%
Vt	2415.5	0.0155	mg/L	0.00008	0.0309	mg/L	0.51%
Znt	361.8	0.00713	mg/L	0.000127	0.0143	mg/L	1.78%

Sequence No.: 23
 Sample ID: 2708020271_2XMS
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 42
 Date Collected: 8/13/2007 20:20:51
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020271_2XMS
 Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020271_2XMS

Analyte	Mean Corrected			Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.	Units		Conc.	Units		
Sca	238983.1	92.7	%	0.56				0.60%
Yr	235573.3	90.4	%	1.24				1.38%
B_f	45364.1	1.72	mg/L	0.008	3.44	mg/L	0.015	0.44%
Ba†	40935.1	0.501	mg/L	0.0009	1.00	mg/L	0.002	0.17%
Be†	82115.4	0.0250	mg/L	0.00001	0.0501	mg/L	0.00001	0.03%
Ca†	2137664.5	217	mg/L	0.7	434	mg/L	1.4	0.32%
Cd†	2735.8	0.119	mg/L	0.0002	0.238	mg/L	0.0003	0.14%
Co†	11136.0	0.491	mg/L	0.0022	0.982	mg/L	0.0045	0.46%
Cr†	41284.5	0.588	mg/L	0.0002	1.18	mg/L	0.000	0.03%
Cu†	214971.8	0.504	mg/L	0.0017	1.01	mg/L	0.003	0.34%
Fe†	4857.3	2.66	mg/L	0.068	5.32	mg/L	0.135	2.54%
K†	37861.4	28.6	mg/L	0.68	57.1	mg/L	1.37	2.39%
Mg†	408796.0	101	mg/L	0.0	201	mg/L	0.1	0.03%
Mn†	7424.1	0.498	mg/L	0.0026	0.995	mg/L	0.0051	0.52%
Na†	1556229.9	462	mg/L	1.1	925	mg/L	2.2	0.24%
Ni†	5237.2	0.240	mg/L	0.0018	0.481	mg/L	0.0037	0.77%
Pb†	2725.1	0.509	mg/L	0.0039	1.02	mg/L	0.008	0.77%
Tl†	1572.9	0.529	mg/L	0.0046	1.06	mg/L	0.009	0.87%
V†	81527.3	0.505	mg/L	0.0001	1.01	mg/L	0.000	0.02%
Zn†	27522.9	0.541	mg/L	0.0028	1.08	mg/L	0.006	0.52%

Sequence No.: 24
 Sample ID: 2708020271_2XMSD
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 43
 Date Collected: 8/13/2007 20:24:37
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020271_2XMSD

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020271_2XMSD

Analyte	Mean Corrected			Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.		
Sca	237294.6	92.0 %	0.03				0.03%
Yr	238965.5	91.7 %	0.21				0.23%
B_t	44702.6	1.70 mg/L	0.008	3.39 mg/L	0.016		0.48%
Bat	40297.1	0.493 mg/L	0.0027	0.986 mg/L	0.0054		0.55%
Bet	80677.3	0.0246 mg/L	0.00016	0.0492 mg/L	0.00031		0.63%
Cat	2106008.1	214 mg/L	0.1	428 mg/L	0.3		0.07%
Cdt	2692.6	0.117 mg/L	0.0001	0.234 mg/L	0.0002		0.09%
Cot	11003.3	0.485 mg/L	0.0013	0.971 mg/L	0.0026		0.26%
Crt	40567.2	0.578 mg/L	0.0022	1.16 mg/L	0.004		0.39%
Cuf	211557.1	0.496 mg/L	0.0029	0.991 mg/L	0.0058		0.58%
Fet	4706.8	2.58 mg/L	0.038	5.15 mg/L	0.075		1.47%
Kt	37109.2	28.0 mg/L	0.02	56.0 mg/L	0.04		0.06%
Mgt	403360.3	99.3 mg/L	0.22	199 mg/L	0.4		0.22%
Mot	7347.9	0.493 mg/L	0.0016	0.985 mg/L	0.0032		0.33%
Nat	1537421.5	457 mg/L	1.4	914 mg/L	2.9		0.32%
Nit	5165.2	0.237 mg/L	0.0013	0.474 mg/L	0.0027		0.57%
Pbt	2647.4	0.495 mg/L	0.0015	0.990 mg/L	0.0029		0.30%
Tlt	1534.7	0.517 mg/L	0.0034	1.03 mg/L	0.007		0.66%
Vt	79992.0	0.496 mg/L	0.0022	0.991 mg/L	0.0045		0.45%
Znt	26938.3	0.529 mg/L	0.0021	1.06 mg/L	0.004		0.40%

Sequence No.: 27
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/13/2007 20:36:30
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	220.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	248794.2	96.5	%	0.82			0.85%
Yr	249401.6	95.7	%	0.42			0.43%
B ₁	63579.0	2.40	mg/L	0.009	2.40	mg/L	0.38%
	QC value within limits for B ₁		Recovery = 96.13%				
Ba ₁	399389.6	4.89	mg/L	0.005	4.89	mg/L	0.09%
	QC value within limits for Ba		Recovery = 97.72%				
Be ₁	6378886.9	1.95	mg/L	0.018	1.95	mg/L	0.94%
	QC value within limits for Be		Recovery = 97.26%				
Ca ₁	507692.5	51.6	mg/L	0.39	51.6	mg/L	0.77%
	QC value within limits for Ca		Recovery = 103.11%				
Cd ₁	58626.3	2.53	mg/L	0.008	2.53	mg/L	0.30%
	QC value within limits for Cd		Recovery = 101.09%				
Co ₁	113534.8	5.01	mg/L	0.012	5.01	mg/L	0.23%
	QC value within limits for Co		Recovery = 100.16%				
Cr ₁	341443.1	4.86	mg/L	0.018	4.86	mg/L	0.36%
	QC value within limits for Cr		Recovery = 97.27%				
Cu ₁	2016821.1	4.72	mg/L	0.013	4.72	mg/L	0.28%
	QC value within limits for Cu		Recovery = 94.49%				
Fe ₁	9168.8	5.02	mg/L	0.116	5.02	mg/L	2.31%
	QC value within limits for Fe		Recovery = 100.38%				
K ₁	65121.8	49.1	mg/L	1.02	49.1	mg/L	2.08%
	QC value within limits for K		Recovery = 98.29%				
Mg ₁	212930.7	52.4	mg/L	0.41	52.4	mg/L	0.78%
	QC value within limits for Mg		Recovery = 104.82%				
Mo ₁	71138.7	4.77	mg/L	0.024	4.77	mg/L	0.49%
	QC value within limits for Mo		Recovery = 95.38%				
Na ₁	174321.3	51.8	mg/L	0.33	51.8	mg/L	0.64%
	QC value within limits for Na		Recovery = 103.60%				
Ni ₁	108518.5	4.98	mg/L	0.014	4.98	mg/L	0.28%
	QC value within limits for Ni		Recovery = 99.63%				
Pb ₁	26965.0	5.04	mg/L	0.033	5.04	mg/L	0.66%
	QC value within limits for Pb		Recovery = 100.80%				
Tl ₁	15561.2	5.24	mg/L	0.023	5.24	mg/L	0.44%
	QC value within limits for Tl		Recovery = 104.74%				
V ₁	786477.7	4.87	mg/L	0.006	4.87	mg/L	0.13%
	QC value within limits for V		Recovery = 97.37%				
Zn ₁	259071.2	5.07	mg/L	0.005	5.07	mg/L	0.09%
	QC value within limits for Zn		Recovery = 101.42%				

All analyte(s) passed QC.

Sequence No.: 28
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 20:39:45
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	263623.8	102	%	0.4			0.39%
Yr	247041.8	94.8	%	2.85			3.01%
B _f	479.7	0.0182	mg/L	0.00101	0.0182 mg/L	0.00101	5.55%
	QC value within limits for B _f	Recovery = Not calculated					
Ba _f	-6.7	-0.00008	mg/L	0.000068	-0.00008 mg/L	0.000068	82.77%
	QC value within limits for Ba	Recovery = Not calculated					
Be _f	59.3	0.00002	mg/L	0.000020	0.00002 mg/L	0.000020	111.08%
	QC value within limits for Be	Recovery = Not calculated					
Ca _f	35.5	0.00361	mg/L	0.001272	0.00361 mg/L	0.001272	35.24%
	QC value within limits for Ca	Recovery = Not calculated					
Cd _f	7.3	0.00031	mg/L	0.000057	0.00031 mg/L	0.000057	18.30%
	QC value within limits for Cd	Recovery = Not calculated					
Co _f	-4.6	-0.00020	mg/L	0.000089	-0.00020 mg/L	0.000089	43.86%
	QC value within limits for Co	Recovery = Not calculated					
Cr _f	3.3	0.00005	mg/L	0.000221	0.00005 mg/L	0.000221	474.68%
	QC value within limits for Cr	Recovery = Not calculated					
Cu _f	51.6	0.00012	mg/L	0.000208	0.00012 mg/L	0.000208	172.27%
	QC value within limits for Cu	Recovery = Not calculated					
Fe _f	2.2	0.00123	mg/L	0.000261	0.00123 mg/L	0.000261	21.24%
	QC value within limits for Fe	Recovery = Not calculated					
K _f	17.6	0.0133	mg/L	0.01857	0.0133 mg/L	0.01857	140.07%
	QC value within limits for K	Recovery = Not calculated					
Mg _f	-0.3	-0.00007	mg/L	0.002011	-0.00007 mg/L	0.002011	>999.9%
	QC value within limits for Mg	Recovery = Not calculated					
Mo _f	9.6	0.00064	mg/L	0.000273	0.00064 mg/L	0.000273	42.55%
	QC value within limits for Mo	Recovery = Not calculated					
Na _f	553.6	0.164	mg/L	0.0072	0.164 mg/L	0.0072	4.40%
	QC value within limits for Na	Recovery = Not calculated					
Ni _f	-3.2	-0.00014	mg/L	0.000129	-0.00014 mg/L	0.000129	89.39%
	QC value within limits for Ni	Recovery = Not calculated					
Pb _f	13.9	0.00259	mg/L	0.000141	0.00259 mg/L	0.000141	5.42%
	QC value within limits for Pb	Recovery = Not calculated					
Tl _f	5.0	0.00167	mg/L	0.000887	0.00167 mg/L	0.000887	52.95%
	QC value within limits for Tl	Recovery = Not calculated					
V _f	-35.8	-0.00022	mg/L	0.000189	-0.00022 mg/L	0.000189	85.86%
	QC value within limits for V	Recovery = Not calculated					
Zn _f	7.0	0.00014	mg/L	0.000113	0.00014 mg/L	0.000113	81.81%
	QC value within limits for Zn	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 39
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/13/2007 21:25:54
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	219.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	250959.4	97.3 %	0.74			0.77%
Yr	240539.3	92.3 %	0.32			0.35%
B _f	62421.1	2.36 mg/L	0.004	2.36 mg/L	0.004	0.15%
	QC value within limits for B _f	Recovery = 94.37%				
Ba _f	391862.3	4.79 mg/L	0.032	4.79 mg/L	0.032	0.68%
	QC value within limits for Ba	Recovery = 95.87%				
Be _f	6261796.2	1.91 mg/L	0.016	1.91 mg/L	0.016	0.83%
	QC value within limits for Be	Recovery = 95.48%				
Ca _f	512459.0	52.0 mg/L	0.12	52.0 mg/L	0.12	0.24%
	QC value within limits for Ca	Recovery = 104.07%				
Cd _f	58695.5	2.53 mg/L	0.034	2.53 mg/L	0.034	1.34%
	QC value within limits for Cd	Recovery = 101.21%				
Co _f	113876.7	5.02 mg/L	0.064	5.02 mg/L	0.064	1.28%
	QC value within limits for Co	Recovery = 100.46%				
Cr _f	336698.7	4.80 mg/L	0.016	4.80 mg/L	0.016	0.33%
	QC value within limits for Cr	Recovery = 95.92%				
Cu _f	1977556.8	4.63 mg/L	0.035	4.63 mg/L	0.035	0.77%
	QC value within limits for Cu	Recovery = 92.66%				
Fe _f	9421.6	5.16 mg/L	0.012	5.16 mg/L	0.012	0.23%
	QC value within limits for Fe	Recovery = 103.14%				
K _f	66466.9	50.2 mg/L	0.62	50.2 mg/L	0.62	1.24%
	QC value within limits for K	Recovery = 100.32%				
Mg _f	215217.4	53.0 mg/L	0.09	53.0 mg/L	0.09	0.16%
	QC value within limits for Mg	Recovery = 105.95%				
Mo _f	70836.2	4.75 mg/L	0.082	4.75 mg/L	0.082	1.72%
	QC value within limits for Mo	Recovery = 94.97%				
Na _f	174994.9	52.0 mg/L	0.27	52.0 mg/L	0.27	0.53%
	QC value within limits for Na	Recovery = 104.00%				
Ni _f	108897.5	5.00 mg/L	0.080	5.00 mg/L	0.080	1.60%
	QC value within limits for Ni	Recovery = 99.98%				
Pb _f	26930.2	5.03 mg/L	0.090	5.03 mg/L	0.090	1.79%
	QC value within limits for Pb	Recovery = 100.67%				
Tl _f	15488.1	5.21 mg/L	0.066	5.21 mg/L	0.066	1.27%
	QC value within limits for Tl	Recovery = 104.25%				
V _f	773191.8	4.79 mg/L	0.018	4.79 mg/L	0.018	0.38%
	QC value within limits for V	Recovery = 95.72%				
Zn _f	254872.8	4.99 mg/L	0.024	4.99 mg/L	0.024	0.49%
	QC value within limits for Zn	Recovery = 99.76%				

All analyte(s) passed QC.

Sequence No.: 40
Sample ID: CCB
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 0
Date Collected: 8/13/2007 21:29:08
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
All 219.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	267268.4	104 %		0.5			0.50%
Yr	247462.2	94.9 %		0.01			0.01%
B_t	332.8	0.0126 mg/L		0.00031	0.0126 mg/L	0.00031	2.44%
	QC value within limits for B_ Recovery = Not calculated						
Bat	-4.7	-0.00006 mg/L		0.000013	-0.00006 mg/L	0.000013	22.32%
	QC value within limits for Ba Recovery = Not calculated						
Be†	82.2	0.00003 mg/L		0.000003	0.00003 mg/L	0.000003	10.64%
	QC value within limits for Be Recovery = Not calculated						
Ca†	37.3	0.00378 mg/L		0.000055	0.00378 mg/L	0.000055	1.45%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	1.1	0.00005 mg/L		0.000029	0.00005 mg/L	0.000029	62.17%
	QC value within limits for Cd Recovery = Not calculated						
Co†	2.0	0.00009 mg/L		0.000220	0.00009 mg/L	0.000220	250.00%
	QC value within limits for Co Recovery = Not calculated						
Cr†	156.1	0.00222 mg/L		0.000151	0.00222 mg/L	0.000151	6.77%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-36.3	-0.00008 mg/L		0.000131	-0.00008 mg/L	0.000131	153.89%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-2.2	-0.00122 mg/L		0.001379	-0.00122 mg/L	0.001379	113.36%
	QC value within limits for Fe Recovery = Not calculated						
K†	-26.5	-0.0200 mg/L		0.06398	-0.0200 mg/L	0.06398	320.26%
	QC value within limits for K Recovery = Not calculated						
Mg†	-0.0	0.00000 mg/L		0.000699	0.00000 mg/L	0.000699	>999.9%
	QC value within limits for Mg Recovery = Not calculated						
Mo†	12.2	0.00082 mg/L		0.000045	0.00082 mg/L	0.000045	5.54%
	QC value within limits for Mo Recovery = Not calculated						
Na†	358.1	0.106 mg/L		0.0090	0.106 mg/L	0.0090	8.44%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-2.2	-0.00010 mg/L		0.000095	-0.00010 mg/L	0.000095	94.57%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	10.2	0.00190 mg/L		0.000472	0.00190 mg/L	0.000472	24.87%
	QC value within limits for Pb Recovery = Not calculated						
Tl†	2.4	0.00082 mg/L		0.000252	0.00082 mg/L	0.000252	30.84%
	QC value within limits for Tl Recovery = Not calculated						
V†	-3.8	-0.00001 mg/L		0.000096	-0.00001 mg/L	0.000096	883.78%
	QC value within limits for V Recovery = Not calculated						
Zn†	-5.5	-0.00011 mg/L		0.000092	-0.00011 mg/L	0.000092	84.39%
	QC value within limits for Zn Recovery = Not calculated						
All analyte(s) passed QC.							

Sequence No.: 41
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 8/13/2007 21:32:33
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	219.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	257948.7	100	%	0.3			0.29%
Yr	243689.0	93.5	%	1.70			1.82%
B_†	30683.3	1.16	mg/L	0.001	1.16 mg/L	0.001	0.05%
	QC value within limits for B_ Recovery = 92.78%						
Ba†	196193.4	2.40	mg/L	0.006	2.40 mg/L	0.006	0.25%
	QC value within limits for Ba Recovery = 96.00%						
Be†	3116016.1	0.950	mg/L	0.0023	0.950 mg/L	0.0023	0.25%
	QC value within limits for Be Recovery = 95.02%						
Ca†	257880.5	26.2	mg/L	0.07	26.2 mg/L	0.07	0.26%
	QC value within limits for Ca Recovery = 104.74%						
Cd†	28775.4	1.24	mg/L	0.008	1.24 mg/L	0.008	0.66%
	QC value within limits for Cd Recovery = 99.24%						
Co†	55834.1	2.46	mg/L	0.007	2.46 mg/L	0.007	0.26%
	QC value within limits for Co Recovery = 98.52%						
Cr†	167504.3	2.39	mg/L	0.003	2.39 mg/L	0.003	0.12%
	QC value within limits for Cr Recovery = 95.44%						
Cu†	972807.9	2.28	mg/L	0.001	2.28 mg/L	0.001	0.03%
	QC value within limits for Cu Recovery = 91.16%						
Fe†	4709.8	2.58	mg/L	0.044	2.58 mg/L	0.044	1.70%
	QC value within limits for Fe Recovery = 103.12%						
K†	32654.5	24.6	mg/L	0.20	24.6 mg/L	0.20	0.81%
	QC value within limits for K Recovery = 98.57%						
Mg†	108826.3	26.8	mg/L	0.00	26.8 mg/L	0.00	0.01%
	QC value within limits for Mg Recovery = 107.15%						
Mo†	34594.3	2.32	mg/L	0.008	2.32 mg/L	0.008	0.35%
	QC value within limits for Mo Recovery = 92.76%						
Na†	87369.0	26.0	mg/L	0.11	26.0 mg/L	0.11	0.43%
	QC value within limits for Na Recovery = 103.85%						
Ni†	54450.0	2.50	mg/L	0.008	2.50 mg/L	0.008	0.30%
	QC value within limits for Ni Recovery = 99.98%						
Pb†	13435.1	2.51	mg/L	0.007	2.51 mg/L	0.007	0.28%
	QC value within limits for Pb Recovery = 100.44%						
Tl†	7728.3	2.60	mg/L	0.002	2.60 mg/L	0.002	0.07%
	QC value within limits for Tl Recovery = 104.03%						
V†	383531.0	2.37	mg/L	0.004	2.37 mg/L	0.004	0.16%
	QC value within limits for V Recovery = 94.96%						
Zn†	127339.3	2.49	mg/L	0.006	2.49 mg/L	0.006	0.26%
	QC value within limits for Zn Recovery = 99.68%						

All analyte(s) passed QC.

Sequence No.: 48
 Sample ID: 2708020317_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 62
 Date Collected: 8/13/2007 22:02:15
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020317_2X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020317_2X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	254584.4	98.7	%	0.97				0.98%
Yr	242022.0	92.8	%	0.64				0.69%
B ₊	20290.5	0.770	mg/L	0.0002	1.54	mg/L	0.000	0.03%
Ba†	812.4	0.00994	mg/L	0.000172	0.0199	mg/L	0.00034	1.73%
Be†	-368.9	-0.00011	mg/L	0.000006	-0.00023	mg/L	0.000013	5.57%
Ca†	786534.8	79.9	mg/L	0.70	160	mg/L	1.4	0.88%
Cd†	16.0	0.00068	mg/L	0.000084	0.00137	mg/L	0.000168	12.31%
Co†	-12.4	-0.00055	mg/L	0.000177	-0.00109	mg/L	0.000353	32.35%
Cr†	564.1	0.00804	mg/L	0.000027	0.0161	mg/L	0.00005	0.33%
Cu†	509.4	0.00119	mg/L	0.000054	0.00238	mg/L	0.000108	4.54%
Fe†	168.8	0.0924	mg/L	0.00197	0.185	mg/L	0.0039	2.13%
K†	7955.6	6.00	mg/L	0.050	12.0	mg/L	0.10	0.83%
Mg†	171451.5	42.2	mg/L	0.37	84.4	mg/L	0.74	0.87%
Mn†	125.4	0.00841	mg/L	0.000170	0.0168	mg/L	0.00034	2.02%
Na†	605052.7	180	mg/L	0.1	360	mg/L	0.2	0.04%
Ni†	-6.0	-0.00028	mg/L	0.000197	-0.00056	mg/L	0.000393	70.79%
Pb†	-11.5	-0.00215	mg/L	0.000372	-0.00431	mg/L	0.000745	17.28%
Tl†	30.8	0.0104	mg/L	0.00026	0.0208	mg/L	0.00053	2.53%
V†	3042.3	0.0188	mg/L	0.00021	0.0375	mg/L	0.00043	1.14%
Zn†	426.4	0.00840	mg/L	0.000214	0.0168	mg/L	0.00043	2.55%

Sequence No.: 49
Sample ID: 2708020319_5X
Analyst: Walter Hsieh
Initial Sample Wt:
Dilution: 5X

Autosampler Location: 63
Date Collected: 8/13/2007 22:06:36
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: 2708020319_5X

Analyte Back Pressure Flow
All 218.0 kPa 0.65 L/min

Mean Data: 2708020319_5X

Analyte	Mean Corrected			Std.Dev.	Sample		RSD
	Intensity	Conc.	Units		Conc.	Units	
Sca	252139.6	97.8	%	0.02			0.02%
Yr	240663.1	92.3	%	1.24			1.34%
B ₊	35141.8	1.33	mg/L	0.004	6.67	mg/L	0.018
B _{at}	595.4	0.00728	mg/L	0.000044	0.0364	mg/L	0.00022
B _{et}	-1596.6	-0.00049	mg/L	0.000015	-0.00243	mg/L	0.000074
Ca _t	1193410.6	121	mg/L	0.2	606	mg/L	1.1
Cd _t	-15.8	-0.00067	mg/L	0.000282	-0.00336	mg/L	0.001409
Co _t	9.1	0.00040	mg/L	0.000045	0.00200	mg/L	0.000227
Crt	159451.4	2.27	mg/L	0.008	11.4	mg/L	0.04
Cu _t	765.2	0.00179	mg/L	0.000157	0.00896	mg/L	0.000786
Fe _t	51.0	0.0279	mg/L	0.00003	0.140	mg/L	0.0001
K _t	6114.6	4.61	mg/L	0.051	23.1	mg/L	0.26
Mg _t	226144.5	55.7	mg/L	0.11	278	mg/L	0.5
Mo _t	48.4	0.00325	mg/L	0.000460	0.0162	mg/L	0.00230
Na _t	1105309.7	328	mg/L	1.1	1640	mg/L	5.3
Ni _t	-3.0	-0.00014	mg/L	0.000118	-0.00069	mg/L	0.000590
Pb _t	-14.9	-0.00278	mg/L	0.000471	-0.0139	mg/L	0.00236
Tl _t	44.4	0.0149	mg/L	0.00085	0.0745	mg/L	0.00423
V _t	-470.3	0.00964	mg/L	0.000108	0.0482	mg/L	0.00054
Zn _t	406.6	0.00801	mg/L	0.000275	0.0401	mg/L	0.00138

Sequence No.: 50
 Sample ID: 2708020318_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 64
 Date Collected: 8/13/2007 22:11:00
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020318_2X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020318_2X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	248891.1	96.5	%	0.06				0.06%
Yr	237562.8	91.1	%	0.45				0.49%
B_t	67811.8	2.57	mg/L	0.012	5.15	mg/L	0.024	0.47%
Bat	1652.3	0.0202	mg/L	0.00018	0.0404	mg/L	0.00035	0.87%
Bet	-820.4	-0.00025	mg/L	0.000014	-0.00050	mg/L	0.000028	5.58%
Caf	1521834.2	155	mg/L	0.4	309	mg/L	0.8	0.24%
Cdt	42.8	0.00183	mg/L	0.000033	0.00366	mg/L	0.000066	1.82%
Cot	-11.4	-0.00050	mg/L	0.000063	-0.00101	mg/L	0.000125	12.48%
Crt	2136.7	0.0304	mg/L	0.00002	0.0609	mg/L	0.00003	0.05%
Cuf	1266.7	0.00296	mg/L	0.000148	0.00593	mg/L	0.000297	5.00%
Fet	68.0	0.0372	mg/L	0.00033	0.0744	mg/L	0.00065	0.88%
Kt	12311.2	9.29	mg/L	0.089	18.6	mg/L	0.18	0.96%
Mgt	404706.5	99.6	mg/L	0.40	199	mg/L	0.8	0.40%
Mot	130.2	0.00873	mg/L	0.000114	0.0175	mg/L	0.00023	1.31%
Naf	1070945.6	318	mg/L	0.5	636	mg/L	1.0	0.15%
Nit	3.9	0.00018	mg/L	0.000225	0.00035	mg/L	0.000450	126.90%
Pbt	-7.4	-0.00139	mg/L	0.001374	-0.00278	mg/L	0.002749	98.93%
Tlt	40.0	0.0135	mg/L	0.00262	0.0270	mg/L	0.00524	19.40%
Vt	3135.2	0.0195	mg/L	0.00003	0.0389	mg/L	0.00007	0.17%
Znt	625.2	0.0123	mg/L	0.00011	0.0246	mg/L	0.00021	0.87%

Sequence No.: 51
 Sample ID: MBLANK
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 65
 Date Collected: 8/13/2007 22:15:26
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MBLANK

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: MBLANK

Analyte	Mean Corrected			Std.Dev.	Sample		RSD
	Intensity	Conc.	Units		Conc.	Units	
Sca	272091.5	106	%	0.1			0.05%
Yr	256622.2	98.4	%	1.32			1.34%
B_t	586.2	0.0223	mg/L	0.00049	0.0223	mg/L	0.00049 2.20%
Bat	-4.1	-0.00005	mg/L	0.000012	-0.00005	mg/L	0.000012 23.69%
Be†	2.0	0.00000	mg/L	0.000022	0.00000	mg/L	0.000022 >999.9%
Ca†	277.1	0.0281	mg/L	0.00098	0.0281	mg/L	0.00098 3.47%
Cd†	-4.2	-0.00018	mg/L	0.000024	-0.00018	mg/L	0.000024 13.12%
Co†	-4.5	-0.00020	mg/L	0.000092	-0.00020	mg/L	0.000092 46.73%
Cr†	49.2	0.00070	mg/L	0.000052	0.00070	mg/L	0.000052 7.49%
Cu†	-134.5	-0.00031	mg/L	0.000012	-0.00031	mg/L	0.000012 3.71%
Fe†	2.7	0.00148	mg/L	0.001573	0.00148	mg/L	0.001573 106.63%
K†	-37.1	-0.0280	mg/L	0.05695	-0.0280	mg/L	0.05695 203.18%
Mg†	7.6	0.00186	mg/L	0.000032	0.00186	mg/L	0.000032 1.74%
Mo†	-5.3	-0.00036	mg/L	0.000081	-0.00036	mg/L	0.000081 22.74%
Na†	1555.6	0.462	mg/L	0.0199	0.462	mg/L	0.0199 4.30%
Ni†	-13.3	-0.00061	mg/L	0.000007	-0.00061	mg/L	0.000007 1.20%
Pb†	10.7	0.00200	mg/L	0.000165	0.00200	mg/L	0.000165 8.25%
Tl†	-4.0	-0.00135	mg/L	0.001537	-0.00135	mg/L	0.001537 113.95%
V†	-34.7	-0.00021	mg/L	0.000041	-0.00021	mg/L	0.000041 19.75%
Zn†	10.7	0.00022	mg/L	0.000133	0.00022	mg/L	0.000133 61.85%

Sequence No.: 52
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/13/2007 22:19:07
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	255497.5		99.1 %	1.19			1.20%
Yr	239926.8		92.0 %	1.11			1.21%
B ₊	61863.0		2.34 mg/L	0.004	2.34 mg/L	0.004	0.17%
	QC value within limits for B ₊ Recovery = 93.53%						
Ba ₊	385327.2		4.71 mg/L	0.008	4.71 mg/L	0.008	0.17%
	QC value within limits for Ba Recovery = 94.28%						
Be ₊	6173198.5		1.88 mg/L	0.024	1.88 mg/L	0.024	1.26%
	QC value within limits for Be Recovery = 94.13%						
Ca ₊	510116.8		51.8 mg/L	0.24	51.8 mg/L	0.24	0.45%
	QC value within limits for Ca Recovery = 103.60%						
Cd ₊	57565.9		2.48 mg/L	0.013	2.48 mg/L	0.013	0.51%
	QC value within limits for Cd Recovery = 99.26%						
Co ₊	111438.6		4.92 mg/L	0.016	4.92 mg/L	0.016	0.33%
	QC value within limits for Co Recovery = 98.31%						
Cr ₊	333813.3		4.75 mg/L	0.011	4.75 mg/L	0.011	0.23%
	QC value within limits for Cr Recovery = 95.10%						
Cu ₊	1942983.5		4.55 mg/L	0.006	4.55 mg/L	0.006	0.13%
	QC value within limits for Cu Recovery = 91.04%						
Fe ₊	9123.1		4.99 mg/L	0.059	4.99 mg/L	0.059	1.18%
	QC value within limits for Fe Recovery = 99.87%						
K ₊	64778.9		48.9 mg/L	0.94	48.9 mg/L	0.94	1.92%
	QC value within limits for K Recovery = 97.77%						
Mg ₊	216044.9		53.2 mg/L	0.21	53.2 mg/L	0.21	0.40%
	QC value within limits for Mg Recovery = 106.36%						
Mo ₊	68779.8		4.61 mg/L	0.017	4.61 mg/L	0.017	0.38%
	QC value within limits for Mo Recovery = 92.22%						
Na ₊	174707.8		51.9 mg/L	0.49	51.9 mg/L	0.49	0.94%
	QC value within limits for Na Recovery = 103.83%						
Ni ₊	105831.7		4.86 mg/L	0.021	4.86 mg/L	0.021	0.43%
	QC value within limits for Ni Recovery = 97.17%						
Pb ₊	26190.7		4.90 mg/L	0.012	4.90 mg/L	0.012	0.25%
	QC value within limits for Pb Recovery = 97.90%						
Tl ₊	15060.8		5.07 mg/L	0.011	5.07 mg/L	0.011	0.22%
	QC value within limits for Tl Recovery = 101.37%						
V ₊	765603.1		1.71 mg/L	0.006	4.74 mg/L	0.006	0.13%
	QC value within limits for V Recovery = 94.78%						
Zn ₊	251112.8		4.91 mg/L	0.010	4.91 mg/L	0.010	0.19%
	QC value within limits for Zn Recovery = 98.30%						
All analyte(s) passed QC.							

Sequence No.: 53
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 22:22:24
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	263714.6	102 %	%	1.2			1.14%
Yr	239977.5	92.0 %	%	0.74			0.80%
B ₊	567.8	0.0216 mg/L	mg/L	0.00031	0.0216 mg/L	0.00031	1.46%
QC value greater than the upper limit for B ₊ Recovery = Not calculated							
Ba ₊	-8.6	-0.00011 mg/L	mg/L	0.000012	-0.00011 mg/L	0.000012	11.31%
QC value within limits for Ba Recovery = Not calculated							
Be ₊	-21.1	-0.00001 mg/L	mg/L	0.000007	-0.00001 mg/L	0.000007	107.25%
QC value within limits for Be Recovery = Not calculated							
Ca ₊	24.5	0.00249 mg/L	mg/L	0.000821	0.00249 mg/L	0.000821	32.99%
QC value within limits for Ca Recovery = Not calculated							
Cd ₊	-0.7	-0.00003 mg/L	mg/L	0.000075	-0.00003 mg/L	0.000075	251.22%
QC value within limits for Cd Recovery = Not calculated							
Co ₊	-1.0	-0.00004 mg/L	mg/L	0.000152	-0.00004 mg/L	0.000152	347.10%
QC value within limits for Co Recovery = Not calculated							
Cr ₊	73.7	0.00105 mg/L	mg/L	0.000077	0.00105 mg/L	0.000077	7.33%
QC value within limits for Cr Recovery = Not calculated							
Cu ₊	-148.7	-0.00035 mg/L	mg/L	0.000157	-0.00035 mg/L	0.000157	45.24%
QC value within limits for Cu Recovery = Not calculated							
Fe ₊	1.2	0.00068 mg/L	mg/L	0.000603	0.00068 mg/L	0.000603	88.62%
QC value within limits for Fe Recovery = Not calculated							
K ₊	-36.8	-0.0278 mg/L	mg/L	0.08694	-0.0278 mg/L	0.08694	313.19%
QC value within limits for K Recovery = Not calculated							
Mg ₊	-1.0	-0.00024 mg/L	mg/L	0.000987	-0.00024 mg/L	0.000987	410.24%
QC value within limits for Mg Recovery = Not calculated							
Mo ₊	11.7	0.00079 mg/L	mg/L	0.000383	0.00079 mg/L	0.000383	48.81%
QC value within limits for Mo Recovery = Not calculated							
Na ₊	702.7	0.209 mg/L	mg/L	0.0283	0.209 mg/L	0.0283	13.55%
QC value within limits for Na Recovery = Not calculated							
Ni ₊	-3.8	-0.00017 mg/L	mg/L	0.000145	-0.00017 mg/L	0.000145	82.97%
QC value within limits for Ni Recovery = Not calculated							
Pb ₊	6.6	0.00124 mg/L	mg/L	0.000894	0.00124 mg/L	0.000894	72.29%
QC value within limits for Pb Recovery = Not calculated							
Tl ₊	-4.7	-0.00159 mg/L	mg/L	0.000931	-0.00159 mg/L	0.000931	58.51%
QC value within limits for Tl Recovery = Not calculated							
V ₊	40.3	0.00025 mg/L	mg/L	0.000013	0.00025 mg/L	0.000013	5.30%
QC value within limits for V Recovery = Not calculated							
Zn ₊	-10.5	-0.00021 mg/L	mg/L	0.000112	-0.00021 mg/L	0.000112	54.32%
QC value within limits for Zn Recovery = Not calculated							
QC Failed. Retry.							

Sequence No.: 54
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 22:24:59
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	263110.2	102	%	2.0			1.99%
Yr	243202.1	93.3	%	0.73			0.78%
B ₋ t	415.3	0.0158	mg/L	0.00007	0.0158 mg/L	0.00007	0.46%
	QC value within limits for B ₋		Recovery = Not calculated				
Ba†	-14.3	-0.00017	mg/L	0.000057	-0.00017 mg/L	0.000057	32.94%
	QC value within limits for Ba		Recovery = Not calculated				
Be†	43.6	0.00001	mg/L	0.000025	0.00001 mg/L	0.000025	187.82%
	QC value within limits for Be		Recovery = Not calculated				
Ca†	2.8	0.00028	mg/L	0.000424	0.00028 mg/L	0.000424	150.58%
	QC value within limits for Ca		Recovery = Not calculated				
Cd†	-0.6	-0.00003	mg/L	0.000006	-0.00003 mg/L	0.000006	21.55%
	QC value within limits for Cd		Recovery = Not calculated				
Co†	-5.4	-0.00024	mg/L	0.000001	-0.00024 mg/L	0.000001	0.53%
	QC value within limits for Co		Recovery = Not calculated				
Cr†	66.5	0.00095	mg/L	0.000081	0.00095 mg/L	0.000081	8.52%
	QC value within limits for Cr		Recovery = Not calculated				
Cu†	-95.0	-0.00022	mg/L	0.000148	-0.00022 mg/L	0.000148	66.37%
	QC value within limits for Cu		Recovery = Not calculated				
Fe†	-0.7	-0.00039	mg/L	0.000304	-0.00039 mg/L	0.000304	77.47%
	QC value within limits for Fe		Recovery = Not calculated				
K†	-59.6	-0.0450	mg/L	0.00580	-0.0450 mg/L	0.00580	12.89%
	QC value within limits for K		Recovery = Not calculated				
Mg†	3.6	0.00089	mg/L	0.000513	0.00089 mg/L	0.000513	57.73%
	QC value within limits for Mg		Recovery = Not calculated				
Mo†	3.6	0.00024	mg/L	0.000002	0.00024 mg/L	0.000002	0.82%
	QC value within limits for Mo		Recovery = Not calculated				
Na†	569.8	0.169	mg/L	0.0080	0.169 mg/L	0.0080	4.74%
	QC value within limits for Na		Recovery = Not calculated				
Ni†	-7.3	-0.00033	mg/L	0.000405	-0.00033 mg/L	0.000405	121.02%
	QC value within limits for Ni		Recovery = Not calculated				
Pb†	10.4	0.00195	mg/L	0.001244	0.00195 mg/L	0.001244	63.82%
	QC value within limits for Pb		Recovery = Not calculated				
Tl†	3.7	0.00124	mg/L	0.001867	0.00124 mg/L	0.001867	150.59%
	QC value within limits for Tl		Recovery = Not calculated				
V†	-53.3	-0.00032	mg/L	0.000224	-0.00032 mg/L	0.000224	69.52%
	QC value within limits for V		Recovery = Not calculated				
Zn†	-15.0	-0.00029	mg/L	0.000110	-0.00029 mg/L	0.000110	37.37%
	QC value within limits for Zn		Recovery = Not calculated				
All analyte(s) passed QC.							

Sequence No.: 55
 Sample ID: LCS
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 66
 Date Collected: 8/13/2007 22:28:27
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: LCS

Analyte	Mean Corrected			Std.Dev.	Sample		
	Intensity	Conc. Units	Calib		Conc. Units	Std.Dev.	RSD
Sca	264996.0	103 %		0.2			0.20%
Yr	245285.8	94.1 %		1.67			1.77%
B ₊	12443.2	0.470 mg/L		0.0021	0.470 mg/L	0.0021	0.45%
Bat	74830.4	0.915 mg/L		0.0033	0.915 mg/L	0.0033	0.36%
Bet	156464.8	0.0477 mg/L		0.00009	0.0477 mg/L	0.00009	0.19%
Cat	489841.2	49.7 mg/L		0.08	49.7 mg/L	0.08	0.16%
Cdt	4918.1	0.214 mg/L		0.0006	0.214 mg/L	0.0006	0.29%
Cot	21889.1	0.966 mg/L		0.0020	0.966 mg/L	0.0020	0.21%
Crt	65294.3	0.930 mg/L		0.0007	0.930 mg/L	0.0007	0.08%
Cut	391302.3	0.917 mg/L		0.0056	0.917 mg/L	0.0056	0.61%
Fet	9289.6	5.08 mg/L		0.107	5.08 mg/L	0.107	2.11%
Kf	25690.6	19.4 mg/L		0.54	19.4 mg/L	0.54	2.80%
Mgt	84657.6	20.8 mg/L		0.56	20.8 mg/L	0.56	2.66%
Mot	13451.6	0.902 mg/L		0.0002	0.902 mg/L	0.0002	0.02%
Naf	167611.9	49.8 mg/L		0.01	49.8 mg/L	0.01	0.02%
Nit	10322.2	0.474 mg/L		0.0008	0.474 mg/L	0.0008	0.18%
Pbt	5217.4	0.975 mg/L		0.0026	0.975 mg/L	0.0026	0.27%
Tlr	3007.2	1.01 mg/L		0.002	1.01 mg/L	0.002	0.16%
Vf	149310.1	0.924 mg/L		0.0012	0.924 mg/L	0.0012	0.13%
Znf	49981.5	0.982 mg/L		0.0035	0.982 mg/L	0.0035	0.36%

Sequence No.: 56
 Sample ID: LCSD
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 67
 Date Collected: 8/13/2007 22:31:30
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: LCSD

Analyte	Mean Corrected			Std.Dev.	Sample		RSD
	Intensity	Conc.	Calib Units		Conc.	Units	
Sca	261723.2	101	%	0.4			0.35%
Yr	240493.3	92.2	%	0.25			0.27%
B_t	12161.4	0.460	mg/L	0.0020	0.460	mg/L	0.43%
Bat	73529.3	0.899	mg/L	0.0057	0.899	mg/L	0.63%
Bet	153512.9	0.0468	mg/L	0.00029	0.0468	mg/L	0.62%
Cat	488655.3	49.6	mg/L	0.03	49.6	mg/L	0.07%
Cdt	4932.8	0.215	mg/L	0.0014	0.215	mg/L	0.63%
Cot	21933.0	0.967	mg/L	0.0086	0.967	mg/L	0.89%
Crt	63988.8	0.911	mg/L	0.0068	0.911	mg/L	0.74%
Cut	386002.9	0.904	mg/L	0.0066	0.904	mg/L	0.73%
Fet	9200.8	5.04	mg/L	0.067	5.04	mg/L	1.33%
Kt	25750.3	19.4	mg/L	0.29	19.4	mg/L	1.49%
Mgt	84367.7	20.8	mg/L	0.37	20.8	mg/L	1.79%
Mot	13462.3	0.902	mg/L	0.0066	0.902	mg/L	0.73%
Nat	168201.3	50.0	mg/L	0.30	50.0	mg/L	0.60%
Nit	10315.7	0.474	mg/L	0.0040	0.474	mg/L	0.83%
Pbt	5262.4	0.984	mg/L	0.0087	0.984	mg/L	0.89%
Tlt	3036.1	1.02	mg/L	0.009	1.02	mg/L	0.87%
Vt	146019.1	0.904	mg/L	0.0059	0.904	mg/L	0.66%
Znt	49080.3	0.964	mg/L	0.0070	0.964	mg/L	0.73%

Sequence No.: 57
 Sample ID: 2708020327_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 68
 Date Collected: 8/13/2007 22:34:33
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020327_2X

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020327_2X

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	241027.7		93.5 %	0.68			0.73%
Yr	231542.9		88.8 %	0.33			0.37%
B_t	55600.5		2.11 mg/L	0.007	4.22 mg/L	0.014	0.33%
Bat	591.9		0.00724 mg/L	0.000014	0.0145 mg/L	0.00003	0.19%
Bet	-1307.5		-0.00040 mg/L	0.000029	-0.00080 mg/L	0.000057	7.16%
Car	2311569.3		235 mg/L	0.3	469 mg/L	0.6	0.14%
Cdt	27.5		0.00118 mg/L	0.000231	0.00236 mg/L	0.000462	19.59%
Cof	-2.6		-0.00011 mg/L	0.000100	-0.00023 mg/L	0.000200	88.06%
Crt	33973.4		0.484 mg/L	0.0015	0.968 mg/L	0.0030	0.31%
Cut	473.9		0.00111 mg/L	0.000111	0.00222 mg/L	0.000223	10.04%
Fet	7.9		0.00432 mg/L	0.002404	0.00864 mg/L	0.004808	55.64%
Kt	14223.3		10.7 mg/L	0.04	21.5 mg/L	0.07	0.33%
Mgt	545822.9		134 mg/L	0.1	269 mg/L	0.2	0.08%
Mot	458.4		0.0307 mg/L	0.00047	0.0615 mg/L	0.00093	1.52%
Nat	1535923.3		456 mg/L	0.1	913 mg/L	0.3	0.03%
Nit	1.4		0.00006 mg/L	0.000469	0.00013 mg/L	0.000938	736.60%
Pbt	-39.3		-0.00735 mg/L	0.000047	-0.0147 mg/L	0.00009	0.64%
Tlt	55.9		0.0188 mg/L	0.00101	0.0377 mg/L	0.00202	5.36%
Vt	3953.2		0.0270 mg/L	0.00027	0.0540 mg/L	0.00054	1.00%
Znt	326.4		0.00643 mg/L	0.000016	0.0129 mg/L	0.00003	0.25%

Sequence No.: 58
 Sample ID: 2708020327_2XMS
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 69
 Date Collected: 8/13/2007 22:38:57
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020327_2XMS

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: 2708020327_2XMS

Analyte	Mean Corrected		Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units		Conc. Units	Units		
Sca	239056.5	92.7 %	0.51				0.55%
Yr	235407.5	90.3 %	0.19				0.21%
B_t	62274.7	2.36 mg/L	0.005	4.73 mg/L		0.009	0.20%
Bat	37853.2	0.463 mg/L	0.0004	0.926 mg/L		0.0008	0.08%
Bet	75492.4	0.0230 mg/L	0.00003	0.0460 mg/L		0.00005	0.12%
Cat	2585423.8	263 mg/L	0.8	525 mg/L		1.6	0.31%
Cdt	2592.4	0.113 mg/L	0.0013	0.225 mg/L		0.0026	1.13%
Cot	10737.7	0.474 mg/L	0.0048	0.947 mg/L		0.0096	1.02%
Crt	66050.6	0.941 mg/L	0.0017	1.88 mg/L		0.003	0.18%
Cut	201255.4	0.471 mg/L	0.0019	0.943 mg/L		0.0039	0.41%
Fet	4580.3	2.51 mg/L	0.013	5.01 mg/L		0.027	0.53%
Kt	28293.0	21.4 mg/L	0.14	42.7 mg/L		0.29	0.67%
Mgt	594567.3	146 mg/L	0.2	293 mg/L		0.3	0.11%
Mot	7225.1	0.484 mg/L	0.0052	0.969 mg/L		0.0104	1.07%
Nat	1654723.8	492 mg/L	1.0	983 mg/L		2.0	0.21%
Nit	5006.3	0.230 mg/L	0.0016	0.460 mg/L		0.0032	0.71%
Pbt	2485.1	0.464 mg/L	0.0053	0.929 mg/L		0.0106	1.14%
Tlt	1452.1	0.489 mg/L	0.0023	0.978 mg/L		0.0047	0.48%
Vt	78545.1	0.489 mg/L	0.0004	0.977 mg/L		0.0008	0.08%
Znt	25565.9	0.502 mg/L	0.0052	1.00 mg/L		0.010	1.04%

Sequence No.: 59
 Sample ID: 2708020327_2XMSD
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 70
 Date Collected: 8/13/2007 22:42:43
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020327_2XMSD
 Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020327_2XMSD

Analyte	Mean Corrected			Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.	Units		Conc.	Units		
Sca	238310.0	92.4	%	1.14				1.24%
Yr	234523.9	90.0	%	0.06				0.07%
B+	61632.9	2.34	mg/L	0.008	4.68	mg/L	0.017	0.36%
Bat	37354.6	0.457	mg/L	0.0013	0.914	mg/L	0.0026	0.29%
Bet	74572.0	0.0227	mg/L	0.00005	0.0455	mg/L	0.00009	0.20%
Cat	2549597.2	259	mg/L	0.1	518	mg/L	0.3	0.05%
Cdt	2572.6	0.112	mg/L	0.0012	0.224	mg/L	0.0023	1.03%
Cot	10618.0	0.468	mg/L	0.0052	0.937	mg/L	0.0104	1.11%
Crt	65380.2	0.931	mg/L	0.0005	1.86	mg/L	0.001	0.05%
Cut	199147.7	0.467	mg/L	0.0013	0.933	mg/L	0.0027	0.29%
Fet	4555.8	2.49	mg/L	0.042	4.99	mg/L	0.083	1.67%
K+	28039.9	21.2	mg/L	0.35	42.3	mg/L	0.71	1.68%
Mgt	586302.0	144	mg/L	0.1	289	mg/L	0.2	0.06%
Mot	7155.7	0.480	mg/L	0.0068	0.959	mg/L	0.0136	1.42%
Nat	1620502.7	482	mg/L	0.4	963	mg/L	0.8	0.08%
Nit	4962.3	0.228	mg/L	0.0033	0.456	mg/L	0.0066	1.45%
Pbt	2465.2	0.461	mg/L	0.0054	0.921	mg/L	0.0108	1.17%
Tlt	1448.1	0.487	mg/L	0.0036	0.975	mg/L	0.0071	0.73%
Vt	77770.3	0.484	mg/L	0.0006	0.968	mg/L	0.0012	0.13%
Znt	25323.3	0.497	mg/L	0.0065	0.995	mg/L	0.0130	1.31%

Sequence No.: 60
 Sample ID: 2708020332_20X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 71
 Data Collected: 8/13/2007 22:46:27
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020332_20X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020332_20X

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	263791.1	102 %		1.1			1.09%
Yr	238860.0	91.6 %		0.03			0.03%
B ₊ t	14658.5	0.556 mg/L		0.0002	11.1 mg/L	0.00	0.03%
Ba _t	176.4	0.00216 mg/L		0.000031	0.0432 mg/L	0.00062	1.44%
Be _t	-781.6	-0.00024 mg/L		0.000024	-0.00477 mg/L	0.000474	9.95%
Ca _t	454755.7	46.2 mg/L		0.34	924 mg/L	6.9	0.75%
Cd _t	-20.6	-0.00088 mg/L		0.000050	-0.0177 mg/L	0.00101	5.70%
Co _t	1.2	0.00005 mg/L		0.000213	0.00107 mg/L	0.004267	397.17%
Crt	74073.2	1.06 mg/L		0.002	21.1 mg/L	0.03	0.16%
Cu _t	-7.4	-0.00002 mg/L		0.000053	-0.00035 mg/L	0.001059	305.20%
Fe _t	4.4	0.00243 mg/L		0.001320	0.0486 mg/L	0.02640	54.32%
K _t	4301.8	3.25 mg/L		0.000	64.9 mg/L	0.01	0.01%
Mg _t	106394.3	26.2 mg/L		0.18	524 mg/L	3.6	0.69%
Mo _t	26.8	0.00180 mg/L		0.000068	0.0359 mg/L	0.00136	3.78%
Na _t	321155.7	95.4 mg/L		0.69	1910 mg/L	13.9	0.73%
Ni _t	-0.9	-0.00004 mg/L		0.000173	-0.00082 mg/L	0.003470	420.81%
Pb _t	-12.6	-0.00236 mg/L		0.000118	-0.0471 mg/L	0.00236	5.00%
Tl _t	26.9	0.00903 mg/L		0.001706	0.181 mg/L	0.0341	18.90%
V _t	-432.3	0.00316 mg/L		0.000104	0.0632 mg/L	0.00207	3.27%
Zn _t	227.2	0.00448 mg/L		0.000119	0.0895 mg/L	0.00238	2.65%

Sequence No.: 61
 Sample ID: 2708020332_20XMS
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 72
 Date Collected: 8/13/2007 22:50:49
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020332_20XMS

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020332_20XMS

Analyte	Mean Corrected			Sample	Std.Dev.	RSD
	Intensity	Conc.	Units			
Sca	262239.1	102	%		2.0	1.97%
Yr	242720.5	93.1	%		0.93	0.99%
B_f	14564.2	0.553	mg/L	11.1	0.0035	0.63%
Bat	3824.5	0.0468	mg/L	0.936	0.00086	1.83%
Bet	6926.3	0.00211	mg/L	0.0422	0.000005	0.26%
Cat	458178.9	46.5	mg/L	930	0.00	0.01%
Cdt	220.6	0.00962	mg/L	0.192	0.000017	0.18%
Cot	1074.4	0.0474	mg/L	0.948	0.00076	1.60%
Crt	74769.3	1.07	mg/L	21.3	0.006	0.52%
Cut	18707.7	0.0438	mg/L	0.877	0.00023	0.53%
Fet	461.3	0.253	mg/L	5.05	0.0031	1.23%
Kf	5428.1	4.10	mg/L	81.9	0.040	0.99%
Mgt	105471.3	26.0	mg/L	519	0.03	0.10%
Mot	687.1	0.0461	mg/L	0.921	0.00128	2.78%
Nat	315375.8	93.7	mg/L	1870	0.29	0.31%
Nit	494.2	0.0227	mg/L	0.454	0.00043	1.89%
Pbt	247.9	0.0463	mg/L	0.927	0.00169	3.65%
Tlt	172.1	0.0579	mg/L	1.16	0.00241	4.16%
Vt	6813.2	0.0478	mg/L	0.956	0.00029	0.60%
Znt	2636.9	0.0518	mg/L	1.04	0.00135	2.61%

Sequence No.: 62
 Sample ID: 2708020332_20XMSD
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 73
 Date Collected: 8/13/2007 22:55:08
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020332_20XMSD

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020332_20XMSD

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	262040.9	102 %		2.3			2.22%
Yr	242536.0	93.0 %		1.21			1.30%
B_†	14890.3	0.565 mg/L		0.0026	11.3 mg/L	0.05	0.46%
Ba†	3937.6	0.0482 mg/L		0.00128	0.963 mg/L	0.0256	2.66%
Be†	7171.3	0.00219 mg/L		0.000018	0.0437 mg/L	0.00036	0.82%
Ca†	470576.8	47.8 mg/L		0.16	956 mg/L	3.2	0.34%
Cd†	226.2	0.00986 mg/L		0.000395	0.197 mg/L	0.0079	4.00%
Co†	1102.2	0.0486 mg/L		0.00149	0.972 mg/L	0.0298	3.07%
Cr†	77096.5	1.10 mg/L		0.006	22.0 mg/L	0.12	0.53%
Cu†	19653.7	0.0460 mg/L		0.00013	0.921 mg/L	0.0025	0.27%
Fe†	469.5	0.257 mg/L		0.0043	5.14 mg/L	0.086	1.68%
K†	5473.4	4.13 mg/L		0.095	82.6 mg/L	1.90	2.30%
Mg†	108063.4	26.6 mg/L		0.09	532 mg/L	1.7	0.33%
Mo†	713.7	0.0478 mg/L		0.00172	0.957 mg/L	0.0345	3.60%
Na†	322542.6	95.8 mg/L		0.03	1920 mg/L	0.6	0.03%
Ni†	510.7	0.0234 mg/L		0.00104	0.469 mg/L	0.0207	4.42%
Pb†	259.3	0.0485 mg/L		0.00082	0.969 mg/L	0.0165	1.70%
Tl†	175.0	0.0589 mg/L		0.00129	1.18 mg/L	0.026	2.18%
V†	7060.3	0.0495 mg/L		0.00012	0.990 mg/L	0.0024	0.25%
Zn†	2707.5	0.0532 mg/L		0.00129	1.06 mg/L	0.026	2.42%

Sequence No.: 63
 Sample ID: 2708020320_20X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 74
 Date Collected: 8/13/2007 22:59:32
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020320_20X

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020320_20X

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	262113.8	102 %		2.0			1.99%
Yr	235332.8	90.3 %		2.93			3.25%
B_+	3770.2	0.143 mg/L		0.0003	2.86 mg/L	0.005	0.18%
Ba+	173.6	0.00212 mg/L		0.000042	0.0425 mg/L	0.00085	2.00%
Be+	-904.7	-0.00028 mg/L		0.000012	-0.00552 mg/L	0.000233	4.22%
Ca+	267768.3	27.2 mg/L		0.03	544 mg/L	0.5	0.10%
Cd+	-12.0	-0.00051 mg/L		0.000174	-0.0102 mg/L	0.00348	33.96%
Co+	12.0	0.00053 mg/L		0.000277	0.0106 mg/L	0.00554	52.20%
Cr+	103344.6	1.47 mg/L		0.002	29.4 mg/L	0.05	0.16%
Cu+	-36.9	-0.00009 mg/L		0.000118	-0.00172 mg/L	0.002358	137.39%
Fe+	93.2	0.0510 mg/L		0.00089	1.02 mg/L	0.018	1.74%
K+	2330.0	1.76 mg/L		0.039	35.2 mg/L	0.78	2.23%
Mg+	59479.1	14.6 mg/L		0.43	293 mg/L	8.5	2.92%
Mo+	31.5	0.00211 mg/L		0.000165	0.0423 mg/L	0.00330	7.80%
Na+	499487.0	148 mg/L		0.3	2970 mg/L	5.8	0.20%
Ni+	-1.0	-0.00005 mg/L		0.000236	-0.00091 mg/L	0.004726	519.85%
Pb+	-1.2	-0.00022 mg/L		0.000735	-0.00433 mg/L	0.014707	339.79%
Ti+	12.6	0.00423 mg/L		0.001986	0.0846 mg/L	0.03972	46.95%
V+	-959.4	0.00222 mg/L		0.000090	0.0443 mg/L	0.00180	4.05%
Zn+	115.6	0.00228 mg/L		0.000020	0.0456 mg/L	0.00041	0.90%

Sequence No.: 64
 Sample ID: 2708020321_20X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 75
 Date Collected: 8/13/2007 23:03:54
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020321_20X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020321_20X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	259618.8	101	%	1.2				1.21%
Yr	235422.3	90.3	%	0.44				0.49%
B ₁	10494.7	0.398	mg/L	0.0001	7.97	mg/L	0.001	0.01%
Ba ₁	129.2	0.00158	mg/L	0.000069	0.0316	mg/L	0.00139	4.39%
Be ₁	-593.4	-0.00018	mg/L	0.000019	-0.00362	mg/L	0.000372	10.27%
Ca ₁	372289.5	37.8	mg/L	0.04	756	mg/L	0.8	0.11%
Cd ₁	-18.9	-0.00081	mg/L	0.000000	-0.0162	mg/L	0.00001	0.04%
Co ₁	-0.6	-0.00003	mg/L	0.000218	-0.00056	mg/L	0.004355	779.29%
Crt	55917.2	0.796	mg/L	0.0001	15.9	mg/L	0.00	0.01%
Cu ₁	93.5	0.00022	mg/L	0.000125	0.00438	mg/L	0.002509	57.33%
Fe ₁	-1.3	-0.00073	mg/L	0.000936	-0.0146	mg/L	0.01873	128.12%
K ₁	2883.5	2.18	mg/L	0.032	43.5	mg/L	0.64	1.47%
Mg ₁	79644.1	19.6	mg/L	0.06	392	mg/L	1.1	0.28%
Mo ₁	28.7	0.00193	mg/L	0.000463	0.0385	mg/L	0.00927	24.06%
Nat	288004.7	85.6	mg/L	0.01	1710	mg/L	0.3	0.02%
Ni ₁	-6.8	-0.00031	mg/L	0.000511	-0.00628	mg/L	0.010224	162.77%
Pb ₁	-7.1	-0.00132	mg/L	0.000221	-0.0264	mg/L	0.00442	16.72%
Tl ₁	22.0	0.00737	mg/L	0.000336	0.147	mg/L	0.0067	4.56%
V ₁	-199.3	0.00317	mg/L	0.000056	0.0634	mg/L	0.00111	1.75%
Zn ₁	118.2	0.00233	mg/L	0.000108	0.0466	mg/L	0.00216	4.63%

Sequence No.: 65
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/13/2007 23:08:15
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	255060.2	98.9 %	%	2.25			2.27%
Yr	238126.6	91.3 %	%	0.83			0.91%
B_t	61303.3	2.32 mg/L	mg/L	0.001	2.32 mg/L	0.001	0.03%
	QC value within limits for B_ Recovery = 92.68%						
Bat	384419.9	4.70 mg/L	mg/L	0.025	4.70 mg/L	0.025	0.53%
	QC value within limits for Ba Recovery = 94.05%						
Be†	6159058.1	1.88 mg/L	mg/L	0.025	1.88 mg/L	0.025	1.32%
	QC value within limits for Be Recovery = 93.91%						
Ca†	511362.4	51.9 mg/L	mg/L	0.02	51.9 mg/L	0.02	0.05%
	QC value within limits for Ca Recovery = 103.85%						
Cd†	57680.7	2.49 mg/L	mg/L	0.045	2.49 mg/L	0.045	1.80%
	QC value within limits for Cd Recovery = 99.46%						
Co†	111920.3	4.94 mg/L	mg/L	0.077	4.94 mg/L	0.077	1.56%
	QC value within limits for Co Recovery = 98.74%						
Cr†	332241.9	4.73 mg/L	mg/L	0.020	4.73 mg/L	0.020	0.42%
	QC value within limits for Cr Recovery = 94.65%						
Cu†	1928918.5	4.52 mg/L	mg/L	0.005	4.52 mg/L	0.005	0.11%
	QC value within limits for Cu Recovery = 90.38%						
Fe†	9289.4	5.08 mg/L	mg/L	0.014	5.08 mg/L	0.014	0.27%
	QC value within limits for Fe Recovery = 101.70%						
K†	66050.3	49.8 mg/L	mg/L	0.25	49.8 mg/L	0.25	0.49%
	QC value within limits for K Recovery = 99.69%						
Mg†	216989.8	53.4 mg/L	mg/L	0.06	53.4 mg/L	0.06	0.12%
	QC value within limits for Mg Recovery = 106.82%						
Mo†	68978.9	4.62 mg/L	mg/L	0.078	4.62 mg/L	0.078	1.68%
	QC value within limits for Mo Recovery = 92.48%						
Na†	173366.4	51.5 mg/L	mg/L	0.15	51.5 mg/L	0.15	0.29%
	QC value within limits for Na Recovery = 103.03%						
Ni†	106118.6	4.87 mg/L	mg/L	0.077	4.87 mg/L	0.077	1.58%
	QC value within limits for Ni Recovery = 97.43%						
Pb†	26236.5	4.90 mg/L	mg/L	0.085	4.90 mg/L	0.085	1.74%
	QC value within limits for Pb Recovery = 98.07%						
Tl†	15094.8	5.08 mg/L	mg/L	0.100	5.08 mg/L	0.100	1.96%
	QC value within limits for Tl Recovery = 101.60%						
V†	761801.9	4.72 mg/L	mg/L	0.017	4.72 mg/L	0.017	0.35%
	QC value within limits for V Recovery = 94.31%						
Zn†	250101.9	4.89 mg/L	mg/L	0.016	4.89 mg/L	0.016	0.33%
	QC value within limits for Zn Recovery = 97.90%						
All analyte(s) passed QC.							

Sequence No.: 66
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 8/13/2007 23:11:30
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Sca	267376.9	104	%	0.4				0.41%
Yr	239780.1	92.0	%	1.15				1.25%
B ₊	436.4	0.0166	mg/L	0.00066	0.0166	mg/L	0.00066	3.98%
	QC value within limits for B ₊ Recovery = Not calculated							
Ba ₊	-10.0	-0.00012	mg/L	0.000022	-0.00012	mg/L	0.000022	17.71%
	QC value within limits for Ba ₊ Recovery = Not calculated							
Be ₊	-33.9	-0.00001	mg/L	0.000025	-0.00001	mg/L	0.000025	237.29%
	QC value within limits for Be ₊ Recovery = Not calculated							
Ca ₊	31.8	0.00323	mg/L	0.001265	0.00323	mg/L	0.001265	39.19%
	QC value within limits for Ca ₊ Recovery = Not calculated							
Cd ₊	0.7	0.00003	mg/L	0.000191	0.00003	mg/L	0.000191	611.25%
	QC value within limits for Cd ₊ Recovery = Not calculated							
Co ₊	-3.9	-0.00017	mg/L	0.000202	-0.00017	mg/L	0.000202	117.61%
	QC value within limits for Co ₊ Recovery = Not calculated							
Cr ₊	154.4	0.00220	mg/L	0.000126	0.00220	mg/L	0.000126	5.73%
	QC value within limits for Cr ₊ Recovery = Not calculated							
Cu ₊	-35.2	-0.00008	mg/L	0.000222	-0.00008	mg/L	0.000222	268.75%
	QC value within limits for Cu ₊ Recovery = Not calculated							
Fe ₊	-2.3	-0.00124	mg/L	0.000848	-0.00124	mg/L	0.000848	68.32%
	QC value within limits for Fe ₊ Recovery = Not calculated							
K ₊	-56.1	-0.0423	mg/L	0.02832	-0.0423	mg/L	0.02832	66.90%
	QC value within limits for K ₊ Recovery = Not calculated							
Mg ₊	-0.0	-0.00001	mg/L	0.000170	-0.00001	mg/L	0.000170	>999.9%
	QC value within limits for Mg ₊ Recovery = Not calculated							
Mo ₊	12.6	0.00085	mg/L	0.000218	0.00085	mg/L	0.000218	25.75%
	QC value within limits for Mo ₊ Recovery = Not calculated							
Na ₊	538.7	0.160	mg/L	0.0011	0.160	mg/L	0.0011	0.68%
	QC value within limits for Na ₊ Recovery = Not calculated							
Ni ₊	-5.1	-0.00023	mg/L	0.000302	-0.00023	mg/L	0.000302	129.34%
	QC value within limits for Ni ₊ Recovery = Not calculated							
Pb ₊	5.8	0.00108	mg/L	0.000262	0.00108	mg/L	0.000262	24.21%
	QC value within limits for Pb ₊ Recovery = Not calculated							
Tl ₊	2.2	0.00073	mg/L	0.000805	0.00073	mg/L	0.000805	110.18%
	QC value within limits for Tl ₊ Recovery = Not calculated							
V ₊	-7.8	-0.00004	mg/L	0.000332	-0.00004	mg/L	0.000332	930.84%
	QC value within limits for V ₊ Recovery = Not calculated							
Zn ₊	-7.9	-0.00015	mg/L	0.000019	-0.00015	mg/L	0.000019	12.31%
	QC value within limits for Zn ₊ Recovery = Not calculated							

All analyte(s) passed QC.

Sequence No.: 67
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 8/13/2007 23:14:55
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	265143.1	103	%	0.7			0.67%
Yr	240774.4	92.4	%	0.25			0.27%
B_t	30337.9	1.15	mg/L	0.000	1.15 mg/L	0.000	0.01%
	QC value within limits for B_ Recovery = 91.73%						
Bat	193030.8	2.36	mg/L	0.007	2.36 mg/L	0.007	0.29%
	QC value within limits for Ba Recovery = 94.46%						
Be†	3055934.9	0.932	mg/L	0.0018	0.932 mg/L	0.0018	0.19%
	QC value within limits for Be Recovery = 93.19%						
Ca†	258177.9	26.2	mg/L	0.04	26.2 mg/L	0.04	0.15%
	QC value within limits for Ca Recovery = 104.86%						
Cd†	28695.6	1.24	mg/L	0.001	1.24 mg/L	0.001	0.09%
	QC value within limits for Cd Recovery = 98.96%						
Co†	55554.3	2.45	mg/L	0.007	2.45 mg/L	0.007	0.28%
	QC value within limits for Co Recovery = 98.02%						
Cr†	165594.2	2.36	mg/L	0.005	2.36 mg/L	0.005	0.20%
	QC value within limits for Cr Recovery = 94.35%						
Cu†	954512.5	2.24	mg/L	0.013	2.24 mg/L	0.013	0.57%
	QC value less than the lower limit for Cu Recovery = 89.45%						
Fe†	4713.7	2.58	mg/L	0.025	2.58 mg/L	0.025	0.97%
	QC value within limits for Fe Recovery = 103.21%						
K†	32831.4	24.8	mg/L	0.07	24.8 mg/L	0.07	0.28%
	QC value within limits for K Recovery = 99.11%						
Mg†	110042.9	27.1	mg/L	0.04	27.1 mg/L	0.04	0.13%
	QC value within limits for Mg Recovery = 108.35%						
Mo†	33949.6	2.28	mg/L	0.001	2.28 mg/L	0.001	0.07%
	QC value within limits for Mo Recovery = 91.04%						
Na†	87123.8	25.9	mg/L	0.16	25.9 mg/L	0.16	0.63%
	QC value within limits for Na Recovery = 103.56%						
Ni†	53627.2	2.46	mg/L	0.004	2.46 mg/L	0.004	0.18%
	QC value within limits for Ni Recovery = 98.47%						
Pb†	13144.2	2.46	mg/L	0.006	2.46 mg/L	0.006	0.25%
	QC value within limits for Pb Recovery = 98.27%						
Tl†	7622.0	2.57	mg/L	0.005	2.57 mg/L	0.005	0.21%
	QC value within limits for Tl Recovery = 102.60%						
V†	378320.9	2.34	mg/L	0.002	2.34 mg/L	0.002	0.10%
	QC value within limits for V Recovery = 93.68%						
Zn†	125530.1	2.46	mg/L	0.007	2.46 mg/L	0.007	0.27%
	QC value within limits for Zn Recovery = 98.27%						
QC Failed. Retry.							

Sequence No.: 68
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 8/13/2007 23:16:29
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	260760.4	101 %		0.9			0.91%
Yr	242020.1	92.8 %		0.56			0.60%
B ₋ t	30642.8	1.16 mg/L		0.005	1.16 mg/L	0.005	0.47%
	QC value within limits for B ₋ Recovery = 92.66%						
Ba†	193983.8	2.37 mg/L		0.008	2.37 mg/L	0.008	0.35%
	QC value within limits for Ba Recovery = 94.92%						
Be†	3063766.1	0.934 mg/L		0.0039	0.934 mg/L	0.0039	0.41%
	QC value within limits for Be Recovery = 93.43%						
Ca†	258980.5	26.3 mg/L		0.00	26.3 mg/L	0.00	0.01%
	QC value within limits for Ca Recovery = 105.19%						
Cd†	28810.0	1.24 mg/L		0.030	1.24 mg/L	0.030	2.38%
	QC value within limits for Cd Recovery = 99.35%						
Co†	55639.1	2.45 mg/L		0.007	2.45 mg/L	0.007	0.29%
	QC value within limits for Co Recovery = 98.17%						
Cr†	165824.8	2.36 mg/L		0.013	2.36 mg/L	0.013	0.54%
	QC value within limits for Cr Recovery = 94.48%						
Cu†	957249.2	2.24 mg/L		0.002	2.24 mg/L	0.002	0.07%
	QC value less than the lower limit for Cu Recovery = 89.70%						
Fe†	4828.8	2.64 mg/L		0.002	2.64 mg/L	0.002	0.08%
	QC value within limits for Fe Recovery = 105.73%						
K†	33740.5	25.5 mg/L		0.10	25.5 mg/L	0.10	0.41%
	QC value within limits for K Recovery = 101.85%						
Mg†	110394.1	27.2 mg/L		0.03	27.2 mg/L	0.03	0.11%
	QC value within limits for Mg Recovery = 108.69%						
Mo†	34154.9	2.29 mg/L		0.046	2.29 mg/L	0.046	2.03%
	QC value within limits for Mo Recovery = 91.59%						
Na†	86965.2	25.8 mg/L		0.08	25.8 mg/L	0.08	0.29%
	QC value within limits for Na Recovery = 103.37%						
Ni†	53807.3	2.47 mg/L		0.055	2.47 mg/L	0.055	2.22%
	QC value within limits for Ni Recovery = 98.80%						
Pb†	13239.8	2.47 mg/L		0.049	2.47 mg/L	0.049	1.98%
	QC value within limits for Pb Recovery = 98.98%						
Tl†	7630.8	2.57 mg/L		0.038	2.57 mg/L	0.038	1.49%
	QC value within limits for Tl Recovery = 102.72%						
V†	379103.6	2.35 mg/L		0.010	2.35 mg/L	0.010	0.43%
	QC value within limits for V Recovery = 93.87%						
Zn†	126264.2	2.47 mg/L		0.009	2.47 mg/L	0.009	0.36%
	QC value within limits for Zn Recovery = 98.84%						
QC Failed. Continue with analysis.							

Sequence No.: 69
 Sample ID: 2708020322_10X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 10X

Autosampler Location: 76
 Date Collected: 8/13/2007 23:19:39
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020322_10X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020322_10X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	258164.8	100	%	1.0				1.00%
Yr	237745.6	91.2	%	0.50				0.55%
B_t	52161.7	1.98	mg/L	0.000	19.8	mg/L	0.00	0.00%
Bat	241.7	0.00296	mg/L	0.000040	0.0296	mg/L	0.000040	1.35%
Bet	-746.9	-0.00023	mg/L	0.000005	-0.00228	mg/L	0.000049	2.15%
Cat	391554.8	39.8	mg/L	0.23	398	mg/L	2.3	0.57%
Cdt	-14.7	-0.00063	mg/L	0.000157	-0.00630	mg/L	0.001565	24.83%
Cot	2.6	0.00011	mg/L	0.000069	0.00113	mg/L	0.000688	61.07%
Crt	66333.2	0.945	mg/L	0.0005	9.45	mg/L	0.005	0.06%
Cut	174.9	0.00041	mg/L	0.000184	0.00409	mg/L	0.001841	44.95%
Fer	-1.8	-0.00099	mg/L	0.000105	-0.00989	mg/L	0.001049	10.61%
Kt	10884.2	8.21	mg/L	0.041	82.1	mg/L	0.41	0.49%
Mgt	103426.6	25.5	mg/L	0.35	255	mg/L	3.5	1.38%
Mot	89.4	0.00599	mg/L	0.000501	0.0599	mg/L	0.00501	8.35%
Nat	440544.5	131	mg/L	0.4	1310	mg/L	4.1	0.31%
Nit	-2.3	-0.00010	mg/L	0.000026	-0.00104	mg/L	0.000263	25.23%
Pbt	-7.2	-0.00134	mg/L	0.001093	-0.0134	mg/L	0.01093	81.46%
Tlt	24.0	0.00806	mg/L	0.000643	0.0806	mg/L	0.00643	7.98%
Vt	55.3	0.00555	mg/L	0.000048	0.0555	mg/L	0.00048	0.86%
Znt	97.2	0.00192	mg/L	0.000200	0.0192	mg/L	0.00200	10.46%

Sequence No.: 70
 Sample ID: 2708020323_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 77
 Date Collected: 8/13/2007 23:24:02
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020323_2X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020323_2X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	248057.1	96.2	%	1.24				1.29%
Yr	237349.0	91.0	%	0.83				0.91%
B_t	16701.9	0.634	mg/L	0.0044	1.27	mg/L	0.009	0.70%
Bat	1043.2	0.7128	mg/L	0.00012	0.0255	mg/L	0.00024	0.94%
Bet	-902.3	-0.00028	mg/L	0.000002	-0.00055	mg/L	0.000004	0.69%
Ca†	1510046.4	153	mg/L	0.3	307	mg/L	0.6	0.21%
Cdt	31.9	0.00138	mg/L	0.000062	0.00275	mg/L	0.000124	4.51%
Cot	73.1	0.00323	mg/L	0.000278	0.00645	mg/L	0.000556	8.61%
Crt	13384.9	0.191	mg/L	0.0011	0.381	mg/L	0.0023	0.60%
Cut	623.8	0.00146	mg/L	0.000116	0.00293	mg/L	0.000233	7.95%
Fet	27.6	0.0151	mg/L	0.00073	0.0302	mg/L	0.00146	4.83%
K†	7208.3	5.44	mg/L	0.083	10.9	mg/L	0.17	1.53%
Mgt	320630.7	78.9	mg/L	0.18	158	mg/L	0.4	0.22%
Mof	1207.2	0.0809	mg/L	0.00100	0.162	mg/L	0.0020	1.23%
Nat	1890109.5	562	mg/L	0.4	1120	mg/L	0.7	0.06%
Nit	51.0	0.00234	mg/L	0.000528	0.00468	mg/L	0.001057	22.59%
Pbr	-33.1	-0.00619	mg/L	0.001471	-0.0124	mg/L	0.00294	23.78%
Tlt	57.7	0.0194	mg/L	0.00301	0.0388	mg/L	0.00602	15.50%
V†	3613.4	0.0233	mg/L	0.00006	0.0466	mg/L	0.00012	0.25%
Zn†	451.2	0.00887	mg/L	0.000109	0.0177	mg/L	0.00022	1.23%

Sequence No.: 71
 Sample ID: 2708020324_5X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 5X

Autosampler Location: 78
 Date Collected: 8/13/2007 23:28:28
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020324_5X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020324_5X

Analyte	Mean Corrected			Std.Dev.	Sample			RSD
	Intensity	Conc.	Calib Units		Conc.	Units	Std.Dev.	
Sca	254317.5	98.6	%	1.34				1.35%
Yr	232480.1	89.2	%	0.74				0.83%
B_†	53135.3	2.02	mg/L	0.002	10.1	mg/L	0.01	0.10%
Ba†	255.6	0.00313	mg/L	0.000115	0.0156	mg/L	0.00058	3.69%
Be†	-1021.6	-0.00031	mg/L	0.000036	-0.00156	mg/L	0.000178	11.40%
Ca†	1328456.4	135	mg/L	0.9	674	mg/L	4.3	0.63%
Cd†	-16.2	-0.00069	mg/L	0.000339	-0.00347	mg/L	0.001696	48.84%
Co†	-8.3	-0.00037	mg/L	0.000075	-0.00184	mg/L	0.000375	20.39%
Cr†	64298.3	0.916	mg/L	0.0007	4.58	mg/L	0.004	0.08%
Cu†	268.6	0.00063	mg/L	0.000092	0.00314	mg/L	0.000461	14.68%
Fe†	22.3	0.0122	mg/L	0.00163	0.0612	mg/L	0.00814	13.31%
K†	7145.1	5.39	mg/L	0.033	27.0	mg/L	0.17	0.62%
Mg†	357896.5	88.1	mg/L	0.75	440	mg/L	3.7	0.85%
Mn†	65.7	0.00440	mg/L	0.000039	0.0220	mg/L	0.00019	0.88%
Na†	602675.1	179	mg/L	1.2	895	mg/L	6.1	0.68%
Ni†	-13.1	-0.00060	mg/L	0.000392	-0.00300	mg/L	0.001959	65.31%
Pb†	-22.6	-0.00422	mg/L	0.001234	-0.0211	mg/L	0.00617	29.28%
Tl†	33.1	0.0111	mg/L	0.00088	0.0557	mg/L	0.00442	7.93%
V†	965.3	0.0110	mg/L	0.00022	0.0550	mg/L	0.00109	1.98%
Zn†	255.9	0.00505	mg/L	0.000102	0.0252	mg/L	0.00051	2.02%

Sequence No.: 72
 Sample ID: 2708020325_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 79
 Date Collected: 8/13/2007 23:32:52
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020325_2X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020325_2X

Analyte	Mean Corrected			Std.Dev.	Sample			RSD
	Intensity	Conc.	Units		Conc.	Units	Std.Dev.	
Sca	242481.5	94.0	%	0.19				0.21%
Yr	230345.4	88.4	%	0.74				0.83%
B ₊	57317.5	2.18	mg/L	0.002	4.35	mg/L	0.004	0.08%
Ba _f	700.6	0.00857	mg/L	0.000149	0.0171	mg/L	0.00030	1.74%
Be _f	-1208.4	-0.00037	mg/L	0.000019	-0.00074	mg/L	0.000037	5.03%
Cat	2568463.8	261	mg/L	1.8	522	mg/L	3.7	0.70%
Cdr	16.9	0.00072	mg/L	0.000090	0.00145	mg/L	0.000180	12.40%
Cof	-3.1	-0.00014	mg/L	0.000043	-0.00027	mg/L	0.000086	31.61%
Crt	32078.3	0.457	mg/L	0.0013	0.914	mg/L	0.0026	0.28%
Cut	781.7	0.00183	mg/L	0.000336	0.00366	mg/L	0.000672	18.36%
Fef	184.0	0.101	mg/L	0.0039	0.201	mg/L	0.0078	3.88%
Kf	17423.3	13.1	mg/L	0.06	26.3	mg/L	0.11	0.43%
Mgf	636719.5	157	mg/L	1.2	313	mg/L	2.4	0.76%
Mof	709.1	0.0475	mg/L	0.00093	0.0951	mg/L	0.00186	1.96%
Naf	2094504.1	622	mg/L	8.7	1240	mg/L	17.3	1.39%
Nit	17.3	0.00079	mg/L	0.000136	0.00158	mg/L	0.000273	17.23%
Pbf	-35.9	-0.00671	mg/L	0.000378	-0.0134	mg/L	0.00076	5.63%
Tlf	49.0	0.0165	mg/L	0.00160	0.0329	mg/L	0.00320	9.71%
Vf	2263.4	0.0165	mg/L	0.00003	0.0329	mg/L	0.00007	0.21%
Znf	429.4	0.00846	mg/L	0.000345	0.0169	mg/L	0.00069	4.07%

Sequence No.: 73
 Sample ID: 2708020326_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 80
 Date Collected: 8/13/2007 23:37:18
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020326_2X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020326_2X

Analyte	Mean Corrected			Std.Dev.	Sample			RSD
	Intensity	Conc.	Units		Conc.	Units	Std.Dev.	
Sca	248374.1	96.3	%	0.45				0.47%
Yr	235162.3	90.2	%	0.28				0.31%
B_t	52887.8	2.01	mg/L	0.001	4.02	mg/L	0.003	0.07%
Bat	499.0	0.00610	mg/L	0.000074	0.0122	mg/L	0.00015	1.22%
Bef	-1214.0	-0.00037	mg/L	0.000014	-0.00074	mg/L	0.000029	3.87%
Cat	2634614.4	268	mg/L	4.1	535	mg/L	8.2	1.53%
Cdt	19.3	0.00083	mg/L	0.000112	0.00165	mg/L	0.000225	13.60%
Cof	-5.3	-0.00023	mg/L	0.000475	-0.00047	mg/L	0.000951	202.86%
Crt	35889.8	0.511	mg/L	0.0022	1.02	mg/L	0.004	0.43%
Cut	408.3	0.00096	mg/L	0.000143	0.00191	mg/L	0.000286	14.99%
Fef	30.7	0.0168	mg/L	0.00146	0.0336	mg/L	0.00291	8.66%
Kt	12758.3	9.63	mg/L	0.065	19.3	mg/L	0.13	0.68%
Mgt	557657.9	137	mg/L	2.0	275	mg/L	3.9	1.43%
Mot	312.3	0.0209	mg/L	0.00045	0.0419	mg/L	0.00091	2.17%
Nar	1418805.1	42.	mg/L	4.9	843	mg/L	9.8	1.16%
Nit	2.5	0.00012	mg/L	0.000437	0.00023	mg/L	0.000875	377.69%
Pbt	-41.1	-0.00769	mg/L	0.000693	-0.0154	mg/L	0.00139	9.01%
Tlt	49.9	0.0168	mg/L	0.00094	0.0337	mg/L	0.00188	5.57%
Vt	4727.1	0.0319	mg/L	0.00001	0.0638	mg/L	0.00003	0.05%
Znt	361.8	0.00713	mg/L	0.000048	0.0143	mg/L	0.00010	0.67%

Sequence No.: 74
 Sample ID: 2708020328_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 81
 Date Collected: 8/13/2007 23:41:42
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020328_2X

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020328_2X

Analyte	Mean Corrected			Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.	Units		Conc.	Units		
Sca	242820.0	94.2	%	1.13				1.20%
Yr	228496.3	87.6	%	0.23				0.27%
B_t	110418.0	4.19	mg/L	0.010	8.38	mg/L	0.020	0.23%
Bat	625.3	0.00765	mg/L	0.000018	0.0153	mg/L	0.000004	0.24%
Bet	-1711.7	-0.00052	mg/L	0.000007	-0.00104	mg/L	0.000014	1.37%
Cat	2905283.3	295	mg/L	2.2	590	mg/L	4.3	0.74%
Cdt	23.3	0.00100	mg/L	0.000167	0.00199	mg/L	0.000334	16.76%
Cot	-1.8	-0.00008	mg/L	0.000135	-0.00015	mg/L	0.000269	174.28%
Crt	82831.3	1.18	mg/L	0.001	2.36	mg/L	0.002	0.07%
Cut	563.0	0.00132	mg/L	0.000082	0.00264	mg/L	0.000163	6.20%
Fet	-11.5	-0.00629	mg/L	0.000279	-0.0126	mg/L	0.00056	4.44%
Kf	13518.7	10.2	mg/L	0.09	20.4	mg/L	0.18	0.90%
Mgt	677853.9	167	mg/L	1.1	334	mg/L	2.2	0.65%
Mot	223.5	0.0150	mg/L	0.00012	0.0300	mg/L	0.00024	0.81%
Naf	1515960.2	450	mg/L	2.7	901	mg/L	5.5	0.61%
Nit	16.9	0.00077	mg/L	0.000190	0.00155	mg/L	0.000381	24.61%
Pbt	-38.9	-0.00726	mg/L	0.000665	-0.0145	mg/L	0.00133	9.16%
Tlt	55.8	0.0188	mg/L	0.00281	0.0376	mg/L	0.00562	14.95%
Vf	3435.0	0.0277	mg/L	0.00001	0.0553	mg/L	0.00003	0.05%
Znt	301.7	0.00594	mg/L	0.000085	0.0119	mg/L	0.00017	1.44%

Sequence No.: 75
 Sample ID: 2708020329_5X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 5X

Autosampler Location: 82
 Date Collected: 8/13/2007 23:46:08
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020329_5X

Analyte Back Pressure Flow
 All 219.0 kPa 0.65 L/min

Mean Data: 2708020329_5X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD	
	Intensity	Conc.			Conc.	Units		Std.Dev.
Sca	254265.0	98.6	%	0.08			0.08%	
Yr	237854.9	91.2	%	0.30			0.33%	
B ₊ t	97729.9	3.71	mg/L	0.020	18.5	mg/L	0.10	0.54%
B ₊ at	415.0	0.00508	mg/L	0.000061	0.0254	mg/L	0.00030	1.20%
Be ₊ t	-1492.0	-0.00045	mg/L	0.000015	-0.00227	mg/L	0.000077	3.38%
Ca ₊ t	1352619.0	137	mg/L	0.8	687	mg/L	3.8	0.56%
Cd ₊ t	-23.9	-0.00102	mg/L	0.000022	-0.00511	mg/L	0.000109	2.13%
Co ₊ t	5.9	0.00026	mg/L	0.000135	0.00131	mg/L	0.000675	51.58%
Crt	133037.0	1.89	mg/L	0.008	9.47	mg/L	0.039	0.41%
Cu ₊ t	171.0	0.00040	mg/L	0.000097	0.00200	mg/L	0.000486	24.27%
Fe ₊ t	0.8	0.00046	mg/L	0.002363	0.00231	mg/L	0.011816	511.17%
K ₊ t	10070.7	7.60	mg/L	0.063	38.0	mg/L	0.32	0.83%
Mg ₊ t	376996.7	92.8	mg/L	0.44	464	mg/L	2.2	0.48%
Mo ₊ t	78.7	0.00528	mg/L	0.000641	0.0264	mg/L	0.00321	12.15%
Na ₊ t	835808.6	248	mg/L	1.2	1240	mg/L	5.9	0.47%
Ni ₊ t	1.2	0.00005	mg/L	0.000171	0.00027	mg/L	0.000857	317.70%
Pb ₊ t	-18.7	-0.00349	mg/L	0.000059	-0.0175	mg/L	0.00030	1.70%
Tl ₊ t	42.3	0.0142	mg/L	0.00065	0.0710	mg/L	0.00326	4.59%
V ₊ t	14.4	0.0105	mg/L	0.00005	0.0527	mg/L	0.00027	0.51%
Zn ₊ t	315.5	0.00622	mg/L	0.000147	0.0311	mg/L	0.00073	2.36%

Sequence No.: 76
 Sample ID: 2708020330_5X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 5X

Autosampler Location: 83
 Date Collected: 8/13/2007 23:50:31
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020330_5X

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: 2708020330_5X

Analyte	Mean Corrected			Sample			RSD
	Intensity	Conc. Units	Calib	Std.Dev.	Conc. Units	Std.Dev.	
Sca	259343.1	101 %		1.4			1.42%
Yr	237612.8	91.1 %		1.06			1.16%
B_t	65042.7	2.47 mg/L		0.002	12.3 mg/L	0.01	0.08%
Bat	292.0	0.00357 mg/L		0.000011	0.0179 mg/L	0.000005	0.31%
Bet	-1043.1	-0.00032 mg/L		0.000002	-0.00159 mg/L	0.000010	0.66%
Cat	1317573.5	134 mg/L		0.4	669 mg/L	2.2	0.33%
Cdt	-15.0	-0.00064 mg/L		0.000357	-0.00321 mg/L	0.001786	55.70%
Cot	3.3	0.00015 mg/L		0.000059	0.00073 mg/L	0.000293	40.05%
Crt	77791.6	1.11 mg/L		0.005	5.54 mg/L	0.023	0.42%
Cut	198.0	0.00046 mg/L		0.000327	0.00232 mg/L	0.001634	70.49%
Fet	14.8	0.00810 mg/L		0.001064	0.0405 mg/L	0.00532	13.13%
Kt	7094.0	5.35 mg/L		0.092	26.8 mg/L	0.46	1.71%
Mgt	369138.8	90.9 mg/L		0.29	454 mg/L	1.5	0.32%
Mot	69.9	0.00469 mg/L		0.000162	0.0234 mg/L	0.00081	3.46%
Nat	637880.0	190 mg/L		1.3	948 mg/L	6.5	0.68%
Nit	-5.2	-0.00024 mg/L		0.000542	-0.00118 mg/L	0.002711	229.11%
Pbt	-24.2	-0.00452 mg/L		0.000928	-0.0226 mg/L	0.00464	20.54%
Tlt	37.1	0.0125 mg/L		0.00153	0.0624 mg/L	0.00763	12.22%
Vt	921.5	0.3118 mg/L		0.00005	0.0589 mg/L	0.00027	0.45%
Znt	191.8	0.00378 mg/L		0.000041	0.0189 mg/L	0.00021	1.10%

Sequence No.: 77
 Sample ID: 2708020331_20X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 20X

Autosampler Location: 84
 Date Collected: 8/13/2007 23:54:58
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020331_20X
 Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020331_20X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	264447.7	103	%	1.2			1.15%
Yr	241957.6	92.8	%	0.41			0.45%
B_t	31669.8	1.20	mg/L	0.000	24.0	mg/L	0.01 0.04%
Bat	155.4	0.00190	mg/L	0.000098	0.0380	mg/L	0.00197 5.17%
Bet	-771.6	-0.00024	mg/L	0.000023	-0.00471	mg/L	0.000454 9.64%
Cat	372508.6	37.8	mg/L	0.01	757	mg/L	0.1 0.02%
Cdt	-16.9	-0.00072	mg/L	0.000184	-0.0145	mg/L	0.00368 25.39%
Cof	1.6	0.00007	mg/L	0.000052	0.00137	mg/L	0.001030 75.18%
Crt	69450.3	0.989	mg/L	0.0006	19.8	mg/L	0.01 0.06%
Cuf	-45.5	-0.00011	mg/L	0.000236	-0.00213	mg/L	0.004712 221.33%
Fer	2.6	0.00141	mg/L	0.001079	0.0282	mg/L	0.02158 76.55%
Kt	4747.4	3.58	mg/L	0.033	71.7	mg/L	0.66 0.92%
Mgt	97068.4	23.9	mg/L	0.07	478	mg/L	1.4 0.28%
Mot	27.1	0.00182	mg/L	0.000179	0.0364	mg/L	0.00357 9.82%
Nat	332191.9	98.7	mg/L	0.50	1970	mg/L	10.1 0.51%
Nit	2.9	0.00013	mg/L	0.000263	0.00263	mg/L	0.005253 199.97%
Pbt	-7.8	-0.00145	mg/L	0.002382	-0.0291	mg/L	0.04764 163.75%
Tlt	21.9	0.00735	mg/L	0.001609	0.147	mg/L	0.0322 21.88%
Vt	-376.9	0.00314	mg/L	0.000133	0.0628	mg/L	0.00266 4.24%
Znt	105.0	0.00207	mg/L	0.000000	0.0414	mg/L	0.00000 0.00%

Sequence No.: 78
 Sample ID: 2708020342_10X
 Analyt: Walter Hsieh
 Initial Sample Wt:
 Dilution: 10X

Autosampler Location: 85
 Date Collected: 8/13/2007 23:59:20
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020342_10X

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: 2708020342_10X

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	257607.2		99.9 %	0.07			0.07%
Yr	237669.8		91.2 %	0.50			0.55%
B _f	8809.0		0.334 mg/L	0.0001	3.34 mg/L	0.001	0.02%
Ba _f	180.3		0.00221 mg/L	0.000044	0.0221 mg/L	0.00044	2.01%
Be _f	-890.0		-0.00027 mg/L	0.000009	-0.00271 mg/L	0.000093	3.41%
Ca _f	58041.5		5.89 mg/L	0.063	58.9 mg/L	0.63	1.07%
Cd _f	44.4		0.00190 mg/L	0.000089	0.0190 mg/L	0.00089	4.68%
Co _f	24.7		0.00109 mg/L	0.000056	0.0109 mg/L	0.00056	5.13%
Cr _f	88308.4		1.26 mg/L	0.000	12.6 mg/L	0.00	0.03%
Cu _f	267.3		0.00063 mg/L	0.000119	0.00627 mg/L	0.001191	19.01%
Fe _f	87.5		0.0479 mg/L	0.00043	0.479 mg/L	0.0043	0.89%
K _f	6055.1		4.57 mg/L	0.023	45.7 mg/L	0.23	0.51%
Mg _f	8753.6		2.15 mg/L	0.030	21.5 mg/L	0.30	1.41%
Mn _f	42.7		0.00286 mg/L	0.000233	0.0286 mg/L	0.00233	8.14%
Na _f	782595.3		233 mg/L	1.4	2330 mg/L	14.2	0.61%
Ni _f	-5.2		-0.00024 mg/L	0.000131	-0.00241 mg/L	0.001313	54.50%
Pb _f	0.2		0.00005 mg/L	0.000155	0.00046 mg/L	0.001546	336.11%
Tl _f	-0.1		-0.00001 mg/L	0.001107	-0.00008 mg/L	0.011073	>999.9%
V _f	754.0		0.0116 mg/L	0.00034	0.116 mg/L	0.0034	2.95%
Zn _f	98.4		0.00194 mg/L	0.000018	0.0194 mg/L	0.00018	0.94%

Sequence No.: 79
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/14/2007 00:03:49
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	256098.4	99.3	%	0.64			0.65%
Yr	242139.6	92.9	%	0.69			0.74%
B_†	61615.1	2.33	mg/L	0.000	2.33 mg/L	0.000	0.01%
	QC value within limits for B_ Recovery = 93.16%						
Ba†	379822.8	4.65	mg/L	0.013	4.65 mg/L	0.013	0.29%
	QC value within limits for Ba Recovery = 92.93%						
Be†	6017582.8	1.84	mg/L	0.008	1.84 mg/L	0.008	0.42%
	QC value within limits for Be Recovery = 91.75%						
Ca†	514030.8	52.2	mg/L	0.00	52.2 mg/L	0.00	0.01%
	QC value within limits for Ca Recovery = 104.39%						
Cd†	57347.0	2.47	mg/L	0.001	2.47 mg/L	0.001	0.04%
	QC value within limits for Cd Recovery = 98.88%						
Co†	110941.5	4.89	mg/L	0.001	4.89 mg/L	0.001	0.02%
	QC value within limits for Co Recovery = 97.88%						
Cr†	331968.4	4.73	mg/L	0.007	4.73 mg/L	0.007	0.15%
	QC value within limits for Cr Recovery = 94.57%						
Cu†	1905913.9	4.46	mg/L	0.008	4.46 mg/L	0.008	0.19%
	QC value less than the lower limit for Cu Recovery = 89.30%						
Fe†	9217.0	5.05	mg/L	0.102	5.05 mg/L	0.102	2.02%
	QC value within limits for Fe Recovery = 100.90%						
K†	65397.3	49.4	mg/L	0.94	49.4 mg/L	0.94	1.90%
	QC value within limits for K Recovery = 98.71%						
Mg†	218974.7	53.9	mg/L	0.10	53.9 mg/L	0.10	0.19%
	QC value within limits for Mg Recovery = 107.80%						
Mo†	67975.1	4.56	mg/L	0.002	4.56 mg/L	0.002	0.03%
	QC value within limits for Mo Recovery = 91.14%						
Na†	175059.6	52.0	mg/L	0.07	52.0 mg/L	0.07	0.14%
	QC value within limits for Na Recovery = 104.04%						
Ni†	104959.0	4.82	mg/L	0.009	4.82 mg/L	0.009	0.19%
	QC value within limits for Ni Recovery = 96.37%						
Pb†	25872.9	4.84	mg/L	0.003	4.84 mg/L	0.003	0.06%
	QC value within limits for Pb Recovery = 96.71%						
Tl†	14888.3	5.01	mg/L	0.019	5.01 mg/L	0.019	0.38%
	QC value within limits for Tl Recovery = 100.21%						
V†	758797.5	4.70	mg/L	0.002	4.70 mg/L	0.002	0.04%
	QC value within limits for V Recovery = 93.94%						
Zn†	248513.8	4.86	mg/L	0.005	4.86 mg/L	0.005	0.10%
	QC value within limits for Zn Recovery = 97.28%						
QC Failed. Retry.							

Sequence No.: 80
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 8/14/2007 00:05:25
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	256984.1	99.6 %	0.42			0.42%
Yr	237686.2	91.2 %	1.79			1.96%
B ₋ t	62416.1	2.36 mg/L	0.004	2.36 mg/L	0.004	0.19%
	QC value within limits for B ₋	Recovery = 94.37%				
Ba _t	381259.1	4.66 mg/L	0.041	4.66 mg/L	0.041	0.88%
	QC value within limits for Ba	Recovery = 93.28%				
Be _t	6053367.6	1.85 mg/L	0.009	1.85 mg/L	0.009	0.47%
	QC value within limits for Be	Recovery = 92.30%				
Ca _t	513599.1	52.2 mg/L	0.11	52.2 mg/L	0.11	0.20%
	QC value within limits for Ca	Recovery = 104.30%				
Cd _t	58225.9	2.51 mg/L	0.024	2.51 mg/L	0.024	0.97%
	QC value within limits for Cd	Recovery = 100.40%				
Co _t	112586.5	4.97 mg/L	0.046	4.97 mg/L	0.046	0.93%
	QC value within limits for Co	Recovery = 99.33%				
Cr _t	332962.3	4.74 mg/L	0.019	4.74 mg/L	0.019	0.40%
	QC value within limits for Cr	Recovery = 94.85%				
Cu _t	1932291.8	4.53 mg/L	0.047	4.53 mg/L	0.047	1.04%
	QC value within limits for Cu	Recovery = 90.54%				
Fe _t	9306.1	5.09 mg/L	0.040	5.09 mg/L	0.040	0.79%
	QC value within limits for Fe	Recovery = 101.88%				
K _t	65549.4	49.5 mg/L	0.64	49.5 mg/L	0.64	1.29%
	QC value within limits for K	Recovery = 98.94%				
Mg _t	218823.3	53.9 mg/L	0.04	53.9 mg/L	0.04	0.07%
	QC value within limits for Mg	Recovery = 107.73%				
Mo _t	69214.5	4.64 mg/L	0.033	4.64 mg/L	0.033	0.72%
	QC value within limits for Mo	Recovery = 92.80%				
Na _t	172936.5	51.4 mg/L	0.25	51.4 mg/L	0.25	0.50%
	QC value within limits for Na	Recovery = 102.78%				
Ni _t	106801.6	4.90 mg/L	0.050	4.90 mg/L	0.050	1.01%
	QC value within limits for Ni	Recovery = 98.06%				
Pb _t	26352.3	4.93 mg/L	0.035	4.93 mg/L	0.035	0.71%
	QC value within limits for Pb	Recovery = 98.50%				
Tl _t	15044.5	5.06 mg/L	0.045	5.06 mg/L	0.045	0.88%
	QC value within limits for Tl	Recovery = 101.26%				
V _t	760314.8	4.71 mg/L	0.023	4.71 mg/L	0.023	0.48%
	QC value within limits for V	Recovery = 94.13%				
Zn _t	250560.1	4.90 mg/L	0.026	4.90 mg/L	0.026	0.53%
	QC value within limits for Zn	Recovery = 98.07%				

All analyte(s) passed QC.

Sequence No.: 81
Sample ID: CCB
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 0
Date Collected: 8/14/2007 00:08:36
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
All 218.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	269767.6	105 %		0.4			0.37%
Yr	244325.4	93.7 %		1.67			1.78%
B _t	806.3	0.0306 mg/L		0.00101	0.0306 mg/L	0.00101	3.29%
QC value greater than the upper limit for B Recovery = Not calculated							
Ba _t	-12.5	-0.00015 mg/L		0.000049	-0.00015 mg/L	0.000049	32.09%
QC value within limits for Ba Recovery = Not calculated							
Be _t	-43.0	-0.00001 mg/L		0.000003	-0.00001 mg/L	0.000003	25.29%
QC value within limits for Be Recovery = Not calculated							
Ca _t	39.0	0.00396 mg/L		0.001388	0.00396 mg/L	0.001388	35.09%
QC value within limits for Ca Recovery = Not calculated							
Cd _t	1.8	0.00008 mg/L		0.000021	0.00008 mg/L	0.000021	27.24%
QC value within limits for Cd Recovery = Not calculated							
Co _t	-4.1	-0.00018 mg/L		0.000088	-0.00018 mg/L	0.000088	48.45%
QC value within limits for Co Recovery = Not calculated							
Cr _t	115.5	0.00165 mg/L		0.000054	0.00165 mg/L	0.000054	3.29%
QC value within limits for Cr Recovery = Not calculated							
Cu _t	-140.4	-0.00033 mg/L		0.000095	-0.00033 mg/L	0.000095	28.98%
QC value within limits for Cu Recovery = Not calculated							
Fe _t	-0.0	-0.00003 mg/L		0.001379	-0.00003 mg/L	0.001379	>999.9%
QC value within limits for Fe Recovery = Not calculated							
K _t	-99.3	-0.0749 mg/L		0.04318	-0.0749 mg/L	0.04318	57.65%
QC value within limits for K Recovery = Not calculated							
Mg _t	2.4	0.00059 mg/L		0.000068	0.00059 mg/L	0.000068	11.44%
QC value within limits for Mg Recovery = Not calculated							
Mo _t	17.6	0.00118 mg/L		0.000363	0.00118 mg/L	0.000363	30.67%
QC value within limits for Mo Recovery = Not calculated							
Na _t	689.4	0.205 mg/L		0.0007	0.205 mg/L	0.0007	0.35%
QC value within limits for Na Recovery = Not calculated							
Ni _t	-9.5	-0.00044 mg/L		0.000383	-0.00044 mg/L	0.000383	87.38%
QC value within limits for Ni Recovery = Not calculated							
Pb _t	13.1	0.00244 mg/L		0.000139	0.00244 mg/L	0.000139	5.69%
QC value within limits for Pb Recovery = Not calculated							
Tl _t	6.0	0.00202 mg/L		0.001123	0.00202 mg/L	0.001123	55.53%
QC value within limits for Tl Recovery = Not calculated							
V _t	-50.4	-0.00030 mg/L		0.000224	-0.00030 mg/L	0.000224	74.58%
QC value within limits for V Recovery = Not calculated							
Zn _t	-1.6	-0.00003 mg/L		0.000009	-0.00003 mg/L	0.000009	30.85%
QC value within limits for Zn Recovery = Not calculated							
QC Failed. Retry.							

Sequence No.: 82
Sample ID: CCB
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 0
Date Collected: 8/14/2007 00:11:14
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
All 218.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	267904.9	104	%	1.0			0.99%
Yr	247948.0	95.1	%	0.23			0.25%
B ₋ †	604.0	0.0229	mg/L	0.00077	0.0229 mg/L	0.00077	3.35%
QC value greater than the upper limit for B Recovery = Not calculated							
Ba†	-10.9	-0.00013	mg/L	0.000033	-0.00013 mg/L	0.000033	24.59%
QC value within limits for Ba Recovery = Not calculated							
Be†	-75.2	-0.00002	mg/L	0.000003	-0.00002 mg/L	0.000003	11.60%
QC value within limits for Be Recovery = Not calculated							
Ca†	32.0	0.00325	mg/L	0.000652	0.00325 mg/L	0.000652	20.06%
QC value within limits for Ca Recovery = Not calculated							
Cd†	-1.6	-0.00007	mg/L	0.000092	-0.00007 mg/L	0.000092	128.77%
QC value within limits for Cd Recovery = Not calculated							
Co†	-3.8	-0.00017	mg/L	0.000139	-0.00017 mg/L	0.000139	82.95%
QC value within limits for Co Recovery = Not calculated							
Cr†	90.3	0.00129	mg/L	0.000078	0.00129 mg/L	0.000078	6.09%
QC value within limits for Cr Recovery = Not calculated							
Cu†	-205.1	-0.00048	mg/L	0.000012	-0.00048 mg/L	0.000012	2.52%
QC value within limits for Cu Recovery = Not calculated							
Fe†	-1.3	-0.00071	mg/L	0.000839	-0.00071 mg/L	0.000839	117.93%
QC value within limits for Fe Recovery = Not calculated							
K†	-84.1	-0.0635	mg/L	0.10102	-0.0635 mg/L	0.10102	159.11%
QC value within limits for K Recovery = Not calculated							
Mg†	2.4	0.00059	mg/L	0.000310	0.00059 mg/L	0.000310	52.33%
QC value within limits for Mg Recovery = Not calculated							
Mo†	4.5	0.00030	mg/L	0.000199	0.00030 mg/L	0.000199	65.76%
QC value within limits for Mo Recovery = Not calculated							
Na†	585.0	0.174	mg/L	0.0073	0.174 mg/L	0.0073	4.18%
QC value within limits for Na Recovery = Not calculated							
Ni†	-8.8	-0.00041	mg/L	0.000291	-0.00041 mg/L	0.000291	71.72%
QC value within limits for Ni Recovery = Not calculated							
Pb†	6.4	0.00120	mg/L	0.000170	0.00120 mg/L	0.000170	14.21%
QC value within limits for Pb Recovery = Not calculated							
Tl†	0.2	0.00006	mg/L	0.000533	0.00006 mg/L	0.000533	821.64%
QC value within limits for Tl Recovery = Not calculated							
V†	34.7	0.00022	mg/L	0.000009	0.00022 mg/L	0.000009	4.09%
QC value within limits for V Recovery = Not calculated							
Zn†	-11.6	-0.00023	mg/L	0.000227	-0.00023 mg/L	0.000227	100.58%
QC value within limits for Zn Recovery = Not calculated							
QC Failed. Retry.							

Sequence No.: 83

Sample ID: CCB

Analyst:

Initial Sample Wt:

Dilution:

Autosampler Location: 0

Date Collected: 8/14/2007 00:13:50

Data Type: Original

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	267902.3	104	%	0.1			0.06%
Yr	244814.2	93.9	%	1.13			1.20%
B ₋ †	480.4	0.0182	mg/L	0.00028	0.0182 mg/L	0.00028	1.55%
QC value within limits for B ₋ Recovery = Not calculated							
Ba†	-3.2	-0.00004	mg/L	0.000029	-0.00004 mg/L	0.000029	75.89%
QC value within limits for Ba Recovery = Not calculated							
Be†	-20.0	-0.00001	mg/L	0.000023	-0.00001 mg/L	0.000023	382.36%
QC value within limits for Be Recovery = Not calculated							
Ca†	53.6	0.00544	mg/L	0.001167	0.00544 mg/L	0.001167	21.44%
QC value within limits for Ca Recovery = Not calculated							
Cd†	-8.0	-0.00034	mg/L	0.000010	-0.00034 mg/L	0.000010	3.05%
QC value within limits for Cd Recovery = Not calculated							
Co†	-3.2	-0.00014	mg/L	0.000015	-0.00014 mg/L	0.000015	10.44%

QC value within limits for Co	Recovery = Not calculated					
Crt	72.6	0.00103 mg/L	0.000002	0.00103 mg/L	0.000002	0.20%
QC value within limits for Cr	Recovery = Not calculated					
Cut	-236.0	-0.00055 mg/L	0.000077	-0.00055 mg/L	0.000077	13.94%
QC value within limits for Cu	Recovery = Not calculated					
Fe†	-0.5	-0.00030 mg/L	0.003873	-0.00030 mg/L	0.003873	>999.9%
QC value within limits for Fe	Recovery = Not calculated					
K†	-102.9	-0.0776 mg/L	0.04931	-0.0776 mg/L	0.04931	63.52%
QC value within limits for K	Recovery = Not calculated					
Mg†	4.8	0.00117 mg/L	0.000431	0.00117 mg/L	0.000431	36.72%
QC value within limits for Mg	Recovery = Not calculated					
Mo†	1.2	0.00008 mg/L	0.000216	0.00008 mg/L	0.000216	279.52%
QC value within limits for Mo	Recovery = Not calculated					
Nat	564.0	0.168 mg/L	0.0104	0.168 mg/L	0.0104	6.18%
QC value within limits for Na	Recovery = Not calculated					
Nit	-12.6	-0.00058 mg/L	0.000181	-0.00058 mg/L	0.000181	31.31%
QC value within limits for Ni	Recovery = Not calculated					
Pbt	6.2	0.00116 mg/L	0.000265	0.00116 mg/L	0.000265	22.87%
QC value within limits for Pb	Recovery = Not calculated					
Tl†	1.1	0.00036 mg/L	0.001900	0.00036 mg/L	0.001900	526.75%
QC value within limits for Tl	Recovery = Not calculated					
V†	-9.2	-0.00005 mg/L	0.000024	-0.00005 mg/L	0.000024	47.09%
QC value within limits for V	Recovery = Not calculated					
Znt	-21.6	-0.00042 mg/L	0.000123	-0.00042 mg/L	0.000123	29.13%
QC value within limits for Zn	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 84
 Sample ID: 2708020343_2X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 86
 Date Collected: 8/14/2007 00:17:15
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020343_2X

Analyte	Back Pressure	Flow
All	218.0 kPa	0.65 L/min

Mean Data: 2708020343_2X

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	278076.2	108 %	%	2.0			1.82%
Yr	253318.2	97.2 %	%	0.99			1.02%
B_t	384.8	0.0146	mg/L	0.00040	0.0292	mg/L	0.00081 2.76%
Ba_t	130.0	0.00159	mg/L	0.000032	0.00318	mg/L	0.000064 2.00%
Be_t	43.0	0.00001	mg/L	0.000002	0.00003	mg/L	0.000003 13.10%
Cat	32739.6	3.32	mg/L	0.002	6.65	mg/L	0.005 0.07%
Cd_t	-5.6	-0.00024	mg/L	0.000129	-0.00048	mg/L	0.000258 53.65%
Co_t	1.3	0.00006	mg/L	0.000158	0.00012	mg/L	0.000315 266.29%
Crt	115.9	0.00165	mg/L	0.000303	0.00330	mg/L	0.000606 18.35%
Cu_t	-17.3	-0.00004	mg/L	0.000096	-0.00008	mg/L	0.000193 239.10%
Fe_t	80.2	0.0439	mg/L	0.00108	0.0878	mg/L	0.00215 2.45%
K_t	-45.5	-0.0343	mg/L	0.02542	-0.0686	mg/L	0.05084 74.08%
Mg_t	458.1	0.113	mg/L	0.0002	0.226	mg/L	0.0005 0.22%
Mo_t	2.8	0.00019	mg/L	0.000136	0.00038	mg/L	0.000272 71.92%
Na_t	1016.5	0.302	mg/L	0.0064	0.604	mg/L	0.0128 2.12%
Ni_t	-3.9	-0.00018	mg/L	0.000050	-0.00036	mg/L	0.000101 27.89%
Pb_t	15.3	0.00286	mg/L	0.000048	0.00571	mg/L	0.000096 1.69%
Tl_t	3.6	0.00120	mg/L	0.000760	0.00241	mg/L	0.001520 63.18%
V_t	75.9	0.00048	mg/L	0.000243	0.00095	mg/L	0.000487 51.12%
Zn_t	90.7	0.00179	mg/L	0.000208	0.00358	mg/L	0.000416 11.64%

Sequence No.: 85
 Sample ID: 2708020344_5X
 Analyst: Walter Hsieh
 Initial Sample Wt:
 Dilution: 5X

Autosampler Location: 87
 Date Collected: 8/14/2007 00:20:57
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2708020344_5X

Analyte Back Pressure Flow
 All 218.0 kPa 0.65 L/min

Mean Data: 2708020344_5X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	260553.5	101	%	1.2			1.23%
Yr	241980.1	92.8		0.17			0.18%
B_f	15717.1	0.597	mg/L	0.0017	2.98	mg/L	0.008
Ba_f	328.4	0.00402	mg/L	0.000048	0.0201	mg/L	0.00024
Be_f	-1430.8	-0.00044	mg/L	0.000006	-0.00218	mg/L	0.000028
Ca_f	111294.3	11.3	mg/L	0.15	56.5	mg/L	0.73
Cd_f	90.6	0.00388	mg/L	0.000091	0.0194	mg/L	0.00046
Co_f	34.2	0.00151	mg/L	0.000091	0.00753	mg/L	0.000457
Cr_f	168064.1	2.39	mg/L	0.000	12.0	mg/L	0.00
Cu_f	314.1	0.00074	mg/L	0.000110	0.00368	mg/L	0.000551
Fe_f	137.6	0.0753	mg/L	0.00001	0.377	mg/L	0.0001
K_f	12189.9	9.20	mg/L	0.137	46.0	mg/L	0.68
Mg_f	16616.1	4.09	mg/L	0.054	20.5	mg/L	0.27
Mn_f	97.5	0.00653	mg/L	0.000035	0.0327	mg/L	0.00018
Na_f	1517164.9	451	mg/L	1.0	2250	mg/L	4.8
Ni_f	-2.1	-0.00010	mg/L	0.000255	-0.00048	mg/L	0.001276
Pb_f	-2.7	-0.00051	mg/L	0.000013	-0.00256	mg/L	0.000067
Tl_f	7.1	0.00242	mg/L	0.000506	0.0121	mg/L	0.00253
V_f	1644.0	0.02338	mg/L	0.00006	0.117	mg/L	0.0003
Zn_f	320.0	0.00631	mg/L	0.000205	0.0315	mg/L	0.00102

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Lot #
 HNO3 R# 100411 HCL R# 100412
 IS = Yttrium(ME0702007)0.75mL + Scandium ME0606006)0.5mL to 1000mL w/ 2% HNO3

Standards	Lot #	Exp. Date	Dilution
Calibration	ME0704023	(05/01/08) /	1:10 ME0704027
(Prepare daily)	ME0704024	(05/01/08) /	1:10
CCV/MCV/ECV	ME0610005	(04/10/08) /	CCV/ECV 1:20 ME0610006 MCV 1:40 ME0610007
(Prepare daily)			
Spike/LCS	ME0606004	(12/13/07) /	1:100 ME0601006
(Prepare daily)	ME0705006	(10/01/07) /	1:100
	ME0704005	(10/04/08) /	1:200
MRL	ME0703010	(09/16/07) /	1:100 ME0603015
(Prepare daily)			
ICSA	ME0705012	(08/23/07) /	
ICSAB	ME0705013	(08/23/07) /	
QC-25 1PPM	ME0705005	(11/09/07) /	
Linearity	ME0705004	(11/09/07) /	
Method Sr/Ti/Sn/SiO2			
Calibration	ME0708004	(02/07/08) /	
CCV/ECV	ME0703007	(09/16/07) /	
Spike/LCS	ME0703006	(09/16/07) /	1:100
(Prepare daily)			
MRL	ME0708005	(02/07/08) /	1:100
(Prepare daily)			
Method Li			
Std/ICV/MRL	ME0703008	(09/16/07) /	1:1000, 200, 40, 10
(Prepare daily)			
LCS/Spike	ME0707002	(01/10/08) /	1:50
(Prepare daily)			
ccv	ME0707001	(10/31/07) /	1:40
(Prepare daily)			

From May 2005: the calibration std for ICP should be ME0505010, 011 not ME0408010
 dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.
 From 10/4/06: the QC-25 lppm solution ref # should be ME0610001 not ME0610002

ME0704023

Initial:

Date:

w 37
4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #1
Date Received/Prepped: 4/23/2007
Date Expired: 5/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid
Amount: 500 mL

ME #: 0704023

By: wbh

Lot #: A2-MEB235010

Certificate: Y

NIST SRM: Varies

Storage: Room Temp

Component	Comment	Conc. Unit:
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		1000 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		100 ug/ml
Be		50 ug/ml
SR		40 ug/ml
Ag		30 ug/ml
		20 ug/ml



195 Lehigh Avenue, Suite 4
Lakewood, New Jersey 08701 - USA
inorganicventures.com

CERTIFICATE OF ANALYSIS

tel: 800.669.6799 732.901.1900
fax: 732.901.1903
info@inorganicventures.com

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
Catalog No.: MWH-ICAP-CAL-1
Lot Number: **A2-MEB235010**
Matrix: 5% HNO₃(abs)

M80704023

1,000.00 µg/mL each:
Ca, K, Mg, Na,
100.00 µg/mL each:
Al, As, Ba, Co, Cr3, Cu, Fe, Mn, Ni, Pb, Se, Tl, V, Zn,
50.00 µg/mL each:
Cd,
40.00 µg/mL each:
Be,
30.00 µg/mL each:
Sr,
20.00 µg/mL each:
Ag

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.1 ± 0.5 µg/mL	Arsenic, As	100.2 ± 0.3 µg/mL	Barium, Ba	99.9 ± 0.2 µg/mL
Beryllium, Be	39.98 ± 0.08 µg/mL	Cadmium, Cd	50.05 ± 0.12 µg/mL	Calcium, Ca	997 ± 3 µg/mL
Chromium+3, Cr3	100.1 ± 0.4 µg/mL	Cobalt, Co	100.1 ± 0.2 µg/mL	Copper, Cu	100.1 ± 0.2 µg/mL
Iron, Fe	100.0 ± 0.2 µg/mL	Lead, Pb	100.1 ± 0.3 µg/mL	Magnesium, Mg	996 ± 3 µg/mL
Manganese, Mn	100.1 ± 0.3 µg/mL	Nickel, Ni	100.1 ± 0.2 µg/mL	Potassium, K	1,003 ± 2 µg/mL
Selenium, Se	100.1 ± 0.2 µg/mL	Silver, Ag	20.03 ± 0.06 µg/mL	Sodium, Na	997 ± 5 µg/mL
Strontium, Sr	29.92 ± 0.18 µg/mL	Thallium, Tl	100.0 ± 0.1 µg/mL	Vanadium, V	100.1 ± 0.3 µg/mL
Zinc, Zn	100.1 ± 0.4 µg/mL				

Certified Density: 1.056 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean

x_i = individual results

n = number of measurements

$\sum s_i$ = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 ASSAY INFORMATION

ELEMENT	METHOD	NIST SRM#	SRM LOT#	ELEMENT	METHOD	NIST SRM#	SRM LOT#
Ag	ICP Assay	3151	992212	Ag	Volhard	999a	999a
Al	ICP Assay	3101a	010808	Al	EDTA	928	928
As	ICP Assay	3103a	010713	As	Gravimetric		See Sec. 4.2
Ba	Gravimetric		See Sec. 4.2	Ba	ICP Assay	3104a	992907
Be	Gravimetric		See Sec. 4.2	Be	ICP Assay	3105a	892707
Ca	EDTA	928	928	Ca	ICP Assay	3109a	000622
Cd	EDTA	928	928	Cd	ICP Assay	3108	890312
Co	ICP Assay	3113	00630	Co	EDTA	928	928
Cr3	ICP Assay	3112a	990607	Cr3	Gravimetric		See Sec. 4.2
Cu	EDTA	928	928	Cu	ICP Assay	3114	891811
Fe	ICP Assay	3126a	000606	Fe	EDTA	928	928
K	ICP Assay	3141a	891312	K	Gravimetric		See Sec. 4.2
Mg	EDTA	928	928	Mg	ICP Assay	3131a	991107
Mn	EDTA	928	928	Mn	ICP Assay	3132	890903
Na	ICP Assay	3152a	990907	Na	Gravimetric		See Sec. 4.2
Ni	EDTA	928	928	Ni	ICP Assay	3136	000612
Pb	ICP Assay	3128	030721	Pb	EDTA	928	928
Se	ICP Assay	3149	992106	Se	Gravimetric		See Sec. 4.2
Sr	ICP Assay	3153a	990906	Sr	EDTA	928	928
Tl	Gravimetric		See Sec. 4.2	Tl	ICP Assay	3158	993012
V	ICP Assay	3165	992706	V	EDTA	928	928
Zn	EDTA	928	928	Zn	ICP Assay	3168a	001402

- 4.2 **BALANCE CALIBRATION** - All balances are checked daily using an in-house procedure. The weights used for testing are annually compared to master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 692476 - Class 1 and 692476A - Class 2. The NIST test number is 822/260017-98. All analytical balances are calibrated every 4 months. The balances are calibrated with a class 1 and/or class 2 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is
- 4.3 **THERMOMETER CALIBRATION** - The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/789543, 217368/P14452, 176240/P14452, 176240. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification
- 4.4 **GLASSWARE CALIBRATION** - An in-house procedure is used to calibrate all Class A Glassware used in the manufacturing and quality control of CRM's.

5.0 **TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES IN µg/mL - N/A**

6.0 **INTENDED USE**

- For the calibration of analytical instruments including but not limited to the following:
ICP-MS, ICP-OES, FAAS, GFAA, XRF, and DCP
- For the validation of analytical methods
- For the preparation of "working reference samples"
- For interference studies and the determination of correction coefficients
- For detection limit and linearity studies
- For additional intended uses, contact Technical Staff

7.0 **INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL**

Storage & Handling - Keep Tightly sealed when not in use. Store and use at $20 \pm 4^\circ\text{C}$. Do Not pipette from the container. Do Not return portions removed from pipetting to container.

Element Specific Information - For specific information regarding any element. Contact technical staff.

Low Silver Note: This solution contains "LOW" levels of Silver. Please store this entire bottle inside a sealed glass jar.

8.0 **HAZARDOUS INFORMATION** - Please refer to the enclosed Material Safety Data sheet for information regarding this CRM.

9.0 **HOMOGENEITY** - This solution was mixed according to in-house procedure IV-MPM-004 and is guaranteed to be homogeneous.

10.0 **QUALITY STANDARD DOCUMENTATION**



10.1 **ISO 9001:2000 Quality Management System Registration - QMI Certificate Number 010105**

Recognized by:

- Registrar Accreditation Board (ANSI-RAB)
- Standards Council of Canada (SCC)
- Dutch Council for Accreditation (RVA)
- Entidad Mexicana de Acreditacion, a.c.(EMA)

Members of IQ Net International Certification Network:

- Argentina (IRAM), Australia (QAS), Austria (ÖQS), Belgium (Avinter), Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)

10.2 **ISO/IEC 17025 - 1999 "General Requirements for the Competence of Testing and Calibration"**

- Chemical Testing - Accredited A2LA Certificate Number 883.01

10.3 **ISO/IEC Guide 34 - 2000 "General Requirements for the Competence of Reference Material Producers"**

- Reference Materials Production - Accredited A2LA Certificate Number 883.02

A2LA Mutual Recognition Agreement Partners:

- Australia (NATA), Austria (BrWA), Belgium (BELTEST) (BKO-OBE), Canada (SCC), Chinese Taipei (CNLA), Czech Republic (NAO), Denmark (DANAK), Finland (FINAS), France (COFRAC), Germany (DAR), Hong Kong (HKAS), Ireland (NAB), Italy (SIT) (SINAL), Japan (JAB) (JNLA), Republic of Korea (KOLAS), The Netherlands (RvA), New Zealand (IANZ), Norway (NA), Portugal (IPQ), Singapore (SAC-SINGLAS), Spain (ENAC), Sweden (SWEDAC), Switzerland (SAS), United Kingdom (UKAS) and United States (NVLAP) (ICBO ES)

10.4 **10CFR50 Appendix B - Nuclear Regulatory Commission - Domestic Licensing of Production and Utilization Facilities**

10.5 **10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance**

10.6 **MIL-STD-45662A (Obsolete/Observed)**

11.0 DATE OF CERTIFICATION AND PERIOD OF VALIDITY

- 11.1 Shelf Life - The period of time during which the concentration of the analyte(s) in a properly packaged, unopened, and unused standard stored under environmentally controlled and monitored conditions will remain within the specified uncertainty range. Shelf life is limited primarily by transpiration (loss of water from the solution) and infrequently, by chemical instability. Transpiration studies of chemically-stable solutions performed at the manufacturer's facility show a CRM shelf-life of twenty one months for solutions packaged in 125-mL low density polyethylene bottles. When stored under special environmental controls that minimize transpiration and instability, the shelf life can be extended past this limit.
- 11.2 Expiration Date - The date after which a CRM should not be used. Routine laboratory use of a CRM increases transpiration losses and the chance of contamination which affect the integrity of the CRM and limit its useful life. Manufacturer concurs with state and federal regulatory agencies' recommendations that solution standards be assigned a one-year expiration date.

Certification Date: April 16, 2007
Expiration Date:

EXPIRES
1/2/2008

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By: Nick Maida, Product Documentation Administrator

Nicholas Maida

Certificate Approved By: Katalin Le, QC Manager

Katalin Le

Certifying Officer: Paul Gaines, PhD., Senior Technical Director

Paul R. Gaines

ME0704024

Initial:

WJ

Date:

4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #2
Date Received/Prepped: 4/23/2007
Date Expired: 5/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid + Trace HF
Amount: 500 mL

ME #: 0704024
By: wbh
Lot #: A2-MEB235011
Certificate: Y
NIST SRM: Varies
Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml



195 Lenigh Avenue, Suite 4
 Lakewood, New Jersey 08701 - USA
 inorganicventures.com

CERTIFICATE OF ANALYSIS

tel: 800.669.6799 - 732.401.1410
 fax: 732.401.1403
 info@inorganicventures.com

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM
 Custom Solution
 Catalog No.: MWH-ICAP-CAL-2
 Lot Number: **A2-MEB235011**
 Matrix: tr. HF, 5% HNO₃(abs)

M70704024

100.00 µg/mL each:
 Mo, Sb, Sn, Ti
 50.00 µg/mL each:
 B

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	50.07 ± 0.28 µg/mL	Molybdenum, Mo	100.3 ± 0.3 µg/mL
Tin, Sn	100.3 ± 0.3 µg/mL	Titanium, Ti	100.5 ± 0.2 µg/mL		

Certified Density: 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean

x_i = individual results

n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

$\sum s_i$ = The summation of all significant estimated errors

(Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

10.0 QUALITY STANDARD DOCUMENTATION



10.1 ISO 9001:2000 Quality Management System Registration - QMI Certificate Number 010105

Recognized by:

Registrar Accreditation Board (ANSI-RAB)

Standards Council of Canada (SCC)

Dutch Council for Accreditation (RVA)

Entidad Mexicana de Acreditacion, a.c.(EMA)

Members of IQ Net International Certification Network:

Argentina (IRAM), Australia (QAS), Austria (OQS), Belgium (Avinter), Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)

10.2 ISO/IEC 17025 - 1999 "General Requirements for the Competence of Testing and Calibration"

- Chemical Testing - Accredited A2LA Certificate Number 883.01

10.3 ISO/IEC Guide 34 - 2000 "General Requirements for the Competence of Reference Material Producers"

- Reference Materials Production - Accredited A2LA Certificate Number 883.02

A2LA Mutual Recognition Agreement Partners:

Australia (NATA), Austria (BmWA), Belgium (BELTEST) (BKO-OBE), Canada (SCC), Chinese Taipei (CNLA), Czech Republic (NAO), Denmark (DANAK), Finland (FINAS), France (COFRAC), Germany (DAR), Hong Kong (HKAS), Ireland (NAB), Italy (SIT) (SINAL), Japan (JAB) (JNLA), Republic of Korea (KOLAS), The Netherlands (RVA), New Zealand (IANZ), Norway (NA), Portugal (IPQ), Singapore (SAC-SINGLAS), Spain (ENAC), Sweden (SWEDAC), Switzerland (SAS), United Kingdom (UKAS) and United States (NVLAP) (ICBO ES)

10.4 10CFR50 Appendix B - Nuclear Regulatory Commission - Domestic Licensing of Production and Utilization Facilities

10.5 10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance

10.6 MIL-STD-45662A (Obsolete/Observed)

11.0 DATE OF CERTIFICATION AND PERIOD OF VALIDITY

11.1 **Shelf Life** - The period of time during which the concentration of the analyte(s) in a properly packaged, unopened, and unused standard stored under environmentally controlled and monitored conditions will remain within the specified uncertainty range. Shelf life is limited primarily by transpiration (loss of water from the solution) and infrequently, by chemical instability. Transpiration studies of chemically-stable solutions performed at the manufacturer's facility show a CRM shelf-life of twenty one months for solutions packaged in 125-mL low density polyethylene bottles. When stored under special environmental controls that minimize transpiration and instability, the shelf life can be extended past this limit.

11.2 **Expiration Date** - The date after which a CRM should not be used. Routine laboratory use of a CRM increases transpiration losses and the chance of contamination which affect the integrity of the CRM and limit its useful life. Manufacturer concurs with state and federal regulatory agencies' recommendations that solution standards be assigned a one-year expiration date.

Certification Date: April 16, 2007

Expiration Date:

EXPIRES
1/2008

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By: Nick Maida, Product Documentation Administrator

Certificate Approved By: Katalin Le, QC Manager

Certifying Officer: Paul Gaines, PhD., Senior Technical Director

4.1 ASSAY INFORMATION

ELEMENT	METHOD	NIST SRM#	SRM LOT#	ELEMENT	METHOD	NIST SRM#	SRM LOT#
B	ICP Assay	3107	991907	B	Gravimetric		See Sec. 4.2
Mo	Gravimetric		See Sec. 4.2	Mo	ICP Assay	3134	891307
Sb	ICP Assay	3102a	990707	Sb	Gravimetric		See Sec. 4.2
Sn	ICP Assay	3161a	993107	Sn	Gravimetric		See Sec. 4.2
Ti	Gravimetric		See Sec. 4.2	Ti	ICP Assay	3162a	992801

4.2 **BALANCE CALIBRATION** - All balances are checked daily using an in-house procedure. The weights used for testing are annually compared to master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 692476 - Class 1 and 692476A - Class 2. The NIST test number is 822/260017-98. All analytical balances are calibrated every 4 months. The balances are calibrated with a class 1 and/or class 2 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is

4.3 **THERMOMETER CALIBRATION** - The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification

4.4 **GLASSWARE CALIBRATION** - An in-house procedure is used to calibrate all Class A Glassware used in the manufacturing and quality control of CRM's.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES IN µg/mL - N/A

6.0 INTENDED USE

For the calibration of analytical instruments including but not limited to the following:
ICP-MS, ICP-OES, FAAS, GFAA, XRF, and DCP

For the validation of analytical methods

For the preparation of "working reference samples"

For interference studies and the determination of correction coefficients

For detection limit and linearity studies

For additional intended uses, contact Technical Staff

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

Storage & Handling - Keep Tightly sealed when not in use. Store and use at 20 ± 4°C. Do Not pipette from the container. Do Not return portions removed from pipetting to container.

Element Specific Information - For specific information regarding any element: Contact technical staff.

HF Note: This standard should not be prepared or stored in glass.

8.0 **HAZARDOUS INFORMATION** - Please refer to the enclosed Material Safety Data sheet for information regarding this CRM.

9.0 **HOMOGENEITY** - This solution was mixed according to in-house procedure IV-MPM-004 and is guaranteed to be homogeneous.

Initial:
Date:

WBH
10/17/06

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV Stock Standard
Date Received/Prepped: 10/17/2006
Date Expired: 4/10/2008
Manufacturer: CPI
Matrix: 5% HNO₃ = tr HF
Amount: 100 mL x 10

ME #: 0610005
By: WBH
Lot #: 06j053
Certificate: Y
NIST SRM: Varius
Storage: Room Temp

Component	Comment	Conc. Unit:
Ag		20 ppm
Al		100 ppm
As		100 ppm
B		100 ppm
Ba		50 ppm
bE		100 ppm
Ca		40 ppm
Cd		1000 ppm
Co		50 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		100 ppm
Mg		1000 ppm
Mn		1000 ppm
Mo		100 ppm
Na		100 ppm
Ni		1000 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		100 ppm
Sn		20 ppm
Ti		20 ppm



USA

5580 Skyline Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

Expiry: 4/10/2008

Certificate of Analysis

Part Number: 4400-061003RH01
Lot Number: 06J053
Shelf Life: 18 months

M70610005

MWH
Custom Multi
5% HNO3 + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.



METALS STANDARD DOCUMENTATION

wbh
6/19/06

Standard:	ICP/ICPMS LCS/SPIKE Solution	ME #: 0606004
Date Received/Prepped:	6/19/2006	By: wbh
Date Expired:	12/13/2006	Lot #: 06F103
Manufacturer:	CPI	Certificate: Y
Matrix:	5% HNO ₃ + 0.1% HF	NIST SRM: 3100 Series
Amount:	10 x 100 mL	Storage: Room Temp

Component	Comment	Conc. Unit:
Iron	CPI P/N: 4400-050314RH01	500 mg/L
Aluminum		200 mg/L
Barium		100 mg/L
Cobalt		100 mg/L
Chromium		100 mg/L
Copper		100 mg/L
Molybdenum		100 mg/L
Strontium		100 mg/L
Titanium		100 mg/L
Vanadium		100 mg/L
Zinc		100 mg/L
Tin		100 mg/L
Silver		100 mg/L
Boron		50 mg/L
Manganese		50 mg/L
Nickel		50 mg/L
Antimony		50 mg/L
Arsenic		50 mg/L
Cadmium		20 mg/L
Lead		20 mg/L
Selenium		20 mg/L
Thallium		20 mg/L
Uraium		20 mg/L
Beryllium		5 mg/L
Tin		100 mg/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 12/13/2007

Certificate of Analysis

Part Number: 4400-050314RH01
Lot Number: 06F103
Shelf Life: 18 months

M70606004

MWH Labs
 5% HNO₃ + 0.1% HF
 #REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.



Initial: WBY
Date: 5/9/07

METALS STANDARD DOCUMENTATION

Standard: ICP Spike solution
Date Received/Prepped: 5/9/2007
Date Expired: 10/1/2007
Manufacturer: MWH-WBH
Matrix: 2% HNO3
Amount: 100mL x 2

ME #: 0705006
By: WBH
Lot #:
Certificate: Y
NIST SRM: 3100 SERIES
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	8.0mL ME0611005/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

Initial:
Date:

WBH
11/1/2006

METALS STANDARD DOCUMENTATION

Standard: Arsenic Stock Std
Date Received/Prepped: 11/1/2006
Date Expired: 10/1/2007
Manufacturer: IV
Matrix: 1.4% HNO3
Amount: 100mL

ME #: 0611005
By: WBH
Lot #: Y-AS02029
Certificate: Y
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	Cat # CGAS1-1	1004 ppm



195 Lenigh Avenue, Suite 4
Lakewood, New Jersey 08701 - USA
inorganicventures.com

CERTIFICATE OF ANALYSIS

tel: 800.669.6199 - 732.901.1900
fax: 732.901.1903
info@inorganicventures.com

1.0 **INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 **DESCRIPTION OF CRM** **1000 µg/mL Arsenic in 1.4% (abs) HNO3**

Catalog Number: CGAS1-1, CGAS1-2, and CGAS1-5
Lot Number: **Y -AS02029**
Starting Material: As Polycrystalline lump
Starting Material Purity (%): 99.999055
Starting Material Lot No: 23115
Matrix: 1.4% (abs) HNO3

MF0611005

3.0 **CERTIFIED VALUES AND UNCERTAINTIES**

Certified Concentration: 1004 ± 2 µg/mL 995 ± 2 µg/g

Certified Density: 1.009 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_1}{n}$$

(\bar{x}) = mean
x1 = individual results
n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_1^2)]^{1/2}}{(n)^{1/2}}$$

$\sum s_1$ = The summation of all significant estimated errors
(Most common are the errors from instrumental measurement weighting, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 **TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS**

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)
This IV product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 **Assay Method #1** **1004 ± 2 µg/mL 995 ± 2 µg/g (Avg 2 runs)**
ICP Assay NIST SRM 3103a Lot Number: 010713
Assay Method #2 **1003 ± 5 µg/mL 994 ± 5 µg/g**
Gravimetric NIST SRM Lot Number: See Sec. 4.2

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

Storage & Handling - Keep tightly sealed when not in use. Store and use at $20 \pm 4^\circ\text{C}$. Do not pipet from container. Do not return portions removed for pipetting to container.

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 74.9216; mix of +3 and +5; 6; H_3AsO_4 and HAsO_2

Chemical Compatibility - Arsenic has no cationic chemistry. It is soluble in HCl , HNO_3 , H_3PO_4 , H_2SO_4 and HF aqueous matrices water and NH_4OH . It is stable with most inorganic anions (forms arsenate when boiled with chromate) but many cationic metals form the insoluble arsenates under pH neutral conditions. When fluorinated and / or under acidic conditions arsenate formation is typically not a problem at moderate to low concentrations.

Stability - 2-100 ppb levels stable for months alone or mixed with other elements at equivalent levels in 1% HNO_3 / LDPE container. 1-10,000 ppm solutions chemically stable for years in 1-5% HNO_3 / LDPE container.

As Containing Samples (Preparation and Solution) - As_2O_3 (soluble in 1:1 H_2O / HNO_3); Oxides (the oxide exists in crystalline and amorphous forms where the amorphous form is more water soluble. The oxides typically dissolve in dilute acidic solutions when boiled); Minerals (One gram of powdered sample is fused in a NiO crucible with 10 grams of a 1:1 mix of K_2CO_3 and KNO_3 and the melt extracted with hot water); Organic Matrices (0.2 to 0.5 grams of the sample are fused with 15 grams of a 1:1 Na_2CO_3 / Na_2O_2 mix in a NiO crucible. The fuseate is extracted with water and acidified with HNO_3)

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Type	Interferences (underlined indicates severe)
ICP-OES 189.042 nm	0.05 / 0.005 $\mu\text{g}/\text{mL}$	1	atom	<u>Cr</u>
ICP-OES 193.696 nm	0.1 / 0.01 $\mu\text{g}/\text{mL}$	1	atom	<u>V, Ge</u>
ICP-OES 228.812 nm	0.1 / 0.01 $\mu\text{g}/\text{mL}$	1	atom	<u>Cd, Pt, Ir, Co</u>
ICP-MS 75 amu	20 ppt	n/a	M+	40Ar35Cl, 59Co16O, 36Ar38Ar1H, 38Ar37Cl, 6Ar39K, 150Nd2+, 150Sm2+

- 8.0 **HAZARDOUS INFORMATION** - Please refer to the enclosed Material Safety Data sheet for information regarding this CRM
- 9.0 **HOMOGENEITY** - This solution was mixed according to procedure IV-MPM-004 and is guaranteed to be homogeneous.

10.0 QUALITY STANDARD DOCUMENTATION



10.1 **ISO 9001:2000 Quality Management System Registration - QMI Certificate Number 010105**

Recognized by:

Registrar Accreditation Board (ANSI-RAB)

Standards Council of Canada (SCC)

Dutch Council for Accreditation (RVA)

Entidad Mexicana de Acreditacion, a.c.(EMA)

Members of **IQ Net International Certification Network:**

Argentina (IRAM), Australia (QAS), Austria (ÖQS), Belgium (Avinter), Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)

10.2 **ISO/IEC 17025 - 1999 "General Requirements for the Competence of Testing and Calibration"**

- Chemical Testing - Accredited A2LA Certificate Number 883.01

10.3 **ISO/IEC Guide 34 - 2000 "General Requirements for the Competence of Reference Material Producers"**

- Reference Materials Production - Accredited A2LA Certificate Number 883.02

A2LA Mutual Recognition Agreement Partners:

Australia (NATA), Austria (BmWA), Belgium (BELTEST) (BKO-OBE), Canada (SCC), Chinese Taipei (CNLA), Czech Republic (NAO), Denmark (DANAK), Finland (FINAS), France (COFRAC), Germany (DAR), Hong Kong (HKAS), Ireland (NAB), Italy (SIT) (SINAL), Japan (JAB) (JNLA), Republic of Korea (KOLAS), The Netherlands (RvA), New Zealand (IANZ), Norway (NA), Portugal (IPQ), Singapore (SAC-SINGLAS), Spain (ENAC), Sweden (SWEDAC), Switzerland (SAS), United Kingdom (UKAS) and United States (NVLAP) (ICBO ES)

10.4 **10CFR50 Appendix B - Nuclear Regulatory Commission**

- Domestic Licensing of Production and Utilization Facilities

10.5 **10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance**

10.6 **MIL-STD-45662A (Obsolete/Observed)**

- 4.2 BALANCE CALIBRATION** - All balances are checked daily using in-house procedure number 6-IMM001. The weights used for testing are annually compared to Gerhart Scale Corporation's master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 692476 - Class 1 and 692476A - Class 2. The NIST test number is 822/260017-98. All analytical balances are calibrated every 4 months by Gerhart Scale Corp. of South Amboy. The balances are calibrated with a class 1 and/or class 2 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is 822/260017-98.
- 4.3 THERMOMETER CALIBRATION** - The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. The in-house procedure No. is 2-QC-001. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification Nos. 92564, 119016, 471047 and NIST test report Nos. 811/258522, 811/2557078, and 236090.
- 4.4 GLASSWARE CALIBRATION** - In-house procedure 3-QC-002 is used to calibrate all Class A glassware used in the manufacture and quality control of CRM's.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP/MS AND ICP-OES IN µg/mL

CRM's solutions are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

Q Al 0.00038	M Dy < 0.01884	Q Li < 0.00002	M Pr < 0.00094	M Te < 0.09418
Q Sb < 0.01000	M Er < 0.01570	M Lu < 0.00126	Q Re < 0.01000	M Tb < 0.00094
s As	M Eu < 0.00942	Q Mg 0.00017	M Rh < 0.00314	M Tl < 0.00314
M Ba < 0.03139	M Gd < 0.00314	Q Mn 0.00005	M Rb < 0.00314	M Th < 0.00314
M Be < 0.00157	M Ga < 0.00314	Q Hg < 0.01200	M Ru < 0.00628	M Tm < 0.00126
M Bi < 0.00126	M Ge < 0.01884	M Mo < 0.00628	M Sm < 0.00314	Q Sn 0.00073
Q B < 0.01200	M Au < 0.00942	M Nd < 0.00628	M Sc < 0.03139	M Ti < 0.15697
M Cd < 0.00942	M Hf < 0.00628	Q Ni < 0.00200	M Se < 0.02511	M W < 0.03139
Q Ca 0.00206	M Ho < 0.00157	Q Nb < 0.00200	Q Si 0.00476	M U < 0.00628
M Ce < 0.01570	M In < 0.03139	n Os	M Ag < 0.00628	M V < 0.00628
M Cs < 0.00094	M Ir < 0.01570	M Pd < 0.01570	Q Na 0.00159	M Yb < 0.00314
M Cr < 0.01570	Q Fe < 0.00110	Q P < 0.00260	M Sr < 0.00157	M Y < 0.12558
M Co < 0.00942	M La < 0.00157	M Pt < 0.00628	Q S < 0.02500	Q Zn 0.00044
M Cu < 0.01884	M Pb < 0.00942	Q K 0.00048	M Ta < 0.02198	M Zr < 0.01570

M - Checked by ICP-MS Q - Checked by ICP-OES i - Spectral Interference n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

- For the calibration of analytical instruments including but not limited to the following: ICP-MS, ICP-OES, FAAS, GFAA, XRF, and DCP
- For the validation of analytical methods
- For the preparation of "working reference samples"
- For interference studies and the determination of correction coefficients
- For detection limit and linearity studies
- For additional intended uses, contact IV Technical Staff

11.0 DATE OF CERTIFICATION AND PERIOD OF VALIDITY

11.1 IV Shelf Life - The period of time during which the concentration of the analyte(s) in a properly packaged, unopened, and unused standard stored under environmentally controlled and monitored conditions will remain within the specified uncertainty range. Shelf life is limited primarily by transpiration (loss of water from the solution) and infrequently, by chemical instability. Transpiration studies (P-SP01020) of chemically-stable solutions performed at Inorganic Ventures Labs indicate a CRM shelf-life of four years for solutions packaged in 500-mL low density polyethylene bottles. When stored under special conditions that minimize transpiration and instability, the shelf life can be extended past this limit.

11.2 Expiration Date - The date after which a CRM should not be used. Routine laboratory use of a CRM increases transpiration losses and the chance of contamination which affect the integrity of the CRM and limit its useful life. Inorganic Ventures Labs concurs with state and federal regulatory agencies' recommendations that solution standards be assigned a one-year expiration date.

Certification Date: September 15, 2005

Expiration Date: **EXPIRES**
1st 2007

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By: Nick Maida, QA Administrator *Nick Maida*

Certificate Approved By: Katalin Le, QC Manager *Katalin Le*

Certifying Officer: Paul Gaines, PhD., Technical Director *Paul R. Gaines*

Initial:
Date:

WBH
4/16/07

METALS STANDARD DOCUMENTATION

Standard: Lead Stock Standard
Date Received/Prepped: 4/16/2007
Date Expired: 10/11/2008
Manufacturer: CPI
Matrix: 2% HNO3
Amount: 100 mL

ME #: 0704013
By: WBH
Lot #: 07A097
Certificate: Y
NIST SRM: 3128
Room temp. storage

Component	Comment	Conc. Unit:
Pb	P/N S4400-1000281	1000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 GS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

P/N S4400-1000281
P/N 4400-1000281
 Single-Element Lead Standard
 Pb in 2% HNO₃
 1000 ± 3 µg/mL

1270704013

Lot # 07A097

Material Source: Lead Metal
 Source Purity: 99.995 %
 Specific Gravity: 1.009 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3128. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 10.3	0.1	Cu 58	0.1	Pb X	0.1	K ND	70	Tl 0.25	0.1
Sb ND	0.1	Dy ND	0.1	Li 2	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.22	0.1	Eu ND	0.1	Mg 1.4	0.2	Rh INT	0.1	Sn ND	0.1
Be 0.58	0.1	Gd ND	0.1	Mn 3.8	1	Rb ND	0.1	Ti 0.58	0.1
Bi 0.7	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.17	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.9	0.1	Si 31	8	Yb ND	0.1
Ca 25	7	Ho ND	0.1	Nb ND	0.1	Ag 6.1	0.1	Y ND	0.1
Ce ND	0.1	I 0.1	0.2	Os ND	0.1	Na 3.5	1	Zn 23	2
Cs 0.26	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr INT	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.





Initial:

Date:

WJY
3/5/07

METALS STANDARD DOCUMENTATION

Standard: Selenium Stock Standard
Date Received/Prepped: 3/5/2007
Date Expired: 8/22/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0703001
By: wbn
Lot #: 6.00E+228
Certificate: Y
NIST SRM: 3148
Storage: Room Temp

Component	Comment	Conc. Unit:
Se	P/N # S4400-1000491	1000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

AUG 22 08

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-1000491
P/N S4400-1000491
 Single-Element Selenium Standard
 Se in 2% HNO₃
 1000 ± 3 µg/mL

ME0703001

Lot # 06E228

Material Source: Selenium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.011 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3148. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 1.8	0.1	Cu 0.4	0.1	Pb 0.3	0.1	K ND	70	Tl 3.6	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Eu ND	0.1	Mg 1.1	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.6	0.1	Sm ND	0.1	U ND	0.1
Br INT	10	Au ND	0.1	Nd ND	0.1	Se X	6	V ND	1
Cd 0.4	0.1	Hf ND	0.1	Ni ND	0.6	Si 40	8	Yb ND	0.1
Ca 5	7	Ho ND	0.1	Nb INT	0.1	Ag 0.8	0.1	Y ND	0.1
Ce ND	0.1	I 0.5	0.2	Os ND	0.1	Na 3.8	1	Zn ND	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr INT	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:
Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Thallium 1000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 2% HNO3
Amount: 100 mL

ME #: 0702006
By: WBH
Lot #: 06H213
Certificate: Y
NIST SRM: 3158
Room temp. storage

Component	Comment	Conc. Unit:
Tl	P/N S4400-1000581	1000 ppm

Aug 16 08



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

M70702006

CERTIFICATE OF ANALYSIS

P/N 4400-1000581
P/N S4400-1000581

Single Element Thallium Standard
Tl in 2% HNO₃
1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal
Source Purity: 99.999%
Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 13.3	0.1	Cu 9.3	0.1	Pb 41	0.1	K ND	70	Tl X	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.37	0.1	Eu ND	0.1	Mg 1.7	0.2	Rh ND	0.1	Sn ND	0.1
Be 0.67	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti 0.45	0.1
Bi 0.12	0.1	Ga ND	0.1	Hg 0.16	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.21	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd 1.6	0.1	Hf ND	0.1	Ni 1.1	0.1	Si 46	8	Yb ND	0.1
Ca 51	7	Ho ND	0.1	Nb ND	0.1	Ag 0.3	0.1	Y ND	0.1
Ce ND	0.1	I 0.4	0.2	Os ND	0.1	Na 3.3	1	Zn 14.7	2
Cs 0.24	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 20	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.



Initial:
Date:

WBJ
4/10/07

METALS STANDARD DOCUMENTATION

Standard: ICP LCS/SPIKE STOCK STD
Date Received/Prepped: 4/10/2007
Date Expired: 10/4/2008
Manufacturer: CPI
Matrix: 5% HNO3
Amount: 100mL

ME #: 0704005
By: wbj
Lot #: 07D019
Certificate: Y
NIST SRM: 3100 series
Storage: Room Temp

Component	Comment	Conc. Unit:
Ca	P/N # 4400-130309	10000 ppm
K	per 500mL DI	4000 ppm
Mg		4000 ppm
Na		10000 ppm



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

CERTIFICATE OF ANALYSIS

P/N: 4400-130309
Lot Number: 07D019
Shelf Life: 18 months
Expiration Date: 10/04/2008

ME 0704005

MWH
Dat MW Standard
 $\mu\text{g/mL} \pm 0.5\%$ in 5% HNO_3

Na 10,000 Ca 10,000 Mg 4,000 K 4,000

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000 $\mu\text{g/mL}$ by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: WJ
Date: 3/16/07

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working stock Solution
Date Received/Prepped: 3/16/2007
Date Expired: 9/16/2007
Manufacturer: MWH-WBH
Matrix: 5% HNO3
Amount: 2X100 mL

ME #: 0703010
By: WBH
Lot #:
Certificate:
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
Al	10mL ME0609001	5 ppm
Sb		5 ppm
As		10 ppm
Ba		2 ppm
Be		0.1 ppm
B		5 ppm
Cd		0.5 ppm
Ca		100 ppm
Cr		1 ppm
Co		5 ppm
Cu		1 ppm
Fe		2 ppm
Pb		2 ppm
Mg		10 ppm
Mn		0.2 ppm
Mo		2 ppm
Ni		2 ppm
K		100 ppm
Se		10 ppm
Ag		1 ppm
Na		100 ppm
Zn		2 ppm
V		0.2 ppm
Tl		10 ppm
Li		0.1 ppm
Ti		2 ppm
Sr		1 ppm
Sn		20 ppm


Initial: WBH
Date: 9/25/06

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Stock Solution
Date Received/Prepped: 9/25/2006
Date Expired: 9/20/2007
Manufacturer: CPI
Matrix: 2% HNO₃ + tr HF
Amount: 100 mL

ME #: 0609001
By: WBH
Lot #: 061162
Certificate: Yes
NIST SRM: 3100 series
Storage: Room Temp

Component	Comment	Conc. Unit:
Al	Part # 4400-060915RH01	50 ppm
Sb		50 ppm
As		100 ppm
Ba		20 ppm
Be		1 ppm
B		50 ppm
Cd		5 ppm
Ca		1000 ppm
Cr		10 ppm
Co		50 ppm
Cu		10 ppm
Fe		20 ppm
Pb		20 ppm
Mg		100 ppm
Mn		2 ppm
Mo		20 ppm
Ni		20 ppm
K		1000 ppm
Se		100 ppm
Ag		10 ppm
Na		1000 ppm
Zn		20 ppm
V		2 ppm
Tl		10 ppm
Li		1 ppm
Ti		20 ppm
Sr		10 ppm
Sn		200 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 9/20/2007

1470609001

Certificate of Analysis

Part Number: 4400-060915RH01
Lot Number: 061162
Shelf Life: 12 months

MWH
 Custom Standard
 2% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Al	50	Pb	20	Zn	20
Sb	50	Li	1	Ti	20
As	100	Mg	100	Sr	10
Ba	20	Mn	2	Sn	200
Be	1	Mo	20		
B	50	Ni	20		
Cd	5	K	1000		
Ca	1000	Se	100		
Cr	10	Ag	10		
Co	50	Na	1000		
Cu	10	TL	100		
Fe	20	V	2		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

colitag

PEAK PERFORMANCE
 CERTIFIED REFERENCE MATERIALS

Actusprep 7000™
 Extraction Manifold

MOD BLOCK

Nu-Phase™
 SPE Disks

ME0705012

Initial:

WBH

Date:

5/16/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Std A (ICSA)
Date Received/Prepped: 5/16/2007
Date Expired: 8/23/2007
Manufacturer: MWH-WBH
Matrix: 5% HNO3
Amount: 500 mL

ME #: 0705012
By: WBH
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Al	25mL ME0603001/500mL	250 ppm
Ca		250 ppm
Fe		100 ppm
Mg		250 ppm

Initial:

WBY

Date:

3/2/06

METALS STANDARD DOCUMENTATION

Standard: ICP ICSA Stock solution
Date Received/Prepped: 3/2/2006
Date Expired: 8/23/2007
Manufacturer: CPI
Matrix: 5% HNO₃
Amount: 500mL

ME #: 0603001
By: WBH
Lot #: 06A078
Certificate: Y
NIST SRM: varius
Storage: Room Temp

Component	Comment	Conc. Unit:
AL	P/N 4400-INTA1-500	5000 mg/L
CA		5000 mg/L
FE		2000 mg/L
MG		2000 mg/L

AUG 23 07



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500

CLP Interferents A Solution
in 5% HNO₃

Lot # 06A078

M80603001

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:
Date:

WBH
5/16/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Std AB (ICSAB)
Date Received/Prepped: 5/16/2007
Date Expired: 8/23/2007
Manufacturer: MWH-WBH
Matrix: 5% HNO₃
Amount: 500 mL

ME #: 0705013
By: WBH
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Al	25mL ME0603001/500mL	250 ppm
Ca		250 ppm
Fe		100 ppm
Mg		250 ppm
Ag	25mL ME0603002/500mL	0.5 ppm
Ba	2.5mL	0.25 ppm
Be	⊕ 5/16/07	0.25 ppm
Cd		0.5 ppm
Co		0.25 ppm
Cr		0.25 ppm
Cu		0.25 ppm
Mn		0.25 ppm
Ni		0.5 ppm
Pb		0.5 ppm
V		0.25 ppm
Zn		0.5 ppm

Initial:

WSY

Date:

3/2/06

METALS STANDARD DOCUMENTATION

Standard: ICP ICSA Stock solution
Date Received/Prepped: 3/2/2006
Date Expired: 8/23/2007
Manufacturer: CPI
Matrix: 5% HNO3
Amount: 500mL

ME #: 0603001
By: WBH
Lot #: 06A078
Certificate: Y
NIST SRM: various
Storage: Room Temp

Component	Comment	Conc. Unit:
AL	P/N 4400-INTA1-500	5000 mg/L
CA		5000 mg/L
FE		2000 mg/L
MG		2000 mg/L



USA

5580 Skylark Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

6 2 3 07

*Innovative Solutions
in Analytical Science and
Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500

**CLP Interferents A Solution
in 5% HNO₃**

Lot # 06A078

M80603001

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:

Date:

WBH
3/2/06

METALS STANDARD DOCUMENTATION

Standard: ICSB Stock Solution
Date Received/Prepped: 3/2/2006
Date Expired: 8/23/2007
Manufacturer: CPI
Matrix: 5% HNO3
Amount: 100 mL

ME #: 0603002
By: WBH
Lot #: 04L149
Certificate: Y
NIST SRM: 3100 series
Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	P/N 4400-INTB1-100	100 ppm
Ba		50 ppm
Be		50 ppm
Cd		100 ppm
Co		50 ppm
Cr		50 ppm
Mn		100 ppm
Ni		50 ppm
Pb		100 ppm
V		50 ppm
Zn		100 ppm
Sb		50 ppm



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

MF0603002

CERTIFICATE OF ANALYSIS

P/N 4400-INTB1-100
CLP Analytes B Solution
in 5% HNO₃

Lot # 04L149

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag 100	Ba 50	Be 50	Cd 100
Co 50	Cr 50	Cu 50	Mn 50
Ni 100	Pb 100	V 50	Zn 100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: WJY
 Date: 5/9/07

METALS STANDARD DOCUMENTATION

Standard: ICP QC-25 1PPM
 Date Received/Prepped: 5/9/2007
 Date Expired: 11/9/2007
 Manufacturer: MWH-DYH
 Matrix: 5% HNO3
 Amount: 500 mL

ME #: 0705005
 By: DYH
 Lot #: VARIOUS
 Certificate:
 NIST SRM:
 Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	5mL ME0608006+ 5mL ME0608007	1 ppm
Al	per 500mL DI	1
B		1
Ba		1
Be		1
Ca		1
Cd		1
Co		1
Cr		1
Cu		1
Fe		1
K		1
Li		10
Mg		1
Mn		1
Mo		1
Na		1
Ni		1
Pb		1
Sb		1
Se		1
Si		1
Sr		0.5
Ti		1
Tl		1
V		1
Zn		1

Initial: WxN
Date: 8/31/06

METALS STANDARD DOCUMENTATION

Standard: QC Check Std 7
Date Received/Prepped: 8/31/2006
Date Expired: 8/31/2007
Manufacturer: Crescent Chemical
Matrix: 5% HNO₃ + tr. HF
Amount: 100 mL

ME #: 0608007
By: WBH
Lot #: 062504K
Certificate: Y
NIST SRM: Various
Room temp. storage

Component	Comment	Conc. Unit:
Ag	QC-007.1	100 ppm
Al		100 ppm
B		100 ppm
Ba		100 ppm
K		100 ppm
Na		1000 ppm
Si		100 ppm
		50 ppm

ME0608007

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 7

CATALOG NO: QC-007.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F⁻

LOT NO.: 062504K

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3101a, 3104a, 3107, 3141a, 3150, 3151, and 3152a.

Concentrations are given in µg/mL unless noted otherwise.

Ag	100	Al	100	B	100	Ba	100	K	1,000±5
Na	100	Si	50.0						

Crescent Chemical Co. Inc.

Julie M. MacAnton

QA Manager

EXPIRES: August 2007

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913

Initial:
Date:

WBH
8/31/06

METALS STANDARD DOCUMENTATION

Standard: QC Check Std 21
Date Received/Prepped: 8/31/2006
Date Expired: 8/31/2007
Manufacturer: Crescent Chemical
Matrix: 5% HNO₃ + tr. Tartaric Acid
Amount: 100 mL

ME #: 0608006
By: WBH
Lot #: 062504J
Certificate: Y
NIST SRM: Various
Room temp. storage

Component	Comment	Conc. Unit:
AS	QC-021.1	100 ppm
Be		100 ppm
Ca		100 ppm
Cd		100 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
Li		100 ppm
Mg		100 ppm
Mn		100 ppm
Mo		100 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Sr		100 ppm
Ti		100 ppm
V		100 ppm
Zn		100 ppm

ME0608006

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 21

CATALOG NO: QC-021.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F/tr. Tartaric Acid

LOT NO.: 062504J

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3102a, 3103a, 3105a, 3108, 3109a, 3112a, 3113, 3114, 3126a, 3128, 3129a, 3131a, 3132, 3134, 3136, 3149, 3153a, 3158, 3162a, 3165, and 3168a.

Concentrations are given in µg/mL unless noted otherwise.

As 100	Be 100	Ca 100	Cd 100	Co 100
Cr 100	Cu 100	Fe 100	Li 100	Mg 100
Mn 100	Mo 100	Ni 100	Pb 100	Sb 100
Se 100	Sr 100	Ti 100	Tl 100	V 100
Zn 100				

Crescent Chemical Co. Inc.

Julie M. MacIntosh
QA Manager

EXPIRES: August 2007

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

Crescent Chemical Co, Inc., 2 Oval Dive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913

Initial:
Date:

WBH
8/9/07

METALS STANDARD DOCUMENTATION

Standard: ICP LINEARITY CHECK
Date Received/Prepped: 5/9/2007
Date Expired: 11/9/2007
Manufacturer: MWH-WBH
Matrix: 5% HNO3
Amount: 500mL

ME #: 0705004
By: WBH
Lot #:
Certificate: Y
NIST SRM: 3100 SERIES
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
CA	15.0mL ME0702002/500mL	300 ppm
K	15.0mL ME0702005/500mL	300 ppm
MG	10.0mL ME0702004/500mL	200 ppm
NA	15.0mL ME0702003/500mL	300 ppm
FE	5.0mL ME0701008/500mL	100 ppm

Initial:
Date:

W 3y
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO₃
Amount: 250 mL

ME #: 0702005
By: WBH
Lot #: 07B056
Certificate: Y
NIST SRM: 3141
Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

AUG 16 08



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 -31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

ME0702005

P/N 4400-10M411
P/N S4400-10M411
Single-Element Potassium Standard
K in 1% HNO₃
10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)
Source Purity: 99.9999%
Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 0.39	0.1	Cu 0.16	0.1	Pb ND	0.1	K X	70	Tl ND	0.1
Sb 0.34	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.14	0.1	Eu ND	0.1	Mg 2.6	0.2	Rh ND	0.1	Sn 0.17	0.1
Be ND	0.1	Gd ND	0.1	Mn 0.93	1	Rb 9.5	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
C ⁺ ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	20	Yb ND	0.1
Ca 82	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na 19	1	Zn 2.9	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:
Date:

WBH
2/12/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 250 mL

ME #: 0702004
By: WBH
Lot #: 07B058
Certificate: Y
NIST SRM: 3131
Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



USA
 5580 Skyline Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M311

P/N S4400-10M311

Single-Element Magnesium Standard

Mg in 4% HNO₃

10,000 ± 30 µg/mL

Lot # 07B058

Handwritten: 470702004

Material Source: Magnesium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 28	0.1	Cu 1.6	0.1	Pb 7.7	0.7	K ND	70	Tl 0.91	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr 0.28	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.28	0.1	Eu ND	0.1	Mg X	0.2	Rh ND	0.1	Sn 0.14	0.1
Be ND	0.1	Gd 0.23	0.1	Mn 19.8	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga 0.18	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd 1.1	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 1	0.1	Si 64	20	Yb ND	0.1
Ca ND	7	Ho ND	0.1	Nb ND	0.1	Ag 0.19	0.1	Y 0.2	0.1
Ce 2.1	0.1	I 1	0.2	Os ND	0.1	Na 7.2	1	Zn ND	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 0.19	0.1	Zr 0.29	0.1
Cr ND	1	Fe 80	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La 0.76	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:
Date:

U31
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702003
By: WBH
Lot #: 07B057
Certificate: Y
NIST SRM: 3152a
Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skyline Boulevard 707.525.5788
Santa Rosa, CA 95403 900.878.7654
www.cpiinternational.com Fax 707.546.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

P/N S4400-10M521

M70702003

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO₃)

Source Purity: 99.99%

Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 1.5	0.1	Cu 0.45	0.1	Pb ND	0.1	K ND	70	Tl ND	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.13	0.1	Eu ND	0.1	Mg 2.3	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Sc ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	8	Yb ND	0.1
Ca 120	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na X	1	Zn 2.9	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr i	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.



Initial:
Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 250 mL

ME #: 0702002
By: WBH
Lot #: 07B065
Certificate: Y
NIST SRM: 3109a
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA
5580 Skyline Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

170702602

Single-Element Calcium Standard
Ca in 4% HNO₃
10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)
Source Purity: 99.997%
Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Tl 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial: W34
Date: 1/27/07

METALS STANDARD DOCUMENTATION

Standard: FE 10000ppm Stock Std
Date Received/Prepped: 1/26/2007
Date Expired: 7/19/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 100 mL

ME #: 701008
By: WBH
Lot #: 061143
Certificate:
NIST SRM: 3126a
Storage: Room Temp

Component	Comment	Conc. Unit:
Fe	PN4400-10M261	10000 PPM



Innovative Solutions
in Analytical Science and
Technology

USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M261
P/N S4400-10M261
Single-Element Iron Standard

Fe in 4% HNO₃
10,000 ± 30 µg/mL

Lot # 06I143

ME0701008

Material Source: Iron Metal
Source Purity: 99.999%
Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al INT	0.1	Cu 6.4	0.1	Pb ND	0.1	K ND	70	Tl 0.18	0.1
Sb 0.35	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Eu ND	0.1	Mg 1.3	0.2	Rh ND	0.1	Sn 0.67	0.1
Be ND	0.1	Gd ND	0.1	Mn INT	1	Rb ND	0.1	Ti 0.21	0.1
Bi ND	0.1	Ga 0.41	0.1	Hg ND	0.2	Ru ND	0.1	W 0.13	0.1
B ND	4	Ge INT	0.1	Mo 4.9	0.1	Sr ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 9.3	0.1	Si INT	8	Yb ND	0.1
Ca 15	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.34	0.2	Os ND	0.1	Na 8	1	Zn 8.6	2
Cs 0.34	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr 3.3	1	Fe X	30	P 28	10	Ta ND	0.1		
Co 12	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



From Page No. _____

LOG #	CLIENT	SX - ID	MIX	VOLUME	COMMENTS
BLANK LCS LCS/D	200.7 DIRECT	08-08-07	JRF		HNO ₃ RE 100411 → (in) HCl RE 100412 → (2.5ml) LCS/SPIKE ME 0606004 ME 0705006 → (0.5ml) ME 0704005 → (0.25ml)
2708020271	KEPPMCLGEE.MP	I-B	AD	50ml → 50ml	
↓ MS ↓ MS		↓			
2708020272		I-AR			
↓ MS ↓ MS		↓			
2709020256		I-O			
0257		I-P			
0258		I-H			
0259		I-U			
0260		I-T			
0261		I-R			
0262		I-F			
0263		I-N			
0264		I-E			
0265		I-M			
0266		I-D			
0267		I-C			
0268		I-S			
0269		I-L			
0270		I-R			
0317		M-92			
0318		M-97			
0319		M-31A			
	200.7 DIRECT	JRF	08-08-07		

To Page No. _____

Witnessed & Understood by me,

Date

Invented by

Date

Recorded by

Project No. _____
Book No. _____

TITLE # 167

LOG #	CLIENT	SX-ID	MIX	VOLUME	COMMENTS	
	200.7 DIGEST	JRF	08-08-07		HNO ₃ R# 100411 → (1ml) HCl R# 100412 → (2.5ml) LCS/SPIKE	
2708020327	KERR MCGEE-MA	I-K	AR	50ml → 50ml	ME 0606004 ME 0705006 → (0.5ml) ME 0704005 → (0.25ml)	
↓ MS		↓				
↓ MSD						
2708020332			I-V			
↓ MS			↓			
↓ MSD						
2708020320			M-50			
0321			M-34			
0322			M-35			
0323			M-19			
0324			M-39			
0325			M-68			
0326			M-1			
0328			I-J			
0329			I-2			
0330			M-67			
0331			I-E			
0342			M-12A			
0343			EB-2			
✓ 0344			MD-2			
2708030224			M-87			
↓ 0225			M-74			
↓ 0226			M-73			
↓ 0227			M-88			
		200.7 DIGEST	08-08-07	08-08-07	JRF	

To Page No. _____

Witnessed & Understood by me,	Date	Invested by 153	Date
		Recorded by	