

## TABLE OF CONTENTS

COVER PAGE.....	2
QC CHECKLIST .....	3
SUMMARY SHEET .....	3
ANALYTICAL SEQUENCE .....	6
RUNLOG.....	12
INITIAL CALIBRATION.....	17
PERIODIC QC.....	35
QC: (MBLK, MRL, LCS1, LCS2).....	45
QC: (MS 2805080991).....	50
PERIODIC QC.....	51
QC: (MSD 2805080991) .....	54
PERIODIC QC.....	55
QC: (MBLK) .....	58
PERIODIC QC.....	59
QC: (MRL, LCS1, LCS2) .....	62
QC: (MS/MSD 2805090119) .....	66
QC: (MS/MSD 2805090120) .....	68
PERIODIC QC.....	70
SAMPLES .....	77
STANDARDS PREPARATION WORKSHEET AND CERTIFICATES OF ANALYSIS.....	96

# Level IV Data Package

MWH Group 240233

**Method: EPA 6010B CRVI**

2805090112  
2805090115  
2805090116  
2805090117  
2805090118  
2805090119  
2805090120  
2805090121  
2805090122  
2805090123  
2805090124  
2805090132  
2805090133  
2805090134  
2805090135  
2805090136  
2805090137  
2805090138

EPA 200.7/6010B QC Check List

Analyst CSK Analysis Date 5-21-08 Reviewer/Date 5-22-08

Instrument PerKin Elmer Optima 4300DV

- All sample analyzed within 6 month holding time
- All sample raw concentration below the high standard or linear range o marked for dilution and rerun

Initial and closing QC

- ICV within +/- 5%
- Linearity check +/- 10%
- ICSAB +/- 20%
- QCS +/- 5%
- MRL +/- 50%

Middle, closing and batch QC

- FilterCheck < 1/2 MRL
- MBLANK < 1/2 MRL
- LCS +/-15%
- MS/MSD +/-30% (200.7) +/- 25% (6010B)
- CCV/MCV/ECV +/- 10%
- ICB/CCB/ECB < 1/2 MRL
- CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/-20%
- RPD between LCS/LCSD is within +/- 20%
- Internal standards 60 TO 125%
- All pH of the samples are < 2
  
- No more than 20 samples per batch
- MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples
- yes QIR needed for failed QC
- NA Special Det Code Cr 6 to 10, Ni 6 to 10 noted on the cover sheet
- NA R value for multi point calibration is > 0.995
- NA Proper MRL check ran for special low MRL samples

Reagent and Standards used for  
Optima 4300 DV  
Updated 03/31/08

Int: CSK  
Date: 5-22-08

Method 200.7/6010

# ICP SUMMARY SHEET

File ID: 080521  
 Date Started: 5/21/08  
 Analyst ID: CSK

## SAMPLE ID

LINEARITY (13:35)	LINEARITY (13:42)	Wash (13:48)
FILTERCHECK (14:17)	2805080195_10XX (14:34)	2805070675_10XX (14:47)
2805070677_10XX (15:00)	2805070678_10XX (15:04)	2805070679_10XX (15:08)
2805190158 (15:25)	D805150230 (15:45)	2805190370 (15:49)
2805200055 (15:53)	2805200009 (15:58)	2805200010 (16:02)
2805200011 (16:06)	2805200012 (16:11)	2805200013 (16:15)
2805200014 (16:19)	2805190004 (16:24)	2805200320 (17:44)
2805200459 (17:47)	2805200453 (17:51)	2805200454 (17:55)
2805200455 (17:59)	2805200215 (18:03)	2805200603 (18:07)
2805200505 (18:11)	2805200507 (18:26)	2805200510 (18:42)
2805200509 (18:53)	2805200511 (18:57)	2805200512 (19:08)
2805200064 (19:12)	2805200068 (19:16)	2805200070 (19:20)
2805200605 (19:24)	2805200607 (19:28)	2805200608 (19:33)
2805200609 (19:37)	2805200280_2 (19:57)	2805090186_2 (20:18)
2805090189_2 (20:28)	2805090181_2 (20:44)	2805090182_2 (20:48)
2805090184_2 (20:53)	2805090185_2 (20:57)	2805100128_2 (21:01)
2805100129_2 (21:06)	2805100130_2 (21:10)	2805100131_2 (21:14)
2805100132_2 (21:36)	2805100135_2 (21:41)	2805100137_2 (21:45)
2805100138_1 (21:49)	2805100139_2 (21:53)	2805100140_2 (21:57)
2805130453_2 (22:02)	2805130454_2 (22:06)	2805140073_2 (22:10)
2805140074_2 (22:14)	<u>2805080677_2</u> <sup>280</sup> <sub>52-08</sub> (22:54)	2805080991_2 (23:03)
2805080656_5 (23:19)	2805080657_1 (23:22)	2805080665_2 (23:25)
2805080666_5 (23:31)	2805080668_5 (23:34)	2805080673_5 (23:37)
2805080681_1 (23:40)	2805080682_1 (23:43)	2805080683_1 (23:46)
2805080684_5 (23:55)	2805080685_5 (23:58)	2805080686_5 (0:01)
2805080688_5 (0:04)	2805080695_2 (0:07)	2805080696_2 (0:10)
<u>2805080697_2</u> (0:13)	2805080999_2 (0:16)	2805090112_5 (0:19)
2805090119_2 (0:45)	2805090120_2 (0:54)	2805090115_5 (1:09)
2805090116_2 (1:12)	2805090117_5 (1:15)	2805090118_5 (1:18)
2805090121_5 (1:21)	2805090122_5 (1:24)	2805090123_5 (1:27)
2805090124_1 (1:30)	2805090132_1 (1:33)	2805090133_1 (1:36)
2805090134_5 (1:47)	2805090135_5 (1:50)	2805090136_1 (1:53)
2805090137_1 (1:56)	2805090138_5 (1:59)	2805090152_5 (2:02)
2805090157_5 (2:05)	2805090177_2 (2:08)	

COMMENT:

- 1) Na bold samples. due to the spike sample 280508195 concentration is too high even with 10X, the result is too low. Need QIR.  
(recovery)
- 2) sample from 2805200280 Are Cr bold samples.
- 3) 2805080677 → 2805080667 } Typing Error  
2805080697 → 2805080997

Analyst: CSK  
 5-22-08

Approved By: WBJ

WBJ MM 6/3/08

BATCH NUMBER for 080521

Test Parameter:

SCA YR AG AL AS B\_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2805080195\_10X

2805080195_10X	2805070675_10X	2805070677_10X
2805070678_10X	2805070679_10X	

Batch ID: 2805190158

2805190158	2805190370	2805200055
2805200009	2805200010	2805200011
2805200012	2805200013	2805200014
2805190004	2805200320	2805200459
2805200453	2805200454	2805200455
2805200215	2805200603	2805200505
2805200507		

Batch ID: 2805200510

2805200510	2805200509	2805200511
2805200512	2805200064	2805200068
2805200070	2805200605	2805200607
2805200608	2805200609	2805200280_2X

Batch ID: 2805090186\_2X

2805090186_2X	2805090189_2X	2805090181_2X
2805090182_2X	2805090184_2X	2805090185_2X
2805100128_2X	2805100129_2X	2805100130_2X
2805100131_2X	2805100132_2X	2805100135_2X
2805100137_2X	2805100138_10X	2805100139_2X
2805100140_2X	2805130453_2X	2805130454_2X
2805140073_2X	2805140074_2X	

Batch ID: 2805080677\_2X

2805080677_2X	2805080991_2X	2805080656_5X
2805080657_10X	2805080665_2X	2805080666_5X
2805080668_5X	2805080673_5X	2805080681_10X
2805080682_10X	2805080683_10X	2805080684_5X
2805080685_5X	2805080686_5X	2805080688_5X
2805080695_2X	2805080696_2X	2805080697_2X
2805080999_2X	2805090112_5X	

Batch ID: 2805090119\_2X

2805090119_2X	2805090120_2X	2805090115_5X
2805090116_2X	2805090117_5X	2805090118_5X

2805090121\_5X  
2805090124\_10X  
2805090134\_5X  
2805090137\_10X  
2805090157\_5X

2805090122\_5X  
2805090132\_10X  
2805090135\_5X  
2805090138\_5X  
2805090177\_2X

2805090123\_5X  
2805090133\_10X  
2805090136\_10X  
2805090152\_5X

Analytical Sequence

Method : 200.7&6010\_080304

Seq.	Loc.	Sample ID	Status
1	15	ICV	
2	9	LINEARITY	
3	10	ICSA	
4	11	ICSAB	
5	0	Wash	
6	12	QC-25 1ppm	
7	4	CCV	
8	0	ICB	
9	20	MRL	
10	16	MRL/2	
11	18	FILTERCHECK	
12	38	MBLANK2007	
13	24	MRL2007	
14	39	LCS2007	
15	40	LCSD2007	
16	41	2805080195_10X	
17	42	2805080195_10XMS	
18	43	2805080195_10XMSD	
19	44	2805070675_10X	
20	4	CCV	
21	0	CCB	QC Passed
22	45	2805070677_10X	Analyzed
23	46	2805070678_10X	Analyzed
24	47	2805070679_10X	Analyzed
25	48	MBLANK	Analyzed
26	21	MRL	Analyzed
27	49	LCS	Analyzed
28	50	LCSD	Analyzed
29	51	2805190158	Analyzed
30	52	2805190158MS	Analyzed
31	53	2805190158MSD	Analyzed
32	4	CCV	QC Passed
33	0	CCB	QC Passed
34	5	MCV	QC Passed
35	54	D805150230	Analyzed
36	55	2805190370	Analyzed
37	56	2805200055	Analyzed
38	57	2805200009	Analyzed
39	58	2805200010	Analyzed
40	59	2805200011	Analyzed
41	60	2805200012	Analyzed
42	61	2805200013	Analyzed
43	62	2805200014	Analyzed
44	63	2805190004	Analyzed
45	4	CCV	QC Passed
46	0	CCB	QC Passed
47	64	2805190004MS	Analyzed
48	65	2805190004MSD	Analyzed
49	66	2805150826	
50	67	2805150100_5X	
51	68	2805150132_5X	
52	69	2805150132_100X	
53	70	2805080368	
54	71	2805160299	
55	72	2805130392	
56	73	2805141006	

\*\*\*\*\*  
Analytical Sequence

Method: 200.7&amp;6010\_080304

Seq.	Loc.	ID	Status
1	15	ICV	
2	9	LINEARITY	
3	10	ICSA	
4	11	ICSAB	
5	0	Wash	
6	12	QC-25 1ppm	
7	4	CCV	
8	0	ICB	
9	20	MRL	
10	16	MRL/2	
11	18	FILTERCHECK	
12	38	MBLANK2007	
13	24	MRL2007	
14	39	LCS2007	
15	40	LCSD2007	
16	41	2805080195_10X	
17	42	2805080195_10XMS	
18	43	2805080195_10XMSD	
19	44	2805070675_10X	
20	4	CCV	
21	0	CCB	
22	45	2805070677_10X	
23	46	2805070678_10X	
24	47	2805070679_10X	
25	48	MBLANK	
26	21	MRL	
27	49	LCS	
28	50	LCSD	
29	51	2805190158	
30	52	2805190158MS	
31	53	2805190158MSD	
32	4	CCV	
33	0	CCB	
34	5	MCV	
35	54	D805150230	
36	55	2805190370	
37	56	2805200055	
38	57	2805200009	
39	58	2805200010	
40	59	2805200011	
41	60	2805200012	
42	61	2805200013	
43	62	2805200014	
44	63	2805190004	
45	4	CCV	
46	0	CCB	
47	64	2805190004MS	
48	65	2805190004MSD	
49	66	2805200320	
50	67	2805200459	
51	68	2805200453	
52	69	2805200454	
53	70	2805200455	
54	71	2805200215	
55	72	2805200603	
56	73	2805200505	
57	4	CCV	
58	0	CCB	
59	5	MCV	
60	74	2805200507	Analyzed
61	75	MBLANK	Analyzed
62	21	MRL	Analyzed
63	76	LCS	Analyzed
64	77	LCSD	Analyzed
65	78	2805200510	Analyzed
66	79	2805200510MS	Analyzed
67	80	2805200510MSD	Analyzed



68	81	2805200509	Analyzed
69	82	2805200511	Analyzed
70	4	CCV	QC Passed
71	0	CCB	QC Passed
72	83	2805200512	Analyzed
73	84	2805200054	Analyzed
74	85	2805200068	Analyzed
75	86	2805200070	Analyzed
76	87	2805200605	Analyzed
77	88	2805200607	Analyzed
78	89	2805200608	Analyzed
79	90	2805200609	Analyzed
80	91	2805200609MS	Analyzed
81	92	2805200609MSD	Analyzed
82	4	CCV	QC Passed
83	0	CCB	QC Passed
84	5	MCV	QC Passed
85	93	2805200280_2X	Analyzed
86	94	MBLANK2007	Analyzed
87	21	MRL	Analyzed
88	95	MRL2007	Analyzed
89	96	LCS2007	Analyzed
90	97	LCSD2007	Analyzed
91	98	2805090186_2X	Analyzed
92	99	2805090186_2XMS	Analyzed
93	100	2805090186_2XMSD	Analyzed
94	101	2805090189_2X	Analyzed
95	4	CCV	QC Passed
96	0	CCB	QC Passed
97	102	2805090189_2XMS	Analyzed
98	103	2805090189_2XMSD	Analyzed
99	104	2805090181_2X	Analyzed
100	105	2805090182_2X	Analyzed
101	106	2805090184_2X	Analyzed
102	107	2805090185_2X	Analyzed
103	108	2805100128_2X	Analyzed
104	109	2805100129_2X	Analyzed
105	110	2805100130_2X	Analyzed
106	111	2805100131_2X	Analyzed
107	4	CCV	QC Failed
108	4	CCV	QC Failed
109	4	CCV	QC Passed
110	0	CCB	QC Failed
111	0	CCB	QC Failed
112	0	CCB	QC Failed
113	5	MCV	QC Passed
114	112	2805100132_2X	Analyzed
115	113	2805100135_2X	Analyzed
116	114	2805100137_2X	Analyzed
117	115	2805100138_10X	Analyzed
118	116	2805100139_2X	Analyzed
119	117	2805100140_2X	Analyzed
120	118	2805130453_2X	Analyzed
121	119	2805130454_2X	Analyzed
122	120	2805140073_2X	Analyzed
123	121	2805140074_2X	Analyzed
124	4	ECV	QC Failed
125	4	ECV	QC Failed
126	0	ECB	QC Passed

## \* Analytical Sequence

Method: 200.7&amp;6010 080304

Seq.	Loc.	ID	Status
1	15	ICV	
2	9	LINEARITY	
3	10	ICSA	
4	11	ICSAB	
5	0	Wash	
6	12	QC-25 1ppm	
7	4	CCV	
8	0	ICB	
9	20	MRL	
10	16	MRL/2	
11	18	FILTERCHECK	
12	38	MBLANK2007	
13	24	MRL2007	
14	39	LCS2007	
15	40	LCSD2007	
16	41	2805080195_10X	
17	42	2805080195_10XMS	
18	43	2805080195_10XMSD	
19	44	2805070675_10X	
20	4	CCV	
21	0	CCB	
22	45	2805070677_10X	
23	46	2805070678_10X	
24	47	2805070679_10X	
25	48	MBLANK	
26	21	MRL	
27	49	LCS	
28	50	LCSD	
29	51	2805190158	
30	52	2805190158MS	
31	53	2805190158MSD	
32	4	CCV	
33	0	CCB	
34	5	MCV	
35	54	D805150230	
36	55	2805190370	
37	56	2805200055	
38	57	2805200009	
39	58	2805200010	
40	59	2805200011	
41	60	2805200012	
42	61	2805200013	
43	62	2805200014	
44	63	2805190004	
45	4	CCV	
46	0	CCB	
47	64	2805190004MS	
48	65	2805190004MSD	
49	66	2805200320	
50	67	2805200459	
51	68	2805200453	
52	69	2805200454	
53	70	2805200455	
54	71	2805200215	
55	72	2805200603	
56	73	2805200505	
57	4	CCV	
58	0	CCB	
59	5	MCV	
60	74	2805200507	
61	75	MBLANK	
62	21	MRL	
63	76	LCS	
64	77	LCSD	
65	78	2805200510	
66	79	2805200510MS	
67	80	2805200510MSD	

68	81	2805200509	
69	82	2805200511	
70	4	CCV	
71	0	CCB	
72	83	2805200512	
73	84	2805200064	
74	85	2805200068	
75	86	2805200070	
76	87	2805200605	
77	88	2805200607	
78	89	2805200608	
79	90	2805200609	
80	91	2805200609MS	
81	92	2805200609MSD	
82	4	CCV	
83	0	CCB	
84	5	MCV	
85	93	2805200280_2X	
86	94	MBLANK2007	
87	21	MRL	
88	95	MRL2007	
89	96	LCS2007	
90	97	LCSD2007	
91	98	2805090186_2X	
92	99	2805090186_2XMS	
93	100	2805090186_2XMSD	
94	101	2805090189_2X	
95	4	CCV	
96	0	CCB	
97	102	2805090189_2XMS	
98	103	2805090189_2XMSD	
99	104	2805090181_2X	
100	105	2805090182_2X	
101	106	2805090184_2X	
102	107	2805090185_2X	
103	108	2805100128_2X	
104	109	2805100129_2X	
105	110	2805100130_2X	
106	111	2805100131_2X	
107	4	CCV	
108	0	CCB	
109	5	MCV	
110	112	2805100132_2X	
111	113	2805100135_2X	
112	114	2805100137_2X	
113	115	2805100138_10X	
114	116	2805100139_2X	
115	117	2805100140_2X	
116	118	2805130453_2X	
117	119	2805130454_2X	
118	120	2805140073_2X	
119	121	2805140074_2X	
120	4	CCV	QC Failed
121	4	CCV	QC Failed
122	4	CCV	QC Failed
123	0	CCB	QC Passed
124	122	MBLANK2007	Analyzed
125	21	MRL	Analyzed
126	24	MRL2007	Analyzed
127	123	LCS2007	Analyzed
128	124	LCSD2007	Analyzed
129	125	2805080677_2X	Analyzed
130	126	2805080677_2XMS	Analyzed
131	127	2805080677_2XMSD	Analyzed
132	128	2805080991_2X	Analyzed
133	129	2805080991_2XMS	Analyzed
134	4	CCV	QC Passed
135	0	CCB	QC Passed
136	5	MCV	QC Passed
137	130	2805080991_2XMSD	Analyzed
138	131	2805080656_5X	Analyzed
139	132	2805080657_10X	Analyzed

140	133	2805080665_2X	Analyzed
141	134	2805080666_5X	Analyzed
142	135	2805080668_5X	Analyzed
143	136	2805080673_5X	Analyzed
144	137	2805080681_10X	Analyzed
145	138	2805080682_10X	Analyzed
146	139	2805080683_10X	Analyzed
147	4	CCV	QC Passed
148	0	CCB	QC Passed
149	140	2805080684_5X	Analyzed
150	141	2805080685_5X	Analyzed
151	142	2805080686_5X	Analyzed
152	143	2805080688_5X	Analyzed
153	144	2805080695_2X	Analyzed
154	145	2805080696_2X	Analyzed
155	146	2805080697_2X	Analyzed
156	147	2805080999_2X	Analyzed
157	148	2805090112_5X	Analyzed
158	38	MBLANK2007	Analyzed
159	4	CCV	QC Passed
160	0	CCB	QC Passed
161	5	MCV	QC Passed
162	21	MRL	Analyzed
163	24	MRL2007	Analyzed
164	39	LCS2007	Analyzed
165	40	LCSD2007	Analyzed
166	41	2805090119_2X	Analyzed
167	42	2805090119_2XMS	Analyzed
168	43	2805090119_2XMSD	Analyzed
169	44	2805090120_2X	Analyzed
170	45	2805090120_2XMS	Analyzed
171	46	2805090120_2XMSD	Analyzed
172	4	CCV	QC Passed
173	0	CCB	QC Passed
174	47	2805090115_5X	Analyzed
175	48	2805090116_2X	Analyzed
176	49	2805090117_5X	Analyzed
177	50	2805090118_5X	Analyzed
178	51	2805090121_5X	Analyzed
179	52	2805090122_5X	Analyzed
180	53	2805090123_5X	Analyzed
181	54	2805090124_10X	Analyzed
182	55	2805090132_10X	Analyzed
183	56	2805090133_10X	Analyzed
184	4	CCV	QC Passed
185	0	CCB	QC Passed
186	5	MCV	QC Passed
187	57	2805090134_5X	Analyzed
188	58	2805090135_5X	Analyzed
189	59	2805090136_10X	Analyzed
190	60	2805090137_10X	Analyzed
191	61	2805090138_5X	Analyzed
192	62	2805090152_5X	Analyzed
193	63	2805090157_5X	Analyzed
194	64	2805090177_2X	Analyzed
195	4	ECV	QC Passed
196	0	ECB	QC Passed

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	5/21/08	13:32	1	9.9459	9.95	95-105	99.4%
LINEARITY	5/21/08	13:35	1	0.0032	.0032		
ICSA	5/21/08	13:39	1	0.0005	0.0005	80-120	
LINEARITY	5/21/08	13:42	1	0.0025	.0025		
ICSAB	5/21/08	13:45	1	.25183	.252	80-120	100%
Wash	5/21/08	13:48	1	0.0000	0.0000		
QC-25 1ppm	5/21/08	13:56	1	.98132	.980		
CCV	5/21/08	13:59	1	4.9141	4.91	90-110	98.2%
MCV	5/21/08	14:03	1	2.5189	2.52	90-110	100%
ICB	5/21/08	14:07	1	0.0001	0.0001		
MRL	5/21/08	14:10	1	0.0103	.0103	50-150	103%
MRL/2	5/21/08	14:14	1	0.0057	.0057		
FILTERCHECK	5/21/08	14:17	1	0.0001	0.0000		
MBLANK2007	5/21/08	14:21	1	-0.0001	ND		
MRL2007	5/21/08	14:25	1	0.0110	.011		
LCS2007	5/21/08	14:28	1	.94745	.947	85-115	94.7%
LCS2007	5/21/08	14:31	1	.98383	.984	85-115	98.3%
2805080195_10X	5/21/08	14:34	10	0.0770	.077		
2805080195_10XMS	5/21/08	14:39	10	1.0525	1.05	[ 0.976]	9.75 Q
2805080195_10XMSD	5/21/08	14:43	10	1.0692	1.07	[ 0.992]	9.92 Q
2805080195_10XT	5/21/08	14:43	10		10.00	70 - 130	
2805070675_10X	5/21/08	14:47	10	0.0003	0.0003		
CCV	5/21/08	14:51	1	5.0401	5.04	90-110	100%
CCB	5/21/08	14:56	1	-0.0001	ND		
2805070677_10X	5/21/08	15:00	10	0.0018	.0018		
2805070678_10X	5/21/08	15:04	10	0.0004	0.0003		
2805070679_10X	5/21/08	15:08	10	0.0009	0.0009		
MBLANK	5/21/08	15:12	1	-0.0000	ND		
MRL	5/21/08	15:16	1	0.0097	.0097	50-150	97.4%
LCS	5/21/08	15:20	1	.94261	.943	85-115	94.2%
LCSD	5/21/08	15:22	1	.92615	.926	85-115	92.6%
2805190158	5/21/08	15:25	1	0.0052	.0052		
2805190158MS	5/21/08	15:29	1	.92652	.927	[ 0.921]	92.1%
2805190158MSD	5/21/08	15:32	1	.93229	.932	[ 0.927]	92.7%
2805190158T	5/21/08	15:32	1		1.00	70 - 130	
CCV	5/21/08	15:35	1	5.0709	5.07	90-110	101%
CCB	5/21/08	15:38	1	-0.0000	ND		
MCV	5/21/08	15:42	1	2.4505	2.45	90-110	98.0%
D805150230	5/21/08	15:45	1	-0.0004	ND		
2805190370	5/21/08	15:49	1	-0.0005	ND		
2805200055	5/21/08	15:53	1	-0.0003	ND		
2805200009	5/21/08	15:58	1	-0.0004	ND		
2805200010	5/21/08	16:02	1	0.0012	.0012		
2805200011	5/21/08	16:06	1	-0.0003	ND		
2805200012	5/21/08	16:11	1	0.0004	0.0003		
2805200013	5/21/08	16:15	1	0.0004	0.0004		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2805200014	5/21/08	16:19	1	0.0008	0.0008		
2805190004	5/21/08	16:24	1	-0.0002	ND		
CCV	5/21/08	16:27	1	5.0236	5.02	90-110	100%
CCB	5/21/08	16:31	1	0.0001	0.0000		
2805190004MS	5/21/08	16:34	1	.92541	.925	[ 0.925]	92.5%
2805190004MSD	5/21/08	16:37	1	.93946	.939	[ 0.939]	93.9%
2805190004T	5/21/08	16:37	1		1.00	70 - 130	
CCV	5/21/08	17:35	1	5.1021	5.1	90-110	102%
CCB	5/21/08	17:38	1	-0.0000	ND		
2805200320	5/21/08	17:44	1	0.0005	0.0004		
2805200459	5/21/08	17:47	1	-0.0004	ND		
2805200453	5/21/08	17:51	1	0.0016	.0016		
2805200454	5/21/08	17:55	1	0.0015	.0015		
2805200455	5/21/08	17:59	1	-0.0004	ND		
2805200215	5/21/08	18:03	1	-0.0004	ND		
2805200603	5/21/08	18:07	1	0.0081	.0081		
2805200505	5/21/08	18:11	1	-0.0007	ND		
CCV	5/21/08	18:15	1	5.0637	5.06	90-110	101%
CCB	5/21/08	18:18	1	0.0001	0.0000		
MCV	5/21/08	18:21	1	2.5868	2.59	90-110	103%
2805200507	5/21/08	18:26	1	-0.0006	ND		
MBLANK	5/21/08	18:30	1	0.0001	0.0001		
MRL	5/21/08	18:34	1	0.0099	.0099	50-150	99.0%
LCS	5/21/08	18:37	1	.97513	.975	85-115	97.5%
LCSD	5/21/08	18:40	1	.95523	.955	85-115	95.5%
2805200510	5/21/08	18:42	1	-0.0002	ND		
2805200510MS	5/21/08	18:46	1	.95724	.957	[ 0.957]	95.7%
2805200510MSD	5/21/08	18:50	1	.97185	.972	[ 0.972]	97.1%
2805200510T	5/21/08	18:50	1		1.00	70 - 130	
2805200509	5/21/08	18:53	1	0.0000	0		
2805200511	5/21/08	18:57	1	-0.0005	ND		
CCV	5/21/08	19:01	1	5.1491	5.15	90-110	102%
CCB	5/21/08	19:05	1	0.0002	0.0001		
2805200512	5/21/08	19:08	1	-0.0006	ND		
2805200064	5/21/08	19:12	1	0.0252	.025		
2805200068	5/21/08	19:16	1	0.0222	.022		
2805200070	5/21/08	19:20	1	0.0228	.023		
2805200605	5/21/08	19:24	1	-0.0003	ND		
2805200607	5/21/08	19:28	1	-0.0002	ND		
2805200608	5/21/08	19:33	1	-0.0003	ND		
2805200609	5/21/08	19:37	1	-0.0003	ND		
2805200609MS	5/21/08	19:41	1	.96268	.963	[ 0.963]	96.2%
2805200609MSD	5/21/08	19:44	1	.95927	.959	[ 0.959]	95.9%
2805200609T	5/21/08	19:44	1		1.00	70 - 130	
CCV	5/21/08	19:47	1	5.2577	5.26	90-110	105%
CCB	5/21/08	19:51	1	0.0001	0.0000		
MCV	5/21/08	19:54	1	2.6140	2.61	90-110	104%
2805200280_2X	5/21/08	19:57	2	.21786	.220		
MBLANK2007	5/21/08	20:02	1	0.0002	0.0002		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
MRL	5/21/08	20:05	1	0.0105	.0105	50-150	104%
MRL2007	5/21/08	20:09	1	0.0115	.0115		
LCS2007	5/21/08	20:12	1	1.0615	1.06	85-115	106%
LCSD2007	5/21/08	20:16	1	.97612	.976	85-115	97.6%
2805090186_2X	5/21/08	20:18	2	0.0001	0.0000		
2805090186_2XMS	5/21/08	20:22	2	.99959	1.00	[ 1.000]	49.9 Q
2805090186_2XMSD	5/21/08	20:25	2	.99619	.996	[ 0.996]	49.8 Q
2805090186_2XT	5/21/08	20:25	2		2.00	70 - 130	99
2805090189_2X	5/21/08	20:28	2	0.0000	0.0000		
CCV	5/21/08	20:31	1	5.2437	5.24	90-110	104%
CCB	5/21/08	20:35	1	0.0001	0.0001		
2805090189_2XMS	5/21/08	20:38	2	1.0084	1.01	[ 1.008]	56.4 Q
2805090189_2XMSD	5/21/08	20:41	2	1.0628	1.06	[ 1.063]	53.1 Q
2805090189_2XT	5/21/08	20:41	2		2.00	70 - 130	106
2805090181_2X	5/21/08	20:44	2	.09997	.1		
2805090182_2X	5/21/08	20:48	2	0.0782	.078		
2805090184_2X	5/21/08	20:53	2	0.0104	.010		
2805090185_2X	5/21/08	20:57	2	.01203	.012		
2805100128_2X	5/21/08	21:01	2	0.0021	.0021		
2805100129_2X	5/21/08	21:06	2	.55388	.550		
2805100130_2X	5/21/08	21:10	2	.91655	.920		
2805100131_2X	5/21/08	21:14	2	.73133	.730		
CCV	5/21/08	21:21	1	5.1896	5.19	90-110	103%
CCB	5/21/08	21:30	1	0.0001	0.0001		
MCV	5/21/08	21:33	1	2.5769	2.58	90-110	103%
2805100132_2X	5/21/08	21:36	2	.49203	.490		
2805100135_2X	5/21/08	21:41	2	0.0009	0.0009		
2805100137_2X	5/21/08	21:45	2	0.0002	0.0002		
2805100138_10X	5/21/08	21:49	10	34.496	34		
2805100139_2X	5/21/08	21:53	2	0.0306	.031		
2805100140_2X	5/21/08	21:57	2	0.0576	.058		
2805130453_2X	5/21/08	22:02	2	.11898	.120		
2805130454_2X	5/21/08	22:06	2	.11419	.110		
2805140073_2X	5/21/08	22:10	2	0.0075	.0075		
2805140074_2X	5/21/08	22:14	2	0.0780	.078		
ECV	5/21/08	22:19	1	5.0827	5.08	90-110	101%
ECB	5/21/08	22:22	1	0.0004	0.0003		
CCV	5/21/08	22:34	1	5.0988	5.1	90-110	101%
CCB	5/21/08	22:37	1	0.0001	0.0001		
MBLANK2007	5/21/08	22:39	1	-0.0001	ND		
MRL	5/21/08	22:42	1	0.0101	.0101	50-150	100%
MRL2007	5/21/08	22:45	1	0.0115	.0115		
LCS2007	5/21/08	22:48	1	.98502	.985	85-115	98.5%
LCSD2007	5/21/08	22:51	1	.95341	.953	85-115	95.3%
2805080607_2X	5/21/08	22:54	2	.32905	.330		
2805080607_2XMS	5/21/08	22:57	2	1.2844	1.28	[ 0.955]	47.7 Q
2805080607_2XMSD	5/21/08	23:00	2	1.2763	1.28	[ 0.947]	47.8 Q
2805080607_2XT	5/21/08	23:00	2		2.00	70 - 130	0.95

CSK  
5-22-08

Sample ID	Date	Time	Dil.	Raw	Rept.	Limit	Comment
2805080991_2X	5/21/08	23:03	2	0.0003	0.0003 ✓		0.957.
2805080991_2XMS	5/21/08	23:06	2	.94970	.95 ✓	[ 0.950]	47.4 Q
CCV	5/21/08	23:09	1	5.0469	5.05	90-110	100%
CCB	5/21/08	23:12	1	0.0001	0.0000		
MCV	5/21/08	23:14	1	2.5375	2.54	90-110	101%
2805080991_2XMSD	5/21/08	23:17	2	1.0000	1.00	[ 1.000]	50.0 Q
2805080991_2XT	5/21/08	23:17	2		2.00 ✓	70 - 130	100%
2805080656_5X	5/21/08	23:19	(5)	7.0694	7.1 ✓		
2805080657_10X	5/21/08	23:22	(10)	31.134	31 ✓		CSK
2805080665_2X	5/21/08	23:25	2	0.0093	.0093 ✓		J-22 of
CCV	5/21/08	23:28	1	4.9527	4.95	90-110	99.0%
2805080666_5X	5/21/08	23:31	(5)	1.3697	1.4 ✓		
2805080668_5X	5/21/08	23:34	(5)	4.9169	4.9 ✓		
2805080673_5X	5/21/08	23:37	(5)	1.1177	1.1 ✓		
2805080681_10X	5/21/08	23:40	(10)	23.713	24 ✓		
2805080682_10X	5/21/08	23:43	(10)	22.215	22 ✓		
2805080683_10X	5/21/08	23:46	(10)	11.408	11 ✓		
CCV	5/21/08	23:49	1	4.9831	4.98	90-110	99.6%
CCB	5/21/08	23:52	1	0.0015	.0015		
2805080684_5X	5/21/08	23:55	(5)	3.0084	3.0 ✓		
2805080685_5X	5/21/08	23:58	(5)	1.2471	1.2 ✓		
2805080686_5X	5/22/08	0:01	(5)	1.2788	1.3 ✓		
2805080688_5X	5/22/08	0:04	(5)	13.751	14 ✓		
2805080695_2X	5/22/08	0:07	2	2.7784	2.8 ✓		
2805080696_2X	5/22/08	0:10	2	1.0708	1.1 ✓		
2805080697_2X	5/22/08	0:13	2	0.0046	.0046 ✓		
2805080999_2X	5/22/08	0:16	2	0.0021	.0021 ✓		
2805090112_5X	5/22/08	0:19	(5)	.90842	.910 ✓		
MBLANK2007	5/22/08	0:22	1	0.0004	0.0003		
CCV	5/22/08	0:25	1	4.9881	4.99	90-110	99.7%
CCB	5/22/08	0:28	1	0.0005	0.0005		
MCV	5/22/08	0:31	1	2.5148	2.51	90-110	100%
MRL	5/22/08	0:33	1	0.0105	.0105	50-150	105%
MRL2007	5/22/08	0:36	1	0.0114	.0114		
LCS2007	5/22/08	0:39	1	.95899	.959 ✓	85-115	95.8%
LCSD2007	5/22/08	0:42	1	.96078	.961 ✓	85-115	96.0%
2805090119_2X	5/22/08	0:45	2	0.0647	.065 ✓		
2805090119_2XMS	5/22/08	0:48	2	1.0532	1.05 ✓	[ 0.989]	49.4 Q
2805090119_2XMSD	5/22/08	0:51	2	1.0129	1.01 ✓	[ 0.948]	47.4 Q
2805090119_2XT	5/22/08	0:51	2		2.00 ✓	70 - 130	95%
2805090120_2X	5/22/08	0:54	2	.18039	.180 ✓		
2805090120_2XMS	5/22/08	0:57	2	1.1300	1.13 ✓	[ 0.950]	47.4 Q
2805090120_2XMSD	5/22/08	1:00	2	1.1673	1.17 ✓	[ 0.987]	49.3 Q
2805090120_2XT	5/22/08	1:00	2		2.00	70 - 130	99% CSK
CCV	5/22/08	1:03	1	4.9408	4.94	90-110	98.8%
CCB	5/22/08	1:06	1	0.0003	0.0003		J-22
2805090115_5X	5/22/08	1:09	(5)	3.7895	3.8 ✓		
2805090116_2X	5/22/08	1:12	2	.88476	.880 ✓		



Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2805090117_5X	5/22/08	1:15	5	2.9111	2.9 ✓		
2805090118_5X	5/22/08	1:18	5	2.7389	2.7 ✓		
2805090121_5X	5/22/08	1:21	5	3.9985	4.0 ✓		
2805090122_5X	5/22/08	1:24	5	4.8871	4.9 ✓		
2805090123_5X	5/22/08	1:27	5	4.2750	4.3 ✓		
2805090124_10X	5/22/08	1:30	10	28.256	28 ✓		
2805090132_10X	5/22/08	1:33	10	25.441	25 ✓		
2805090133_10X	5/22/08	1:36	10	29.583	30 ✓		
CCV	5/22/08	1:39	1	4.8736	4.87	90-110	97.4%
CCB	5/22/08	1:42	1	0.0023	.0023 ✓		
MCV	5/22/08	1:45	1	2.4881	2.49	90-110	99.5%
2805090134_5X	5/22/08	1:47	5	2.7698	2.8 ✓		
2805090135_5X	5/22/08	1:50	5	3.9205	3.9 ✓		
2805090136_10X	5/22/08	1:53	10	19.735	20 ✓		
2805090137_10X	5/22/08	1:56	10	35.180	35 ✓		
2805090138_5X	5/22/08	1:59	5	1.0802	1.1 ✓		
2805090152_5X	5/22/08	2:02	5	0.0836	.084 ✓		
2805090157_5X	5/22/08	2:05	5	.10458	.1 ✓		
2805090177_2X	5/22/08	2:08	2	0.0032	.0032 ✓		
ECV	5/22/08	2:11	1	4.8729	4.87	90-110	97.4%
ECB	5/22/08	2:14	1	0.0008	0.0008 ✓		

Nebulizer Parameters: Hg\_ReAlign

Analyte Back Pressure Flow
All 151.0 kPa 0.55 L/min

5/21/2008 10:24:47 Hg ReAlign... Actual peak offset, (nm): 0.004
Drift (nm): -0.001 Slit adjustment: -3

Nebulizer Parameters: Hg\_ReAlign

Analyte Back Pressure Flow
All 153.0 kPa 0.54 L/min

5/21/2008 10:27:57 Hg ReAlign... Actual peak offset (nm): 0.003
Drift (nm): -0.000 Slit adjustment: -2

Align View XY Axial for analyte Mn 257.610

Table with 3 columns: X-position, Y-position, Intensity. Contains 45 rows of data points for Mn 257.610.

5/21/2008 10:38:21 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 655112.7 for Axial viewing
Y viewing position set to 15.0 mm having Peak intensity 655112.7 for Axial viewing

Align View X Radial for analyte Mn 257.610

X-position	Y-position	Intensity
-7.0	15.0	4638.3
-6.5	15.0	5483.0
-6.0	15.0	6472.4
-5.5	15.0	7203.1
-5.0	15.0	8662.0
-4.5	15.0	11757.3
-4.0	15.0	16839.0
-3.5	15.0	27103.6
-3.0	15.0	39792.9
-2.5	15.0	59090.8
-2.0	15.0	81263.7
-1.5	15.0	100606.6
-1.0	15.0	129495.6
-0.5	15.0	145229.3
0.0	15.0	175944.4
0.5	15.0	165437.1
1.0	15.0	143398.9
1.5	15.0	113605.4
2.0	15.0	88922.5
2.5	15.0	67743.2
3.0	15.0	46092.6
3.5	15.0	29105.3
4.0	15.0	21177.4
4.5	15.0	18198.0
5.0	15.0	16472.6
5.5	15.0	15198.6
6.0	15.0	13964.2
6.5	15.0	11192.5
7.0	15.0	9697.9

5/21/2008 10:40:34 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 175944.4 for Radial viewing

Cdt	35.6	0.40	1.13%	[0.00] mg/L
Cot	-44.4	4.77	10.76%	[0.00] mg/L
Crt	106.1	0.26	0.25%	[0.00] mg/L
Cut	3240.3	38.11	1.18%	[0.00] mg/L
Fet	-27.0	0.28	1.04%	[0.00] mg/L
Kt	-15.7	76.92	489.51%	[0.00] mg/L
Mgt	-267.7	1.42	0.53%	[0.00] mg/L
Mnt	337.5	6.14	1.82%	[0.00] mg/L
Mot	15.9	0.28	1.77%	[0.00] mg/L
Naf	-148.5	0.42	0.28%	[0.00] mg/L
Nit	-52.4	6.66	12.73%	[0.00] mg/L
Pbt	-16.6	1.33	8.02%	[0.00] mg/L
Sbt	7.0	0.98	13.92%	[0.00] mg/L
Set	2.0	1.12	55.53%	[0.00] mg/L
Tlr	-20.1	0.85	4.22%	[0.00] mg/L
Vt	130.5	9.13	6.99%	[0.00] mg/L
Znt	147.7	2.75	1.87%	[0.00] mg/L

=====  
**Analysis Begun**

Start Time: 5/21/2008 13:22:34                      Plasma On Time: 5/21/2008 10:24:22  
 Logged In Analyst: Charley Kay                      Technique: ICP Continuous  
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
 Batch ID: 080521  
 Results Data Set: 080521  
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
**Method Loaded**

Method Name: 200.7&6010\_080304

Method Last Saved: 5/2/2008 09:27:36

IEC File: IEC080304.iec

MSF File:

Method Description: 200.7/6010\_080304

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B_	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Mo	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a

=====  
 Sequence No.: 1  
 Sample ID: Calib Blank 1  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/21/2008 13:22:51  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

=====  
**Nebulizer Parameters: Calib Blank 1**

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

=====  
**Mean Data: Calib Blank 1**

Analyte	Mean Corrected Intensity	Std.Dev	RSD	Calib Conc.	Units
Sca	404742.0	3110.56	0.77%	100	%
Yr	323061.0	1244.48	0.39%	100	%
Ag†	235.4	1.16	0.49%	[0.00]	mg/L
Al†	88.0	68.58	77.94%	[0.00]	mg/L
As†	2.9	2.41	84.24%	[0.00]	mg/L
B_†	179.7	1.62	0.90%	[0.00]	mg/L
Ba†	-22.6	2.01	8.87%	[0.00]	mg/L
Be†	-3677.2	27.04	0.74%	[0.00]	mg/L
Ca†	2045.8	19.00	0.93%	[0.00]	mg/L

User canceled analysis.

=====  
Analysis Begun

Start Time: 5/21/2008 13:26:01                      Plasma On Time: 5/21/2008 10:24:22  
Logged In Analyst: Charley Kay                      Technique: ICP Continuous  
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
Batch ID: 080521  
Results Data Set: 080521  
Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 2    Autosampler Location: 15  
Sample ID: Standard 2                                      Date Collected: 5/21/2008 13:26:02  
Analyst:    Data Type: Original  
Initial Sample Wt:    Initial Sample Vol:  
Dilution:    Sample Prep Vol:

-----  
Nebulizer Parameters: Standard 2  
Analyte                      Back Pressure              Flow  
All                              213.0 kPa                      0.65 L/min

-----  
Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	361802.3	405.03	0.11%	89.4 %
Yr	312337.6	2027.85	0.65%	96.7 %
Agf	486827.1	591.45	0.12%	[2] mg/L
Alf	69574.3	10.95	0.02%	[10] mg/L
Ast	15399.7	230.12	1.49%	[10] mg/L
B_f	129037.4	190.22	0.15%	[5.02] mg/L
Baf	390340.1	766.48	0.20%	[10] mg/L
Bet	9919934.5	65580.02	0.66%	[4.01] mg/L
Caf	1604780.8	11420.38	0.71%	[100] mg/L
Cdf	104724.5	158.83	0.15%	[5.01] mg/L
Cof	142491.5	89.67	0.06%	[10] mg/L
Crt	409531.9	1027.09	0.25%	[9.97] mg/L
Cuf	2976990.1	4196.62	0.14%	[10] mg/L
Fef	34668.3	212.06	0.61%	[9.98] mg/L
Kf	101208.5	286.78	0.28%	[100] mg/L
Mgf	994485.5	7891.06	0.79%	[100] mg/L
Mnf	4695790.6	1839.99	0.04%	[10] mg/L
Mof	96679.2	215.34	0.22%	[9.98] mg/L
Naf	265148.9	628.35	0.24%	[100] mg/L
Nif	181525.8	435.76	0.24%	[10] mg/L
Pbf	36984.2	409.01	1.11%	[10] mg/L
Sbf	15482.8	301.03	1.94%	[10] mg/L
Sef	9413.1	124.80	1.33%	[10] mg/L
Tlf	21205.0	249.41	1.18%	[10] mg/L
Vf	1457745.8	1431.46	0.10%	[10] mg/L
Znf	372517.7	290.32	0.08%	[10] mg/L

-----  
Calibration Summary

Analyte	Stds	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	243400	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	6957	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	1540	0.00000	1.000000	
B_	1	Lin, Calc Int	-0.0	25700	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	39030	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	2474000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	16050	0.00000	1.000000	
Cd	1	Lin, Calc Int	0.0	20900	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	14250	0.00000	1.000000	
Cr	1	Lin, Calc Int	0.0	41080	0.00000	1.000000	
Cu	1	Lin, Calc Int	0.0	297700	0.00000	1.000000	

Fe	1	Lin, Calc Int	0.0	3474	0.00000	1.000000
K	1	Lin, Calc Int	0.0	1012	0.00000	1.000000
Mg	1	Lin, Calc Int	0.0	9945	0.00000	1.000000
Mn	1	Lin, Calc Int	0.0	469600	0.00000	1.000000
Mo	1	Lin, Calc Int	0.0	9687	0.00000	1.000000
Na	1	Lin, Calc Int	0.0	2651	0.00000	1.000000
Ni	1	Lin, Calc Int	0.0	18150	0.00000	1.000000
Pb	1	Lin, Calc Int	0.0	3698	0.00000	1.000000
Sb	1	Lin, Calc Int	0.0	1548	0.00000	1.000000
Se	1	Lin, Calc Int	0.0	941.3	0.00000	1.000000
Tl	1	Lin, Calc Int	0.0	2120	0.00000	1.000000
V	1	Lin, Calc Int	0.0	145800	0.00000	1.000000
Zn	1	Lin, Calc Int	0.0	37250	0.00000	1.000000

=====  
Analytical Sequence

Method: 200.7&6010\_080304

Seq.	Loc.	ID	Status
1	0	Calib Blank 1	Applied
2	15	Standard 2	Applied



=====  
 Analysis Bequn

Start Time: 5/21/2008 13:32:27                      Plasma On Time: 5/21/2008 10:24:22  
 Logged In Analyst: Charley Kay                      Technique: ICP Continuous  
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
 Batch ID: 080521  
 Results Data Set: 080521  
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
 Sequence No.: 1                                      Autosampler Location: 15  
 Sample ID: ICV                                      Date Collected: 5/21/2008 13:32:29  
 Analyst:    Data Type: Original  
 Initial Sample Wt:                                  Initial Sample Vol:  
 Dilution:    Sample Prep Vol:

-----  
 Nebulizer Parameters: ICV  
 Analyte                      Back Pressure              Flow  
 All                              213.0 kPa                      0.65 L/min

-----  
 Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	362587.4	89.6 %		0.04			0.05%
Yr	315456.9	97.6 %		1.14			1.17%
Ag†	485642.8	2.00 mg/L		0.004	2.00 mg/L	0.004	0.19%
	QC value within limits for Ag		Recovery = 99.94%				
Al†	69860.3	9.68 mg/L		0.016	9.68 mg/L	0.016	0.16%
	QC value within limits for Al		Recovery = 96.78%				
As†	15306.2	9.94 mg/L		0.017	9.94 mg/L	0.017	0.17%
	QC value within limits for As		Recovery = 99.39%				
B_†	129039.7	4.93 mg/L		0.018	4.93 mg/L	0.018	0.36%
	QC value within limits for B_		Recovery = 98.65%				
Ba†	390661.0	10.0 mg/L		0.01	10.0 mg/L	0.01	0.12%
	QC value within limits for Ba		Recovery = 100.08%				
Be†	9972093.4	4.03 mg/L		0.003	4.03 mg/L	0.003	0.07%
	QC value within limits for Be		Recovery = 100.81%				
Ca†	1596667.1	99.5 mg/L		1.83	99.5 mg/L	1.83	1.84%
	QC value within limits for Ca		Recovery = 99.49%				
Cd†	104725.9	4.90 mg/L		0.007	4.90 mg/L	0.007	0.14%
	QC value within limits for Cd		Recovery = 98.06%				
Co†	141774.9	9.95 mg/L		0.046	9.95 mg/L	0.046	0.46%
	QC value within limits for Co		Recovery = 99.50%				
Cr†	408544.3	9.95 mg/L		0.032	9.95 mg/L	0.032	0.32%
	QC value within limits for Cr		Recovery = 99.46%				
Cu†	2977285.9	10.0 mg/L		0.01	10.0 mg/L	0.01	0.07%
	QC value within limits for Cu		Recovery = 100.10%				
Fe†	35176.0	10.1 mg/L		0.03	10.1 mg/L	0.03	0.26%
	QC value within limits for Fe		Recovery = 101.26%				
K†	102412.2	101 mg/L		0.3	101 mg/L	0.3	0.33%
	QC value within limits for K		Recovery = 101.19%				
Mg†	997974.9	100 mg/L		1.7	100 mg/L	1.7	1.74%
	QC value within limits for Mg		Recovery = 100.37%				
Mn†	4690644.2	9.99 mg/L		0.008	9.99 mg/L	0.008	0.08%
	QC value within limits for Mn		Recovery = 99.87%				
Mo†	96457.1	9.96 mg/L		0.017	9.96 mg/L	0.017	0.17%
	QC value within limits for Mo		Recovery = 99.57%				
Na†	267764.5	101 mg/L		0.0	101 mg/L	0.0	0.03%
	QC value within limits for Na		Recovery = 100.99%				
Ni†	181333.4	9.99 mg/L		0.002	9.99 mg/L	0.002	0.02%
	QC value within limits for Ni		Recovery = 99.89%				
Pb†	36821.4	9.96 mg/L		0.003	9.96 mg/L	0.003	0.03%
	QC value within limits for Pb		Recovery = 99.56%				
Sb†	15402.3	9.99 mg/L		0.006	9.99 mg/L	0.006	0.06%
	QC value within limits for Sb		Recovery = 99.88%				
Se†	9459.5	10.1 mg/L		0.08	10.1 mg/L	0.08	0.75%

QC value within limits for Se	Recovery = 100.74%						
Tlt 21087.4	10.0 mg/L	0.02	10.0 mg/L	0.02	0.24%		
QC value within limits for Tl	Recovery = 100.01%						
Vt 1455765.5	10.0 mg/L	0.01	10.0 mg/L	0.01	0.13%		
QC value within limits for V	Recovery = 100.39%						
Znt 372213.7	9.91 mg/L	0.015	9.91 mg/L	0.015	0.15%		
QC value within limits for Zn	Recovery = 99.06%						
All analyte(s) passed QC.							

Sequence No.: 2  
 Sample ID: LINEARITY  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 9  
 Date Collected: 5/21/2008 13:35:45  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

## Nebulizer Parameters: LINEARITY

Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

## Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	347772.8	85.9 %		0.54			0.63%
Yr	307541.4	95.2 %		0.61			0.64%
Agf	-5400.6	0.0158 mg/L		0.00085	0.0158 mg/L	0.00085	5.37%
	QC value within limits for Ag	Recovery = Not calculated					
Alf	-40.3	-0.00599 mg/L		0.005506	-0.00599 mg/L	0.005506	91.98%
	QC value within limits for Al	Recovery = Not calculated					
Asf	-176.1	-0.114 mg/L		0.0003	-0.114 mg/L	0.0003	0.30%
	QC value within limits for As	Recovery = Not calculated					
B_f	471.1	0.0184 mg/L		0.00092	0.0184 mg/L	0.00092	5.01%
	QC value within limits for B_	Recovery = Not calculated					
Baf	63.0	0.00161 mg/L		0.000003	0.00161 mg/L	0.000003	0.21%
	QC value within limits for Ba	Recovery = Not calculated					
Bef	-694.6	-0.00028 mg/L		0.000025	-0.00028 mg/L	0.000025	9.09%
	QC value within limits for Be	Recovery = Not calculated					
Caf	4777890.4	298 mg/L		0.4	298 mg/L	0.4	0.14%
	QC value within limits for Ca	Recovery = 99.24%					
Cdf	-9.8	0.00112 mg/L		0.000001	0.00112 mg/L	0.000001	0.13%
	QC value within limits for Cd	Recovery = Not calculated					
Cof	35.4	0.00249 mg/L		0.000032	0.00249 mg/L	0.000032	1.28%
	QC value within limits for Co	Recovery = Not calculated					
Crf	131.6	0.00320 mg/L		0.000053	0.00320 mg/L	0.000053	1.64%
	QC value within limits for Cr	Recovery = Not calculated					
Cuf	-2760.9	-0.00927 mg/L		0.000358	-0.00927 mg/L	0.000358	3.86%
	QC value within limits for Cu	Recovery = Not calculated					
Fef	358678.9	103 mg/L		0.7	103 mg/L	0.7	0.67%
	QC value within limits for Fe	Recovery = 103.25%					
Kf	335556.4	332 mg/L		0.2	332 mg/L	0.2	0.06%
	QC value greater than the upper limit for K	Recovery = 110.52%					
Mgf	1890292.0	190 mg/L		0.9	190 mg/L	0.9	0.48%
	QC value within limits for Mg	Recovery = Not calculated					
Mnf	-1061.5	-0.00653 mg/L		0.000023	-0.00653 mg/L	0.000023	0.35%
	QC value within limits for Mn	Recovery = Not calculated					
Mof	51.1	0.00527 mg/L		0.000349	0.00527 mg/L	0.000349	6.62%
	QC value within limits for Mo	Recovery = Not calculated					
Naf	841880.5	318 mg/L		0.9	318 mg/L	0.9	0.29%
	QC value within limits for Na	Recovery = 105.84%					
Nif	10.7	0.00059 mg/L		0.000424	0.00059 mg/L	0.000424	72.14%
	QC value within limits for Ni	Recovery = Not calculated					
Pbf	-4.5	-0.00123 mg/L		0.000529	-0.00123 mg/L	0.000529	43.04%
	QC value within limits for Pb	Recovery = Not calculated					
Sbf	19.7	0.0127 mg/L		0.00179	0.0127 mg/L	0.00179	14.06%
	QC value within limits for Sb	Recovery = Not calculated					
Sef	-91.9	0.158 mg/L		0.0012	0.158 mg/L	0.0012	0.79%
	QC value within limits for Se	Recovery = Not calculated					
Tlf	23.6	0.0111 mg/L		0.00132	0.0111 mg/L	0.00132	11.96%
	QC value within limits for Tl	Recovery = Not calculated					
Vf	-252.5	0.00233 mg/L		0.000149	0.00233 mg/L	0.000149	6.37%
	QC value within limits for V	Recovery = Not calculated					
Znf	809.0	0.0217 mg/L		0.00019	0.0217 mg/L	0.00019	0.89%
	QC value within limits for Zn	Recovery = Not calculated					

QC Failed. Continue with analysis.

Sequence No.: 3  
 Sample ID: ICSA  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 10  
 Date Collected: 5/21/2008 13:39:28  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	358424.6	88.6	%	0.02			0.02%
Yr	317955.6	98.4	%	0.73			0.75%
Ag†	-5296.3	0.0154	mg/L	0.00078	0.0154 mg/L	0.00078	5.04%
	QC value within limits for Ag Recovery = Not calculated						
Al†	1718089.2	247	mg/L	3.0	247 mg/L	3.0	1.20%
	QC value within limits for Al Recovery = 98.78%						
As†	-355.1	-0.231	mg/L	0.0031	-0.231 mg/L	0.0031	1.35%
	QC value within limits for As Recovery = Not calculated						
B_†	-101.8	-0.00388	mg/L	0.000419	-0.00388 mg/L	0.000419	10.79%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	75.1	0.00192	mg/L	0.000076	0.00192 mg/L	0.000076	3.95%
	QC value within limits for Ba Recovery = Not calculated						
Be†	-956.0	-0.00039	mg/L	0.000013	-0.00039 mg/L	0.000013	3.32%
	QC value within limits for Be Recovery = Not calculated						
Ca†	3973417.1	248	mg/L	4.3	248 mg/L	4.3	1.75%
	QC value within limits for Ca Recovery = 99.04%						
Cd†	-31.5	0.00169	mg/L	0.000180	0.00169 mg/L	0.000180	10.66%
	QC value within limits for Cd Recovery = Not calculated						
Co†	14.8	0.00104	mg/L	0.000114	0.00104 mg/L	0.000114	10.99%
	QC value within limits for Co Recovery = Not calculated						
Cr†	22.6	0.00055	mg/L	0.000097	0.00055 mg/L	0.000097	17.68%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-3117.3	-0.0105	mg/L	0.00004	-0.0105 mg/L	0.00004	0.35%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	351167.4	101	mg/L	1.6	101 mg/L	1.6	1.58%
	QC value within limits for Fe Recovery = 101.09%						
K†	260.7	0.258	mg/L	0.0194	0.258 mg/L	0.0194	7.53%
	QC value within limits for K Recovery = Not calculated						
Mg†	2325901.4	234	mg/L	4.0	234 mg/L	4.0	1.72%
	QC value within limits for Mg Recovery = 93.62%						
Mn†	-1360.5	-0.00815	mg/L	0.000019	-0.00815 mg/L	0.000019	0.23%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	4.8	0.00049	mg/L	0.000386	0.00049 mg/L	0.000386	78.62%
	QC value within limits for Mo Recovery = Not calculated						
Na†	421.7	0.159	mg/L	0.0061	0.159 mg/L	0.0061	3.85%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-9.5	-0.00052	mg/L	0.000485	-0.00052 mg/L	0.000485	92.97%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	-143.4	-0.0388	mg/L	0.00036	-0.0388 mg/L	0.00036	0.93%
	QC value within limits for Pb Recovery = Not calculated						
Sb†	14.5	0.00935	mg/L	0.006247	0.00935 mg/L	0.006247	66.80%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-116.1	0.127	mg/L	0.0139	0.127 mg/L	0.0139	10.95%
	QC value within limits for Se Recovery = Not calculated						
Tl†	15.5	0.00728	mg/L	0.001572	0.00728 mg/L	0.001572	21.61%
	QC value within limits for Tl Recovery = Not calculated						
V†	-388.0	0.00131	mg/L	0.000476	0.00131 mg/L	0.000476	36.47%
	QC value within limits for V Recovery = Not calculated						
Zn†	623.1	0.0167	mg/L	0.00001	0.0167 mg/L	0.00001	0.05%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

User canceled analysis.

Analysis Begun

Start Time: 5/21/2008 13:42:45

Plasma On Time: 5/21/2008 10:24:22

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif

Batch ID: 080521

Results Data Set: 080521

Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 2

Autosampler Location: 9

Sample ID: LINEARITY

Date Collected: 5/21/2008 13:42:46

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	348437.1	86.1 %	0.09			0.11%
Yr	306929.2	95.0 %	0.78			0.82%
Ag†	-5392.8	0.0147 mg/L	0.00062	0.0147 mg/L	0.00062	4.20%
	QC value within limits for Ag	Recovery = Not calculated				
Al†	-17.6	-0.00250 mg/L	0.011554	-0.00250 mg/L	0.011554	461.72%
	QC value within limits for Al	Recovery = Not calculated				
As†	-185.4	-0.120 mg/L	0.0027	-0.120 mg/L	0.0027	2.26%
	QC value within limits for As	Recovery = Not calculated				
B_†	-111.9	-0.00428 mg/L	0.000452	-0.00428 mg/L	0.000452	10.56%
	QC value within limits for B_	Recovery = Not calculated				
Ba†	41.3	0.00106 mg/L	0.000040	0.00106 mg/L	0.000040	3.78%
	QC value within limits for Ba	Recovery = Not calculated				
Be†	-1190.0	-0.00048 mg/L	0.000027	-0.00048 mg/L	0.000027	5.53%
	QC value within limits for Be	Recovery = Not calculated				
Ca†	4648811.1	290 mg/L	7.3	290 mg/L	7.3	2.51%
	QC value within limits for Ca	Recovery = 96.56%				
Cd†	-28.5	0.00031 mg/L	0.000224	0.00031 mg/L	0.000224	72.63%
	QC value within limits for Cd	Recovery = Not calculated				
Co†	26.2	0.00184 mg/L	0.000169	0.00184 mg/L	0.000169	9.19%
	QC value within limits for Co	Recovery = Not calculated				
Cr†	102.4	0.00249 mg/L	0.000146	0.00249 mg/L	0.000146	5.85%
	QC value within limits for Cr	Recovery = Not calculated				
Cu†	-2888.1	-0.00970 mg/L	0.000057	-0.00970 mg/L	0.000057	0.58%
	QC value within limits for Cu	Recovery = Not calculated				
Fe†	348384.1	100 mg/L	2.7	100 mg/L	2.7	2.66%
	QC value within limits for Fe	Recovery = 100.29%				
K†	324186.6	320 mg/L	8.5	320 mg/L	8.5	2.66%
	QC value within limits for K	Recovery = 106.77%				
Mg†	1844638.6	186 mg/L	4.3	186 mg/L	4.3	2.33%
	QC value within limits for Mg	Recovery = Not calculated				
Mn†	-1326.0	-0.00699 mg/L	0.000176	-0.00699 mg/L	0.000176	2.52%
	QC value within limits for Mn	Recovery = Not calculated				
Mo†	-6.7	-0.00069 mg/L	0.000082	-0.00069 mg/L	0.000082	11.84%
	QC value within limits for Mo	Recovery = Not calculated				
Na†	816823.7	308 mg/L	6.5	308 mg/L	6.5	2.10%
	QC value within limits for Na	Recovery = 102.69%				
Ni†	5.4	0.00030 mg/L	0.000779	0.00030 mg/L	0.000779	261.47%
	QC value within limits for Ni	Recovery = Not calculated				
Pb†	-4.6	-0.00123 mg/L	0.001837	-0.00123 mg/L	0.001837	149.12%
	QC value within limits for Pb	Recovery = Not calculated				
Sb†	18.1	0.0117 mg/L	0.00089	0.0117 mg/L	0.00089	7.67%
	QC value within limits for Sb	Recovery = Not calculated				

Se†	-110.4	0.131 mg/L	0.0059	0.131 mg/L	0.0059	4.49%
QC value within limits for Se Recovery = Not calculated						
Tl†	23.7	0.0112 mg/L	0.00007	0.0112 mg/L	0.00007	0.66%
QC value within limits for Tl Recovery = Not calculated						
V†	-312.9	0.00180 mg/L	0.000662	0.00180 mg/L	0.000662	36.83%
QC value within limits for V Recovery = Not calculated						
Zn†	748.8	0.0201 mg/L	0.00021	0.0201 mg/L	0.00021	1.04%
QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

User canceled analysis.

=====  
Analysis Begun

Start Time: 5/21/2008 13:45:12 Plasma On Time: 5/21/2008 10:24:22  
Logged In Analyst: Charley Kay Technique: ICP Continuous  
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
Batch ID: 080521  
Results Data Set: 080521  
Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 4 Autosampler Location: 11  
Sample ID: ICSAB Date Collected: 5/21/2008 13:45:13  
Analyst: Data Type: Original  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:

-----  
Nebulizer Parameters: ICSAB

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

-----  
Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Conc. Units	Std.Dev.	RSD
Sca	357376.1	88.3 %		0.61			0.70%
Yr	317999.3	98.4 %		0.36			0.36%
Agf	19948.3	0.120 mg/L		0.0004	0.120 mg/L	0.0004	0.34%
Alf	1749676.1	251 mg/L		0.6	251 mg/L	0.6	0.22%
Ast	-360.9	-0.234 mg/L		0.0020	-0.234 mg/L	0.0020	0.85%
B_tf	-232.0	-0.0112 mg/L		0.00107	-0.0112 mg/L	0.00107	9.58%
Baf	10384.5	0.266 mg/L		0.0016	0.266 mg/L	0.0016	0.58%
Bef	618416.3	0.250 mg/L		0.0006	0.250 mg/L	0.0006	0.26%
Caf	4001503.7	249 mg/L		0.2	249 mg/L	0.2	0.08%
Cdf	10543.5	0.508 mg/L		0.0039	0.508 mg/L	0.0039	0.77%
Cof	3399.0	0.239 mg/L		0.0027	0.239 mg/L	0.0027	1.12%
Crf	10344.3	0.252 mg/L		0.0016	0.252 mg/L	0.0016	0.64%
Cuf	73608.6	0.247 mg/L		0.0003	0.247 mg/L	0.0003	0.11%
Fef	355603.7	102 mg/L		0.5	102 mg/L	0.5	0.47%
Kf	396.4	0.392 mg/L		0.0218	0.392 mg/L	0.0218	5.56%
Mgf	2336643.2	235 mg/L		0.2	235 mg/L	0.2	0.06%
Mnf	119085.9	0.248 mg/L		0.0009	0.248 mg/L	0.0009	0.38%
Mof	-7.7	-0.00080 mg/L		0.000178	-0.00080 mg/L	0.000178	22.36%
Naf	817.9	0.308 mg/L		0.0105	0.308 mg/L	0.0105	3.39%
Nif	8562.5	0.472 mg/L		0.0028	0.472 mg/L	0.0028	0.60%
Pbf	1696.8	0.459 mg/L		0.0016	0.459 mg/L	0.0016	0.35%
Sbf	10.9	0.00802 mg/L		0.001379	0.00802 mg/L	0.001379	17.19%

Se†	-120.1	0.126 mg/L	0.0098	0.126 mg/L	0.0098	7.76%
QC value greater than the upper limit for Se Recovery = Not calculated						
Tl†	10.9	0.00734 mg/L	0.006489	0.00734 mg/L	0.006489	88.37%
QC value within limits for Tl Recovery = Not calculated						
V†	35928.8	0.252 mg/L	0.0003	0.252 mg/L	0.0003	0.12%
QC value within limits for V Recovery = 100.68%						
Zn†	20275.8	0.541 mg/L	0.0011	0.541 mg/L	0.0011	0.20%
QC value within limits for Zn Recovery = 108.13%						
QC Failed. Continue with analysis.						



Sequence No.: 5  
 Sample ID: Wash  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/21/2008 13:48:58  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: Wash  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: Wash

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	404664.2	100.0 %	%	0.71			0.71%
Yr	338245.1	105 %	%	0.0			0.02%
Agf	13.0	0.00005 mg/L	mg/L	0.000083	0.00005 mg/L	0.000083	154.09%
QC value within limits for Ag		Recovery = Not calculated					
Alt	-24.4	-0.00350 mg/L	mg/L	0.004636	-0.00350 mg/L	0.004636	132.41%
QC value within limits for Al		Recovery = Not calculated					
Ast	-0.6	-0.00038 mg/L	mg/L	0.001467	-0.00038 mg/L	0.001467	387.12%
QC value within limits for As		Recovery = Not calculated					
B_t	47.0	0.00183 mg/L	mg/L	0.000061	0.00183 mg/L	0.000061	3.33%
QC value within limits for B_		Recovery = Not calculated					
Bat	3.6	0.00009 mg/L	mg/L	0.000054	0.00009 mg/L	0.000054	58.19%
QC value within limits for Ba		Recovery = Not calculated					
Bet	125.9	0.00005 mg/L	mg/L	0.000019	0.00005 mg/L	0.000019	37.52%
QC value within limits for Be		Recovery = Not calculated					
Cat	61.4	0.00382 mg/L	mg/L	0.000379	0.00382 mg/L	0.000379	9.92%
QC value within limits for Ca		Recovery = Not calculated					
Cdt	3.9	0.00019 mg/L	mg/L	0.000016	0.00019 mg/L	0.000016	8.32%
QC value within limits for Cd		Recovery = Not calculated					
Cot	2.9	0.00020 mg/L	mg/L	0.000271	0.00020 mg/L	0.000271	132.60%
QC value within limits for Co		Recovery = Not calculated					
Crt	0.5	0.00001 mg/L	mg/L	0.000072	0.00001 mg/L	0.000072	656.21%
QC value within limits for Cr		Recovery = Not calculated					
Cut	14.3	0.00005 mg/L	mg/L	0.000013	0.00005 mg/L	0.000013	27.65%
QC value within limits for Cu		Recovery = Not calculated					
Fet	8.6	0.00248 mg/L	mg/L	0.001136	0.00248 mg/L	0.001136	45.86%
QC value within limits for Fe		Recovery = Not calculated					
Kf	89.6	0.0886 mg/L	mg/L	0.05562	0.0886 mg/L	0.05562	62.81%
QC value within limits for K		Recovery = Not calculated					
Mgf	11.3	0.00114 mg/L	mg/L	0.000240	0.00114 mg/L	0.000240	21.03%
QC value within limits for Mg		Recovery = Not calculated					
Mnf	-81.2	-0.00017 mg/L	mg/L	0.000005	-0.00017 mg/L	0.000005	3.14%
QC value within limits for Mn		Recovery = Not calculated					
Mot	-0.6	-0.00006 mg/L	mg/L	0.000019	-0.00006 mg/L	0.000019	29.23%
QC value within limits for Mo		Recovery = Not calculated					
Nat	91.3	0.0344 mg/L	mg/L	0.00536	0.0344 mg/L	0.00536	15.56%
QC value within limits for Na		Recovery = Not calculated					
Nit	1.0	0.00006 mg/L	mg/L	0.000031	0.00006 mg/L	0.000031	54.00%
QC value within limits for Ni		Recovery = Not calculated					
Pbf	8.3	0.00223 mg/L	mg/L	0.000339	0.00223 mg/L	0.000339	15.19%
QC value within limits for Pb		Recovery = Not calculated					
Sbf	0.8	0.00050 mg/L	mg/L	0.001182	0.00050 mg/L	0.001182	234.05%
QC value within limits for Sb		Recovery = Not calculated					
Set	-7.1	-0.00753 mg/L	mg/L	0.002737	-0.00753 mg/L	0.002737	36.33%
QC value within limits for Se		Recovery = Not calculated					
Tlf	1.6	0.00073 mg/L	mg/L	0.001551	0.00073 mg/L	0.001551	211.41%
QC value within limits for Tl		Recovery = Not calculated					
Vt	55.2	0.00038 mg/L	mg/L	0.000061	0.00038 mg/L	0.000061	16.07%
QC value within limits for V		Recovery = Not calculated					
Znf	11.0	0.00029 mg/L	mg/L	0.000070	0.00029 mg/L	0.000070	23.67%
QC value within limits for Zn		Recovery = Not calculated					

All analyte(s) passed QC.



QC value within limits for Cu	Recovery = 101.38%				
Fe†	3685.9	1.06 mg/L	0.021	1.06 mg/L	0.021 1.94%
QC value within limits for Fe	Recovery = 106.11%				
K†	9594.2	9.48 mg/L	0.098	9.48 mg/L	0.098 1.03%
QC value within limits for K	Recovery = 94.80%				
Mg†	10601.4	1.07 mg/L	0.012	1.07 mg/L	0.012 1.08%
QC value within limits for Mg	Recovery = 106.79%				
Mn†	489840.3	1.04 mg/L	0.002	1.04 mg/L	0.002 0.21%
QC value within limits for Mn	Recovery = 104.31%				
Mo†	9249.9	0.955 mg/L	0.0055	0.955 mg/L	0.0055 0.58%
QC value within limits for Mo	Recovery = 95.49%				
Na†	2566.5	0.968 mg/L	0.0153	0.968 mg/L	0.0153 1.58%
QC value within limits for Na	Recovery = 96.80%				
Ni†	19310.4	1.06 mg/L	0.001	1.06 mg/L	0.001 0.11%
QC value within limits for Ni	Recovery = 106.38%				
Pb†	3859.7	1.04 mg/L	0.011	1.04 mg/L	0.011 1.06%
QC value within limits for Pb	Recovery = 104.36%				
Sb†	1440.4	0.934 mg/L	0.0032	0.934 mg/L	0.0032 0.34%
QC value within limits for Sb	Recovery = 93.40%				
Se†	887.3	0.945 mg/L	0.0120	0.945 mg/L	0.0120 1.27%
QC value within limits for Se	Recovery = 94.53%				
Tl†	2278.8	1.08 mg/L	0.011	1.08 mg/L	0.011 0.98%
QC value within limits for Tl	Recovery = 108.05%				
V†	138313.2	0.954 mg/L	0.0000	0.954 mg/L	0.0000 0.00%
QC value within limits for V	Recovery = 95.40%				
Zn†	38361.3	1.02 mg/L	0.001	1.02 mg/L	0.001 0.10%
QC value within limits for Zn	Recovery = 102.07%				

All analyte(s) passed QC.

Sequence No.: 7  
 Sample ID: CCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/21/2008 13:59:53  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

## Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

## Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	380635.8	94.0 %		1.43			1.52%
Yr	341197.5	106 %		0.8			0.71%
Agf	238099.1	0.980 mg/L	Recovery = 98.01%	0.0007	0.980 mg/L	0.0007	0.07%
QC value within limits for Ag							
Alf	35168.4	4.88 mg/L	Recovery = 97.51%	0.028	4.88 mg/L	0.028	0.57%
QC value within limits for Al							
Asf	7328.3	4.76 mg/L	Recovery = 95.17%	0.045	4.76 mg/L	0.045	0.95%
QC value within limits for As							
B_l	63626.2	2.43 mg/L	Recovery = 97.22%	0.003	2.43 mg/L	0.003	0.13%
QC value within limits for B_							
Baf	194481.8	4.98 mg/L	Recovery = 99.65%	0.009	4.98 mg/L	0.009	0.19%
QC value within limits for Ba							
Be_f	4983695.5	2.02 mg/L	Recovery = 100.76%	0.007	2.02 mg/L	0.007	0.35%
QC value within limits for Be							
Ca_f	790044.3	49.2 mg/L	Recovery = 98.46%	0.56	49.2 mg/L	0.56	1.14%
QC value within limits for Ca							
Cd_f	41553.6	1.94 mg/L	Recovery = 96.87%	0.009	1.94 mg/L	0.009	0.46%
QC value within limits for Cd							
Co_f	70663.9	4.96 mg/L	Recovery = 99.18%	0.008	4.96 mg/L	0.008	0.15%
QC value within limits for Co							
Cr_f	201856.3	4.91 mg/L	Recovery = 98.28%	0.030	4.91 mg/L	0.030	0.61%
QC value within limits for Cr							
Cu_f	1524795.9	5.13 mg/L	Recovery = 102.53%	0.018	5.13 mg/L	0.018	0.36%
QC value within limits for Cu							
Fe_f	18477.1	5.32 mg/L	Recovery = 106.38%	0.008	5.32 mg/L	0.008	0.16%
QC value within limits for Fe							
K_f	50942.8	50.3 mg/L	Recovery = 100.67%	0.22	50.3 mg/L	0.22	0.44%
QC value within limits for K							
Mg_f	506620.9	51.0 mg/L	Recovery = 101.90%	0.43	51.0 mg/L	0.43	0.85%
QC value within limits for Mg							
Mn_f	2341350.7	4.98 mg/L	Recovery = 99.70%	0.010	4.98 mg/L	0.010	0.20%
QC value within limits for Mn							
Mo_f	47683.1	4.92 mg/L	Recovery = 98.44%	0.006	4.92 mg/L	0.006	0.12%
QC value within limits for Mo							
Na_f	134954.4	50.9 mg/L	Recovery = 101.80%	0.40	50.9 mg/L	0.40	0.78%
QC value within limits for Na							
Ni_f	91282.2	5.03 mg/L	Recovery = 100.57%	0.004	5.03 mg/L	0.004	0.08%
QC value within limits for Ni							
Pb_f	18495.3	5.00 mg/L	Recovery = 100.02%	0.047	5.00 mg/L	0.047	0.95%
QC value within limits for Pb							
Sb_f	7507.8	4.87 mg/L	Recovery = 97.38%	0.055	4.87 mg/L	0.055	1.14%
QC value within limits for Sb							
Se_f	4687.7	4.99 mg/L	Recovery = 99.86%	0.074	4.99 mg/L	0.074	1.48%
QC value within limits for Se							
Tl_f	11009.1	5.22 mg/L	Recovery = 104.40%	0.044	5.22 mg/L	0.044	0.84%
QC value within limits for Tl							
V_f	716133.8	4.94 mg/L	Recovery = 98.77%	0.010	4.94 mg/L	0.010	0.21%
QC value within limits for V							
Zn_f	189102.4	5.03 mg/L	Recovery = 100.66%	0.005	5.03 mg/L	0.005	0.10%
QC value within limits for Zn							

All analyte(s) passed QC.

User canceled analysis.

=====  
Analysis Begun

Start Time: 5/21/2008 14:03:40

Plasma On Time: 5/21/2008 10:24:22

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif

Batch ID: 080521

Results Data Set: 080521

Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 34

Autosampler Location: 5

Sample ID: MCV

Date Collected: 5/21/2008 14:03:41

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

-----  
Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

-----  
Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	388942.0	96.1	%	0.01			
Yr	331990.9	103	%	0.7			0.01%
Ag†	116274.4	0.479	mg/L	0.0008	0.479 mg/L	0.0008	0.71%
	QC value within limits for Ag	Recovery =	95.73%				0.17%
Al†	17412.4	2.41	mg/L	0.015	2.41 mg/L	0.015	0.60%
	QC value within limits for Al	Recovery =	96.48%				
As†	3640.6	2.36	mg/L	0.004	2.36 mg/L	0.004	0.17%
	QC value within limits for As	Recovery =	94.56%				
B_†	31868.9	1.22	mg/L	0.006	1.22 mg/L	0.006	0.49%
	QC value within limits for B_	Recovery =	97.44%				
Ba†	100086.4	2.56	mg/L	0.006	2.56 mg/L	0.006	0.23%
	QC value within limits for Ba	Recovery =	102.56%				
Be†	2515251.4	1.02	mg/L	0.011	1.02 mg/L	0.011	1.12%
	QC value within limits for Be	Recovery =	101.71%				
Ca†	406337.0	25.3	mg/L	0.23	25.3 mg/L	0.23	0.92%
	QC value within limits for Ca	Recovery =	101.28%				
Cd†	20592.8	0.960	mg/L	0.0010	0.960 mg/L	0.0010	0.10%
	QC value within limits for Cd	Recovery =	96.04%				
Co†	36568.7	2.57	mg/L	0.007	2.57 mg/L	0.007	0.25%
	QC value within limits for Co	Recovery =	102.66%				
Cr†	103469.9	2.52	mg/L	0.002	2.52 mg/L	0.002	0.09%
	QC value within limits for Cr	Recovery =	100.76%				
Cu†	731531.2	2.46	mg/L	0.017	2.46 mg/L	0.017	0.70%
	QC value within limits for Cu	Recovery =	98.39%				
Fe†	9132.4	2.63	mg/L	0.020	2.63 mg/L	0.020	0.75%
	QC value within limits for Fe	Recovery =	105.16%				
K†	24748.8	24.5	mg/L	0.05	24.5 mg/L	0.05	0.19%
	QC value within limits for K	Recovery =	97.81%				
Mg†	255314.6	25.7	mg/L	0.08	25.7 mg/L	0.08	0.33%
	QC value within limits for Mg	Recovery =	102.71%				
Mn†	1196943.2	2.55	mg/L	0.025	2.55 mg/L	0.025	1.00%
	QC value within limits for Mn	Recovery =	101.94%				
Mo†	24117.4	2.49	mg/L	0.001	2.49 mg/L	0.001	0.06%
	QC value within limits for Mo	Recovery =	99.58%				
Na†	64730.0	24.4	mg/L	0.03	24.4 mg/L	0.03	0.13%
	QC value within limits for Na	Recovery =	97.65%				
Ni†	47469.1	2.62	mg/L	0.011	2.62 mg/L	0.011	0.43%
	QC value within limits for Ni	Recovery =	104.60%				
Pb†	9557.2	2.58	mg/L	0.005	2.58 mg/L	0.005	0.20%
	QC value within limits for Pb	Recovery =	103.37%				
Sb†	3675.4	2.38	mg/L	0.005	2.38 mg/L	0.005	0.21%
	QC value within limits for Sb	Recovery =	95.34%				

Se†	2327.1	2.48 mg/L	0.013	2.48 mg/L	0.013	0.54%
QC value within limits for Se Recovery = 99.15%						
Tl†	5648.6	2.68 mg/L	0.000	2.68 mg/L	0.000	0.01%
QC value within limits for Tl Recovery = 107.12%						
V†	358453.9	2.47 mg/L	0.004	2.47 mg/L	0.004	0.16%
QC value within limits for V Recovery = 98.89%						
Zn†	96145.9	2.56 mg/L	0.008	2.56 mg/L	0.008	0.32%
QC value within limits for Zn Recovery = 102.34%						
All analyte(s) passed QC.						



Se†	-0.8	-0.00088	mg/L	0.002553	-0.00088	mg/L	0.002553	290.63%
QC value within limits for Se								
Recovery =								
Not calculated								
Tl†	2.2	0.00103	mg/L	0.000841	0.00103	mg/L	0.000841	81.82%
QC value within limits for Tl								
Recovery =								
Not calculated								
V†	18.7	0.00013	mg/L	0.000063	0.00013	mg/L	0.000063	48.76%
QC value within limits for V								
Recovery =								
Not calculated								
Zn†	8.7	0.00023	mg/L	0.000081	0.00023	mg/L	0.000081	35.15%
QC value within limits for Zn								
Recovery =								
Not calculated								

All analyte(s) passed QC.



Sequence No.: 124  
 Sample ID: ECV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/21/2008 22:18:24  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

## Nebulizer Parameters: ECV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

## Mean Data: ECV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	369310.8	91.2 %	0.22			0.24%
Yr	329728.4	102 %	0.3			0.27%
Ca	829200.8	51.7 mg/L	0.58	51.7 mg/L	0.58	1.12%
	QC value within limits for Ca Recovery = 103.34%					
Cr	210014.5	5.11 mg/L	0.016	5.11 mg/L	0.016	0.32%
	QC value within limits for Cr Recovery = 102.26%					
Fe	20074.6	5.78 mg/L	0.002	5.78 mg/L	0.002	0.04%
	QC value greater than the upper limit for Fe Recovery = 115.58%					
Mn	2413150.7	5.14 mg/L	0.009	5.14 mg/L	0.009	0.18%
	QC value within limits for Mn Recovery = 102.78%					
Na	145271.7	54.8 mg/L	0.00	54.8 mg/L	0.00	0.00%
	QC value within limits for Na Recovery = 109.58%					
QC Failed. Retry.						

Sequence No.: 125  
 Sample ID: ECV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/21/2008 22:19:49  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

## Nebulizer Parameters: ECV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

## Mean Data: ECV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	372221.5	92.0 %	0.42			0.46%
Yr	330252.7	102 %	0.4			0.36%
Ca	842102.9	52.5 mg/L	0.02	52.5 mg/L	0.02	0.04%
	QC value within limits for Ca Recovery = 104.95%					
Cr	208779.6	5.08 mg/L	0.033	5.08 mg/L	0.033	0.64%
	QC value within limits for Cr Recovery = 101.65%					
Fe	20147.1	5.80 mg/L	0.066	5.80 mg/L	0.066	1.14%
	QC value greater than the upper limit for Fe Recovery = 116.00%					
Mn	2418127.9	5.15 mg/L	0.001	5.15 mg/L	0.001	0.02%
	QC value within limits for Mn Recovery = 102.99%					
Na	144824.6	54.6 mg/L	0.13	54.6 mg/L	0.13	0.23%
	QC value within limits for Na Recovery = 109.24%					
QC Failed. Continue with analysis.						

Sequence No.: 126  
 Sample ID: ECB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/21/2008 22:22:45  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: ECB

Analyte Back Pressure Flow  
 All 214.0 kPa 0.65 L/min

Mean Data: ECB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	396550.5	98.0	%	1.32			1.35%
Yr	347346.7	108	%	1.1			0.98%
Ca†	62.3	0.00388	mg/L	0.000082	0.00388 mg/L	0.000082	2.11%
	QC value within limits for Ca Recovery = Not calculated						
Cr†	14.5	0.00035	mg/L	0.000088	0.00035 mg/L	0.000088	24.81%
	QC value within limits for Cr Recovery = Not calculated						
Fe†	-3.7	-0.00105	mg/L	0.000093	-0.00105 mg/L	0.000093	8.79%
	QC value within limits for Fe Recovery = Not calculated						
Mn†	-3.7	-0.00001	mg/L	0.000037	-0.00001 mg/L	0.000037	468.35%
	QC value within limits for Mn Recovery = Not calculated						
Na†	290.1	0.109	mg/L	0.0138	0.109 mg/L	0.0138	12.60%
	QC value within limits for Na Recovery = Not calculated						

All analyte(s) passed QC.

=====  
Analysis Begun

Start Time: 5/21/2008 22:30:26 Plasma On Time: 5/21/2008 10:24:22  
Logged In Analyst: Charley Kay Technique: ICP Continuous  
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
Batch ID: 080521  
Results Data Set: 080521  
Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 1 Autosampler Location: 15  
Sample ID: ICV Date Collected: 5/21/2008 22:30:27  
Analyst: Data Type: Original  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:  
User canceled analysis.

=====  
Analysis Begun

Start Time: 5/21/2008 22:31:27 Plasma On Time: 5/21/2008 10:24:22  
Logged In Analyst: Charley Kay Technique: ICP Continuous  
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif  
Batch ID: 080521  
Results Data Set: 080521  
Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 120 Autosampler Location: 4  
Sample ID: CCV Date Collected: 5/21/2008 22:31:27  
Analyst: Data Type: Original  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:

=====  
Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

=====  
Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	375575.2	92.8	%	0.08			0.09%
Yr	336674.9	104	%	0.1			0.09%
Ca†	831478.3	51.8	mg/L	0.16	51.8 mg/L	0.16	0.31%
	QC value within limits for Ca		Recovery = 103.63%				
Cr†	205235.6	5.00	mg/L	0.009	5.00 mg/L	0.009	0.19%
	QC value within limits for Cr		Recovery = 99.93%				
Fe†	20068.3	5.78	mg/L	0.017	5.78 mg/L	0.017	0.29%
	QC value greater than the upper limit for Fe		Recovery = 115.54%				
Mn†	2372391.5	5.05	mg/L	0.000	5.05 mg/L	0.000	0.01%
	QC value within limits for Mn		Recovery = 101.04%				
Na†	143267.7	54.0	mg/L	0.22	54.0 mg/L	0.22	0.41%
	QC value within limits for Na		Recovery = 108.07%				
QC Failed. Retry.							

=====  
Sequence No.: 121 Autosampler Location: 4  
Sample ID: CCV Date Collected: 5/21/2008 22:32:51  
Analyst: Data Type: Original  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:

=====  
Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
All 214.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	372761.9	92.1 %		0.19			0.20%
Yr	331633.2	103 %		0.1			0.11%
Ca†	825400.5	51.4 mg/L		0.14	51.4 mg/L	0.14	0.28%
	QC value within limits for Ca		Recovery = 102.87%				
Cr†	207030.9	5.04 mg/L		0.028	5.04 mg/L	0.028	0.56%
	QC value within limits for Cr		Recovery = 100.80%				
Fe†	19630.7	5.65 mg/L		0.046	5.65 mg/L	0.046	0.81%
	QC value greater than the upper limit for Fe		Recovery = 113.02%				
Mn†	2391639.8	5.09 mg/L		0.001	5.09 mg/L	0.001	0.01%
	QC value within limits for Mn		Recovery = 101.86%				
Na†	141890.9	53.5 mg/L		0.16	53.5 mg/L	0.16	0.30%
	QC value within limits for Na		Recovery = 107.03%				

QC Failed. Retry.

Sequence No.: 122  
Sample ID: CCV  
Analyst:  
Initial Sample Wt:  
Dilution:

Autosampler Location: 4  
Date Collected: 5/21/2008 22:34:09  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	368790.5	91.1 %		0.14			0.16%
Yr	330365.0	102 %		1.1			1.10%
Ca†	840081.4	52.3 mg/L		0.16	52.3 mg/L	0.16	0.31%
	QC value within limits for Ca		Recovery = 104.70%				
Cr†	209442.8	5.10 mg/L		0.002	5.10 mg/L	0.002	0.03%
	QC value within limits for Cr		Recovery = 101.98%				
Fe†	19999.6	5.76 mg/L		0.000	5.76 mg/L	0.000	0.01%
	QC value greater than the upper limit for Fe		Recovery = 115.15%				
Mn†	2413381.0	5.14 mg/L		0.003	5.14 mg/L	0.003	0.06%
	QC value within limits for Mn		Recovery = 102.79%				
Na†	144758.4	54.6 mg/L		0.09	54.6 mg/L	0.09	0.16%
	QC value within limits for Na		Recovery = 109.19%				

QC Failed. Continue with analysis.

Sequence No.: 123  
Sample ID: CCB  
Analyst:  
Initial Sample Wt:  
Dilution:

Autosampler Location: 0  
Date Collected: 5/21/2008 22:37:04  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	396375.1	97.9 %		0.38			0.39%
Yr	346445.4	107 %		1.1			1.05%
Cr†	5.7	0.00014 mg/L		0.000052	0.00014 mg/L	0.000052	37.93%
	QC value within limits for Cr Recovery = Not calculated						
Mn†	-4.3	-0.00001 mg/L		0.000019	-0.00001 mg/L	0.000019	209.14%
	QC value within limits for Mn Recovery = Not calculated						
Na†	179.1	0.0676 mg/L		0.01028	0.0676 mg/L	0.01028	15.22%
	QC value within limits for Na Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 124  
 Sample ID: MBLANK2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 122  
 Date Collected: 5/21/2008 22:39:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MBLANK2007

Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: MBLANK2007

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity				Conc. Units			
Sca	408850.3		101 %	0.3				0.26%
Yr	356261.5		110 %	0.3				0.27%
Cr†	-2.4	-0.00006	mg/L	0.000096	-0.00006	mg/L	0.000096	166.18%
Mn†	-75.8	-0.00016	mg/L	0.000020	-0.00016	mg/L	0.000020	12.19%
Na†	212.7	0.0802	mg/L	0.00169	0.0802	mg/L	0.00169	2.10%

Sequence No.: 125  
 Sample ID: MRL  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 21  
 Date Collected: 5/21/2008 22:42:45  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MRL  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	399421.0	98.7 %	%	0.10			0.10%
Yr	354149.7	110 %	%	0.6			0.56%
Crt	413.9	0.0101 mg/L	mg/L	0.00017	0.0101 mg/L	0.00017	1.69%
Mnt	948.3	0.00202 mg/L	mg/L	0.000012	0.00202 mg/L	0.000012	0.62%
Naf	2895.1	1.09 mg/L	mg/L	0.014	1.09 mg/L	0.014	1.31%

Sequence No.: 126  
Sample ID: MRL2007  
Analyst:  
Initial Sample Wt:  
Dilution: 1X

Autosampler Location: 24  
Date Collected: 5/21/2008 22:45:40  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

## Nebulizer Parameters: MRL2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

## Mean Data: MRL2007

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	400442.8	98.9 %		1.15			1.16%
Yr	351399.0	109 %		0.2			0.17%
Crt	472.6	0.0115 mg/L		0.00017	0.0115 mg/L	0.00017	1.44%
Mnt	1044.8	0.00223 mg/L		0.000036	0.00223 mg/L	0.000036	1.61%
Nat	3276.4	1.24 mg/L		0.018	1.24 mg/L	0.018	1.46%



Sequence No.: 127  
 Sample ID: LCS2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 123  
 Date Collected: 5/21/2008 22:48:38  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: LCS2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: LCS2007

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	377194.1	93.2 %	0.36			0.39%
Yr	341442.2	106 %	0.5			0.48%
Crt	40461.2	0.985 mg/L	0.0016	0.985 mg/L	0.0016	0.16%
Mnt	237187.6	0.505 mg/L	0.0013	0.505 mg/L	0.0013	0.25%
Nat	143155.7	54.0 mg/L	0.09	54.0 mg/L	0.09	0.16%

Sequence No.: 128  
 Sample ID: LCSD2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 124  
 Date Collected: 5/21/2008 22:51:37  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: LCSD2007

Analyte	Back Pressure	Flow
All	214.0 kPa	0.65 L/min

Mean Data: LCSD2007

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity				Conc. Units			
Sca	382162.8		94.4 %	0.29				0.30%
Yr	346983.5		107 %	0.8				0.75%
Crt	39162.7		0.953 mg/L	0.0019	0.953 mg/L	0.0019		0.20%
Mnt	229917.9		0.490 mg/L	0.0011	0.490 mg/L	0.0011		0.22%
Nat	136476.2		51.5 mg/L	0.01	51.5 mg/L	0.01		0.01%

Sequence No.: 133  
 Sample ID: 2805080991\_2XMS  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 129  
 Date Collected: 5/21/2008 23:06:39  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805080991\_2XMS  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805080991\_2XMS

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	387962.0		95.9 %	0.39			0.40%
Yr	340106.6		105 %	0.3			0.25%
Crt	19505.2		0.475 mg/L	0.0011	0.950 mg/L	0.0021	0.22%
Mnt	115599.8		0.246 mg/L	0.0009	0.492 mg/L	0.0017	0.35%
Naf	67133.7		25.3 mg/L	0.04	50.6 mg/L	0.09	0.18%

Sequence No.: 134  
 Sample ID: CCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/21/2008 23:09:02  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	370427.5	91.5 %	0.36			0.39%
Yr	327891.7	101 %	0.3			0.26%
Cr	207309.9	5.05 mg/L	0.039	5.05 mg/L	0.039	0.78%
	QC value within limits for Cr Recovery = 100.94%					
Mn	2405119.0	5.12 mg/L	0.017	5.12 mg/L	0.017	0.34%
	QC value within limits for Mn Recovery = 102.44%					
Na	144749.3	54.6 mg/L	0.09	54.6 mg/L	0.09	0.16%
	QC value within limits for Na Recovery = 109.18%					
All analyte(s) passed QC.						

Sequence No.: 135  
 Sample ID: CCB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/21/2008 23:12:00  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	214.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	395239.1	97.7	%	0.18			0.19%
Yr	341600.1	106	%	0.1			0.08%
Cr†	3.5	0.00009	mg/L	0.000015	0.00009 mg/L	0.000015	17.16%
	QC value within limits for Cr Recovery = Not calculated						
Mn†	-40.8	-0.00009	mg/L	0.000019	-0.00009 mg/L	0.000019	21.62%
	QC value within limits for Mn Recovery = Not calculated						
Na†	254.0	0.0958	mg/L	0.00797	0.0958 mg/L	0.00797	8.32%
	QC value within limits for Na Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 136  
Sample ID: MCV  
Analyst:  
Initial Sample Wt:  
Dilution:

Autosampler Location: 5  
Date Collected: 5/21/2008 23:14:42  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	373869.0	92.4	%	0.18			0.20%
Yr	329292.7	102	%	1.4			1.39%
Cr†	104235.0	2.54	mg/L	0.011	2.54 mg/L	0.011	0.44%
	QC value within limits for Cr		Recovery =	101.50%			
Mn†	1224501.9	2.61	mg/L	0.004	2.61 mg/L	0.004	0.14%
	QC value within limits for Mn		Recovery =	104.31%			
Na†	70636.9	26.6	mg/L	0.08	26.6 mg/L	0.08	0.30%
	QC value within limits for Na		Recovery =	106.56%			

All analyte(s) passed QC.

Sequence No.: 137  
 Sample ID: 2805080991\_2XMSD  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 130  
 Date Collected: 5/21/2008 23:17:00  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805080991\_2XMSD  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805080991\_2XMSD

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	389418.9	96.2 %	0.34			0.35%
Yr	337327.2	104 %	0.1			0.07%
Cr†	20539.8	0.500 mg/L	0.0038	1.00 mg/L	0.008	0.77%
Mn†	121192.3	0.258 mg/L	0.0008	0.516 mg/L	0.0016	0.32%
Nat	69437.0	26.2 mg/L	0.04	52.4 mg/L	0.08	0.14%

User canceled analysis.

=====  
Analysis Begun

Start Time: 5/21/2008 23:28:08

Plasma On Time: 5/21/2008 10:24:22

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080521.sif

Batch ID: 080521

Results Data Set: 080521

Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====  
Sequence No.: 147

Autosampler Location: 4

Sample ID: CCV

Date Collected: 5/21/2008 23:28:09

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

-----  
Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

-----  
Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	375047.2	92.7 %		1.00			1.08%
Yr	334101.6	103 %		1.9			1.85%
Cr	203441.0	4.95 mg/L		0.020	4.95 mg/L	0.020	0.41%
QC value within limits for Cr Recovery = 99.05%							
Mn	2354893.2	5.01 mg/L		0.020	5.01 mg/L	0.020	0.39%
QC value within limits for Mn Recovery = 100.30%							
Na	139252.3	52.5 mg/L		0.16	52.5 mg/L	0.16	0.30%
QC value within limits for Na Recovery = 105.04%							
All analyte(s) passed QC.							



Sequence No.: 147  
Sample ID: CCV  
Analyst:  
Initial Sample Wt:  
Dilution:

Autosampler Location: 4  
Date Collected: 5/21/2008 23:49:31  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	374167.1	92.4 %	0.71			0.76%
Yr	326818.7	101 %	0.6			0.58%
Crt	204690.7	4.98 mg/L	0.007	4.98 mg/L	0.007	0.14%
QC value within limits for Cr		Recovery = 99.66%				
Mnt	2375506.3	5.06 mg/L	0.004	5.06 mg/L	0.004	0.09%
QC value within limits for Mn		Recovery = 101.18%				
Nat	141534.5	53.4 mg/L	0.02	53.4 mg/L	0.02	0.04%
QC value within limits for Na		Recovery = 106.76%				

All analyte(s) passed QC.

Sequence No.: 148  
 Sample ID: CCB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/21/2008 23:52:29  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	396712.2	98.0 %		0.64			0.66%
Yr	341714.7	106 %		0.2			0.17%
Cr	59.8	0.00145 mg/L		0.000100	0.00145 mg/L	0.000100	6.91%
	QC value within limits for Cr	Recovery =	Not calculated				
Mn	-49.3	-0.00010 mg/L		0.000017	-0.00010 mg/L	0.000017	16.19%
	QC value within limits for Mn	Recovery =	Not calculated				
Na	232.1	0.0875 mg/L		0.00462	0.0875 mg/L	0.00462	5.27%
	QC value within limits for Na	Recovery =	Not calculated				

All analyte(s) passed QC.

Sequence No.: 158  
 Sample ID: MBLANK2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 38  
 Date Collected: 5/22/2008 00:22:31  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MBLANK2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MBLANK2007

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Sca	407090.4	101	%	0.0				0.03%
Yr	353875.5	110	%	0.5				0.41%
Crt	15.8	0.00039	mg/L	0.000036	0.00039	mg/L	0.000036	9.25%
Mnt	-109.5	-0.00023	mg/L	0.000016	-0.00023	mg/L	0.000016	6.78%
Nat	528.6	0.199	mg/L	0.0094	0.199	mg/L	0.0094	4.69%

Sequence No.: 159  
 Sample ID: CCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/22/2008 00:25:26  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	370103.1	91.4 %	%	0.63			0.69%
Yr	328178.7	102 %	%	0.8			0.82%
Crt	204896.8	4.99 mg/L	mg/L	0.008	4.99 mg/L	0.008	0.16%
QC value within limits for Cr		Recovery = 99.76%					
Mnt	2375595.0	5.06 mg/L	mg/L	0.007	5.06 mg/L	0.007	0.14%
QC value within limits for Mn		Recovery = 101.18%					
Naf	142408.1	53.7 mg/L	mg/L	0.16	53.7 mg/L	0.16	0.30%
QC value within limits for Na		Recovery = 107.42%					

All analyte(s) passed QC.

Sequence No.: 160  
 Sample ID: CCB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/22/2008 00:28:24  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. %	Calib. Units	Std.Dev.	Sample Conc. mg/L	Units	Std.Dev.	RSD
Sca	396715.4	98.0	%	0.45				0.45%
Yr	343208.7	106	%	0.8				0.73%
Cr	21.7	0.00053	mg/L	0.000061	0.00053	mg/L	0.000061	11.61%
QC value within limits for Cr Recovery = Not calculated								
Mn	-39.1	-0.00008	mg/L	0.000022	-0.00008	mg/L	0.000022	25.89%
QC value within limits for Mn Recovery = Not calculated								
Na	260.9	0.0984	mg/L	0.00551	0.0984	mg/L	0.00551	5.60%
QC value within limits for Na Recovery = Not calculated								
All analyte(s) passed QC.								

Sequence No.: 161  
 Sample ID: MCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 5  
 Date Collected: 5/22/2008 00:31:06  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	381147.9	94.2 %		0.80			0.85%
Yr	331248.8	103 %		0.8			0.74%
Cr†	103302.6	2.51 mg/L		0.004	2.51 mg/L	0.004	0.16%
	QC value within limits for Cr		Recovery = 100.60%				
Mn†	1211047.4	2.58 mg/L		0.002	2.58 mg/L	0.002	0.08%
	QC value within limits for Mn		Recovery = 103.16%				
Na†	68786.5	25.9 mg/L		0.26	25.9 mg/L	0.26	1.00%
	QC value within limits for Na		Recovery = 103.77%				

All analyte(s) passed QC.

Sequence No.: 162  
 Sample ID: MRL  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 21  
 Date Collected: 5/22/2008 00:33:24  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	395749.0	97.8 %	0.17			0.17%
Yr	343913.4	106 %	1.2			1.14%
Crt	432.5	0.0105 mg/L	0.00006	0.0105 mg/L	0.00006	0.52%
Mnt	963.7	0.00205 mg/L	0.000010	0.00205 mg/L	0.000010	0.46%
Nat	2876.0	1.08 mg/L	0.001	1.08 mg/L	0.001	0.05%

Sequence No.: 163  
 Sample ID: MRL2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 24  
 Date Collected: 5/22/2008 00:36:20  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MRL2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MRL2007

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	400390.9	98.9 %	%	0.52			0.53%
Yr	346330.6	107 %	%	1.5			1.42%
Crf	469.2	0.0114 mg/L	mg/L	0.00024	0.0114 mg/L	0.00024	2.12%
Mnt	1022.7	0.00218 mg/L	mg/L	0.000019	0.00218 mg/L	0.000019	0.89%
Nat	3146.8	1.19 mg/L	mg/L	0.018	1.19 mg/L	0.018	1.55%



Sequence No.: 164  
 Sample ID: LCS2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 39  
 Date Collected: 5/22/2008 00:39:18  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: LCS2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: LCS2007

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	383616.3	94.8 %	%	1.42			1.50%
Yr	334469.1	104 %	%	0.5			0.51%
Crt	39392.1	0.959 mg/L	mg/L	0.0089	0.959 mg/L	0.0089	0.93%
Mnt	233866.5	0.498 mg/L	mg/L	0.0035	0.498 mg/L	0.0035	0.71%
Nat	140189.3	52.9 mg/L	mg/L	0.02	52.9 mg/L	0.02	0.03%

Sequence No.: 165  
 Sample ID: LCSD2007  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 1X

Autosampler Location: 40  
 Date Collected: 5/22/2008 00:42:14  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: LCSD2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: LCSD2007

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	388611.0	96.0 %	0.00			0.00%
Yr	334349.6	103 %	0.7			0.67%
Crt	39465.5	0.961 mg/L	0.0034	0.961 mg/L	0.0034	0.35%
Mnt	234443.1	0.499 mg/L	0.0006	0.499 mg/L	0.0006	0.11%
Nat	138969.5	52.4 mg/L	0.09	52.4 mg/L	0.09	0.18%

Sequence No.: 167  
 Sample ID: 2805090119\_2XMS  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 42  
 Date Collected: 5/22/2008 00:48:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090119\_2XMS  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090119\_2XMS

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	377347.2	93.2 %	%	0.30			0.32%
Yr	329710.6	102 %	%	1.5			1.47%
Crt	21632.1	0.527 mg/L	mg/L	0.0012	1.05 mg/L	0.002	0.23%
Mnt	123135.6	0.262 mg/L	mg/L	0.0017	0.524 mg/L	0.0034	0.65%
Nat	301096.9	114 mg/L	mg/L	0.0	227 mg/L	0.1	0.03%

Sequence No.: 168  
 Sample ID: 2805090119\_2XMSD  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 43  
 Date Collected: 5/22/2008 00:51:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090119\_2XMSD  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090119\_2XMSD

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	374944.8	92.6 %		0.22			0.24%
Yr	334636.7	104 %		0.1			0.09%
Crt	20805.0	0.506 mg/L		0.0014	1.01 mg/L	0.003	0.28%
Mnt	117881.6	0.251 mg/L		0.0010	0.502 mg/L	0.0020	0.40%
Nat	292865.8	110 mg/L		0.2	221 mg/L	0.4	0.16%

Sequence No.: 170  
 Sample ID: 2805090120\_2XMS  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 45  
 Date Collected: 5/22/2008 00:57:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090120\_2XMS  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090120\_2XMS

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	365066.9	90.2 %	%	0.60			0.67%
Yr	322422.7	99.8 %	%	0.87			0.87%
Cri	23209.5	0.565 mg/L	mg/L	0.0056	1.13 mg/L	0.011	0.98%
Mnt	120908.8	0.257 mg/L	mg/L	0.0016	0.515 mg/L	0.0031	0.60%
Nat	512010.7	193 mg/L	mg/L	0.2	386 mg/L	0.4	0.09%

Sequence No.: 171  
 Sample ID: 2805090120\_2XMSD  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 46  
 Date Collected: 5/22/2008 01:00:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090120\_2XMSD  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090120\_2XMSD

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	362303.7	89.5 %	%	0.29			0.32%
Yr	321721.3	99.6 %	%	0.02			0.02%
Cri	23974.5	0.584 mg/L	mg/L	0.0034	1.17 mg/L	0.007	0.58%
Mnf	126545.2	0.269 mg/L	mg/L	0.0006	0.539 mg/L	0.0013	0.24%
Nat	527769.8	199 mg/L	mg/L	1.5	398 mg/L	2.9	0.73%

Sequence No.: 172  
Sample ID: CCV  
Analyst:  
Initial Sample Wt:  
Dilution:

Autosampler Location: 4  
Date Collected: 5/22/2008 01:03:43  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	371701.0	91.8 %		0.86			0.93%
Yr	329948.9	102 %		0.0			0.01%
Cr	202952.0	4.94 mg/L		0.034	4.94 mg/L	0.034	0.68%
	QC value within limits for Cr		Recovery = 98.82%				
Mn	2362929.7	5.03 mg/L		0.010	5.03 mg/L	0.010	0.20%
	QC value within limits for Mn		Recovery = 100.64%				
Na	143668.6	54.2 mg/L		0.04	54.2 mg/L	0.04	0.07%
	QC value within limits for Na		Recovery = 108.37%				
All analyte(s) passed QC.							

Sequence No.: 173  
 Sample ID: CCB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/22/2008 01:06:42  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	391544.8	96.7 %		1.45			1.50%
Yr	342606.9	106 %		0.1			0.14%
Cr	12.7	0.00031 mg/L		0.000047	0.00031 mg/L	0.000047	15.34%
	QC value within limits for Cr Recovery = Not calculated						
Mn	-40.3	-0.00009 mg/L		0.000005	-0.00009 mg/L	0.000005	6.14%
	QC value within limits for Mn Recovery = Not calculated						
Na	184.2	0.0695 mg/L		0.00324	0.0695 mg/L	0.00324	4.66%
	QC value within limits for Na Recovery = Not calculated						
All analyte(s) passed QC.							



Sequence No.: 184  
 Sample ID: CCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/22/2008 01:39:40  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	373204.2	92.2 %	%	0.41				0.45%
Yr	324550.6	100 %	%	0.6				0.62%
Cr†	200190.6	4.87 mg/L	mg/L	0.022	4.87 mg/L		0.022	0.44%
	QC value within limits for Cr		Recovery = 97.47%					
Mn†	2354258.9	5.01 mg/L	mg/L	0.006	5.01 mg/L		0.006	0.13%
	QC value within limits for Mn		Recovery = 100.27%					
Na†	141259.2	53.3 mg/L	mg/L	0.04	53.3 mg/L		0.04	0.07%
	QC value within limits for Na		Recovery = 106.55%					
All analyte(s) passed QC.								

Sequence No.: 185  
 Sample ID: CCB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/22/2008 01:42:39  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	395528.9	97.7	%	0.21			0.21%
Yr	337200.5	104	%	0.0			0.01%
Cr†	93.5	0.00228	mg/L	0.000003	0.00228 mg/L	0.000003	0.13%
	QC value within limits for Cr Recovery = Not calculated						
Mn†	-71.7	-0.00015	mg/L	0.000005	-0.00015 mg/L	0.000005	3.10%
	QC value within limits for Mn Recovery = Not calculated						
Na†	291.9	0.110	mg/L	0.0180	0.110 mg/L	0.0180	16.31%
	QC value within limits for Na Recovery = Not calculated						
All analyte(s) passed QC.							

Sequence No.: 186  
 Sample ID: MCV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 5  
 Date Collected: 5/22/2008 01:45:21  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	382869.6	94.6 %	%	0.72			0.76%
Yr	325214.2	101 %	%	1.6			1.63%
Crt	102203.5	2.49 mg/L	mg/L	0.012	2.49 mg/L	0.012	0.47%
	QC value within limits for Cr Recovery = 99.53%						
Mnt	1210343.1	2.58 mg/L	mg/L	0.007	2.58 mg/L	0.007	0.29%
	QC value within limits for Mn Recovery = 103.10%						
Naf	70174.8	26.5 mg/L	mg/L	0.04	26.5 mg/L	0.04	0.17%
	QC value within limits for Na Recovery = 105.86%						
All analyte(s) passed QC.							

Sequence No.: 195  
 Sample ID: ECV  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 4  
 Date Collected: 5/22/2008 02:11:48  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: ECV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ECV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Sca	368037.1	90.9	%	0.80				0.88%
Yr	323335.8	100	%	0.4				0.40%
Cr	200164.7	4.87	mg/L	0.026	4.87	mg/L	0.026	0.54%
	QC value within limits for Cr		Recovery =	97.46%				
Mn	2361303.5	5.03	mg/L	0.014	5.03	mg/L	0.014	0.28%
	QC value within limits for Mn		Recovery =	100.57%				
Na	144010.3	54.3	mg/L	0.01	54.3	mg/L	0.01	0.02%
	QC value within limits for Na		Recovery =	108.63%				

All analyte(s) passed QC.

Sequence No.: 196  
 Sample ID: ECB  
 Analyst:  
 Initial Sample Wt:  
 Dilution:

Autosampler Location: 0  
 Date Collected: 5/22/2008 02:14:46  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: ECB

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: ECB

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		
	Intensity	Conc.			Conc.	Units	Std.Dev.
Sca	397081.4	98.1	%	0.04			0.04%
Yr	332487.5	103	%	0.2			0.15%
Cr†	33.4	0.00081	mg/L	0.000156	0.00081 mg/L	0.000156	19.17%
	QC value within limits for Cr		Recovery =	Not calculated			
Mn†	-76.5	-0.00016	mg/L	0.000024	-0.00016 mg/L	0.000024	14.44%
	QC value within limits for Mn		Recovery =	Not calculated			
Na†	318.6	0.120	mg/L	0.0004	0.120 mg/L	0.0004	0.35%
	QC value within limits for Na		Recovery =	Not calculated			

All analyte(s) passed QC.

Sequence No.: 132  
 Sample ID: 2805080991\_2X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 128  
 Date Collected: 5/21/2008 23:03:41  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805080991\_2X  
 Analyte Back Pressure Flow  
 All 214.0 kPa 0.65 L/min

Mean Data: 2805080991\_2X

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity				Conc.	Units		
Sca	401163.3		99.1 %	0.76				0.77%
Yr	347599.4		108 %	0.8				0.72%
Crt	6.5	0.00016	mg/L	0.000121	0.00032	mg/L	0.000242	76.39%
Mnt	-56.9	-0.00012	mg/L	0.000013	-0.00024	mg/L	0.000026	10.93%
Naf	478.1	0.180	mg/L	0.0007	0.361	mg/L	0.0013	0.37%

Sequence No.: 157  
 Sample ID: 2805090112\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 148  
 Date Collected: 5/22/2008 00:19:27  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090112\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090112\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	360944.5	89.2 %	%	0.45			0.50%
Yr	331352.2	103 %	%	0.5			0.50%
Crt	7463.0	0.182 mg/L	mg/L	0.0019	0.908 mg/L	0.0095	1.05%
Mnr	2037.8	0.00434 mg/L	mg/L	0.000037	0.0217 mg/L	0.00018	0.84%
Nat	741785.0	280 mg/L	mg/L	0.2	1400 mg/L	1.0	0.07%

Sequence No.: 166  
 Sample ID: 2805090119\_2X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 41  
 Date Collected: 5/22/2008 00:45:10  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090119\_2X  
 Analyte Back Pressure Flow  
 All 212.0 kPa 0.65 L/min

Mean Data: 2805090119\_2X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	382316.3	94.5 %		1.46			1.55%
Yr	332178.9	103 %		0.8			0.79%
Crt	1329.5	0.0324 mg/L		0.00033	0.0647 mg/L	0.00066	1.02%
Mni	309.7	0.00066 mg/L		0.000035	0.00132 mg/L	0.000070	5.34%
Nat	228292.3	86.1 mg/L		0.46	172 mg/L	0.9	0.54%



Sequence No.: 169  
Sample ID: 2805090120\_2X  
Analyst:  
Initial Sample Wt:  
Dilution: 2X

Autosampler Location: 44  
Date Collected: 5/22/2008 00:54:46  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: 2805090120\_2X  
Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: 2805090120\_2X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	369959.4	91.4 %		0.40			0.44%
Yr	325478.2	101 %		0.3			0.33%
Crt	3705.0	0.0902 mg/L		0.00011	0.180 mg/L	0.0002	0.12%
Mnt	2136.3	0.00455 mg/L		0.000002	0.00910 mg/L	0.000004	0.04%
Nat	464703.9	175 mg/L		0.1	351 mg/L	0.1	0.03%

Sequence No.: 174  
 Sample ID: 2805090115\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 47  
 Date Collected: 5/22/2008 01:09:24  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090115\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090115\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	372277.8	92.0 %	1.01			1.10%
Yr	328377.7	102 %	0.8			0.80%
Crt	31131.9	0.758 mg/L	0.0144	3.79 mg/L	0.072	1.90%
Mnt	12183.8	0.0259 mg/L	0.00033	0.130 mg/L	0.0016	1.26%
Nat	318639.0	120 mg/L	0.2	601 mg/L	1.0	0.16%

Sequence No.: 175  
 Sample ID: 2805090116\_2X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 2X

Autosampler Location: 48  
 Date Collected: 5/22/2008 01:12:26  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090116\_2X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090116\_2X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	338057.7	83.5 %		0.96			1.15%
Yr	310369.1	96.1 %		0.85			0.89%
Crt	18171.5	0.442 mg/L		0.0022	0.885 mg/L	0.0044	0.50%
Mnf	5732.4	0.0122 mg/L		0.00004	0.0244 mg/L	0.00009	0.37%
Naf	2001431.9	755 mg/L		10.1	1510 mg/L	20.2	1.34%

Sequence No.: 176  
 Sample ID: 2805090117\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 49  
 Date Collected: 5/22/2008 01:15:32  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090117\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090117\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	377775.6	93.3 %	%	0.68			0.73%
Yr	329418.4	102 %	%	0.8			0.74%
Crt	23915.6	0.582 mg/L	mg/L	0.0007	2.91 mg/L	0.004	0.12%
Mnt	4447.5	0.00947 mg/L	mg/L	0.000114	0.0474 mg/L	0.00057	1.21%
Nat	320306.3	121 mg/L	mg/L	0.3	604 mg/L	1.5	0.25%

Sequence No.: 177  
 Sample ID: 2805090118\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 50  
 Date Collected: 5/22/2008 01:18:34  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090118\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090118\_5X

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	370331.4	91.5 %	0.03			0.04%
Yr	326585.9	101 %	0.9			0.92%
Crt	22501.3	0.548 mg/L	0.0019	2.74 mg/L	0.009	0.34%
Mnt	2412.9	0.00514 mg/L	0.000029	0.0257 mg/L	0.00014	0.55%
Nat	374474.8	141 mg/L	0.1	706 mg/L	0.5	0.07%

Sequence No.: 178  
Sample ID: 2805090121\_5X  
Analyst:  
Initial Sample Wt:  
Dilution: 5X

Autosampler Location: 51  
Date Collected: 5/22/2008 01:21:34  
Data Type: Original  
Initial Sample Vol:  
Sample Prep Vol:

Nebulizer Parameters: 2805090121\_5X  
Analyte Back Pressure Flow  
All 213.0 kPa 0.65 L/min

Mean Data: 2805090121\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	364241.0	90.0 %		0.25			0.28%
Yr	324299.6	100 %		1.2			1.17%
Crt	32849.0	0.800 mg/L		0.0051	4.00 mg/L	0.026	0.54%
Mnt	1302.4	0.00277 mg/L		0.000011	0.0139 mg/L	0.00005	0.39%
Nat	583567.8	220 mg/L		0.0	1100 mg/L	0.2	0.02%

Sequence No.: 179  
 Sample ID: 2805090122\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 52  
 Date Collected: 5/22/2008 01:24:35  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090122\_5X

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: 2805090122\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	359700.2	88.9 %	%	0.82			0.93%
Yr	317951.0	98.4 %	%	0.91			0.92%
Crt	40149.2	0.977 mg/L	mg/L	0.0049	4.89 mg/L	0.024	0.50%
Mn†	3627.6	0.00773 mg/L	mg/L	0.000069	0.0386 mg/L	0.00034	0.89%
Na†	635441.2	240 mg/L	mg/L	0.3	1200 mg/L	1.5	0.12%

Sequence No.: 180  
 Sample ID: 2805090123\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 53  
 Date Collected: 5/22/2008 01:27:36  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090123\_5X

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: 2805090123\_5X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	358804.0	88.7 %	%	0.68			0.77%
Yr	320035.8	99.1 %	%	0.45			0.45%
Cr†	35120.4	0.855 mg/L	mg/L	0.0090	4.28 mg/L	0.045	1.05%
Mnt	10446.0	0.0222 mg/L	mg/L	0.00015	0.111 mg/L	0.0007	0.67%
Nat	686839.6	259 mg/L	mg/L	0.5	1300 mg/L	2.7	0.21%



Sequence No.: 181  
 Sample ID: 2805090124\_10X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 10X

Autosampler Location: 54  
 Date Collected: 5/22/2008 01:30:37  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090124\_10X

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: 2805090124\_10X

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	360353.6	89.0 %	0.37			0.41%
Yr	319839.8	99.0 %	1.48			1.49%
Cr†	116067.2	2.83 mg/L	0.016	28.3 mg/L	0.16	0.55%
M†	4581.2	0.00976 mg/L	0.000063	0.0976 mg/L	0.00063	0.64%
Na†	662017.3	250 mg/L	1.8	2500 mg/L	17.7	0.71%

Sequence No.: 182  
 Sample ID: 2805090132\_10X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 10X

Autosampler Location: 55  
 Date Collected: 5/22/2008 01:33:38  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090132\_10X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090132\_10X

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	363568.5	89.8	%	0.04			0.05%
Yr	316726.8	98.0	%	0.95			0.97%
Cr†	104506.5	2.54	mg/L	0.008	25.4 mg/L	0.08	0.31%
Mn†	272.6	0.00058	mg/L	0.000035	0.00580 mg/L	0.000348	6.00%
Nat	615686.7	232	mg/L	1.4	2320 mg/L	14.1	0.61%

=====  
 Sequence No.: 183  
 Sample ID: 2805090133\_10X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 10X

=====  
 Autosampler Location: 56  
 Date Collected: 5/22/2008 01:36:39  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

-----  
 Nebulizer Parameters: 2805090133\_10X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

-----  
 Mean Data: 2805090133\_10X

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	362029.8	89.4 %	0.23			0.26%
Yr	319267.0	98.8 %	0.92			0.93%
Crt	121519.3	2.96 mg/L	0.030	29.6 mg/L	0.30	1.03%
Mnf	559.2	0.00119 mg/L	0.000007	0.0119 mg/L	0.00007	0.58%
Nat	706476.5	266 mg/L	0.6	2660 mg/L	5.6	0.21%

Sequence No.: 187  
 Sample ID: 2805090134\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 57  
 Date Collected: 5/22/2008 01:47:39  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090134\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090134\_5X

Analyte	Mean Corrected		Calib.		Sample		RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.		
Sca	370040.7	91.4 %	0.32				0.35%
Yr	325674.0	101 %	0.5				0.53%
Crt	22755.1	0.554 mg/L	0.0033		2.77 mg/L	0.016	0.59%
Mnt	2590.7	0.00552 mg/L	0.000035		0.0276 mg/L	0.00017	0.63%
Nat	630890.6	238 mg/L	1.9		1190 mg/L	9.7	0.82%

Sequence No.: 188  
 Sample ID: 2805090135\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 58  
 Date Collected: 5/22/2008 01:50:39  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090135\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090135\_5X

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	368360.2	91.0 %		0.29			0.32%
Yr	321575.8	99.5 %		0.58			0.58%
Crf	32208.4	0.784 mg/L		0.0040	3.92 mg/L	0.020	0.51%
Mnt	1022.2	0.00218 mg/L		0.000050	0.0109 mg/L	0.00025	2.31%
Nat	701426.8	265 mg/L		1.2	1320 mg/L	5.9	0.45%

Sequence No.: 189  
 Sample ID: 2805090136\_10X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 10X

Autosampler Location: 59  
 Date Collected: 5/22/2008 01:53:40  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090136\_10X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090136\_10X

Analyte	Mean Corrected		Calib.		Sample		Std.Dev.	RSD
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units		
Sca	369944.2	91.4	%	0.48				0.52%
Yr	320904.7	99.3	%	1.60				1.61%
Crt	81065.3	1.97	mg/L	0.015	19.7	mg/L	0.15	0.76%
Mnt	298.0	0.00063	mg/L	0.000041	0.00635	mg/L	0.000408	6.43%
Nat	482817.7	182	mg/L	0.3	1820	mg/L	2.8	0.15%

Sequence No.: 190  
 Sample ID: 2805090137\_10X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 10X

Autosampler Location: 60  
 Date Collected: 5/22/2008 01:56:42  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090137\_10X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090137\_10X

Analyte	Mean Corrected		Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
	Intensity						
Sca	359700.7	88.9 %		0.37			0.42%
Yr	317894.2	98.4 %		0.54			0.55%
Crt	144507.0	3.52 mg/L		0.021	35.2 mg/L	0.21	0.60%
Mnt	11310.2	0.0241 mg/L		0.00003	0.241 mg/L	0.0003	0.12%
Nat	761384.0	287 mg/L		0.0	2870 mg/L	0.3	0.01%

Sequence No.: 191  
 Sample ID: 2805090138\_5X  
 Analyst:  
 Initial Sample Wt:  
 Dilution: 5X

Autosampler Location: 61  
 Date Collected: 5/22/2008 01:59:43  
 Data Type: Original  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 2805090138\_5X  
 Analyte Back Pressure Flow  
 All 213.0 kPa 0.65 L/min

Mean Data: 2805090138\_5X

Analyte	Mean Corrected		Calib.		Sample		Std.Dev.	RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	RSD		
Sca	376197.7	92.9 %	0.34				0.37%	
Yr	331187.3	103 %	0.7				0.68%	
Crt	8874.3	0.216 mg/L	0.0020	1.08 mg/L	0.010	0.94%		
Mnt	18488.9	0.0394 mg/L	0.00025	0.197 mg/L	0.0013	0.64%		
Nat	339879.7	128 mg/L	0.4	641 mg/L	1.8	0.27%		



**Standard  
Preparation  
Worksheet  
&  
Certificate of  
Analysis**

Reagent Lot #  
 HNO3 R# 100450 HCL R# 100446  
 IS = Yttrium(ME0709008)0.75mL + Scandium (ME0710007)0.5mL to 1000mL w/ 2% HNO3

Standards	Lot #	Exp. Date	Dilution
Calibration	ME0712001	(12/01/08)	1:10 ME0801001
(Prepare daily)	ME0712001	(12/01/08)	1:10
CCV/MCV/ECV	ME0710008	(04/17/09)	CCV/ECV 1:20 ME0801002
(Prepare daily)			MCV 1:40 ME0801003
Spike/LCS	ME0709009	(03/11/09)	1:100 ME0801005
(Prepare daily)	ME0801004	(07/11/08)	1:100
	ME0803001	(08/13/08)	1:50
MRL	ME0801007	(07/11/08)	1:100 ME0801008
(Prepare daily)			
ICSA	ME0712003	(06/01/08)	
ICSAB	ME0712004	(06/01/08)	
QCS	ME0610005	(04/10/08)	
1ppm Check	ME0801010	(07/11/08)	
Linearity	ME0805001	(11/05/08)	

Method Sr/Ti/Sn/SiO2

Calibration	ME0801012	(07/11/08)	
CCV/ECV	ME0803011	(09/30/08)	
QCS	ME0801012	(07/11/08)	
Spike/LCS	ME0803012	(09/30/08)	1:100
(Prepare daily)			
MRL	ME0801014	(07/11/08)	1:100
(Prepare daily)			
Method Li			
Std/ICV/MRL	ME0801009	(07/11/08)	1:1000, 200, 40, 10
(Prepare daily)			
QCS	ME0801011	(07/11/08)	1:10
(Prepare daily)			
LCS/Spike	ME0801011	(07/11/08)	1:50
(Prepare daily)			
ccv	ME0801011	(07/11/08)	1:40
(Prepare daily)			

From May 2005: the calibration std for ICP should be ME0505010,011 not ME0408010

dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.

Initial:  
Date:

WJ  
11/1/08

## METALS STANDARD DOCUMENTATION

**Standard:** ICP Calibration STD  
**Date Received/Prepped:** Prep Daily  
**Date Expired:** 12/1/2008  
**Manufacturer:** MWH-wbh  
**Matrix:** 2% HNO3 + 5% HCl  
**Amount:**

**ME #:** 0801001  
**By:** wbh  
**Lot #:**  
**Certificate:** NO  
**NIST SRM:**  
**Storage:** Room Temp

Component	Comment	Conc. Unit:
Mo	1:10 ME0712002	10 ug/ml
Sb		10 ug/ml
Sn		10 ug/ml
Ti		10 ug/ml
B		5 ug/ml
Ca	1:10 ME0712001	100 ug/ml
K		100 ug/ml
Mg		100 ug/ml
Na		100 ug/ml
Al		10 ug/ml
As		10 ug/ml
Ba		10 ug/ml
Co		10 ug/ml
Cr		10 ug/ml
Cu		10 ug/ml
Fe		10 ug/ml
Mn		10 ug/ml
Ni		10 ug/ml
Pb		10 ug/ml
Se		10 ug/ml
Tl		10 ug/ml
V		10 ug/ml
Zn		10 ug/ml
Cd		5 ug/ml
Be		4 ug/ml
SR		3 ug/ml
Ag		2 ug/ml

Initial: STE  
Date: 12/1/07

## METALS STANDARD DOCUMENTATION

**Standard:** ICPCalibration Stock Std #1  
**Date Received/Prepped:** 12/1/2007  
**Date Expired:** 12/1/2008  
**Manufacturer:** Inorganic Ventures  
**Matrix:** 5% Nitric Acid  
**Amount:** 500 mL

**ME #:** 0712001  
**By:** STE  
**Lot #:** A2-MEB243151  
**Certificate:** Y  
**NIST SRM:** Varies  
**Storage:** Room Temp

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		100 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		50 ug/ml
Be		40 ug/ml
SR		30 ug/ml
Ag		20 ug/ml

**1.0 INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

**2.0 DESCRIPTION OF CRM** Custom Solution  
 Catalog No.: MWH-ICAP-CAL-1  
 Lot Number: **A2-MEB243151**  
 Matrix: 5% HNO<sub>3</sub>(abs)

ME 0712001

1,000.00 µg/mL ea:

Ca, K, Mg, Na,

100.00 µg/mL ea:

Al, As, Ba, Co, Cr<sub>3</sub>, Cu, Fe, Mn, Ni, Pb, Se, Ti, V, Zn,

50.00 µg/mL ea:

Cd,

40.00 µg/mL ea:

Be,

30.00 µg/mL ea:

Sr,

20.00 µg/mL ea:

Ag

### 3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.4 ± 0.3 µg/mL	Arsenic, As	100.1 ± 0.3 µg/mL	Barium, Ba	99.6 ± 0.4 µg/mL
Beryllium, Be	40.04 ± 0.08 µg/mL	Cadmium, Cd	50.15 ± 0.12 µg/mL	Calcium, Ca	1,000 ± 2 µg/mL
Chromium+3, Cr <sub>3</sub>	100.0 ± 0.2 µg/mL	Cobalt, Co	99.9 ± 0.2 µg/mL	Copper, Cu	100.0 ± 0.2 µg/mL
Iron, Fe	99.6 ± 0.1 µg/mL	Lead, Pb	100.0 ± 0.3 µg/mL	Magnesium, Mg	1,000 ± 4 µg/mL
Manganese, Mn	100.0 ± 0.3 µg/mL	Nickel, Ni	100.0 ± 0.3 µg/mL	Potassium, K	1,001 ± 5 µg/mL
Selenium, Se	100.0 ± 0.2 µg/mL	Silver, Ag	20.04 ± 0.02 µg/mL	Sodium, Na	1,002 ± 5 µg/mL
Strontium, Sr	30.04 ± 0.18 µg/mL	Thallium, Tl	99.7 ± 0.1 µg/mL	Vanadium, V	100.0 ± 0.3 µg/mL
Zinc, Zn	100.0 ± 0.3 µg/mL				

**Certified Density:** 1.055 g/mL (measured at 22° C)

Initial:  
Date:

STE  
12/15/07

## METALS STANDARD DOCUMENTATION

**Standard:** ICPCalibration Stock Std #2  
**Date Received/Prepped:** 12/1/2007  
**Date Expired:** 12/1/2008  
**Manufacturer:** Inorganic Ventures  
**Matrix:** 5% Nitric Acid + Trace HF  
**Amount:** 500 mL

**ME #:** 0712002  
**By:** STE  
**Lot #:** A2-MEB243152  
**Certificate:** Y  
**NIST SRM:** Varies  
**Storage:** Room Temp

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml
Sb		100 ug/ml
Sn		100 ug/ml
Ti		100 ug/ml
B		50 ug/ml
Mo		100 ug/ml

**1.0 INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

**2.0 DESCRIPTION OF CRM** Custom Solution  
 Catalog No.: MWH-ICAP-CAL-2  
 Lot Number: **A2-MEB243152**  
 Matrix: tr. HF, 5% HNO<sub>3</sub>(abs)

**ME 0712 002**

100.00 µg/mL ea:  
 Mo, Sb, Sn, Ti,  
 50.00 µg/mL ea:  
 B

**3.0 CERTIFIED VALUES AND UNCERTAINTIES**

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	49.85 ± 0.20 µg/mL	Molybdenum, Mo	100.2 ± 0.3 µg/mL
Tin, Sn	100.2 ± 0.3 µg/mL	Titanium, Ti	100.4 ± 0.2 µg/mL		

**Certified Density:** 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

( $\bar{x}$ ) = mean  
 $x_i$  = individual results  
 $n$  = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

$\sum s_i$  = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

**4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS**

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.



Initial:

STE

Date:

10/19/07

### METALS STANDARD DOCUMENTATION

**Standard:** ICP CCV/MCV/QCS Stock Standard  
**Date Received/Prepped:** 10/19/2007  
**Date Expired:** 4/17/2009  
**Manufacturer:** CPI  
**Matrix:** 5% HNO3 = tr HF  
**Amount:** 100 mL x 10

**ME #:** 0710008  
**By:** STE  
**Lot #:** 07J154  
**Certificate:** Y  
**NIST SRM:** Various  
**Storage:** Room Temp

Component	Comment	Conc. Unit:
Ag	P/N 4400-061003RH01	20 ppm
Al		100 ppm
As		100 ppm
B		50 ppm
Ba		100 ppm
bE		40 ppm
Ca		1000 ppm
Cd		50 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		1000 ppm
Mg		1000 ppm
Mn		100 ppm
Mo		100 ppm
Na		1000 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		20 ppm
Sn		20 ppm
Ti		20 ppm



**USA**  
 5580 Skylane Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 96  
 The Netherlands www.cpiinternational.com

*Innovative Solutions  
 in Analytical Science and  
 Technology*

Expiry: 4/17/2009

# Certificate of Analysis

**Part Number:** 4400-061003RH01  
**Lot Number:** 07J154  
**Shelf Life:** 18 months

ME 0710008

07/10/09  
 STE

MWH  
 Custom Multi  
 5% HNO<sub>3</sub> + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000 µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at [www.cpiinternational.com](http://www.cpiinternational.com).

105

Initial: W3y  
 Date: 4/11/09

## METALS STANDARD DOCUMENTATION

<b>Standard:</b>	ICP MCV Working Standard	<b>ME #:</b> 0801003
<b>Date Received/Prepped:</b>	Daily	<b>By:</b> Wbh
<b>Date Expired:</b>	4/17/2009	<b>Lot #:</b>
<b>Manufacturer:</b>	CPI	<b>Certificate:</b> Y
<b>Matrix:</b>	2% HNO <sub>3</sub> + 5% HCL	<b>NIST SRM:</b> Various
<b>Amount:</b>	100 mL	<b>Storage:</b> Room Temp

Component	Comment	Conc. Unit:
Ag	5ml ME0710008 / 100mL	0.5 ppm
Al		2.5 ppm
As		2.5 ppm
B		1.25 ppm
Ba		2.5 ppm
bE		1 ppm
Ca		25 ppm
Cd		1 ppm
Co		2.5 ppm
Cr		2.5 ppm
Cu		2.5 ppm
Fe		2.5 ppm
K		25 ppm
Mg		25 ppm
Mn		2.5 ppm
Mo		2.5 ppm
Na		25 ppm
Ni		2.5 ppm
Pb		2.5 ppm
Sb		2.5 ppm
Se		2.5 ppm
Tl		2.5 ppm
V		2.5 ppm
Zn		2.5 ppm
Sr		0.5 ppm
Sn		0.5 ppm
Ti		0.5 ppm

Initial:

Date:

WJH  
4/17/09

### METALS STANDARD DOCUMENTATION

**Standard:** ICP CCV/ECV Working Standard  
**Date Received/Prepped:** Daily  
**Date Expired:** 4/17/2009  
**Manufacturer:** CPI  
**Matrix:** 2% HNO<sub>3</sub> + 5% HCL  
**Amount:** 100 mL

**ME #:** 0801002  
**By:** Wbh  
**Lot #:**  
**Certificate:** Y  
**NIST SRM:** Various  
**Storage:** Room Temp

Component	Comment	Conc. Unit:
Ag	5ml ME0710008 / 100mL	1 ppm
Al		5 ppm
As		5 ppm
B		2.5 ppm
Ba		5 ppm
bE		2 ppm
Ca		50 ppm
Cd		2 ppm
Co		5 ppm
Cr		5 ppm
Cu		5 ppm
Fe		5 ppm
K		50 ppm
Mg		50 ppm
Mn		5 ppm
Mo		5 ppm
Na		50 ppm
Ni		5 ppm
Pb		5 ppm
Sb		5 ppm
Se		5 ppm
Tl		5 ppm
V		5 ppm
Zn		5 ppm
Sr		1 ppm
Sn		1 ppm
Ti		1 ppm

Initial:

WBH

Date:

7/11/08

### METALS STANDARD DOCUMENTATION

**Standard:** Spike and LCS std for ICP  
**Date Received/Prepped:** prep daily  
**Date Expired:** 7/11/2008  
**Manufacturer:** MWH-wbh  
**Matrix:** 2% HNO3 + 5% HCl  
**Amount:**

**ME #:** 0801005  
**By:** WBH  
**Lot #:**  
**Certificate:** N  
**NIST SRM:**  
**Storage:** Room Temp

Component	Comment	Conc. Unit:
CA	1:200 ME0709007	50 ppm
K		20 ppm
MG		20 ppm
NA		50 ppm
Iron	1:100 ME0709009	5 mg/L
Aluminum		2 mg/L
Barium		1 mg/L
Cobalt		1 mg/L
Chromium		1 mg/L
Copper		1 mg/L
Molybdenum		1 mg/L
Strontium		1 mg/L
Titanium		1 mg/L
Vanadium		1 mg/L
Zinc		1 mg/L
Tin		1 mg/L
Silver		0.5 mg/L
Boron		0.5 mg/L
Manganese		0.5 mg/L
Nickel		0.5 mg/L
Antimony		0.5 mg/L
Arsenic		0.2 mg/L
Cadmium		0.2 mg/L
Lead		0.2 mg/L
Selenium		0.2 mg/L
Thallium		0.2 mg/L
Uranium		0.2 mg/L
Beryllium		0.05 mg/L
AS	1:100 ME0801004	0.8 mg/L
PB		0.8 mg/L
SE		0.8 mg/L
TL		0.8 mg/L

Initial:

Wbh

Date:

7/11/08

## METALS STANDARD DOCUMENTATION

Standard: ICP Spike Solution  
Date Received/Prepped: 1/11/2008  
Date Expired: 7/11/2008  
Manufacturer: MWH-wbh  
Matrix: 2% HNO<sub>3</sub>  
Amount: 100mL

ME #: 0801004  
By: Wbh  
Lot #: VARIOUS  
Certificate:  
NIST SRM:  
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	8.0mL ME07090023/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

Initial: wbh  
Date: 3/12/08

## METALS STANDARD DOCUMENTATION

**Standard:** ICP LCS/SPIKE Solution  
**Date Received/Prepped:** 3/12/2008  
**Date Expired:** 8/12/2008  
**Manufacturer:** MWH-wbh  
**Matrix:** 2% HNO<sub>3</sub> + 5% HCl  
**Amount:** 100mL

**ME #:** 0803001  
**By:** wbh  
**Lot #:**  
**Certificate:** NO  
**NIST SRM:**  
**Storage:** Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	1:4 ME0702002	5000 ug/ml
K	1:10 ME0702005	1000 ug/ml
Mg	1:10 ME0702004	1000 ug/ml
Na	1:4 ME0702003	5000 ug/ml

Initial: STE  
Date: 9/13/07

## METALS STANDARD DOCUMENTATION

**Standard:** ICP CCV/MCV Stock Standard  
**Date Received/Prepped:** 9/13/2007  
**Date Expired:** 3/11/2009  
**Manufacturer:** CPI International  
**Matrix:** 5% HNO3 AND 0.1% HF  
**Amount:** 100 mL

**ME #:** 0709009  
**By:** STE  
**Lot #:** 071040  
**Certificate:**  
**NIST SRM:**  
**Storage:** Room Temp.

Component	Comment	Conc. Unit:
Fe	P/N 4400-050314RH01	500
Al	(10 bottles)	200
Ba		100
Co		100
Cr		100
Cu		100
Mo		100
Sr		100
Ti		100
V		100
Zn		100
Ag		50
Ba		50
Mn		50
Ni		50
Sb		50
As		50
Cd		20
Pb		20
Se		20
Tl		20
Sn		20
Be		100
U		5
		20





**USA**

5580 Skylane Boulevard 707.525.5788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.545.7901

**EUROPE**

P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

*Innovative Solutions  
in Analytical Science and  
Technology*

Expiry: 3/11/2009

# Certificate of Analysis

**Part Number:** 4400-050314RH01  
**Lot Number:** 071040  
**Shelf Life:** 18 months

ME C7 09 009

MWH Labs  
5% HNO3 + 0.1% HF  
#REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

W 37  
2/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** Potassium 10000ppm Stock Std  
**Date Received/Prepped:** 2/20/2007  
**Date Expired:** 8/16/2008  
**Manufacturer:** CPI  
**Matrix:** 1% HNO3  
**Amount:** 250 mL

**ME #:** 0702005  
**By:** WBH  
**Lot #:** 07B056  
**Certificate:** Y  
**NIST SRM:** 3141  
Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
K	P/N 4400-10M411	10000 ppm



Initial:

Date:

WBH  
2/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** Magnesium 10000ppm Stock Std      **ME #:** 0702004  
**Date Received/Prepped:** 2/20/2007      **By:** WBH  
**Date Expired:** 8/16/2008      **Lot #:** 07B058  
**Manufacturer:** CPI      **Certificate:** Y  
**Matrix:** 4% HNO3      **NIST SRM:** 3131  
**Amount:** 250 mL      Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Mg	P/N 4400-10M311	10000 ppm



**USA**  
 5580 Skylane Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com

Innovative Solutions  
 in Analytical Science and  
 Technology

## CERTIFICATE OF ANALYSIS

470702004

**P/N 4400-10M311**  
**P/N S4400-10M311**  
 Single-Element Magnesium Standard  
 Mg in 4% HNO<sub>3</sub>  
 10,000 ± 30 µg/mL

Lot # 07B058

Material Source: Magnesium Metal  
 Source Purity: 99.99%  
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	28	0.1	Cu	1.6	0.1	Pb	7.7	0.7	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	0.28	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.28	0.1	Eu	ND	0.1	Mg	X	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	0.23	0.1	Mn	19.8	1	Rb	ND	0.1
Bi	ND	0.1	Ga	0.18	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	1.1	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	1	0.1	Si	64	20
Ca	ND	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.19	0.1
Ce	2.1	0.1	I	1	0.2	Os	ND	0.1	Na	7.2	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	0.19	0.1
Cr	ND	1	Fe	80	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	0.76	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: W3y  
Date: 2/20/07

## METALS STANDARD DOCUMENTATION

<b>Standard:</b>	Sodium 10000ppm Stock Std	<b>ME #:</b> 0702003
<b>Date Received/Prepped:</b>	2/20/2007	<b>By:</b> WBH
<b>Date Expired:</b>	8/16/2008	<b>Lot #:</b> 07B057
<b>Manufacturer:</b>	CPI	<b>Certificate:</b> Y
<b>Matrix:</b>	1% HNO3	<b>NIST SRM:</b> 3152a
<b>Amount:</b>	250 mL	Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



**USA**  
 5580 Skyline Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 GS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiiinternational.com

*Innovative Solutions  
 in Analytical Science and  
 Technology*

## CERTIFICATE OF ANALYSIS

*M70702003*

**P/N 4400-10M521**  
**P/N S4400-10M521**  
 Single-Element Sodium Standard  
 Na in 1% HNO<sub>3</sub>  
 10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO<sub>3</sub>)  
 Source Purity: 99.99%  
 Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 1.5	0.1	Cu 0.45	0.1	Pb ND	0.1	K ND	70	Ti ND	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.13	0.1	Eu ND	0.1	Mg 2.3	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Tl ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	8	Yb ND	0.1
Ca 120	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na X	1	Zn 2.9	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial: WBH  
Date: 2/20/07

### METALS STANDARD DOCUMENTATION

**Standard:** Calcium 10000ppm Stock Std      **ME #:** 0702002  
**Date Received/Prepped:** 2/20/2007      **By:** WBH  
**Date Expired:** 8/16/2008      **Lot #:** 07B065  
**Manufacturer:** CPI      **Certificate:** Y  
**Matrix:** 4% HNO3      **NIST SRM:** 3109a  
**Amount:** 250 mL      Room temp. storage

Component	Comment	Conc. Unit:
Ca	P/N 4400-10M91	10000 ppm





**USA**  
 5580 Skyline Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

AUG 16 08

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com

*Innovative Solutions  
 in Analytical Science and  
 Technology*

## CERTIFICATE OF ANALYSIS

**P/N 4400-10M91**  
**P/N S4400-10M91**

180702602

Single-Element Calcium Standard  
 Ca in 4% HNO<sub>3</sub>  
 10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO<sub>3</sub>)  
 Source Purity: 99.997%  
 Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.



	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	7	0.1	Cu	1.7	0.1	Pb	0.23	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	1.5	0.1	Eu	ND	0.1	Mg	38	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	1.5	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	3	0.1	Si	47	8
Ca	X	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	0.27	0.2	Os	ND	0.1	Na	11.6	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	55	0.1
Cr	ND	1	Fe	INT	30	P	ND	10	Ta	ND	0.1
Co	INT	0.1	La	0.41	0.1	Pt	ND	0.1	Te	ND	0.1

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

120



Initial:

Date:

*WBH*

*2/20/07*

## METALS STANDARD DOCUMENTATION

**Standard:** Thallium 1000ppm Stock Std  
**Date Received/Prepped:** 2/20/2007  
**Date Expired:** 8/16/2008  
**Manufacturer:** CPI  
**Matrix:** 2% HNO3  
**Amount:** 100 mL

**ME #:** 0702006  
**By:** WBH  
**Lot #:** 06H213  
**Certificate:** Y  
**NIST SRM:** 3158  
Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Tl	P/N S4400-1000581	1000 ppm

Aug 16 08



USA  
5580 Skylane Boulevard 707.525.6788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.545.7901

EUROPE  
P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

Innovative Solutions  
in Analytical Science and  
Technology

M70702006

# CERTIFICATE OF ANALYSIS

**P/N 4400-1000581**

**P/N S4400-1000581**

Single Element Thallium Standard

Tl in 2% HNO<sub>3</sub>

1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal  
Source Purity: 99.999%  
Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>Ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 13.3	0.1	Cu 9.3	0.1	Pb 41	0.1	K ND	70	Tl X	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.37	0.1	Eu ND	0.1	Mg 1.7	0.2	Rh ND	0.1	Sn ND	0.1
Be 0.67	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti 0.45	0.1
Bi 0.12	0.1	Ga ND	0.1	Hg 0.16	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.21	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd 1.6	0.1	Hf ND	0.1	Ni 1.1	0.1	Si 46	8	Yb ND	0.1
Ca 51	7	Ho ND	0.1	Nb ND	0.1	Ag 0.3	0.1	Y ND	0.1
Ce ND	0.1	I 0.4	0.2	Os ND	0.1	Na 33	1	Zn 14.7	2
Cs 0.24	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 20	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.

METALS STANDARD DOCUMENTATION

123

Initial: WJH  
Date: 3/5/07



Standard:	Selenium Stock Standard	ME #:	0703001
Date Received/Prepped:	3/5/2007	By:	wjh
Date Expired:	8/22/2008	Lot #:	6.00E+228
Manufacturer:	CPI	Certificate:	Y
Matrix:	2% HNO3	NIST SRM:	3148
Amount:	100 mL	Storage:	Room Temp
Component	Comment	Conc. Unit:	1000 ppm
Se	P/N # S4400-1000491		



Initial:

*WBH*

Date:

*4/16/07*

### METALS STANDARD DOCUMENTATION

**Standard:** Lead Stock Standard  
**Date Received/Prepped:** 4/16/2007  
**Date Expired:** 10/11/2008  
**Manufacturer:** CPI  
**Matrix:** 2% HNO<sub>3</sub>  
**Amount:** 100 mL

**ME #:** 0704013  
**By:** WBH  
**Lot #:** 07A097  
**Certificate:** Y  
**NIST SRM:** 3128  
Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Pb	P/N S4400-1000281	1000 ppm



Innovative Solutions  
in Analytical Science and  
Technology

**USA**  
5580 Skylane Boulevard 707.525.5788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.546.7901

**EUROPE**  
P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

## CERTIFICATE OF ANALYSIS

**P/N S4400-1000281**  
**P/N 4400-1000281**  
Single-Element Lead Standard  
Pb in 2% HNO<sub>3</sub>  
1000 ± 3 µg/mL

1270704013

Lot # 07A097

Material Source: Lead Metal  
Source Purity: 99.995 %  
Specific Gravity: 1.009 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3128. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	10.3	0.1	Cu	58	0.1	Pb	X	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	2	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.22	0.1	Eu	ND	0.1	Mg	1.4	0.2	Rh	IN	0.1
Be	0.58	0.1	Gd	ND	0.1	Mn	3.8	1	Rb	ND	0.1
Bi	0.7	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.17	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.9	0.1	Si	31	8
Ca	25	7	Ho	ND	0.1	Nb	ND	0.1	Ag	6.1	0.1
Ce	ND	0.1	I	0.1	0.2	Os	ND	0.1	Na	3.5	1
Cs	0.26	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1
Cr	ND	1	Fe	ND	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

ME0709023

Initial:

STE

Date:

9/24/07

## METALS STANDARD DOCUMENTATION

**Standard:** As Stock Standard  
**Date Received/Prepped:** 9/24/2007  
**Date Expired:** 10/1/2008  
**Manufacturer:** Inorganic Ventures  
**Matrix:** 1.4% HNO3  
**Amount:** 100 mL X2

**ME #:** 0709023  
**By:** STE  
**Lot #:** A2-AS02035  
**Certificate:** Y  
**NIST SRM:**  
**Storage:** Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
As	PN: CGAS1-1	1000 ug/ml



**1.0 INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

**2.0 DESCRIPTION OF CRM**      **1000 µg/mL Arsenic in 1.4% (abs) HNO<sub>3</sub>**

Catalog Number:            CGAS1-1, CGAS1-2, and CGAS1-5  
 Lot Number:                 **A2-AS02035**  
 Starting Material:         As Polycrystalline lump  
 Starting Material Purity (%):    99.998288  
 Starting Material Lot No:    23444  
 Matrix:                      1.4% (abs) HNO<sub>3</sub>

**ME 0709023**

**3.0 CERTIFIED VALUES AND UNCERTAINTIES**

**Certified Concentration:**    1000 ± 6 µg/mL

**Certified Density:**            1.010 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

( $\bar{x}$ ) = mean  
 $x_i$  = individual results  
 $n$  = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

$\sum s_i$  = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

**4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS**

- "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)
- This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

**4.1 Assay Method #1**      **1000 ± 6 µg/mL**  
 ICP Assay NIST SRM 3103a Lot Number: 010713

**Assay Method #2**      **1001 ± 5 µg/mL**  
 Gravimetric NIST SRM Lot Number: See Sec. 4.2

Date: W37  
7/11/08

## METALS STANDARD DOCUMENTATION

**Standard:** ICP MRL Working Stock Solution **ME #:** 0801008  
**Date Received/Prepped:** Daily **By:** Wbh  
**Date Expired:** 7/11/2008 **Lot #:**  
**Manufacturer:** MWH-wbh **Certificate:** Y  
**5% HNO3** 2% HNO3 + 5% Hcl **NIST SRM:**  
**Amount:** 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	0.1mL ME0801007 / 10mL	0.05 ppm
Sb		0.05 ppm
As		0.1 ppm
Ba		0.02 ppm
Be		0.001 ppm
Ba		0.05 ppm
Cd		0.005 ppm
Ca		1 ppm
Cr		0.01 ppm
Co		0.05 ppm
Cu		0.01 ppm
Fe		0.02 ppm
Pb		0.02 ppm
Li		0.1 ppm
Mg		0.002 ppm
Mn		0.02 ppm
Mo		0.02 ppm
Ni		1 ppm
K		0.1 ppm
Se		0.01 ppm
Ag		1 ppm
Na		0.02 ppm
Tl		0.002 ppm
V		0.1 ppm
Zn		0.001 ppm
Ti		0.02 ppm
Sr		0.01 ppm
Sn		0.2 ppm

Date: WZ  
1/11/08

### METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working Stock Solu ME #: 0801007  
Date Received/Prepped: 1/11/2008 By: Wbh  
Date Expired: 7/11/2008 Lot #:  
Manufacturer: MWH-wbh Certificate: Y  
Matrix: 5% HNO3 NIST SRM:  
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	10mL ME0709020 / 100mL	5 ppm
Sb		5 ppm
As		10 ppm
Ba		2 ppm
Be		0.1 ppm
Ba		5 ppm
Cd		0.5 ppm
Ca		100 ppm
Cr		1 ppm
Co		5 ppm
Cu		1 ppm
Fe		2 ppm
Pb		2 ppm
Li		10 ppm
Mg		0.2 ppm
Mn		2 ppm
Mo		2 ppm
Ni		100 ppm
K		10 ppm
Se		1 ppm
Ag		100 ppm
Na		2 ppm
Tl		0.2 ppm
V		10 ppm
Zn		0.1 ppm
Ti		2 ppm
Sr		1 ppm
Sn		20 ppm

Initial: SRE  
Date: 9/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** ICP MRL Stock Standard **ME #:** 0709020  
**Date Received/Prepped:** 9/20/2007 **By:**  
**Date Expired:** 9/18/2008 **Lot #:** 061162  
**Manufacturer:** CPI **Certificate:** Y  
**Matrix:** 2% HNO<sub>3</sub> + tr HF **NIST SRM:**  
**Amount:** 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-060915RHO1	50 ppm
Sb		50 ppm
As		100 ppm
Ba		20 ppm
Be		1 ppm
Ba		50 ppm
Cd		5 ppm
Ca		1000 ppm
Cr		10 ppm
Co		50 ppm
Cu		10 ppm
Fe		20 ppm
Pb		20 ppm
Li		1 ppm
Mg		100 ppm
Mn		2 ppm
Mo		20 ppm
Ni		20 ppm
K		1000 ppm
Se		100 ppm
Ag		10 ppm
Na		1000 ppm
Tl		100 ppm
V		2 ppm
Zn		20 ppm
Ti		20 ppm
Sr		10 ppm
Sn		200 ppm



**USA**  
 5580 Skylane Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com

*Innovative Solutions  
 in Analytical Science and  
 Technology*

ME0709020

Expiry 9/18/2008

# Certificate of Analysis

**Part Number: 4400-060915RH01**  
**Lot Number: 061162**  
**Shelf Life: 12 months**

MWH  
 Custom Standard  
 2% HNO<sub>3</sub> + tr HF

Concentrations in ug/mL ± 0.5%

Al	50	Pb	20	Zn	20
Si	50	Li	1	Ti	20
As	100	Mg	100	Sr	10
Ba	20	Mn	2	Sn	200
Be	1	Mo	20		
B	50	Ni	20		
Cd	5	K	1000		
Ca	1000	Se	100		
Cr	10	Ag	10		
Co	50	Na	1000		
Cu	10	TL	100		
Fe	20	V	2		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at [www.cpiinternational.com](http://www.cpiinternational.com).

# METALS STANDARD DOCUMENTATION

**Standard:** INTERFERENCE CHECK STANDAF **ME #:** 0806001  
**Date Received/Prepped:** 6/1/2008 **By:** CSK  
**Date Expired:** 12/1/2008 **Lot #:** various  
**Manufacturer:** MWH-CSK **Certificate:**  
**Matrix:** 5% HNO3 **NIST SRM:**  
**Amount:** 500ML **Storage:** Room Temp

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
AL	P/N 4400-INTA1-500	250 ppm
CA	25ML IN 500ML	250 ppm
FE		100 ppm
MG		250 ppm
AG		
BA		
BE		
CD		
CO		
CR		
CU		
MN		
NI		
PB		
V		
ZN		

Initial:

CSK

Date:

6-1-08

### METALS STANDARD DOCUMENTATION

**Standard:** INTERFERENCE CHECK STANDAF  
**Date Received/Prepped:** 6/1/2008  
**Date Expired:** 12/1/2008  
**Manufacturer:** MWH-CSK  
**Matrix:** 5% HNO3  
**Amount:** 500 mL


**ME #:** 0806002  
**By:** CSK  
**Lot #:** various  
**Certificate:**  
**NIST SRM:**  
**Storage:** Room Temp.

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500(25ML)	250 ppm
Ca	P/N 4400-INTB1-100(2.5ML)	250 ppm
Fe		100 ppm
Mg		250 ppm
AG		0.5 ppm
BA		0.25 ppm
BE		0.25 ppm
CD		0.5 ppm
Co		0.25 ppm
Cr3		0.25 ppm
Cu		0.5 ppm
Mn		0.25 ppm
Ni		0.25 ppm
Pb		0.25 ppm
V		0.25 ppm
Zn		0.5 ppm

Ti  
 V  
 Zn  
 Si

100 ppm  
 100 ppm  
 100 ppm  
 50 ppm

*Handwritten signature and date: 6/1/08*



**USA**  
 5580 Skylane Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com



*Innovative Solutions  
 in Analytical Science and  
 Technology*

ME 070500

## CERTIFICATE OF ANALYSIS

**P/N 4400-INTB1-100**  
 CLP Analytes B Solution  
 in 5% HNO<sub>3</sub>

Lot # 07c256

Material Source: Metals and Salts  
 Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag	100	Ba	50	Be	50	Cd	100
Co	50	Cr	50	Cu	50	Mn	50
Ni	100	Pb	100	V	50	Zn	100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial: STE  
Date: 8/27/07

## METALS STANDARD DOCUMENTATION

**Standard:** CLP Interferents A Solution **ME #:** 0708009  
**Date Received/Prepped:** 8/27/2007 **By:** STE  
**Date Expired:** 2/15/2009 **Lot #:** 07E175  
**Manufacturer:** CPI International **Certificate:**  
**Matrix:** 5% HNO3 **NIST SRM:**  
**Amount:** 500 mL **Storage:** Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Al	P/N 4400-INTA1-500	5000 ug/mL
Ca		5000 ug/mL
Fe		2000 ug/mL
Mg		5000 ug/mL

FEB 15 09



**USA**

5580 Skylane Boulevard 707.525.6788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.545.7901

**EUROPE**

P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

*Innovative Solutions  
in Analytical Science and  
Technology*

## **CERTIFICATE OF ANALYSIS**

**P/N 4400-INTA1-500**

CLP Interferents A Solution  
in 5% HNO<sub>3</sub>

ME 0705009

Lot # 07E175

Material Source: Metals and Salts  
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

137

**colitag™**  
The First On-Demand Water Test System

**PEAK PERFORMANCE**  
CERTIFIED REFERENCE MATERIALS

**Accu\*prep 7000™**  
Extraction Manifold

**MOD BLOCK**

**Vu\*Phase™**  
PE Blocks

Initial: Wbh  
 Date: 1/11/08

## METALS STANDARD DOCUMENTATION

Standard: ICP 1 PPM CHECK  
 Date Received/Prepped: 1/11/2008  
 Date Expired: 7/11/2008  
 Manufacturer: MWH-wbh  
 Matrix: 5% HNO<sub>3</sub>  
 Amount: 500 mL

ME #: 0801010  
 By: Wbh  
 Lot #: VARIOUS  
 Certificate:  
 NIST SRM:  
 Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	5mL ME0708011 /500mL	1 mg/L
Al		1 mg/L
B		1 mg/L
Ba		1 mg/L
K		10 mg/L
Na		1 mg/L
Si		1 mg/L
As	5mL ME0708012 /500mL	1 mg/L
Be		1 mg/L
Ca		1 mg/L
Cd		1 mg/L
Co		1 mg/L
Cr		1 mg/L
Cu		1 mg/L
Fe		1 mg/L
Li		1 mg/L
Mg		1 mg/L
Mn		1 mg/L
Mo		1 mg/L
Ni		1 mg/L
Pb		1 mg/L
Sb		1 mg/L
Se		1 mg/L
Sr		1 mg/L
Ti		1 mg/L
Tl		1 mg/L
V		1 mg/L
Zn		1 mg/L

Initial:

STE

Date:

8/27/07

## METALS STANDARD DOCUMENTATION

<b>Standard:</b>	QC Check Standard 21	<b>ME #:</b> 0708012
<b>Date Received/Prepped:</b>	8/27/2007	<b>By:</b> STE
<b>Date Expired:</b>	8/31/2008	<b>Lot #:</b> 074438H
<b>Manufacturer:</b>	Crescent Chemical Co. Inc.	<b>Certificate:</b>
<b>Matrix:</b>	5% HNO <sub>3</sub> /tr. F/tr Tartaric Acid	<b>NIST SRM:</b>
<b>Amount:</b>	100 mL	Room temp. storage

Component	Comment	Conc. Unit:
As	Catalog No: QC-021.1	100 ug/mL
Be		100 ug/mL
Ca		100 ug/mL
Cd		100 ug/mL
Co		100 ug/mL
Cu		100 ug/mL
Fe		100 ug/mL
Li		100 ug/mL
Mg		100 ug/mL
Mn		100 ug/mL
Mo		100 ug/mL
Ni		100 ug/mL
Pb		100 ug/mL
Sb		100 ug/mL
Se		100 ug/mL
Sr		100 ug/mL
Ti		100 ug/mL
Tl		100 ug/mL
V		100 ug/mL
Zn		100 ug/mL

ME 0708012

*Laboratory Report - Certificate of Analysis*

**Environmental Multielement Standard**

**QC Check Standard 21**

**CATALOG NO: QC-021.1**

**CONTENTS: See Below**

**MATRIX:** 5% HNO<sub>3</sub>/tr. F/tr. Tartaric Acid

**LOT NO.:** 074438H

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3102a, 3103a, 3105a, 3108, 3109a, 3112a, 3113, 3114, 3126a, 3128, 3129a, 3131a, 3132, 3134, 3136, 3149, 3153a, 3158, 3162a, 3165, and 3168a.

Concentrations are given in µg/mL unless noted otherwise.

As 100	Be 100	Ca 100	Cd 100	Co 100
Cr 100	Cu 100	Fe 100	Li 100	Mg 100
Mn 100	Mo 100	Ni 100	Pb 100	Sb 100
Se 100	Sr 100	Ti 100	Tl 100	V 100
Zn 100				

**Crescent Chemical Co. Inc.**

Julie M. MacIntosh  
QA Manager

**EXPIRES:** August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

*Crescent Chemical Co, Inc., 2 Oval Dive, Islandia, NY 11749  
(516) 348-0333 - Fax (516) 348-0913*

ME0705011

Initial:

STE

Date:

8/27/07

### METALS STANDARD DOCUMENTATION

**Standard:** QC Check Standard 7  
**Date Received/Prepped:** 8/27/2007  
**Date Expired:** 8/31/2007  
**Manufacturer:** Crescent Chemical Co. Inc.  
**Matrix:** 5% HNO3  
**Amount:** 100

ME #: 0705011

By: STE

Lot #: 074438I

Certificate:

NIST SRM:

Storage: Room Temp.

Component	Comment	Conc. Unit:
Ag	Catalog No: QC-007.1	100 ug/mL
Al		100 ug/mL
B		100 ug/mL
Ba		100 ug/mL
K		1000 +/- 5 ug/mL
Na		100 ug/mL
Si		50 ug/mL

ME 07 08 011

*Laboratory Report - Certificate of Analysis*

**Environmental Multielement Standard**

**QC Check Standard 7**

**CATALOG NO: QC-007.1**

**CONTENTS: See Below**

**MATRIX: 5% HNO<sub>3</sub>/tr. F<sup>-</sup>**

**LOT NO.: 074438I**

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3101a, 3104a, 3107, 3141a, 3150, 3151, and 3152a.

Concentrations are given in µg/mL unless noted otherwise.

Ag 100	Al 100	B 100	Ba 100	K 1,000±5
Na 100	Si 50.0			

**Crescent Chemical Co. Inc.**

Julie M. MacIntosh  
QA Manager

**EXPIRES:** August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

*Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749  
(516) 348-0333 - Fax (516) 348-0913*

Initial: CSK  
Date: 05/05/08

## METALS STANDARD DOCUMENTATION

**Standard:** ICP LINEARITY CHECK  
**Date Received/Prepped:** 5/5/2008  
**Date Expired:** 11/5/2008  
**Manufacturer:** MWH-STE  
**Matrix:** 5% HNO3  
**Amount:** 500ML

**ME #:** 0805001  
**By:** CSK  
**Lot #:** VARIOUS  
**Certificate:**  
**NIST SRM:**  
**Storage:** Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
CA	15.0ML ME0702002/500ML	300 ppm
K	15.0ML ME0702005/500ML	300 ppm
MG	10.0ML ME0702004/500ML	200 ppm
NA	15.0ML ME0702003/500ML	300 ppm
FE	5.0ML ME0701008/500ML	100 ppm



Initial:

Date:

W 37  
2/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** Potassium 10000ppm Stock Std  
**Date Received/Prepped:** 2/20/2007  
**Date Expired:** 8/16/2008  
**Manufacturer:** CPI  
**Matrix:** 1% HNO<sub>3</sub>  
**Amount:** 250 mL

**ME #:** 0702005  
**By:** WBH  
**Lot #:** 07B056  
**Certificate:** Y  
**NIST SRM:** 3141  
Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
K	P/N 4400-10M411	10000 ppm

AUG 16 08



**USA**  
 5580 Skylane Boulevard 707.525.5788  
 Santa Rosa, CA 95403 800.878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com

*Innovative Solutions  
 in Analytical Science and  
 Technology*

## CERTIFICATE OF ANALYSIS

M20702005

**P/N 4400-10M411**  
**P/N S4400-10M411**  
 Single-Element Potassium Standard  
 K in 1% HNO<sub>3</sub>  
 10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO<sub>3</sub>)  
 Source Purity: 99.999%  
 Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	0.39	0.1	Cu	0.16	0.1	Pb	ND	0.1	K	X	70
Sb	0.34	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.14	0.1	Eu	ND	0.1	Mg	2.6	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	0.93	1	Rb	9.5	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	20
Ca	82	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	19	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1
									Tl	ND	0.1
									Th	ND	0.1
									Tm	ND	0.1
									Sn	0.17	0.1
									Ti	ND	0.1
									W	ND	0.1
									U	ND	0.1
									V	ND	1
									Yb	ND	0.1
									Y	ND	0.1
									Zn	2.9	1
									Zr	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:


WBH  
2/19/07

## METALS STANDARD DOCUMENTATION

**Standard:** Magnesium 10000ppm Stock Std  
**Date Received/Prepped:** 2/20/2007  
**Date Expired:** 8/16/2008  
**Manufacturer:** CPI  
**Matrix:** 4% HNO3  
**Amount:** 250 mL

**ME #:** 0702004  
**By:** WBH  
**Lot #:** 07B058  
**Certificate:** Y  
**NIST SRM:** 3131  
Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Mg	P/N 4400-10M311	10000 ppm



**USA**  
 5580 Skylane Boulevard 707 525.5788  
 Santa Rosa, CA 95403 800 878.7654  
 www.cpiinternational.com Fax 707.545.7901

**EUROPE**  
 P.O. Box 2704 +31 20 638 05 97  
 1000 CS Amsterdam Fax +31 20 420 28 36  
 The Netherlands www.cpiinternational.com

Innovative Solutions  
 in Analytical Science and  
 Technology

## CERTIFICATE OF ANALYSIS

**P/N 4400-10M311**  
**P/N S4400-10M311**  
 Single-Element Magnesium Standard  
 Mg in 4% HNO<sub>3</sub>  
 10,000 ± 30 µg/mL

470702004

Lot # 07B058

Material Source: Magnesium Metal  
 Source Purity: 99.99%  
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	28	0.1	Cu	1.6	0.1	Pb	7.7	0.7	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	0.28	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.28	0.1	Eu	ND	0.1	Mg	X	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	0.23	0.1	Mn	19.8	1	Rb	ND	0.1
Bi	ND	0.1	Ga	0.18	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	1.1	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	1	0.1	Si	64	20
Ca	ND	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.19	0.1
Ce	2.1	0.1	I	1	0.2	Os	ND	0.1	Na	7.2	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	0.19	0.1
Cr	ND	1	Fe	80	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	0.76	0.1	Pt	ND	0.1	Te	ND	0.1
									Tl	0.91	0.1
									Th	ND	0.1
									Tm	ND	0.1
									Sn	0.14	0.1
									Ti	ND	0.1
									W	ND	0.1
									U	ND	0.1
									V	ND	1
									Yb	ND	0.1
									Y	0.2	0.1
									Zn	ND	1
									Zr	0.29	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:

Date:

UBJ  
2/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** Sodium 10000ppm Stock Std **ME #:** 0702003  
**Date Received/Prepped:** 2/20/2007 **By:** WBH  
**Date Expired:** 8/16/2008 **Lot #:** 07B057  
**Manufacturer:** CPI **Certificate:** Y  
**Matrix:** 1% HNO3 **NIST SRM:** 3152a  
**Amount:** 250 mL Room temp. storage

<b>Component</b>	<b>Comment</b>	<b>Conc. Unit:</b>
Na	P/N 4400-10M521	10000 ppm



Initial: WBH  
Date: 2/20/07

## METALS STANDARD DOCUMENTATION

**Standard:** Calcium 10000ppm Stock Std  
**Date Received/Prepped:** 2/20/2007  
**Date Expired:** 8/16/2008  
**Manufacturer:** CPI  
**Matrix:** 4% HNO3  
**Amount:** 250 mL

**ME #:** 0702002  
**By:** WBH  
**Lot #:** 07B065  
**Certificate:** Y  
**NIST SRM:** 3109a  
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions  
in Analytical Science and  
Technology

**USA**  
5680 Skylane Boulevard 707.525.5788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

**EUROPE**  
P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

## CERTIFICATE OF ANALYSIS

**P/N 4400-10M91**  
**P/N S4400-10M91**

170702602

Single-Element Calcium Standard  
Ca in 4% HNO<sub>3</sub>  
10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO<sub>3</sub>)  
Source Purity: 99.997%  
Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Tl 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



Initial:

Date:

W34  
1/27/07

## METALS STANDARD DOCUMENTATION

Standard: FE 10000ppm Stock Std  
Date Received/Prepped: 1/26/2007  
Date Expired: 7/19/2008  
Manufacturer: CPI  
Matrix: 4% HNO3  
Amount: 100 mL

ME #: 701008  
By: WBH  
Lot #: 061143  
Certificate:  
NIST SRM: 3126a  
Storage: Room Temp

Component	Comment	Conc. Unit:
Fe	PN4400-10M261	10000 PPM



**USA**

5580 Skylane Boulevard 707.525.5788  
Santa Rosa, CA 95403 800.878.7654  
www.cpiinternational.com Fax 707.545.7901

**EUROPE**

P.O. Box 2704 +31 20 638 05 97  
1000 CS Amsterdam Fax +31 20 420 28 36  
The Netherlands www.cpiinternational.com

Innovative Solutions  
in Analytical Science and  
Technology

# CERTIFICATE OF ANALYSIS

MTE070100X

**P/N 4400-10M261**  
**P/N S4400-10M261**  
Single-Element Iron Standard  
Fe in 4% HNO<sub>3</sub>  
10,000 ± 30 µg/mL

Lot # 06I143

Material Source: Iron Metal  
Source Purity: 99.999%  
Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL					
Al	INT	0.1	Cu	6.4	0.1	Pb	ND	0.1	K	ND	70	Tl	0.18	0.1
Sb	0.35	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1	Th	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1	Tm	ND	0.1
Ba	ND	0.1	Eu	ND	0.1	Mg	1.3	0.2	Rh	ND	0.1	Sn	0.67	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	INT	1	Rb	ND	0.1	Ti	0.21	0.1
Bi	ND	0.1	Ga	0.41	0.1	Hg	ND	0.2	Ru	ND	0.1	W	0.13	0.1
B	ND	4	Ge	INT	0.1	Mo	4.9	0.1	Sm	ND	0.1	U	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6	V	ND	1
Cd	ND	0.1	Hf	ND	0.1	Ni	9.3	0.1	Si	INT	8	Yb	ND	0.1
Ca	15	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1	Y	ND	0.1
Ce	ND	0.1	I	0.34	0.2	Os	ND	0.1	Na	8	1	Zn	8.6	2
Cs	0.34	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1	Zr	ND	0.1
Cr	3.3	1	Fe	X	30	P	28	10	Ta	ND	0.1			
Co	12	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1			

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.