

TABLE OF CONTENTS

COVER PAGE	1
QC CHECKLIST	2
SUMMARY SHEET	3
RUNLOG	5
INITIAL CALIBRATION	7
PERIODIC QC	22
QC: (MBLK, MRL, LCS, LCSD)	30
QC: (MS/MSD 2806190595)	34
PERIODIC QC	36
QC: (MS/MSD 2806250566)	38
PERIODIC QC	40
SAMPLES	46
STANDARDS PREPARATION WORKSHEET AND CERTIFICATES OF ANALYSIS	49

Level IV Data Package

MWH Group 245247

Method: EPA 200.7

2806240538

2806240540

EPA 200.7/6010B QC Check List

Analyst CSK Analysis Date 6-30-08 Reviewer/Date 7-1-08

Instrument PerKin Elmer Optima 4300DV

- All sample analyzed within 6 month holding time
- All sample raw concentration below the high standard or linear range o marked for dilution and rerun

- Initial and closing QC
- ICV within +/- 5%
 - Linearity check +/- 10%
 - ICSAB +/- 20%
 - QCS +/- 5%
 - MRL +/- 50%

- Middle, closing and batch QC
- FilterCheck < 1/2 MRL
 - MBLANK < 1/2 MRL
 - LCS +/-15%
 - MS/MSD +/-30%(200.7) +/- 25% (6010B)
 - CCV/MCV/ECV +/- 10%
 - ICB/CCB/ECB < 1/2 MRL
 - CCB ran after the CCV

- General QC
- RPD between MS/MSD is within +/-20%
 - RPD between LCS/LCSD is within +/- 20%
 - Internal standards 60 TO 125%
 - All pH of the samples are < 2

- No more than 20 samples per batch
- MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

- NA QIR needed for failed QC
- NA Special Det Code noted on the cover sheet
- NA R value for multi point calibration is > 0.995
- NA Proper MRL check ran for special low MRL samples

Reagent and Standards used for
Optima 4300 DV
Updated 03/31/08

Method 200.7/6010

Int: CSK
Date: 7-1-08

ICP SUMMARY SHEET

File ID: 080630A
 Date Started: 6/30/08
 Analyst ID: CSK

SAMPLE ID

LINEARITY	(22:02)	FILTERCHECK	(22:30)	2806190595	(22:46)
2806250566	(22:55)	2806180344	(23:12)	2806180345	(23:16)
2806190596	(23:19)	2806200356	(23:23)	2806200358	(23:27)
2806240538_2	(23:31)	2806240540_2	(23:35)	2806250254	(23:39)
2806250255	(23:52)	2806250256	(23:56)	2806250257	(0:00)
2806250258	(0:04)	2806250259	(0:08)	2806250260	(0:12)
2806250143	(0:15)	2806250147	(0:19)	2806250159	(0:22)
2806260194_2	(0:26)	2806250685	(0:54)	2806250695	(1:05)
2806250165	(1:25)	2806250176	(1:29)	2806250200	(1:32)
2806250206	(1:36)	2806250223	(1:40)	2806250684	(1:43)
2806250686	(1:46)	2806250687	(1:50)	2806250688	(1:54)
2806250689	(2:04)	2806250693	(2:08)	2806250696	(2:12)
2806250697	(2:16)	2806250698	(2:21)	2806250699	(2:24)
2806250700	(2:28)	2806250701	(2:32)	2806260200	(2:36)

WB/11/17/08

COMMENT:

Analyst: CSK
 7-1-08

Approved By: W311

Peer Reviewed: BLR

BLR 7/1/08
 7/2/08

BATCH NUMBER for 080630A

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2806190595

2806190595	2806250566	2806180344
2806180345	2806190596	2806200356
2806200358	2806240538_2X	2806240540_2X
2806250254	2806250255	2806250256
2806250257	2806250258	2806250259
2806250260	2806250143	2806250147
2806250159	2806260194_2X	

Batch ID: 2806250685

2806250685	2806250695	2806250165
2806250176	2806250200	2806250206
2806250223	2806250684	2806250686
2806250687	2806250688	2806250689
2806250693	2806250696	2806250697
2806250698	2806250699	2806250700
2806250701	2806260200	

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	6/30/08	21:48	1	9.8001	9.8	95-105	98.0%
QC-25 lppm	6/30/08	21:53	1	.96278	.960		
CCV	6/30/08	21:56	1	4.9894	4.99	90-110	99.7%
MCV	6/30/08	21:59	1	2.4908	2.49	90-110	99.6%
LINEARITY	6/30/08	22:02	1	0.0022	.0022		
ICSA	6/30/08	22:10	1	0.0001	0.0000	80-120	
ICSAB	6/30/08	22:13	1	.24168	.242	80-120	96.6%
ICB	6/30/08	22:16	1	-0.0002	ND		
MRL	6/30/08	22:19	1	0.0094	.0094	50-150	94.2%
FILTERCHECK	6/30/08	22:30	1	0.0006	0.0005		
MBLANK2007	6/30/08	22:34	1	-0.0004	ND		
MRL2007	6/30/08	22:37	1	0.0093	.0093		
LCS2007	6/30/08	22:41	1	.93157	.932	85-115	93.1%
LCSD2007	6/30/08	22:43	1	.89896	.899	85-115	89.8%
2806190595	6/30/08	22:46	1	-0.0004	ND		
2806190595MS	6/30/08	22:49	1	.93728	.937	[0.937]	93.7%
2806190595MSD	6/30/08	22:52	1	.98157	.982	[0.982]	98.1%
2806190595T	6/30/08	22:52	1		1.00	70 - 130	
2806250566	6/30/08	22:55	1	0.0004	0.0003		
CCV	6/30/08	22:59	1	4.9442	4.94	90-110	98.8%
CCB	6/30/08	23:02	1	-0.0003	ND		
2806250566MS	6/30/08	23:06	1	.96164	.962	[0.962]	96.1%
2806250566MSD	6/30/08	23:09	1	.94186	.942	[0.942]	94.1%
2806250566T	6/30/08	23:09	1		1.00	70 - 130	
2806180344	6/30/08	23:12	1	-0.0002	ND		
2806180345	6/30/08	23:16	1	-0.0006	ND		
2806190596	6/30/08	23:19	1	-0.0002	ND		
2806200356	6/30/08	23:23	1	0.0004	0.0004		
2806200358	6/30/08	23:27	1	-0.0005	ND		
2806240538_2X	6/30/08	23:31	2	0.0076	.0076		
2806240540_2X	6/30/08	23:35	2	0.0216	.022		
2806250254	6/30/08	23:39	1	0.0031	.0031		
CCV	6/30/08	23:43	1	4.9000	4.9	90-110	98.0%
CCB	6/30/08	23:46	1	-0.0003	ND		
MCV	6/30/08	23:49	1	2.4495	2.45	90-110	97.9%
2806250255	6/30/08	23:52	1	0.0059	.0059		
2806250256	6/30/08	23:56	1	-0.0000	ND		
2806250257	7/1/08	0:00	1	0.0062	.0062		
2806250258	7/1/08	0:04	1	0.0177	.018		
2806250259	7/1/08	0:08	1	0.0109	.011		
2806250260	7/1/08	0:12	1	0.0284	.028		
2806250143	7/1/08	0:15	1	-0.0003	ND		
2806250147	7/1/08	0:19	1	0.0002	0.0001		
2806250159	7/1/08	0:22	1	0.0001	0.0001		
2806260194_2X	7/1/08	0:26	2	.26327	.260		
CCV	7/1/08	0:29	1	4.8428	4.84	90-110	96.8%

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
CCB	7/1/08	0:35	1	-0.0004	ND		
MBLANK2007	7/1/08	0:39	1	-0.0005	ND		
MRL	7/1/08	0:42	1	0.0093	.0093	50-150	93.0%
MRL2007	7/1/08	0:46	1	0.0088	.0088		
LCS2007	7/1/08	0:49	1	.90306	.903	85-115	90.3%
LCSD2007	7/1/08	0:52	1	.89948	.899	85-115	89.9%
2806250685	7/1/08	0:54	1	0.0004	0.0004		
2806250685MS	7/1/08	0:59	1	.89705	.897	[0.897]	89.7%
2806250685MSD	7/1/08	1:02	1	.91068	.911	[0.911]	91.0%
2806250685T	7/1/08	1:02	1		1.00	70 - 130	
2806250695	7/1/08	1:05	1	0.0117	.012		
2806250695MS	7/1/08	1:09	1	.90052	.901	[0.889]	88.8%
CCV	7/1/08	1:12	1	4.7877	4.79	90-110	95.7%
CCB	7/1/08	1:15	1	-0.0003	ND		
MCV	7/1/08	1:19	1	2.4209	2.42	90-110	96.8%
2806250695MSD	7/1/08	1:22	1	.88333	.883	[0.872]	87.1%
2806250695T	7/1/08	1:22	1		1.00	70 - 130	
2806250165	7/1/08	1:25	1	-0.0001	ND		
2806250176	7/1/08	1:29	1	0.0045	.0045		
2806250200	7/1/08	1:32	1	0.0000	0.0000		
2806250206	7/1/08	1:36	1	0.0003	0.0002		
2806250223	7/1/08	1:40	1	0.0004	0.0004		
2806250684	7/1/08	1:43	1	0.0021	.0021		
2806250686	7/1/08	1:46	1	0.0005	0.0004		
2806250687	7/1/08	1:50	1	0.0008	0.0008		
2806250688	7/1/08	1:54	1	0.0002	0.0002		
CCV	7/1/08	1:57	1	4.7677	4.77	90-110	95.3%
CCB	7/1/08	2:01	1	-0.0004	ND		
2806250689	7/1/08	2:04	1	0.0003	0.0002		
2806250693	7/1/08	2:08	1	0.0008	0.0007		
2806250696	7/1/08	2:12	1	0.0003	0.0002		
2806250697	7/1/08	2:16	1	0.0011	.0011		
2806250698	7/1/08	2:21	1	0.0006	0.0006		
2806250699	7/1/08	2:24	1	0.0007	0.0007		
2806250700	7/1/08	2:28	1	0.0004	0.0004		
2806250701	7/1/08	2:32	1	0.0001	0.0000		
2806260200	7/1/08	2:36	1	0.0014	.0014		
ECV	7/1/08	2:39	1	4.8155	4.82	90-110	96.3%
ECB	7/1/08	2:42	1	-0.0004	ND		

=====
Analysis Begun

Start Time: 6/30/2008 21:42:41

Plasma On Time: 6/30/2008 12:36:04

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif

Batch ID: 080630A

Results Data Set: 080630

Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====
Method Loaded

Method Name: 200.7&6010_080304

Method Last Saved: 5/2/2008 09:27:36

IEC File: IEC080304.iec

MSF File:

Method Description: 200.7/6010_080304

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
MO	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a

=====
Sequence No.: 1

Autosampler Location: 0

Sample ID: Calib Blank 1

Date Collected: 6/30/2008 21:42:57

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

=====
Nebulizer Parameters: Calib Blank 1

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

=====
Mean Data: Calib Blank 1

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc.	Units
Sca	417193.4	2582.54	0.62%	100	%
Yr	490636.0	6698.55	1.37%	100	%
Ag†	424.6	115.10	27.11%	[0.00]	mg/L
Al†	70.7	5.17	7.31%	[0.00]	mg/L
As†	-2.2	4.00	184.68%	[0.00]	mg/L
B†	80.2	8.39	10.47%	[0.00]	mg/L
Ba†	-27.5	0.64	2.34%	[0.00]	mg/L
Be†	-3600.2	0.77	0.02%	[0.00]	mg/L
Cat	3566.0	67.64	1.90%	[0.00]	mg/L

Cdt	35.3	2.81	7.95%	[0.00] mg/L
Cot	-47.7	2.48	5.19%	[0.00] mg/L
Crt	122.8	5.64	4.59%	[0.00] mg/L
Cut	3893.7	80.02	2.06%	[0.00] mg/L
Fet	-106.2	4.17	3.93%	[0.00] mg/L
Kt	131.3	7.30	5.56%	[0.00] mg/L
Mgt	-542.3	11.88	2.19%	[0.00] mg/L
Mnt	330.9	1.27	0.38%	[0.00] mg/L
Mot	10.5	1.69	16.03%	[0.00] mg/L
Nat	-151.5	7.37	4.87%	[0.00] mg/L
Nit	-46.7	2.63	5.62%	[0.00] mg/L
Pbt	-20.8	1.74	8.38%	[0.00] mg/L
Sbt	7.4	1.68	22.73%	[0.00] mg/L
Set	-1.8	2.59	141.91%	[0.00] mg/L
Tlt	-21.3	3.98	18.73%	[0.00] mg/L
Vt	137.9	21.76	15.78%	[0.00] mg/L
Znt	138.9	2.66	1.91%	[0.00] mg/L

Fe	1	Lin, Calc Int	0.0	11390	0.00000	1.000000
K	1	Lin, Calc Int	0.0	1199	0.00000	1.000000
Mg	1	Lin, Calc Int	0.0	21340	0.00000	1.000000
Mn	1	Lin, Calc Int	0.0	495400	0.00000	1.000000
Mo	1	Lin, Calc Int	0.0	9979	0.00000	1.000000
Na	1	Lin, Calc Int	0.0	5154	0.00000	1.000000
Ni	1	Lin, Calc Int	0.0	18810	0.00000	1.000000
Pb	1	Lin, Calc Int	0.0	3866	0.00000	1.000000
Sb	1	Lin, Calc Int	0.0	1563	0.00000	1.000000
Se	1	Lin, Calc Int	0.0	985.7	0.00000	1.000000
Tl	1	Lin, Calc Int	0.0	2150	0.00000	1.000000
V	1	Lin, Calc Int	0.0	155200	0.00000	1.000000
Zn	1	Lin, Calc Int	0.0	38340	0.00000	1.000000

=====
Analysis Begun

Start Time: 6/30/2008 21:48:39

Plasma On Time: 6/30/2008 12:36:04

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif

Batch ID: 080630A

Results Data Set: 080630

Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====
Sequence No.: 1

Autosampler Location: 15

Sample ID: ICV

Date Collected: 6/30/2008 21:48:40

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	371774.2	89.1	%	0.42			0.47%
Yr	460366.6	93.8	%	0.06			0.06%
Agf	500392.4	1.98	mg/L	0.006	1.98 mg/L	0.006	0.28%
	QC value within limits for Ag	Recovery	= 98.90%				
Alf	76550.8	9.48	mg/L	0.026	9.48 mg/L	0.026	0.27%
	QC value less than the lower limit for Al	Recovery	= 94.77%				
Asf	15133.5	9.78	mg/L	0.114	9.78 mg/L	0.114	1.16%
	QC value within limits for As	Recovery	= 97.79%				
B_f	135186.0	4.94	mg/L	0.008	4.94 mg/L	0.008	0.15%
	QC value within limits for B_	Recovery	= 98.77%				
Baf	367776.2	9.83	mg/L	0.027	9.83 mg/L	0.027	0.27%
	QC value within limits for Ba	Recovery	= 98.33%				
Be_f	10724653.7	3.96	mg/L	0.035	3.96 mg/L	0.035	0.88%
	QC value within limits for Be	Recovery	= 99.00%				
Ca_f	2985248.4	98.9	mg/L	0.98	98.9 mg/L	0.98	0.99%
	QC value within limits for Ca	Recovery	= 98.90%				
Cdf	105326.6	4.83	mg/L	0.008	4.83 mg/L	0.008	0.17%
	QC value within limits for Cd	Recovery	= 96.58%				
Cof	155367.8	9.88	mg/L	0.021	9.88 mg/L	0.021	0.22%
	QC value within limits for Co	Recovery	= 98.77%				
Crf	452582.9	9.80	mg/L	0.011	9.80 mg/L	0.011	0.11%
	QC value within limits for Cr	Recovery	= 98.00%				
Cuf	2928943.7	9.84	mg/L	0.029	9.84 mg/L	0.029	0.30%
	QC value within limits for Cu	Recovery	= 98.42%				
Fe_f	112474.9	9.87	mg/L	0.052	9.87 mg/L	0.052	0.53%
	QC value within limits for Fe	Recovery	= 98.74%				
Kf	118478.0	98.8	mg/L	0.35	98.8 mg/L	0.35	0.35%
	QC value within limits for K	Recovery	= 98.82%				
Mgf	2104829.4	98.7	mg/L	1.04	98.7 mg/L	1.04	1.05%
	QC value within limits for Mg	Recovery	= 98.67%				
Mnf	4884760.0	9.86	mg/L	0.002	9.86 mg/L	0.002	0.02%
	QC value within limits for Mn	Recovery	= 98.57%				
Mof	98386.0	9.86	mg/L	0.025	9.86 mg/L	0.025	0.26%
	QC value within limits for Mo	Recovery	= 98.59%				
Naf	507646.8	98.5	mg/L	0.31	98.5 mg/L	0.31	0.31%
	QC value within limits for Na	Recovery	= 98.49%				
Nif	186382.4	9.91	mg/L	0.009	9.91 mg/L	0.009	0.09%
	QC value within limits for Ni	Recovery	= 99.08%				
Pbf	38075.0	9.85	mg/L	0.092	9.85 mg/L	0.092	0.93%
	QC value within limits for Pb	Recovery	= 98.48%				
Sbf	15299.6	9.83	mg/L	0.060	9.83 mg/L	0.060	0.61%
	QC value within limits for Sb	Recovery	= 98.26%				
Se_f	9663.3	9.83	mg/L	0.107	9.83 mg/L	0.107	1.09%

Ni†	19590.2	1.04 mg/L	0.003	1.04 mg/L	0.003	0.32%
	QC value within limits for Ni	Recovery = 104.14%				
Pb†	3958.5	1.02 mg/L	0.004	1.02 mg/L	0.004	0.40%
	QC value within limits for Pb	Recovery = 102.39%				
Sb†	1427.5	0.917 mg/L	0.0053	0.917 mg/L	0.0053	0.58%
	QC value within limits for Sb	Recovery = 91.67%				
Se†	904.4	0.920 mg/L	0.0002	0.920 mg/L	0.0002	0.02%
	QC value within limits for Se	Recovery = 92.00%				
Tl†	2230.7	1.04 mg/L	0.003	1.04 mg/L	0.003	0.32%
	QC value within limits for Tl	Recovery = 104.29%				
V†	144396.6	0.935 mg/L	0.0009	0.935 mg/L	0.0009	0.10%
	QC value within limits for V	Recovery = 93.53%				
Zn†	38675.4	1.000 mg/L	0.0015	1.000 mg/L	0.0015	0.15%
	QC value within limits for Zn	Recovery = 99.98%				

All analyte(s) passed QC.

Sequence No.: 8
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/30/2008 21:56:52
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	387573.7	92.9	%	0.11			0.12%
Yr	473380.9	96.5	%	0.47			0.48%
Ag†	241597.0	0.955	mg/L	0.0013	0.955 mg/L	0.0013	0.13%
	QC value within limits for Ag	Recovery = 95.51%					
As†	7350.4	4.75	mg/L	0.007	4.75 mg/L	0.007	0.15%
	QC value within limits for As	Recovery = 94.99%					
B_†	67027.8	2.45	mg/L	0.014	2.45 mg/L	0.014	0.55%
	QC value within limits for B_	Recovery = 97.94%					
Ba†	185489.2	4.96	mg/L	0.010	4.96 mg/L	0.010	0.19%
	QC value within limits for Ba	Recovery = 99.19%					
Be†	5457754.4	2.02	mg/L	0.016	2.02 mg/L	0.016	0.79%
	QC value within limits for Be	Recovery = 100.76%					
Ca†	1509816.6	50.0	mg/L	0.03	50.0 mg/L	0.03	0.07%
	QC value within limits for Ca	Recovery = 100.04%					
Cd†	42123.0	1.92	mg/L	0.006	1.92 mg/L	0.006	0.32%
	QC value within limits for Cd	Recovery = 96.16%					
Co†	79627.3	5.06	mg/L	0.025	5.06 mg/L	0.025	0.49%
	QC value within limits for Co	Recovery = 101.25%					
Cr†	230417.3	4.99	mg/L	0.020	4.99 mg/L	0.020	0.40%
	QC value within limits for Cr	Recovery = 99.79%					
Cu†	1445995.3	4.86	mg/L	0.002	4.86 mg/L	0.002	0.04%
	QC value within limits for Cu	Recovery = 97.18%					
Fe†	57271.8	5.03	mg/L	0.027	5.03 mg/L	0.027	0.53%
	QC value within limits for Fe	Recovery = 100.55%					
K†	57753.0	48.2	mg/L	0.01	48.2 mg/L	0.01	0.02%
	QC value within limits for K	Recovery = 96.34%					
Mg†	1061084.7	49.7	mg/L	0.01	49.7 mg/L	0.01	0.02%
	QC value within limits for Mg	Recovery = 99.48%					
Mn†	2474785.0	4.99	mg/L	0.012	4.99 mg/L	0.012	0.25%
	QC value within limits for Mn	Recovery = 99.88%					
Mo†	48993.6	4.91	mg/L	0.030	4.91 mg/L	0.030	0.61%
	QC value within limits for Mo	Recovery = 98.19%					
Na†	248140.4	48.1	mg/L	0.03	48.1 mg/L	0.03	0.05%
	QC value within limits for Na	Recovery = 96.29%					
Ni†	95685.5	5.09	mg/L	0.008	5.09 mg/L	0.008	0.15%
	QC value within limits for Ni	Recovery = 101.73%					
Pb†	19507.8	5.05	mg/L	0.011	5.05 mg/L	0.011	0.21%
	QC value within limits for Pb	Recovery = 100.92%					
Sb†	7513.4	4.83	mg/L	0.005	4.83 mg/L	0.005	0.09%
	QC value within limits for Sb	Recovery = 96.51%					
Se†	4833.5	4.92	mg/L	0.000	4.92 mg/L	0.000	0.01%
	QC value within limits for Se	Recovery = 98.33%					
Tl†	11066.5	5.17	mg/L	0.006	5.17 mg/L	0.006	0.11%
	QC value within limits for Tl	Recovery = 103.48%					
V†	762861.4	4.94	mg/L	0.009	4.94 mg/L	0.009	0.18%
	QC value within limits for V	Recovery = 98.82%					
Zn†	195906.6	5.07	mg/L	0.010	5.07 mg/L	0.010	0.20%
	QC value within limits for Zn	Recovery = 101.32%					

All analyte(s) passed QC.

User canceled analysis.

=====
Analysis Begun

Start Time: 6/30/2008 21:59:37

Plasma On Time: 6/30/2008 12:36:04

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif

Batch ID: 080630A

Results Data Set: 080630

Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 35

Autosampler Location: 5

Sample ID: MCV

Date Collected: 6/30/2008 21:59:39

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	408212.0	97.8	%	0.77			0.79%
Yr	483998.3	98.6	%	0.59			0.60%
Ag†	120651.0	0.477	mg/L	0.0015	0.477 mg/L	0.0015	0.32%
	QC value within limits for Ag	Recovery = 95.40%					
As†	3644.0	2.35	mg/L	0.026	2.35 mg/L	0.026	1.11%
	QC value within limits for As	Recovery = 94.19%					
B_f	33964.9	1.24	mg/L	0.003	1.24 mg/L	0.003	0.21%
	QC value within limits for B_f	Recovery = 99.28%					
Ba†	93630.7	2.50	mg/L	0.011	2.50 mg/L	0.011	0.45%
	QC value within limits for Ba	Recovery = 100.14%					
Be†	2696329.0	0.996	mg/L	0.0180	0.996 mg/L	0.0180	1.80%
	QC value within limits for Be	Recovery = 99.56%					
Ca†	754953.2	25.0	mg/L	0.07	25.0 mg/L	0.07	0.27%
	QC value within limits for Ca	Recovery = 100.05%					
Cd†	20709.5	0.945	mg/L	0.0084	0.945 mg/L	0.0084	0.89%
	QC value within limits for Cd	Recovery = 94.54%					
Co†	40163.9	2.55	mg/L	0.002	2.55 mg/L	0.002	0.06%
	QC value within limits for Co	Recovery = 102.14%					
Cr†	115032.3	2.49	mg/L	0.021	2.49 mg/L	0.021	0.86%
	QC value within limits for Cr	Recovery = 99.64%					
Cu†	719074.6	2.42	mg/L	0.047	2.42 mg/L	0.047	1.95%
	QC value within limits for Cu	Recovery = 96.66%					
Fe†	28691.2	2.52	mg/L	0.015	2.52 mg/L	0.015	0.61%
	QC value within limits for Fe	Recovery = 100.75%					
K†	28751.1	24.0	mg/L	0.07	24.0 mg/L	0.07	0.29%
	QC value within limits for K	Recovery = 95.92%					
Mg†	533561.7	25.0	mg/L	0.06	25.0 mg/L	0.06	0.22%
	QC value within limits for Mg	Recovery = 100.05%					
Mn†	1236289.1	2.49	mg/L	0.036	2.49 mg/L	0.036	1.46%
	QC value within limits for Mn	Recovery = 99.79%					
Mo†	24257.9	2.43	mg/L	0.021	2.43 mg/L	0.021	0.86%
	QC value within limits for Mo	Recovery = 97.24%					
Na†	123546.2	24.0	mg/L	0.04	24.0 mg/L	0.04	0.18%
	QC value within limits for Na	Recovery = 95.88%					
Ni†	48569.3	2.58	mg/L	0.004	2.58 mg/L	0.004	0.15%
	QC value within limits for Ni	Recovery = 103.27%					
Pb†	9823.3	2.54	mg/L	0.022	2.54 mg/L	0.022	0.88%
	QC value within limits for Pb	Recovery = 101.63%					
Sb†	3737.8	2.40	mg/L	0.024	2.40 mg/L	0.024	1.02%
	QC value within limits for Sb	Recovery = 96.02%					
Se†	2371.9	2.41	mg/L	0.026	2.41 mg/L	0.026	1.06%
	QC value within limits for Se	Recovery = 96.51%					

Tl†	5586.7	2.61 mg/L	0.026	2.61 mg/L	0.026	0.98%
	QC value within limits for Tl	Recovery = 104.48%				
V†	379434.4	2.46 mg/L	0.007	2.46 mg/L	0.007	0.29%
	QC value within limits for V	Recovery = 98.30%				
Zn†	98538.1	2.55 mg/L	0.013	2.55 mg/L	0.013	0.51%
	QC value within limits for Zn	Recovery = 101.92%				

All analyte(s) passed QC.

User canceled analysis.

=====
Analysis Begun

Start Time: 6/30/2008 22:02:27
Logged In Analyst: Charley Kay
Spectrometer Model: Optima 4300 DV, S/N 077N2121801

Plasma On Time: 6/30/2008 12:36:04
Technique: ICP Continuous

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif
Batch ID: 080630A
Results Data Set: 080630
Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====
Sequence No.: 3
Sample ID: LINEARITY
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 9
Date Collected: 6/30/2008 22:02:29
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte Back Pressure Flow
All 212.0 kPa 0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	362531.6	86.9	%	0.76			0.87%
Yr	450312.7	91.8	%	0.25			0.28%
Ag†	-4907.7	0.0145	mg/L	0.00028	0.0145 mg/L	0.00028	1.92%
	QC value within limits for Ag	Recovery =	Not calculated				
As†	-135.6	-0.0876	mg/L	0.00325	-0.0876 mg/L	0.00325	3.71%
	QC value within limits for As	Recovery =	Not calculated				
B_†	352.9	0.0132	mg/L	0.00459	0.0132 mg/L	0.00459	34.75%
	QC value within limits for B_	Recovery =	Not calculated				
Ba†	46.0	0.00123	mg/L	0.000081	0.00123 mg/L	0.000081	6.56%
	QC value within limits for Ba	Recovery =	Not calculated				
Be†	-788.4	-0.00029	mg/L	0.000012	-0.00029 mg/L	0.000012	4.21%
	QC value within limits for Be	Recovery =	Not calculated				
Ca†	8480455.9	281	mg/L	0.9	281 mg/L	0.9	0.30%
	QC value within limits for Ca	Recovery =	93.65%				
Cd†	-17.1	0.00042	mg/L	0.000208	0.00042 mg/L	0.000208	49.38%
	QC value within limits for Cd	Recovery =	Not calculated				
Co†	37.3	0.00237	mg/L	0.000105	0.00237 mg/L	0.000105	4.42%
	QC value within limits for Co	Recovery =	Not calculated				
Cr†	102.5	0.00222	mg/L	0.000013	0.00222 mg/L	0.000013	0.59%
	QC value within limits for Cr	Recovery =	Not calculated				
Cu†	-3196.0	-0.0107	mg/L	0.00020	-0.0107 mg/L	0.00020	1.84%
	QC value within limits for Cu	Recovery =	Not calculated				
Fe†	1048663.5	92.1	mg/L	0.36	92.1 mg/L	0.36	0.40%
	QC value within limits for Fe	Recovery =	92.06%				
K†	360155.1	300	mg/L	0.3	300 mg/L	0.3	0.12%
	QC value within limits for K	Recovery =	100.13%				
Mg†	3791783.7	178	mg/L	0.5	178 mg/L	0.5	0.28%
	QC value within limits for Mg	Recovery =	Not calculated				
Mn†	-1121.9	-0.00626	mg/L	0.000030	-0.00626 mg/L	0.000030	0.49%
	QC value within limits for Mn	Recovery =	Not calculated				
Mo†	44.0	0.00441	mg/L	0.001315	0.00441 mg/L	0.001315	29.83%
	QC value within limits for Mo	Recovery =	Not calculated				
Na†	1480197.0	287	mg/L	0.6	287 mg/L	0.6	0.21%
	QC value within limits for Na	Recovery =	95.73%				
Ni†	0.8	0.00005	mg/L	0.000582	0.00005 mg/L	0.000582	>999.9%
	QC value within limits for Ni	Recovery =	Not calculated				
Pb†	-11.9	-0.00308	mg/L	0.000343	-0.00308 mg/L	0.000343	11.15%
	QC value within limits for Pb	Recovery =	Not calculated				
Sb†	14.8	0.00943	mg/L	0.000825	0.00943 mg/L	0.000825	8.75%
	QC value within limits for Sb	Recovery =	Not calculated				
Se†	-109.4	0.117	mg/L	0.0032	0.117 mg/L	0.0032	2.78%
	QC value within limits for Se	Recovery =	Not calculated				

Tl†	27.5	0.0127 mg/L	0.00379	0.0127 mg/L	0.00379	29.75%
QC value within limits for Tl Recovery = Not calculated						
V†	-534.5	0.00018 mg/L	0.000289	0.00018 mg/L	0.000289	161.94%
QC value within limits for V Recovery = Not calculated						
Zn†	984.0	0.0257 mg/L	0.00030	0.0257 mg/L	0.00030	1.18%
QC value within limits for Zn Recovery = Not calculated						
All analyte(s) passed QC.						

Sequence No.: 4
 Sample ID: ICESA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 6/30/2008 22:06:09
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICESA

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: ICESA

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	374215.5	89.7	%	0.96			1.07%
Yr	459046.7	93.6	%	0.31			0.33%
Ag†	-4759.7	0.0152	mg/L	0.00034	0.0152 mg/L	0.00034	2.26%
	QC value within limits for Ag	Recovery =	Not calculated				
As†	-320.8	-0.207	mg/L	0.0007	-0.207 mg/L	0.0007	0.35%
	QC value within limits for As	Recovery =	Not calculated				
B_†	-97.5	-0.00354	mg/L	0.000549	-0.00354 mg/L	0.000549	15.50%
	QC value within limits for B_	Recovery =	Not calculated				
Ba†	70.0	0.00187	mg/L	0.000042	0.00187 mg/L	0.000042	2.27%
	QC value within limits for Ba	Recovery =	Not calculated				
Be†	-681.2	-0.00025	mg/L	0.000015	-0.00025 mg/L	0.000015	6.02%
	QC value within limits for Be	Recovery =	Not calculated				
Ca†	7208128.8	239	mg/L	0.9	239 mg/L	0.9	0.36%
	QC value within limits for Ca	Recovery =	95.52%				
Cd†	-26.1	0.00165	mg/L	0.000079	0.00165 mg/L	0.000079	4.79%
	QC value within limits for Cd	Recovery =	Not calculated				
Co†	17.2	0.00110	mg/L	0.000339	0.00110 mg/L	0.000339	30.98%
	QC value within limits for Co	Recovery =	Not calculated				
Cr†	9.0	0.00019	mg/L	0.000209	0.00019 mg/L	0.000209	107.69%
	QC value within limits for Cr	Recovery =	Not calculated				
Cu†	-3344.4	-0.0112	mg/L	0.00008	-0.0112 mg/L	0.00008	0.69%
	QC value within limits for Cu	Recovery =	Not calculated				
Fe†	1052509.3	92.4	mg/L	0.29	92.4 mg/L	0.29	0.32%
	QC value within limits for Fe	Recovery =	92.40%				
K†	325.8	0.272	mg/L	0.0441	0.272 mg/L	0.0441	16.24%
	QC value within limits for K	Recovery =	Not calculated				
Mg†	4737557.0	222	mg/L	0.5	222 mg/L	0.5	0.23%
	QC value less than the lower limit for Mg	Recovery =	88.88%				
Mn†	-1213.9	-0.00744	mg/L	0.000025	-0.00744 mg/L	0.000025	0.34%
	QC value within limits for Mn	Recovery =	Not calculated				
Mo†	12.5	0.00125	mg/L	0.000277	0.00125 mg/L	0.000277	22.08%
	QC value within limits for Mo	Recovery =	Not calculated				
Na†	581.3	0.113	mg/L	0.0035	0.113 mg/L	0.0035	3.14%
	QC value within limits for Na	Recovery =	Not calculated				
Ni†	-11.5	-0.00061	mg/L	0.000285	-0.00061 mg/L	0.000285	46.42%
	QC value within limits for Ni	Recovery =	Not calculated				
Pb†	-148.4	-0.0384	mg/L	0.00142	-0.0384 mg/L	0.00142	3.70%
	QC value within limits for Pb	Recovery =	Not calculated				
Sb†	8.5	0.00543	mg/L	0.009254	0.00543 mg/L	0.009254	170.35%
	QC value within limits for Sb	Recovery =	Not calculated				
Se†	-111.3	0.116	mg/L	0.0011	0.116 mg/L	0.0011	0.94%
	QC value within limits for Se	Recovery =	Not calculated				
Tl†	15.7	0.00729	mg/L	0.001059	0.00729 mg/L	0.001059	14.53%
	QC value within limits for Tl	Recovery =	Not calculated				
V†	-545.1	0.00011	mg/L	0.000520	0.00011 mg/L	0.000520	464.10%
	QC value within limits for V	Recovery =	Not calculated				
Zn†	706.2	0.0184	mg/L	0.00031	0.0184 mg/L	0.00031	1.69%
	QC value within limits for Zn	Recovery =	Not calculated				

QC Failed. Continue with analysis.

Tl†	10.5	0.00485 mg/L	0.001693	0.00485 mg/L	0.001693	34.93%
QC value within limits for Tl		Recovery =	Not calculated			
V†	-574.3	-0.00007 mg/L	0.000550	-0.00007 mg/L	0.000550	800.78%
QC value within limits for V		Recovery =	Not calculated			
Zn†	713.6	0.0186 mg/L	0.00009	0.0186 mg/L	0.00009	0.47%
QC value within limits for Zn		Recovery =	Not calculated			

QC Failed. Continue with analysis.

User canceled analysis.

Analysis Begun

Start Time: 6/30/2008 22:13:50

Plasma On Time: 6/30/2008 12:36:04

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif

Batch ID: 080630A

Results Data Set: 080630

Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 5

Autosampler Location: 11

Sample ID: ICSAB

Date Collected: 6/30/2008 22:13:51

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	373126.6	89.4	%	0.81			0.91%
Yr	457819.9	93.3	%	0.01			0.01%
Ag†	92106.4	0.398	mg/L	0.0010	0.398 mg/L	0.0010	0.25%
QC value less than the lower limit for Ag				Recovery = 79.59%			
As†	-333.2	-0.215	mg/L	0.0018	-0.215 mg/L	0.0018	0.82%
QC value less than the lower limit for As				Recovery = Not calculated			
B_†	-241.7	-0.0111	mg/L	0.00110	-0.0111 mg/L	0.00110	9.93%
QC value within limits for B_				Recovery = Not calculated			
Ba†	9573.5	0.256	mg/L	0.0008	0.256 mg/L	0.0008	0.32%
QC value within limits for Ba				Recovery = 102.39%			
Be†	645400.5	0.238	mg/L	0.0005	0.238 mg/L	0.0005	0.19%
QC value within limits for Be				Recovery = 95.30%			
Ca†	7344480.4	243	mg/L	0.2	243 mg/L	0.2	0.08%
QC value within limits for Ca				Recovery = 97.33%			
Cd†	10305.5	0.486	mg/L	0.0046	0.486 mg/L	0.0046	0.95%
QC value within limits for Cd				Recovery = 97.29%			
Co†	3627.9	0.231	mg/L	0.0011	0.231 mg/L	0.0011	0.46%
QC value within limits for Co				Recovery = 92.26%			
Cr†	11161.1	0.242	mg/L	0.0009	0.242 mg/L	0.0009	0.38%
QC value within limits for Cr				Recovery = 96.67%			
Cu†	71220.6	0.239	mg/L	0.0004	0.239 mg/L	0.0004	0.17%
QC value within limits for Cu				Recovery = 95.73%			
Fe†	1068754.3	93.8	mg/L	0.49	93.8 mg/L	0.49	0.52%
QC value within limits for Fe				Recovery = 93.82%			
K†	131.1	0.109	mg/L	0.0392	0.109 mg/L	0.0392	35.87%
QC value within limits for K				Recovery = Not calculated			
Mg†	4839493.9	227	mg/L	0.6	227 mg/L	0.6	0.26%
QC value within limits for Mg				Recovery = 90.80%			
Mn†	121265.9	0.240	mg/L	0.0006	0.240 mg/L	0.0006	0.25%
QC value within limits for Mn				Recovery = 95.87%			
Mo†	1.1	0.00011	mg/L	0.000147	0.00011 mg/L	0.000147	129.98%
QC value within limits for Mo				Recovery = Not calculated			
Na†	342.0	0.0664	mg/L	0.00592	0.0664 mg/L	0.00592	8.93%
QC value within limits for Na				Recovery = Not calculated			
Ni†	8531.2	0.453	mg/L	0.0040	0.453 mg/L	0.0040	0.88%
QC value within limits for Ni				Recovery = 90.70%			
Pb†	1722.8	0.446	mg/L	0.0034	0.446 mg/L	0.0034	0.75%
QC value within limits for Pb				Recovery = 89.12%			
Sb†	10.2	0.00750	mg/L	0.007026	0.00750 mg/L	0.007026	93.68%
QC value within limits for Sb				Recovery = Not calculated			
Se†	-115.9	0.115	mg/L	0.0092	0.115 mg/L	0.0092	7.99%
QC value greater than the upper limit for Se				Recovery = Not calculated			

Tlt	17.2	0.0101 mg/L	0.00448	0.0101 mg/L	0.00448	44.24%
QC value within limits for Tl Recovery = Not calculated						
Vf	36910.9	0.243 mg/L	0.0007	0.243 mg/L	0.0007	0.28%
QC value within limits for V Recovery = 97.06%						
Znf	20395.1	0.528 mg/L	0.0012	0.528 mg/L	0.0012	0.23%
QC value within limits for Zn Recovery = 105.69%						
QC Failed. Continue with analysis.						

User canceled analysis.

=====
Analysis Begun

Start Time: 6/30/2008 22:16:32 Plasma On Time: 6/30/2008 12:36:04
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif
 Batch ID: 080630A
 Results Data Set: 080630
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====
 Sequence No.: 9 Autosampler Location: 0
 Sample ID: ICB Date Collected: 6/30/2008 22:16:33
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min.

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	421541.6	101	%	0.9			0.88%
Yr	499507.5	102	%	0.3			0.27%
Ag†	276.7	0.00110	mg/L	0.000236	0.00110 mg/L	0.000236	21.54%
QC value within limits for Ag		Recovery	= Not calculated				
As†	8.7	0.00564	mg/L	0.001491	0.00564 mg/L	0.001491	26.46%
QC value within limits for As		Recovery	= Not calculated				
B_f	44.9	0.00166	mg/L	0.000099	0.00166 mg/L	0.000099	5.94%
QC value within limits for B_f		Recovery	= Not calculated				
Bar	3.2	0.00008	mg/L	0.000109	0.00008 mg/L	0.000109	128.37%
QC value within limits for Ba		Recovery	= Not calculated				
Be†	159.1	0.00006	mg/L	0.000023	0.00006 mg/L	0.000023	39.22%
QC value within limits for Be		Recovery	= Not calculated				
Cat	502.1	0.0166	mg/L	0.00396	0.0166 mg/L	0.00396	23.82%
QC value within limits for Ca		Recovery	= Not calculated				
Cd†	0.5	-0.00006	mg/L	0.000060	-0.00006 mg/L	0.000060	107.09%
QC value within limits for Cd		Recovery	= Not calculated				
Co†	6.3	0.00040	mg/L	0.000027	0.00040 mg/L	0.000027	6.62%
QC value within limits for Co		Recovery	= Not calculated				
Cr†	-11.4	-0.00025	mg/L	0.000044	-0.00025 mg/L	0.000044	17.84%
QC value within limits for Cr		Recovery	= Not calculated				
Cu†	277.9	0.00093	mg/L	0.000032	0.00093 mg/L	0.000032	3.43%
QC value within limits for Cu		Recovery	= Not calculated				
Fe†	111.7	0.00980	mg/L	0.001899	0.00980 mg/L	0.001899	19.38%
QC value within limits for Fe		Recovery	= Not calculated				
K†	60.9	0.0508	mg/L	0.00616	0.0508 mg/L	0.00616	12.12%
QC value within limits for K		Recovery	= Not calculated				
Mg†	356.8	0.0167	mg/L	0.00407	0.0167 mg/L	0.00407	24.34%
QC value within limits for Mg		Recovery	= Not calculated				
Mn†	-114.5	-0.00023	mg/L	0.000013	-0.00023 mg/L	0.000013	5.54%
QC value within limits for Mn		Recovery	= Not calculated				
Mo†	1.7	0.00017	mg/L	0.000325	0.00017 mg/L	0.000325	189.63%
QC value within limits for Mo		Recovery	= Not calculated				
Na†	33.7	0.00654	mg/L	0.008413	0.00654 mg/L	0.008413	128.59%
QC value within limits for Na		Recovery	= Not calculated				
Ni†	-0.8	-0.00004	mg/L	0.000134	-0.00004 mg/L	0.000134	302.45%
QC value within limits for Ni		Recovery	= Not calculated				
Pb†	8.3	0.00215	mg/L	0.000622	0.00215 mg/L	0.000622	28.97%
QC value within limits for Pb		Recovery	= Not calculated				
Sb†	-2.6	-0.00163	mg/L	0.003103	-0.00163 mg/L	0.003103	190.24%
QC value within limits for Sb		Recovery	= Not calculated				
Se†	-2.5	-0.00256	mg/L	0.003236	-0.00256 mg/L	0.003236	126.51%
QC value within limits for Se		Recovery	= Not calculated				

Tlt	2.4	0.00111 mg/L	0.002490	0.00111 mg/L	0.002490	224.35%
QC value within limits for Tl		Recovery =	Not calculated			
Vf	17.2	0.00011 mg/L	0.000116	0.00011 mg/L	0.000116	105.63%
QC value within limits for V		Recovery =	Not calculated			
Znf	-15.9	-0.00041 mg/L	0.000041	-0.00041 mg/L	0.000041	9.94%
QC value within limits for Zn		Recovery =	Not calculated			

All analyte(s) passed QC.

Sequence No.: 10
 Sample ID: MRL
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 20
 Date Collected: 6/30/2008 22:19:57
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	422650.5	101	%	2.1			2.05%
Yr	496896.1	101	%	0.3			0.34%
Ag†	2490.0	0.00983	mg/L	0.000005	0.00983 mg/L	0.000005	0.05%
	QC value within limits for Ag Recovery = 98.33%						
As†	140.0	0.0904	mg/L	0.00043	0.0904 mg/L	0.00043	0.48%
	QC value within limits for As Recovery = 90.45%						
B_†	1271.2	0.0471	mg/L	0.00076	0.0471 mg/L	0.00076	1.61%
	QC value within limits for B_ Recovery = 94.14%						
Ba†	738.2	0.0197	mg/L	0.00023	0.0197 mg/L	0.00023	1.15%
	QC value within limits for Ba Recovery = 98.69%						
Be†	2706.9	0.00100	mg/L	0.000027	0.00100 mg/L	0.000027	2.67%
	QC value within limits for Be Recovery = 100.06%						
Ca†	30020.8	0.995	mg/L	0.0030	0.995 mg/L	0.0030	0.30%
	QC value within limits for Ca Recovery = 99.46%						
Cd†	124.1	0.00471	mg/L	0.000365	0.00471 mg/L	0.000365	7.75%
	QC value within limits for Cd Recovery = 94.28%						
Co†	782.1	0.0497	mg/L	0.00058	0.0497 mg/L	0.00058	1.16%
	QC value within limits for Co Recovery = 99.45%						
Cr†	435.1	0.00942	mg/L	0.000136	0.00942 mg/L	0.000136	1.45%
	QC value within limits for Cr Recovery = 94.21%						
Cu†	2959.6	0.00997	mg/L	0.000452	0.00997 mg/L	0.000452	4.54%
	QC value within limits for Cu Recovery = 99.68%						
Fe†	263.9	0.0232	mg/L	0.00069	0.0232 mg/L	0.00069	2.99%
	QC value within limits for Fe Recovery = 115.82%						
K†	1158.2	0.966	mg/L	0.0154	0.966 mg/L	0.0154	1.59%
	QC value within limits for K Recovery = 96.60%						
Mg†	2304.2	0.108	mg/L	0.0007	0.108 mg/L	0.0007	0.66%
	QC value within limits for Mg Recovery = 108.04%						
Mn†	915.4	0.00185	mg/L	0.000043	0.00185 mg/L	0.000043	2.34%
	QC value within limits for Mn Recovery = 92.26%						
Mo†	190.3	0.0191	mg/L	0.00014	0.0191 mg/L	0.00014	0.71%
	QC value within limits for Mo Recovery = 95.35%						
Na†	4781.9	0.928	mg/L	0.0058	0.928 mg/L	0.0058	0.63%
	QC value within limits for Na Recovery = 92.78%						
Ni†	381.2	0.0203	mg/L	0.00012	0.0203 mg/L	0.00012	0.57%
	QC value within limits for Ni Recovery = 101.32%						
Pb†	85.7	0.0222	mg/L	0.00046	0.0222 mg/L	0.00046	2.08%
	QC value within limits for Pb Recovery = 110.82%						
Sb†	63.4	0.0406	mg/L	0.00112	0.0406 mg/L	0.00112	2.76%
	QC value within limits for Sb Recovery = 81.12%						
Se†	89.9	0.0913	mg/L	0.00234	0.0913 mg/L	0.00234	2.56%
	QC value within limits for Se Recovery = 91.27%						
Tl†	234.3	0.109	mg/L	0.0019	0.109 mg/L	0.0019	1.75%
	QC value within limits for Tl Recovery = 108.78%						
V†	304.3	0.00201	mg/L	0.000034	0.00201 mg/L	0.000034	1.66%
	QC value within limits for V Recovery = 100.70%						
Zn†	767.0	0.0198	mg/L	0.00004	0.0198 mg/L	0.00004	0.19%
	QC value within limits for Zn Recovery = 99.16%						

All analyte(s) passed QC.

User canceled analysis.

=====
Analysis Begun

Start Time: 6/30/2008 22:23:26

Plasma On Time: 6/30/2008 12:36:04

Logged In Analyst: Charley Kay

Technique: ICP Continuous

Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif

Batch ID: 080630A

Results Data Set: 080630

Results Library: C:\pe\Charley Kay\Results\Results.mdb

=====
Sequence No.: 12

Autosampler Location: 18

Sample ID: FILTERCHECK

Date Collected: 6/30/2008 22:23:27

Analyst:

Data Type: Original

Initial Sample Wt:

Initial Sample Vol:

Dilution: 1X

Sample Prep Vol:

Nebulizer Parameters: FILTERCHECK

Analyte

Back Pressure

Flow

All

212.0 kPa

0.65 L/min

Plasma has been extinguished

Vf	78.5	0.00051 mg/L	0.000221	0.00051 mg/L	0.000221	43.42%
Znf	-29.9	-0.00078 mg/L	0.000139	-0.00078 mg/L	0.000139	17.90%

=====
Analysis Begun

Start Time: 6/30/2008 22:29:35
Logged In Analyst: Charley Kay
Spectrometer Model: Optima 4300 DV, S/N 077N2121801

Plasma On Time: 6/30/2008 22:27:31
Technique: ICP Continuous
Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif
Batch ID: 080630A
Results Data Set: 080630
Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 1
Sample ID: ICV
Analyst:
Initial Sample Wt:
Dilution:
User canceled analysis.

Autosampler Location: 15
Date Collected: 6/30/2008 22:29:54
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

=====
Analysis Begun

Start Time: 6/30/2008 22:30:22
Logged In Analyst: Charley Kay
Spectrometer Model: Optima 4300 DV, S/N 077N2121801

Plasma On Time: 6/30/2008 22:27:31
Technique: ICP Continuous

Sample Information File: C:\pe\Charley Kay\Sample Information\080630A.sif
Batch ID: 080630A
Results Data Set: 080630
Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 11
Sample ID: FILTERCHECK
Analyst:
Initial Sample Wt:
Dilution: 1X

Autosampler Location: 18
Date Collected: 6/30/2008 22:30:22
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

=====
Nebulizer Parameters: FILTERCHECK

Analyte Back Pressure Flow
All 212.0 kPa 0.65 L/min

=====
Mean Data: FILTERCHECK

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	382647.5	91.7	%	0.35				0.38%
Yr	454520.2	92.6	%	0.62				0.67%
Agf	27.8	0.00011	mg/L	0.000158	0.00011	mg/L	0.000158	145.87%
Ast	9.3	0.00598	mg/L	0.000564	0.00598	mg/L	0.000564	9.43%
B_t	816.9	0.0304	mg/L	0.00396	0.0304	mg/L	0.00396	13.03%
Baf	-10.0	-0.00027	mg/L	0.000019	-0.00027	mg/L	0.000019	7.05%
Bef	-750.9	-0.00028	mg/L	0.000043	-0.00028	mg/L	0.000043	15.49%
Caf	-202.3	-0.00670	mg/L	0.001927	-0.00670	mg/L	0.001927	28.75%
Cdf	10.8	0.00042	mg/L	0.000134	0.00042	mg/L	0.000134	31.82%
Cof	-9.2	-0.00058	mg/L	0.000088	-0.00058	mg/L	0.000088	15.09%
Crf	26.3	0.00057	mg/L	0.000074	0.00057	mg/L	0.000074	12.89%
Cuf	498.3	0.00167	mg/L	0.000176	0.00167	mg/L	0.000176	10.50%
Fef	-40.5	-0.00356	mg/L	0.000353	-0.00356	mg/L	0.000353	9.91%
Kf	73.1	0.0610	mg/L	0.02722	0.0610	mg/L	0.02722	44.66%
Mgf	-262.6	-0.0123	mg/L	0.00111	-0.0123	mg/L	0.00111	9.04%
Mnf	81.5	0.00016	mg/L	0.000003	0.00016	mg/L	0.000003	1.94%
Mof	6.7	0.00067	mg/L	0.000052	0.00067	mg/L	0.000052	7.76%
Naf	-96.4	-0.0187	mg/L	0.00159	-0.0187	mg/L	0.00159	8.52%
Nif	-11.5	-0.00061	mg/L	0.000170	-0.00061	mg/L	0.000170	27.72%
Pbf	-17.4	-0.00451	mg/L	0.000234	-0.00451	mg/L	0.000234	5.19%
Sbf	0.9	0.00060	mg/L	0.003392	0.00060	mg/L	0.003392	566.82%
Sef	2.8	0.00282	mg/L	0.006925	0.00282	mg/L	0.006925	245.55%
Tlf	-0.9	-0.00041	mg/L	0.000761	-0.00041	mg/L	0.000761	184.93%

Sequence No.: 12
Sample ID: MBLANK2007
Analyst:
Initial Sample Wt:
Dilution: 1X

Autosampler Location: 38
Date Collected: 6/30/2008 22:34:01
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: MBLANK2007
Analyte Back Pressure Flow
All 213.0 kPa 0.65 L/min

Mean Data: MBLANK2007

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	427748.4	103 %	%	0.3			0.27%
Yr	500136.3	102 %	%	0.5			0.52%
Ag†	7.3	0.00003	mg/L	0.000018	0.00003	mg/L	0.000018 61.51%
As†	6.5	0.00423	mg/L	0.000927	0.00423	mg/L	0.000927 21.93%
B†	535.5	0.0199	mg/L	0.00126	0.0199	mg/L	0.00126 6.33%
Ba†	-0.5	-0.00001	mg/L	0.000089	-0.00001	mg/L	0.000089 704.64%
Be†	90.6	0.00003	mg/L	0.000031	0.00003	mg/L	0.000031 91.80%
Ca†	2262.4	0.0750	mg/L	0.00001	0.0750	mg/L	0.00001 0.02%
Cd†	3.5	0.00011	mg/L	0.000026	0.00011	mg/L	0.000026 24.22%
Co†	3.6	0.00023	mg/L	0.000105	0.00023	mg/L	0.000105 45.91%
Cr†	-20.5	-0.00044	mg/L	0.000012	-0.00044	mg/L	0.000012 2.78%
Cu†	-96.9	-0.00033	mg/L	0.000009	-0.00033	mg/L	0.000009 2.73%
Fe†	10.2	0.00089	mg/L	0.000569	0.00089	mg/L	0.000569 63.64%
K†	39.7	0.0331	mg/L	0.00665	0.0331	mg/L	0.00665 20.11%
Mg†	-0.6	-0.00003	mg/L	0.000839	-0.00003	mg/L	0.000839 >999.9%
Mn†	-46.8	-0.00009	mg/L	0.000002	-0.00009	mg/L	0.000002 2.00%
Mo†	0.5	0.00005	mg/L	0.000007	0.00005	mg/L	0.000007 14.73%
Na†	243.1	0.0472	mg/L	0.00603	0.0472	mg/L	0.00603 12.78%
Ni†	-2.3	-0.00012	mg/L	0.000304	-0.00012	mg/L	0.000304 246.17%
Pb†	-1.2	-0.00030	mg/L	0.000444	-0.00030	mg/L	0.000444 147.44%
Sb†	-0.3	-0.00017	mg/L	0.002704	-0.00017	mg/L	0.002704 >999.9%
Se†	0.1	0.00006	mg/L	0.001092	0.00006	mg/L	0.001092 >999.9%
Tl†	-0.2	-0.00008	mg/L	0.002821	-0.00008	mg/L	0.002821 >999.9%
V†	27.0	0.00017	mg/L	0.000181	0.00017	mg/L	0.000181 105.25%
Zn†	14.0	0.00037	mg/L	0.000099	0.00037	mg/L	0.000099 27.11%

Sequence No.: 13
Sample ID: MRL2007
Analyst:
Initial Sample Wt:
Dilution: 1X

Autosampler Location: 24
Date Collected: 6/30/2008 22:37:38
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: MRL2007

Analyte Back Pressure Flow
All 213.0 kPa 0.65 L/min

Mean Data: MRL2007

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	427067.3	102 %	%	0.7			0.65%
Yr	506710.7	103 %	%	1.0			0.95%
Agf	2314.1	0.00914	mg/L	0.000050	0.00914 mg/L	0.000050	0.54%
B_f	1427.6	0.0530	mg/L	0.00024	0.0530 mg/L	0.00024	0.45%
Baf	749.2	0.0200	mg/L	0.00041	0.0200 mg/L	0.00041	2.05%
Bet	2779.7	0.00103	mg/L	0.000004	0.00103 mg/L	0.000004	0.38%
Cat	30547.1	1.01	mg/L	0.006	1.01 mg/L	0.006	0.58%
Cdf	129.3	0.00606	mg/L	0.000106	0.00606 mg/L	0.000106	1.75%
Crf	429.1	0.00929	mg/L	0.000024	0.00929 mg/L	0.000024	0.26%
Cuf	2818.4	0.00946	mg/L	0.000042	0.00946 mg/L	0.000042	0.45%
Fef	233.3	0.0205	mg/L	0.00006	0.0205 mg/L	0.00006	0.29%
Kf	1134.9	0.947	mg/L	0.0263	0.947 mg/L	0.0263	2.78%
Mgf	2158.5	0.101	mg/L	0.0003	0.101 mg/L	0.0003	0.27%
Mnf	898.5	0.00181	mg/L	0.000014	0.00181 mg/L	0.000014	0.80%
Mof	188.5	0.0189	mg/L	0.00038	0.0189 mg/L	0.00038	2.03%
Naf	5020.3	0.974	mg/L	0.0073	0.974 mg/L	0.0073	0.75%
Nif	387.6	0.0206	mg/L	0.00032	0.0206 mg/L	0.00032	1.53%
Vf	312.3	0.00207	mg/L	0.000096	0.00207 mg/L	0.000096	4.66%
Znf	816.9	0.0211	mg/L	0.00022	0.0211 mg/L	0.00022	1.03%

Sequence No.: 14
 Sample ID: LCS2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 39
 Date Collected: 6/30/2008 22:41:17
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS2007

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: LCS2007

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	413601.2	99.1 %	%	0.89				0.90%
Yr	504595.7	103 %	%	0.0				0.01%
Agf	113543.4	0.450 mg/L	mg/L	0.0008	0.450 mg/L		0.0008	0.17%
B_i	12275.1	0.450 mg/L	mg/L	0.0039	0.450 mg/L		0.0039	0.86%
Bat	34584.0	0.925 mg/L	mg/L	0.0087	0.925 mg/L		0.0087	0.94%
Bet	128820.4	0.0477 mg/L	mg/L	0.00020	0.0477 mg/L		0.00020	0.43%
Caf	1424423.6	47.2 mg/L	mg/L	0.05	47.2 mg/L		0.05	0.12%
Cdt	4088.6	0.192 mg/L	mg/L	0.0010	0.192 mg/L		0.0010	0.51%
Crt	43021.1	0.932 mg/L	mg/L	0.0088	0.932 mg/L		0.0088	0.94%
Cuf	275903.1	0.927 mg/L	mg/L	0.0076	0.927 mg/L		0.0076	0.82%
Fef	54501.4	4.78 mg/L	mg/L	0.017	4.78 mg/L		0.017	0.36%
Kf	22274.3	18.6 mg/L	mg/L	0.00	18.6 mg/L		0.00	0.01%
Mgf	408330.4	19.1 mg/L	mg/L	0.02	19.1 mg/L		0.02	0.09%
Mnt	235099.4	0.474 mg/L	mg/L	0.0011	0.474 mg/L		0.0011	0.23%
Mof	9062.8	0.908 mg/L	mg/L	0.0106	0.908 mg/L		0.0106	1.17%
Naf	235481.7	45.7 mg/L	mg/L	0.07	45.7 mg/L		0.07	0.16%
Nif	9117.6	0.485 mg/L	mg/L	0.0063	0.485 mg/L		0.0063	1.30%
Vf	143825.6	0.932 mg/L	mg/L	0.0024	0.932 mg/L		0.0024	0.26%
Znf	37269.1	0.967 mg/L	mg/L	0.0093	0.967 mg/L		0.0093	0.97%

Sequence No.: 15
 Sample ID: LCSD2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 40
 Date Collected: 6/30/2008 22:43:42
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD2007
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: LCSD2007

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
SCa	415294.3	99.5 %	%	0.56				0.56%
Yr	497797.2	101 %	%	0.0				0.01%
Agt	110356.6	0.437 mg/L	mg/L	0.0006	0.437 mg/L		0.0006	0.14%
B_t	11830.0	0.434 mg/L	mg/L	0.0054	0.434 mg/L		0.0054	1.25%
Baf	33659.7	0.900 mg/L	mg/L	0.0042	0.900 mg/L		0.0042	0.47%
Bel	125263.9	0.0464 mg/L	mg/L	0.00021	0.0464 mg/L		0.00021	0.45%
Car	1394773.7	46.2 mg/L	mg/L	0.05	46.2 mg/L		0.05	0.10%
Cdr	3958.1	0.185 mg/L	mg/L	0.0021	0.185 mg/L		0.0021	1.12%
Crf	41515.2	0.899 mg/L	mg/L	0.0051	0.899 mg/L		0.0051	0.57%
Cut	268101.3	0.900 mg/L	mg/L	0.0024	0.900 mg/L		0.0024	0.26%
Fel	52435.5	4.60 mg/L	mg/L	0.006	4.60 mg/L		0.006	0.13%
Kt	21909.4	18.3 mg/L	mg/L	0.12	18.3 mg/L		0.12	0.64%
Mgt	395447.2	18.5 mg/L	mg/L	0.00	18.5 mg/L		0.00	0.01%
Mnt	229187.5	0.462 mg/L	mg/L	0.0004	0.462 mg/L		0.0004	0.09%
Mof	8767.4	0.879 mg/L	mg/L	0.0055	0.879 mg/L		0.0055	0.62%
Nat	232208.2	45.1 mg/L	mg/L	0.29	45.1 mg/L		0.29	0.64%
Nit	8761.1	0.466 mg/L	mg/L	0.0069	0.466 mg/L		0.0069	1.49%
Vt	139672.5	0.905 mg/L	mg/L	0.0029	0.905 mg/L		0.0029	0.32%
Znt	35948.0	0.933 mg/L	mg/L	0.0097	0.933 mg/L		0.0097	1.04%

Sequence No.: 17
 Sample ID: 2806190595MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 42
 Date Collected: 6/30/2008 22:49:08
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806190595MS
 Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Mean Data: 2806190595MS

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Conc. Units	Sample Std.Dev.	RSD
Sca	407510.0	97.7	%	0.54			0.55%
Yr	494980.4	101	%	1.6			1.57%
Ag†	113405.2	0.450	mg/L	0.0007	0.450 mg/L	0.0007	0.15%
B_†	14810.7	0.544	mg/L	0.0071	0.544 mg/L	0.0071	1.31%
Ba†	36645.6	0.980	mg/L	0.0064	0.980 mg/L	0.0064	0.65%
Be†	129236.1	0.0478	mg/L	0.00021	0.0478 mg/L	0.00021	0.45%
Ca†	2159417.1	71.5	mg/L	0.81	71.5 mg/L	0.81	1.13%
Cd†	4159.8	0.195	mg/L	0.0006	0.195 mg/L	0.0006	0.30%
Cr†	43284.9	0.937	mg/L	0.0031	0.937 mg/L	0.0031	0.33%
Cu†	308914.2	1.04	mg/L	0.002	1.04 mg/L	0.002	0.15%
Fe†	64430.0	5.66	mg/L	0.021	5.66 mg/L	0.021	0.37%
K†	24140.0	20.1	mg/L	0.26	20.1 mg/L	0.26	1.27%
Mg†	492884.8	23.1	mg/L	0.16	23.1 mg/L	0.16	0.67%
Mn†	269236.0	0.543	mg/L	0.0007	0.543 mg/L	0.0007	0.12%
Mo†	9197.1	0.922	mg/L	0.0063	0.922 mg/L	0.0063	0.69%
Nat	365305.1	70.9	mg/L	0.93	70.9 mg/L	0.93	1.31%
Ni†	9031.5	0.480	mg/L	0.0037	0.480 mg/L	0.0037	0.77%
V†	143126.5	0.927	mg/L	0.0039	0.927 mg/L	0.0039	0.43%
Zn†	39348.9	1.02	mg/L	0.009	1.02 mg/L	0.009	0.87%

Sequence No.: 18
 Sample ID: 2806190595MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 43
 Date Collected: 6/30/2008 22:52:14
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806190595MSD
 Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Mean Data: 2806190595MSD

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	405385.9	97.2	%	0.21			0.22%
Yr	493479.5	101	%	1.5			1.54%
Ag†	116814.0	0.463	mg/L	0.0003	0.463 mg/L	0.0003	0.06%
B_†	15394.1	0.565	mg/L	0.0037	0.565 mg/L	0.0037	0.66%
Ba†	38053.8	1.02	mg/L	0.006	1.02 mg/L	0.006	0.62%
Be†	133025.7	0.0492	mg/L	0.00011	0.0492 mg/L	0.00011	0.23%
Ca†	2212570.8	73.3	mg/L	1.47	73.3 mg/L	1.47	2.00%
Cd†	4281.6	0.201	mg/L	0.0004	0.201 mg/L	0.0004	0.19%
Cr†	45330.5	0.982	mg/L	0.0038	0.982 mg/L	0.0038	0.39%
Cu†	317632.9	1.07	mg/L	0.002	1.07 mg/L	0.002	0.15%
Fe†	66094.3	5.80	mg/L	0.027	5.80 mg/L	0.027	0.47%
K†	24632.7	20.5	mg/L	0.10	20.5 mg/L	0.10	0.47%
Mg†	501265.5	23.5	mg/L	0.10	23.5 mg/L	0.10	0.42%
Mn†	276508.7	0.558	mg/L	0.0002	0.558 mg/L	0.0002	0.03%
Mo†	9593.4	0.961	mg/L	0.0012	0.961 mg/L	0.0012	0.13%
Na†	371725.1	72.1	mg/L	0.28	72.1 mg/L	0.28	0.39%
Ni†	9378.9	0.499	mg/L	0.0001	0.499 mg/L	0.0001	0.01%
V†	147521.0	0.956	mg/L	0.0020	0.956 mg/L	0.0020	0.21%
Zn†	40804.4	1.06	mg/L	0.001	1.06 mg/L	0.001	0.10%

Sequence No.: 20
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/30/2008 22:59:37
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	397605.0	95.3	%	0.45			0.47%
Yr	486921.1	99.2	%	0.85			0.85%
Ag†	243142.3	0.961	mg/L	0.0031	0.961 mg/L	0.0031	0.32%
	QC value within limits for Ag	Recovery =	96.12%				
B_†	67165.6	2.46	mg/L	0.014	2.46 mg/L	0.014	0.56%
	QC value within limits for B_	Recovery =	98.57%				
Ba†	188573.5	5.04	mg/L	0.006	5.04 mg/L	0.006	0.13%
	QC value within limits for Ba	Recovery =	100.84%				
Be†	5489626.9	2.03	mg/L	0.009	2.03 mg/L	0.009	0.42%
	QC value within limits for Be	Recovery =	101.34%				
Ca†	1529489.6	50.7	mg/L	0.07	50.7 mg/L	0.07	0.14%
	QC value within limits for Ca	Recovery =	101.34%				
Cd†	42671.8	2.00	mg/L	0.001	2.00 mg/L	0.001	0.03%
	QC value within limits for Cd	Recovery =	99.95%				
Cr†	228330.3	4.94	mg/L	0.018	4.94 mg/L	0.018	0.36%
	QC value within limits for Cr	Recovery =	98.88%				
Cu†	1470102.8	4.94	mg/L	0.045	4.94 mg/L	0.045	0.92%
	QC value within limits for Cu	Recovery =	98.74%				
Fe†	58076.7	5.10	mg/L	0.013	5.10 mg/L	0.013	0.26%
	QC value within limits for Fe	Recovery =	101.97%				
K†	59107.0	49.3	mg/L	0.21	49.3 mg/L	0.21	0.43%
	QC value within limits for K	Recovery =	98.60%				
Mg†	1065794.1	50.0	mg/L	0.02	50.0 mg/L	0.02	0.05%
	QC value within limits for Mg	Recovery =	99.93%				
Mn†	2500523.3	5.05	mg/L	0.047	5.05 mg/L	0.047	0.92%
	QC value within limits for Mn	Recovery =	100.92%				
Mo†	49074.8	4.92	mg/L	0.013	4.92 mg/L	0.013	0.27%
	QC value within limits for Mo	Recovery =	98.36%				
Na†	248030.7	48.1	mg/L	0.02	48.1 mg/L	0.02	0.03%
	QC value within limits for Na	Recovery =	96.24%				
Ni†	97283.8	5.17	mg/L	0.009	5.17 mg/L	0.009	0.18%
	QC value within limits for Ni	Recovery =	103.43%				
V†	767856.4	4.97	mg/L	0.018	4.97 mg/L	0.018	0.37%
	QC value within limits for V	Recovery =	99.46%				
Zn†	198055.8	5.12	mg/L	0.011	5.12 mg/L	0.011	0.21%
	QC value within limits for Zn	Recovery =	102.43%				

All analyte(s) passed QC.

Sequence No.: 21
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/30/2008 23:02:47
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	427950.8	103	%	0.4			0.42%
Yr	499430.7	102	%	0.8			0.76%
Ag†	-62.2	-0.00025	mg/L	0.000085	-0.00025 mg/L	0.000085	34.73%
	QC value within limits for Ag Recovery = Not calculated						
B_†	304.9	0.0113	mg/L	0.00065	0.0113 mg/L	0.00065	5.76%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	3.0	0.00008	mg/L	0.000033	0.00008 mg/L	0.000033	41.08%
	QC value within limits for Ba Recovery = Not calculated						
Be†	133.3	0.00005	mg/L	0.000018	0.00005 mg/L	0.000018	37.28%
	QC value within limits for Be Recovery = Not calculated						
Ca†	59.7	0.00198	mg/L	0.001197	0.00198 mg/L	0.001197	60.46%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	7.4	0.00035	mg/L	0.000188	0.00035 mg/L	0.000188	54.08%
	QC value within limits for Cd Recovery = Not calculated						
Cr†	-14.8	-0.00032	mg/L	0.000012	-0.00032 mg/L	0.000012	3.63%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-201.7	-0.00068	mg/L	0.000018	-0.00068 mg/L	0.000018	2.63%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	5.3	0.00047	mg/L	0.000443	0.00047 mg/L	0.000443	95.00%
	QC value within limits for Fe Recovery = Not calculated						
K†	11.7	0.00976	mg/L	0.003167	0.00976 mg/L	0.003167	32.45%
	QC value within limits for K Recovery = Not calculated						
Mg†	-4.5	-0.00021	mg/L	0.000152	-0.00021 mg/L	0.000152	71.62%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-99.0	-0.00020	mg/L	0.000004	-0.00020 mg/L	0.000004	1.90%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	24.1	0.00241	mg/L	0.000078	0.00241 mg/L	0.000078	3.22%
	QC value within limits for Mo Recovery = Not calculated						
Na†	76.9	0.0149	mg/L	0.00634	0.0149 mg/L	0.00634	42.50%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-3.8	-0.00020	mg/L	0.000196	-0.00020 mg/L	0.000196	97.94%
	QC value within limits for Ni Recovery = Not calculated						
V†	27.3	0.00018	mg/L	0.000020	0.00018 mg/L	0.000020	11.60%
	QC value within limits for V Recovery = Not calculated						
Zn†	-36.3	-0.00095	mg/L	0.000106	-0.00095 mg/L	0.000106	11.22%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 22
 Sample ID: 2806250566MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 45
 Date Collected: 6/30/2008 23:06:10
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806250566MS

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: 2806250566MS

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	394789.5	94.6	%	0.24				0.26%
Yr	481414.4	98.1	%	1.21				1.23%
Ag†	116787.5	0.463	mg/L	0.0010	0.463	mg/L	0.0010	0.21%
B_†	19430.2	0.716	mg/L	0.0024	0.716	mg/L	0.0024	0.33%
Ba†	40066.7	1.07	mg/L	0.007	1.07	mg/L	0.007	0.67%
Be†	133657.9	0.0495	mg/L	0.00010	0.0495	mg/L	0.00010	0.21%
Ca†	2064981.6	68.4	mg/L	1.48	68.4	mg/L	1.48	2.17%
Cd†	4217.2	0.198	mg/L	0.0009	0.198	mg/L	0.0009	0.47%
Cr†	44409.7	0.962	mg/L	0.0048	0.962	mg/L	0.0048	0.50%
Cu†	294014.4	0.987	mg/L	0.0025	0.987	mg/L	0.0025	0.26%
Fe†	56959.3	5.00	mg/L	0.042	5.00	mg/L	0.042	0.83%
K†	39302.9	32.8	mg/L	0.11	32.8	mg/L	0.11	0.34%
Mg†	629845.3	29.5	mg/L	0.05	29.5	mg/L	0.05	0.19%
Mn†	246282.6	0.496	mg/L	0.0004	0.496	mg/L	0.0004	0.08%
Mo†	9410.0	0.943	mg/L	0.0019	0.943	mg/L	0.0019	0.21%
Na†	557990.1	108	mg/L	0.4	108	mg/L	0.4	0.39%
Ni†	9332.9	0.496	mg/L	0.0014	0.496	mg/L	0.0014	0.29%
V†	148543.2	0.962	mg/L	0.0018	0.962	mg/L	0.0018	0.18%
Zn†	41731.4	1.08	mg/L	0.005	1.08	mg/L	0.005	0.44%

Sequence No.: 23
 Sample ID: 2806250566MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 46
 Date Collected: 6/30/2008 23:09:16
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806250566MSD
 Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Mean Data: 2806250566MSD

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Conc. Units	Sample Std.Dev.	RSD
Sca	404163.5	96.9	%	0.65			0.67%
Yr	488544.3	99.6	%	1.77			1.78%
Ag†	115887.9	0.459	mg/L	0.0024	0.459 mg/L	0.0024	0.52%
B_†	18983.6	0.699	mg/L	0.0017	0.699 mg/L	0.0017	0.25%
Ba†	39148.1	1.05	mg/L	0.000	1.05 mg/L	0.000	0.00%
Be†	131543.2	0.0487	mg/L	0.00053	0.0487 mg/L	0.00053	1.09%
Ca†	2036959.3	67.5	mg/L	0.86	67.5 mg/L	0.86	1.27%
Cd†	4169.6	0.195	mg/L	0.0012	0.195 mg/L	0.0012	0.60%
Cr†	43496.5	0.942	mg/L	0.0023	0.942 mg/L	0.0023	0.24%
Cu†	292889.2	0.984	mg/L	0.0032	0.984 mg/L	0.0032	0.33%
Fe†	56260.9	4.94	mg/L	0.040	4.94 mg/L	0.040	0.81%
K†	38428.8	32.1	mg/L	0.42	32.1 mg/L	0.42	1.31%
Mg†	616969.5	28.9	mg/L	0.14	28.9 mg/L	0.14	0.47%
Mn†	243071.8	0.490	mg/L	0.0038	0.490 mg/L	0.0038	0.78%
Mo†	9282.5	0.930	mg/L	0.0005	0.930 mg/L	0.0005	0.06%
Na†	546063.9	106	mg/L	1.1	106 mg/L	1.1	1.04%
Ni†	9127.3	0.485	mg/L	0.0014	0.485 mg/L	0.0014	0.29%
V†	146368.4	0.948	mg/L	0.0089	0.948 mg/L	0.0089	0.94%
Zn†	40831.0	1.06	mg/L	0.001	1.06 mg/L	0.001	0.14%

Sequence No.: 32
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/30/2008 23:43:19
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	401400.6	96.2	%	1.22			1.26%
Yr	487223.5	99.3	%	0.83			0.83%
Ag†	242465.5	0.959	mg/L	0.0021	0.959 mg/L	0.0021	0.22%
	QC value within limits for Ag		Recovery = 95.85%				
B_†	67200.3	2.47	mg/L	0.001	2.47 mg/L	0.001	0.03%
	QC value within limits for B_		Recovery = 98.64%				
Ba†	188680.4	5.04	mg/L	0.006	5.04 mg/L	0.006	0.13%
	QC value within limits for Ba		Recovery = 100.89%				
Be†	5345151.0	1.97	mg/L	0.026	1.97 mg/L	0.026	1.33%
	QC value within limits for Be		Recovery = 98.68%				
Ca†	1514609.9	50.2	mg/L	0.07	50.2 mg/L	0.07	0.13%
	QC value within limits for Ca		Recovery = 100.36%				
Cd†	42506.4	1.99	mg/L	0.005	1.99 mg/L	0.005	0.26%
	QC value within limits for Cd		Recovery = 99.56%				
Cr†	226289.8	4.90	mg/L	0.013	4.90 mg/L	0.013	0.26%
	QC value within limits for Cr		Recovery = 98.00%				
Cu†	1454268.2	4.88	mg/L	0.001	4.88 mg/L	0.001	0.02%
	QC value within limits for Cu		Recovery = 97.68%				
Fe†	57256.5	5.03	mg/L	0.011	5.03 mg/L	0.011	0.22%
	QC value within limits for Fe		Recovery = 100.53%				
K†	59724.6	49.8	mg/L	0.05	49.8 mg/L	0.05	0.10%
	QC value within limits for K		Recovery = 99.63%				
Mg†	1056969.2	49.5	mg/L	0.01	49.5 mg/L	0.01	0.02%
	QC value within limits for Mg		Recovery = 99.10%				
Mn†	2444277.1	4.93	mg/L	0.004	4.93 mg/L	0.004	0.09%
	QC value within limits for Mn		Recovery = 98.65%				
Mo†	48736.3	4.88	mg/L	0.005	4.88 mg/L	0.005	0.11%
	QC value within limits for Mo		Recovery = 97.68%				
Na†	248436.0	48.2	mg/L	0.03	48.2 mg/L	0.03	0.06%
	QC value within limits for Na		Recovery = 96.40%				
Ni†	96188.9	5.11	mg/L	0.024	5.11 mg/L	0.024	0.48%
	QC value within limits for Ni		Recovery = 102.26%				
V†	761928.4	4.93	mg/L	0.016	4.93 mg/L	0.016	0.33%
	QC value within limits for V		Recovery = 98.69%				
Zn†	196451.8	5.08	mg/L	0.019	5.08 mg/L	0.019	0.37%
	QC value within limits for Zn		Recovery = 101.60%				

All analyte(s) passed QC.

Sequence No.: 33
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/30/2008 23:46:29
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	428441.1	103 %		0.4			0.35%
Yr	503502.0	103 %		1.4			1.34%
Ag†	-83.6	-0.00033	mg/L	0.000248	-0.00033 mg/L	0.000248	75.14%
	QC value within limits for Ag Recovery = Not calculated						
B_†	497.3	0.0185	mg/L	0.00126	0.0185 mg/L	0.00126	6.84%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	2.1	0.00006	mg/L	0.000013	0.00006 mg/L	0.000013	23.84%
	QC value within limits for Ba Recovery = Not calculated						
Be†	189.2	0.00007	mg/L	0.000027	0.00007 mg/L	0.000027	38.14%
	QC value within limits for Be Recovery = Not calculated						
Ca†	18.8	0.00062	mg/L	0.001327	0.00062 mg/L	0.001327	213.22%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	2.9	0.00014	mg/L	0.000103	0.00014 mg/L	0.000103	75.54%
	QC value within limits for Cd Recovery = Not calculated						
Cr†	-12.9	-0.00028	mg/L	0.000081	-0.00028 mg/L	0.000081	29.11%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-283.6	-0.00095	mg/L	0.000192	-0.00095 mg/L	0.000192	20.17%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	8.1	0.00071	mg/L	0.000334	0.00071 mg/L	0.000334	47.10%
	QC value within limits for Fe Recovery = Not calculated						
K†	50.4	0.0421	mg/L	0.00071	0.0421 mg/L	0.00071	1.69%
	QC value within limits for K Recovery = Not calculated						
Mg†	20.3	0.00095	mg/L	0.000554	0.00095 mg/L	0.000554	58.19%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	112.8	0.00023	mg/L	0.000016	0.00023 mg/L	0.000016	7.07%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	20.9	0.00209	mg/L	0.000049	0.00209 mg/L	0.000049	2.35%
	QC value within limits for Mo Recovery = Not calculated						
Na†	490.6	0.0952	mg/L	0.00146	0.0952 mg/L	0.00146	1.53%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-0.1	0.00000	mg/L	0.000157	0.00000 mg/L	0.000157	>999.9%
	QC value within limits for Ni Recovery = Not calculated						
V†	9.3	0.00006	mg/L	0.000113	0.00006 mg/L	0.000113	190.58%
	QC value within limits for V Recovery = Not calculated						
Zn†	-46.0	-0.00120	mg/L	0.000123	-0.00120 mg/L	0.000123	10.25%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 34
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 6/30/2008 23:49:52
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	416072.8	99.7	%	0.31			0.31%
Yr	494889.7	101	%	1.0			0.96%
Ag†	120529.5	0.477	mg/L	0.0006	0.477 mg/L	0.0006	0.12%
	QC value within limits for Ag Recovery = 95.30%						
B_†	33090.4	1.21	mg/L	0.006	1.21 mg/L	0.006	0.45%
	QC value within limits for B_ Recovery = 97.12%						
Ba†	94846.7	2.54	mg/L	0.001	2.54 mg/L	0.001	0.03%
	QC value within limits for Ba Recovery = 101.44%						
Be†	2687420.3	0.992	mg/L	0.0003	0.992 mg/L	0.0003	0.03%
	QC value within limits for Be Recovery = 99.23%						
Ca†	764809.5	25.3	mg/L	0.01	25.3 mg/L	0.01	0.03%
	QC value within limits for Ca Recovery = 101.35%						
Cd†	21081.2	0.988	mg/L	0.0031	0.988 mg/L	0.0031	0.31%
	QC value within limits for Cd Recovery = 98.75%						
Cr†	113123.3	2.45	mg/L	0.021	2.45 mg/L	0.021	0.85%
	QC value within limits for Cr Recovery = 97.98%						
Cu†	729865.3	2.45	mg/L	0.001	2.45 mg/L	0.001	0.04%
	QC value within limits for Cu Recovery = 98.04%						
Fe†	28943.8	2.54	mg/L	0.003	2.54 mg/L	0.003	0.13%
	QC value within limits for Fe Recovery = 101.64%						
K†	29316.1	24.5	mg/L	0.15	24.5 mg/L	0.15	0.63%
	QC value within limits for K Recovery = 97.81%						
Mg†	534861.1	25.1	mg/L	0.01	25.1 mg/L	0.01	0.02%
	QC value within limits for Mg Recovery = 100.29%						
Mn†	1241427.9	2.51	mg/L	0.004	2.51 mg/L	0.004	0.15%
	QC value within limits for Mn Recovery = 100.21%						
Mo†	24237.0	2.43	mg/L	0.014	2.43 mg/L	0.014	0.58%
	QC value within limits for Mo Recovery = 97.15%						
Na†	122570.1	23.8	mg/L	0.11	23.8 mg/L	0.11	0.45%
	QC value within limits for Na Recovery = 95.12%						
Ni†	48625.5	2.58	mg/L	0.015	2.58 mg/L	0.015	0.57%
	QC value within limits for Ni Recovery = 103.39%						
V†	378939.6	2.45	mg/L	0.009	2.45 mg/L	0.009	0.36%
	QC value within limits for V Recovery = 98.17%						
Zn†	98507.8	2.55	mg/L	0.008	2.55 mg/L	0.008	0.30%
	QC value within limits for Zn Recovery = 101.89%						

All analyte(s) passed QC.

Sequence No.: 45
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 7/1/2008 00:29:58
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	403188.6	96.6	%	0.43			0.44%
Yr	477599.9	97.3	%	0.16			0.16%
Ag†	239996.9	0.949	mg/L	0.0006	0.949 mg/L	0.0006	0.06%
	QC value within limits for Ag	Recovery =	94.88%				
B_†	66570.2	2.44	mg/L	0.004	2.44 mg/L	0.004	0.14%
	QC value within limits for B_	Recovery =	97.70%				
Ba†	189649.0	5.07	mg/L	0.021	5.07 mg/L	0.021	0.41%
	QC value within limits for Ba	Recovery =	101.41%				
Be†	5291472.2	1.95	mg/L	0.001	1.95 mg/L	0.001	0.06%
	QC value within limits for Be	Recovery =	97.69%				
Ca†	1503160.0	49.8	mg/L	0.01	49.8 mg/L	0.01	0.02%
	QC value within limits for Ca	Recovery =	99.60%				
Cd†	42190.2	1.98	mg/L	0.002	1.98 mg/L	0.002	0.09%
	QC value within limits for Cd	Recovery =	98.82%				
Cr†	223646.2	4.84	mg/L	0.013	4.84 mg/L	0.013	0.27%
	QC value within limits for Cr	Recovery =	96.86%				
Cu†	1452150.6	4.88	mg/L	0.028	4.88 mg/L	0.028	0.57%
	QC value within limits for Cu	Recovery =	97.53%				
Fe†	56239.3	4.94	mg/L	0.007	4.94 mg/L	0.007	0.14%
	QC value within limits for Fe	Recovery =	98.74%				
K†	60679.9	50.6	mg/L	0.15	50.6 mg/L	0.15	0.30%
	QC value within limits for K	Recovery =	101.22%				
Mg†	1037455.4	48.6	mg/L	0.21	48.6 mg/L	0.21	0.44%
	QC value within limits for Mg	Recovery =	97.27%				
Mn†	2432644.9	4.91	mg/L	0.022	4.91 mg/L	0.022	0.45%
	QC value within limits for Mn	Recovery =	98.18%				
Mo†	48357.7	4.85	mg/L	0.010	4.85 mg/L	0.010	0.20%
	QC value within limits for Mo	Recovery =	96.92%				
Na†	248625.5	48.2	mg/L	0.09	48.2 mg/L	0.09	0.19%
	QC value within limits for Na	Recovery =	96.48%				
Ni†	95703.9	5.09	mg/L	0.000	5.09 mg/L	0.000	0.00%
	QC value within limits for Ni	Recovery =	101.75%				
V†	755143.6	4.89	mg/L	0.006	4.89 mg/L	0.006	0.13%
	QC value within limits for V	Recovery =	97.81%				
Zn†	195010.4	5.04	mg/L	0.005	5.04 mg/L	0.005	0.10%
	QC value within limits for Zn	Recovery =	100.85%				

All analyte(s) passed QC.

Sequence No.: 46
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 7/1/2008 00:33:11
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	429303.0	103 %		1.1			1.08%
Yr	493294.6	101 %		0.8			0.83%
Agf	-116.2	-0.00046	mg/L	0.000097	-0.00046 mg/L	0.000097	21.22%
	QC value within limits for Ag Recovery = Not calculated						
B_†	654.4	0.0243	mg/L	0.00101	0.0243 mg/L	0.00101	4.13%
	QC value greater than the upper limit for B_ Recovery = Not calculated						
Baf	2.9	0.00008	mg/L	0.000000	0.00008 mg/L	0.000000	0.24%
	QC value within limits for Ba Recovery = Not calculated						
Be†	122.0	0.00004	mg/L	0.000011	0.00004 mg/L	0.000011	24.46%
	QC value within limits for Be Recovery = Not calculated						
Ca†	13.4	0.00044	mg/L	0.000786	0.00044 mg/L	0.000786	176.99%
	QC value within limits for Ca Recovery = Not calculated						
Cd†	4.3	0.00020	mg/L	0.000190	0.00020 mg/L	0.000190	95.38%
	QC value within limits for Cd Recovery = Not calculated						
Crt	-17.7	-0.00038	mg/L	0.000011	-0.00038 mg/L	0.000011	2.98%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-321.5	-0.00108	mg/L	0.000213	-0.00108 mg/L	0.000213	19.77%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	11.1	0.00097	mg/L	0.000907	0.00097 mg/L	0.000907	93.45%
	QC value within limits for Fe Recovery = Not calculated						
K†	14.2	0.0119	mg/L	0.01861	0.0119 mg/L	0.01861	156.87%
	QC value within limits for K Recovery = Not calculated						
Mg†	22.2	0.00104	mg/L	0.000161	0.00104 mg/L	0.000161	15.42%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	5.1	0.00001	mg/L	0.000011	0.00001 mg/L	0.000011	110.27%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	19.8	0.00199	mg/L	0.000559	0.00199 mg/L	0.000559	28.13%
	QC value within limits for Mo Recovery = Not calculated						
Na†	426.9	0.0828	mg/L	0.00480	0.0828 mg/L	0.00480	5.79%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-3.6	-0.00019	mg/L	0.000060	-0.00019 mg/L	0.000060	31.52%
	QC value within limits for Ni Recovery = Not calculated						
V†	26.0	0.00017	mg/L	0.000004	0.00017 mg/L	0.000004	2.46%
	QC value within limits for V Recovery = Not calculated						
Zn†	-53.4	-0.00139	mg/L	0.000147	-0.00139 mg/L	0.000147	10.54%
	QC value within limits for Zn Recovery = Not calculated						
QC Failed. Retry.							

Sequence No.: 47
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 7/1/2008 00:35:44
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	428040.6	103 %		0.8			0.83%

Yr	486786.7	99.2 %	1.36			1.37%
Agf	-142.6	-0.00056 mg/L	0.000044	-0.00056 mg/L	0.000044	7.85%
	QC value within limits for Ag	Recovery =	Not calculated			
B_t	429.3	0.0160 mg/L	0.00052	0.0160 mg/L	0.00052	3.26%
	QC value within limits for B_	Recovery =	Not calculated			
Baf	2.7	0.00007 mg/L	0.000155	0.00007 mg/L	0.000155	215.85%
	QC value within limits for Ba	Recovery =	Not calculated			
Bef	77.2	0.00003 mg/L	0.000013	0.00003 mg/L	0.000013	44.93%
	QC value within limits for Be	Recovery =	Not calculated			
Caf	38.5	0.00127 mg/L	0.001156	0.00127 mg/L	0.001156	90.68%
	QC value within limits for Ca	Recovery =	Not calculated			
Cdf	3.5	0.00016 mg/L	0.000053	0.00016 mg/L	0.000053	32.92%
	QC value within limits for Cd	Recovery =	Not calculated			
Crf	-17.2	-0.00037 mg/L	0.000168	-0.00037 mg/L	0.000168	44.87%
	QC value within limits for Cr	Recovery =	Not calculated			
Cuf	-396.5	-0.00133 mg/L	0.000014	-0.00133 mg/L	0.000014	1.08%
	QC value within limits for Cu	Recovery =	Not calculated			
Fef	1.8	0.00015 mg/L	0.000507	0.00015 mg/L	0.000507	327.49%
	QC value within limits for Fe	Recovery =	Not calculated			
Kf	29.2	0.0244 mg/L	0.00269	0.0244 mg/L	0.00269	11.05%
	QC value within limits for K	Recovery =	Not calculated			
Mgf	2.8	0.00013 mg/L	0.000020	0.00013 mg/L	0.000020	14.75%
	QC value within limits for Mg	Recovery =	Not calculated			
Mnf	-26.1	-0.00005 mg/L	0.000016	-0.00005 mg/L	0.000016	30.44%
	QC value within limits for Mn	Recovery =	Not calculated			
Mof	8.9	0.00089 mg/L	0.000099	0.00089 mg/L	0.000099	11.01%
	QC value within limits for Mo	Recovery =	Not calculated			
Naf	404.3	0.0784 mg/L	0.00661	0.0784 mg/L	0.00661	8.42%
	QC value within limits for Na	Recovery =	Not calculated			
Nif	-1.2	-0.00006 mg/L	0.000202	-0.00006 mg/L	0.000202	316.05%
	QC value within limits for Ni	Recovery =	Not calculated			
Vf	23.1	0.00015 mg/L	0.000035	0.00015 mg/L	0.000035	23.62%
	QC value within limits for V	Recovery =	Not calculated			
Znf	-52.9	-0.00138 mg/L	0.000115	-0.00138 mg/L	0.000115	8.33%
	QC value within limits for Zn	Recovery =	Not calculated			

All analyte(s) passed QC.

Sequence No.: 16
 Sample ID: 2806190595
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 41
 Date Collected: 6/30/2008 22:46:07
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806190595
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: 2806190595

Analyte	Mean Corrected		Calib. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	425348.1	102 %	%	0.3				0.25%
Yr	513023.6	105 %	%	0.4				0.35%
Ag†	-16.7	0.00025	mg/L	0.000591	0.00025	mg/L	0.000591	235.37%
B †	2091.8	0.0768	mg/L	0.00034	0.0768	mg/L	0.00034	0.44%
Ba†	821.7	0.0220	mg/L	0.00012	0.0220	mg/L	0.00012	0.53%
Be†	23.9	0.00001	mg/L	0.000001	0.00001	mg/L	0.000001	9.89%
Cat	744173.0	24.7	mg/L	0.04	24.7	mg/L	0.04	0.16%
Cd†	-2.1	-0.00010	mg/L	0.000083	-0.00010	mg/L	0.000083	82.07%
Cr†	-19.4	-0.00042	mg/L	0.000079	-0.00042	mg/L	0.000079	18.81%
Cu†	35980.6	0.121	mg/L	0.0001	0.121	mg/L	0.0001	0.12%
Fe†	9814.3	0.862	mg/L	0.0065	0.862	mg/L	0.0065	0.75%
K†	1778.3	1.48	mg/L	0.006	1.48	mg/L	0.006	0.42%
Mg†	86115.7	4.04	mg/L	0.023	4.04	mg/L	0.023	0.56%
Mn†	33286.3	0.0671	mg/L	0.00019	0.0671	mg/L	0.00019	0.29%
Mo†	49.0	0.00491	mg/L	0.000051	0.00491	mg/L	0.000051	1.04%
Nat	134295.4	26.1	mg/L	0.16	26.1	mg/L	0.16	0.62%
Ni†	-9.8	-0.00052	mg/L	0.000494	-0.00052	mg/L	0.000494	94.68%
V†	32.3	0.00024	mg/L	0.000014	0.00024	mg/L	0.000014	5.73%
Zn†	1725.8	0.0450	mg/L	0.00000	0.0450	mg/L	0.00000	0.01%

Sequence No.: 29
 Sample ID: 2806240538_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 52
 Date Collected: 6/30/2008 23:31:30
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806240538_2X
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: 2806240538_2X

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	370158.2	88.7	%	0.61			0.69%
Yr	469905.3	95.8	%	0.51			0.53%
Agf	-368.0	-0.00075	mg/L	0.000143	-0.00150 mg/L	0.000286	19.07%
B_tf	43996.0	1.64	mg/L	0.014	3.27 mg/L	0.028	0.84%
Baf	510.8	0.0137	mg/L	0.00005	0.0273 mg/L	0.00009	0.34%
Bef	-935.3	-0.00034	mg/L	0.000017	-0.00069 mg/L	0.000033	4.84%
Ca	5961023.9	197	mg/L	0.4	395 mg/L	0.9	0.22%
Cdf	-14.6	-0.00068	mg/L	0.000144	-0.00136 mg/L	0.000287	21.06%
Crf	174.8	0.00379	mg/L	0.000022	0.00757 mg/L	0.000043	0.57%
Cuf	617.7	0.00208	mg/L	0.000103	0.00416 mg/L	0.000206	4.96%
Fef	21704.1	1.91	mg/L	0.013	3.81 mg/L	0.026	0.68%
Kf	18457.4	15.4	mg/L	0.07	30.8 mg/L	0.14	0.46%
Mgf	1984882.7	93.0	mg/L	0.25	186 mg/L	0.5	0.27%
Mnf	248925.8	0.500	mg/L	0.0005	1.00 mg/L	0.001	0.10%
Mof	1070.5	0.107	mg/L	0.0007	0.215 mg/L	0.0015	0.68%
Naf	3417796.0	663	mg/L	3.1	1330 mg/L	6.2	0.47%
Nif	523.5	0.0278	mg/L	0.00027	0.0557 mg/L	0.00054	0.98%
Vf	2340.9	0.0152	mg/L	0.00003	0.0304 mg/L	0.00006	0.21%
Znf	310.6	0.00789	mg/L	0.000183	0.0158 mg/L	0.00037	2.31%

Sequence No.: 30
 Sample ID: 2806240540_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 53
 Date Collected: 6/30/2008 23:35:13
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2806240540_2X
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: 2806240540_2X

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	366394.9	87.8	%	0.96			1.09%
Yr	461429.5	94.0	%	0.12			0.12%
Agf	-208.7	-0.00082	mg/L	0.000185	-0.00163 mg/L	0.000370	22.63%
B_i	46071.1	1.71	mg/L	0.012	3.42 mg/L	0.023	0.68%
Baf	478.2	0.0128	mg/L	0.00005	0.0256 mg/L	0.00010	0.41%
Bel	-973.5	-0.00036	mg/L	0.000004	-0.00072 mg/L	0.000009	1.22%
Cal	6073167.4	201	mg/L	1.5	402 mg/L	2.9	0.73%
Cdf	-3.2	-0.00015	mg/L	0.000128	-0.00030 mg/L	0.000255	84.94%
Crf	498.0	0.0108	mg/L	0.00002	0.0216 mg/L	0.00003	0.15%
Cuf	758.3	0.00255	mg/L	0.000084	0.00511 mg/L	0.000168	3.29%
Fef	204.6	0.0180	mg/L	0.00036	0.0359 mg/L	0.00071	1.98%
Kf	19546.0	16.3	mg/L	0.00	32.6 mg/L	0.01	0.02%
Mgf	2027882.9	95.0	mg/L	0.88	190 mg/L	1.8	0.93%
Mnf	227501.1	0.457	mg/L	0.0038	0.914 mg/L	0.0076	0.83%
Mof	385.3	0.0386	mg/L	0.00109	0.0772 mg/L	0.00218	2.82%
Naf	3401359.6	660	mg/L	5.4	1320 mg/L	10.8	0.81%
Nif	122.8	0.00653	mg/L	0.000244	0.0131 mg/L	0.00049	3.74%
Vf	3652.4	0.0236	mg/L	0.00032	0.0472 mg/L	0.00064	1.36%
Znf	269.9	0.00697	mg/L	0.000060	0.0139 mg/L	0.00012	0.86%

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Lot #
 HNO3 R# 100450 HCL R# 100446
 IS = Yttrium(ME0709008)0.75mL + Scandium (ME0710007)0.5mL to 1000mL w/ 2% HNO3

Standards	Lot #	Exp. Date	Dilution
Calibration (Prepare daily)	ME0712001	(12/01/08)	1:10 ME0801001
	ME0712001	(12/01/08)	1:10
CCV/MCV/ECV (Prepare daily)	ME0710008	(04/17/09)	CCV/ECV 1:20 ME0801002
			MCV 1:40 ME0801003
Spike/LCS (Prepare daily)	ME0709009	(03/11/09)	1:100 ME0801005
	ME0801004	(07/11/08)	1:100
	ME0803001	(08/13/08)	1:50
MRL (Prepare daily)	ME0801007	(07/11/08)	1:100 ME0801008
ICSA	ME0806001	(12/01/08)	
ICSAB	ME0806002	(12/01/08)	
QCS	ME0610005	(04/10/08)	Not in use.
1ppm Check	ME0801010	(07/11/08)	
Linearity	ME0805001	(08/16/08)	
Method Sr/Ti/Sn/SiO2			
Calibration	ME0801012	(07/11/08)	
CCV/ECV	ME0803011	(09/30/08)	
QCS	ME0801012	(07/11/08)	
Spike/LCS (Prepare daily)	ME0803012	(09/30/08)	1:100
MRL (Prepare daily)	ME0801014	(07/11/08)	1:100
Method Li			
Std/ICV/MRL (Prepare daily)	ME0801009	(07/11/08)	1:1000, 200, 40, 10
QCS (Prepare daily)	ME0801011	(07/11/08)	1:10
LCS/Spike (Prepare daily)	ME0801011	(07/11/08)	1:50
ccv (Prepare daily)	ME0801011	(07/11/08)	1:40

From May 2005: the calibration std for 50 should be ME0505010,011 not ME0408010

dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.

Initial: wbh
Date: 12/1/08

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration STD ME #: 0801001
Date Received/Prepped: Prep Daily By: wbh
Date Expired: 12/1/2008 Lot #:
Manufacturer: MWH-wbh Certificate: NO
Matrix: 2% HNO3 + 5% HCl NIST SRM:
Amount: Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	1:10 ME0712002	10 ug/ml
Sb		10 ug/ml
Sn		10 ug/ml
Ti		10 ug/ml
B		5 ug/ml
Ca	1:10 ME0712001	100 ug/ml
K		100 ug/ml
Mg		100 ug/ml
Na		100 ug/ml
Al		10 ug/ml
As		10 ug/ml
Ba		10 ug/ml
Co		10 ug/ml
Cr		10 ug/ml
Cu		10 ug/ml
Fe		10 ug/ml
Mn		10 ug/ml
Ni		10 ug/ml
Pb		10 ug/ml
Se		10 ug/ml
Tl		10 ug/ml
V		10 ug/ml
Zn		10 ug/ml
Cd		5 ug/ml
Be		4 ug/ml
SR		3 ug/ml
Ag		2 ug/ml

Initial: STE
Date: 12/1/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #1
Date Received/Prepped: 12/1/2007
Date Expired: 12/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid
Amount: 500 mL

ME #: 0712001
By: STE
Lot #: A2-MEB243151
Certificate: Y
NIST SRM: Varies
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		100 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		50 ug/ml
Be		40 ug/ml
SR		30 ug/ml
Ag		20 ug/ml

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-1
 Lot Number: **A2-MEB243151**
 Matrix: 5% HNO₃(abs)

ME 0712001

1,000.00 µg/mL ea:
 Ca, K, Mg, Na,
 100.00 µg/mL ea:
 Al, As, Ba, Co, Cr₃, Cu, Fe, Mn, Ni, Pb, Se, Tl, V, Zn,
 50.00 µg/mL ea:
 Cd,
 40.00 µg/mL ea:
 Be,
 30.00 µg/mL ea:
 Sr,
 20.00 µg/mL ea:
 Ag

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.4 ± 0.3 µg/mL	Arsenic, As	100.1 ± 0.3 µg/mL
Beryllium, Be	40.04 ± 0.08 µg/mL	Cadmium, Cd	50.15 ± 0.12 µg/mL
Chromium+3, Cr ₃	100.0 ± 0.2 µg/mL	Cobalt, Co	99.9 ± 0.2 µg/mL
Iron, Fe	99.6 ± 0.1 µg/mL	Lead, Pb	100.0 ± 0.3 µg/mL
Manganese, Mn	100.0 ± 0.3 µg/mL	Nickel, Ni	100.0 ± 0.3 µg/mL
Selenium, Se	100.0 ± 0.2 µg/mL	Silver, Ag	20.04 ± 0.02 µg/mL
Strontium, Sr	30.04 ± 0.18 µg/mL	Thallium, Tl	99.7 ± 0.1 µg/mL
Zinc, Zn	100.0 ± 0.3 µg/mL		

Certified Density: 1.055 g/mL (measured at 22° C)

Initial:
Date:

STE
12/1/07

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration Stock Std #2
Date Received/Prepped: 12/1/2007
Date Expired: 12/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid + Trace HF
Amount: 500 mL

ME #: 0712002
By: STE
Lot #: A2-MEB243152
Certificate: Y
NIST SRM: Varies
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml
Sb		100 ug/ml
Sn		100 ug/ml
Ti		100 ug/ml
B		50 ug/ml
Mo		100 ug/ml

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-2
 Lot Number: **A2-MEB243152**
 Matrix: tr. HF, 5% HNO₃(abs)

ME 0712 C02

100.00 µg/mL ea:
 Mo, Sb, Sn, Ti
 50.00 µg/mL ea:
 B

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	49.85 ± 0.20 µg/mL	Molybdenum, Mo	100.2 ± 0.3 µg/mL
Tin, Sn	100.2 ± 0.3 µg/mL	Titanium, Ti	100.4 ± 0.2 µg/mL		

Certified Density: 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean
 x_i = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2\{(\sum s_i)^2\}^{1/2}}{(n)^{1/2}}$$

$\sum s_i$ = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

Initial: STE
Date: 10/19/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV/QCS Stock Standard
Date Received/Prepped: 10/19/2007
Date Expired: 4/17/2009
Manufacturer: CPI
Matrix: 5% HNO3 = tr HF
Amount: 100 mL x 10

ME #: 0710008
By: STE
Lot #: 07J154
Certificate: Y
NIST SRM: Various
Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	P/N 4400-061003RH01	20 ppm
Al		100 ppm
As		100 ppm
B		50 ppm
Ba		100 ppm
bE		40 ppm
Ca		1000 ppm
Cd		50 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		1000 ppm
Mg		1000 ppm
Mn		100 ppm
Mo		100 ppm
Na		1000 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		20 ppm
Sn		20 ppm
Ti		20 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 4/17/2009

Certificate of Analysis

Part Number: 4400-061003RH01
Lot Number: 07J154
Shelf Life: 18 months

ME 0710008
 10/13/08
 STE

MWH
 Custom Multi
 5% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiiinternational.com.

Initial: WBY

Date: 11/1/08

METALS STANDARD DOCUMENTATION

Standard: ICP MCV Working Standard
Date Received/Prepped: Daily
Date Expired: 4/17/2009
Manufacturer: CPI
Matrix: 2% HNO₃ + 5% HCL
Amount: 100 mL

ME #: 0801003
By: Wbh
Lot #:
Certificate: Y
NIST SRM: Various
Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	5ml ME0710008 / 100mL	0.5 ppm
Al		2.5 ppm
As		2.5 ppm
B		1.25 ppm
Ba		2.5 ppm
bE		1 ppm
Ca		25 ppm
Cd		1 ppm
Co		2.5 ppm
Cr		2.5 ppm
Cu		2.5 ppm
Fe		2.5 ppm
K		25 ppm
Mg		25 ppm
Mn		2.5 ppm
Mo		2.5 ppm
Na		25 ppm
Ni		2.5 ppm
Pb		2.5 ppm
Sb		2.5 ppm
Se		2.5 ppm
Tl		2.5 ppm
V		2.5 ppm
Zn		2.5 ppm
Sr		0.5 ppm
Sn		0.5 ppm
Ti		0.5 ppm

Initial:

Date:

WBJ
4/14/09

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/ECV Working Standard
 Date Received/Prepped: Daily
 Date Expired: 4/17/2009
 Manufacturer: CPI
 Matrix: 2% HNO₃ + 5% HCL
 Amount: 100 mL

ME #: 0801002
 By: Wbh
 Lot #:
 Certificate: Y
 NIST SRM: Various
 Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	5ml ME0710008 / 100mL	1 ppm
Al		5 ppm
As		5 ppm
B		2.5 ppm
Ba		5 ppm
bE		2 ppm
Ca		50 ppm
Cd		2 ppm
Co		5 ppm
Cr		5 ppm
Cu		5 ppm
Fe		5 ppm
K		50 ppm
Mg		50 ppm
Mn		5 ppm
Mo		5 ppm
Na		50 ppm
Ni		5 ppm
Pb		5 ppm
Sb		5 ppm
Se		5 ppm
Tl		5 ppm
V		5 ppm
Zn		5 ppm
Sr		1 ppm
Sn		1 ppm
Ti		1 ppm

Initial:

Date:

WBH
7/11/08

METALS STANDARD DOCUMENTATION

Standard: Spike and LCS std for ICP
Date Received/Prepped: prep daily
Date Expired: 7/11/2008
Manufacturer: MWH-wbh
Matrix: 2% HNO3 + 5% HCl
Amount:

ME #: 0801005
By: WBH
Lot #:
Certificate: N
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
CA	1:200 ME0709007	50 ppm
K		20 ppm
MG		20 ppm
NA		50 ppm
Iron	1:100 ME0709009	5 mg/L
Aluminum		2 mg/L
Barium		1 mg/L
Cobalt		1 mg/L
Chromium		1 mg/L
Copper		1 mg/L
Molybdenum		1 mg/L
Strontium		1 mg/L
Titanium		1 mg/L
Vanadium		1 mg/L
Zinc		1 mg/L
Tin		1 mg/L
Silver		0.5 mg/L
Boron		0.5 mg/L
Manganese		0.5 mg/L
Nickel		0.5 mg/L
Antimony		0.5 mg/L
Arsenic		0.2 mg/L
Cadmium		0.2 mg/L
Lead		0.2 mg/L
Selenium		0.2 mg/L
Thallium		0.2 mg/L
Uraium		0.2 mg/L
Beryllium		0.05 mg/L
AS	1:100 ME0801004	0.8 mg/L
PB		0.8 mg/L
SE		0.8 mg/L
TL		0.8 mg/L

Initial:

Date:

W36

11/11/08

METALS STANDARD DOCUMENTATION

Standard: ICP Spike Solution
Date Received/Prepped: 1/11/2008
Date Expired: 7/11/2008
Manufacturer: MWH-wbh
Matrix: 2% HNO₃
Amount: 100mL

ME #: 0801004
By: Wbh
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	8.0mL ME07090023/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

Initial:
Date:

wbh
2/25/08

METALS STANDARD DOCUMENTATION

Standard: ICP LCS/SPIKE Solution
Date Received/Prepped: 2/25/2008
Date Expired: 7/25/2008
Manufacturer: MWH-wbh
Matrix: 2% HNO₃ + 5% HCl
Amount: 100mL

ME #: 0802001
By: wbh
Lot #:
Certificate: NO
NIST SRM:
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	1:4 ME0702002	5000 ug/ml
K	1:10 ME0702005	1000 ug/ml
Mg	1:10 ME0702004	1000 ug/ml
Na	1:4 ME0702003	5000 ug/ml

Initial: STE
Date: 9/13/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV Stock Standard
Date Received/Prepped: 9/13/2007
Date Expired: 3/11/2009
Manufacturer: CPI International
Matrix: 5% HNO₃ AND 0.1% HF
Amount: 100 mL

ME #: 0709009
By: STE
Lot #: 071040
Certificate:
NIST SRM:
Storage: Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Fe	P/N 4400-050314RH01	500
Al	(10 bottles)	200
Ba		100
Co		100
Cr		100
Cu		100
Mo		100
Sr		100
Ti		100
V		100
Zn		50
Ag		50
Ba		50
Mn		50
Ni		50
Sb		50
As		20
Cd		20
Pb		20
Se		20
Tl		20
Sn		100
Be		5
U		20



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 3/11/2009

Certificate of Analysis

Part Number: 4400-050314RH01
Lot Number: 071040
Shelf Life: 18 months

ME C7 C9 C09

MWH Labs
 5% HNO₃ + 0.1% HF
 #REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702005
By: WBH
Lot #: 07B056
Certificate: Y
NIST SRM: 3141
 Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

AUG 18 08



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

M80702005

P/N 4400-10M411
P/N S4400-10M411
 Single-Element Potassium Standard
 K in 1% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)
 Source Purity: 99.999%
 Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 0.39	0.1	Cu 0.16	0.1	Pb ND	0.1	K X	70	Tl ND	0.1
Sb 0.34	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.14	0.1	Eu ND	0.1	Mg 2.6	0.2	Rh ND	0.1	Sn 0.17	0.1
Be ND	0.1	Gd ND	0.1	Mn 0.93	1	Rb 9.5	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
C? ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	20	Yb ND	0.1
Ca 82	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na 19	1	Zn 2.9	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

WBH

Date:

2/14/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO₃
Amount: 250 mL

ME #: 0702004
By: WBH
Lot #: 07B058
Certificate: Y
NIST SRM: 3131
Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



USA
 5660 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M311

470702004

P/N S4400-10M311

Single-Element Magnesium Standard

Mg in 4% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B058

Material Source: Magnesium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 28	0.1	Cu 1.6	0.1	Pb 7.7	0.7	K ND	70	Tl 0.91	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr 0.28	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.28	0.1	Eu ND	0.1	Mg X	0.2	Rh ND	0.1	Sn 0.14	0.1
Be ND	0.1	Gd 0.23	0.1	Mn 19.8	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga 0.18	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd 1.1	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 1	0.1	Si 64	20	Yb ND	0.1
Ca ND	7	Ho ND	0.1	Nb ND	0.1	Ag 0.19	0.1	Y 0.2	0.1
Ce 2.1	0.1	I 1	0.2	Os ND	0.1	Na 7.2	1	Zn ND	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 0.19	0.1	Zr 0.29	0.1
Cr ND	1	Fe 80	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La 0.76	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

WBH

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702003
By: WBH
Lot #: 07B057
Certificate: Y
NIST SRM: 3152a
Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

M70702003

Material Source: Sodium Nitrate (NaNO₃)
Source Purity: 99.99%
Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>					
Al	1.5	0.1	Cu	0.45	0.1	Pb	ND	0.1	K	ND	70	Tl	ND	0.1
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1	Th	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1	Tm	ND	0.1
Ba	0.13	0.1	Eu	ND	0.1	Mg	2.3	0.2	Rh	ND	0.1	Sn	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1	Ti	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1	W	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1	U	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6	V	ND	1
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	8	Yb	ND	0.1
Ca	120	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1	Y	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	X	1	Zn	2.9	2
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1	Zr	ND	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1			
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1			

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.



Initial:

WBH

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 250 mL

ME #: 0702002
By: WBH
Lot #: 07B065
Certificate: Y
NIST SRM: 3109a
 Room temp. storage

Component	Comment	Conc. Unit:
Ca	P/N 4400-10M91	10000 ppm



USA
 5580 Skylark Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

180702602

Single-Element Calcium Standard
 Ca in 4% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)
 Source Purity: 99.997%
 Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Tl 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.



73

Accu-prep 7000™
 Extraction RapiFold



Yu-Phase™
 IPE Disks

Initial:

Date:

WBY
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Thallium 1000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 2% HNO3
Amount: 100 mL

ME #: 0702006
By: WBH
Lot #: 06H213
Certificate: Y
NIST SRM: 3158
Room temp. storage

Component	Comment	Conc. Unit:
Tl	P/N S4400-1000581	1000 ppm

Aug 16 08



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

MF0702006

P/N 4400-1000581
P/N S4400-1000581

Single Element Thallium Standard
Tl in 2% HNO₃
1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal
Source Purity: 99.999%
Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.




<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>Ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 13.3	0.1	Cu 9.3	0.1	Pb 41	0.1	K ND	70	Tl X	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.37	0.1	Eu ND	0.1	Mg 1.7	0.2	Rh ND	0.1	Sn ND	0.1
Be 0.67	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti 0.45	0.1
Bi 0.12	0.1	Ga ND	0.1	Hg 0.16	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.21	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd 1.6	0.1	Hf ND	0.1	Ni 1.1	0.1	Si 46	8	Yb ND	0.1
Ca 51	7	Ho ND	0.1	Nb ND	0.1	Ag 0.3	0.1	Y ND	0.1
Ce ND	0.1	I 0.4	0.2	Os ND	0.1	Na 3.3	1	Zn 14.7	2
Cs 0.24	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 20	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.

75





Initial:
Date:

W/34
3/5/07

METALS STANDARD DOCUMENTATION

Standard: Selenium Stock Standard
Date Received/Prepped: 3/5/2007
Date Expired: 8/22/2008
Manufacturer: CPI
Matrix: 2% HNO3
Amount: 100 mL

ME #: 0703001
By: wbh
Lot #: 6.00E+228
Certificate: Y
NIST SRM: 3148
Storage: Room Temp

Component	Comment	Conc. Unit:
Se	P/N # S4400-1000491	1000 ppm



USA
 5580 Skyline Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-1000491

P/N S4400-1000491

Single-Element Selenium Standard

Se in 2% HNO₃

1000 ± 3 µg/mL

Lot # 06E228

M70703001

Material Source: Selenium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.011 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3148. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 1.8	0.1	Cu 0.4	0.1	Pb 0.3	0.1	K ND	70	Tl 3.6	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Fu ND	0.1	Mg 1.1	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.6	0.1	Sm ND	0.1	U ND	0.1
Br INT	10	Au ND	0.1	Nd ND	0.1	Se X	6	V ND	1
Cd 0.4	0.1	Hf ND	0.1	Ni 0.6	0.1	Si 40	8	Yb ND	0.1
Ca 5	7	Ho ND	0.1	Nb INT	0.1	Ag 0.8	0.1	Y ND	0.1
Ce ND	0.1	I 0.5	0.2	Os ND	0.1	Na 3.8	1	Zn ND	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr INT	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:
Date:

WBH
4/16/07

METALS STANDARD DOCUMENTATION

Standard: Lead Stock Standard
Date Received/Prepped: 4/16/2007
Date Expired: 10/11/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0704013
By: WBH
Lot #: 07A097
Certificate: Y
NIST SRM: 3128
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Pb	P/N S4400-1000281	1000 ppm

ME0709023

Initial:

STE

Date:

9/24/07

METALS STANDARD DOCUMENTATION

Standard: As Stock Standard
Date Received/Prepped: 9/24/2007
Date Expired: 10/1/2008
Manufacturer: Inorganic Ventures
Matrix: 1.4% HNO3
Amount: 100 mL X2

ME #: 0709023
By: STE
Lot #: A2-AS02035
Certificate: Y
NIST SRM:
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
As	PN: CGAS1-1	1000 ug/ml

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM **1000 µg/mL Arsenic in 1.4% (abs) HNO₃**

Catalog Number: CGAS1-1, CGAS1-2, and CGAS1-5
 Lot Number: **A2-AS02035**
 Starting Material: As Polycrystalline lump
 Starting Material Purity (%): 99.998288
 Starting Material Lot No: 23444
 Matrix: 1.4% (abs) HNO₃

ME 0709 023

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Concentration: 1000 ± 6 µg/mL

Certified Density: 1.010 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean

x_i = individual results

n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

$\sum s_i$ = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 Assay Method #1 **1000 ± 6 µg/mL**
 ICP Assay NIST SRM 3103a Lot Number: 010713

Assay Method #2 **1001 ± 5 µg/mL**
 Gravimetric NIST SRM Lot Number: See Sec. 4.2

Date: WBY
7/11/08

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working Stock Solution ME #: 0801008
Date Received/Prepped: Daily By: Wbh
Date Expired: 7/11/2008 Lot #:
Manufacturer: MWH-wbh Certificate: Y
5% HNO3 2% HNO3 + 5% Hcl NIST SRM:
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	0.1mL ME0801007 / 10mL	0.05 ppm
Sb		0.05 ppm
As		0.1 ppm
Ba		0.02 ppm
Be		0.001 ppm
Ba		0.05 ppm
Cd		0.005 ppm
Ca		1 ppm
Cr		0.01 ppm
Co		0.05 ppm
Cu		0.01 ppm
Fe		0.02 ppm
Pb		0.02 ppm
Li		0.1 ppm
Mg		0.002 ppm
Mn		0.02 ppm
Mo		0.02 ppm
Ni		1 ppm
K		0.1 ppm
Se		0.01 ppm
Ag		1 ppm
Na		0.02 ppm
Tl		0.002 ppm
V		0.1 ppm
Zn		0.001 ppm
Ti		0.02 ppm
Sr		0.01 ppm
Sn		0.2 ppm

Date:

WZ
1/11/08

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working Stock Solu ME #: 0801007
Date Received/Prepped: 1/11/2008 By: Wbh
Date Expired: 7/11/2008 Lot #:
Manufacturer: MWH-wbh Certificate: Y
Matrix: 5% HNO3 NIST SRM:
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	10mL ME0709020 / 100mL	5 ppm
Sb		5 ppm
As		10 ppm
Ba		2 ppm
Be		0.1 ppm
Ba		5 ppm
Cd		0.5 ppm
Ca		100 ppm
Cr		1 ppm
Co		5 ppm
Cu		1 ppm
Fe		2 ppm
Pb		2 ppm
Li		10 ppm
Mg		0.2 ppm
Mn		2 ppm
Mo		2 ppm
Ni		100 ppm
K		10 ppm
Se		1 ppm
Ag		100 ppm
Na		2 ppm
Tl		0.2 ppm
V		10 ppm
Zn		0.1 ppm
Ti		2 ppm
Sr		1 ppm
Sn		20 ppm

Initial:
Date:

STE
9/20/07

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Stock Standard
Date Received/Prepped: 9/20/2007
Date Expired: 9/18/2008
Manufacturer: CPI
Matrix: 2% HNO₃ + tr HF
Amount: 100 mL

ME #: 0709020
By:
Lot #: 061162
Certificate: Y
NIST SRM:
Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-060915RHO1	50 ppm
Sb		50 ppm
As		100 ppm
Ba		20 ppm
Be		1 ppm
Ba		50 ppm
Cd		5 ppm
Ca		1000 ppm
Cr		10 ppm
Co		50 ppm
Cu		10 ppm
Fe		20 ppm
Pb		20 ppm
Li		1 ppm
Mg		100 ppm
Mn		2 ppm
Mo		20 ppm
Ni		20 ppm
K		1000 ppm
Se		100 ppm
Ag		10 ppm
Na		1000 ppm
Tl		100 ppm
V		2 ppm
Zn		20 ppm
Ti		20 ppm
Sr		10 ppm
Sn		200 ppm

**USA**

5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

ME0709020

Expiry 9/18/2008

Certificate of Analysis

Part Number: 4400-060915RH01
Lot Number: 061162
Shelf Life: 12 months

MWH
 Custom Standard
 2% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Al	50	Pb	20	Zn	20
Si	50	Li	1	Ti	20
As	100	Mg	100	Sr	10
Ba	20	Mn	2	Sn	200
Be	1	Mo	20		
B	50	Ni	20		
Cd	5	K	1000		
Ca	1000	Se	100		
Cr	10	Ag	10		
Co	50	Na	1000		
Cu	10	TL	100		
Fe	20	V	2		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: CSK
 Date: 6-1-08

METALS STANDARD DOCUMENTATION

Standard:	INTERFERENCE CHECK STANDAF	ME #: 0806002
Date Received/Prepped:	6/1/2008	By: CSK
Date Expired:	12/1/2008	Lot #: various
Manufacturer:	MWH-CSK	Certificate:
Matrix:	5% HNO3	NIST SRM:
Amount:	500 mL	Storage: Room Temp.

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500(25ML)	250 ppm
Ca	P/N 4400-INTB1-100(2.5ML)	250 ppm
Fe		100 ppm
Mg		250 ppm
AG		0.5 ppm
BA		0.25 ppm
BE		0.25 ppm
CD		0.5 ppm
Co		0.25 ppm
Cr3		0.25 ppm
Cu		0.5 ppm
Mn		0.25 ppm
Ni		0.25 ppm
Pb		0.25 ppm
V		0.25 ppm
Zn		0.5 ppm

Tl
V
Zn
Si

~~100 ppm~~
~~100 ppm~~
~~100 ppm~~
~~50 ppm~~

6/1/08

METALS STANDARD DOCUMENTATION

Standard: INTERFERENCE CHECK STANDAF
Date Received/Prepped: 6/1/2008
Date Expired: 12/1/2008
Manufacturer: MWH-CSK
Matrix: 5% HNO3
Amount: 500ML

ME #: 0806001
By: CSK
Lot #: various
Certificate:
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
AL	P/N 4400-INTA1-500	250 ppm
CA	25ML IN 500ML	250 ppm
FE		100 ppm
MG		250 ppm
AG		
BA		
BE		
CD		
CO		
CR		
CU		
MN		
NI		
PB		
V		
ZN		

Initial:

STE

Date:

8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Analytes B Solution
Date Received/Prepped: 8/27/2007
Date Expired: 2/15/2009
Manufacturer: CPI International
Matrix: 5% HNO₃
Amount: 100 mL

ME #: 0708010
By: STE
Lot #: 07c256
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Ag	P/N 4400-INTB1-100	100 ug/L
Ba		50 ug/L
Be		50 ug/L
Cd		100 ug/L
Co		50 ug/L
Cr		50 ug/L
Cu		50 ug/L
Mn		50 ug/L
Ni		100 ug/L
Pb		100 ug/L
V		50 ug/L
Zn		100 ug/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 GS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

FEB 15 09

*Innovative Solutions
 in Analytical Science and
 Technology*

ME 070 500

CERTIFICATE OF ANALYSIS

P/N 4400-INTB1-100

CLP Analytes B Solution
 in 5% HNO₃

Lot # 07c256

Material Source: Metals and Salts
 Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag	100	Ba	50	Be	50	Cd	100
Co	50	Cr	50	Cu	50	Mn	50
Ni	100	Pb	100	V	50	Zn	100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: STE
Date: 8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Interferents A Solution **ME #:** 0708009
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 2/15/2009 **Lot #:** 07E175
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL **Storage:** Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Al	P/N 4400-INTA1-500	5000 ug/mL
Ca		5000 ug/mL
Fe		2000 ug/mL
Mg		5000 ug/mL

FEB 15 09



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500

CLP Interferents A Solution
in 5% HNO₃

ME 0708009

Lot # 07E175

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: Wbh
 Date: 1/11/08

METALS STANDARD DOCUMENTATION

Standard: ICP 1 PPM CHECK
 Date Received/Prepped: 1/11/2008
 Date Expired: 7/11/2008
 Manufacturer: MWH-wbh
 Matrix: 5% HNO₃
 Amount: 500 mL

ME #: 0801010
 By: Wbh
 Lot #: VARIOUS
 Certificate:
 NIST SRM:
 Storage: Room Tem

Component	Comment	Conc. Unit:
Ag	5mL ME0708011 /500mL	1 mg/L
Al		1 mg/L
B		1 mg/L
Ba		1 mg/L
K		10 mg/L
Na		1 mg/L
Si		1 mg/L
As	5mL ME0708012 /500mL	1 mg/L
Be		1 mg/L
Ca		1 mg/L
Cd		1 mg/L
Co		1 mg/L
Cr		1 mg/L
Cu		1 mg/L
Fe		1 mg/L
Li		1 mg/L
Mg		1 mg/L
Mn		1 mg/L
Mo		1 mg/L
Ni		1 mg/L
Pb		1 mg/L
Sb		1 mg/L
Se		1 mg/L
Sr		1 mg/L
Ti		1 mg/L
Tl		1 mg/L
V		1 mg/L
Zn		1 mg/L

Initial: STE
 Date: 8/27/07

METALS STANDARD DOCUMENTATION

Standard: QC Check Standard 21
Date Received/Prepped: 8/27/2007
Date Expired: 8/31/2008
Manufacturer: Crescent Chemical Co. Inc.
Matrix: 5% HNO₃/tr. F/tr Tartaric Acid
Amount: 100 mL

ME #: 0708012
By: STE
Lot #: 074438H
Certificate:
NIST SRM:
 Room temp. storage

Component	Comment	Conc. Unit:
As	Catalog No: QC-021.1	100 ug/mL
Be		100 ug/mL
Ca		100 ug/mL
Cd		100 ug/mL
Co		100 ug/mL
Cu		100 ug/mL
Fe		100 ug/mL
Li		100 ug/mL
Mg		100 ug/mL
Mn		100 ug/mL
Mo		100 ug/mL
Ni		100 ug/mL
Pb		100 ug/mL
Sb		100 ug/mL
Se		100 ug/mL
Sr		100 ug/mL
Ti		100 ug/mL
Tl		100 ug/mL
V		100 ug/mL
Zn		100 ug/mL

ME 0708012

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 21

CATALOG NO: QC-021.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F/tr. Tartaric Acid

LOT NO.: 074438H

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3102a, 3103a, 3105a, 3108, 3109a, 3112a, 3113, 3114, 3126a, 3128, 3129a, 3131a, 3132, 3134, 3136, 3149, 3153a, 3158, 3162a, 3165, and 3168a.

Concentrations are given in µg/mL unless noted otherwise.

As 100	Be 100	Ca 100	Cd 100	Co 100
Cr 100	Cu 100	Fe 100	Li 100	Mg 100
Mn 100	Mo 100	Ni 100	Pb 100	Sb 100
Se 100	Sr 100	Ti 100	Tl 100	V 100
Zn 100				

Crescent Chemical Co. Inc.

Julie M. MacIntosh
QA Manager

EXPIRES: August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913

ME0705011

Initial:
Date:

STE
8/31/07

METALS STANDARD DOCUMENTATION

Standard: QC Check Standard 7
Date Received/Prepped: 8/27/2007
Date Expired: 8/31/2007
Manufacturer: Crescent Chemical Co. Inc.
Matrix: 5% HNO3
Amount: 100

ME #: 0705011
By: STE
Lot #: 074438I
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Ag	Catalog No: QC-007.1	100 ug/mL
Al		100 ug/mL
B		100 ug/mL
Ba		100 ug/mL
K		1000 +/- 5 ug/mL
Na		100 ug/mL
Si		50 ug/mL

ME 07 08 011

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 7

CATALOG NO: QC-007.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F⁻

LOT NO.: 074438I

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3101a, 3104a, 3107, 3141a, 3150, 3151, and 3152a.

Concentrations are given in µg/mL unless noted otherwise.

Ag 100	Al 100	B 100	Ba 100	K 1,000±5
Na 100	Si 50.0			

Crescent Chemical Co. Inc.

Julie M. MacIntosh
QA Manager

EXPIRES: August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

*Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913*

Initial: CSK
Date: 05/05/08

METALS STANDARD DOCUMENTATION

Standard: ICP LINEARITY CHECK
Date Received/Prepped: 5/5/2008
Date Expired: 11/5/2008
Manufacturer: MWH-STE
Matrix: 5% HNO3
Amount: 500ML

ME #: 0805001
By: CSK
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
CA	15.0ML ME0702002/500ML	300 ppm
K	15.0ML ME0702005/500ML	300 ppm
MG	10.0ML ME0702004/500ML	200 ppm
NA	15.0ML ME0702003/500ML	300 ppm
FE	5.0ML ME0701008/500ML	100 ppm

Initial:

Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702005
By: WBH
Lot #: 07B056
Certificate: Y
NIST SRM: 3141
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
K	P/N 4400-10M411	10000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

ME0702005

P/N 4400-10M411
P/N S4400-10M411
 Single-Element Potassium Standard
 K in 1% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)
 Source Purity: 99.999%
 Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 0.39	0.1	Cu 0.16	0.1	Pb ND	0.1	K X	70	Tl ND	0.1
Sb 0.34	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.14	0.1	Eu ND	0.1	Mg 2.6	0.2	Rh ND	0.1	Sn 0.17	0.1
Be ND	0.1	Gd ND	0.1	Mn 0.93	1	Rb 9.5	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
C ²⁺ ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	20	Yb ND	0.1
Ca 82	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na 19	1	Zn 2.9	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH
2/19/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std **ME #:** 0702004
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B058
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3131
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M311

P/N S4400-10M311

Single-Element Magnesium Standard

Mg in 4% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B058

M 70702004

Material Source: Magnesium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 28	0.1	Cu 1.6	0.1	Pb 7.7	0.7	K ND	70	Tl 0.91	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr 0.28	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.28	0.1	Eu ND	0.1	Mg X	0.2	Rh ND	0.1	Sn 0.14	0.1
Be ND	0.1	Gd 0.23	0.1	Mn 19.8	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga 0.18	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd 1.1	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 1	0.1	Si 64	20	Yb ND	0.1
Ca ND	7	Ho ND	0.1	Nb ND	0.1	Ag 0.19	0.1	Y 0.2	0.1
Ce 2.1	0.1	I 1	0.2	Os ND	0.1	Na 7.2	1	Zn ND	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 0.19	0.1	Zr 0.29	0.1
Cr ND	1	Fe 80	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La 0.76	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

U31

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std **ME #:** 0702003
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B057
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO3 **NIST SRM:** 3152a
Amount: 250 mL **Room temp. storage**

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



USA
5580 Skylane Boulevard 707.526.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

M70702003

Material Source: Sodium Nitrate (NaNO₃)
Source Purity: 99.99%
Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	1.5	0.1	Cu	0.45	0.1	Pb	ND	0.1	K	ND	70	Tl	ND	0.1
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1	Th	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1	Tm	ND	0.1
Ba	0.13	0.1	Eu	ND	0.1	Mg	2.3	0.2	Rh	ND	0.1	Sn	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1	Ti	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1	W	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1	U	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6	V	ND	1
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	8	Yb	ND	0.1
Ca	120	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1	Y	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	X	1	Zn	2.9	2
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1	Zr	ND	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1			
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1			

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial:

WBH

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 250 mL

ME #: 0702002
By: WBH
Lot #: 07B065
Certificate: Y
NIST SRM: 3109a
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

1780702602

Single-Element Calcium Standard
Ca in 4% HNO₃
10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)
Source Purity: 99.997%
Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Tl 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

105



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
In Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M261

P/N S4400-10M261

Single-Element Iron Standard

Fe in 4% HNO3

10,000 ± 30 µg/mL

Lot # 06I143

Handwritten ID: MTE070100X

Material Source: Iron Metal
Source Purity: 99.999%
Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

Table with 10 columns: Element, ppb, DL, Element, ppb, DL, Element, ppb, DL, Element, ppb, DL. Lists various elements and their concentrations/detection limits.

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

From Page No.	LOG #	CURRENT	SX #	MIP	VOLUME	COMMENTS
		200.7 DIGEST	06-30-08	JRF		HND3 100452 HLI 100456 LCS / SPK
		BLANK LCS LCS				ME 0709009
	2806190595	XXXXXXXXXX	18E48	AQ	SOML → SOML	ME 08030013 ME 0802001 → 0.
	↓ MS	↓	↓			
	↓ MSD	↓				
	2806250566	XXXXXXXXXX	FINAL EFFLUENT			
	↓ MS	↓	↓			
	↓ MSD	↓				
	2806180344	XXXXXXXXXX				
	↓ 0345	XXXXXXXXXX				
	2806190596	↓	GF 56			
	2806200356	XXXX	100228			
	↓ 0358	↓	100243			
	2806240538	RETRACTEE	EFFLUENT			
	↓ 0540	↓	INFLOW			
	2806250254	X5 XXXXXX	MW-2			
	↓ 0255	X5 ↓	MW-3			
	↓ 0256	X5 ↓	MW-4			
	↓ 0257	X5 ↓	MW-5			
	↓ 0258	X5 ↓	MW-6			
	↓ 0259	X5 ↓	MW-7			
	↓ 0260	X5 ↓	MW-8			
	2806250143	XXXXXXXXXX	NCL-3			
	↓ 0147	↓	DW-CUL-1			
	↓ 0159	↓	DW-TMP-1			
	2806260194	RETRACTEE	M-100	6010 ←		6010 ←
		200.7 DIGEST	06-30-08	JRF		
		0806220				

Witnessed & Understood by me,	Date	Invented by	Date
108		Recorded by	

Project No. _____

Book No. _____

TITLE _____

From Page No.	LOG #	LOG #	SK - SD	FIX	ROWING	COMMENTS
		2007	06-30-08	JRP		MND's 100452
	BLANK	01685				HCI 100456
	LCS					LCS/SPK
	LCSD					ME 0709009
280625	0685	[REDACTED]	1207118	AQ	50M → 50M	ME 08030013 →
	MS					ME 0802001 →
	MSD					
280625	0695	[REDACTED]	1207115			MRL
	MS					ME 070920 →
	MSD					
280625	0165	[REDACTED]	DWCTH-1			HOTBLOCK #1
	0176		CB#6			INT: 110/96
	0200		M200A			FIN:
	0206		CB#17			
	0223		BST-2CB#3			
280625	0684	[REDACTED]	1207117			
	0686		1207119			
	0687		1207120			
	0688		1207121			
	0689		1207098			
	0693		1207114			
	0696		1207116			
	0697		1207124			
	0698		1207122			
	0699		1207123			
	0700		1208388			
	0701		1208306			
280620	0200	[REDACTED]	BY POINT			
	2007		06-30-08	JRP		
	Digest					

080630

To Page

Witnessed & Understood by me,

Date

Invented by

Date

Recorded by