



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

ERM
2525 Natomas Park Drive, Suite 350
Sacramento, CA 95833
ATTN: Ms. Maria Barajas-Albalawi

August 14, 2008

SUBJECT: BRC Tronox Parcel F, Data Validation

Dear Ms. Barajas-Albalawi

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on July 31, 2008. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 19215:

<u>SDG #</u>	<u>Fraction</u>
IRF1297, IRF1298	2,2'-/4,4'-Dichlorobenzil, Hexavalent Chromium & Chlorite

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, October 2004
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996; update IIIA, April 1998; IIIB, November 2004; Update IV, February 2007

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

LDC #19215 (ERM-Sacramento / BRC Tronox, Parcel F)

80/20

LDC	SDG#	DATE REC'D	(3) DATE DUE	2,2',4,4'-DCB (8270C)		Chlorite (300.1)		Cr(VI) (7196A)		W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S			
				W	S	W	S	W	S																	W	S	W
Matrix: Water/Soil																												
A	IRF1297	07/31/08	08/21/08	0	3	0	3	0	3																			
A	IRF1297	07/31/08	08/21/08	0	2	0	2	0	2																			
B	IRF1298	07/31/08	08/21/08	0	5	0	5	0	5																			
Total				T/LR	0	10	0	10	0	10	0	10	0	10	0	10	0	10	0	10	0	10	0	10	0	10	0	30

Shaded cells indicate Level IV validation (all other cells are Level III validation). These sample counts do not include MS/MSD, and DUPs

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: BRC Tronox Parcel F
Collection Date: June 10, 2008
LDC Report Date: August 12, 2008
Matrix: Soil
Parameters: 2,2'-/4,4'-Dichlorobenzil
Validation Level: EPA Level III & IV
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): IRF1297

Sample Identification

TSB-FJ-06-02-10'
TSB-FJ-06-02-20'**
TSB-FJ-06-02-30'
TSB-FR-02-02-10'**
TSB-FR-02-02-10'-FD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for 2,2'-/4,4'-Dichlorobenzil.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals. All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% .

Average relative response factors (RRF) for all target compounds were within validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 25.0% .

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration RRF values were within validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 2,2'-/4,4'-Dichlorobenzil was found in the method blanks.

No field blanks were identified in this SDG.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples TSB-FR-02-02-10'** and TSB-FR-02-02-10'-FD were identified as field duplicates. No 2,2'-/4,4'-Dichlorobenzil were detected in any of the samples.

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Data Qualification Summary - SDG IRF1297**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Laboratory Blank Data Qualification Summary - SDG
IRF1297**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Field Blank Data Qualification Summary - SDG IRF1297**

No Sample Data Qualified in this SDG

LDC #: 19215A2
 SDG #: IRF1297
 Laboratory: Test America

VALIDATION COMPLETENESS WORKSHEET
 Level III/IV

Date: 8/11/08
 Page: 1 of 1
 Reviewer: SVZ
 2nd Reviewer: J

METHOD: GC/MS 2,2'-/4,4'-Dichlorobenzil (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 6/10/08
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	non CCC / non SPCC
IV.	Continuing calibration/ICV	A	ICV = 25% ↓
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client Spec (Parcel 5)
VIII.	Laboratory control samples	A	ICS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	Not reviewed for Level III validation.
XII.	Compound quantitation/CRQLs	A	Not reviewed for Level III validation.
XIII.	Tentatively identified compounds (TICs)	N	Not reviewed for Level III validation.
XIV.	System performance	A	Not reviewed for Level III validation.
XV.	Overall assessment of data	A	
XVI.	Field duplicates	ND	D = 4, 5
XVII.	Field blanks	N	

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet
 ND = No compounds detected
 R = Rinsate
 FB = Field blank
 D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: Soil ** Indicates sample underwent Level IV validation

1	TSB-FJ-06-02-10'	11		21		31	
2	TSB-FJ-06-02-20'**	12		22		32	
3	TSB-FJ-06-02-30'	13		23		33	
4	TSB-FR-02-02-10'**	14		24		34	
5	TSB-FR-02-02-10'-FD	15		25		35	
6	8# 16058 Bk1	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

LDC #: 19215 AY
 SDG #: su level

VALIDATION FINDINGS CHECKLIST

Page: 1 of 7
 Reviewer: JN
 2nd Reviewer: [Signature]

Method: Semivolatiles (EPA SW 846 Method 8270C)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Cooler temperature criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
II. GC/MS Instrument Performance Check				
Were the DFTPP performance results reviewed and found to be within the specified criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all samples analyzed within the 12 hour clock criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a curve fit used for evaluation?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Did the initial calibration meet the curve fit acceptance criteria of > 0.990?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Were all percent relative standard deviations (%RSD) ≤ 30% and relative response factors (RRF) > 0.05?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) ≤ 25% and relative response factors (RRF) ≥ 0.05?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
V. Blanks				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a method blank analyzed for each matrix and concentration?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
If 2 or more base neutral or acid surrogates were outside QC limits, was a reanalysis performed to confirm %R?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
If any %R was less than 10 percent, was a reanalysis performed to confirm %R?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
VII. Matrix Spike/Matrix spike duplicate				
Were a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD. Soil / Water.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Was a MS/MSD analyzed every 20 samples of each matrix?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

LDC #: 19215 A2
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: JW
 2nd Reviewer: [Signature]

Validation Area	Yes	No	NA	Findings/Comments
Was an LCS analyzed per extraction batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the performance evaluation (PE) samples within the acceptance limits?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
X. Internal standards				
Were internal standard area counts within -50% or +100% of the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were retention times within + 30 seconds from the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XI. Target compound identification				
Were relative retention times (RRT's) within + 0.06 RRT units of the standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were chromatogram peaks verified and accounted for?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 10 percent relative intensity) in the reference spectrum evaluated in sample spectrum?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were relative intensities of the major ions within ± 20% between the sample and the reference spectra?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
XIV. System performance				
System performance was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field duplicates.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
XVII. Field blanks				
Field blanks were identified in this SDG.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field blanks.	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_u)/(A_u)(C_x)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$
 A_x = Area of compound, A_u = Area of associated internal standard
 C_x = Concentration of compound, C_u = Concentration of internal standard
 S = Standard deviation of the RRFs, X = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (SD std)	RRF (SD std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD		
1	CAL	4/09/08	Phenol (1st internal standard)	1.095	1.093	1.076	1.076	7.21	7.21		
			Naphthalene (2nd internal standard)								
	MS 45		Fluorene (3rd internal standard)								
			Pentachlorophenol (4th internal standard)								
			Bis(2-ethylhexyl)phthalate (5th internal standard)								
			Benzo(a)pyrene (6th internal standard)								
2			Phenol (1st internal standard)								
			Naphthalene (2nd internal standard)								
			Fluorene (3rd internal standard)								
			Pentachlorophenol (4th internal standard)								
			Bis(2-ethylhexyl)phthalate (5th internal standard)								
			Benzo(a)pyrene (6th internal standard)								
3			Phenol (1st internal standard)								
			Naphthalene (2nd internal standard)								
			Fluorene (3rd internal standard)								
			Pentachlorophenol (4th internal standard)								
			Bis(2-ethylhexyl)phthalate (5th internal standard)								
			Benzo(a)pyrene (6th internal standard)								

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = $100 \times (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$ Where: ave. RRF = initial calibration average RRF
 RRF = continuing calibration RRF
 $\text{RRF} = (A_x)(C_x) / (A_s)(C_s)$
 A_x = Area of compound, A_s = Area of associated internal standard
 C_x = Concentration of compound, C_s = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Average RRF (Initial)	Reported		Recalculated	
					RRF (CC)	%D	RRF (CC)	%D
1	SSTD050	6/1/08	Ethanol (1st internal standard) 4'-DCBrazil	1.076	1.096	1.9	1.8	
			Naphthalene (2nd internal standard)					
			Fluorene (3rd internal standard)					
			Pentachlorophenol (4th internal standard)					
			Bis(2-ethylhexyl)phthalate (5th internal standard)					
			Benzo(a)pyrene (6th internal standard)					
2			Phenol (1st internal standard)					
			Naphthalene (2nd internal standard)					
			Fluorene (3rd internal standard)					
			Pentachlorophenol (4th internal standard)					
			Bis(2-ethylhexyl)phthalate (5th internal standard)					
			Benzo(a)pyrene (6th internal standard)					
3			Phenol (1st internal standard)					
			Naphthalene (2nd internal standard)					
			Fluorene (3rd internal standard)					
			Pentachlorophenol (4th internal standard)					
			Bis(2-ethylhexyl)phthalate (5th internal standard)					
			Benzo(a)pyrene (6th internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 19 215A2
 SDG #: See Cover

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: SF/SS * 100

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: # 2

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5	50	36.25	72	72.5	0
2-Fluorobiphenyl	↓	37.59	75	75	↓
Terphenyl-d14		41.92	84	84	
Phenol-d5	100	93.99	74	74	↓
2-Fluorophenol	↓	74.96	75	75	
2,4,6-Tribromophenol		84.40	84	84	
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample/Laboratory Control Sample Duplicates Results Verification

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SC/SA)$ Where: SSC = Spike concentration
 SA = Spike added

RPD = $100 * \frac{|LCS - LCSD|}{LCS + LCSD}$ LCS = Laboratory control sample percent recovery LCSD = Laboratory control sample duplicate percent recovery

LCS/LCSD samples: 8F 1658 - blk 1

Compound	Spike Added (ug/kg)		Spike Concentration (ug/kg)		LCS Percent Recovery		LCSD Percent Recovery		LCS/LCSD RPD	
	LCS	LCSD	LCS	LCSD	Reported	Recalc	Reported	Recalc	Reported	Recalculated
2,2',4,4'-Dichlorobenzil Benzol	3330	NA	2780	NA	87	83				
N-Nitroso-di-n-propylamine										
4-Chloro-3-methylphenol										
Acenaphthene										
Pentachlorophenol										
Pyrene										

Comments: Refer to Laboratory Control Sample/Laboratory Control Sample Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 19215 A-7
 SDG #: see cover

VALIDATION FINDINGS WORKSHEET

Sample Calculation Verification

Page: 1 of 1
 Reviewer: JVL
 2nd reviewer: 9

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

Y N N/A
 Y N N/A

Were all reported results recalculated and verified for all level IV samples?
 Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(V_i)(DF)(2.0)}{(A_s)(RRF)(V_s)(V_i)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_s = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- V_s = Volume or weight of sample extract in milliliters (ml) or grams (g).
- V_i = Volume of extract injected in microliters (ul)
- V_t = Volume of the concentrated extract in microliters (ul)
- Df = Dilution Factor.
- %S = Percent solids, applicable to soil and solid matrices only.
- 2.0 = Factor of 2 to account for GPC cleanup

Example:

Sample I.D. _____, ND:

$$\text{Conc.} = \frac{() () () () ()}{() () () () ()}$$

=

#	Sample ID	Compound	Reported Concentration ()	Calculated Concentration ()	Qualification

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: BRC Tronox Parcel F
Collection Date: June 10, 2008
LDC Report Date: August 12, 2008
Matrix: Soil
Parameters: 2,2'-/4,4'-Dichlorobenzil
Validation Level: EPA Level III
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): IRF1298

Sample Identification

TSB-FR-02-02-20'
TSB-FR-02-02-30'
TSB-FJ-02-02-10'
TSB-FJ-02-02-20'
TSB-FJ-02-02-30'

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for 2,2'-/4,4'-Dichlorobenzil.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals. All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% .

Average relative response factors (RRF) for all target compounds were within validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 25.0% .

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration RRF values were within validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 2,2'-/4,4'-Dichlorobenzil was found in the method blanks.

No field blanks were identified in this SDG.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Data Qualification Summary - SDG IRF1298**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Laboratory Blank Data Qualification Summary - SDG
IRF1298**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
2,2'-/4,4'-Dichlorobenzil - Field Blank Data Qualification Summary - SDG IRF1298**

No Sample Data Qualified in this SDG

LDC #: 19215B2
 SDG #: IRF1298
 Laboratory: Test America

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 8/1/08
 Page: 1 of 1
 Reviewer: *SPC*
 2nd Reviewer: *[Signature]*

METHOD: GC/MS 2,2'-/4,4'-Dichlorobenzil (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 6/10/08
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	non CCC / non SPEC
IV.	Continuing calibration/ICV	A	ICV $\leq 25\%$ ↓
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client spec (Parcel 6)
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	N	

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

Soil

1	TSB-FR-02-02-20'	11		21		31	
2	TSB-FR-02-02-30'	12		22		32	
3	TSB-FJ-02-02-10'	13		23		33	
4	TSB-FJ-02-02-20'	14		24		34	
5	TSB-FJ-02-02-30'	15		25		35	
6	8F16058-Blk1	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: BRC Tronox Parcel F
Collection Date: June 10, 2008
LDC Report Date: August 11, 2008
Matrix: Soil
Parameters: Hexavalent Chromium & Chlorite
Validation Level: EPA Level III & IV
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): IRF1297

Sample Identification

TSB-FJ-06-02-10'
TSB-FJ-06-02-20'**
TSB-FJ-06-02-30'
TSB-FR-02-02-10'**
TSB-FR-02-02-10'-FD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 5 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 300.1 for Chlorite and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section X.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No chlorite or hexavalent chromium was found in the initial, continuing and preparation blanks.

No field blanks were identified in this SDG.

IV. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the 300.1 method. All surrogate recoveries (%R) were within QC limits with the following exceptions:

Sample	Surrogate	%R (Limits)	Analyte	Flag	A or P
TSB-FJ-06-02-20***	Dichloroacetate	87.8 (90-115)	Chlorite	J- (all detects) UJ (all non-detects)	P

V. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VII. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

X. Field Duplicates

Samples TSB-FR-02-02-10'** and TSB-FR-02-02-10'-FD were identified as field duplicates. No chlorite or hexavalent chromium was detected in any of the samples.

**BRC Tronox Parcel F
Hexavalent Chromium & Chlorite - Data Qualification Summary - SDG IRF1297**

SDG	Sample	Analyte	Flag	A or P	Reason
IRF1297	TSB-FJ-06-02-20**	Chlorite	J- (all detects) UJ (all non-detects)	P	Surrogate recovery (%R)

**BRC Tronox Parcel F
Hexavalent Chromium & Chlorite - Laboratory Blank Data Qualification Summary -
SDG IRF1297**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
Hexavalent Chromium & Chlorite - Field Blank Data Qualification Summary - SDG
IRF1297**

No Sample Data Qualified in this SDG

LDC #: 19215A6

VALIDATION COMPLETENESS WORKSHEET

Date: 8/6/08

SDG #: IRF1297

Level III/IV

Page: 1 of 1

Laboratory: Test America

Reviewer: al

2nd Reviewer: W

METHOD: (Analyte) Chlorite (EPA Method 300.1), Hexavalent Chromium (EPA SW846 Method 7196A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 6/10/08
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Surrogate Spikes	SW	
V	Matrix Spike/Matrix Spike Duplicates	A	} for IRF1299
VI.	Duplicates	N	
VII.	Laboratory control samples	A	LCS
VIII.	Sample result verification	A	Not reviewed for Level III validation.
IX.	Overall assessment of data	A	
X.	Field duplicates	ND	D: 4 + 5
XI	Field blanks	N	

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

AW soil

1	TSB-FJ-06-02-10'	11		21		31	
2	TSB-FJ-06-02-20'**	12		22		32	
3	TSB-FJ-06-02-30'	13		23		33	
4	TSB-FR-02-02-10'**	14		24		34	
5	TSB-FR-02-02-10'-FD	15		25		35	
6	PB	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC #: 10215AG
 SDG #: IR 81257

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: CI
 2nd Reviewer: W

Method: Inorganics (EPA Method 8000)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	1			
Cooler temperature criteria was met.	1			
II. Calibration				
Were all instruments calibrated daily, each set-up time?	1			
Were the proper number of standards used?	1			
Were all initial calibration correlation coefficients > 0.995?	1			
Were all initial and continuing calibration verification %Rs within the 90-110% QC limits?	1			
Were titrant checks performed as required? (Level IV only)			1	
Were balance checks performed as required? (Level IV only)			1	
III. Blanks				
Was a method blank associated with every sample in this SDG?	1			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.		1		
IV. Matrix spike/Matrix spike duplicates and Duplicates				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.			1	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.			1	
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of ≤ CRDL (≤ 2X CRDL for soil) was used for samples that were ≤ 5X the CRDL, including when only one of the duplicate sample values were < 5X the CRDL.			1	
V. Laboratory control samples				
Was an LCS analyzed for this SDG?	1			
Was an LCS analyzed per extraction batch?	1			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% (85-115% for Method 300.0) QC limits?	1			
VI. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		1		
Were the performance evaluation (PE) samples within the acceptance limits?			1	

LDC #: 19215AL
 SDG #: 17E1297

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: CS
 2nd Reviewer: [Signature]

Validation Area	Yes	No	NA	Findings/Comments
VII. Sample result verification				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	1			
Were detection limits < RL?	1			
VIII. Overall assessment of data				
Overall assessment of data was found to be acceptable.	1			
IX. Field duplicates				
Field duplicate pairs were identified in this SDG.	1	2		
Target analytes were detected in the field duplicates.		1		
X. Field blanks				
Field blanks were identified in this SDG.		1		
Target analytes were detected in the field blanks.			1	

LDC #: 19215AL
 SDG #: 12P1297

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1
 Reviewer: GA
 2nd reviewer: W

All circled methods are applicable to each sample.

Sample ID	Parameter
1-5	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR⁶⁺ Chloride _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺ _____

Comments: _____

LDC #: 1621546
 SDG #: 1621546

VALIDATION FINDINGS WORKSHEET Surrogate Recovery

Page: 1 of 1
 Reviewer: SA
 2nd Reviewer: L

METHOD: Chlorite (EPA 300.1)

Are surrogates required by the method? Yes or No
 Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 N N/A Were surrogates spiked into all samples and blanks?
 Y N/A Did all surrogate recoveries (%R) meet the QC limits?

#	Date	Lab ID/Reference	Column	Surrogate Compound	%R (Limits)	Associated Samples	Qualifications
		2	Nst Spiked	A	87.8 (90-115)	2	J-1051P
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		
					()		

Letter Designation	Surrogate Compound	Recovery QC Limits (Soil)	Recovery QC Limits (Water)	Comments
A	Dichloroacetate			
B				

LDC #: 10215AC
 SDG #: 1125123

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: Inorganics, Method See Com

The correlation coefficient (r) for the calibration of Chlorite was recalculated. Calibration date: 6/7/08

An initial or continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration of each analyte in the ICV or CCV source

Type of Analysis	Analyte	Conc (mg/L) (units)	Area (units)	Recalculated		Reported		Acceptable (Y/N)
				r	%R	r	%R	
Initial calibration Calibration verification	Blank	0	2136					
	Standard 1	20	411221					
	Standard 2	100	2267407.8					
	Standard 3	200	4654440					
	Standard 4	400	9599573.3	0.999687		0.999687		Y
	Standard 5							
	Standard 6							
Calibration verification	Chlorite	200.0		104.8		NR		Y
	Cr6+	0.300		102.9		NR		Y
Calibration verification								

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

LDC #: 19215A6
 SDG #: 1851253

METHOD: Inorganics, Method See Com

Percent recoveries (%R) for a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$
 Where, Found = concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).
 True = concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$
 Where, S = Original sample concentration
 D = Duplicate sample concentration

Sample ID	Type of Analysis	Element	Found / S (units)	True / D (units)	Recalculated		Reported		Acceptable (Y/N)
					%R / RPD	%R / RPD			
8P23074-BS1	Laboratory control sample	Cu	0.3438	0.4	86	86			Y
	Matrix spike sample		(SSR-SR)						
	Duplicate sample								

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 19215AG
SDG #: 18F12A7

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
Reviewer: AJ
2nd reviewer: L

METHOD: Inorganics, Method Su Con

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- N N/A Have results been reported and calculated correctly?
- N N/A Are results within the calibrated range of the instruments?
- N N/A Are all detection limits below the CRQL?

Compound (analyte) results for _____ reported with a positive detect were recalculated and verified using the following equation:

Concentration =

Recalculation:

$$\text{Chlorit: } \frac{(4.088 \text{ g/L})(0.04 \text{ L})(5)}{(0.004 \text{ g})(0.80)} = 255.5 \text{ mg/kg}$$

#	Sample ID	Analyte	Reported Concentration (mg/kg)	Calculated Concentration (mg/kg)	Acceptable (Y/N)
	2	Chlorit	250	255.5	Y

Note: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: BRC Tronox Parcel F
Collection Date: June 10, 2008
LDC Report Date: August 6, 2008
Matrix: Soil
Parameters: Chlorite & Hexavalent Chromium
Validation Level: EPA Level III
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): IRF1298

Sample Identification

TSB-FR-02-02-20'
TSB-FR-02-02-30'
TSB-FJ-02-02-10'
TSB-FJ-02-02-20'
TSB-FJ-02-02-30'

Introduction

This data review covers 5 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 300.1 for Chlorite and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No chlorite or hexavalent chromium were found in the initial, continuing and preparation blanks.

No field blanks were identified in this SDG.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

**BRC Tronox Parcel F
Chlorite & Hexavalent Chromium - Data Qualification Summary - SDG IRF1298**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
Chlorite & Hexavalent Chromium - Laboratory Blank Data Qualification Summary -
SDG IRF1298**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel F
Chlorite & Hexavalent Chromium - Field Blank Data Qualification Summary - SDG
IRF1298**

No Sample Data Qualified in this SDG

LDC #: 19215B6
 SDG #: IRF1298
 Laboratory: Test America

VALIDATION COMPLETENESS WORKSHEET
 Level III

Date: 8/6/08
 Page: 1 of 1
 Reviewer: AI
 2nd Reviewer: [Signature]

METHOD: (Analyte) Chlorite (EPA Method 300.1), Hexavalent Chromium (EPA SW846 Method 7196A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>6/10/08</u>
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Surrogate Spikes	A	
V	Matrix Spike/Matrix Spike Duplicates	A	} <u>from IRF1298</u>
VI.	Duplicates	N	
VII.	Laboratory control samples	A	<u>LLS</u>
VIII.	Sample result verification	N	
IX.	Overall assessment of data	A	
X.	Field duplicates	N	
XI	Field blanks	N	

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples:

Am soil

1	TSB-FR-02-02-20'	11		21		31	
2	TSB-FR-02-02-30'	12		22		32	
3	TSB-FJ-02-02-10'	13		23		33	
4	TSB-FJ-02-02-20'	14		24		34	
5	TSB-FJ-02-02-30'	15		25		35	
6	PB	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC #: 19215A6
SDG #: 12F1298

VALIDATION FINDINGS WORKSHEET Sample Specific Analysis Reference

Page: 1 of 1
Reviewer: CA
2nd reviewer: ✓

All circled methods are applicable to each sample.

Sample ID	Parameter
1-5	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC <u>CR⁶⁺</u> Chlorite
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁶⁺

Comments: _____

