

TABLE OF CONTENTS

| | |
|---|----------|
| Cover Page | 2 |
| QC Checklist..... | 3 |
| Bench Sheet..... | 4 |
| Standards Preparation Worksheet and Certificates of Analysis | 5 |

Level IV Data Package

MWH Group 212263

Method: 7196 CRVI

Sample No.:

2708030231

2708030232

2708030233

2708030234

2708030235

SM 7196A QC Check List

Cr-VI

Analyst WBN

Analysis date 8/3/07 Reviewer/Date WV 8/6/07

Instrument: HACH DR/4000V

- All sample analyzed within 24 Hrs. holding time
- All sample raw concentration below the high standard or linear range of this batch if not, overranged samples marked for dilution and rerun

ALL QC

- All calibration verifications within +/- 10%
- LCS/LCSD within +/- 15%
- MS/MSD within +/-30%
- MRL within +/- 50%
- CCV within +/- 10%
- CCB/Blank <1/2 MRL
- RPD between MS/MSD is within +/-20%
- All pH of the samples are 7

- No more than 20 samples per batch
- MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples
- QIR needed for failed QC
- Special Det Code noted on the cover sheet

3 Aug 07

Walter Hsieh 8/13/07

| Instrument: Hach DR 4000V | | | | Cr-VI by EPA Method 7196A | | | | | | | |
|---------------------------------------|-----------|-------|-----------------|---|---------|-------------------------------------|----------|--------------------------------|----|------------------------|------|
| Analyst: <u>Walter Hsieh</u> | | | | Date: <u>8/3/07</u> | | Page: <u>1 of 1</u> | | Start time: <u>11:15</u> | | End time: <u>12:00</u> | |
| Cal. Stock Std: Use Built-In Curve in | | | | Cal. Stock Std exp. Date: <u>08/10/2007</u> | | Cal. Working Std: <u>WH070511-1</u> | | <u>12:15 → 12:1</u> | | | |
| LCS Stock Std: <u>HACH DR4000V</u> | | | | LCS Stock Std exp. Date: <u>08/10/2007</u> | | LCS Working Std: <u>WH070511-2</u> | | | | | |
| Correlation Coeff: <u>—</u> | | | | Slope: <u>—</u> | | Y-intercept: <u>—</u> | | LCS True Value: <u>0.05PPM</u> | | | |
| Sample | Dilution | | Turb Bik | | | Calc. | Reported | Sampling | pH | Comments | |
| | sample | total | (same dilution) | | res abs | | | | | | mg/L |
| | mls | mls | abs | abs | | mg/L | mg/L | | | | |
| Blank | 10 | 10 | 0.000 | | | 0.000 | 0.000 | | | | |
| Std 1 (0.005 ppm) | 10 | 10 | 0.010 | | | 0.005 | 0.005 | | | (100) | |
| Std 2 (0.020 ppm) | 10 | 10 | 0.038 | | | 0.020 | 0.020 | | | (100) | |
| Std 3 (0.050 ppm) | 10 | 10 | 0.092 | | | 0.050 | 0.050 | | | (100) | |
| Std 4 (0.200 ppm) | 10 | 10 | 0.360 | | | 0.197 | 0.197 | | | (98.5) | |
| Std 5 (0.500 ppm) | 10 | 10 | 0.883 | | | 0.483 | 0.483 | | | (96.6) | |
| LCS - 1 (0.05 ppm) | 10 | 10 | 0.076 | | | 0.052 | 0.052 | | | (100) Limit: 85 - 115% | |
| Sample Log # | 5 | 10 | 0.012 | | | 0.008 | 0.016 | | 7 | 13:52 | |
| 1 2708030247 | Discharge | 1.0 | 10 | 0.009 | | 0.005 | 0.005 | 8/1 | 9 | 11:51 | |
| 2 2708030248 | West | 0.5 | 10 | 0.178 | | 0.097 | 1.94 | | | 11:52 | |
| 3 2708030248ms | West MS | 0.5 | 10 | 0.264 | | 0.144 | 2.88 | | | 11:53 2.88-1.94 (74) | |
| 4 2708030248msd | West MS | 0.5 | 10 | 0.264 | | 0.145 | 2.90 | | | 11:54 2.90-1.94 (76) | |
| 5 2708030249 | EAST | 0.1 | 10 | 0.304 | | 0.166 | 16.6 | | | 11:55 1.03 (96) | |
| 6 2708030239 | SW-11 | 10 | 10 | 0.037 | | 0.020 | 0.020 | | | 11:56 | |
| 7 2708010343 | EB-2 | 10 | 10 | 0.002 | | 0.002 | 0.005 | 8/1 | 7 | 11:57 | |
| 8 2708010344 | MD-2 | 5 | 10 | | | OR | | | | | |
| 9 2708010344 | MD-2 | 1 | 10 | | | OR | | | | | |
| 10 2708010342 | M12A | 0.1 | 10 | 0.211 | | 0.143 | 14.3 | | | 11:58 | |
| 11 2708010344 | MD-2 | 0.1 | 10 | 0.260 | | 0.142 | 14.2 | | | 11:59 | |
| 12 LCS (0.05) | | | 10 | 0.092 | | 0.050 | 0.050 | | | (100) (100) | |
| 13 2708030232 | M56 | 0.5 | 10 | 0.350 | | 0.191 | 38.2 | 8/1 | 7 | 13:32 H1 1107 | |
| 14 2708030231 | M84 | 10 | 10 | 0.165 | | 0.090 | 0.090 | | | 13:33 H3 0905 | |
| 15 2708030234 | M11 | 0.5 | 10 | 0.260 | | 0.142 | 28.4 | | | 13:34 H1 1135 | |
| 16 2708030235 | M10 | 10 | 10 | 0.645 | | 0.350 | 0.350 | | | 13:35 H1 1235 | |
| 17 2708030232 | M100 | 10 | 10 | 0.398 | | 0.218 | 0.22 | | | 13:36 H3 10:30 | |
| 18 LCS (0.05) | | | 10 | 0.095 | | 0.052 | 0.052 | | | (100) | |
| 19 Blk | | | 10 | 0.000 | | 0.000 | 0.000 | | | | |
| 20 20 ppb | | | 10 | 0.037 | | 0.020 | 0.020 | | | (100) | |
| LCS - 2 (0.05 ppm) | | | 10 | 0.093 | | 0.051 | 0.051 | | | (100) Limit: 85 - 115% | |
| 0.020 ppm/CCV | | | 10 | 0.036 | | 0.020 | 0.020 | | | (110) Limit: 90-110% | |
| Blank/CCB | | | 10 | 0.000 | | 0.000 | 0.000 | | | | |

F F F F

H3

Analytical Reviewed By: Walter Hsieh QC Reviewed By: UV 8/6/07 Color Reagent: HACH Perma Chem Chroma Ver3
 Entered By: Walter Hsieh EXP. Date: 09/30/09 Lot #: A7045

REV. 02/06/07 wbn

H3 → 5x. Rec'd Past MT.
 all 4 5x's were sampled at 10:00. Tod Ex. Arrived past 10:00 the next day.
 4
 8/31/07

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Preparation Documentation

Reagent: 5 ppm Cr16 STD
 Date Received/Prepped: 2 1 6 1 0 7 1 1 1
 Date Expired: 5 1 6 1 0 7 1 1 1
 Manufacturer: WBM
 Storage Condition: Room Temp

MW #: WH070206-1
 By: WBN
 Matrix: AB 1/1 H₂SO₄
 Amount: 100 ml
 Lot #: _____

| Component | Comment | Standard | Concentration |
|--------------|--------------------|----------|---------------|
| <u>0.5ml</u> | <u># ME0604003</u> | | <u>5ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

Reagent: 5 ppm Cr16 QC
 Date Received/Prepped: 2 1 6 1 0 7 1 1 1
 Date Expired: 5 1 6 1 0 7 1 1 1
 Manufacturer: WBN
 Storage Condition: Room Temp

MW #: WH070206-2
 By: WBN
 Matrix: AB 1/1 H₂SO₄
 Amount: 100 ml
 Lot #: _____

| Component | Comment | Standard | Concentration |
|--------------|-----------------|----------|---------------|
| <u>0.5ml</u> | <u>R 201081</u> | | <u>5ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

Reagent: 5 ppm Cr16 Cl. STD.
 Date Received/Prepped: 5 1 1 1 0 7 1 1 1
 Date Expired: 8 1 1 0 1 0 7 1 1 1
 Manufacturer: WBN
 Storage Condition: Room Temp

MW #: WH070511-1
 By: WBN
 Matrix: AB 1/1 H₂SO₄
 Amount: 100 ml
 Lot #: _____

| Component | Comment | Standard | Concentration |
|--------------|-----------------|----------|---------------|
| <u>0.5ml</u> | <u>R 201632</u> | | <u>5ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

Reagent Preparation Documentation

Reagent: Cr⁶⁺ QC Std.
 Date Received/Prepped: 5 1 11 1 07 1 1 1
 Date Expired: 8 1 10 1 07 1 1
 Manufacturer: WGH
 Storage Condition: Room Temp

MW #: WH070511-2
 By: WSH
 Matrix: Ag/1% H₂SO₄
 Amount: 100ml
 Lot #: _____

| Component | Comment | Standard | Concentration |
|------------------------------|------------------|----------|---------------|
| <u>Cr⁶⁺ 0.5ml</u> | <u>Rh 201090</u> | | <u>5 ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

Reagent: 5 ppm Cr⁶⁺ Cal Std
 Date Received/Prepped: 8 1 9 1 07 1 1
 Date Expired: 11 1 9 1 07 1 1
 Manufacturer: FAS
 Storage Condition: Room Temp

MW #: Fj 080907-1
 By: Fj
 Matrix: aqueous 1% H₂SO₄
 Amount: 100 mL
 Lot #: _____

| Component | Comment | Standard | Concentration |
|-------------------------------|-----------------|----------|---------------|
| <u>Cr⁶⁺ 0.5 ml</u> | <u>R 201090</u> | | <u>5 ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

Reagent: Cr⁶⁺ QC std
 Date Received/Prepped: 8 1 9 1 07 1 1 1
 Date Expired: 11 1 9 1 07 1 1 1
 Manufacturer: FAS
 Storage Condition: Room Temp

MW #: Fj 080907-2
 By: Fj
 Matrix: aqueous 1% H₂SO₄
 Amount: 100 mL
 Lot #: _____

| Component | Comment | Standard | Concentration |
|-------------------------------|-----------------|----------|---------------|
| <u>Cr⁶⁺ 0.5 mL</u> | <u>R 201632</u> | | <u>5 ppm</u> |
| | | | |
| | | | |
| | | | |
| | | | |

Comment: _____

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM **1000 µg/mL Chromium (+6) in H2O**

R# 201632

Catalog Number: CGCR(6)1-1, CGCR(6)1-2, and CGCR(6)1-5
 Lot Number: **Z-CR02152**
 Starting Material: (NH₄)₂Cr₂O₇
 Starting Material Purity (%): 99.989259
 Starting Material Lot No: F04N14
 Matrix: H₂O

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Concentration: 1000 ± 3 µg/mL

Certified Density: 0.999 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

| | | |
|--|---|--|
| $\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$ | $\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$ | <p>(\bar{x}) = mean x_i = individual results n = number of measurements $\sum s_i$ = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)</p> |
|--|---|--|

The independent samples t-test was used to determine if there is agreement between the above assay methods at the 95% confidence interval. Both methods were compared and showed agreement within the stated uncertainties. This agreement is a confirmation of the accuracy of this CRM.

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

- "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)
- This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 Assay Method #1 **1000 ± 3 µg/mL**
 Redox NIST SRM 136e Lot Number: 980702

Assay Method #2 **1001 ± 4 µg/mL**
 ICP Assay NIST SRM 3112a Lot Number: 990607

- 4.2 **BALANCE CALIBRATION** - All balances are checked daily using an in-house procedure. The weights used for testing are annually compared to master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 692476 - Class 1 and 692476A - Class 2. The NIST test number is 822/260017-98. All analytical balances are calibrated every 4 months. The balances are calibrated with a class 1 and/or class 2 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is 822/260017-98.
- 4.3 **THERMOMETER CALIBRATION** - The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification Nos. 92564, 119016, 471047 and NIST test report Nos. 811/258522, 811/2557078, and 236090.
- 4.4 **GLASSWARE CALIBRATION** - An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM's.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP/MS AND ICP-OES IN µg/mL

CRM's solutions are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

| | | | | |
|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|
| <u>Q</u> Al < 0.00090 | <u>M</u> Dy < 0.01185 | <u>Q</u> Li < 0.00002 | <u>M</u> Pr < 0.00059 | <u>M</u> Te < 0.05923 |
| <u>M</u> Sb < 0.00099 | <u>M</u> Er < 0.00987 | <u>M</u> Lu < 0.00079 | <u>M</u> Re < 0.00197 | <u>M</u> Tb < 0.00059 |
| <u>M</u> As < 0.01974 | <u>M</u> Eu < 0.00592 | <u>Q</u> Mg < 0.00030 | <u>M</u> Rh < 0.00197 | <u>M</u> Tl < 0.00197 |
| <u>M</u> Ba < 0.01974 | <u>M</u> Gd < 0.00197 | <u>M</u> Mn < 0.00790 | <u>M</u> Rb < 0.00197 | <u>M</u> Th < 0.00197 |
| <u>M</u> Be < 0.00099 | <u>M</u> Ga < 0.00197 | <u>Q</u> Hg < 0.01500 | <u>M</u> Ru < 0.00395 | <u>M</u> Tm < 0.00079 |
| <u>M</u> Bi < 0.00079 | <u>M</u> Ge < 0.01185 | <u>M</u> Mo < 0.00395 | <u>M</u> Sm < 0.00197 | <u>M</u> Sn < 0.00987 |
| <u>Q</u> B < 0.01000 | <u>M</u> Au < 0.00592 | <u>M</u> Nd < 0.00395 | <u>M</u> Sc < 0.01974 | <u>Q</u> Ti < 0.00100 |
| <u>M</u> Cd < 0.00592 | <u>M</u> Hf < 0.00395 | <u>M</u> Ni < 0.01579 | <u>M</u> Se < 0.01579 | <u>M</u> W < 0.01974 |
| <u>Q</u> Ca 0.00027 | <u>M</u> Ho < 0.00099 | <u>M</u> Nb < 0.00099 | <u>Q</u> Si < 0.20000 | <u>M</u> U < 0.00395 |
| <u>M</u> Ce < 0.00987 | <u>M</u> In < 0.01974 | <u>n</u> Os | <u>M</u> Ag < 0.00395 | <u>Q</u> V < 0.02000 |
| <u>M</u> Cs < 0.00059 | <u>M</u> Ir < 0.00987 | <u>M</u> Pd < 0.00987 | <u>Q</u> Na 0.00300 | <u>M</u> Yb < 0.00197 |
| <u>s</u> Cr | <u>Q</u> Fe < 0.01000 | <u>Q</u> P < 0.04000 | <u>M</u> Sr < 0.00099 | <u>M</u> Y < 0.07897 |
| <u>M</u> Co < 0.00592 | <u>M</u> La < 0.00099 | <u>M</u> Pt < 0.00395 | <u>i</u> S | <u>Q</u> Zn < 0.00400 |
| <u>M</u> Cu < 0.01185 | <u>M</u> Pb < 0.00592 | <u>Q</u> K 0.32485 | <u>M</u> Ta < 0.01382 | <u>M</u> Zr < 0.00987 |

M - Checked by ICP-MS Q - Checked by ICP-OES i - Spectral Interference n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

For the calibration of analytical instruments including but not limited to the following:
ICP-MS, ICP-OES, FAAS, GFAA, XRF, and DCP

For the validation of analytical methods

For the preparation of "working reference samples"

For interference studies and the determination of correction coefficients

For detection limit and linearity studies

For additional intended uses, contact Technical Staff

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

Storage & Handling - Keep tightly sealed when not in use. Store and use at $20 \pm 4^\circ\text{C}$. Do not pipet from container. Do not return portions removed for pipetting to container.

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 51.9961; +3; 6; Cr2O7²⁻

Chemical Compatibility - Stable in HCl, HNO₃, H₂SO₄, HF, H₃PO₄. Avoid basic media. Stable with most metals (Except Ba and Pb) and inorganic anions in acidic media.

Stability - 2-100 ppb levels stable for months in 1% HNO₃ / LDPE container as Total Cr, however, the stability of this oxidation state is unknown. 1-10,000 ppm solutions chemically stable for years in 0 - 1% HNO₃ / LDPE container.

Cr Containing Samples (Preparation and Solution) - Metal (soluble in HCl); Oxides/Ores (Chrome ore/oxides are very difficult to dissolve. The following procedures [A-D] are commonly used:

A. Fusion with KHSO₄ and extraction with hot KCl. The residue fused with Na₂CO₃ and KClO₃, 3:1.

B. Fusion with NaKSO₄ and NaF, 2:1.

C. Fusion with magnesia or lime and sodium or potassium carbonates, 4:1.

D. Fusion with Na₂O₂ or NaOH and KNO₃ or NaOH and Na₂O₂.

Nickel, iron, copper, or silver crucibles should be used for A, B and C);

Organic Matrices (Ash at 450EC followed by one of the fusion methods above or sulfuric/hydrogen peroxide acid digestions may be applicable to non oxide containing samples).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

| Technique/Line | Estimated D.L. | Order | Type | Interferences (underlined indicates severe) |
|--------------------|----------------------|-------|------|--|
| ICP-OES 205.552 nm | 0.006 / 0.0008 µg/mL | 1 | ion | <u>Os</u> |
| ICP-OES 284.325 nm | 0.008 / 0.0007 µg/mL | 1 | ion | |
| ICP-OES 276.654 nm | 0.01 / 0.001 µg/mL | 1 | ion | Cu, Ta, <u>V</u> |
| ICP-MS 52 amu | 40 ppt | n/a | M+ | 36S16O, 36Ar16O - The 50Cr, 53Cr, 54Cr lines suffer from many more potential interferences from sulfur, chlorine and argon compounds of oxygen, nitrogen and carbon. |

8.0 HAZARDOUS INFORMATION - Please refer to the enclosed Material Safety Data sheet for information regarding this CRM.

9.0 HOMOGENEITY - This solution was mixed according to an in house procedure and is guaranteed to be homogeneous.

10.0 QUALITY STANDARD DOCUMENTATION



10.1 ISO 9001:2000 Quality Management System Registration - QMI Certificate Number 010105

Recognized by:

Registrar Accreditation Board (ANSI-RAB)

Standards Council of Canada (SCC)

Dutch Council for Accreditation (RVA)

Entidad Mexicana de Acreditacion, a.c.(EMA)

Members of IQ Net International Certification Network:

Argentina (IRAM), Australia (QAS), Austria (OQS), Belgium (Avinter), Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JOA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)

10.2 ISO/IEC 17025 - 1999 "General Requirements for the Competence of Testing and Calibration"

- Chemical Testing - Accredited A2LA Certificate Number 883.01

10.3 ISO/IEC Guide 34 - 2000 "General Requirements for the Competence of Reference Material Producers"

- Reference Materials Production - Accredited A2LA Certificate Number 883.02

A2LA Mutual Recognition Agreement Partners:

Australia (NATA), Austria (BmWA), Belgium (BELTEST) (BKO-OBE), Canada (SCC), Chinese Taipei (CNLA), Czech Republic (NAO), Denmark (DANAK), Finland (FINAS), France (COFRAC), Germany (DAR), Hong Kong (HKAS), Ireland (NAB), Italy (SIT) (SINAL), Japan (JAB) (JNLA), Republic of Korea (KOLAS), The Netherlands (RVA), New Zealand (IANZ), Norway (NA), Portugal (IPQ), Singapore (SAC-SINGLAS), Spain (ENAC), Sweden (SWEDAC), Switzerland (SAS), United Kingdom (UKAS) and United States (NVLAP) (ICBO ES)

10.4 10CFR50 Appendix B - Nuclear Regulatory Commission

- Domestic Licensing of Production and Utilization Facilities

10.5 10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance

10.6 MIL-STD-45662A (Obsolete/Observed)

11.0 DATE OF CERTIFICATION AND PERIOD OF VALIDITY

11.1 Shelf Life - The period of time during which the concentration of the analyte(s) in a properly packaged, unopened, and unused standard stored under environmentally controlled and monitored conditions will remain within the specified uncertainty range. Shelf life is limited primarily by transpiration (loss of water from the solution) and infrequently, by chemical instability. Transpiration studies of chemically-stable solutions performed at the manufacturer's facility show a CRM shelf-life of twenty one months for solutions packaged in 125-mL low density polyethylene bottles. When stored under special conditions that minimize transpiration and instability, the shelf life can be extended past this limit.

11.2 Expiration Date - The date after which a CRM should not be used. Routine laboratory use of a CRM increases transpiration losses and the chance of contamination which affect the integrity of the CRM and limit its useful life. Manufacturer concurs with state and federal regulatory agencies' recommendations that solution standards be assigned a one-year expiration date.

Certification Date: December 06, 2006

Expiration Date: **EXPIRES**
01/02/2007

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By: Nick Maida, Product Documentation Administrator *Nickolas Maida*

Certificate Approved By: Katalin Le, QC Manager *Katalin Le*

Certifying Officer: Paul Gaines, PhD., Senior Technical Director *Paul R. Gaines*