



**LABORATORY DATA CONSULTANTS, INC.**

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

ERM  
2525 Natomas Park Drive, Suite 350  
Sacramento, CA 95833  
ATTN: Ms. Maria Barajas-Albalawi

January 22, 2008

SUBJECT: BRC Tronox Parcel C/D/F/G, Data Validation

Dear Ms. Barajas-Albalawi

Enclosed are the final validation reports for the fraction listed below. These SDGs were received on January 4, 2008. Attachment 1 is a summary of the samples that were reviewed for each analysis.

**LDC Project # 18054:**

<b><u>SDG #</u></b>	<b><u>Fraction</u></b>
F7K270268, F7K280229, F7K290114, F7K290369	Volatiles

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996; update IIIA, April 1998

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto  
Operations Manager/Senior Chemist



**BRC Tronox Parcel C/D/F/G  
Data Validation Reports  
LDC# 18054**

Volatiles

LDC

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** BRC Tronox Parcel C/D/F/G  
**Collection Date:** November 21, 2007  
**LDC Report Date:** January 15, 2008  
**Matrix:** Air  
**Parameters:** Volatiles  
**Validation Level:** EPA Level III  
**Laboratory:** TestAmerica, Inc.

**Sample Delivery Group (SDG):** F7K270268

**Sample Identification**

TSB-CR-07  
TSB-CR-06  
TSB-DR-06  
TSB-CJ-08

## Introduction

This data review covers 4 air samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method TO-14A for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.

J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.

J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.

UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds with the following exceptions:

Date	Compound	%RSD	Associated Samples	Flag	A or P
11/27/07	4-Ethyltoluene	32.177	All samples in SDG F7K270268	J (all detects) UJ (all non-detects)	P

Average relative response factors (RRF) for all volatile target compounds and system monitoring compounds were within validation criteria.

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
12/03/07	4-Ethyltoluene 1,2,4-Trimethylbenzene	32.4 34.0	All samples in SDG F7K270268	J+ (all detects) J+ (all detects)	P

All of the continuing calibration RRF values were within validation criteria.

## V. Blanks

Method blank analyses were performed at the required frequency. No volatile contaminants were found in the method blanks.

No field blanks were identified in this SDG.

**VI. Surrogate Spikes**

Surrogates were not required by the method.

**VII. Matrix Spike/Matrix Spike Duplicates**

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

**VIII. Laboratory Control Samples (LCS)**

Although laboratory control samples were not required by the method, laboratory control samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

**IX. Regional Quality Assurance and Quality Control**

Not applicable.

**X. Internal Standards**

All internal standard areas and retention times were within QC limits.

**XI. Target Compound Identifications**

Raw data were not reviewed for this SDG.

**XII. Compound Quantitation and CRQLs**

Raw data were not reviewed for this SDG.

**XIII. Tentatively Identified Compounds (TICs)**

Raw data were not reviewed for this SDG.

**XIV. System Performance**

Raw data were not reviewed for this SDG.

**XV. Overall Assessment of Data**

Data flags are summarized at the end of this report if data has been qualified.

**XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**BRC Tronox Parcel C/D/F/G  
Volatiles - Data Qualification Summary - SDG F7K270268**

SDG	Sample	Compound	Flag	A or P	Reason
F7K270268	TSB-CR-07 TSB-CR-06 TSB-DR-06 TSB-CJ-08	4-Ethyltoluene	J (all detects) UJ (all non-detects)	P	Initial calibration (%RSD)
F7K270268	TSB-CR-07 TSB-CR-06 TSB-DR-06 TSB-CJ-08	4-Ethyltoluene 1,2,4-Trimethylbenzene	J+ (all detects) J+ (all detects)	P	Continuing calibration (%D)

**BRC Tronox Parcel C/D/F/G  
Volatiles - Laboratory Blank Data Qualification Summary - SDG F7K270268**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel C/D/F/G  
Volatiles - Field Blank Data Qualification Summary - SDG F7K270268**

No Sample Data Qualified in this SDG



LDC #: 18054A48  
 SDG #: F7K270268  
 Laboratory: Test America

**VALIDATION COMPLETENESS WORKSHEET**

Level III

Date: 1/14/08  
 Page: 1 of 1  
 Reviewer: JYB  
 2nd Reviewer: [Signature]

**METHOD:** GC/MS Volatiles (EPA Method TO-14A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>11/21/07</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	SW	
IV.	Continuing calibration/ICV	SW	
V.	Blanks	A	
VI.	Surrogate spikes	N	
VII.	Matrix spike/Matrix spike duplicates	N	
VIII.	Laboratory control samples	A	<u>LCS / D</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	N	

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet  
 ND = No compounds detected  
 R = Rinsate  
 FB = Field blank  
 D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples: Air

1	TSB-CR-07	11	21	31
2	TSB-CR-06	12	22	32
3	TSB-DR-06	13	23	33
4	TSB-CJ-08	14	24	34
5	<u>7338609 MB</u>	15	25	35
6		16	26	36
7		17	27	37
8		18	28	38
9		19	29	39
10		20	30	40

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method TO-14/TO-14A)

A. Chloromethane*	S. Trichloroethene	KK. Trichlorofluoromethane	CCC. tert-Butylbenzene	UUU. Benzyl chloride
B. Bromomethane	T. Dibromochloromethane	LL. Methyl-tert-butyl ether	DDD. 1,2,4-Trimethylbenzene	VVV. 4-Ethyltoluene
C. Vinyl chloride**	U. 1,1,2-Trichloroethane	MM. 1,2-Dibromo-3-chloropropane	EEE. sec-Butylbenzene	WWW. Ethanol
D. Chloroethane	V. Benzene	NN. Diethyl ether	FFF. 1,3-Dichlorobenzene	XXX. Ethyl ether
E. Methylene chloride	W. trans-1,3-Dichloropropene	OO. 2,2-Dichloropropane	GGG. p-Isopropyltoluene	YYY. tert-Butanol
F. Acetone	X. Bromoform*	PP. Bromochloromethane	HHH. 1,4-Dichlorobenzene	ZZZ. tert-Butyl alcohol
G. Carbon disulfide	Y. 4-Methyl-2-pentanone	QQ. 1,1-Dichloropropene	III. n-Butylbenzene	AAAA. Ethyl tert-butyl ether
H. 1,1-Dichloroethene**	Z. 2-Hexanone	RR. Dibromomethane	JJJ. 1,2-Dichlorobenzene	BBBB. tert-Amyl methyl ether
I. 1,1-Dichloroethane*	AA. Tetrachloroethene	SS. 1,3-Dichloropropane	KKK. 1,2,4-Trichlorobenzene	CCCC. 1-Chlorohexane
J. 1,2-Dichloroethene, total	BB. 1,1,2,2-Tetrachloroethane*	TT. 1,2-Dibromoethane	LLL. Hexachlorobutadiene	DDDD. Isopropyl alcohol
K. Chloroform**	CC. Toluene**	UU. 1,1,1,2-Tetrachloroethane	MMM. Naphthalene	EEEE. Acetonitrile
L. 1,2-Dichloroethane	DD. Chlorobenzene*	VV. Isopropylbenzene	NNN. 1,2,3-Trichlorobenzene	FFFF. Acrolein
M. 2-Butanone	EE. Ethylbenzene**	WW. Bromobenzene	OOO. 1,3,5-Trichlorobenzene	GGGG. Acrylonitrile
N. 1,1,1-Trichloroethane	FF. Styrene	XX. 1,2,3-Trichloropropane	PPP. trans-1,2-Dichloroethene	HHHH. 1,4-Dioxane
O. Carbon tetrachloride	GG. Xylenes, total	YY. n-Propylbenzene	QQQ. cis-1,2-Dichloroethene	IIII. Isobutyl alcohol
P. Bromodichloromethane	HH. Vinyl acetate	ZZ. 2-Chlorotoluene	RRR. m,p-Xylenes	JJJJ. Methacrylonitrile
Q. 1,2-Dichloropropane**	II. 2-Chloroethylvinyl ether	AAA. 1,3,5-Trimethylbenzene	SSS. o-Xylene	KKKK. Propionitrile
R. cis-1,3-Dichloropropene	JJ. Dichlorodifluoromethane	BBB. 4-Chlorotoluene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane	LLLL

\* = System performance check compounds (SPCC) for RRF ; \*\* = Calibration check compounds (CCC) for %RSD.

LDC #: 18054 A-48

SDG #: See Copy

# VALIDATION FINDINGS WORKSHEET

## Initial Calibration

Page: 1 of 1  
Reviewer: JVL  
2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA TO-14/TO-14A)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Did the laboratory perform a 5 point calibration prior to sample analysis?

Y N N/A Were all percent relative standard deviations (%RSD)  $\leq 30\%$  and relative response factors (RRF)  $\geq 0.05$ ?

#	Date	Standard ID	Compound	Finding %RSD (Limit: $\leq 30.0\%$ )	Finding RRF (Limit: $\geq 0.05$ )	Associated Samples	Qualifications
	11/27/07	1CAL	V V Y	32.177		All + B1K	J / UJ / P

METHOD: GC/MS VOA (EPA TO-14/TO-14A)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

N  N/A Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?

N  N/A Were all percent differences (%D)  $\leq$  30% and relative response factors (RRF)  $\geq$  0.05?

#	Date	Standard ID	Compound	Finding %D (Limit: $\leq$ 30.0%)	Finding RRF (Limit: $\geq$ 0.05)	Associated Samples	Qualifications
	12/02/07	CC12033	VVV (F)	32.4		All + B1K ↓	J+ acts / P ↓
			DDD (F)	34.0			

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** BRC Tronox Parcel C/D/F/G  
**Collection Date:** November 27, 2007  
**LDC Report Date:** January 15, 2008  
**Matrix:** Air  
**Parameters:** Volatiles  
**Validation Level:** EPA Level III  
**Laboratory:** TestAmerica, Inc.

**Sample Delivery Group (SDG):** F7K280229

**Sample Identification**

TSB-CR-02  
TSB-CR-03  
TSB-DR-02  
TSB-DJ-01  
TSB-DR-04  
TSB-DR-03  
TSB-DR-05

## Introduction

This data review covers 7 air samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method TO-14A for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.

J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.

J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.

UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

Average relative response factors (RRF) for all volatile target compounds and system monitoring compounds were within validation criteria.

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

All of the continuing calibration RRF values were within validation criteria.

## V. Blanks

Method blank analyses were performed at the required frequency. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
7351601-MB	12/14/07	Methylene chloride	1.6 ppb(v/v)	All samples in SDG F7K280229

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
TSB-CR-02	Methylene chloride	7.4 ppb(v/v)	7.4U ppb(v/v)
TSB-DR-02	Methylene chloride	2.6 ppb(v/v)	2.6U ppb(v/v)
TSB-DJ-01	Methylene chloride	3.6 ppb(v/v)	3.6U ppb(v/v)
TSB-DR-04	Methylene chloride	2.1 ppb(v/v)	2.1U ppb(v/v)
TSB-DR-03	Methylene chloride	1.6 ppb(v/v)	2.0U ppb(v/v)
TSB-DR-05	Methylene chloride	2.6 ppb(v/v)	2.6U ppb(v/v)

No field blanks were identified in this SDG.

#### **VI. Surrogate Spikes**

Surrogates were not required by the method.

#### **VII. Matrix Spike/Matrix Spike Duplicates**

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

#### **VIII. Laboratory Control Samples (LCS)**

Although laboratory control samples were not required by the method, laboratory control samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

#### **IX. Regional Quality Assurance and Quality Control**

Not applicable.

#### **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

#### **XI. Target Compound Identifications**

Raw data were not reviewed for this SDG.



## **XII. Compound Quantitation and CRQLs**

Raw data were not reviewed for this SDG.

## **XIII. Tentatively Identified Compounds (TICs)**

Raw data were not reviewed for this SDG.

## **XIV. System Performance**

Raw data were not reviewed for this SDG.

## **XV. Overall Assessment of Data**

Data flags are summarized at the end of this report if data has been qualified.

## **XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Data Qualification Summary - SDG F7K280229**

No Sample Data Qualified in this SDG

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Laboratory Blank Data Qualification Summary - SDG F7K280229**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
F7K280229	TSB-CR-02	Methylene chloride	7.4U ppb(v/v)	A
F7K280229	TSB-DR-02	Methylene chloride	2.6U ppb(v/v)	A
F7K280229	TSB-DJ-01	Methylene chloride	3.6U ppb(v/v)	A
F7K280229	TSB-DR-04	Methylene chloride	2.1U ppb(v/v)	A
F7K280229	TSB-DR-03	Methylene chloride	2.0U ppb(v/v)	A
F7K280229	TSB-DR-05	Methylene chloride	2.6U ppb(v/v)	A

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Field Blank Data Qualification Summary - SDG F7K280229**

No Sample Data Qualified in this SDG

LDC #: 18054B48  
 SDG #: F7K280229  
 Laboratory: Test America

**VALIDATION COMPLETENESS WORKSHEET**

Level III

Date: 1/14/08  
 Page: 1 of 1  
 Reviewer: JTB  
 2nd Reviewer: [Signature]

**METHOD:** GC/MS Volatiles (EPA Method TO-14A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11/27/07
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	
IV.	Continuing calibration/ICV	A	
V.	Blanks	SW	
VI.	Surrogate spikes	N	
VII.	Matrix spike/Matrix spike duplicates	N	
VIII.	Laboratory control samples	A	LCS 1D
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	N	

Note: A = Acceptable      ND = No compounds detected      D = Duplicate  
 N = Not provided/applicable      R = Rinsate      TB = Trip blank  
 SW = See worksheet      FB = Field blank      EB = Equipment blank

Validated Samples:

Air

1	TSB-CR-02	11	7351601 MB	21	31
2	TSB-CR-03	12	7351608 MB (K)	22	32
3	TSB-DR-02	13		23	33
4	TSB-DJ-01	14		24	34
5	TSB-DR-04	15		25	35
6	TSB-DR-03	16		26	36
7	TSB-DR-05	17		27	37
8		18		28	38
9		19		29	39
10		20		30	40

# TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method TO-14/TO-14A)

A. Chloromethane*	S. Trichloroethene	KK. Trichlorofluoromethane	CCC. tert-Butylbenzene	UUU. Benzyl chloride
B. Bromomethane	T. Dibromochloromethane	LL. Methyl-tert-butyl ether	DDD. 1,2,4-Trimethylbenzene	VVV. 4-Ethyltoluene
C. Vinyl chloride**	U. 1,1,2-Trichloroethane	MM. 1,2-Dibromo-3-chloropropane	EEE. sec-Butylbenzene	WWW. Ethanol
D. Chloroethane	V. Benzene	NN. Diethyl ether	FFF. 1,3-Dichlorobenzene	XXX. Ethyl ether
E. Methylene chloride	W. trans-1,3-Dichloropropene	OO. 2,2-Dichloropropane	GGG. p-Isopropyltoluene	YYY. tert-Butanol
F. Acetone	X. Bromoform*	PP. Bromochloromethane	HHH. 1,4-Dichlorobenzene	ZZZ. tert-Butyl alcohol
G. Carbon disulfide	Y. 4-Methyl-2-pentanone	QQ. 1,1-Dichloropropene	III. n-Butylbenzene	AAAA. Ethyl tert-butyl ether
H. 1,1-Dichloroethene**	Z. 2-Hexanone	RR. Dibromomethane	JJJ. 1,2-Dichlorobenzene	BBBB. tert-Amyl methyl ether
I. 1,1-Dichloroethane*	AA. Tetrachloroethene	SS. 1,3-Dichloropropane	KKK. 1,2,4-Trichlorobenzene	CCCC. 1-Chlorohexane
J. 1,2-Dichloroethene, total	BB. 1,1,2,2-Tetrachloroethane*	TT. 1,2-Dibromoethane	LLL. Hexachlorobutadiene	DDDD. Isopropyl alcohol
K. Chloroform**	CC. Toluene**	UU. 1,1,1,2-Tetrachloroethane	MMM. Naphthalene	EEEE. Acetonitrile
L. 1,2-Dichloroethane	DD. Chlorobenzene*	VV. Isopropylbenzene	NNN. 1,2,3-Trichlorobenzene	FFFF. Acrolein
M. 2-Butanone	EE. Ethylbenzene**	WW. Bromobenzene	OOO. 1,3,5-Trichlorobenzene	GGGG. Acrylonitrile
N. 1,1,1-Trichloroethane	FF. Styrene	XX. 1,2,3-Trichloropropane	PPP. trans-1,2-Dichloroethene	HHHH. 1,4-Dioxane
O. Carbon tetrachloride	GG. Xylenes, total	YY. n-Propylbenzene	QQQ. cis-1,2-Dichloroethene	IIII. Isobutyl alcohol
P. Bromodichloromethane	HH. Vinyl acetate	ZZ. 2-Chlorotoluene	RRR. m,p-Xylenes	JJJJ. Methacrylonitrile
Q. 1,2-Dichloropropane**	II. 2-Chloroethylvinyl ether	AAA. 1,3,5-Trimethylbenzene	SSS. o-Xylene	KKKK. Propionitrile
R. cis-1,3-Dichloropropene	JJ. Dichlorodifluoromethane	BBB. 4-Chlorotoluene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane	LLLL

\* = System performance check compounds (SPCC) for RRF ; \*\* = Calibration check compounds (CCC) for %RSD.



**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** BRC Tronox Parcel C/D/F/G  
**Collection Date:** November 26, 2007  
**LDC Report Date:** January 15, 2008  
**Matrix:** Air  
**Parameters:** Volatiles  
**Validation Level:** EPA Level III  
**Laboratory:** TestAmerica, Inc.

**Sample Delivery Group (SDG):** F7K290114

**Sample Identification**

TSB-CJ-04  
TSB-CJ-07  
TSB-CJ-03  
TSB-CJ-05  
TSB-CJ-02  
TSB-CJ-01  
TSB-CJ-01(FD)  
TSB-CR-01

## Introduction

This data review covers 8 air samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method TO-14A for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds with the following exceptions:

Date	Compound	%RSD	Associated Samples	Flag	A or P
11/27/07	4-Ethyltoluene	32.177	All samples in SDG F7K290114	J (all detects) UJ (all non-detects)	P

Average relative response factors (RRF) for all volatile target compounds and system monitoring compounds were within validation criteria.

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

All of the continuing calibration RRF values were within validation criteria.

## V. Blanks

Method blank analyses were performed at the required frequency. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
7339519-MB	12/4/07	Methylene chloride	1.2 ppb(v/v)	All samples in SDG F7K290114



Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
TSB-CJ-04	Methylene chloride	1.5 ppb(v/v)	2.0U ppb(v/v)
TSB-CJ-07	Methylene chloride	1.6 ppb(v/v)	2.0U ppb(v/v)
TSB-CJ-03	Methylene chloride	2.5 ppb(v/v)	2.5U ppb(v/v)
TSB-CJ-05	Methylene chloride	2.5 ppb(v/v)	2.5U ppb(v/v)
TSB-CJ-02	Methylene chloride	8.2 ppb(v/v)	8.2U ppb(v/v)
TSB-CJ-01	Methylene chloride	7.7 ppb(v/v)	7.7U ppb(v/v)
TSB-CJ-01 (FD)	Methylene chloride	7.8 ppb(v/v)	7.8U ppb(v/v)
TSB-CR-01	Methylene chloride	10 ppb(v/v)	10U ppb(v/v)

No field blanks were identified in this SDG.

## VI. Surrogate Spikes

Surrogates were not required by the method.

## VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

## VIII. Laboratory Control Samples (LCS)

Although laboratory control samples were not required by the method, laboratory control samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## IX. Regional Quality Assurance and Quality Control

Not applicable.

## X. Internal Standards

All internal standard areas and retention times were within QC limits.

## XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

## XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

## XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

## XIV. System Performance

Raw data were not reviewed for this SDG.

## XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

## XVI. Field Duplicates

Samples TSB-CJ-01 and TSB-CJ-01(FD) were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ppb(v/v))		RPD (Limits)	Difference (Limits)	Flag	A or P
	TSB-CJ-01	TSB-CJ-01(FD)				
Chloromethane	1.4	1.6	-	0.2 ( $\leq 4.0$ )	-	-
Vinyl chloride	4.0	4.4	-	0.4 ( $\leq 3.0$ )	-	-
Chloroethane	35	39	11 ( $\leq 50$ )	-	-	-
Acetone	32	41	-	9 ( $\leq 10.0$ )	-	-
Methylene chloride	7.7	7.8	-	0.1 ( $\leq 2.0$ )	-	-
1,1-Dichloroethane	56	64	13 ( $\leq 50$ )	-	-	-
Benzene	1.7	2.1	-	0.4 ( $\leq 3.0$ )	-	-

Compound	Concentration (ppb(v/v))		RPD (Limits)	Difference (Limits)	Flag	A or P
	TSB-CJ-01	TSB-CJ-01(FD)				
1,2-Dichloroethane	11	13	17 ( $\leq 50$ )	-	-	-
Trichloroethene	3.2	4.0	-	0.8 ( $\leq 2.0$ )	-	-
Bromodichloromethane	5.2	6.4	-	1.2 ( $\leq 2.0$ )	-	-
Toluene	1.7	2.0	-	0.3 ( $\leq 2.0$ )	-	-
1,1,2-Trichloroethane	3.4	3.9	-	0.5 ( $\leq 2.0$ )	-	-
Tetrachloroethene	210	230	9 ( $\leq 50$ )	-	-	-
Chlorobenzene	2.0U	1.2	-	0.8 ( $\leq 2.0$ )	-	-
1,3-Dichlorobenzene	1.4	1.6	-	0.2 ( $\leq 2.0$ )	-	-
Chloroform	3100	3700	18 ( $\leq 50$ )	-	-	-

**BRC Tronox Parcel C/D/F/G  
Volatiles - Data Qualification Summary - SDG F7K290114**

SDG	Sample	Compound	Flag	A or P	Reason
F7K290114	TSB-CJ-04 TSB-CJ-07 TSB-CJ-03 TSB-CJ-05 TSB-CJ-02 TSB-CJ-01 TSB-CJ-01 (FD) TSB-CR-01	4-Ethyltoluene	J (all detects) UJ (all non-detects)	P	Initial calibration (%RSD)

**BRC Tronox Parcel C/D/F/G  
Volatiles - Laboratory Blank Data Qualification Summary - SDG F7K290114**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
F7K290114	TSB-CJ-04	Methylene chloride	2.0U ppb(v/v)	A
F7K290114	TSB-CJ-07	Methylene chloride	2.0U ppb(v/v)	A
F7K290114	TSB-CJ-03	Methylene chloride	2.5U ppb(v/v)	A
F7K290114	TSB-CJ-05	Methylene chloride	2.5U ppb(v/v)	A
F7K290114	TSB-CJ-02	Methylene chloride	8.2U ppb(v/v)	A
F7K290114	TSB-CJ-01	Methylene chloride	7.7U ppb(v/v)	A
F7K290114	TSB-CJ-01 (FD)	Methylene chloride	7.8U ppb(v/v)	A
F7K290114	TSB-CR-01	Methylene chloride	10U ppb(v/v)	A

**BRC Tronox Parcel C/D/F/G  
Volatiles - Field Blank Data Qualification Summary - SDG F7K290114**

No Sample Data Qualified in this SDG

LDC #: 18054C48  
 SDG #: F7K290114  
 Laboratory: Test America

**VALIDATION COMPLETENESS WORKSHEET**

Level III

Date: 1/17/08  
 Page: 1 of 1  
 Reviewer: JVL  
 2nd Reviewer:

**METHOD:** GC/MS Volatiles (EPA Method TO-14A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11/26/07
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	SW	
IV.	Continuing calibration/ICV	A	
V.	Blanks	SW	
VI.	Surrogate spikes	N	
VII.	Matrix spike/Matrix spike duplicates	N	
VIII.	Laboratory control samples	A	LCS /b
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 6, 7
XVII.	Field blanks	N	

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet

ND = No compounds detected  
 R = Rinsate  
 FB = Field blank

D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples:

Air

1	1	TSB-CJ-04	11	7339519 MB	21	31
2	1	TSB-CJ-07	12	7340411 MB	22	32
3	1	TSB-CJ-03	13		23	33
4	1	TSB-CJ-05	14		24	34
5	1/2	TSB-CJ-02	15		25	35
6	1/2	TSB-CJ-01	16		26	36
7	1/2	TSB-CJ-01(FD)	17		27	37
8	1/2	TSB-CR-01	18		28	38
9			19		29	39
10			20		30	40

# TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method TO-14/TO-14A)

A. Chloromethane*	S. Trichloroethene	KK. Trichlorofluoromethane	CCC. tert-Butylbenzene	UUU. Benzyl chloride
B. Bromomethane	T. Dibromochloromethane	LL. Methyl-tert-butyl ether	DDD. 1,2,4-Trimethylbenzene	VVV. 4-Ethyltoluene
C. Vinyl chloride**	U. 1,1,2-Trichloroethane	MM. 1,2-Dibromo-3-chloropropane	EEE. sec-Butylbenzene	WWW. Ethanol
D. Chloroethane	V. Benzene	NN. Diethyl ether	FFF. 1,3-Dichlorobenzene	XXX. Ethyl ether
E. Methylene chloride	W. trans-1,3-Dichloropropene	OO. 2,2-Dichloropropane	GGG. p-Isopropyltoluene	YYY. tert-Butanol
F. Acetone	X. Bromoform*	PP. Bromochloromethane	HHH. 1,4-Dichlorobenzene	ZZZ. tert-Butyl alcohol
G. Carbon disulfide	Y. 4-Methyl-2-pentanone	QQ. 1,1-Dichloropropene	III. n-Butylbenzene	AAAA. Ethyl tert-butyl ether
H. 1,1-Dichloroethene**	Z. 2-Hexanone	RR. Dibromomethane	JJJ. 1,2-Dichlorobenzene	BBBB. tert-Amyl methyl ether
I. 1,1-Dichloroethane*	AA. Tetrachloroethene	SS. 1,3-Dichloropropane	KKK. 1,2,4-Trichlorobenzene	CCCC. 1-Chlorohexane
J. 1,2-Dichloroethene, total	BB. 1,1,2,2-Tetrachloroethane*	TT. 1,2-Dibromoethane	LLL. Hexachlorobutadiene	DDDD. Isopropyl alcohol
K. Chloroform**	CC. Toluene**	UU. 1,1,1,2-Tetrachloroethane	MMM. Naphthalene	EEEE. Acetonitrile
L. 1,2-Dichloroethane	DD. Chlorobenzene*	VV. Isopropylbenzene	NNN. 1,2,3-Trichlorobenzene	FFFF. Acrolein
M. 2-Butanone	EE. Ethylbenzene**	WW. Bromobenzene	OOO. 1,3,5-Trichlorobenzene	GGGG. Acrylonitrile
N. 1,1,1-Trichloroethane	FF. Styrene	XX. 1,2,3-Trichloropropane	PPP. trans-1,2-Dichloroethene	HHHH. 1,4-Dioxane
O. Carbon tetrachloride	GG. Xylenes, total	YY. n-Propylbenzene	QQQ. cis-1,2-Dichloroethene	IIII. Isobutyl alcohol
P. Bromodichloromethane	HH. Vinyl acetate	ZZ. 2-Chlorotoluene	RRR. m,p-Xylenes	JJJJ. Methacrylonitrile
Q. 1,2-Dichloropropane**	II. 2-Chloroethylvinyl ether	AAA. 1,3,5-Trimethylbenzene	SSS. o-Xylene	KKKK. Propionitrile
R. cis-1,3-Dichloropropene	JJ. Dichlorodifluoromethane	BBB. 4-Chlorotoluene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane	LLLL

\* = System performance check compounds (SPCC) for RRF ; \*\* = Calibration check compounds (CCC) for %RSD.

## VALIDATION FINDINGS WORKSHEET

### Initial Calibration

Page: 1 of 1  
 Reviewer: JVZ  
 2nd Reviewer: \_\_\_\_\_

LDC #: 18057 C 48  
 SDG #: See cover

METHOD: GC/MS VOA (EPA TO-14/TO-14A)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Did the laboratory perform a 5 point calibration prior to sample analysis?

Y N N/A Were all percent relative standard deviations (%RSD)  $\leq$  30% and relative response factors (RRF)  $\geq$  0.05?

#	Date	Standard ID	Compound	Finding %RSD (Limit: $\leq$ 30.0%)	Finding RRF (Limit: $\geq$ 0.05)	Associated Samples	Qualifications
	<u>11/27/07</u>	<u>1CAL</u>	<u>VVV</u>	<u>32.177</u>		<u>All + BIK</u>	<u>J/MS/P</u>

LDC #: 18057 C48  
 SDG #: See cover  
**VALIDATION FINDINGS WORKSHEET**  
**Blanks**

Page: 1 of 1  
 Reviewer: JVZ  
 2nd Reviewer: \_\_\_\_\_

METHOD: GC/MS VOA (EPA TO-14/TO-14A)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".  
 Y N N/A  
 Was a method blank associated with every sample in this SDG? Y

Y N N/A  
 Was a method blank analyzed at least once every 12 hours for each matrix and concentration? Y

Y N N/A  
 Was there contamination in the method blanks? If yes, please see the qualifications below. Y

Blank analysis date: 12/04/07

Conc. units: µg/L (Y/N)

Associated Samples: A/1

Compound	Blank ID	Sample Identification												
		1	2	3	4	5	6	7	8					
	7339519MB													
E	1.2	1.5/2.04	1.6/2.04	2.5/U	2.5/U	8.2/U	7.7/U	7.8/U	10/U					
CRQL														

Blank analysis date: \_\_\_\_\_  
 Conc. units: \_\_\_\_\_

Compound	Blank ID	Sample Identification												
		1	2	3	4	5	6	7	8					
CRQL														

Associated Samples: \_\_\_\_\_

CIRCLED RESULTS WERE NOT QUALIFIED. ALL RESULTS NOT CIRCLED WERE QUALIFIED BY THE FOLLOWING STATEMENT:

Note: Common contaminants such as Methylene chloride, Acetone, 2-Butanone, Carbon disulfide and TICs that were detected in samples within ten times the associated method blank concentration were qualified as not detected, "U". Other contaminants within five times the method blank concentration were also qualified as not detected, "U".



LDC#: 18054C48  
 SDG#: See Cover

**VALIDATION FINDINGS WORKSHEET**  
**Field Duplicates**

Page:      of       
 Reviewer:             
 2nd Reviewer:           

METHOD: GCMS VOA (EPA TO-14A)

Y N NA Were field duplicate pairs identified in this SDG?  
Y N NA Were target analytes detected in the field duplicate pairs?

Compound	Concentration (ppb(v/v))		( $\leq 50$ )	Difference	Diff Limits	Qualifications (Parent Only)
	6	7	RPD			
Chloromethane	1.4	1.6		0.2	( $\leq 4.0$ )	
Vinyl Chloride	4.0	4.4		0.4	( $\leq 3.0$ )	
Chloroethane	35	39	11			
Acetone	32	41		9	( $\leq 10.0$ )	
Methylene chloride	7.7	7.8		0.1	( $\leq 2.0$ )	
1,1-Dichloroethane	56	64	13			
Benzene	1.7	2.1		0.4	( $\leq 3.0$ )	
1,2-Dichloroethane	11	13	17			
Trichloroethene	3.2	4.0	<del>22</del>	0.8	( $\leq 2.0$ )	
Bromodichloromethane	5.2	6.4		1.2	↓	
Toluene	1.7	2.0		0.3		
1,1,2-Trichloroethane	3.4	3.9		0.5		
Tetrachloroethene	210	230	9			
Chlorobenzene	2.0U	1.2		0.8	( $\leq 2.0$ )	
1,3-Dichlorobenzene	1.4	1.6		0.2	↓	
<sup>1000</sup> Chloroform	3100	3700	18			

V:\FIELD DUPLICATES\18054C48.wpd

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** BRC Tronox Parcel C/D/F/G  
**Collection Date:** November 28, 2007  
**LDC Report Date:** January 21, 2008  
**Matrix:** Air  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** TestAmerica, Inc.  
**Sample Delivery Group (SDG):** F7K290369

**Sample Identification**

TSB-CR-04  
TSB-CR-05  
TSB-CJ-06  
TSB-DR-01  
TSB-DR-01(FD)

## Introduction

This data review covers 5 air samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method TO-14A for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds with the following exceptions:

Date	Compound	%RSD	Associated Samples	Flag	A or P
11/27/07	4-Ethyltoluene	32.177	All samples in SDG F7K290369	J (all detects) UJ (all non-detects)	P

Average relative response factors (RRF) for all volatile target compounds and system monitoring compounds were within validation criteria.

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
12/5/07	1,2,4-Trichlorobenzene	39.7	TSB-DR-01 (FD) 7340411MB	J- (all detects) UJ (all non-detects)	P

All of the continuing calibration RRF values were within validation criteria.

## V. Blanks

Method blank analyses were performed at the required frequency. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
7339521-MB	12/4/07	Methylene chloride	1.2 ppb(v/v)	TSB-CR-04 TSB-CR-05 TSB-CJ-06 TSB-DR-01
7340411-MB	12/5/07	Methylene chloride	2.7 ppb(v/v)	TSB-DR-01 (FD)

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
TSB-CR-04	Methylene chloride	3.2 ppb(v/v)	3.2U ppb(v/v)
TSB-CR-05	Methylene chloride	2.8 ppb(v/v)	2.8U ppb(v/v)
TSB-CJ-06	Methylene chloride	5.2 ppb(v/v)	5.2U ppb(v/v)
TSB-DR-01	Methylene chloride	8.9 ppb(v/v)	8.9U ppb(v/v)
TSB-DR-01 (FD)	Methylene chloride	11 ppb(v/v)	11U ppb(v/v)

No field blanks were identified in this SDG.

## VI. Surrogate Spikes

Surrogates were not required by the method.

## VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

## VIII. Laboratory Control Samples (LCS)

Although laboratory control samples were not required by the method, laboratory control samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## IX. Regional Quality Assurance and Quality Control

Not applicable.

## X. Internal Standards

All internal standard areas and retention times were within QC limits.

## XI. Target Compound Identifications

All target compound identifications were within validation criteria.

## XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

## XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

## XIV. System Performance

The system performance was acceptable.

## XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

## XVI. Field Duplicates

Samples TSB-DR-01 and TSB-DR-01(FD) were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ppb(v/v))		RPD (Limits)	Difference (Limits)	Flag	A or P
	TSB-DR-01	TSB-DR-01(FD)				
Chloromethane	4.0U	1.6	-	2.4 ( $\leq 4.0$ )	-	-
Bromomethane	4.0U	2.5	-	1.5 ( $\leq 4.0$ )	-	-
Chloroethane	20	27	30 ( $\leq 50$ )	-	-	-
Acetone	34	36	-	2 ( $\leq 10.0$ )	-	-
Methylene chloride	8.9	11	-	2.1 ( $\leq 2.0$ )	J (all detects)	A

Compound	Concentration (ppb(v/v))		RPD (Limits)	Difference (Limits)	Flag	A or P
	TSB-DR-01	TSB-DR-01(FD)				
1,1-Dichloroethane	1.6	1.9	-	0.3 ( $\leq 2.0$ )	-	-
Chloroform	33	40	19 ( $\leq 50$ )	-	-	-
Benzene	3.0U	1.6	-	1.4 ( $\leq 3.0$ )	-	-
Trichloroethene	1.1	1.3	-	0.2 ( $\leq 2.0$ )	-	-
Toluene	1	2.0U	-	1 ( $\leq 2.0$ )	-	-
Tetrachloroethene	13	15	14 ( $\leq 50$ )	-	-	-

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Data Qualification Summary - SDG F7K290369**

SDG	Sample	Compound	Flag	A or P	Reason
F7K290369	TSB-CR-04 TSB-CR-05 TSB-CJ-06 TSB-DR-01 TSB-DR-01 (FD)	4-Ethyltoluene	J (all detects) UJ (all non-detects)	P	Initial calibration (%RSD)
F7K290369	TSB-DR-01 (FD)	1,2,4-Trichlorobenzene	J- (all detects) UJ (all non-detects)	P	Continuing calibration (%D)
F7K290369	TSB-DR-01 TSB-DR-01 (FD)	Methylene chloride	J (all detects)	A	Field duplicates (Difference)

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Laboratory Blank Data Qualification Summary - SDG F7K290369**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
F7K290369	TSB-CR-04	Methylene chloride	3.2U ppb(v/v)	A
F7K290369	TSB-CR-05	Methylene chloride	2.8U ppb(v/v)	A
F7K290369	TSB-CJ-06	Methylene chloride	5.2U ppb(v/v)	A
F7K290369	TSB-DR-01	Methylene chloride	8.9U ppb(v/v)	A
F7K290369	TSB-DR-01 (FD)	Methylene chloride	11U ppb(v/v)	A

**BRC Tronox Parcel C/D/F/G**  
**Volatiles - Field Blank Data Qualification Summary - SDG F7K290369**

No Sample Data Qualified in this SDG



LDC #: 18054D48  
 SDG #: F7K290369  
 Laboratory: Test America

**VALIDATION COMPLETENESS WORKSHEET**  
 Level IV

Date: 1/14/08  
 Page: 1 of 1  
 Reviewer: JVL  
 2nd Reviewer: [Signature]

**METHOD:** GC/MS Volatiles (EPA Method TO-14A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11/28/07
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	SW	
IV.	Continuing calibration/ICV	SW	
V.	Blanks	SW	
VI.	Surrogate spikes	N	
VII.	Matrix spike/Matrix spike duplicates	N	
VIII.	Laboratory control samples	A	UCS 10
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	
XII.	Compound quantitation/CRQLs	A	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	A	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 4.5
XVII.	Field blanks	N	

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet  
 ND = No compounds detected  
 R = Rinsate  
 FB = Field blank  
 D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples:

Air

1	TSB-CR-04	11	7339521 MB	21	31
2	TSB-CR-05	12	7340411 MB	22	32
3	TSB-CJ-06	13		23	33
4	TSB-DR-01	14		24	34
5	TSB-DR-01(FD)	15		25	35
6		16		26	36
7		17		27	37
8		18		28	38
9		19		29	39
10		20		30	40

VALIDATION FINDINGS CHECKLIST

IDC #: 18054 D48  
 SDG #: See label

Method: Volatiles (EPA Method TO-14/TO-14A)

Validation Area	Yes	No	NA	Findings/Comments
<b>I. Technical holding times</b>				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Canister pressure criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>II. GC/MS Instrument performance check</b>				
Were the BFB performance results reviewed and found to be within the specified criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all samples analyzed within the 12 hour clock criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>III. Initial calibration</b>				
Did the laboratory perform a 5 point calibration prior to sample analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) $\leq$ 30% and relative response factors (RRF) $\geq$ 0.05?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>IV. Continuing calibration</b>				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) $\leq$ 30% and relative response factors (RRF) $\geq$ 0.05?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>V. Blanks</b>				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a method blank analyzed at least once every 12 hours for each matrix and concentration?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>VI. Surrogate spikes</b>				
Were all surrogate %R within QC limits?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
<b>VII. Matrix spike/Matrix spike duplicates</b>				
Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>VIII. Laboratory control samples</b>				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was an LCS analyzed per analytical batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

# VALIDATION FINDINGS CHECKLIST

Page: 7 of 2  
 Reviewer: JV  
 2nd Reviewer: [Signature]

LDC #: 18054 D48  
 SDG #: See Cover

Validation Area	Yes	No	NA	Findings/Comments
<b>IX. Regional Quality Assurance and Quality Control</b>				
Were performance evaluation (PE) samples performed?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the performance evaluation (PE) samples within the acceptance limits?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>X. Internal standards</b>				
Were internal standard area counts within +/-40% from the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were retention times within +/- 30.0 seconds from the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XI. Target compound identification</b>				
Were relative retention times (RRT's) within $\pm 0.06$ RRT units of the standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were chromatogram peaks verified and accounted for?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XII. Compound quantitation/CRQLs</b>				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XIII. Tentatively identified compounds (TICs)</b>				
Were the major ions (> 10 percent relative intensity) in the reference spectrum evaluated in sample spectrum?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were relative intensities of the major ions within $\pm 20\%$ between the sample and the reference spectra?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>XIV. System performance</b>				
System performance was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XV. Overall assessment of data</b>				
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XVI. Field duplicates</b>				
Field duplicate pairs were identified in this SDG.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field duplicates.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>XVII. Field blanks</b>				
Field blanks were identified in this SDG.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field blanks.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

# TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method TO-14/TO-14A)

A. Chloromethane*	S. Trichloroethene	KK. Trichlorofluoromethane	CCC. tert-Butylbenzene	UUU. Benzyl chloride
B. Bromomethane	T. Dibromochloromethane	LL. Methyl-tert-butyl ether	DDD. 1,2,4-Trimethylbenzene	VVV. 4-Ethyltoluene
C. Vinyl chloride**	U. 1,1,2-Trichloroethane	MM. 1,2-Dibromo-3-chloropropane	EEE. sec-Butylbenzene	WWW. Ethanol
D. Chloroethane	V. Benzene	NN. Diethyl ether	FFF. 1,3-Dichlorobenzene	XXX. Ethyl ether
E. Methylene chloride	W. trans-1,3-Dichloropropene	OO. 2,2-Dichloropropane	GGG. p-Isopropyltoluene	YYY. tert-Butanol
F. Acetone	X. Bromoform*	PP. Bromochloromethane	HHH. 1,4-Dichlorobenzene	ZZZ. tert-Butyl alcohol
G. Carbon disulfide	Y. 4-Methyl-2-pentanone	QQ. 1,1-Dichloropropene	III. n-Butylbenzene	AAA. Ethyl tert-butyl ether
H. 1,1-Dichloroethene**	Z. 2-Hexanone	RR. Dibromomethane	JJJ. 1,2-Dichlorobenzene	BBB. tert-Amyl methyl ether
I. 1,1-Dichloroethane*	AA. Tetrachloroethene	SS. 1,3-Dichloropropane	KKK. 1,2,4-Trichlorobenzene	CCCC. 1-Chlorohexane
J. 1,2-Dichloroethene, total	BB. 1,1,2,2-Tetrachloroethane*	TT. 1,2-Dibromoethane	LLL. Hexachlorobutadiene	DDDD. Isopropyl alcohol
K. Chloroform**	CC. Toluene**	UU. 1,1,1,2-Tetrachloroethane	MMM. Naphthalene	EEEE. Acetonitrile
L. 1,2-Dichloroethane	DD. Chlorobenzene*	VV. Isopropylbenzene	NNN. 1,2,3-Trichlorobenzene	FFFF. Acrolein
M. 2-Butanone	EE. Ethylbenzene**	WW. Bromobenzene	OOO. 1,3,5-Trichlorobenzene	GGGG. Acrylonitrile
N. 1,1,1-Trichloroethane	FF. Styrene	XX. 1,2,3-Trichloropropane	PPP. trans-1,2-Dichloroethene	HHHH. 1,4-Dioxane
O. Carbon tetrachloride	GG. Xylenes, total	YY. n-Propylbenzene	QQQ. cis-1,2-Dichloroethene	IIII. Isobutyl alcohol
P. Bromodichloromethane	HH. Vinyl acetate	ZZ. 2-Chlorotoluene	RRR. m,p-Xylenes	JJJJ. Methacrylonitrile
Q. 1,2-Dichloropropane**	II. 2-Chloroethylvinyl ether	AAA. 1,3,5-Trimethylbenzene	SSS. o-Xylene	KKKK. Propionitrile
R. cis-1,3-Dichloropropene	JJ. Dichlorodifluoromethane	BBB. 4-Chlorotoluene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane	LLLL

\* = System performance check compounds (SPCC) for RRF ; \*\* = Calibration check compounds (CCC) for %RSD.







LDC#: 18054048  
 SDG#: See Cover

**VALIDATION FINDINGS WORKSHEET**  
Field Duplicates

Page: 1 of 1  
 Reviewer: JVB  
 2nd Reviewer: J

**METHOD:** GCMS VOA (EPA TO-14A)

Y N NA Were field duplicate pairs identified in this SDG?  
Y N NA Were target analytes detected in the field duplicate pairs?

Compound	Concentration ( ppb(v/v) )		(<50) RPD	Difference	Diff Limits	Qualifications (Parent Only)
	4	5				
Chloromethane	4.0U	1.6		2.4	$\leq 4.0$	
Bromomethane	4.0U	2.5		1.5	↓	
Chloroethane	20	27	30			
Acetone	34	36		2	$\leq 10.0$	
Methylene chloride	8.9	11		2.1	$\leq 2.0$	J detx / A
1,1-Dichloroethane	1.6	1.9		0.3	↓	
Chloroform	33	40	19			
Benzene	3.0U	1.6		1.4	$\leq 3.0$	
Trichloroethene	1.1	1.3		0.2	$\leq 2.0$	
Toluene	1.0	2.0U		1	↓	
Tetrachloroethene	13	15	14			

V:\FIELD DUPLICATES\18054D48.wpd



**VALIDATION FINDINGS WORKSHEET**  
**Initial Calibration Calculation Verification**

LDC #: 18 054 048  
 SDG #: 1 Sep 2007

Page: 1 of 1  
 Reviewer: JMG  
 2nd Reviewer: R

**METHOD:** GC/MS VOA (EPA TO-14/TO-14A)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_b)/(A_b)(C_x)$   
 average RRF = sum of the RRFs/number of standards  
 $\%RSD = 100 * (S/X)$

$A_x$  = Area of compound,  
 $C_x$  = Concentration of compound,  
 $S$  = Standard deviation of the RRFs  
 $X$  = Mean of the RRFs

$A_b$  = Area of associated internal standard  
 $C_b$  = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (SD std)	RRF (SD std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD		
1	ICAL	11/27/07	Methylene chloride (1st internal standard)	0.4468	0.4468	0.44707	0.44705	16.489	16.687		
			Trichlorethene (2nd internal standard)	0.49229	0.49229	0.49029	0.49029	13.011	13.011		
			Toluene (3rd internal standard)	1.48856	1.48856	1.25272	1.25272	10.157	10.157		
2			Methylene chloride (1st internal standard)								
			Trichlorethene (2nd internal standard)								
			Toluene (3rd internal standard)								
3			Methylene chloride (1st internal standard)								
			Trichlorethene (2nd internal standard)								
			Toluene (3rd internal standard)								
4			Methylene chloride (1st internal standard)								
			Trichlorethene (2nd internal standard)								
			Toluene (3rd internal standard)								

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

**VALIDATION FINDINGS WORKSHEET**  
**Continuing Calibration Results Verification**

LDC #: 18054 D48  
 SDG #: See Cover

**METHOD: GC/MS VOA (EPA TO-14/TO14A)**

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference =  $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$   
 RRF =  $(A_x)(C_s) / (A_s)(C_x)$   
 Where: ave. RRF = initial calibration average RRF  
 RRF = continuing calibration RRF  
 A<sub>x</sub> = Area of compound, A<sub>s</sub> = Area of associated internal standard  
 C<sub>x</sub> = Concentration of compound, C<sub>s</sub> = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Average RRF (Initial)	Reported		Recalculated	
					RRF (CC)	%D	RRF (CC)	%D
1	CC 12041	12/04/07	Methylene chloride (1st internal standard)	0.44704	0.49259	10.2	0.49259	10.2
			Trichlorethene (2nd internal standard)	0.49029	0.48657	0.8	0.48657	0.8
			Toluene (3rd internal standard)	1.35272	1.48933	10.1	1.48933	10.1
2	CC 12053	12/05/07	Methylene chloride (1st internal standard)	0.44704	0.45459	1.7	0.45459	1.7
			Trichlorethene (2nd internal standard)	0.49029	0.46351	5.5	0.46351	5.5
			Toluene (3rd internal standard)	1.35272	1.43247	5.9	1.43247	5.9
3			Methylene chloride (1st internal standard)					
			Trichlorethene (2nd internal standard)					
			Toluene (3rd internal standard)					
4			Methylene chloride (1st internal standard)					
			Trichlorethene (2nd internal standard)					
			Toluene (3rd internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 18054 D48  
SDG #: See Cover

### VALIDATION FINDINGS WORKSHEET

#### Sample Calculation Verification

Page: 1 of 1  
Reviewer: SVZ  
2nd reviewer: J

METHOD: GC/MS VOA (EPA Method TO-15)

- Y N N/A     Were all reported results recalculated and verified for all level IV samples?  
Y N N/A     Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

Concentration = 
$$\frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_o)(\%S)}$$

- $A_x$  = Area of the characteristic ion (EICP) for the compound to be measured
- $A_s$  = Area of the characteristic ion (EICP) for the specific internal standard
- $I_s$  = Amount of internal standard added in nanograms (ng)
- RRF = Relative response factor of the calibration standard.
- $V_o$  = Volume or weight of sample pruged in milliliters (ml) or grams (g).
- Df = Dilution factor.
- %S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. 1 , AA:

Conc. = 
$$\frac{(50990)(60)(23.90)(500)}{(1036232)(0.91807)(12.200)(940)}$$
  
= 2.6786  
≈ 2.7 ppb (v/v)

#	Sample ID	Compound	Reported Concentration ( )	Calculated Concentration ( )	Qualification