

TABLE OF CONTENTS

COVER PAGE.....	1
QC CHECKLIST.....	2
SUMMARY SHEET.....	3
RUNLOG.....	5
INITIAL CALIBRATION.....	7
PERIODIC QC.....	22
QC: (MBLK, MRL, LCS1, LCS2).....	33
SAMPLE (2711190059).....	38
QC: (MS/MSD 2711190059).....	39
PERIODIC QC.....	41
SAMPLE (2711200235).....	44
QC: (MS/MSD 2711200235).....	45
PERIODIC QC.....	47
SAMPLE (2711200572).....	49
CLOSING QC.....	50
STANDARDS PREPARATION WORKSHEET AND CERTIFICATES OF ANALYSIS.....	53

Level IV Data Package

MWH Group 223000

Method: EPA 200.7

2711200572

EPA 200.7/6010B QC Check List

Analyst W31 Analysis Date 12/3/07 Reviewer/Date WME/254A

Instrument PerKin Elmer Optima 4300DV

- All sample analyzed within 6 month holding time
- All sample raw concentration below the high standard or linear range o marked for dilution and rerun

Initial and closing QC

- ICV within +/- 5%
- Linearity check +/- 10%
- ICSAB +/- 20%
- QCS +/- 5%
- MRL +/- 50%

Middle, closing and batch QC

- FilterCheck < 1/2 MRL
- MBLANK < 1/2 MRL
- LCS +/-15%
- MS/MSD +/-30% (200.7) +/- 25% (6010B)
- CCV/MCV/ECV +/- 10%
- ICB/CCB/ECB < 1/2 MRL
- CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/-20%
- RPD between LCS/LCSD is within +/- 20%
- Internal standards +/-20%
- All pH of the samples are < 2

- No more than 20 samples per batch
- MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

- NA QIR needed for failed QC
- Special Det Code noted on the cover sheet
- R value for multi point calibration is > 0.995
- NA Proper MRL check ran for special low MRL samples

Reagent and Standards used for
Optima 4300 DV
Updated 12/01/06

Int: W31
Date: 12/3/07

Method 200.7/6010

ICP SUMMARY SHEET

File ID: 071203a
Date Started: 12/3/07
Analyst ID: wbh

SAMPLE ID

QCS1	(21:50)	QCS2	(21:54)	QCS3	(21:57)
LINEARITY	(22:01)	N711140386_2	(22:51)	N711140388_2	(23:03)
N711140446_2	(23:08)	N711140450_2	(23:20)	N711160270_2	(23:24)
2711190059	(23:48)	2711200235_2	(12:12)	2711200228_2	(12:25)
2711200229_2	(12:29)	2711200230_2	(12:33)	2711200231_2	(12:37)
2711200232_2	(12:42)	2711200233_2	(12:46)	2711200234_2	(12:50)
2711200236_2	(1:01)	2711200237_2	(1:05)	2711200238_2	(1:10)
2711200239_2	(1:14)	2711200240_2	(1:18)	2711080235_2	(1:22)
2711200572_2	(1:27)	2711270096_2	(1:32)	2711270097_2	(1:37)
2711270098_2	(1:41)				

COMMENT:

B#1 167 Reported BAP LC/LCID
(No. 523)
12/4/07

Analyst: wbh

Approved By: MW 12/4/07

BATCH NUMBER for 071203a

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CO CR CU FE K MG MN MO NA NI PB

Batch ID: N711140386_2X

Batch ID: 2711190059

2711190059	2711200235_2X	2711200228_2X
2711200229_2X	2711200230_2X	2711200231_2X
2711200232_2X	2711200233_2X	2711200234_2X
2711200236_2X	2711200237_2X	2711200238_2X
2711200239_2X	2711200240_2X	2711080235_2X
2711200572_2X	2711270096_2X	2711270097_2X
2711270098_2X		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	12/3/07	21:46	1	10.271	10.3	95-105	102%
QCS1	12/3/07	21:50	1	10.532	11		
QCS2	12/3/07	21:54	1	10.396	10		
QCS3	12/3/07	21:57	1	10.373	10		
LINEARITY	12/3/07	22:01	1	0.0019	.0019		
ICSA	12/3/07	22:05	1	-0.0004	ND	80-120	
ICSAB	12/3/07	22:14	1	.24392	.244	80-120	97.5%
CCV	12/3/07	22:18	1	5.0494	5.05	90-110	100%
ICB	12/3/07	22:22	1	-0.0001	ND		
MRL	12/3/07	22:26	1	0.0109	.0109	50-150	108%
MBLANK2007	12/3/07	22:30	1	0.0032	.0032		
MRL2007	12/3/07	22:34	1	-0.0003	ND		
MRL	12/3/07	22:39	1	0.0104	.0104	50-150	103%
LCS	12/3/07	22:43	1	0.0008	0.0008	85-115	8.45% Q
LCSD	12/3/07	22:47	1	-0.0005	ND	85-115	-4.55% Q
N711140386_2X	12/3/07	22:51	2	0.0199	.020		
N711140386_2XMS	12/3/07	22:55	2	1.9862	1.99	[1.966]	98.3%
N711140386_2XMSD	12/3/07	22:59	2	2.0101	2.01	[1.990]	99.5%
N711140386_2XT	12/3/07	22:59	2		2.00	70 - 130	
N711140388_2X	12/3/07	23:03	2	0.0019	.0019		
N711140446_2X	12/3/07	23:08	2	2.0711	2.1		
CCV	12/3/07	23:12	1	5.0063	5.01	90-110	100%
CCB	12/3/07	23:16	1	0.0001	0.0000		
N711140450_2X	12/3/07	23:20	2	2.0472	2.0		
N711160270_2X	12/3/07	23:24	2	0.0124	.012		
MBLANK2007	12/3/07	23:29	1	-0.0001	ND		
MRL	12/3/07	23:33	1	0.0103	.0103	50-150	102%
MRL2007	12/3/07	23:37	1	0.0097	.0097		
LCS	12/3/07	23:41	1	.97201	.972	85-115	97.2%
LCSD	12/3/07	23:45	1	.98516	.985	85-115	98.5%
2711190059	12/3/07	23:48	1	0.0004	0.0003		
2711190059MS	12/3/07	23:53	1	1.0508	1.05	[1.051]	105%
2711190059MSD	12/3/07	23:56	1	.98298	.983	[0.983]	98.2%
2711190059T	12/3/07	23:56	1		1.00	70 - 130	
CCV	12/4/07	12:00	1	5.0138	5.01	90-110	100%
CCB	12/4/07	12:04	1	-0.0002	ND		
MCV	12/4/07	12:08	1	2.4998	2.5	90-110	99.9%
2711200235_2X	12/4/07	12:12	2	-0.0001	ND		
2711200235_2XMS	12/4/07	12:16	2	1.0089	1.01	[1.009]	50.4%
2711200235_2XMSD	12/4/07	12:21	2	1.0050	1.01	[1.005]	50.2%
2711200235_2XT	12/4/07	12:21	2		2.00	70 - 130	
2711200228_2X	12/4/07	12:25	2	0.0003	0.0003		
2711200229_2X	12/4/07	12:29	2	0.0002	0.0002		
2711200230_2X	12/4/07	12:33	2	0.0022	.0022		
2711200231_2X	12/4/07	12:37	2	0.0006	0.0006		
2711200232_2X	12/4/07	12:42	2	0.0003	0.0002		
2711200233_2X	12/4/07	12:46	2	0.0001	0.0000		

101
 101
 9/24/07

File ID: 071203a

CR

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2711200234_2X	12/4/07	12:50	2	-0.0002	ND		
CCV	12/4/07	12:54	1	5.0870	5.09	90-110	101%
CCB	12/4/07	12:57	1	-0.0001	ND		
2711200236_2X	12/4/07	1:01	2	0.0006	0.0006		
2711200237_2X	12/4/07	1:05	2	0.0031	.0031		
2711200238_2X	12/4/07	1:10	2	0.0005	0.0005		
2711200239_2X	12/4/07	1:14	2	0.0005	0.0004		
2711200240_2X	12/4/07	1:18	2	0.0001	0.0001		
2711080235_2X	12/4/07	1:22	2	.89491	.890		
2711200572_2X	12/4/07	1:27	2	0.0058	.0058		
2711270096_2X	12/4/07	1:32	2	0.0128	.013		
2711270097_2X	12/4/07	1:37	2	0.0046	.0046		
2711270098_2X	12/4/07	1:41	2	0.0056	.0056		
ECV	12/4/07	1:46	1	5.0953	5.1	90-110	101%
ECB	12/4/07	1:49	1	0.0002	0.0002		

=====
Analysis Begun

Start Time: 12/3/2007 8:44:38 PM Plasma On Time: 12/3/2007 6:19:37 PM
Logged In Analyst: Administrator Technique: ICP Continuous
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Owner\Sample Information\071203.sif
Batch ID: 071203
Results Data Set: 071203A
Results Library: C:\pe\Owner\Results\Results.mdb

=====
Method Loaded

Method Name: 200.7&6010_071003 Method Last Saved: 12/3/2007 12:53:14 PM
IEC File: 070703.iec MSF File:
Method Description: 200.7/6010_070703

=====
Sequence No.: 1 Autosampler Location: 0
Sample ID: Calib Blank 1 Date Collected: 12/3/2007 8:44:38 PM
Analyst: Data Type: Original
Initial Sample Wt: Initial Sample Vol:
Dilution: Sample Prep Vol:

=====
Nebulizer Parameters: Calib Blank 1

Analyte Back Pressure Flow
All 246.0 kPa 0.65 L/min

=====
Mean Data: Calib Blank 1

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	455714.7	1390.36	0.31%	100 %
Yr	247525.5	818.64	0.33%	100 %
Ag†	335.4	49.01	14.61%	[0.00] mg/L
Al†	34.7	9.57	27.60%	[0.00] mg/L
As†	6.0	1.11	18.36%	[0.00] mg/L
B_†	166.4	3.57	2.15%	[0.00] mg/L
Ba†	-32.8	1.49	4.55%	[0.00] mg/L
Be†	-4091.0	23.55	0.58%	[0.00] mg/L
Ca†	671.7	2.94	0.44%	[0.00] mg/L
Co†	-64.1	0.14	0.22%	[0.00] mg/L
Cr†	313.5	1.16	0.37%	[0.00] mg/L
Cu†	2082.8	73.27	3.52%	[0.00] mg/L
Fe†	-7.7	0.40	5.18%	[0.00] mg/L
K†	207.0	85.95	41.53%	[0.00] mg/L
Mg†	-46.8	2.79	5.96%	[0.00] mg/L
Mn†	97.1	2.93	3.02%	[0.00] mg/L
Mo†	24.8	3.45	13.90%	[0.00] mg/L
Na†	-374.7	69.62	18.58%	[0.00] mg/L
Ni†	-62.2	5.30	8.52%	[0.00] mg/L
Pb†	-13.6	13.12	96.21%	[0.00] mg/L
Sb†	3.4	4.37	127.86%	[0.00] mg/L
Se†	-2.8	2.68	97.18%	[0.00] mg/L
Tl†	-37.1	3.66	9.88%	[0.00] mg/L
V†	189.1	21.62	11.43%	[0.00] mg/L
Zn†	353.2	12.27	3.47%	[0.00] mg/L

Sequence No.: 2
Sample ID: Standard 2
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 15
Date Collected: 12/3/2007 8:48:45 PM
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
All 246.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc. Units	Calib
Sca	408467.5	2595.98	0.64%	89.6	%
Yr	233206.0	1851.70	0.79%	94.2	%
Ag†	579581.1	343.39	0.06%	[2]	mg/L
Al†	46531.5	551.21	1.18%	[10]	mg/L
As†	22004.9	203.28	0.92%	[10]	mg/L
B_†	163007.2	647.65	0.40%	[5.02]	mg/L
Ba†	749232.4	1292.05	0.17%	[10]	mg/L
Be†	12462592.3	41635.36	0.33%	[4.01]	mg/L
Ca†	727311.1	504.59	0.07%	[100]	mg/L
Co†	270354.4	427.90	0.16%	[10]	mg/L
Cr†	813513.3	211.28	0.03%	[9.97]	mg/L
Cu†	3945164.6	3939.12	0.10%	[10]	mg/L
Fe†	10860.6	73.38	0.68%	[9.98]	mg/L
K†	122965.9	1914.40	1.56%	[100]	mg/L
Mg†	308003.0	1125.56	0.37%	[100]	mg/L
Mn†	5808277.4	11463.27	0.20%	[10]	mg/L
Mo†	133236.4	94.27	0.07%	[9.98]	mg/L
Na†	361934.4	4196.63	1.16%	[100]	mg/L
Ni†	231659.0	248.81	0.11%	[10]	mg/L
Pb†	50172.0	321.98	0.64%	[10]	mg/L
Sb†	20865.8	138.17	0.66%	[10]	mg/L
Se†	13806.4	90.09	0.65%	[10]	mg/L
Tl†	28658.9	258.01	0.90%	[10]	mg/L
V†	1775472.3	91.66	0.01%	[10]	mg/L
Zn†	502238.7	25.28	0.01%	[10]	mg/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	289800	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	4653	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	2200	0.00000	1.000000	
B_	1	Lin, Calc Int	0.0	32470	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	74920	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	3108000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	7273	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	27040	0.00000	1.000000	
Cr	1	Lin, Calc Int	0.0	81600	0.00000	1.000000	
Cu	1	Lin, Calc Int	-0.0	394500	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	1088	0.00000	1.000000	
K	1	Lin, Calc Int	-0.0	1230	0.00000	1.000000	
Mg	1	Lin, Calc Int	-0.0	3080	0.00000	1.000000	
Mn	1	Lin, Calc Int	0.0	580800	0.00000	1.000000	
Mo	1	Lin, Calc Int	0.0	13350	0.00000	1.000000	
Na	1	Lin, Calc Int	0.0	3619	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	23170	0.00000	1.000000	
Pb	1	Lin, Calc Int	0.0	5017	0.00000	1.000000	
Sb	1	Lin, Calc Int	0.0	2087	0.00000	1.000000	
Se	1	Lin, Calc Int	0.0	1381	0.00000	1.000000	
Tl	1	Lin, Calc Int	0.0	2866	0.00000	1.000000	
V	1	Lin, Calc Int	0.0	177500	0.00000	1.000000	
Zn	1	Lin, Calc Int	0.0	50220	0.00000	1.000000	

Sequence No.: 3
Sample ID: ICV
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 15
Date Collected: 12/3/2007 8:52:30 PM
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
All 246.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	871870.8	191 %	7.3			3.80%
Yr	227595.9	91.9 %	6.21			6.75%
Ag†	4698.6	0.0162 mg/L	0.02287	0.0162 mg/L	0.02287	141.06%
	QC value less than the lower limit for Ag Recovery = 0.81%					
Al†	43241.0	9.29 mg/L	0.897	9.29 mg/L	0.897	9.65%
	QC value less than the lower limit for Al Recovery = 92.93%					
As†	30.7	0.0195 mg/L	0.01727	0.0195 mg/L	0.01727	88.48%
	QC value less than the lower limit for As Recovery = 0.20%					
B_†	2159.6	0.0663 mg/L	0.05079	0.0663 mg/L	0.05079	76.58%
	QC value less than the lower limit for B_ Recovery = 1.33%					
Ba†	6914.0	0.0923 mg/L	0.12691	0.0923 mg/L	0.12691	137.53%
	QC value less than the lower limit for Ba Recovery = 0.92%					
Be†	1507703.4	0.485 mg/L	0.6810	0.485 mg/L	0.6810	140.37%
	QC value less than the lower limit for Be Recovery = 12.13%					
Ca†	700209.4	96.3 mg/L	8.54	96.3 mg/L	8.54	8.88%
	QC value within limits for Ca Recovery = 96.27%					
Co†	2453.6	0.0908 mg/L	0.12366	0.0908 mg/L	0.12366	136.25%
	QC value less than the lower limit for Co Recovery = 0.91%					
Cr†	7344.2	0.0900 mg/L	0.12769	0.0900 mg/L	0.12769	141.86%
	QC value less than the lower limit for Cr Recovery = 0.90%					
Cu†	33248.3	0.0844 mg/L	0.12111	0.0844 mg/L	0.12111	143.56%
	QC value less than the lower limit for Cu Recovery = 0.84%					
Fe†	10110.2	9.29 mg/L	0.916	9.29 mg/L	0.916	9.86%
	QC value less than the lower limit for Fe Recovery = 92.90%					
K†	118492.7	96.4 mg/L	8.74	96.4 mg/L	8.74	9.07%
	QC value within limits for K Recovery = 96.36%					
Mg†	297723.0	96.7 mg/L	8.84	96.7 mg/L	8.84	9.14%
	QC value within limits for Mg Recovery = 96.66%					
Mn†	53993.1	0.0930 mg/L	0.12869	0.0930 mg/L	0.12869	138.43%
	QC value less than the lower limit for Mn Recovery = 0.93%					
Mo†	1650.5	0.124 mg/L	0.1565	0.124 mg/L	0.1565	126.56%
	QC value less than the lower limit for Mo Recovery = 1.24%					
Na†	347017.8	95.9 mg/L	8.67	95.9 mg/L	8.67	9.04%
	QC value within limits for Na Recovery = 95.88%					
Ni†	2055.5	0.0887 mg/L	0.12311	0.0887 mg/L	0.12311	138.75%
	QC value less than the lower limit for Ni Recovery = 0.89%					
Pb†	109.8	0.0219 mg/L	0.01850	0.0219 mg/L	0.01850	84.51%
	QC value less than the lower limit for Pb Recovery = 0.22%					
Sb†	32.6	0.0142 mg/L	0.01308	0.0142 mg/L	0.01308	92.19%
	QC value less than the lower limit for Sb Recovery = 0.14%					
Se†	24.6	0.0359 mg/L	0.02526	0.0359 mg/L	0.02526	70.28%
	QC value less than the lower limit for Se Recovery = 0.36%					
Tl†	127.7	0.0448 mg/L	0.03640	0.0448 mg/L	0.03640	81.20%
	QC value less than the lower limit for Tl Recovery = 0.45%					
V†	15730.3	0.0891 mg/L	0.12393	0.0891 mg/L	0.12393	139.10%
	QC value less than the lower limit for V Recovery = 0.89%					
Zn†	3976.7	0.0786 mg/L	0.11436	0.0786 mg/L	0.11436	145.55%
	QC value less than the lower limit for Zn Recovery = 0.79%					
QC Failed. Retry.						

Sequence No.: 4
Sample ID: ICV
Analyst:
Initial Sample Wt:

Autosampler Location: 15
Date Collected: 12/3/2007 8:54:16 PM
Data Type: Original
Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	865888.5	190 %		11.3			5.96%
Yr	444379.8	180 %		39.4			21.93%
Agt	-151.7	-0.00052 mg/L		0.000864	-0.00052 mg/L	0.000864	165.09%
	QC value less than the lower limit for Ag Recovery = -0.03%						
Alf	-8.1	-0.00174 mg/L		0.000374	-0.00174 mg/L	0.000374	21.48%
	QC value less than the lower limit for Al Recovery = -0.02%						
Asf	-0.6	-0.00025 mg/L		0.000820	-0.00025 mg/L	0.000820	331.79%
	QC value less than the lower limit for As Recovery = -0.00%						
B_f	186.7	0.00575 mg/L		0.000945	0.00575 mg/L	0.000945	16.45%
	QC value less than the lower limit for B_ Recovery = 0.11%						
Baf	20.5	0.00027 mg/L		0.000044	0.00027 mg/L	0.000044	16.18%
	QC value less than the lower limit for Ba Recovery = 0.00%						
Be_f	3154.8	0.00102 mg/L		0.000131	0.00102 mg/L	0.000131	12.95%
	QC value less than the lower limit for Be Recovery = 0.03%						
Ca_f	-235.8	-0.0324 mg/L		0.00960	-0.0324 mg/L	0.00960	29.62%
	QC value less than the lower limit for Ca Recovery = -0.03%						
Co_f	34.3	0.00127 mg/L		0.000116	0.00127 mg/L	0.000116	9.13%
	QC value less than the lower limit for Co Recovery = 0.01%						
Cr_f	-206.5	-0.00253 mg/L		0.000031	-0.00253 mg/L	0.000031	1.23%
	QC value less than the lower limit for Cr Recovery = -0.03%						
Cu_f	-1379.2	-0.00349 mg/L		0.000880	-0.00349 mg/L	0.000880	25.18%
	QC value less than the lower limit for Cu Recovery = -0.03%						
Fe_f	4.2	0.00387 mg/L		0.000061	0.00387 mg/L	0.000061	1.58%
	QC value less than the lower limit for Fe Recovery = 0.04%						
K_f	-84.7	-0.0689 mg/L		0.02990	-0.0689 mg/L	0.02990	43.42%
	QC value less than the lower limit for K Recovery = -0.07%						
Mg_f	27.8	0.00902 mg/L		0.001982	0.00902 mg/L	0.001982	21.96%
	QC value less than the lower limit for Mg Recovery = 0.01%						
Mn_f	-73.4	-0.00013 mg/L		0.000025	-0.00013 mg/L	0.000025	19.55%
	QC value less than the lower limit for Mn Recovery = -0.00%						
Mo_f	-5.5	-0.00042 mg/L		0.000132	-0.00042 mg/L	0.000132	31.74%
	QC value less than the lower limit for Mo Recovery = -0.00%						
Na_f	759.3	0.210 mg/L		0.0449	0.210 mg/L	0.0449	21.39%
	QC value less than the lower limit for Na Recovery = 0.21%						
Ni_f	36.2	0.00156 mg/L		0.000181	0.00156 mg/L	0.000181	11.57%
	QC value less than the lower limit for Ni Recovery = 0.02%						
Pb_f	15.0	0.00298 mg/L		0.000523	0.00298 mg/L	0.000523	17.52%
	QC value less than the lower limit for Pb Recovery = 0.03%						
Sb_f	2.4	0.00119 mg/L		0.000302	0.00119 mg/L	0.000302	25.42%
	QC value less than the lower limit for Sb Recovery = 0.01%						
Se_f	4.3	0.00310 mg/L		0.001208	0.00310 mg/L	0.001208	38.93%
	QC value less than the lower limit for Se Recovery = 0.03%						
Tl_f	25.1	0.00876 mg/L		0.000165	0.00876 mg/L	0.000165	1.88%
	QC value less than the lower limit for Tl Recovery = 0.09%						
V_f	-116.0	-0.00067 mg/L		0.000240	-0.00067 mg/L	0.000240	35.98%
	QC value less than the lower limit for V Recovery = -0.01%						
Zn_f	-284.6	-0.00568 mg/L		0.000117	-0.00568 mg/L	0.000117	2.07%
	QC value less than the lower limit for Zn Recovery = -0.06%						
QC Failed. Recalibrate and continue with analysis.							

Sequence No.: 5
 Sample ID: Calib Blank 1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 12/3/2007 8:58:22 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: Calib Blank 1

Analyte	Mean Corrected		RSD	Calib	
	Intensity	Std.Dev.		Conc.	Units
Sca	448520.1	7059.80	1.57%	100	%
Yr	245607.7	564.36	0.23%	100	%
Ag†	347.9	57.45	16.51%	[0.00]	mg/L
Al†	44.8	9.49	21.19%	[0.00]	mg/L
As†	9.2	1.87	20.37%	[0.00]	mg/L
B†	612.8	16.39	2.67%	[0.00]	mg/L
Ba†	-24.9	4.46	17.95%	[0.00]	mg/L
Be†	-3984.6	80.98	2.03%	[0.00]	mg/L
Ca†	669.4	3.40	0.51%	[0.00]	mg/L
Co†	-60.6	0.49	0.81%	[0.00]	mg/L
Cr†	324.1	5.64	1.74%	[0.00]	mg/L
Cu†	2206.6	56.09	2.54%	[0.00]	mg/L
Fe†	-6.8	0.06	0.89%	[0.00]	mg/L
K†	215.7	62.44	28.94%	[0.00]	mg/L
Mg†	-41.8	3.15	7.54%	[0.00]	mg/L
Mn†	135.5	2.62	1.93%	[0.00]	mg/L
Mo†	108.1	9.76	9.03%	[0.00]	mg/L
Nat	-365.3	41.37	11.33%	[0.00]	mg/L
Ni†	-53.5	0.32	0.60%	[0.00]	mg/L
Pb†	-10.4	7.35	70.83%	[0.00]	mg/L
Sb†	7.8	0.14	1.80%	[0.00]	mg/L
Se†	-0.1	5.04	>999.9%	[0.00]	mg/L
Tl†	-13.4	6.66	49.52%	[0.00]	mg/L
V†	194.1	57.57	29.66%	[0.00]	mg/L
Zn†	384.8	3.59	0.93%	[0.00]	mg/L

Sequence No.: 6
 Sample ID: Standard 2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:02:13 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	796063.8	120932.78	15.19%	177 %
Yr	358868.5	5352.52	1.49%	146 %
Ag†	-136.5	104.00	76.19%	[2] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Al†	-25.2	2.93	11.63%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
As†	-5.4	1.58	29.02%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
B_†	-472.1	18.80	3.98%	[5.02] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Ba†	8.3	1.41	17.01%	[10] mg/L
Be†	2968.6	129.31	4.36%	[4.01] mg/L
Ca†	-115.1	30.80	26.75%	[100] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Co†	28.8	5.13	17.85%	[10] mg/L
Cr†	-198.2	16.91	8.53%	[9.97] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Cu†	-1453.3	52.24	3.59%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Fe†	2.2	2.95	133.14%	[9.98] mg/L
K†	31.0	31.65	102.22%	[100] mg/L
Mg†	15.0	0.42	2.78%	[100] mg/L
Mn†	-124.4	0.67	0.54%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Mo†	-96.0	0.62	0.64%	[9.98] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Na†	583.7	101.09	17.32%	[100] mg/L
Ni†	23.6	2.47	10.44%	[10] mg/L
Pb†	8.4	1.37	16.22%	[10] mg/L
Sb†	-2.7	2.03	75.53%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Se†	-1.5	0.82	54.66%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Tl†	-2.0	5.50	280.03%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
V†	-115.0	19.28	16.77%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Zn†	-325.4	8.41	2.58%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ba	1	Lin, Calc Int	0.0	0.8269	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	740.3	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	2.877	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	0.2223	0.00000	1.000000	
K	1	Lin, Calc Int	0.0	0.3096	0.00000	1.000000	
Mg	1	Lin, Calc Int	0.0	0.1495	0.00000	1.000000	
Na	1	Lin, Calc Int	0.0	5.837	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	2.363	0.00000	1.000000	
Pb	1	Lin, Calc Int	0.0	0.8446	0.00000	1.000000	

Sequence No.: 7
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:07:04 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	741576.5	165 %	1.3			0.81%
Yr	109729.9	44.7 %	11.04			24.72%
Saturated within auto integration window (code 4)						
Ag†	-326.7				19.83	6.07%
Unable to evaluate QC.						
Al†	-28.4				56.11	197.43%
Unable to evaluate QC.						
As†	-5.1				1.91	37.08%
Unable to evaluate QC.						
B_†	-491.8				0.23	0.05%
Unable to evaluate QC.						
Ba†	6.5	7.83 mg/L	0.787	7.83 mg/L	0.787	10.05%
QC value less than the lower limit for Ba Recovery = 78.29%						
Be†	3317.0	4.48 mg/L	0.025	4.48 mg/L	0.025	0.56%
QC value greater than the upper limit for Be Recovery = 112.02%						
Ca†	1272.5				401.43	31.55%
Unable to evaluate QC.						
Co†	25.2	8.75 mg/L	0.663	8.75 mg/L	0.663	7.57%
QC value less than the lower limit for Co Recovery = 87.52%						
Cr†	-173.8				21.71	12.49%
Unable to evaluate QC.						
Cu†	-1745.7				10.36	0.59%
Unable to evaluate QC.						
Fe†	-8.2	-36.9 mg/L	2.03	-36.9 mg/L	2.03	5.50%
QC value less than the lower limit for Fe Recovery = -369.50%						
K†	441.1	1420 mg/L	617.2	1420 mg/L	617.2	43.32%
QC value greater than the upper limit for K Recovery = 1424.62%						
Mg†	-50.8	-340 mg/L	162.9	-340 mg/L	162.9	47.90%
QC value less than the lower limit for Mg Recovery = -340.07%						
Mn†	-120.4				1.44	1.19%
Unable to evaluate QC.						
Mot	-92.0				4.75	5.16%
Unable to evaluate QC.						
Na†	1714.5	294 mg/L	108.0	294 mg/L	108.0	36.78%
QC value greater than the upper limit for Na Recovery = 293.75%						
Ni†	20.9	8.83 mg/L	0.466	8.83 mg/L	0.466	5.27%
QC value less than the lower limit for Ni Recovery = 88.27%						
Pb†	5.8	6.87 mg/L	0.517	6.87 mg/L	0.517	7.52%
QC value less than the lower limit for Pb Recovery = 68.68%						
Sb†	-1.6				2.86	175.75%
Unable to evaluate QC.						
Se†	1.3				0.69	52.63%
Unable to evaluate QC.						
Tl†	-3.2				3.12	97.26%
Unable to evaluate QC.						
V†	-141.5				4.31	3.05%
Unable to evaluate QC.						
Zn†	-320.1				4.77	1.49%
Unable to evaluate QC.						
QC Failed. Retry.						

Sequence No.: 8
 Sample ID: ICV
 Analyst:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:09:55 PM
 Data Type: Original

Initial Sample Wt:
Dilution:

Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
All 247.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	589665.3	131 %	86.8			66.05%
Yr	410975.9	167 %	14.7			8.78%
Ag†	288.4				542.21	188.03%
	Unable to evaluate QC.					
Al†	-38.0				11.83	31.13%
	Unable to evaluate QC.					
As†	-2.3				6.10	264.22%
	Unable to evaluate QC.					
B_†	-421.9				121.35	28.76%
	Unable to evaluate QC.					
Ba†	-1.0	-1.26 mg/L	18.793	-1.26 mg/L	18.793	>999.9%
	QC value less than the lower limit for Ba Recovery = -12.55%					
Be†	1699.6	2.30 mg/L	2.415	2.30 mg/L	2.415	105.20%
	QC value less than the lower limit for Be Recovery = 57.40%					
Ca†	-140.5				70.25	50.01%
	Unable to evaluate QC.					
Co†	1.4	0.497 mg/L	13.6774	0.497 mg/L	13.6774	>999.9%
	QC value less than the lower limit for Co Recovery = 4.97%					
Cr†	-112.8				134.62	119.39%
	Unable to evaluate QC.					
Cu†	-599.2				1280.54	213.72%
	Unable to evaluate QC.					
Fe†	1.3	6.05 mg/L	6.938	6.05 mg/L	6.938	114.62%
	QC value less than the lower limit for Fe Recovery = 60.53%					
K†	-22.6	-73.0 mg/L	36.04	-73.0 mg/L	36.04	49.40%
	QC value less than the lower limit for K Recovery = -72.95%					
Mg†	15.5	104 mg/L	16.5	104 mg/L	16.5	15.95%
	QC value within limits for Mg Recovery = 103.70%					
Mn†	-106.8				18.71	17.52%
	Unable to evaluate QC.					
Mo†	-90.0				10.23	11.36%
	Unable to evaluate QC.					
Na†	1009.3	173 mg/L	17.4	173 mg/L	17.4	10.06%
	QC value greater than the upper limit for Na Recovery = 172.93%					
Ni†	4.0	1.67 mg/L	12.591	1.67 mg/L	12.591	752.57%
	QC value less than the lower limit for Ni Recovery = 16.73%					
Pb†	9.7	11.4 mg/L	2.05	11.4 mg/L	2.05	17.94%
	QC value greater than the upper limit for Pb Recovery = 114.42%					
Sb†	1.8				7.76	431.77%
	Unable to evaluate QC.					
Se†	0.1				0.25	185.32%
	Unable to evaluate QC.					
Tl†	-13.9				14.90	107.11%
	Unable to evaluate QC.					
V†	-49.2				134.84	274.04%
	Unable to evaluate QC.					
Zn†	-286.2				62.09	21.69%
	Unable to evaluate QC.					
QC Failed. Recalibrate and continue with analysis.						

Sequence No.: 9
 Sample ID: Calib Blank 1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 12/3/2007 9:14:41 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: Calib Blank 1

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	448745.2	881.22	0.20%	100 %
Yr	247232.1	5147.49	2.08%	100 %
Ag†	300.8	37.26	12.40%	[0.00] mg/L
Al†	59.7	15.87	26.59%	[0.00] mg/L
As†	2.2	1.34	60.72%	[0.00] mg/L
B_†	287.4	4.15	1.44%	[0.00] mg/L
Ba†	-31.3	0.31	0.98%	[0.00] mg/L
Be†	-3987.3	37.13	0.93%	[0.00] mg/L
Ca†	659.1	11.49	1.74%	[0.00] mg/L
Co†	-59.9	1.98	3.30%	[0.00] mg/L
Cr†	322.8	0.06	0.02%	[0.00] mg/L
Cu†	2103.3	12.43	0.59%	[0.00] mg/L
Fe†	-9.5	1.92	20.10%	[0.00] mg/L
K†	188.0	5.53	2.94%	[0.00] mg/L
Mg†	-47.0	2.76	5.88%	[0.00] mg/L
Mn†	113.5	1.70	1.50%	[0.00] mg/L
Mo†	40.1	2.65	6.61%	[0.00] mg/L
Na†	-383.7	4.06	1.06%	[0.00] mg/L
Ni†	-56.5	3.74	6.62%	[0.00] mg/L
Pb†	-10.5	3.54	33.59%	[0.00] mg/L
Sb†	7.6	0.33	4.33%	[0.00] mg/L
Se†	-3.8	0.74	19.39%	[0.00] mg/L
Tl†	-25.2	4.71	18.73%	[0.00] mg/L
V†	174.0	9.77	5.61%	[0.00] mg/L
Zn†	361.8	0.24	0.07%	[0.00] mg/L

Sequence No.: 10
 Sample ID: Standard 2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:18:32 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	531144.1	5815.88	1.09%	118 %
Saturated within auto integration window (code 4)				
Yr	303277.8	4013.83	1.32%	123 %
Ag†	232.7	30.30	13.02%	[2] mg/L
Al†	-57.6	12.33	21.42%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
As†	-0.2	0.02	6.92%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
B_†	-135.1	7.87	5.82%	[5.02] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Ba†	6.1	0.08	1.30%	[10] mg/L
Be†	2152.3	38.50	1.79%	[4.01] mg/L
Ca†	-155.3	355.34	228.86%	[100] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Co†	8.8	1.97	22.47%	[10] mg/L
Cr†	-130.4	0.94	0.72%	[9.97] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Cu†	-692.8	64.90	9.37%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Fe†	3.9	0.21	5.37%	[9.98] mg/L
K†	31.8	3.08	9.70%	[100] mg/L
Mg†	12.5	1.50	11.99%	[100] mg/L
Mn†	-98.7	4.50	4.56%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Mo†	-21.7	0.86	3.95%	[9.98] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Na†	832.3	53.02	6.37%	[100] mg/L
Ni†	11.8	4.92	41.76%	[10] mg/L
Pb†	6.5	1.12	17.24%	[10] mg/L
Sb†	1.3	0.38	30.62%	[10] mg/L
Se†	3.7	0.91	24.38%	[10] mg/L
Tl†	-4.0	3.28	82.22%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
V†	-57.0	39.31	68.94%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				
Zn†	-282.2	0.02	0.01%	[10] mg/L
No calibration curve because standard intensity and concentration values are not in the same order.				

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	116.3	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	0.6125	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	536.7	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	0.8759	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	0.3913	0.00000	1.000000	
K	1	Lin, Calc Int	0.0	0.3175	0.00000	1.000000	
Mg	1	Lin, Calc Int	0.0	0.1253	0.00000	1.000000	
Na	1	Lin, Calc Int	0.0	8.323	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	1.179	0.00000	1.000000	
Pb	1	Lin, Calc Int	0.0	0.6509	0.00000	1.000000	
Sb	1	Lin, Calc Int	0.0	0.1255	0.00000	1.000000	
Se	1	Lin, Calc Int	0.0	0.3731	0.00000	1.000000	

Sequence No.: 11
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:23:18 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
 All 247.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	717839.4	160 %	84.7			52.94%
Yr	446161.5	180 %	13.6			7.54%
Ag†	135.7	1.17 mg/L	1.438	1.17 mg/L	1.438	123.29%
QC value less than the lower limit for Ag Recovery = 58.32%						
Al†	-59.1				21.33	36.06%
Unable to evaluate QC.						
As†	1.1				2.89	269.19%
Unable to evaluate QC.						
B_†	-162.5				72.80	44.81%
Unable to evaluate QC.						
Ba†	11.7	19.1 mg/L	15.78	19.1 mg/L	15.78	82.42%
QC value greater than the upper limit for Ba Recovery = 191.49%						
Be†	2363.3	4.40 mg/L	1.466	4.40 mg/L	1.466	33.30%
QC value greater than the upper limit for Be Recovery = 110.08%						
Cat	-210.1				44.24	21.06%
Unable to evaluate QC.						
Co†	16.6	18.9 mg/L	28.08	18.9 mg/L	28.08	148.29%
QC value greater than the upper limit for Co Recovery = 189.39%						
Crt	-163.4				85.27	52.19%
Unable to evaluate QC.						
Cu†	-917.6				595.76	64.93%
Unable to evaluate QC.						
Fe†	6.2	16.0 mg/L	3.16	16.0 mg/L	3.16	19.81%
QC value greater than the upper limit for Fe Recovery = 159.65%						
K†	-81.5	-257 mg/L	141.1	-257 mg/L	141.1	54.93%
QC value less than the lower limit for K Recovery = -256.79%						
Mg†	23.5	188 mg/L	22.0	188 mg/L	22.0	11.74%
QC value greater than the upper limit for Mg Recovery = 187.64%						
Mnt	-93.9				9.75	10.39%
Unable to evaluate QC.						
Mo†	-23.1				9.97	43.18%
Unable to evaluate QC.						
Na†	596.6	71.7 mg/L	8.96	71.7 mg/L	8.96	12.50%
QC value less than the lower limit for Na Recovery = 71.68%						
Ni†	10.2	8.68 mg/L	19.256	8.68 mg/L	19.256	221.96%
QC value less than the lower limit for Ni Recovery = 86.75%						
Pb†	5.1	7.89 mg/L	1.427	7.89 mg/L	1.427	18.08%
QC value less than the lower limit for Pb Recovery = 78.89%						
Sb†	-1.4	-11.0 mg/L	16.20	-11.0 mg/L	16.20	147.18%
QC value less than the lower limit for Sb Recovery = -110.07%						
Se†	2.3	6.27 mg/L	2.708	6.27 mg/L	2.708	43.20%
QC value less than the lower limit for Se Recovery = 62.68%						
Tl†	4.0				10.26	255.58%
Unable to evaluate QC.						
V†	-60.3				56.99	94.52%
Unable to evaluate QC.						
Zn†	-290.9				35.91	12.35%
Unable to evaluate QC.						
QC Failed. Retry.						

Sequence No.: 12
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:26:09 PM
 Data Type: Original
 Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	247.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	636272.1	142 %	57.0			40.18%
Yr	355063.5	144 %	46.3			32.22%
Ag†	88.2	0.758 mg/L	0.7564	0.758 mg/L	0.7564	99.79%
	QC value less than the lower limit for Ag Recovery = 37.90%					
Al†	-35.9				27.76	77.41%
	Unable to evaluate QC.					
As†	0.5				0.63	118.58%
	Unable to evaluate QC.					
B_†	-159.6				47.80	29.94%
	Unable to evaluate QC.					
Ba†	7.7	12.5 mg/L	18.23	12.5 mg/L	18.23	145.45%
	QC value greater than the upper limit for Ba Recovery = 125.33%					
Be†	2291.3	4.27 mg/L	1.434	4.27 mg/L	1.434	33.58%
	QC value greater than the upper limit for Be Recovery = 106.72%					
Cat	-41.7				245.49	589.12%
	Unable to evaluate QC.					
Co†	14.2	16.2 mg/L	22.08	16.2 mg/L	22.08	136.17%
	QC value greater than the upper limit for Co Recovery = 162.12%					
Crt	-142.5				70.27	49.30%
	Unable to evaluate QC.					
Cu†	-933.3				515.52	55.24%
	Unable to evaluate QC.					
Fe†	4.1	10.6 mg/L	0.25	10.6 mg/L	0.25	2.36%
	QC value greater than the upper limit for Fe Recovery = 105.67%					
K†	2.2	6.80 mg/L	173.366	6.80 mg/L	173.366	>999.9%
	QC value less than the lower limit for K Recovery = 6.80%					
Mg†	14.3	114 mg/L	95.3	114 mg/L	95.3	83.36%
	QC value greater than the upper limit for Mg Recovery = 114.29%					
Mnt	-103.1				2.83	2.74%
	Unable to evaluate QC.					
Mo†	-23.7				6.27	26.45%
	Unable to evaluate QC.					
Na†	1131.2	136 mg/L	29.9	136 mg/L	29.9	21.99%
	QC value greater than the upper limit for Na Recovery = 135.90%					
Ni†	14.3	12.1 mg/L	13.59	12.1 mg/L	13.59	112.04%
	QC value greater than the upper limit for Ni Recovery = 121.33%					
Pb†	8.4	13.0 mg/L	8.64	13.0 mg/L	8.64	66.60%
	QC value greater than the upper limit for Pb Recovery = 129.69%					
Sb†	-1.6	-12.9 mg/L	3.87	-12.9 mg/L	3.87	29.92%
	QC value less than the lower limit for Sb Recovery = -129.31%					
Se†	1.6	4.44 mg/L	4.187	4.44 mg/L	4.187	94.26%
	QC value less than the lower limit for Se Recovery = 44.42%					
Tl†	-0.9				9.14	981.66%
	Unable to evaluate QC.					
V†	-71.1				26.44	37.18%
	Unable to evaluate QC.					
Zn†	-287.4				27.90	9.71%
	Unable to evaluate QC.					

Max reruns exceeded. Stop analysis.

User canceled analysis.

Analysis Begun

Start Time: 12/3/2007 9:38:58 PM Plasma On Time: 12/3/2007 6:19:37 PM
Logged In Analyst: Administrator Technique: ICP Continuous
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Owner\Sample Information\071203.sif
Batch ID: 071203
Results Data Set: 071203A
Results Library: C:\pe\Owner\Results\Results.mdb

Sequence No.: 1 Autosampler Location: 0
Sample ID: Calib Blank 1 Date Collected: 12/3/2007 9:38:58 PM
Analyst: Data Type: Original
Initial Sample Wt: Initial Sample Vol:
Dilution: Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1
Analyte Back Pressure Flow
All 246.0 kPa 0.65 L/min

Mean Data: Calib Blank 1

Table with 5 columns: Analyte, Mean Corrected Intensity, Std.Dev., RSD, and Calib Conc. Units. Lists various elements like Sca, Yr, Ag, Al, As, B, Ba, Be, Ca, Co, Cr, Cu, Fe, K, Mg, Mn, Mo, Na, Ni, Pb, Sb, Se, Tl, V, Zn with their respective intensity, standard deviation, relative standard deviation, and calibration concentration.

Sequence No.: 2
 Sample ID: Standard 2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:42:49 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected	Std.Dev.	RSD	Conc.	Units
Sca	399424.5	2903.71	0.73%	90.4	%
Yr	228351.7	661.75	0.29%	93.8	%
Ag†	554193.7	558.29	0.10%	[2]	mg/L
Al†	44733.8	278.29	0.62%	[10]	mg/L
As†	20846.4	268.93	1.29%	[10]	mg/L
B_†	154743.4	149.10	0.10%	[5.02]	mg/L
Ba†	722323.9	2370.32	0.33%	[10]	mg/L
Be†	12164741.1	154019.37	1.27%	[4.01]	mg/L
Ca†	708904.4	178.25	0.03%	[100]	mg/L
Co†	260195.5	591.00	0.23%	[10]	mg/L
Cr†	782048.2	1950.33	0.25%	[9.97]	mg/L
Cu†	3754803.0	6209.30	0.17%	[10]	mg/L
Fe†	10443.2	78.67	0.75%	[9.98]	mg/L
K†	120436.3	1671.59	1.39%	[100]	mg/L
Mg†	300605.9	427.19	0.14%	[100]	mg/L
Mn†	5589475.8	17267.47	0.31%	[10]	mg/L
Mo†	128090.4	170.70	0.13%	[9.98]	mg/L
Na†	351501.0	5603.50	1.59%	[100]	mg/L
Ni†	222279.3	214.70	0.10%	[10]	mg/L
Pb†	48312.5	556.01	1.15%	[10]	mg/L
Sb†	19938.2	217.84	1.09%	[10]	mg/L
Se†	13081.0	182.52	1.40%	[10]	mg/L
Tl†	27502.9	346.96	1.26%	[10]	mg/L
V†	1705312.3	3013.23	0.18%	[10]	mg/L
Zn†	482288.8	427.88	0.09%	[10]	mg/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Rslope
Ag	1	Lin, Calc Int	0.0	277100	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	4473	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	2085	0.00000	1.000000	
B_	1	Lin, Calc Int	-0.0	30830	0.00000	1.000000	
Ba	1	Lin, Calc Int	-0.0	72230	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	3034000	0.00000	1.000000	
Ca	1	Lin, Calc Int	-0.0	7089	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	26020	0.00000	1.000000	
Cr	1	Lin, Calc Int	0.0	78440	0.00000	1.000000	
Cu	1	Lin, Calc Int	0.0	375500	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	1046	0.00000	1.000000	
K	1	Lin, Calc Int	-0.0	1204	0.00000	1.000000	
Mg	1	Lin, Calc Int	0.0	3006	0.00000	1.000000	
Mn	1	Lin, Calc Int	0.0	558900	0.00000	1.000000	
Mo	1	Lin, Calc Int	0.0	12830	0.00000	1.000000	
Na	1	Lin, Calc Int	0.0	3515	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	22230	0.00000	1.000000	
Pb	1	Lin, Calc Int	-0.0	4831	0.00000	1.000000	
Sb	1	Lin, Calc Int	0.0	1994	0.00000	1.000000	
Se	1	Lin, Calc Int	-0.0	1308	0.00000	1.000000	
Tl	1	Lin, Calc Int	-0.0	2750	0.00000	1.000000	
V	1	Lin, Calc Int	-0.0	170500	0.00000	1.000000	
Zn	1	Lin, Calc Int	-0.0	48230	0.00000	1.000000	

Sequence No.: 3
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 12/3/2007 9:46:33 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	387456.6	87.7 %		0.18			0.20%
Yr	223249.5	91.7 %		0.70			0.77%
Ag†	569697.7	2.06 mg/L		0.007	2.06 mg/L	0.007	0.34%
	QC value within limits for Ag	Recovery = 102.80%					
Al†	45970.1	10.3 mg/L		0.24	10.3 mg/L	0.24	2.33%
	QC value within limits for Al	Recovery = 102.76%					
As†	21840.1	10.5 mg/L		0.01	10.5 mg/L	0.01	0.12%
	QC value within limits for As	Recovery = 104.76%					
B_†	159924.4	5.17 mg/L		0.013	5.17 mg/L	0.013	0.26%
	QC value within limits for B_	Recovery = 103.34%					
Ba†	745090.5	10.3 mg/L		0.03	10.3 mg/L	0.03	0.33%
	QC value within limits for Ba	Recovery = 103.15%					
Be†	12479079.0	4.11 mg/L		0.036	4.11 mg/L	0.036	0.88%
	QC value within limits for Be	Recovery = 102.84%					
Ca†	727907.2	103 mg/L		0.6	103 mg/L	0.6	0.54%
	QC value within limits for Ca	Recovery = 102.68%					
Co†	268380.3	10.3 mg/L		0.04	10.3 mg/L	0.04	0.40%
	QC value within limits for Co	Recovery = 103.15%					
Cr†	805720.5	10.3 mg/L		0.04	10.3 mg/L	0.04	0.42%
	QC value within limits for Cr	Recovery = 102.72%					
Cu†	3874732.4	10.3 mg/L		0.03	10.3 mg/L	0.03	0.32%
	QC value within limits for Cu	Recovery = 103.29%					
Fe†	10724.1	10.2 mg/L		0.21	10.2 mg/L	0.21	2.02%
	QC value within limits for Fe	Recovery = 102.48%					
K†	123821.2	103 mg/L		1.5	103 mg/L	1.5	1.42%
	QC value within limits for K	Recovery = 102.81%					
Mg†	309341.2	103 mg/L		0.6	103 mg/L	0.6	0.62%
	QC value within limits for Mg	Recovery = 102.91%					
Mn†	5765256.3	10.3 mg/L		0.04	10.3 mg/L	0.04	0.40%
	QC value within limits for Mn	Recovery = 103.14%					
Mo†	132179.1	10.3 mg/L		0.03	10.3 mg/L	0.03	0.34%
	QC value within limits for Mo	Recovery = 102.99%					
Na†	360384.7	103 mg/L		0.9	103 mg/L	0.9	0.86%
	QC value within limits for Na	Recovery = 102.53%					
Ni†	228924.1	10.3 mg/L		0.04	10.3 mg/L	0.04	0.37%
	QC value within limits for Ni	Recovery = 102.99%					
Pb†	50227.8	10.4 mg/L		0.01	10.4 mg/L	0.01	0.09%
	QC value within limits for Pb	Recovery = 103.96%					
Sb†	20691.2	10.2 mg/L		0.06	10.2 mg/L	0.06	0.55%
	QC value within limits for Sb	Recovery = 102.15%					
Se†	13625.1	10.4 mg/L		0.01	10.4 mg/L	0.01	0.14%
	QC value within limits for Se	Recovery = 104.36%					
Tl†	28713.2	10.5 mg/L		0.04	10.5 mg/L	0.04	0.38%
	QC value within limits for Tl	Recovery = 104.69%					
V†	1756209.0	10.4 mg/L		0.04	10.4 mg/L	0.04	0.41%
	QC value within limits for V	Recovery = 103.55%					
Zn†	497248.1	10.2 mg/L		0.04	10.2 mg/L	0.04	0.35%
	QC value within limits for Zn	Recovery = 102.40%					

All analyte(s) passed QC.

Sequence No.: 4
 Sample ID: QCS1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 12/3/2007 9:50:18 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS1

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: QCS1

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	407002.4	92.1 %		0.51			0.55%
Yr	234737.5	96.5 %		0.43			0.45%
Ag†	585202.2	2.11 mg/L		0.005	2.11 mg/L	0.005	0.25%
	QC value greater than the upper limit for Ag Recovery = 105.60%						
Al†	46923.4	10.5 mg/L		0.03	10.5 mg/L	0.03	0.32%
	QC value within limits for Al Recovery = 104.89%						
As†	22086.1	10.6 mg/L		0.02	10.6 mg/L	0.02	0.17%
	QC value greater than the upper limit for As Recovery = 105.94%						
B_†	165428.3	5.35 mg/L		0.030	5.35 mg/L	0.030	0.56%
	QC value greater than the upper limit for B_ Recovery = 106.90%						
Ba†	760842.1	10.5 mg/L		0.02	10.5 mg/L	0.02	0.17%
	QC value greater than the upper limit for Ba Recovery = 105.33%						
Be†	12935832.5	4.26 mg/L		0.050	4.26 mg/L	0.050	1.17%
	QC value greater than the upper limit for Be Recovery = 106.60%						
Ca†	750809.9	106 mg/L		0.1	106 mg/L	0.1	0.10%
	QC value greater than the upper limit for Ca Recovery = 105.91%						
Co†	277405.8	10.7 mg/L		0.00	10.7 mg/L	0.00	0.01%
	QC value greater than the upper limit for Co Recovery = 106.61%						
Cr†	826204.8	10.5 mg/L		0.03	10.5 mg/L	0.03	0.30%
	QC value greater than the upper limit for Cr Recovery = 105.33%						
Cu†	4015750.8	10.7 mg/L		0.04	10.7 mg/L	0.04	0.39%
	QC value greater than the upper limit for Cu Recovery = 107.05%						
Fe†	11138.1	10.6 mg/L		0.09	10.6 mg/L	0.09	0.89%
	QC value greater than the upper limit for Fe Recovery = 106.44%						
K†	124540.2	103 mg/L		0.3	103 mg/L	0.3	0.28%
	QC value within limits for K Recovery = 103.41%						
Mg†	315267.8	105 mg/L		0.3	105 mg/L	0.3	0.32%
	QC value within limits for Mg Recovery = 104.88%						
Mn†	5911032.9	10.6 mg/L		0.01	10.6 mg/L	0.01	0.11%
	QC value greater than the upper limit for Mn Recovery = 105.75%						
Mo†	134949.3	10.5 mg/L		0.02	10.5 mg/L	0.02	0.22%
	QC value greater than the upper limit for Mo Recovery = 105.14%						
Na†	369074.5	105 mg/L		0.4	105 mg/L	0.4	0.43%
	QC value within limits for Na Recovery = 105.00%						
Ni†	236321.1	10.6 mg/L		0.01	10.6 mg/L	0.01	0.07%
	QC value greater than the upper limit for Ni Recovery = 106.32%						
Pb†	51741.7	10.7 mg/L		0.04	10.7 mg/L	0.04	0.35%
	QC value greater than the upper limit for Pb Recovery = 107.10%						
Sb†	20899.3	10.3 mg/L		0.07	10.3 mg/L	0.07	0.66%
	QC value within limits for Sb Recovery = 103.15%						
Se†	14211.0	10.9 mg/L		0.03	10.9 mg/L	0.03	0.32%
	QC value greater than the upper limit for Se Recovery = 108.85%						
Tl†	29808.0	10.9 mg/L		0.04	10.9 mg/L	0.04	0.39%
	QC value greater than the upper limit for Tl Recovery = 108.68%						
V†	1801892.6	10.6 mg/L		0.04	10.6 mg/L	0.04	0.35%
	QC value greater than the upper limit for V Recovery = 106.24%						
Zn†	514973.1	10.6 mg/L		0.02	10.6 mg/L	0.02	0.23%
	QC value greater than the upper limit for Zn Recovery = 106.05%						
QC Failed. Continue with analysis.							

Sequence No.: 5
 Sample ID: QCS2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 12/3/2007 9:54:03 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS2

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: QCS2

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	398752.2	90.3 %		0.36			0.40%
Yr	229244.8	94.2 %		0.52			0.55%
Ag†	577390.5	2.08 mg/L		0.001	2.08 mg/L	0.001	0.05%
	QC value within limits for Ag Recovery = 104.19%						
Al†	46470.6	10.4 mg/L		0.13	10.4 mg/L	0.13	1.27%
	QC value within limits for Al Recovery = 103.88%						
As†	21376.0	10.3 mg/L		0.01	10.3 mg/L	0.01	0.09%
	QC value within limits for As Recovery = 102.53%						
B_†	162220.5	5.24 mg/L		0.003	5.24 mg/L	0.003	0.06%
	QC value within limits for B_ Recovery = 104.83%						
Ba†	750669.6	10.4 mg/L		0.01	10.4 mg/L	0.01	0.12%
	QC value within limits for Ba Recovery = 103.92%						
Be†	12646794.1	4.17 mg/L		0.001	4.17 mg/L	0.001	0.01%
	QC value within limits for Be Recovery = 104.22%						
Ca†	744141.4	105 mg/L		0.2	105 mg/L	0.2	0.18%
	QC value within limits for Ca Recovery = 104.97%						
Co†	272752.2	10.5 mg/L		0.00	10.5 mg/L	0.00	0.02%
	QC value within limits for Co Recovery = 104.83%						
Cr†	815484.8	10.4 mg/L		0.01	10.4 mg/L	0.01	0.09%
	QC value within limits for Cr Recovery = 103.96%						
Cu†	3957646.3	10.6 mg/L		0.03	10.6 mg/L	0.03	0.26%
	QC value greater than the upper limit for Cu Recovery = 105.50%						
Fe†	10992.1	10.5 mg/L		0.15	10.5 mg/L	0.15	1.46%
	QC value greater than the upper limit for Fe Recovery = 105.05%						
K†	124155.1	103 mg/L		1.5	103 mg/L	1.5	1.42%
	QC value within limits for K Recovery = 103.09%						
Mg†	313419.3	104 mg/L		0.2	104 mg/L	0.2	0.20%
	QC value within limits for Mg Recovery = 104.26%						
Mn†	5831103.4	10.4 mg/L		0.01	10.4 mg/L	0.01	0.10%
	QC value within limits for Mn Recovery = 104.32%						
Mo†	133181.1	10.4 mg/L		0.02	10.4 mg/L	0.02	0.23%
	QC value within limits for Mo Recovery = 103.77%						
Na†	368212.2	105 mg/L		1.3	105 mg/L	1.3	1.22%
	QC value within limits for Na Recovery = 104.75%						
Ni†	232606.6	10.5 mg/L		0.00	10.5 mg/L	0.00	0.02%
	QC value within limits for Ni Recovery = 104.65%						
Pb†	50632.3	10.5 mg/L		0.02	10.5 mg/L	0.02	0.15%
	QC value within limits for Pb Recovery = 104.80%						
Sb†	20379.2	10.1 mg/L		0.02	10.1 mg/L	0.02	0.17%
	QC value within limits for Sb Recovery = 100.56%						
Se†	13767.4	10.5 mg/L		0.02	10.5 mg/L	0.02	0.22%
	QC value greater than the upper limit for Se Recovery = 105.45%						
Tl†	29124.1	10.6 mg/L		0.03	10.6 mg/L	0.03	0.24%
	QC value greater than the upper limit for Tl Recovery = 106.19%						
V†	1778014.1	10.5 mg/L		0.01	10.5 mg/L	0.01	0.11%
	QC value within limits for V Recovery = 104.84%						
Zn†	507760.1	10.5 mg/L		0.01	10.5 mg/L	0.01	0.06%
	QC value within limits for Zn Recovery = 104.57%						
QC Failed. Continue with analysis.							

Sequence No.: 6
 Sample ID: QCS3
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 12/3/2007 9:57:48 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS3

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: QCS3

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	401608.9	90.9 %		0.03			0.03%
Yr	230260.3	94.6 %		0.83			0.87%
Ag†	575951.1	2.08 mg/L		0.001	2.08 mg/L	0.001	0.07%
	QC value within limits for Ag Recovery = 103.93%						
Al†	46909.3	10.5 mg/L		0.03	10.5 mg/L	0.03	0.32%
	QC value within limits for Al Recovery = 104.86%						
As†	21605.5	10.4 mg/L		0.24	10.4 mg/L	0.24	2.30%
	QC value within limits for As Recovery = 103.64%						
B_†	162108.5	5.24 mg/L		0.003	5.24 mg/L	0.003	0.05%
	QC value within limits for B_ Recovery = 104.76%						
Ba†	748357.1	10.4 mg/L		0.02	10.4 mg/L	0.02	0.17%
	QC value within limits for Ba Recovery = 103.60%						
Be†	12676734.1	4.18 mg/L		0.030	4.18 mg/L	0.030	0.72%
	QC value within limits for Be Recovery = 104.47%						
Ca†	741887.5	105 mg/L		0.1	105 mg/L	0.1	0.05%
	QC value within limits for Ca Recovery = 104.65%						
Co†	272245.1	10.5 mg/L		0.02	10.5 mg/L	0.02	0.20%
	QC value within limits for Co Recovery = 104.63%						
Cr†	813716.7	10.4 mg/L		0.01	10.4 mg/L	0.01	0.14%
	QC value within limits for Cr Recovery = 103.74%						
Cu†	3946741.8	10.5 mg/L		0.01	10.5 mg/L	0.01	0.11%
	QC value greater than the upper limit for Cu Recovery = 105.21%						
Fe†	11132.5	10.6 mg/L		0.03	10.6 mg/L	0.03	0.27%
	QC value greater than the upper limit for Fe Recovery = 106.39%						
K†	123565.3	103 mg/L		0.5	103 mg/L	0.5	0.46%
	QC value within limits for K Recovery = 102.60%						
Mg†	312340.0	104 mg/L		0.2	104 mg/L	0.2	0.15%
	QC value within limits for Mg Recovery = 103.90%						
Mn†	5817988.1	10.4 mg/L		0.02	10.4 mg/L	0.02	0.17%
	QC value within limits for Mn Recovery = 104.09%						
Mo†	132905.6	10.4 mg/L		0.01	10.4 mg/L	0.01	0.08%
	QC value within limits for Mo Recovery = 103.55%						
Na†	365767.5	104 mg/L		0.1	104 mg/L	0.1	0.09%
	QC value within limits for Na Recovery = 104.06%						
Ni†	232632.9	10.5 mg/L		0.02	10.5 mg/L	0.02	0.16%
	QC value within limits for Ni Recovery = 104.66%						
Pb†	50921.4	10.5 mg/L		0.18	10.5 mg/L	0.18	1.66%
	QC value greater than the upper limit for Pb Recovery = 105.40%						
Sb†	20442.8	10.1 mg/L		0.22	10.1 mg/L	0.22	2.17%
	QC value within limits for Sb Recovery = 100.89%						
Se†	13860.8	10.6 mg/L		0.24	10.6 mg/L	0.24	2.29%
	QC value greater than the upper limit for Se Recovery = 106.17%						
Tl†	29270.4	10.7 mg/L		0.16	10.7 mg/L	0.16	1.50%
	QC value greater than the upper limit for Tl Recovery = 106.72%						
V†	1774661.6	10.5 mg/L		0.01	10.5 mg/L	0.01	0.14%
	QC value within limits for V Recovery = 104.64%						
Zn†	507307.2	10.4 mg/L		0.00	10.4 mg/L	0.00	0.01%
	QC value within limits for Zn Recovery = 104.47%						
	QC Failed. Continue with analysis.						

Sequence No.: 7
 Sample ID: LINEARITY
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 9
 Date Collected: 12/3/2007 10:01:33 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	377794.3	85.5 %		0.86			1.01%
Yr	216203.4	88.9 %		0.52			0.58%
Ag†	-9801.6	-0.0354 mg/L		0.00013	-0.0354 mg/L	0.00013	0.37%
QC value within limits for Ag		Recovery = Not calculated					
Al†	-26.7	-0.00598 mg/L		0.005846	-0.00598 mg/L	0.005846	97.82%
QC value within limits for Al		Recovery = Not calculated					
As†	-141.0	-0.00654 mg/L		0.000410	-0.00654 mg/L	0.000410	6.28%
QC value within limits for As		Recovery = Not calculated					
B_†	1170.9	0.0380 mg/L		0.00353	0.0380 mg/L	0.00353	9.31%
QC value within limits for B_		Recovery = Not calculated					
Ba†	138.4	0.00192 mg/L		0.000194	0.00192 mg/L	0.000194	10.15%
QC value within limits for Ba		Recovery = Not calculated					
Be†	-1041.9	-0.00034 mg/L		0.000045	-0.00034 mg/L	0.000045	12.98%
QC value within limits for Be		Recovery = Not calculated					
Ca†	2137352.9	302 mg/L		4.3	302 mg/L	4.3	1.44%
QC value within limits for Ca		Recovery = 100.50%					
Co†	66.5	0.00255 mg/L		0.000153	0.00255 mg/L	0.000153	5.97%
QC value within limits for Co		Recovery = Not calculated					
Cr†	145.9	0.00186 mg/L		0.000117	0.00186 mg/L	0.000117	6.29%
QC value within limits for Cr		Recovery = Not calculated					
Cu†	-4335.3	-0.0115 mg/L		0.00030	-0.0115 mg/L	0.00030	2.56%
QC value within limits for Cu		Recovery = Not calculated					
Fe†	105849.7	101 mg/L		0.2	101 mg/L	0.2	0.15%
QC value within limits for Fe		Recovery = 101.15%					
K†	380079.7	316 mg/L		2.8	316 mg/L	2.8	0.89%
QC value within limits for K		Recovery = 105.20%					
Mg†	593970.7	198 mg/L		2.4	198 mg/L	2.4	1.22%
QC value within limits for Mg		Recovery = 98.80%					
Mn†	1983.7	0.00355 mg/L		0.000052	0.00355 mg/L	0.000052	1.46%
QC value within limits for Mn		Recovery = Not calculated					
Mo†	130.1	0.0101 mg/L		0.00222	0.0101 mg/L	0.00222	21.92%
QC value within limits for Mo		Recovery = Not calculated					
Na†	1104709.0	314 mg/L		3.5	314 mg/L	3.5	1.11%
QC value within limits for Na		Recovery = 104.76%					
Ni†	5.8	0.00026 mg/L		0.000057	0.00026 mg/L	0.000057	21.82%
QC value within limits for Ni		Recovery = Not calculated					
Pb†	-12.5	-0.00258 mg/L		0.001733	-0.00258 mg/L	0.001733	67.15%
QC value within limits for Pb		Recovery = Not calculated					
Sb†	54.2	0.0272 mg/L		0.00303	0.0272 mg/L	0.00303	11.17%
QC value within limits for Sb		Recovery = Not calculated					
Se†	-329.3	-0.0549 mg/L		0.01365	-0.0549 mg/L	0.01365	24.88%
QC value within limits for Se		Recovery = Not calculated					
Tl†	76.4	0.0278 mg/L		0.00679	0.0278 mg/L	0.00679	24.45%
QC value within limits for Tl		Recovery = Not calculated					
V†	-541.2	-0.00316 mg/L		0.000308	-0.00316 mg/L	0.000308	9.74%
QC value within limits for V		Recovery = Not calculated					
Zn†	976.0	0.0202 mg/L		0.00014	0.0202 mg/L	0.00014	0.68%
QC value within limits for Zn		Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 8
 Sample ID: ICSA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 12/3/2007 10:05:43 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	396929.4	89.9 %		0.66			0.74%
Yr	225653.1	92.7 %		0.15			0.16%
Ag†	-9773.1	-0.0353 mg/L		0.00035	-0.0353 mg/L	0.00035	0.99%
	QC value within limits for Ag	Recovery = Not calculated					
Al†	1101821.8	246 mg/L		0.6	246 mg/L	0.6	0.25%
	QC value within limits for Al	Recovery = 98.52%					
As†	-427.5	-0.145 mg/L		0.0012	-0.145 mg/L	0.0012	0.82%
	QC value within limits for As	Recovery = Not calculated					
B_†	327.3	0.0106 mg/L		0.00203	0.0106 mg/L	0.00203	19.12%
	QC value within limits for B_	Recovery = Not calculated					
Ba†	164.8	0.00228 mg/L		0.000084	0.00228 mg/L	0.000084	3.68%
	QC value within limits for Ba	Recovery = Not calculated					
Be†	-1000.2	-0.00033 mg/L		0.000002	-0.00033 mg/L	0.000002	0.67%
	QC value within limits for Be	Recovery = Not calculated					
Ca†	1777610.3	251 mg/L		0.4	251 mg/L	0.4	0.16%
	QC value within limits for Ca	Recovery = 100.30%					
Co†	35.2	0.00135 mg/L		0.000071	0.00135 mg/L	0.000071	5.24%
	QC value within limits for Co	Recovery = Not calculated					
Cr†	-34.5	-0.00044 mg/L		0.000037	-0.00044 mg/L	0.000037	8.39%
	QC value within limits for Cr	Recovery = Not calculated					
Cu†	-4796.3	-0.0128 mg/L		0.00008	-0.0128 mg/L	0.00008	0.61%
	QC value within limits for Cu	Recovery = Not calculated					
Fe†	103836.9	99.2 mg/L		0.05	99.2 mg/L	0.05	0.05%
	QC value within limits for Fe	Recovery = 99.23%					
K†	235.5	0.196 mg/L		0.0634	0.196 mg/L	0.0634	32.42%
	QC value within limits for K	Recovery = Not calculated					
Mg†	724905.3	241 mg/L		0.3	241 mg/L	0.3	0.14%
	QC value within limits for Mg	Recovery = 96.46%					
Mn†	2197.3	0.00393 mg/L		0.000005	0.00393 mg/L	0.000005	0.12%
	QC value within limits for Mn	Recovery = Not calculated					
Mo†	45.2	0.00352 mg/L		0.000130	0.00352 mg/L	0.000130	3.69%
	QC value within limits for Mo	Recovery = Not calculated					
Na†	660.2	0.188 mg/L		0.0072	0.188 mg/L	0.0072	3.86%
	QC value within limits for Na	Recovery = Not calculated					
Ni†	-21.7	-0.00098 mg/L		0.000296	-0.00098 mg/L	0.000296	30.23%
	QC value within limits for Ni	Recovery = Not calculated					
Pb†	-183.2	-0.0379 mg/L		0.00140	-0.0379 mg/L	0.00140	3.68%
	QC value within limits for Pb	Recovery = Not calculated					
Sb†	36.9	0.0185 mg/L		0.00075	0.0185 mg/L	0.00075	4.04%
	QC value within limits for Sb	Recovery = Not calculated					
Se†	-349.8	-0.0743 mg/L		0.00326	-0.0743 mg/L	0.00326	4.39%
	QC value within limits for Se	Recovery = Not calculated					
Tl†	43.2	0.0157 mg/L		0.00633	0.0157 mg/L	0.00633	40.34%
	QC value within limits for Tl	Recovery = Not calculated					
V†	-583.8	-0.00343 mg/L		0.000046	-0.00343 mg/L	0.000046	1.35%
	QC value within limits for V	Recovery = Not calculated					
Zn†	688.4	0.0143 mg/L		0.00017	0.0143 mg/L	0.00017	1.16%
	QC value within limits for Zn	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 9
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 11
 Date Collected: 12/3/2007 10:09:52 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	393544.1	89.1 %	0.20			0.23%
Yr	222609.1	91.5 %	1.15			1.25%
Ag†	138462.8	0.500 mg/L	0.0005	0.500 mg/L	0.0005	0.10%
	QC value within limits for Ag Recovery = 99.94%					
Al†	1113081.3	249 mg/L	4.0	249 mg/L	4.0	1.61%
	QC value within limits for Al Recovery = 99.53%					
As†	-439.2	-0.151 mg/L	0.0036	-0.151 mg/L	0.0036	2.37%
	QC value less than the lower limit for As Recovery = Not calculated					
B_†	126.7	0.00363 mg/L	0.000864	0.00363 mg/L	0.000864	23.82%
	QC value within limits for B_ Recovery = Not calculated					
Ba†	18558.7	0.257 mg/L	0.0002	0.257 mg/L	0.0002	0.09%
	QC value within limits for Ba Recovery = 102.77%					
Be†	736900.7	0.243 mg/L	0.0002	0.243 mg/L	0.0002	0.09%
	QC value within limits for Be Recovery = 97.17%					
Ca†	1798259.9	254 mg/L	4.1	254 mg/L	4.1	1.61%
	QC value within limits for Ca Recovery = 101.47%					
Co†	6161.5	0.237 mg/L	0.0002	0.237 mg/L	0.0002	0.08%
	QC value within limits for Co Recovery = 94.72%					
Cr†	19108.6	0.244 mg/L	0.0009	0.244 mg/L	0.0009	0.37%
	QC value within limits for Cr Recovery = 97.44%					
Cu†	90490.0	0.241 mg/L	0.0001	0.241 mg/L	0.0001	0.04%
	QC value within limits for Cu Recovery = 96.49%					
Fe†	103922.1	99.3 mg/L	0.21	99.3 mg/L	0.21	0.21%
	QC value within limits for Fe Recovery = 99.31%					
K†	248.8	0.207 mg/L	0.0247	0.207 mg/L	0.0247	11.95%
	QC value within limits for K Recovery = Not calculated					
Mg†	733325.7	244 mg/L	3.6	244 mg/L	3.6	1.46%
	QC value within limits for Mg Recovery = 97.58%					
Mn†	144278.2	0.258 mg/L	0.0007	0.258 mg/L	0.0007	0.28%
	QC value within limits for Mn Recovery = 103.25%					
Mo†	22.9	0.00179 mg/L	0.000841	0.00179 mg/L	0.000841	47.08%
	QC value within limits for Mo Recovery = Not calculated					
Na†	386.0	0.110 mg/L	0.0162	0.110 mg/L	0.0162	14.74%
	QC value within limits for Na Recovery = Not calculated					
Ni†	10235.6	0.460 mg/L	0.0008	0.460 mg/L	0.0008	0.17%
	QC value within limits for Ni Recovery = 92.10%					
Pb†	2217.2	0.459 mg/L	0.0005	0.459 mg/L	0.0005	0.12%
	QC value within limits for Pb Recovery = 91.79%					
Sb†	38.9	0.0156 mg/L	0.00545	0.0156 mg/L	0.00545	34.84%
	QC value within limits for Sb Recovery = Not calculated					
Se†	-327.6	-0.0571 mg/L	0.02222	-0.0571 mg/L	0.02222	38.89%
	QC value within limits for Se Recovery = Not calculated					
Tl†	19.1	0.00764 mg/L	0.002414	0.00764 mg/L	0.002414	31.61%
	QC value within limits for Tl Recovery = Not calculated					
V†	41187.8	0.243 mg/L	0.0007	0.243 mg/L	0.0007	0.27%
	QC value within limits for V Recovery = 97.15%					
Zn†	25480.7	0.525 mg/L	0.0012	0.525 mg/L	0.0012	0.23%
	QC value within limits for Zn Recovery = 105.04%					
QC Failed. Retry.						

Sequence No.: 10
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:

Autosampler Location: 11
 Date Collected: 12/3/2007 10:12:05 PM
 Data Type: Original
 Initial Sample Vol:

Dilution:

Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	388690.9	88.0	%	0.76			0.87%
Yr	223610.2	91.9	%	0.69			0.75%
Ag†	137559.6	0.496	mg/L	0.0022	0.496	mg/L	0.0022 0.44%
	QC value within limits for Ag Recovery = 99.29%						
Al†	1092254.6	244	mg/L	1.1	244	mg/L	1.1 0.46%
	QC value within limits for Al Recovery = 97.67%						
As†	-439.2	-0.151	mg/L	0.0034	-0.151	mg/L	0.0034 2.28%
	QC value less than the lower limit for As Recovery = Not calculated						
B_†	15.6	0.00003	mg/L	0.000168	0.00003	mg/L	0.000168 663.53%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	18740.2	0.259	mg/L	0.0020	0.259	mg/L	0.0020 0.75%
	QC value within limits for Ba Recovery = 103.78%						
Be†	732972.3	0.242	mg/L	0.0002	0.242	mg/L	0.0002 0.09%
	QC value within limits for Be Recovery = 96.65%						
Ca†	1773762.0	250	mg/L	0.2	250	mg/L	0.2 0.08%
	QC value within limits for Ca Recovery = 100.08%						
Co†	6198.0	0.238	mg/L	0.0026	0.238	mg/L	0.0026 1.08%
	QC value within limits for Co Recovery = 95.28%						
Cr†	19294.5	0.246	mg/L	0.0021	0.246	mg/L	0.0021 0.85%
	QC value within limits for Cr Recovery = 98.39%						
Cu†	90198.3	0.240	mg/L	0.0008	0.240	mg/L	0.0008 0.33%
	QC value within limits for Cu Recovery = 96.18%						
Fe†	103758.5	99.2	mg/L	0.44	99.2	mg/L	0.44 0.44%
	QC value within limits for Fe Recovery = 99.16%						
K†	211.0	0.175	mg/L	0.0108	0.175	mg/L	0.0108 6.14%
	QC value within limits for K Recovery = Not calculated						
Mg†	724958.0	241	mg/L	0.6	241	mg/L	0.6 0.26%
	QC value within limits for Mg Recovery = 96.47%						
Mn†	143995.0	0.258	mg/L	0.0003	0.258	mg/L	0.0003 0.13%
	QC value within limits for Mn Recovery = 103.05%						
Mo†	22.6	0.00176	mg/L	0.001047	0.00176	mg/L	0.001047 59.37%
	QC value within limits for Mo Recovery = Not calculated						
Na†	477.1	0.136	mg/L	0.0088	0.136	mg/L	0.0088 6.48%
	QC value within limits for Na Recovery = Not calculated						
Ni†	10335.4	0.465	mg/L	0.0034	0.465	mg/L	0.0034 0.74%
	QC value within limits for Ni Recovery = 92.99%						
Pb†	2226.2	0.461	mg/L	0.0032	0.461	mg/L	0.0032 0.69%
	QC value within limits for Pb Recovery = 92.16%						
Sb†	31.6	0.0119	mg/L	0.00490	0.0119	mg/L	0.00490 41.17%
	QC value within limits for Sb Recovery = Not calculated						
Se†	-345.7	-0.0713	mg/L	0.00753	-0.0713	mg/L	0.00753 10.55%
	QC value within limits for Se Recovery = Not calculated						
Tl†	33.2	0.0128	mg/L	0.00129	0.0128	mg/L	0.00129 10.12%
	QC value within limits for Tl Recovery = Not calculated						
V†	41031.1	0.242	mg/L	0.0009	0.242	mg/L	0.0009 0.35%
	QC value within limits for V Recovery = 96.79%						
Zn†	25734.6	0.530	mg/L	0.0058	0.530	mg/L	0.0058 1.09%
	QC value within limits for Zn Recovery = 106.09%						
QC Failed. Retry.							

Sequence No.: 11
Sample ID: ICSAB
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 11
Date Collected: 12/3/2007 10:14:11 PM
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte	Back Pressure	Flow
All	246.0 kPa	0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	390098.0	88.3 %	1.01			1.15%
Yr	222050.7	91.3 %	0.27			0.29%
Ag†	134351.9	0.485 mg/L	0.0056	0.485 mg/L	0.0056	1.15%
	QC value within limits for Ag Recovery = 96.97%					
Al†	1111937.6	249 mg/L	0.3	249 mg/L	0.3	0.12%
	QC value within limits for Al Recovery = 99.43%					
As†	-440.1	-0.151 mg/L	0.0062	-0.151 mg/L	0.0062	4.09%
	QC value less than the lower limit for As Recovery = Not calculated					
B_†	-65.5	-0.00260 mg/L	0.000429	-0.00260 mg/L	0.000429	16.50%
	QC value within limits for B_ Recovery = Not calculated					
Ba†	18605.1	0.258 mg/L	0.0035	0.258 mg/L	0.0035	1.37%
	QC value within limits for Ba Recovery = 103.03%					
Be†	733029.8	0.242 mg/L	0.0001	0.242 mg/L	0.0001	0.05%
	QC value within limits for Be Recovery = 96.65%					
Ca†	1797029.0	253 mg/L	0.4	253 mg/L	0.4	0.15%
	QC value within limits for Ca Recovery = 101.40%					
Co†	6159.8	0.237 mg/L	0.0027	0.237 mg/L	0.0027	1.16%
	QC value within limits for Co Recovery = 94.70%					
Cr†	19133.1	0.244 mg/L	0.0032	0.244 mg/L	0.0032	1.30%
	QC value within limits for Cr Recovery = 97.57%					
Cu†	90629.1	0.242 mg/L	0.0014	0.242 mg/L	0.0014	0.57%
	QC value within limits for Cu Recovery = 96.64%					
Fe†	103722.7	99.1 mg/L	0.14	99.1 mg/L	0.14	0.14%
	QC value within limits for Fe Recovery = 99.12%					
K†	188.7	0.157 mg/L	0.0061	0.157 mg/L	0.0061	3.89%
	QC value within limits for K Recovery = Not calculated					
Mg†	733370.3	244 mg/L	0.1	244 mg/L	0.1	0.05%
	QC value within limits for Mg Recovery = 97.59%					
Mn†	143792.5	0.257 mg/L	0.0004	0.257 mg/L	0.0004	0.15%
	QC value within limits for Mn Recovery = 102.90%					
Mo†	16.8	0.00131 mg/L	0.000209	0.00131 mg/L	0.000209	15.99%
	QC value within limits for Mo Recovery = Not calculated					
Na†	434.2	0.124 mg/L	0.0111	0.124 mg/L	0.0111	8.96%
	QC value within limits for Na Recovery = Not calculated					
Ni†	10253.3	0.461 mg/L	0.0058	0.461 mg/L	0.0058	1.25%
	QC value within limits for Ni Recovery = 92.26%					
Pb†	2222.3	0.460 mg/L	0.0083	0.460 mg/L	0.0083	1.81%
	QC value within limits for Pb Recovery = 92.00%					
Sb†	34.9	0.0136 mg/L	0.00004	0.0136 mg/L	0.00004	0.32%
	QC value within limits for Sb Recovery = Not calculated					
Se†	-334.7	-0.0630 mg/L	0.00073	-0.0630 mg/L	0.00073	1.15%
	QC value within limits for Se Recovery = Not calculated					
Tl†	47.1	0.0178 mg/L	0.00392	0.0178 mg/L	0.00392	21.97%
	QC value within limits for Tl Recovery = Not calculated					
V†	41057.5	0.242 mg/L	0.0001	0.242 mg/L	0.0001	0.03%
	QC value within limits for V Recovery = 96.84%					
Zn†	25546.4	0.527 mg/L	0.0062	0.527 mg/L	0.0062	1.17%
	QC value within limits for Zn Recovery = 105.31%					
QC Failed. Continue with analysis.						

Sequence No.: 12
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 12/3/2007 10:18:14 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	245.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	410396.6	92.9 %		0.44			0.47%
Yr	223987.4	92.1 %		0.40			0.44%
Ag†	274716.9	0.991 mg/L		0.0021	0.991 mg/L	0.0021	0.21%
	QC value within limits for Ag Recovery = 99.14%						
Al†	22928.6	5.13 mg/L		0.013	5.13 mg/L	0.013	0.26%
	QC value within limits for Al Recovery = 102.51%						
As†	10107.0	4.85 mg/L		0.021	4.85 mg/L	0.021	0.43%
	QC value within limits for As Recovery = 96.96%						
B_†	76184.3	2.46 mg/L		0.012	2.46 mg/L	0.012	0.50%
	QC value within limits for B_ Recovery = 98.44%						
Ba†	368654.3	5.10 mg/L		0.022	5.10 mg/L	0.022	0.44%
	QC value within limits for Ba Recovery = 102.07%						
Be†	6194123.8	2.04 mg/L		0.029	2.04 mg/L	0.029	1.41%
	QC value within limits for Be Recovery = 102.09%						
Ca†	362533.8	51.1 mg/L		0.05	51.1 mg/L	0.05	0.09%
	QC value within limits for Ca Recovery = 102.28%						
Co†	134012.0	5.15 mg/L		0.024	5.15 mg/L	0.024	0.46%
	QC value within limits for Co Recovery = 103.01%						
Cr†	396083.0	5.05 mg/L		0.020	5.05 mg/L	0.020	0.39%
	QC value within limits for Cr Recovery = 100.99%						
Cu†	1898859.9	5.06 mg/L		0.013	5.06 mg/L	0.013	0.25%
	QC value within limits for Cu Recovery = 101.24%						
Fe†	5476.1	5.23 mg/L		0.081	5.23 mg/L	0.081	1.55%
	QC value within limits for Fe Recovery = 104.66%						
K†	58827.6	48.8 mg/L		0.32	48.8 mg/L	0.32	0.66%
	QC value within limits for K Recovery = 97.69%						
Mg†	155115.1	51.6 mg/L		0.04	51.6 mg/L	0.04	0.07%
	QC value within limits for Mg Recovery = 103.20%						
Mn†	2881736.1	5.16 mg/L		0.022	5.16 mg/L	0.022	0.42%
	QC value within limits for Mn Recovery = 103.11%						
Mo†	64608.1	5.03 mg/L		0.031	5.03 mg/L	0.031	0.62%
	QC value within limits for Mo Recovery = 100.68%						
Na†	177729.2	50.6 mg/L		0.14	50.6 mg/L	0.14	0.27%
	QC value within limits for Na Recovery = 101.13%						
Ni†	114571.3	5.15 mg/L		0.021	5.15 mg/L	0.021	0.41%
	QC value within limits for Ni Recovery = 103.09%						
Pb†	24836.9	5.14 mg/L		0.012	5.14 mg/L	0.012	0.22%
	QC value within limits for Pb Recovery = 102.82%						
Sb†	9687.6	4.78 mg/L		0.017	4.78 mg/L	0.017	0.36%
	QC value within limits for Sb Recovery = 95.57%						
Se†	6575.2	5.04 mg/L		0.013	5.04 mg/L	0.013	0.27%
	QC value within limits for Se Recovery = 100.73%						
Tl†	14483.1	5.28 mg/L		0.005	5.28 mg/L	0.005	0.09%
	QC value within limits for Tl Recovery = 105.60%						
V†	856743.2	5.05 mg/L		0.016	5.05 mg/L	0.016	0.31%
	QC value within limits for V Recovery = 101.04%						
Zn†	249785.4	5.14 mg/L		0.021	5.14 mg/L	0.021	0.41%
	QC value within limits for Zn Recovery = 102.88%						
All analyte(s) passed QC.							

Sequence No.: 13
 Sample ID: ICB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 12/3/2007 10:22:31 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
 All 246.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	433996.3	98.2 %		0.10			0.10%
Yr	236534.7	97.2 %		0.68			0.70%
Ag†	-27.6	-0.00010 mg/L		0.000080	-0.00010 mg/L	0.000080	80.48%
	QC value within limits for Ag	Recovery = Not calculated					
Al†	31.4	0.00702 mg/L		0.000632	0.00702 mg/L	0.000632	8.99%
	QC value within limits for Al	Recovery = Not calculated					
As†	30.4	0.0146 mg/L		0.00221	0.0146 mg/L	0.00221	15.16%
	QC value within limits for As	Recovery = Not calculated					
B_†	501.1	0.0163 mg/L		0.00097	0.0163 mg/L	0.00097	5.95%
	QC value within limits for B_	Recovery = Not calculated					
Ba†	6.7	0.00009 mg/L		0.000133	0.00009 mg/L	0.000133	144.61%
	QC value within limits for Ba	Recovery = Not calculated					
Be†	12.0	0.00000 mg/L		0.000043	0.00000 mg/L	0.000043	>999.9%
	QC value within limits for Be	Recovery = Not calculated					
Ca†	-82.4	-0.0116 mg/L		0.00110	-0.0116 mg/L	0.00110	9.43%
	QC value within limits for Ca	Recovery = Not calculated					
Co†	-0.1	0.00000 mg/L		0.000168	0.00000 mg/L	0.000168	>999.9%
	QC value within limits for Co	Recovery = Not calculated					
Cr†	-8.1	-0.00010 mg/L		0.000039	-0.00010 mg/L	0.000039	37.49%
	QC value within limits for Cr	Recovery = Not calculated					
Cu†	200.7	0.00053 mg/L		0.000093	0.00053 mg/L	0.000093	17.38%
	QC value within limits for Cu	Recovery = Not calculated					
Fe†	3.4	0.00328 mg/L		0.000553	0.00328 mg/L	0.000553	16.88%
	QC value within limits for Fe	Recovery = Not calculated					
K†	79.8	0.0663 mg/L		0.01896	0.0663 mg/L	0.01896	28.61%
	QC value within limits for K	Recovery = Not calculated					
Mg†	-2.6	-0.00085 mg/L		0.000797	-0.00085 mg/L	0.000797	93.54%
	QC value within limits for Mg	Recovery = Not calculated					
Mn†	20.6	0.00004 mg/L		0.000001	0.00004 mg/L	0.000001	1.73%
	QC value within limits for Mn	Recovery = Not calculated					
Mo†	61.3	0.00478 mg/L		0.000321	0.00478 mg/L	0.000321	6.71%
	QC value within limits for Mo	Recovery = Not calculated					
Na†	41.6	0.0118 mg/L		0.00193	0.0118 mg/L	0.00193	16.33%
	QC value within limits for Na	Recovery = Not calculated					
Ni†	1.4	0.00007 mg/L		0.000039	0.00007 mg/L	0.000039	59.28%
	QC value within limits for Ni	Recovery = Not calculated					
Pb†	26.8	0.00555 mg/L		0.000432	0.00555 mg/L	0.000432	7.77%
	QC value within limits for Pb	Recovery = Not calculated					
Sb†	25.1	0.0126 mg/L		0.00154	0.0126 mg/L	0.00154	12.20%
	QC value within limits for Sb	Recovery = Not calculated					
Se†	-1.5	-0.00116 mg/L		0.002258	-0.00116 mg/L	0.002258	194.56%
	QC value within limits for Se	Recovery = Not calculated					
Tl†	8.3	0.00301 mg/L		0.000601	0.00301 mg/L	0.000601	19.96%
	QC value within limits for Tl	Recovery = Not calculated					
V†	7.3	0.00004 mg/L		0.000167	0.00004 mg/L	0.000167	394.50%
	QC value within limits for V	Recovery = Not calculated					
Zn†	-55.6	-0.00115 mg/L		0.000039	-0.00115 mg/L	0.000039	3.41%
	QC value within limits for Zn	Recovery = Not calculated					
All analyte(s) passed QC.							

Sequence No.: 29
 Sample ID: MBLANK2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 57
 Date Collected: 12/3/2007 11:29:17 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MBLANK2007

Analyte Back Pressure Flow
 All 249.0 kPa 0.65 L/min

Mean Data: MBLANK2007

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	419829.8	95.0 %	%	0.54			0.57%
Yr	226145.2	92.9 %	%	0.28			0.30%
Ag†	-24.8	-0.00009 mg/L	mg/L	0.000217	-0.00009 mg/L	0.000217	242.59%
Al†	-32.9	-0.00735 mg/L	mg/L	0.004817	-0.00735 mg/L	0.004817	65.58%
As†	7.0	0.00338 mg/L	mg/L	0.000143	0.00338 mg/L	0.000143	4.23%
B†	305.3	0.00991 mg/L	mg/L	0.000797	0.00991 mg/L	0.000797	8.04%
Ba†	3.1	0.00004 mg/L	mg/L	0.000097	0.00004 mg/L	0.000097	223.21%
Be†	-93.6	-0.00003 mg/L	mg/L	0.000011	-0.00003 mg/L	0.000011	36.02%
Ca†	-41.6	-0.00587 mg/L	mg/L	0.002126	-0.00587 mg/L	0.002126	36.25%
Co†	-2.9	-0.00011 mg/L	mg/L	0.000064	-0.00011 mg/L	0.000064	56.92%
Cr†	-11.6	-0.00015 mg/L	mg/L	0.000067	-0.00015 mg/L	0.000067	45.34%
Cu†	65.1	0.00017 mg/L	mg/L	0.000008	0.00017 mg/L	0.000008	4.85%
Fe†	1.9	0.00186 mg/L	mg/L	0.003087	0.00186 mg/L	0.003087	166.26%
K†	62.7	0.0520 mg/L	mg/L	0.03812	0.0520 mg/L	0.03812	73.26%
Mg†	3.8	0.00126 mg/L	mg/L	0.001107	0.00126 mg/L	0.001107	87.98%
Mn†	46.8	0.00008 mg/L	mg/L	0.000002	0.00008 mg/L	0.000002	1.86%
Mo†	9.9	0.00077 mg/L	mg/L	0.000140	0.00077 mg/L	0.000140	18.15%
Na†	689.1	0.196 mg/L	mg/L	0.0111	0.196 mg/L	0.0111	5.66%
Ni†	-0.7	-0.00003 mg/L	mg/L	0.000024	-0.00003 mg/L	0.000024	76.95%
Pb†	7.5	0.00155 mg/L	mg/L	0.000921	0.00155 mg/L	0.000921	59.28%
Sb†	7.1	0.00357 mg/L	mg/L	0.000504	0.00357 mg/L	0.000504	14.11%
Se†	6.1	0.00463 mg/L	mg/L	0.000019	0.00463 mg/L	0.000019	0.42%
Tl†	3.8	0.00138 mg/L	mg/L	0.000742	0.00138 mg/L	0.000742	53.83%
V†	-16.9	-0.00010 mg/L	mg/L	0.000151	-0.00010 mg/L	0.000151	150.89%
Zn†	-141.4	-0.00293 mg/L	mg/L	0.000093	-0.00293 mg/L	0.000093	3.17%

Sequence No.: 30
 Sample ID: MRL
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 21
 Date Collected: 12/3/2007 11:33:22 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte	Back Pressure	Flow
All	249.0 kPa	0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units			Conc. Units	Units		
Sca	437985.4	99.1 %	%	0.70				0.70%
Yr	237596.0	97.6 %	%	1.01				1.03%
Ag†	2780.4	0.0100 mg/L	mg/L	0.00020	0.0100 mg/L	0.00020		1.97%
Al†	216.2	0.0483 mg/L	mg/L	0.01032	0.0483 mg/L	0.01032		21.35%
As†	199.4	0.0957 mg/L	mg/L	0.00043	0.0957 mg/L	0.00043		0.45%
B†	1701.6	0.0551 mg/L	mg/L	0.00041	0.0551 mg/L	0.00041		0.74%
Ba†	1530.3	0.0212 mg/L	mg/L	0.00003	0.0212 mg/L	0.00003		0.14%
Be†	3135.6	0.00103 mg/L	mg/L	0.000026	0.00103 mg/L	0.000026		2.56%
Cat	7403.1	1.04 mg/L	mg/L	0.001	1.04 mg/L	0.001		0.13%
Co†	1402.4	0.0539 mg/L	mg/L	0.00033	0.0539 mg/L	0.00033		0.61%
Cr†	805.8	0.0103 mg/L	mg/L	0.00006	0.0103 mg/L	0.00006		0.62%
Cu†	3957.8	0.0106 mg/L	mg/L	0.00009	0.0106 mg/L	0.00009		0.88%
Fe†	25.2	0.0241 mg/L	mg/L	0.00100	0.0241 mg/L	0.00100		4.17%
K†	1270.6	1.05 mg/L	mg/L	0.013	1.05 mg/L	0.013		1.24%
Mg†	330.9	0.110 mg/L	mg/L	0.0003	0.110 mg/L	0.0003		0.28%
Mn†	1313.7	0.00235 mg/L	mg/L	0.000025	0.00235 mg/L	0.000025		1.08%
Mo†	265.6	0.0207 mg/L	mg/L	0.00040	0.0207 mg/L	0.00040		1.91%
Nat	4029.8	1.15 mg/L	mg/L	0.008	1.15 mg/L	0.008		0.72%
Ni†	479.9	0.0216 mg/L	mg/L	0.00003	0.0216 mg/L	0.00003		0.13%
Pb†	111.2	0.0230 mg/L	mg/L	0.00033	0.0230 mg/L	0.00033		1.43%
Sb†	91.0	0.0454 mg/L	mg/L	0.00047	0.0454 mg/L	0.00047		1.04%
Se†	130.6	0.0999 mg/L	mg/L	0.00259	0.0999 mg/L	0.00259		2.60%
Tl†	323.5	0.118 mg/L	mg/L	0.0011	0.118 mg/L	0.0011		0.95%
V†	311.2	0.00188 mg/L	mg/L	0.000247	0.00188 mg/L	0.000247		13.11%
Zn†	824.3	0.0169 mg/L	mg/L	0.00016	0.0169 mg/L	0.00016		0.95%

Sequence No.: 31
 Sample ID: MRL2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 58
 Date Collected: 12/3/2007 11:37:26 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL2007

Analyte	Back Pressure	Flow
All	249.0 kPa	0.65 L/min

Mean Data: MRL2007

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units			Conc. Units	Units		
Sca	431024.3	97.6 %		0.99				1.02%
Yr	233444.6	95.9 %		1.43				1.49%
B†	1584.5	0.0513 mg/L		0.00032	0.0513 mg/L		0.00032	0.62%
Ba†	1483.3	0.0205 mg/L		0.00017	0.0205 mg/L		0.00017	0.84%
Be†	2948.2	0.00097 mg/L		0.000013	0.00097 mg/L		0.000013	1.34%
Ca†	7239.0	1.02 mg/L		0.009	1.02 mg/L		0.009	0.85%
Co†	1382.2	0.0531 mg/L		0.00043	0.0531 mg/L		0.00043	0.81%
Cr†	758.8	0.00967 mg/L		0.000066	0.00967 mg/L		0.000066	0.68%
Cu†	4050.4	0.0108 mg/L		0.00005	0.0108 mg/L		0.00005	0.42%
Fe†	25.8	0.0247 mg/L		0.00085	0.0247 mg/L		0.00085	3.46%
K†	1246.1	1.03 mg/L		0.022	1.03 mg/L		0.022	2.15%
Mg†	319.8	0.106 mg/L		0.0027	0.106 mg/L		0.0027	2.50%
Mn†	1231.7	0.00220 mg/L		0.000015	0.00220 mg/L		0.000015	0.68%
Mo†	254.3	0.0198 mg/L		0.00030	0.0198 mg/L		0.00030	1.52%
Na†	3852.5	1.10 mg/L		0.016	1.10 mg/L		0.016	1.46%
Ni†	467.1	0.0210 mg/L		0.00008	0.0210 mg/L		0.00008	0.39%
Pb†	109.9	0.0227 mg/L		0.00107	0.0227 mg/L		0.00107	4.72%
V†	294.7	0.00178 mg/L		0.000034	0.00178 mg/L		0.000034	1.90%
Zn†	826.1	0.0170 mg/L		0.00007	0.0170 mg/L		0.00007	0.42%

Sequence No.: 32
 Sample ID: LCS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 59
 Date Collected: 12/3/2007 11:41:30 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS

Analyte	Back Pressure	Flow
All	249.0 kPa	0.65 L/min

Mean Data: LCS

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	411418.3	93.1 %	%	0.38			0.41%
Yr	224516.3	92.3 %	%	0.69			0.75%
B _f	14432.5	0.466 mg/L	mg/L	0.0041	0.466 mg/L	0.0041	0.87%
Ba _f	70362.9	0.974 mg/L	mg/L	0.0040	0.974 mg/L	0.0040	0.41%
Be _f	149599.9	0.0493 mg/L	mg/L	0.00038	0.0493 mg/L	0.00038	0.78%
Ca _f	357841.0	50.5 mg/L	mg/L	0.12	50.5 mg/L	0.12	0.23%
Co _f	25795.5	0.991 mg/L	mg/L	0.0049	0.991 mg/L	0.0049	0.50%
Cr _f	76245.3	0.972 mg/L	mg/L	0.0070	0.972 mg/L	0.0070	0.72%
Cu _f	371299.9	0.990 mg/L	mg/L	0.0051	0.990 mg/L	0.0051	0.51%
Fe _f	5246.0	5.01 mg/L	mg/L	0.136	5.01 mg/L	0.136	2.71%
K _f	23694.6	19.7 mg/L	mg/L	0.59	19.7 mg/L	0.59	2.98%
Mg _f	61736.1	20.5 mg/L	mg/L	0.44	20.5 mg/L	0.44	2.12%
Mn _f	279927.5	0.501 mg/L	mg/L	0.0025	0.501 mg/L	0.0025	0.50%
Mo _f	12272.7	0.956 mg/L	mg/L	0.0040	0.956 mg/L	0.0040	0.42%
Na _f	178275.0	50.7 mg/L	mg/L	0.01	50.7 mg/L	0.01	0.02%
Ni _f	11029.8	0.496 mg/L	mg/L	0.0039	0.496 mg/L	0.0039	0.79%
Pb _f	4760.9	0.985 mg/L	mg/L	0.0086	0.985 mg/L	0.0086	0.87%
V _f	164849.7	0.972 mg/L	mg/L	0.0025	0.972 mg/L	0.0025	0.26%
Zn _f	48225.4	0.997 mg/L	mg/L	0.0048	0.997 mg/L	0.0048	0.48%

Sequence No.: 33
 Sample ID: LCSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 60
 Date Collected: 12/3/2007 11:45:02 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD

Analyte	Back Pressure	Flow
All	249.0 kPa	0.65 L/min

Mean Data: LCSD

Analyte	Mean Corrected			Sample			RSD
	Intensity	Conc. Units	Calib	Std.Dev.	Conc. Units	Std.Dev.	
Sca	406189.8	92.0 %	%	0.08			0.09%
Yr	221032.2	90.8 %	%	0.38			0.41%
B†	14659.9	0.474 mg/L	mg/L	0.0060	0.474 mg/L	0.0060	1.26%
Ba†	71208.9	0.986 mg/L	mg/L	0.0087	0.986 mg/L	0.0087	0.89%
Be†	149869.1	0.0494 mg/L	mg/L	0.00007	0.0494 mg/L	0.00007	0.14%
Ca†	358679.2	50.6 mg/L	mg/L	0.02	50.6 mg/L	0.02	0.05%
Co†	26127.9	1.00 mg/L	mg/L	0.008	1.00 mg/L	0.008	0.82%
Cr†	77276.6	0.985 mg/L	mg/L	0.0068	0.985 mg/L	0.0068	0.69%
Cu†	372733.2	0.994 mg/L	mg/L	0.0056	0.994 mg/L	0.0056	0.56%
Fe†	5284.9	5.05 mg/L	mg/L	0.154	5.05 mg/L	0.154	3.05%
K†	23892.4	19.8 mg/L	mg/L	0.42	19.8 mg/L	0.42	2.10%
Mg†	62039.5	20.6 mg/L	mg/L	0.46	20.6 mg/L	0.46	2.24%
Mn†	281962.4	0.504 mg/L	mg/L	0.0013	0.504 mg/L	0.0013	0.27%
Mo†	12424.8	0.968 mg/L	mg/L	0.0121	0.968 mg/L	0.0121	1.25%
Na†	178646.6	50.8 mg/L	mg/L	0.03	50.8 mg/L	0.03	0.05%
Ni†	11151.6	0.502 mg/L	mg/L	0.0007	0.502 mg/L	0.0007	0.13%
Pb†	4825.1	0.999 mg/L	mg/L	0.0082	0.999 mg/L	0.0082	0.82%
V†	165458.7	0.976 mg/L	mg/L	0.0017	0.976 mg/L	0.0017	0.18%
Zn†	48919.9	1.01 mg/L	mg/L	0.008	1.01 mg/L	0.008	0.83%

Sequence No.: 34
 Sample ID: 2711190059
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 61
 Date Collected: 12/3/2007 11:48:35 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711190059

Analyte Back Pressure Flow
 All 249.0 kPa 0.65 L/min

Mean Data: 2711190059

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	394814.0	89.4 %		0.21			0.23%
Yr	215805.8	88.7 %		0.34			0.38%
B+	13230.3	0.429 mg/L		0.0035	0.429 mg/L	0.0035	0.81%
Ba+	4993.6	0.0691 mg/L		0.00026	0.0691 mg/L	0.00026	0.38%
Be+	-611.5	-0.00020 mg/L		0.000026	-0.00020 mg/L	0.000026	12.67%
Ca+	363948.1	51.3 mg/L		0.35	51.3 mg/L	0.35	0.69%
Co+	38.3	0.00147 mg/L		0.000150	0.00147 mg/L	0.000150	10.21%
Cr+	28.9	0.00037 mg/L		0.000152	0.00037 mg/L	0.000152	41.27%
Cu+	656.4	0.00175 mg/L		0.000161	0.00175 mg/L	0.000161	9.20%
Fe+	136.4	0.130 mg/L		0.0020	0.130 mg/L	0.0020	1.53%
K+	7350.1	6.10 mg/L		0.006	6.10 mg/L	0.006	0.09%
Mg+	63839.9	21.2 mg/L		0.09	21.2 mg/L	0.09	0.42%
Mn+	3175.9	0.00568 mg/L		0.000008	0.00568 mg/L	0.000008	0.13%
Mo+	51.2	0.00399 mg/L		0.000014	0.00399 mg/L	0.000014	0.35%
Na+	619762.4	176 mg/L		1.1	176 mg/L	1.1	0.65%
Ni+	80.1	0.00360 mg/L		0.000466	0.00360 mg/L	0.000466	12.93%
Pb+	6.5	0.00134 mg/L		0.000644	0.00134 mg/L	0.000644	48.00%
V+	1627.0	0.00954 mg/L		0.000207	0.00954 mg/L	0.000207	2.17%
Zn+	506.1	0.0105 mg/L		0.00008	0.0105 mg/L	0.00008	0.74%

Sequence No.: 35
 Sample ID: 2711190059MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 62
 Date Collected: 12/3/2007 11:53:25 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711190059MS

Analyte Back Pressure Flow
 All 250.0 kPa 0.65 L/min

Mean Data: 2711190059MS

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	384700.7		87.1 %	0.20			0.23%
Yr	210312.0		86.4 %	0.18			0.21%
B_t	28934.4		0.936 mg/L	0.0017	0.936 mg/L	0.0017	0.18%
Ba_t	81413.2		1.13 mg/L	0.003	1.13 mg/L	0.003	0.24%
Be_t	160596.0		0.0529 mg/L	0.00007	0.0529 mg/L	0.00007	0.13%
Ca_t	729053.8		103 mg/L	0.0	103 mg/L	0.0	0.04%
Co_t	27732.2		1.07 mg/L	0.005	1.07 mg/L	0.005	0.47%
Cr_t	82431.8		1.05 mg/L	0.005	1.05 mg/L	0.005	0.48%
Cu_t	401540.5		1.07 mg/L	0.003	1.07 mg/L	0.003	0.24%
Fe_t	5820.9		5.56 mg/L	0.048	5.56 mg/L	0.048	0.86%
K_t	32828.6		27.3 mg/L	0.19	27.3 mg/L	0.19	0.68%
Mg_t	127329.1		42.4 mg/L	0.06	42.4 mg/L	0.06	0.14%
Mn_t	305247.8		0.546 mg/L	0.0013	0.546 mg/L	0.0013	0.24%
Mo_t	13369.7		1.04 mg/L	0.005	1.04 mg/L	0.005	0.47%
Na_t	789698.4		225 mg/L	1.2	225 mg/L	1.2	0.56%
Ni_t	11764.8		0.529 mg/L	0.0039	0.529 mg/L	0.0039	0.73%
Pb_t	5041.5		1.04 mg/L	0.003	1.04 mg/L	0.003	0.27%
V_t	178349.5		1.05 mg/L	0.001	1.05 mg/L	0.001	0.12%
Zn_t	53435.1		1.10 mg/L	0.006	1.10 mg/L	0.006	0.56%

Sequence No.: 36
 Sample ID: 2711190059MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 63
 Date Collected: 12/3/2007 11:56:57 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711190059MSD
 Analyte Back Pressure Flow
 All 250.0 kPa 0.65 L/min

Mean Data: 2711190059MSD

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Conc. Units	Sample	Std.Dev.	RSD
Sca	403164.2	91.3 %		0.59				0.64%
Yr	223132.5	91.7 %		1.06				1.16%
B†	27630.3	0.894 mg/L		0.0085	0.894 mg/L		0.0085	0.95%
Ba†	76267.0	1.06 mg/L		0.006	1.06 mg/L		0.006	0.53%
Be†	152823.2	0.0504 mg/L		0.00010	0.0504 mg/L		0.00010	0.20%
Ca†	692816.4	97.7 mg/L		0.13	97.7 mg/L		0.13	0.13%
Co†	25977.6	0.998 mg/L		0.0011	0.998 mg/L		0.0011	0.11%
Cr†	77105.6	0.983 mg/L		0.0055	0.983 mg/L		0.0055	0.56%
Cu†	382557.7	1.02 mg/L		0.001	1.02 mg/L		0.001	0.10%
Fe†	5547.2	5.30 mg/L		0.122	5.30 mg/L		0.122	2.30%
K†	31049.4	25.8 mg/L		0.65	25.8 mg/L		0.65	2.51%
Mg†	120674.3	40.1 mg/L		0.02	40.1 mg/L		0.02	0.05%
Mn†	289861.5	0.519 mg/L		0.0015	0.519 mg/L		0.0015	0.28%
Mo†	12516.3	0.975 mg/L		0.0073	0.975 mg/L		0.0073	0.74%
Na†	759491.0	216 mg/L		0.0	216 mg/L		0.0	0.02%
Ni†	10985.6	0.494 mg/L		0.0024	0.494 mg/L		0.0024	0.49%
Pb†	4747.9	0.983 mg/L		0.0060	0.983 mg/L		0.0060	0.62%
V†	169313.3	0.998 mg/L		0.0001	0.998 mg/L		0.0001	0.01%
Zn†	50047.9	1.03 mg/L		0.008	1.03 mg/L		0.008	0.76%

Sequence No.: 37
Sample ID: CCV
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 4
Date Collected: 12/4/2007 12:00:30 AM
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
All 250.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	407007.7	92.1 %	0.27			0.30%
Yr	225074.6	92.5 %	0.26			0.29%
B_†	75446.2	2.44 mg/L	0.021	2.44 mg/L	0.021	0.88%
	QC value within limits for B_ Recovery = 97.49%					
Baf	363551.7	5.03 mg/L	0.045	5.03 mg/L	0.045	0.89%
	QC value within limits for Ba Recovery = 100.66%					
Bef	6093680.6	2.01 mg/L	0.008	2.01 mg/L	0.008	0.38%
	QC value within limits for Be Recovery = 100.44%					
Ca†	359991.5	50.8 mg/L	0.02	50.8 mg/L	0.02	0.04%
	QC value within limits for Ca Recovery = 101.56%					
Co†	131982.4	5.07 mg/L	0.044	5.07 mg/L	0.044	0.87%
	QC value within limits for Co Recovery = 101.45%					
Cr†	393287.2	5.01 mg/L	0.029	5.01 mg/L	0.029	0.58%
	QC value within limits for Cr Recovery = 100.28%					
Cu†	1890600.0	5.04 mg/L	0.008	5.04 mg/L	0.008	0.16%
	QC value within limits for Cu Recovery = 100.80%					
Fe†	5383.3	5.14 mg/L	0.020	5.14 mg/L	0.020	0.40%
	QC value within limits for Fe Recovery = 102.89%					
K†	59907.8	49.7 mg/L	0.05	49.7 mg/L	0.05	0.10%
	QC value within limits for K Recovery = 99.48%					
Mg†	155086.5	51.6 mg/L	0.03	51.6 mg/L	0.03	0.06%
	QC value within limits for Mg Recovery = 103.18%					
Mn†	2849251.7	5.10 mg/L	0.007	5.10 mg/L	0.007	0.14%
	QC value within limits for Mn Recovery = 101.95%					
Mo†	63617.7	4.96 mg/L	0.048	4.96 mg/L	0.048	0.97%
	QC value within limits for Mo Recovery = 99.13%					
Na†	179033.1	50.9 mg/L	0.16	50.9 mg/L	0.16	0.31%
	QC value within limits for Na Recovery = 101.87%					
Ni†	112575.1	5.06 mg/L	0.038	5.06 mg/L	0.038	0.75%
	QC value within limits for Ni Recovery = 101.29%					
Pb†	24667.9	5.11 mg/L	0.043	5.11 mg/L	0.043	0.85%
	QC value within limits for Pb Recovery = 102.12%					
V†	845328.5	4.98 mg/L	0.007	4.98 mg/L	0.007	0.13%
	QC value within limits for V Recovery = 99.69%					
Zn†	246171.7	5.07 mg/L	0.041	5.07 mg/L	0.041	0.81%
	QC value within limits for Zn Recovery = 101.39%					

All analyte(s) passed QC.

Sequence No.: 38
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 12/4/2007 12:04:09 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 250.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	429678.4	97.3 %		0.46			0.47%
Yr	234706.1	96.5 %		0.30			0.31%
B_i	399.8	0.0130 mg/L		0.00082	0.0130 mg/L	0.00082	6.32%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	12.8	0.00018 mg/L		0.000037	0.00018 mg/L	0.000037	20.93%
	QC value within limits for Ba Recovery = Not calculated						
Be†	-125.2	-0.00004 mg/L		0.000005	-0.00004 mg/L	0.000005	11.28%
	QC value within limits for Be Recovery = Not calculated						
Ca†	-105.9	-0.0149 mg/L		0.00162	-0.0149 mg/L	0.00162	10.81%
	QC value within limits for Ca Recovery = Not calculated						
Co†	4.9	0.00019 mg/L		0.000379	0.00019 mg/L	0.000379	203.28%
	QC value within limits for Co Recovery = Not calculated						
Cr†	-16.1	-0.00021 mg/L		0.000125	-0.00021 mg/L	0.000125	60.81%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	17.3	0.00005 mg/L		0.000110	0.00005 mg/L	0.000110	236.95%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-0.8	-0.00072 mg/L		0.001081	-0.00072 mg/L	0.001081	150.43%
	QC value within limits for Fe Recovery = Not calculated						
K†	79.6	0.0661 mg/L		0.03521	0.0661 mg/L	0.03521	53.30%
	QC value within limits for K Recovery = Not calculated						
Mg†	0.5	0.00017 mg/L		0.001900	0.00017 mg/L	0.001900	>999.9%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	33.0	0.00006 mg/L		0.000014	0.00006 mg/L	0.000014	23.76%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	44.1	0.00343 mg/L		0.000352	0.00343 mg/L	0.000352	10.26%
	QC value within limits for Mo Recovery = Not calculated						
Na†	82.5	0.0235 mg/L		0.01088	0.0235 mg/L	0.01088	46.36%
	QC value within limits for Na Recovery = Not calculated						
Ni†	-0.9	-0.00004 mg/L		0.000021	-0.00004 mg/L	0.000021	51.02%
	QC value within limits for Ni Recovery = Not calculated						
Pb†	4.5	0.00093 mg/L		0.001246	0.00093 mg/L	0.001246	133.87%
	QC value within limits for Pb Recovery = Not calculated						
V†	-66.6	-0.00039 mg/L		0.000230	-0.00039 mg/L	0.000230	58.74%
	QC value within limits for V Recovery = Not calculated						
Zn†	-160.0	-0.00332 mg/L		0.000029	-0.00332 mg/L	0.000029	0.88%
	QC value within limits for Zn Recovery = Not calculated						
All analyte(s) passed QC.							

Sequence No.: 39
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 12/4/2007 12:08:00 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	249.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	392232.5	88.8 %		0.17			0.20%
Yr	208936.5	85.9 %		1.04			1.21%
B_†	37378.4	1.21 mg/L		0.004	1.21 mg/L	0.004	0.32%
	QC value within limits for B_	Recovery = 96.59%					
Ba†	183354.5	2.54 mg/L		0.002	2.54 mg/L	0.002	0.07%
	QC value within limits for Ba	Recovery = 101.54%					
Be†	3044901.8	1.00 mg/L		0.000	1.00 mg/L	0.000	0.00%
	QC value within limits for Be	Recovery = 100.37%					
Ca†	180916.2	25.5 mg/L		0.06	25.5 mg/L	0.06	0.24%
	QC value within limits for Ca	Recovery = 102.08%					
Co†	66977.8	2.57 mg/L		0.002	2.57 mg/L	0.002	0.07%
	QC value within limits for Co	Recovery = 102.97%					
Cr†	196088.9	2.50 mg/L		0.001	2.50 mg/L	0.001	0.04%
	QC value within limits for Cr	Recovery = 99.99%					
Cu†	942387.0	2.51 mg/L		0.004	2.51 mg/L	0.004	0.14%
	QC value within limits for Cu	Recovery = 100.49%					
Fe†	2703.2	2.58 mg/L		0.022	2.58 mg/L	0.022	0.85%
	QC value within limits for Fe	Recovery = 103.33%					
K†	29974.5	24.9 mg/L		0.11	24.9 mg/L	0.11	0.44%
	QC value within limits for K	Recovery = 99.55%					
Mg†	79057.2	26.3 mg/L		0.03	26.3 mg/L	0.03	0.10%
	QC value within limits for Mg	Recovery = 105.20%					
Mn†	1441308.1	2.58 mg/L		0.004	2.58 mg/L	0.004	0.15%
	QC value within limits for Mn	Recovery = 103.14%					
Mo†	32040.9	2.50 mg/L		0.007	2.50 mg/L	0.007	0.27%
	QC value within limits for Mo	Recovery = 99.86%					
Na†	90418.8	25.7 mg/L		0.15	25.7 mg/L	0.15	0.57%
	QC value within limits for Na	Recovery = 102.89%					
Ni†	57319.1	2.58 mg/L		0.007	2.58 mg/L	0.007	0.29%
	QC value within limits for Ni	Recovery = 103.15%					
Pb†	12494.8	2.59 mg/L		0.007	2.59 mg/L	0.007	0.27%
	QC value within limits for Pb	Recovery = 103.45%					
V†	420903.3	2.48 mg/L		0.001	2.48 mg/L	0.001	0.05%
	QC value within limits for V	Recovery = 99.28%					
Zn†	124360.6	2.56 mg/L		0.007	2.56 mg/L	0.007	0.28%
	QC value within limits for Zn	Recovery = 102.44%					

All analyte(s) passed QC.

Sequence No.: 40
 Sample ID: 2711200235_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 64
 Date Collected: 12/4/2007 12:12:13 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711200235_2X

Analyte Back Pressure Flow
 All 249.0 kPa 0.65 L/min

Mean Data: 2711200235_2X

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	409518.6	92.7 %		0.53			0.57%
Yr	216456.0	89.0 %		0.76			0.85%
B _f	2519.1	0.0817 mg/L		0.00017	0.163 mg/L	0.0003	0.20%
Ba _f	2392.1	0.0331 mg/L		0.00015	0.0662 mg/L	0.00031	0.46%
Be _f	-366.7	-0.00012 mg/L		0.000027	-0.00024 mg/L	0.000054	22.27%
Ca _f	73279.1	10.3 mg/L		0.03	20.7 mg/L	0.06	0.30%
Co _f	-4.7	-0.00018 mg/L		0.000179	-0.00036 mg/L	0.000358	98.72%
Cr _f	-3.2	-0.00004 mg/L		0.000124	-0.00008 mg/L	0.000249	306.05%
Cu _f	583.5	0.00155 mg/L		0.000007	0.00311 mg/L	0.000014	0.46%
Fe _f	43.1	0.0412 mg/L		0.00218	0.0824 mg/L	0.00436	5.29%
K _f	1951.4	1.62 mg/L		0.014	3.24 mg/L	0.029	0.89%
Mg _f	21424.9	7.13 mg/L		0.021	14.3 mg/L	0.04	0.29%
Mn _f	2337.4	0.00418 mg/L		0.000035	0.00836 mg/L	0.000069	0.83%
Mo _f	43.4	0.00338 mg/L		0.000198	0.00676 mg/L	0.000395	5.85%
Na _f	102896.3	29.3 mg/L		0.13	58.5 mg/L	0.26	0.44%
Ni _f	9.5	0.00043 mg/L		0.000005	0.00086 mg/L	0.000010	1.22%
Pb _f	2.5	0.00052 mg/L		0.000480	0.00103 mg/L	0.000960	93.14%
V _f	309.2	0.00181 mg/L		0.000010	0.00363 mg/L	0.000020	0.55%
Zn _f	-60.3	-0.00125 mg/L		0.000046	-0.00250 mg/L	0.000091	3.65%

```

=====
Sequence No.: 41                               Autosampler Location: 65
Sample ID: 2711200235_2XMS                    Date Collected: 12/4/2007 12:16:17 AM
Analyst:                                       Data Type: Original
Initial Sample Wt:                            Initial Sample Vol:
Dilution: 2X                                 Sample Prep Vol:
=====
    
```

```

-----
Nebulizer Parameters: 2711200235_2XMS
Analyte      Back Pressure      Flow
All          248.0 kPa           0.65 L/min
-----
    
```

```

-----
Mean Data: 2711200235_2XMS
-----

```

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity	Conc.			Conc.	Units	
Sca	396901.6	89.8 %	%	0.67			0.75%
Yr	213045.2	87.6 %	%	0.51			0.58%
B _f	9980.9	0.323	mg/L	0.0028	0.645	mg/L	0.0057 0.88%
Ba _f	39512.4	0.547	mg/L	0.0052	1.09	mg/L	0.010 0.95%
Be _f	77812.9	0.0257	mg/L	0.00009	0.0513	mg/L	0.00017 0.33%
Ca _f	260736.2	36.8	mg/L	0.03	73.6	mg/L	0.05 0.07%
Co _f	13507.3	0.519	mg/L	0.0041	1.04	mg/L	0.008 0.78%
Cr _f	39569.8	0.504	mg/L	0.0001	1.01	mg/L	0.000 0.01%
Cu _f	195657.4	0.522	mg/L	0.0021	1.04	mg/L	0.004 0.40%
Fe _f	2831.9	2.71	mg/L	0.021	5.41	mg/L	0.043 0.79%
K _f	14610.6	12.1	mg/L	0.13	24.3	mg/L	0.26 1.08%
Mg _f	54223.1	18.0	mg/L	0.00	36.1	mg/L	0.01 0.02%
Mn _f	149896.8	0.268	mg/L	0.0006	0.536	mg/L	0.0013 0.24%
Mo _f	6449.4	0.502	mg/L	0.0034	1.00	mg/L	0.007 0.69%
Na _f	195718.7	55.7	mg/L	0.11	111	mg/L	0.2 0.19%
Ni _f	5728.7	0.258	mg/L	0.0024	0.515	mg/L	0.0047 0.91%
Pb _f	2492.5	0.516	mg/L	0.0023	1.03	mg/L	0.005 0.44%
V _f	85878.5	0.506	mg/L	0.0013	1.01	mg/L	0.003 0.25%
Zn _f	25542.5	0.528	mg/L	0.0049	1.06	mg/L	0.010 0.92%

Sequence No.: 42
 Sample ID: 2711200235_2XMSD
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 66
 Date Collected: 12/4/2007 12:21:01 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711200235_2XMSD

Analyte	Back Pressure	Flow
All	248.0 kPa	0.65 L/min

Mean Data: 2711200235_2XMSD

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units			Conc. Units	Units		
Sca	408217.7	92.4 %	%	0.32				0.34%
Yr	222164.0	91.3 %	%	0.92				1.01%
B †	9938.5	0.321 mg/L	mg/L	0.0021	0.643 mg/L		0.0041	0.64%
Ba †	39104.4	0.541 mg/L	mg/L	0.0018	1.08 mg/L		0.004	0.32%
Be †	77368.6	0.0255 mg/L	mg/L	0.00003	0.0510 mg/L		0.00005	0.10%
Ca †	251080.8	35.4 mg/L	mg/L	0.12	70.8 mg/L		0.24	0.34%
Co †	13458.9	0.517 mg/L	mg/L	0.0023	1.03 mg/L		0.005	0.45%
Cr †	39416.4	0.503 mg/L	mg/L	0.0005	1.01 mg/L		0.001	0.11%
Cu †	195176.2	0.520 mg/L	mg/L	0.0001	1.04 mg/L		0.000	0.02%
Fe †	2802.7	2.68 mg/L	mg/L	0.015	5.36 mg/L		0.030	0.56%
K †	13934.1	11.6 mg/L	mg/L	0.06	23.1 mg/L		0.11	0.48%
Mg †	52124.1	17.3 mg/L	mg/L	0.09	34.7 mg/L		0.18	0.51%
Mn †	149162.4	0.267 mg/L	mg/L	0.0003	0.534 mg/L		0.0007	0.13%
Mo †	6418.2	0.500 mg/L	mg/L	0.0004	1.00 mg/L		0.001	0.09%
Na †	190303.0	54.1 mg/L	mg/L	0.24	108 mg/L		0.5	0.45%
Ni †	5703.5	0.257 mg/L	mg/L	0.0006	0.513 mg/L		0.0011	0.22%
Pb †	2470.7	0.511 mg/L	mg/L	0.0045	1.02 mg/L		0.009	0.89%
V †	85408.7	0.504 mg/L	mg/L	0.0002	1.01 mg/L		0.000	0.05%
Zn †	25193.8	0.521 mg/L	mg/L	0.0009	1.04 mg/L		0.002	0.18%

Sequence No.: 50
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 12/4/2007 12:54:16 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte Back Pressure Flow
 All 249.0 kPa 0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	380554.8	86.1 %		0.19			0.22%
Yr	204450.1	84.0 %		0.33			0.39%
B_†	76297.3	2.46 mg/L		0.022	2.46 mg/L	0.022	0.91%
	QC value within limits for B_ Recovery = 98.59%						
Ba†	368629.7	5.10 mg/L		0.043	5.10 mg/L	0.043	0.84%
	QC value within limits for Ba Recovery = 102.07%						
Be†	6143895.7	2.03 mg/L		0.001	2.03 mg/L	0.001	0.03%
	QC value within limits for Be Recovery = 101.26%						
Ca†	362543.1	51.1 mg/L		0.03	51.1 mg/L	0.03	0.05%
	QC value within limits for Ca Recovery = 102.28%						
Co†	134087.5	5.15 mg/L		0.039	5.15 mg/L	0.039	0.75%
	QC value within limits for Co Recovery = 103.07%						
Cr†	399031.5	5.09 mg/L		0.040	5.09 mg/L	0.040	0.79%
	QC value within limits for Cr Recovery = 101.74%						
Cu†	1911231.7	5.09 mg/L		0.002	5.09 mg/L	0.002	0.03%
	QC value within limits for Cu Recovery = 101.90%						
Fe†	5484.8	5.24 mg/L		0.013	5.24 mg/L	0.013	0.24%
	QC value within limits for Fe Recovery = 104.83%						
K†	61217.3	50.8 mg/L		0.02	50.8 mg/L	0.02	0.05%
	QC value within limits for K Recovery = 101.66%						
Mg†	158811.4	52.8 mg/L		0.04	52.8 mg/L	0.04	0.08%
	QC value within limits for Mg Recovery = 105.66%						
Mn†	2864439.1	5.12 mg/L		0.006	5.12 mg/L	0.006	0.12%
	QC value within limits for Mn Recovery = 102.49%						
Mo†	64438.7	5.02 mg/L		0.035	5.02 mg/L	0.035	0.69%
	QC value within limits for Mo Recovery = 100.41%						
Na†	183643.2	52.2 mg/L		0.14	52.2 mg/L	0.14	0.27%
	QC value within limits for Na Recovery = 104.49%						
Ni†	114167.4	5.14 mg/L		0.043	5.14 mg/L	0.043	0.84%
	QC value within limits for Ni Recovery = 102.72%						
Pb†	24876.9	5.15 mg/L		0.056	5.15 mg/L	0.056	1.09%
	QC value within limits for Pb Recovery = 102.98%						
V†	850178.2	5.01 mg/L		0.017	5.01 mg/L	0.017	0.34%
	QC value within limits for V Recovery = 100.27%						
Zn†	249714.3	5.14 mg/L		0.045	5.14 mg/L	0.045	0.88%
	QC value within limits for Zn Recovery = 102.85%						
All analyte(s) passed QC.							

Sequence No.: 51
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 12/4/2007 12:57:55 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 249.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	405839.9	91.9 %	0.83			0.90%
Yr	212174.8	87.2 %	0.85			0.97%
B _f	413.9	0.0134 mg/L	0.00109	0.0134 mg/L	0.00109	8.11%
	QC value within limits for B _f Recovery = Not calculated					
Ba _f	3.8	0.00005 mg/L	0.000027	0.00005 mg/L	0.000027	50.84%
	QC value within limits for Ba Recovery = Not calculated					
Be _f	-200.6	-0.00007 mg/L	0.000011	-0.00007 mg/L	0.000011	16.10%
	QC value within limits for Be Recovery = Not calculated					
Ca _f	-121.0	-0.0171 mg/L	0.00056	-0.0171 mg/L	0.00056	3.26%
	QC value within limits for Ca Recovery = Not calculated					
Co _f	0.6	0.00002 mg/L	0.000382	0.00002 mg/L	0.000382	>999.9%
	QC value within limits for Co Recovery = Not calculated					
Cr _f	-11.0	-0.00014 mg/L	0.000044	-0.00014 mg/L	0.000044	31.45%
	QC value within limits for Cr Recovery = Not calculated					
Cu _f	48.5	0.00013 mg/L	0.000016	0.00013 mg/L	0.000016	12.43%
	QC value within limits for Cu Recovery = Not calculated					
Fe _f	-0.4	-0.00035 mg/L	0.000509	-0.00035 mg/L	0.000509	146.10%
	QC value within limits for Fe Recovery = Not calculated					
K _f	43.2	0.0358 mg/L	0.02508	0.0358 mg/L	0.02508	69.96%
	QC value within limits for K Recovery = Not calculated					
Mg _f	-1.7	-0.00055 mg/L	0.002801	-0.00055 mg/L	0.002801	505.63%
	QC value within limits for Mg Recovery = Not calculated					
Mn _f	15.2	0.00003 mg/L	0.000002	0.00003 mg/L	0.000002	8.77%
	QC value within limits for Mn Recovery = Not calculated					
Mo _f	38.3	0.00299 mg/L	0.000533	0.00299 mg/L	0.000533	17.85%
	QC value within limits for Mo Recovery = Not calculated					
Na _f	121.3	0.0345 mg/L	0.00345	0.0345 mg/L	0.00345	9.99%
	QC value within limits for Na Recovery = Not calculated					
Ni _f	-19.1	-0.00086 mg/L	0.000382	-0.00086 mg/L	0.000382	44.53%
	QC value within limits for Ni Recovery = Not calculated					
Pb _f	8.7	0.00181 mg/L	0.000008	0.00181 mg/L	0.000008	0.42%
	QC value within limits for Pb Recovery = Not calculated					
V _f	1.7	0.00001 mg/L	0.000231	0.00001 mg/L	0.000231	>999.9%
	QC value within limits for V Recovery = Not calculated					
Zn _f	-153.1	-0.00317 mg/L	0.000052	-0.00317 mg/L	0.000052	1.65%
	QC value within limits for Zn Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 58
 Sample ID: 2711200572_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 80
 Date Collected: 12/4/2007 1:27:35 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711200572_2X

Analyte Back Pressure Flow
 All 248.0 kPa 0.65 L/min

Mean Data: 2711200572_2X

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units			Conc. Units			
Sca	336469.2	76.2 %	%	0.05			0.06%	
Yr	178311.0	73.3 %	%	0.24			0.33%	
B _†	53098.1	1.72 mg/L	mg/L	0.013	3.45 mg/L		0.027	
Ba _†	1027.6	0.0142 mg/L	mg/L	0.00011	0.0285 mg/L		0.00021	
Be _†	-1233.0	-0.00041 mg/L	mg/L	0.000014	-0.00081 mg/L		0.000028	
Ca _†	1497387.9	211 mg/L	mg/L	1.4	422 mg/L		2.9	
Co _†	135.3	0.00520 mg/L	mg/L	0.000194	0.0104 mg/L		0.00039	
Cr _†	226.8	0.00289 mg/L	mg/L	0.000177	0.00578 mg/L		0.000354	
Cu _†	630.0	0.00168 mg/L	mg/L	0.000116	0.00337 mg/L		0.000233	
Fe _†	355.1	0.339 mg/L	mg/L	0.0046	0.679 mg/L		0.0092	
K _†	18389.7	15.3 mg/L	mg/L	0.10	30.5 mg/L		0.20	
Mg _†	308496.7	103 mg/L	mg/L	0.2	205 mg/L		0.4	
Mn _†	274987.5	0.492 mg/L	mg/L	0.0028	0.984 mg/L		0.0056	
Mo _†	1188.2	0.0926 mg/L	mg/L	0.00048	0.185 mg/L		0.0010	
Na _†	2702018.3	769 mg/L	mg/L	8.6	1540 mg/L		17.2	
Ni _†	371.4	0.0167 mg/L	mg/L	0.00019	0.0334 mg/L		0.00039	
Pb _†	-33.9	-0.00701 mg/L	mg/L	0.000215	-0.0140 mg/L		0.00043	
V _†	3473.0	0.0204 mg/L	mg/L	0.00013	0.0408 mg/L		0.00025	
Zn _†	216.2	0.00437 mg/L	mg/L	0.000207	0.00874 mg/L		0.000413	

Sequence No.: 62
 Sample ID: ECV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 12/4/2007 1:46:06 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ECV

Analyte Back Pressure Flow
 All 247.0 kPa 0.65 L/min

Mean Data: ECV

Analyte	Mean Corrected Intensity	Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	372008.1	84.2 %	0.59			0.70%
Yr	195882.9	80.5 %	0.59			0.74%
B_†	74856.6	2.42 mg/L	0.024	2.42 mg/L	0.024	1.01%
	QC value within limits for B_ Recovery = 96.72%					
Ba†	368447.8	5.10 mg/L	0.047	5.10 mg/L	0.047	0.92%
	QC value within limits for Ba Recovery = 102.02%					
Be†	6193667.0	2.04 mg/L	0.010	2.04 mg/L	0.010	0.47%
	QC value within limits for Be Recovery = 102.08%					
Ca†	366675.2	51.7 mg/L	0.04	51.7 mg/L	0.04	0.09%
	QC value within limits for Ca Recovery = 103.45%					
Co†	134278.9	5.16 mg/L	0.038	5.16 mg/L	0.038	0.74%
	QC value within limits for Co Recovery = 103.21%					
Cr†	399679.7	5.10 mg/L	0.037	5.10 mg/L	0.037	0.72%
	QC value within limits for Cr Recovery = 101.91%					
Cu†	1874544.2	5.00 mg/L	0.000	5.00 mg/L	0.000	0.01%
	QC value within limits for Cu Recovery = 99.95%					
Fe†	5590.5	5.34 mg/L	0.032	5.34 mg/L	0.032	0.60%
	QC value within limits for Fe Recovery = 106.85%					
K†	62243.2	51.7 mg/L	0.27	51.7 mg/L	0.27	0.51%
	QC value within limits for K Recovery = 103.36%					
Mg†	160184.7	53.3 mg/L	0.06	53.3 mg/L	0.06	0.11%
	QC value within limits for Mg Recovery = 106.57%					
Mn†	2886835.0	5.16 mg/L	0.002	5.16 mg/L	0.002	0.04%
	QC value within limits for Mn Recovery = 103.30%					
Mo†	64368.2	5.02 mg/L	0.045	5.02 mg/L	0.045	0.89%
	QC value within limits for Mo Recovery = 100.30%					
Na†	185038.6	52.6 mg/L	0.42	52.6 mg/L	0.42	0.80%
	QC value within limits for Na Recovery = 105.28%					
Ni†	114371.9	5.15 mg/L	0.043	5.15 mg/L	0.043	0.84%
	QC value within limits for Ni Recovery = 102.91%					
Pb†	24950.6	5.16 mg/L	0.035	5.16 mg/L	0.035	0.68%
	QC value within limits for Pb Recovery = 103.29%					
V†	855266.0	5.04 mg/L	0.011	5.04 mg/L	0.011	0.21%
	QC value within limits for V Recovery = 100.87%					
Zn†	248329.8	5.11 mg/L	0.041	5.11 mg/L	0.041	0.80%
	QC value within limits for Zn Recovery = 102.28%					

All analyte(s) passed QC.

Sequence No.: 63
Sample ID: ECB
Analyst:
Initial Sample Wt:
Dilution:

Autosampler Location: 0
Date Collected: 12/4/2007 1:49:46 AM
Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: ECB

Analyte Back Pressure Flow
All 246.0 kPa 0.65 L/min

Mean Data: ECB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	392973.1	89.0 %		0.01			0.01%
Yr	199854.2	82.1 %		0.13			0.15%
B _f	486.1	0.0158 mg/L		0.00074	0.0158 mg/L	0.00074	4.70%
	QC value within limits for B _f Recovery = Not calculated						
Ba _f	-3.0	-0.00004 mg/L		0.000013	-0.00004 mg/L	0.000013	32.33%
	QC value within limits for Ba Recovery = Not calculated						
Be _f	-93.0	-0.00003 mg/L		0.000022	-0.00003 mg/L	0.000022	71.86%
	QC value within limits for Be Recovery = Not calculated						
Ca _f	-130.1	-0.0184 mg/L		0.00003	-0.0184 mg/L	0.00003	0.18%
	QC value within limits for Ca Recovery = Not calculated						
Co _f	-9.0	-0.00035 mg/L		0.000191	-0.00035 mg/L	0.000191	55.13%
	QC value within limits for Co Recovery = Not calculated						
Cr _f	17.6	0.00022 mg/L		0.000177	0.00022 mg/L	0.000177	78.82%
	QC value within limits for Cr Recovery = Not calculated						
Cu _f	-32.2	-0.00009 mg/L		0.000139	-0.00009 mg/L	0.000139	161.77%
	QC value within limits for Cu Recovery = Not calculated						
Fe _f	-2.4	-0.00230 mg/L		0.004850	-0.00230 mg/L	0.004850	211.05%
	QC value within limits for Fe Recovery = Not calculated						
K _f	142.4	0.118 mg/L		0.0053	0.118 mg/L	0.0053	4.45%
	QC value within limits for K Recovery = Not calculated						
Mg _f	-11.6	-0.00386 mg/L		0.000278	-0.00386 mg/L	0.000278	7.22%
	QC value within limits for Mg Recovery = Not calculated						
Mn _f	19.6	0.00003 mg/L		0.000004	0.00003 mg/L	0.000004	11.38%
	QC value within limits for Mn Recovery = Not calculated						
Mo _f	36.6	0.00285 mg/L		0.000180	0.00285 mg/L	0.000180	6.32%
	QC value within limits for Mo Recovery = Not calculated						
Na _f	296.9	0.0845 mg/L		0.00872	0.0845 mg/L	0.00872	10.32%
	QC value within limits for Na Recovery = Not calculated						
Ni _f	-8.3	-0.00037 mg/L		0.000247	-0.00037 mg/L	0.000247	66.20%
	QC value within limits for Ni Recovery = Not calculated						
Pb _f	6.4	0.00133 mg/L		0.000334	0.00133 mg/L	0.000334	25.09%
	QC value within limits for Pb Recovery = Not calculated						
V _f	-54.8	-0.00032 mg/L		0.000207	-0.00032 mg/L	0.000207	64.60%
	QC value within limits for V Recovery = Not calculated						
Zn _f	-170.6	-0.00353 mg/L		0.000085	-0.00353 mg/L	0.000085	2.40%
	QC value within limits for Zn Recovery = Not calculated						
All analyte(s) passed QC.							

Analytical Sequence

Method: 200.7&6010_071003

Seq.	Loc.	ID	Status
1	0	Calib Blank 1	Applied
2	15	Standard 2	Applied
3	15	ICV	QC Passed
4	1	QCS1	QC Failed
5	1	QCS2	QC Failed
6	1	QCS3	QC Failed
7	9	LINEARITY	QC Passed
8	10	ICSA	QC Passed
9	11	ICSAB	QC Failed
10	11	ICSAB	QC Failed
11	11	ICSAB	QC Failed
12	4	CCV	QC Passed
13	0	ICB	QC Passed
14	20	MRL	QC Passed
15	46	MBLANK2007	Analyzed
16	47	MRL2007	Analyzed
17	21	MRL	Analyzed
18	48	LCS	Analyzed
19	49	LCS D	Analyzed
20	50	N711140386_2X	Analyzed
21	51	N711140386_2XMS	Analyzed
22	52	N711140386_2XMSD	Analyzed
23	53	N711140388_2X	Analyzed
24	54	N711140446_2X	Analyzed
25	4	CCV	QC Passed
26	0	CCB	QC Passed
27	55	N711140450_2X	Analyzed
28	56	N711160270_2X	Analyzed
29	57	MBLANK2007	Analyzed
30	21	MRL	Analyzed
31	58	MRL2007	Analyzed
32	59	LCS	Analyzed
33	60	LCS D	Analyzed
34	61	2711190059	Analyzed
35	62	2711190059MS	Analyzed
36	63	2711190059MSD	Analyzed
37	4	CCV	QC Passed
38	0	CCB	QC Passed
39	5	MCV	QC Passed
40	64	2711200235_2X	Analyzed
41	65	2711200235_2XMS	Analyzed
42	66	2711200235_2XMSD	Analyzed
43	67	2711200228_2X	Analyzed
44	68	2711200229_2X	Analyzed
45	69	2711200230_2X	Analyzed
46	70	2711200231_2X	Analyzed
47	71	2711200232_2X	Analyzed
48	72	2711200233_2X	Analyzed
49	73	2711200234_2X	Analyzed
50	4	CCV	QC Passed
51	0	CCB	QC Passed
52	74	2711200236_2X	Analyzed
53	75	2711200237_2X	Analyzed
54	76	2711200238_2X	Analyzed
55	77	2711200239_2X	Analyzed
56	78	2711200240_2X	Analyzed
57	79	2711080235_2X	Analyzed
58	80	2711200572_2X	Analyzed
59	81	2711270096_2X	Analyzed
60	82	2711270097_2X	Analyzed
61	83	2711270098_2X	Analyzed
62	4	ECV	QC Passed
63	0	ECB	QC Passed

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Lot #

HNO3 R# 1004132 HCL R# 100427
IS = Yttrium(ME0709008)0.75mL + Scandium (ME0710007)0.5mL to 1000mL w/ 2% HNO3

Standards	Lot #	Exp. Date	Dilution
Calibration	ME0704023	(05/01/08)	1:10 ME0704027
(Prepare daily)	ME0704024	(05/01/08)	1:10
CCV/MCV/ECV	ME0710008	(04/17/09)	CCV/ECV MCV
(Prepare daily)			1:20 ME0710008 1:40 ME0710008
Spike/LCS	ME0606004	(12/13/07)	1:100 ME0601006
(Prepare daily)	ME0710001	(04/02/08)	1:100
	ME0708001	(02/03/09)	1:200
MRL	ME0709021	(03/20/08)	1:100 ME0603015
(Prepare daily)			
ICSA	ME0712003	(6/1/2008)	
ICSAB	ME0712004	(6/1/2008)	
QCS	ME0610005	(04/10/08)	
Linearity	ME0711002	(05/06/08)	

Method Sr/Ti/Sn/SiO2

Calibration	ME0708004	(02/07/08)	
CCV/ECV	ME0711003	(02/06/08)	
QCS	ME0711004	(02/06/08)	
Spike/LCS	ME0712006	(6/1/2008)	1:100
(Prepare daily)			
MRL	ME0712005	(6/1/2008)	1:100
(Prepare daily)			
Method Li			
Std/ICV/MRL	ME0711001	(05/01/08)	1:1000, 200, 40, 10
(Prepare daily)			
QCS	ME0707002	(01/10/08)	1:1000
(Prepare daily)			
LCS/Spike	ME0707002	(01/10/08)	1:50
(Prepare daily)			
ccv	ME0707002	(01/10/08)	1:40
(Prepare daily)			

From May 2005: the calibration std for ICP should be ME0505010, 011 not ME0408010
dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.

ME0704027

Initial:

Date:

wjh
4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration STD
Date Received/Prepped: Prep Daily
Date Expired: 5/1/2008
Manufacturer: MWH-wbh
Matrix: 2% HNO3 + 5% HCl
Amount:

ME #: 07040027
By: wbh
Lot #:
Certificate: NO
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	1:10 ME0704024	10 ug/ml
Sb		10 ug/ml
Sn		10 ug/ml
Ti		10 ug/ml
B		10 ug/ml
Ca	1:10 ME0704023	5 ug/ml
K		100 ug/ml
Mg		100 ug/ml
Na		100 ug/ml
Al		100 ug/ml
As		10 ug/ml
Ba		10 ug/ml
Co		10 ug/ml
Cr		10 ug/ml
Cu		10 ug/ml
Fe		10 ug/ml
Mn		10 ug/ml
Ni		10 ug/ml
Pb		10 ug/ml
Se		10 ug/ml
Tl		10 ug/ml
V		10 ug/ml
Zn		10 ug/ml
Cd		10 ug/ml
Be		5 ug/ml
SR		4 ug/ml
Ag		3 ug/ml
		2 ug/ml

ME0704024

Initial:

WBH

Date:

4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #2
Date Received/Prepped: 4/23/2007
Date Expired: 5/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid + Trace HF
Amount: 500 mL

ME #: 0704024
By: wbh
Lot #: A2-MEB235011
Certificate: Y
NIST SRM: Varies
Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml

ME0704023

Initial:
Date:

lv37
4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #1 **ME #:** 0704023
Date Received/Prepped: 4/23/2007 **By:** wbh
Date Expired: 5/1/2008 **Lot #:** A2-MEB235010
Manufacturer: Inorganic Ventures **Certificate:** Y
Matrix: 5% Nitric Acid **NIST SRM:** Varies
Amount: 500 mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		100 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		50 ug/ml
Be		40 ug/ml
SR		30 ug/ml
Ag		20 ug/ml

17750 1st Avenue, Suite 4
 Littleton, Colorado 80120-1074
 303.973.1111 ext. 211

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-2
 Lot Number: **A2-MEB235011**
 Matrix: tr. HF, 5% HNO₃(abs)

M70704024

100.00 µg/mL each:
 Mo, Sb, Sn, Ti,
 50.00 µg/mL each:
 B

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	50.07 ± 0.28 µg/mL	Molybdenum, Mo	100.3 ± 0.3 µg/mL
Tin, Sn	100.3 ± 0.3 µg/mL	Titanium, Ti	100.5 ± 0.2 µg/mL		

Certified Density: 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

\bar{x} = mean

x_i = individual results

n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2(\sum s_i)^2}{(n)^{1/2}}$$

$\sum s_i$ = The summation of all significant estimated errors

Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term "in-house std." is specified.

17100 Independence Avenue, Suite 100

Columbia, Maryland 21046

Phone: 410-286-1000

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-1
 Lot Number: **A2-MEB235010**
 Matrix: 5% HNO₃(abs)

M80704023

1,000.00 µg/mL each:
 Ca, K, Mg, Na,
 100.00 µg/mL each:
 Al, As, Ba, Co, Cr₃, Cu, Fe, Mn, Ni, Pb, Se, Tl, V, Zn,
 50.00 µg/mL each:
 Cd,
 40.00 µg/mL each:
 Be,
 30.00 µg/mL each:
 Sr,
 20.00 µg/mL each:
 Ag

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.1 ± 0.5 µg/mL	Arsenic, As	100.2 ± 0.3 µg/mL	Barium, Ba	99.9 ± 0.2 µg/mL
Beryllium, Be	39.98 ± 0.08 µg/mL	Cadmium, Cd	50.05 ± 0.12 µg/mL	Calcium, Ca	997 ± 3 µg/mL
Chromium+3, Cr ₃	100.1 ± 0.4 µg/mL	Cobalt, Co	100.1 ± 0.2 µg/mL	Copper, Cu	100.1 ± 0.2 µg/mL
Iron, Fe	100.0 ± 0.2 µg/mL	Lead, Pb	100.1 ± 0.3 µg/mL	Magnesium, Mg	996 ± 3 µg/mL
Manganese, Mn	100.1 ± 0.3 µg/mL	Nickel, Ni	100.1 ± 0.2 µg/mL	Potassium, K	1,003 ± 2 µg/mL
Selenium, Se	100.1 ± 0.2 µg/mL	Silver, Ag	20.03 ± 0.06 µg/mL	Sodium, Na	997 ± 5 µg/mL
Strontium, Sr	29.92 ± 0.18 µg/mL	Thallium, Tl	100.0 ± 0.1 µg/mL	Vanadium, V	100.1 ± 0.3 µg/mL
Zinc, Zn	100.1 ± 0.4 µg/mL				

Certified Density: 1.056 g/mL (measured at 22° C)

Initial: STE
Date: 10/19/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV/QCS Stock Standard **ME #:** 0710008
Date Received/Prepped: 10/19/2007 **By:** STE
Date Expired: 4/17/2009 **Lot #:** 07J154
Manufacturer: CPI **Certificate:** Y
Matrix: 5% HNO3 = tr HF **NIST SRM:** Various
Amount: 100 mL x 10 **Storage:** Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ag	P/N 4400-061003RH01	20 ppm
Al		100 ppm
As		100 ppm
B		50 ppm
Ba		100 ppm
bE		40 ppm
Ca		1000 ppm
Cd		50 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		1000 ppm
Mg		1000 ppm
Mn		100 ppm
Mo		100 ppm
Na		1000 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		20 ppm
Sn		20 ppm
Ti		20 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 4/17/2009

Certificate of Analysis

Part Number: 4400-061003RH01
Lot Number: 07J154
Shelf Life: 18 months

ME 0710008
 received 10/19/09
 STE

MWH
 Custom Multi
 5% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: STE
Date: 10/2/07

METALS STANDARD DOCUMENTATION

Standard: ICP Spike Solution **ME #:** 0710001
Date Received/Prepped: 10/2/2007 **By:** STE
Date Expired: 4/2/2008 **Lot #:** VARIOUS
Manufacturer: MWH-STE **Certificate:**
Matrix: H2O **NIST SRM:**
Amount: 125mL **Storage:** Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
AS	8.0mL ME0611005/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

METALS STANDARD DOCUMENTATION

0433
6/19/06

Standard: ICP/ICPMS LCS/SPIKE Solution **ME #:** 0606004
Date Received/Prepped: 6/19/2006 **By:** wbh
Date Expired: 12/13/2006 **Lot #:** 06F103
Manufacturer: CPI **Certificate:** Y
Matrix: 5% HNO₃ + 0.1% HF **NIST SRM:** 3100 Series
Amount: 10 x 100 mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Iron	CPI P/N: 4400-050314RH01	500 mg/L
Aluminum		200 mg/L
Barium		100 mg/L
Cobalt		100 mg/L
Chromium		100 mg/L
Copper		100 mg/L
Molybdenum		100 mg/L
Strontium		100 mg/L
Titanium		100 mg/L
Vanadium		100 mg/L
Zinc		100 mg/L
Tin		100 mg/L
Silver		50 mg/L
Boron		50 mg/L
Manganese		50 mg/L
Nickel		50 mg/L
Antimony		50 mg/L
Arsenic		20 mg/L
Cadmium		20 mg/L
Lead		20 mg/L
Selenium		20 mg/L
Thallium		20 mg/L
Uraium		20 mg/L
Beryllium		5 mg/L
Tin		100 mg/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 12/13/2007

Certificate of Analysis

Part Number: 4400-050314RH01
Lot Number: 06F103
Shelf Life: 18 months

M7060604

MWH Labs
 5% HNO₃ + 0.1% HF
 #REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

WBH
11/1/06

METALS STANDARD DOCUMENTATION

Standard: Arsenic Stock Std
Date Received/Prepped: 11/1/2006
Date Expired: 10/1/2007
Manufacturer: IV
Matrix: 1.4% HNO3
Amount: 100mL

ME #: 0611005
By: WBH
Lot #: Y-AS02029
Certificate: Y
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	Cat # CGAS1-1	1004 ppm

1.0 **INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 **DESCRIPTION OF CRM** **1000 µg/mL Arsenic in 1.4% (abs) HNO₃**

Catalog Number: CGAS1-1, CGAS1-2, and CGAS1-5
 Lot Number: **Y-AS02029**
 Starting Material: As Polycrystalline lump
 Starting Material Purity (%): 99.999055
 Starting Material Lot No: 23115
 Matrix: 1.4% (abs) HNO₃

170611005

3.0 **CERTIFIED VALUES AND UNCERTAINTIES**

Certified Concentration: 1004 ± 2 µg/mL 995 ± 2 µg/g

Certified Density: 1.009 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_1}{n}$$

(\bar{x}) = mean
 x₁ = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_1^2)]^{1/2}}{(n)}$$

∑s₁ = The summation of all significant estimated errors
 (Most common are the errors from instrumental measurement weighting, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 **TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS**

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This IV product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 **Assay Method #1** 1004 ± 2 µg/mL 995 ± 2 µg/g (Avg 2 runs)
 ICP Assay NIST SRM 3103a Lot Number: 010713

Assay Method #2 1003 ± 5 µg/mL 994 ± 5 µg/g
 Gravimetric NIST SRM Lot Number: See Sec. 4.2

Initial:

WBH

Date:

4/16/07

METALS STANDARD DOCUMENTATION

Standard: Lead Stock Standard
Date Received/Prepped: 4/16/2007
Date Expired: 10/11/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0704013
By: WBH
Lot #: 07A097
Certificate: Y
NIST SRM: 3128
Room temp. storage

Component	Comment	Conc. Unit:
Pb	P/N S4400-1000281	1000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 -31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N S4400-1000281
P/N 4400-1000281
 Single-Element Lead Standard
 Pb in 2% HNO₃
 1000 ± 3 µg/mL

1270704013

Lot # 07A097

Material Source: Lead Metal
 Source Purity: 99.995 %
 Specific Gravity: 1.009 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3128. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 10.3	0.1	Cu 58	0.1	Pb X	0.1	K ND	70	Tl 0.25	0.1
Sb ND	0.1	Dy ND	0.1	Li 2	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.22	0.1	Eu ND	0.1	Mg 1.4	0.2	Rh IN	0.1	Sn ND	0.1
Be 0.58	0.1	Gd ND	0.1	Mn 3.8	1	Rb ND	0.1	Ti 0.58	0.1
Bi 0.7	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.17	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.9	0.1	Si 31	8	Yb ND	0.1
Ca 25	7	Ho ND	0.1	Nb ND	0.1	Ag 6.1	0.1	Y ND	0.1
Ce ND	0.1	I 0.1	0.2	Os ND	0.1	Na 3.5	1	Zn 23	2
Cs 0.26	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr INT	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial:

WBY

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Thallium 1000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0702006
By: WBH
Lot #: 06H213
Certificate: Y
NIST SRM: 3158
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
TI	P/N S4400-1000581	1000 ppm

Aug 16 08



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

MF0702006

P/N 4400-1000581
P/N S4400-1000581

Single Element Thallium Standard
 Tl in 2% HNO₃
 1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal
 Source Purity: 99.999%
 Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>Ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>		<u>ppb</u>	<u>DL</u>
Al	13.3	0.1	Cu	9.3	0.1	Pb	41	0.1	K	ND	70	Tl	X	0.1
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1	Th	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1	Tm	ND	0.1
Ba	0.37	0.1	Eu	ND	0.1	Mg	1.7	0.2	Rh	ND	0.1	Sn	ND	0.1
Be	0.67	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1	Ti	0.45	0.1
Bi	0.12	0.1	Ga	ND	0.1	Hg	0.16	0.2	Ru	ND	0.1	W	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.21	0.1	Sm	ND	0.1	U	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6	V	ND	1
Cd	1.6	0.1	Hf	ND	0.1	Ni	1.1	0.1	Si	46	8	Yb	ND	0.1
Ca	51	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.3	0.1	Y	ND	0.1
Ce	ND	0.1	I	0.4	0.2	Os	ND	0.1	Na	3.3	1	Zn	14.7	2
Cs	0.24	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1	Zr	ND	0.1
Cr	ND	1	Fe	ND	30	P	20	10	Ta	ND	0.1			
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1			

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.



Initial:
Date:

W/34
3/5/07

METALS STANDARD DOCUMENTATION

Standard: Selenium Stock Standard
Date Received/Prepped: 3/5/2007
Date Expired: 8/22/2008
Manufacturer: CPI
Matrix: 2% HNO3
Amount: 100 mL

ME #: 0703001
By: wbh
Lot #: 6.00E+228
Certificate: Y
NIST SRM: 3148
Storage: Room Temp

Component	Comment	Conc. Unit:
Se	P/N # S4400-1000491	1000 ppm



USA
 5580 Skylane Boulevard 707 525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707 545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-1000491

P/N S4400-1000491

Single-Element Selenium Standard

Se in 2% HNO₃

1000 ± 3 µg/mL

Lot # 06E228

M70703001

Material Source: Selenium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.011 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3148. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 1.8	0.1	Cu 0.4	0.1	Pb 0.3	0.1	K ND	70	Tl 3.6	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Eu ND	0.1	Mg 1.1	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo 0.6	0.1	Sm ND	0.1	U ND	0.1
Br INT	10	Au ND	0.1	Nd ND	0.1	Se X	6	V ND	1
Cd 0.4	0.1	Hf ND	0.1	Ni 0.6	0.1	Si 40	8	Yb ND	0.1
Ca 5	7	Ho ND	0.1	Nb INT	0.1	Ag 0.8	0.1	Y ND	0.1
Ce ND	0.1	I 0.5	0.2	Os ND	0.1	Na 3.8	1	Zn ND	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr INT	0.1
Cr ND	1	Fe ND	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pr ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: DYH
Date: 8/8/07

METALS STANDARD DOCUMENTATION

Standard: Dat MW Standard
Date Received/Prepped: 8/8/2007
Date Expired: 2/3/2009
Manufacturer: CPI International
Matrix: 5% HNO₃
Amount: 100mL

ME #: 0708001
By: DYH
Lot #: 07H025
Certificate:
NIST SRM:
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Na	P/N 4400-130309	10000 ug/mL
Ca		10000 ug/mL
Mg		4000 ug/mL
K		4000 ug/mL

**USA**

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

ME0708001

CERTIFICATE OF ANALYSIS

**P/N: 4400-130309
Lot Number: 07H025
Shelf Life: 18 months
Expiration Date: 02/03/2009**

MWH
Dat MW Standard
 $\mu\text{g/mL} \pm 0.5\%$ in 5% HNO_3

Na 10,000 Ca 10,000 Mg 4,000 K 4,000

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at $1000\mu\text{g/mL}$ by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: STE
Date: 12/01/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSAB) **ME #:** 0712004
Date Received/Prepped: 12/1/2007 **By:** STE
Date Expired: 6/1/2008 **Lot #:**
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL **Room temp. storage**

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500 (25 mL)	250 ppm
Ca	P/N 4400-INTB1-100 (2.5 mL)	250 ppm
Fe		100 ppm
Mg		250 ppm
Ag		0.5 ppm
Ba		0.25 ppm
Be		0.25 ppm
Cd		0.5 ppm
Co		0.25 ppm
Cr		0.25 ppm
Cu		0.25 ppm
Mn		0.25 ppm
Ni		0.5 ppm
Pb		0.5 ppm
V		0.25 ppm
Zn		0.5 ppm

Initial: STE
Date: 12/1/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSA) **ME #:** 0712003
Date Received/Prepped: 12/1/2007 **By:** STE
Date Expired: 6/1/2008 **Lot #:**
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	250 ppm
Ca	25 mL in 500 mL	250 ppm
Fe		100 ppm
Mg		250 ppm

Initial:

STE

Date:

8/29/07

METALS STANDARD DOCUMENTATION

Standard: CLP Analytes B Solution
Date Received/Prepped: 8/27/2007
Date Expired: 2/15/2009
Manufacturer: CPI International
Matrix: 5% HNO₃
Amount: 100 mL

ME #: 0708010
By: STE
Lot #: 07c256
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Ag	P/N 4400-INTB1-100	100 ug/L
Ba		50 ug/L
Be		50 ug/L
Cd		100 ug/L
Co		50 ug/L
Cr		50 ug/L
Cu		50 ug/L
Mn		50 ug/L
Ni		100 ug/L
Pb		100 ug/L
V		50 ug/L
Zn		100 ug/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com



*Innovative Solutions
 in Analytical Science and
 Technology*

ME 070 800

CERTIFICATE OF ANALYSIS

P/N 4400-INTB1-100

CLP Analytes B Solution
 in 5% HNO₃

Lot # 07c256

Material Source: Metals and Salts
 Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag	100	Ba	50	Be	50	Cd	100
Co	50	Cr	50	Cu	50	Mn	50
Ni	100	Pb	100	V	50	Zn	100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: STE
Date: 8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Interferents A Solution **ME #:** 0708009
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 2/15/2009 **Lot #:** 07E175
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL **Storage:** Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Al	P/N 4400-INTA1-500	5000 ug/mL
Ca		5000 ug/mL
Fe		2000 ug/mL
Mg		5000 ug/mL

FEB 15 09



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax: 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500

CLP Interferents A Solution
in 5% HNO₃

ME 0708009

Lot # 07E175

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

80

colitag
The Analytical Solution

PEAK PERFORMANCE
CERTIFIED REFERENCE MATERIALS

Accu-prep 1000™
reaction Activated

MOD BLOCK

Waters™
PREP TUBES

Initial: STE
Date: 11/6/07

METALS STANDARD DOCUMENTATION

Standard: ICP LINEARITY CHECK
Date Received/Prepped 11/6/2007
Date Expired: ^{STE} ~~5/6/2008~~ 2/6/2008
Manufacturer: MWH-STE
Matrix: 5% HNO3
Amount: 500 mL

ME #: 0711002
By: STE
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Ca	15.0 mL ME0702002/ 500 mL	300 ppm
K	15.0 mL ME0702005/ 500 mL	300 ppm
Mg	10.0 mL ME0702004/ 500 mL	200 ppm
Na	15 mL ME0702003/ 500 mL	300 ppm
Fe	5.0 mL ME0701008/ 500 mL	100 ppm

Initial:

Date:

W34
1/26/07

METALS STANDARD DOCUMENTATION

Standard: FE 10000ppm Stock Std
Date Received/Prepped: 1/26/2007
Date Expired: 7/19/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 100 mL

ME #: 701008
By: WBH
Lot #: 061143
Certificate:
NIST SRM: 3126a
Storage: Room Temp

Component	Comment	Conc. Unit:
Fe	PN4400-10M261	10000 PPM



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 JUL 19 98 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M261
P/N S4400-10M261
 Single-Element Iron Standard
 Fe in 4% HNO₃
 10,000 ± 30 µg/mL

M7E0701008

Lot # 06I143

Material Source: Iron Metal
 Source Purity: 99.999%
 Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al INT	0.1	Cu 6.4	0.1	Pb ND	0.1	K ND	70	Ti 0.18	0.1
Sb 0.35	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Eu ND	0.1	Mg 1.3	0.2	Rh ND	0.1	Sn 0.67	0.1
Be ND	0.1	Gd ND	0.1	Mn INT	1	Rb ND	0.1	Tl 0.21	0.1
Bi ND	0.1	Ga 0.41	0.1	Hg ND	0.2	Ru ND	0.1	W 0.13	0.1
B ND	4	Ge INT	0.1	Mo 4.9	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 9.3	0.1	Si INT	8	Yb ND	0.1
Ca 15	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.34	0.2	Os ND	0.1	Na 8	1	Zn 8.6	2
Cs 0.34	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr 3.3	1	Fe X	30	P 28	10	Ta ND	0.1		
Co 12	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702005
By: WBH
Lot #: 07B056
Certificate: Y
NIST SRM: 3141
Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

AUG 16 '08



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

M20702005

P/N 4400-10M411
P/N S4400-10M411
Single-Element Potassium Standard
K in 1% HNO₃
10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)
Source Purity: 99.999%
Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 0.39	0.1	Cu 0.16	0.1	Pb ND	0.1	K X	70	Tl ND	0.1
Sb 0.34	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.14	0.1	Eu ND	0.1	Mg 2.6	0.2	Rh ND	0.1	Sn 0.17	0.1
Be ND	0.1	Gd ND	0.1	Mn 0.93	1	Rb 9.5	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Ca 82	7	Hf ND	0.1	Ni 0.4	0.1	Si 50	20	Yb ND	0.1
Ce ND	0.1	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Cs ND	0.1	I ND	0.2	Os ND	0.1	Na 19	1	Zn 2.9	1
Cr ND	1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Co ND	0.1	Fe ND	30	P 18	10	Ta ND	0.1		
		La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH
2/14/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 4% HNO₃
Amount: 250 mL

ME #: 0702004
By: WBH
Lot #: 07B058
Certificate: Y
NIST SRM: 3131
Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Mg	P/N 4400-10M311	10000 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

M70702004

P/N 4400-10M311
P/N S4400-10M311
 Single-Element Magnesium Standard
 Mg in 4% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B058

Material Source: Magnesium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>	<u>ppb</u>	<u>DL</u>
Al 28	0.1	Cu 1.6	0.1	Pb 7.7	0.7	K ND	70	Tl 0.91	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr 0.28	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.28	0.1	Eu ND	0.1	Mg X	0.2	Rh ND	0.1	Sn 0.14	0.1
Be ND	0.1	Gd 0.23	0.1	Mn 19.8	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga 0.18	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd 1.1	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 1	0.1	Si 64	20	Yb ND	0.1
Ca ND	7	Ho ND	0.1	Nb ND	0.1	Ag 0.19	0.1	Y 0.2	0.1
Ce 2.1	0.1	I 1	0.2	Os ND	0.1	Na 7.2	1	Zn ND	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 0.19	0.1	Zr 0.29	0.1
Cr ND	1	Fe 80	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La 0.76	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO3
Amount: 250 mL

ME #: 0702003
By: WBH
Lot #: 07B057
Certificate: Y
NIST SRM: 3152a
Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
In Analytical Science and
Technology

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

M70702003

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO₃)
Source Purity: 99.99%
Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 1.5	0.1	Cu 0.45	0.1	Pb ND	0.1	K ND	70	Tl ND	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.13	0.1	Eu ND	0.1	Mg 2.3	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	8	Yb ND	0.1
Ca 120	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na X	1	Zn 2.9	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.



Initial: WBH
Date: 2/20/07

METALS STANDARD DOCUMENTATION

Standard:	Calcium 10000ppm Stock Std	ME #: 0702002
Date Received/Prepped:	2/20/2007	By: WBH
Date Expired:	8/16/2008	Lot #: 07B065
Manufacturer:	CPI	Certificate: Y
Matrix:	4% HNO3	NIST SRM: 3109a
Amount:	250 mL	Room temp. storage

Component	Comment	Conc. Unit:
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA
5580 Skylark Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

180702602

Single-Element Calcium Standard

Ca in 4% HNO₃
10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)
Source Purity: 99.997%
Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Ti 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Tl ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

91

colitag™

PEAK PERFORMANCE
ANALYTICAL MATERIALS

Accu-prep 7000™
Extraction Manifold

MOD BLOCK

Two-Phase™
PE Diluents