

TABLE OF CONTENTS

COVER PAGE.....	1
QC CHECKLIST	2
SUMMARY SHEET	3
ANALYTICAL SEQUENCE	5
INITIAL CALIBRATION.....	8
PERIODIC QC.....	21
QC: (MBLANK, MRL, LCS1, LCS2).....	23
QC: (MS/MSD 2711210438)	27
PERIODIC QC.....	29
QC: (MS/MSD 2711210440)	31
SAMPLE (2711200049).....	33
PERIODIC QC.....	34
SAMPLES	37
STANDARDS PREPARATION WORKSHEET AND CERTIFICATES OF ANALYSIS.....	40

Level IV Data Package

MWH Group 223000

Method: EPA 200.7

2711200573

EPA 200.7/6010B QC Check List

Analyst STE Analysis Date 11/21/09

Reviewer/Date MME/12/20/09

Instrument PerKin Elmer Optima 4300DV

All sample analyzed within 6 month holding time

All sample raw concentration below the high standard or linear and rerun
changed samples marked for dilution

Initial and closing QC

- ICV within +/- 5%
- Linearity check +/- 10%
- ICSAB +/- 20%
- QCS +/- 5%
- MRL +/- 50%

Middle, closing and batch QC

- n/a FilterCheck < 1/2 MRL
- MBLANK < 1/2 MRL
- LCS +/-15%
- MS/MSD +/-30%(200.7) +/- 25% (6010B)
- CCV/MCV/ECV +/- 10%
- ICB/CCB/ECB < 1/2 MRL
- CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/-20%
- RPD between LCS/LCSD is within +/- 20%
- Internal standards +/-20%
- All pH of the samples are < 2

No more than 20 samples per batch

MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

QIR needed for failed QC

n/a Special Det Code noted on the cover sheet

n/a R value for multi point calibration is > 0.995

n/a Proper MRL check ran for special low MRL samples

Reagent and Standards used for
Optima 4300 DV
Updated 11/06/07

Int: STE
Date: 11/21/09

Method 200.7/6010

ICP SUMMARY SHEET

File ID: 071121d
 Date Started: 11/21/07
 Analyst ID: ste

SAMPLE ID

QCS1	(19:48)	QCS2	(19:51)	QCS3	(19:55)
LINEARITY	(19:58)	WASH	(20:11)	FILTER CHECK	(20:38)
2711200845	(20:46)	B2711160108	(20:59)	B2711160128	(21:03)
B2711160129	(21:07)	B2711160130	(21:11)	B2711160131	(21:15)
B2711160132	(21:19)	B2711160133	(21:23)	B2711160134	(21:27)
B2711160135	(21:32)	2711200328	(21:44)	B2711160136	(21:57)
B2711160137	(22:01)	2711200221	(22:05)	2711200224	(22:10)
2711200225	(22:15)	2711200226	(22:19)	2711200227	(22:24)
2711200232	(22:29)	2711200233	(22:34)	2711210438	(23:07)
B2711160138	(23:20)	B2711160139	(23:24)	B2711160140	(23:29)
B2711160141	(23:33)	B2711160142	(23:37)	B2711160143	(23:42)
B2711160144	(23:46)	B2711160145	(23:50)	B2711160146	(23:55)
2711210440	(12:07)	B2711160147	(12:20)	2711200220	(12:25)
2711200222	(12:30)	2711200223	(12:35)	2711200228	(12:39)
2711200234	(12:44)	2711200572_1	(12:48)	2711200573_1	(12:53)
2711200049	(12:58)	2711210442	(1:31)	2711160138	(1:43)
2711160139	(1:47)	2711160140	(1:52)	2711160141	(1:56)
2711160142	(2:00)	2711160143	(2:04)	2711160144	(2:08)
2711160145	(2:13)	2711160146	(2:18)	2711210129	(2:30)
2711020164	(2:43)	2711020165	(2:48)	2711020166	(2:52)
2711080062	(2:56)	2711160071	(3:00)	F2711140388	(3:04)
2711210016	(3:09)	2711210046	(3:14)	2711210047	(3:19)
2711200738	(3:52)				

COMMENT:

Analyst: STE

Approved By: MJE 11/21/07

BATCH NUMBER for 071121d

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2711200845

2711200845	2711200328	2711200221
2711200224	2711200225	2711200226
2711200227	2711200232	2711200233

Batch ID: 2711210438

2711210438	2711210440	2711200220
2711200222	2711200223	2711200228
2711200234	2711200572_10X	2711200573_10X
2711200049		

Batch ID: 2711210442

2711210442	2711160138	2711160139
2711160140	2711160141	2711160142
2711160143	2711160144	2711160145
2711160146	2711210129	2711020164
2711020165	2711020166	2711080062
2711160071	2711210016	2711210046
2711210047		

Batch ID:

2711200738

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	11/21/07	19:38	1	10.026	10	95-105	100%
ICB	11/21/07	19:42	1	-0.0001	ND		
QCS1	11/21/07	19:48	1	10.188	10		
QCS2	11/21/07	19:51	1	10.083	10		100%
QCS3	11/21/07	19:55	1	10.042	10		
LINEARITY	11/21/07	19:58	1	0.0015	.0015		
ICSA	11/21/07	20:02	1	-0.0007	ND	80-120	
ICSAB	11/21/07	20:07	1	.24603	.246	80-120	98.4%
WASH	11/21/07	20:11	1	0.0000	0.0000		
CCV	11/21/07	20:15	1	5.0887	5.09	90-110	101%
MBLANK	11/21/07	20:19	1	-0.0002	ND ✓		
MRL	11/21/07	20:25	1	0.0097	.0097	50-150	97.4%
LCS	11/21/07	20:29	1	1.0408	1.04	85-115	104%
LCSD	11/21/07	20:33	1	1.0401	1.04	85-115	104%
FILTER CHECK	11/21/07	20:38	1	-0.0028	ND		
MRL/2	11/21/07	20:42	1	-0.0028	ND		
2711200845	11/21/07	20:46	1	0.0000	0.0000		
2711200845MS	11/21/07	20:51	1	1.0087	1.01	[1.009]	100%
2711200845MSD	11/21/07	20:55	1	.99631	.996	[0.996]	99.6%
2711200845T	11/21/07	20:55	1		1.00	70 - 130	
B2711160108	11/21/07	20:59	1	-0.0010	ND		
B2711160128	11/21/07	21:03	1	-0.0008	ND		
B2711160129	11/21/07	21:07	1	-.00107	ND		
B2711160130	11/21/07	21:11	1	-0.0007	ND		
B2711160131	11/21/07	21:15	1	-0.0006	ND		
B2711160132	11/21/07	21:19	1	-0.0010	ND		
B2711160133	11/21/07	21:23	1	-0.0007	ND		
B2711160134	11/21/07	21:27	1	-0.0008	ND		
B2711160135	11/21/07	21:32	1	-0.0009	ND		
CCV	11/21/07	21:36	1	5.0260	5.03	90-110	100%
CCB	11/21/07	21:40	1	-0.0001	ND ✓		
2711200328	11/21/07	21:44	1	0.0068	.0068		
2711200328MS	11/21/07	21:49	1	1.0238	1.02	[1.017]	101%
2711200328MSD	11/21/07	21:53	1	1.0111	1.01	[1.004]	100%
2711200328T	11/21/07	21:53	1		1.00	70 - 130	
B2711160136	11/21/07	21:57	1	-0.0010	ND		
B2711160137	11/21/07	22:01	1	-0.0010	ND		
2711200221	11/21/07	22:05	1	-0.0010	ND		
2711200224	11/21/07	22:10	1	-0.0010	ND		
2711200225	11/21/07	22:15	1	-0.0008	ND		
2711200226	11/21/07	22:19	1	-0.0011	ND		
2711200227	11/21/07	22:24	1	-0.0011	ND		
2711200232	11/21/07	22:29	1	-0.0011	ND		
2711200233	11/21/07	22:34	1	-0.0009	ND		
CCV	11/21/07	22:39	1	5.0536	5.05	90-110	101%
CCB	11/21/07	22:43	1	-0.0001	ND ✓		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
MCV	11/21/07	22:47	1	2.6341	2.63	90-110	105%
MBLANK	11/21/07	22:51	1	-0.0000	ND ✓		
MRL	11/21/07	22:55	1	0.0095	.0095	50-150	95.4%
LCS	11/21/07	22:59	1	1.0305	1.03	85-115	103%
LCSD	11/21/07	23:03	1	1.0344	1.03	85-115	103%
2711210438	11/21/07	23:07	1	-0.0000	ND		
2711210438MS	11/21/07	23:12	1	1.0077	1.01	[1.008]	100%
2711210438MSD	11/21/07	23:16	1	1.0124	1.01	[1.012]	101%
2711210438T	11/21/07	23:16	1		1.00	70 - 130	
B2711160138	11/21/07	23:20	1	-0.0010	ND		
B2711160139	11/21/07	23:24	1	-0.0011	ND		
B2711160140	11/21/07	23:29	1	-0.0011	ND		
B2711160141	11/21/07	23:33	1	-0.0009	ND		
B2711160142	11/21/07	23:37	1	-0.0011	ND		
B2711160143	11/21/07	23:42	1	-0.0010	ND		
B2711160144	11/21/07	23:46	1	-0.0008	ND		
B2711160145	11/21/07	23:50	1	-0.0007	ND		
B2711160146	11/21/07	23:55	1	-0.0008	ND		
CCV	11/21/07	23:59	1	5.0417	5.04	90-110	100%
CCB	11/22/07	12:03	1	-0.0001	ND ✓		
2711210440	11/22/07	12:07	1	-0.0002	ND		
2711210440MS	11/22/07	12:12	1	1.0371	1.04	[1.037]	103%
2711210440MSD	11/22/07	12:16	1	1.0173	1.02	[1.017]	101%
2711210440T	11/22/07	12:16	1		1.00	70 - 130	
B2711160147	11/22/07	12:20	1	-0.0007	ND		
2711200220	11/22/07	12:25	1	-0.0009	ND		
2711200222	11/22/07	12:30	1	-0.0009	ND		
2711200223	11/22/07	12:35	1	-0.0009	ND		
2711200228	11/22/07	12:39	1	-0.0009	ND		
2711200234	11/22/07	12:44	1	-0.0011	ND		
2711200572_10X ✓	11/22/07	12:48	10	-0.0043	ND		
2711200573_10X ✓	11/22/07	12:53	10	0.0135	.014		
2711200049	11/22/07	12:58	1	-0.0009	ND		
CCV	11/22/07	1:03	1	5.0882	5.09	90-110	101%
CCB	11/22/07	1:07	1	0.0001	0.0001 ✓		
MCV	11/22/07	1:11	1	2.6000	2.6	90-110	104%
MBLANK	11/22/07	1:15	1	-0.0001	ND ✓		
MRL	11/22/07	1:19	1	0.0095	.0095	50-150	95.4%
LCS	11/22/07	1:23	1	1.0370	1.04	85-115	103%
LCSD	11/22/07	1:27	1	1.0351	1.04	85-115	103%
2711210442	11/22/07	1:31	1	0.0004	0.0003		
2711210442MS	11/22/07	1:35	1	1.0247	1.02	[1.025]	102%
2711210442MSD	11/22/07	1:39	1	1.0240	1.02	[1.024]	102%
2711210442T	11/22/07	1:39	1		1.00	70 - 130	
2711160138	11/22/07	1:43	1	-0.0010	ND		
2711160139	11/22/07	1:47	1	-0.0012	ND		
2711160140	11/22/07	1:52	1	-0.0011	ND		
2711160141	11/22/07	1:56	1	-0.0012	ND		
2711160142	11/22/07	2:00	1	-0.0011	ND		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2711160143	11/22/07	2:04	1	-0.0011	ND		
2711160144	11/22/07	2:08	1	.01457	.015		
2711160145	11/22/07	2:13	1	-0.0010	ND		
2711160146	11/22/07	2:18	1	-0.0011	ND		
CCV	11/22/07	2:22	1	5.1083	5.11	90-110	102%
CCB	11/22/07	2:26	1	-0.0001	ND ✓		
2711210129	11/22/07	2:30	1	-0.0012	ND		
2711210129MS	11/22/07	2:35	1	1.0219	1.02	[1.022]	102%
2711210129MSD	11/22/07	2:39	1	1.0228	1.02	[1.023]	102%
2711210129T	11/22/07	2:39	1		1.00	70 - 130	
2711020164	11/22/07	2:43	1	-0.0013	ND		
2711020165	11/22/07	2:48	1	-0.0010	ND		
2711020166	11/22/07	2:52	1	-0.0004	ND		
2711080062	11/22/07	2:56	1	-0.0010	ND		
2711160071	11/22/07	3:00	1	-0.0010	ND		
F2711140388	11/22/07	3:04	1	-0.0014	ND		
2711210016	11/22/07	3:09	1	-0.0014	ND		
2711210046	11/22/07	3:14	1	-0.0010	ND		
2711210047	11/22/07	3:19	1	-0.0011	ND		
CCV	11/22/07	3:24	1	5.1223	5.12	90-110	102%
CCB	11/22/07	3:28	1	-0.0001	ND ✓		
MCV	11/22/07	3:32	1	2.6506	2.65	90-110	106%
MBLANK	11/22/07	3:36	1	-0.0001	ND		
MRL	11/22/07	3:40	1	0.0099	.0099	50-150	98.9%
LCS	11/22/07	3:44	1	1.0404	1.04	85-115	104%
LCSD	11/22/07	3:48	1	1.0399	1.04	85-115	103%
2711200738	11/22/07	3:52	1	-0.0011	ND		
2711200738MS	11/22/07	3:57	1	1.0277	1.03	[1.028]	102%

} n/a
} ONR

=====
Analysis Begun

Start Time: 11/21/2007 7:31:13 PM Plasma On Time: 12:00:00 AM
 Logged In Analyst: Administrator Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Owner\Sample Information\NOVEMBER_11_2007_SIF\071121D.sif
 Batch ID: 071121D
 Results Data Set: 071121D
 Results Library: C:\pe\Owner\Results\Results.mdb

=====
Method Loaded

Method Name: Fullist-syljohn
 IEC File: 070703.iec
 Method Description: ICP FULLLIST

Method Last Saved: 11/19/2007 9:16:55 PM
 MSF File:

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Mo	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a

Sequence No.: 1
 Sample ID: Calib Blank 1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 11/21/2007 7:31:13 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

=====
Nebulizer Parameters: Calib Blank 1

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

=====
Mean Data: Calib Blank 1

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	425260.2	2764.35	0.65%	100.0 %
Yr	206125.5	4208.75	2.04%	100 %
Ag†	232.1	30.86	13.30%	[0.00] mg/L
Al†	57.5	11.01	19.16%	[0.00] mg/L
As†	9.8	1.11	11.30%	[0.00] mg/L
B†	964.9	75.51	7.83%	[0.00] mg/L
Ba†	-12.6	3.59	28.45%	[0.00] mg/L
Be†	-3652.2	15.90	0.44%	[0.00] mg/L
Ca†	436.2	2.81	0.65%	[0.00] mg/L

Cdt	54.4	1.13	2.07%	[0.00]	mg/L
Cot	-40.9	1.18	2.89%	[0.00]	mg/L
Crt	243.7	3.27	1.34%	[0.00]	mg/L
Cut	2460.7	60.81	2.47%	[0.00]	mg/L
Fet	0.8	1.07	128.42%	[0.00]	mg/L
Kt	283.3	46.40	16.38%	[0.00]	mg/L
Mgt	-8.1	1.18	14.54%	[0.00]	mg/L
Mnt	188.3	4.28	2.27%	[0.00]	mg/L
Mot	104.4	9.48	9.07%	[0.00]	mg/L
Nat	75.6	2.89	3.82%	[0.00]	mg/L
Nit	-37.9	2.40	6.32%	[0.00]	mg/L
Pbt	2.1	2.01	95.13%	[0.00]	mg/L
Sbt	17.0	0.83	4.86%	[0.00]	mg/L
Set	3.7	1.84	49.75%	[0.00]	mg/L
Tlt	4.3	1.60	36.87%	[0.00]	mg/L
Vt	131.5	29.75	22.62%	[0.00]	mg/L
Znt	150.0	3.97	2.65%	[0.00]	mg/L

Sequence No.: 2
 Sample ID: Standard 2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 11/21/2007 7:35:09 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Standard 2

Analyte Back Pressure Flow
 All 225.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Calib Conc. Units
Sca	378659.1	1487.94	0.39%	89.0 %
Yr	196473.5	105.22	0.05%	95.3 %
Ag†	473967.7	1280.38	0.27%	[2] mg/L
Al†	39672.4	227.63	0.57%	[10] mg/L
As†	15723.4	35.40	0.23%	[10] mg/L
B_†	121536.5	366.54	0.30%	[5.02] mg/L
Ba†	561405.5	1425.60	0.25%	[10] mg/L
Be†	9212937.2	37753.50	0.41%	[4.01] mg/L
Ca†	381367.7	197.34	0.05%	[100] mg/L
Cd†	104134.4	286.93	0.28%	[5.01] mg/L
Co†	195410.1	685.11	0.35%	[10] mg/L
Cr†	602424.7	394.94	0.07%	[9.97] mg/L
Cu†	3539830.3	28901.50	0.82%	[10] mg/L
Fe†	4453.7	5.45	0.12%	[9.98] mg/L
K†	103457.2	1248.01	1.21%	[100] mg/L
Mg†	155171.2	183.70	0.12%	[100] mg/L
Mn†	4387681.8	12370.77	0.28%	[10] mg/L
Mo†	101373.4	198.36	0.20%	[9.98] mg/L
Na†	373519.1	2607.11	0.70%	[100] mg/L
Ni†	165067.8	176.22	0.11%	[10] mg/L
Pb†	35090.8	198.70	0.57%	[10] mg/L
Sb†	16176.3	22.07	0.14%	[10] mg/L
Se†	9697.7	35.78	0.37%	[10] mg/L
Tl†	20587.8	65.40	0.32%	[10] mg/L
V†	1399247.7	3404.02	0.24%	[10] mg/L
Zn†	365051.0	398.43	0.11%	[10] mg/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	237000	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	3967	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	1572	0.00000	1.000000	
B_	1	Lin, Calc Int	0.0	24210	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	56140	0.00000	1.000000	
Be	1	Lin, Calc Int	0.0	2297000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	3814	0.00000	1.000000	
Cd	1	Lin, Calc Int	0.0	20790	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	19540	0.00000	1.000000	
Cr	1	Lin, Calc Int	-0.0	60420	0.00000	1.000000	
Cu	1	Lin, Calc Int	-0.0	354000	0.00000	1.000000	
Fe	1	Lin, Calc Int	0.0	446.3	0.00000	1.000000	
K	1	Lin, Calc Int	0.0	1035	0.00000	1.000000	
Mg	1	Lin, Calc Int	0.0	1552	0.00000	1.000000	
Mn	1	Lin, Calc Int	0.0	438800	0.00000	1.000000	
Mo	1	Lin, Calc Int	0.0	10160	0.00000	1.000000	
Na	1	Lin, Calc Int	-0.0	3735	0.00000	1.000000	
Ni	1	Lin, Calc Int	0.0	16510	0.00000	1.000000	
Pb	1	Lin, Calc Int	0.0	3509	0.00000	1.000000	
Sb	1	Lin, Calc Int	0.0	1618	0.00000	1.000000	
Se	1	Lin, Calc Int	0.0	969.8	0.00000	1.000000	
Tl	1	Lin, Calc Int	0.0	2059	0.00000	1.000000	
V	1	Lin, Calc Int	0.0	139900	0.00000	1.000000	
Zn	1	Lin, Calc Int	0.0	36510	0.00000	1.000000	

Sequence No.: 3
 Sample ID: ICV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 15
 Date Collected: 11/21/2007 7:38:49 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	376749.4	88.6	%	0.22			0.25%
Yr	195185.3	94.7	%	1.08			1.14%
Ag†	474813.0	2.00	mg/L	0.001	2.00 mg/L	0.001	0.07%
	QC value within limits for Ag		Recovery = 100.18%				
Al†	39784.4	10.0	mg/L	0.01	10.0 mg/L	0.01	0.13%
	QC value within limits for Al		Recovery = 100.28%				
As†	15689.8	9.98	mg/L	0.031	9.98 mg/L	0.031	0.31%
	QC value within limits for As		Recovery = 99.78%				
B_†	122399.8	5.04	mg/L	0.008	5.04 mg/L	0.008	0.15%
	QC value within limits for B_		Recovery = 100.71%				
Ba†	564004.8	10.0	mg/L	0.03	10.0 mg/L	0.03	0.25%
	QC value within limits for Ba		Recovery = 100.46%				
Be†	9307059.8	4.05	mg/L	0.019	4.05 mg/L	0.019	0.48%
	QC value within limits for Be		Recovery = 101.27%				
Ca†	380857.1	99.9	mg/L	0.52	99.9 mg/L	0.52	0.53%
	QC value within limits for Ca		Recovery = 99.87%				
Cd†	104586.7	4.90	mg/L	0.014	4.90 mg/L	0.014	0.28%
	QC value within limits for Cd		Recovery = 97.94%				
Co†	196355.9	10.0	mg/L	0.03	10.0 mg/L	0.03	0.31%
	QC value within limits for Co		Recovery = 100.48%				
Cr†	605854.8	10.0	mg/L	0.00	10.0 mg/L	0.00	0.03%
	QC value within limits for Cr		Recovery = 100.27%				
Cu†	3538626.8	10.0	mg/L	0.02	10.0 mg/L	0.02	0.22%
	QC value within limits for Cu		Recovery = 100.06%				
Fe†	4457.6	9.99	mg/L	0.132	9.99 mg/L	0.132	1.32%
	QC value within limits for Fe		Recovery = 99.89%				
K†	103614.3	100	mg/L	0.2	100 mg/L	0.2	0.21%
	QC value within limits for K		Recovery = 100.15%				
Mg†	155665.5	100	mg/L	0.5	100 mg/L	0.5	0.46%
	QC value within limits for Mg		Recovery = 100.32%				
Mn†	4404401.8	10.0	mg/L	0.03	10.0 mg/L	0.03	0.31%
	QC value within limits for Mn		Recovery = 100.38%				
Mo†	101609.6	10.0	mg/L	0.02	10.0 mg/L	0.02	0.20%
	QC value within limits for Mo		Recovery = 100.03%				
Na†	374898.0	100	mg/L	0.2	100 mg/L	0.2	0.19%
	QC value within limits for Na		Recovery = 100.37%				
Ni†	166018.7	10.1	mg/L	0.01	10.1 mg/L	0.01	0.11%
	QC value within limits for Ni		Recovery = 100.58%				
Pb†	34912.7	9.95	mg/L	0.022	9.95 mg/L	0.022	0.22%
	QC value within limits for Pb		Recovery = 99.49%				
Sb†	16072.4	9.78	mg/L	0.005	9.78 mg/L	0.005	0.05%
	QC value within limits for Sb		Recovery = 97.77%				
Se†	9647.5	9.97	mg/L	0.032	9.97 mg/L	0.032	0.33%
	QC value within limits for Se		Recovery = 99.68%				
Tl†	20482.4	9.98	mg/L	0.037	9.98 mg/L	0.037	0.37%
	QC value within limits for Tl		Recovery = 99.77%				
V†	1406276.6	10.1	mg/L	0.01	10.1 mg/L	0.01	0.10%
	QC value within limits for V		Recovery = 101.06%				
Zn†	366445.8	9.97	mg/L	0.017	9.97 mg/L	0.017	0.17%
	QC value within limits for Zn		Recovery = 99.70%				

All analyte(s) passed QC.

Sequence No.: 4
 Sample ID: ICB
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 0
 Date Collected: 11/21/2007 7:42:29 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
 All 225.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc.	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	431402.7	101	%	1.0			0.94%
Yr	208654.0	101	%	0.9			0.85%
Ag†	-9.8	-0.00004	mg/L	0.000066	-0.00004 mg/L	0.000066	158.46%
	QC value within limits for Ag	Recovery = Not calculated					
Al†	-12.5	-0.00315	mg/L	0.000511	-0.00315 mg/L	0.000511	16.21%
	QC value within limits for Al	Recovery = Not calculated					
As†	1.4	0.00088	mg/L	0.000987	0.00088 mg/L	0.000987	111.65%
	QC value within limits for As	Recovery = Not calculated					
B_†	135.6	0.00560	mg/L	0.003463	0.00560 mg/L	0.003463	61.86%
	QC value within limits for B_	Recovery = Not calculated					
Ba†	3.5	0.00006	mg/L	0.000030	0.00006 mg/L	0.000030	49.29%
	QC value within limits for Ba	Recovery = Not calculated					
Be†	39.5	0.00002	mg/L	0.000042	0.00002 mg/L	0.000042	244.17%
	QC value within limits for Be	Recovery = Not calculated					
Ca†	-7.7	-0.00201	mg/L	0.000171	-0.00201 mg/L	0.000171	8.48%
	QC value within limits for Ca	Recovery = Not calculated					
Cd†	-0.9	-0.00006	mg/L	0.000051	-0.00006 mg/L	0.000051	85.77%
	QC value within limits for Cd	Recovery = Not calculated					
Co†	-1.2	-0.00006	mg/L	0.000040	-0.00006 mg/L	0.000040	65.44%
	QC value within limits for Co	Recovery = Not calculated					
Cr†	-5.3	-0.00009	mg/L	0.000137	-0.00009 mg/L	0.000137	155.60%
	QC value within limits for Cr	Recovery = Not calculated					
Cu†	-48.8	-0.00014	mg/L	0.000170	-0.00014 mg/L	0.000170	123.12%
	QC value within limits for Cu	Recovery = Not calculated					
Fe†	-1.9	-0.00435	mg/L	0.002031	-0.00435 mg/L	0.002031	46.64%
	QC value within limits for Fe	Recovery = Not calculated					
K†	-34.9	-0.0337	mg/L	0.02397	-0.0337 mg/L	0.02397	71.10%
	QC value within limits for K	Recovery = Not calculated					
Mg†	-0.9	-0.00055	mg/L	0.002156	-0.00055 mg/L	0.002156	393.27%
	QC value within limits for Mg	Recovery = Not calculated					
Mn†	13.2	0.00003	mg/L	0.000008	0.00003 mg/L	0.000008	26.96%
	QC value within limits for Mn	Recovery = Not calculated					
Mo†	-17.7	-0.00175	mg/L	0.000837	-0.00175 mg/L	0.000837	47.95%
	QC value within limits for Mo	Recovery = Not calculated					
Na†	-33.1	-0.00885	mg/L	0.010047	-0.00885 mg/L	0.010047	113.54%
	QC value within limits for Na	Recovery = Not calculated					
Ni†	4.7	0.00029	mg/L	0.000089	0.00029 mg/L	0.000089	31.22%
	QC value within limits for Ni	Recovery = Not calculated					
Pb†	4.9	0.00139	mg/L	0.000139	0.00139 mg/L	0.000139	10.01%
	QC value within limits for Pb	Recovery = Not calculated					
Sb†	-3.1	-0.00191	mg/L	0.001166	-0.00191 mg/L	0.001166	60.98%
	QC value within limits for Sb	Recovery = Not calculated					
Se†	4.5	0.00463	mg/L	0.002825	0.00463 mg/L	0.002825	60.99%
	QC value within limits for Se	Recovery = Not calculated					
Tl†	-0.9	-0.00043	mg/L	0.002404	-0.00043 mg/L	0.002404	554.95%
	QC value within limits for Tl	Recovery = Not calculated					
V†	7.6	0.00005	mg/L	0.000223	0.00005 mg/L	0.000223	414.35%
	QC value within limits for V	Recovery = Not calculated					
Zn†	-3.4	-0.00010	mg/L	0.000149	-0.00010 mg/L	0.000149	156.23%
	QC value within limits for Zn	Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 5
 Sample ID: QCS1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 11/21/2007 7:46:25 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS1

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: QCS1

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std. Dev.	Sample Conc. Units	Std. Dev.	RSD
Sca	381449.2	89.7 %		0.42			0.47%
Yr	198166.1	96.1 %		1.55			1.62%
Ag†	484781.8	2.05 mg/L		0.000	2.05 mg/L	0.000	0.02%
	QC value within limits for Ag	Recovery = 102.28%					
Al†	40379.3	10.2 mg/L		0.07	10.2 mg/L	0.07	0.64%
	QC value within limits for Al	Recovery = 101.78%					
As†	15416.5	9.80 mg/L		0.113	9.80 mg/L	0.113	1.15%
	QC value within limits for As	Recovery = 98.04%					
B_†	121877.6	5.01 mg/L		0.002	5.01 mg/L	0.002	0.04%
	QC value within limits for B_	Recovery = 100.27%					
Ba†	565502.1	10.1 mg/L		0.01	10.1 mg/L	0.01	0.07%
	QC value within limits for Ba	Recovery = 100.73%					
Be†	9394208.0	4.09 mg/L		0.036	4.09 mg/L	0.036	0.88%
	QC value within limits for Be	Recovery = 102.22%					
Ca†	386070.4	101 mg/L		0.6	101 mg/L	0.6	0.61%
	QC value within limits for Ca	Recovery = 101.23%					
Cd†	85641.8	3.99 mg/L		0.004	3.99 mg/L	0.004	0.11%
	QC value less than the lower limit for Cd	Recovery = 79.78%					
Co†	198346.1	10.2 mg/L		0.01	10.2 mg/L	0.01	0.06%
	QC value within limits for Co	Recovery = 101.50%					
Cr†	606835.8	10.0 mg/L		0.02	10.0 mg/L	0.02	0.22%
	QC value within limits for Cr	Recovery = 100.43%					
Cu†	3548134.4	10.0 mg/L		0.15	10.0 mg/L	0.15	1.52%
	QC value within limits for Cu	Recovery = 100.33%					
Fe†	4556.0	10.2 mg/L		0.06	10.2 mg/L	0.06	0.60%
	QC value within limits for Fe	Recovery = 102.09%					
K†	103699.3	100 mg/L		0.7	100 mg/L	0.7	0.68%
	QC value within limits for K	Recovery = 100.23%					
Mg†	156443.5	101 mg/L		0.5	101 mg/L	0.5	0.54%
	QC value within limits for Mg	Recovery = 100.82%					
Mn†	4429021.2	10.1 mg/L		0.01	10.1 mg/L	0.01	0.09%
	QC value within limits for Mn	Recovery = 100.94%					
Mo†	101650.7	10.0 mg/L		0.01	10.0 mg/L	0.01	0.06%
	QC value within limits for Mo	Recovery = 100.07%					
Na†	379497.3	102 mg/L		0.6	102 mg/L	0.6	0.56%
	QC value within limits for Na	Recovery = 101.60%					
Ni†	167194.2	10.1 mg/L		0.01	10.1 mg/L	0.01	0.06%
	QC value within limits for Ni	Recovery = 101.29%					
Pb†	35525.2	10.1 mg/L		0.10	10.1 mg/L	0.10	0.97%
	QC value within limits for Pb	Recovery = 101.24%					
Sb†	15749.4	9.58 mg/L		0.100	9.58 mg/L	0.100	1.04%
	QC value within limits for Sb	Recovery = 95.77%					
Se†	9825.6	10.2 mg/L		0.05	10.2 mg/L	0.05	0.54%
	QC value within limits for Se	Recovery = 101.52%					
Tl†	21019.7	10.2 mg/L		0.10	10.2 mg/L	0.10	0.96%
	QC value within limits for Tl	Recovery = 102.38%					
V†	1410745.7	10.1 mg/L		0.01	10.1 mg/L	0.01	0.13%
	QC value within limits for V	Recovery = 101.38%					
Zn†	371418.1	10.1 mg/L		0.00	10.1 mg/L	0.00	0.01%
	QC value within limits for Zn	Recovery = 101.05%					
QC Failed. Retry.							

Sequence No.: 6
 Sample ID: QCS1

Autosampler Location: 1
 Date Collected: 11/21/2007 7:48:09 PM

Analyst:
Initial Sample Wt:
Dilution:

Data Type: Original
Initial Sample Vol:
Sample Prep Vol:

Nebulizer Parameters: QCS1

Analyte Back Pressure Flow
All 225.0 kPa 0.65 L/min

Mean Data: QCS1

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	376367.8	88.5 %	0.79			0.89%
Yr	194622.5	94.4 %	0.59			0.63%
Ag†	490762.6	2.07 mg/L	0.002	2.07 mg/L	0.002	0.09%
	QC value within limits for Ag Recovery = 103.54%					
Al†	40343.0	10.2 mg/L	0.01	10.2 mg/L	0.01	0.13%
	QC value within limits for Al Recovery = 101.69%					
As†	15806.0	10.1 mg/L	0.02	10.1 mg/L	0.02	0.21%
	QC value within limits for As Recovery = 100.52%					
B_†	124697.8	5.13 mg/L	0.004	5.13 mg/L	0.004	0.08%
	QC value within limits for B_ Recovery = 102.60%					
Ba†	571432.2	10.2 mg/L	0.02	10.2 mg/L	0.02	0.23%
	QC value within limits for Ba Recovery = 101.79%					
Be†	9653308.0	4.20 mg/L	0.009	4.20 mg/L	0.009	0.21%
	QC value greater than the upper limit for Be Recovery = 105.04%					
Ca†	392461.5	103 mg/L	0.3	103 mg/L	0.3	0.33%
	QC value within limits for Ca Recovery = 102.91%					
Cd†	86546.2	4.03 mg/L	0.013	4.03 mg/L	0.013	0.31%
	QC value less than the lower limit for Cd Recovery = 80.57%					
Co†	200634.2	10.3 mg/L	0.03	10.3 mg/L	0.03	0.28%
	QC value within limits for Co Recovery = 102.67%					
Cr†	615607.7	10.2 mg/L	0.00	10.2 mg/L	0.00	0.00%
	QC value within limits for Cr Recovery = 101.88%					
Cu†	3629156.0	10.3 mg/L	0.03	10.3 mg/L	0.03	0.33%
	QC value within limits for Cu Recovery = 102.62%					
Fe†	4651.2	10.4 mg/L	0.12	10.4 mg/L	0.12	1.16%
	QC value within limits for Fe Recovery = 104.23%					
K†	103291.9	99.8 mg/L	0.35	99.8 mg/L	0.35	0.35%
	QC value within limits for K Recovery = 99.84%					
Mg†	158482.3	102 mg/L	0.0	102 mg/L	0.0	0.03%
	QC value within limits for Mg Recovery = 102.13%					
Mn†	4478877.0	10.2 mg/L	0.02	10.2 mg/L	0.02	0.18%
	QC value within limits for Mn Recovery = 102.08%					
Mo†	103085.3	10.1 mg/L	0.01	10.1 mg/L	0.01	0.11%
	QC value within limits for Mo Recovery = 101.49%					
Na†	377590.1	101 mg/L	0.5	101 mg/L	0.5	0.48%
	QC value within limits for Na Recovery = 101.09%					
Ni†	169403.5	10.3 mg/L	0.00	10.3 mg/L	0.00	0.00%
	QC value within limits for Ni Recovery = 102.63%					
Pb†	36192.8	10.3 mg/L	0.08	10.3 mg/L	0.08	0.82%
	QC value within limits for Pb Recovery = 103.14%					
Sb†	16104.8	9.79 mg/L	0.020	9.79 mg/L	0.020	0.21%
	QC value within limits for Sb Recovery = 97.94%					
Se†	10046.9	10.4 mg/L	0.02	10.4 mg/L	0.02	0.23%
	QC value within limits for Se Recovery = 103.80%					
Tl†	21416.9	10.4 mg/L	0.05	10.4 mg/L	0.05	0.51%
	QC value within limits for Tl Recovery = 104.31%					
V†	1429098.0	10.3 mg/L	0.00	10.3 mg/L	0.00	0.01%
	QC value within limits for V Recovery = 102.70%					
Zn†	376634.5	10.2 mg/L	0.02	10.2 mg/L	0.02	0.19%
	QC value within limits for Zn Recovery = 102.47%					
QC Failed. Continue with analysis.						

Sequence No.: 7
 Sample ID: QCS2
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 11/21/2007 7:51:43 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS2

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: QCS2

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	376962.0	88.6 %	0.06			0.07%
Yr	196153.2	95.2 %	0.32			0.33%
Al†	39390.6	9.93 mg/L	0.078	9.93 mg/L	0.078	0.79%
	QC value within limits for Al Recovery = 99.29%					
B_†	122798.7	5.05 mg/L	0.004	5.05 mg/L	0.004	0.07%
	QC value within limits for B_ Recovery = 101.03%					
Ba†	565049.1	10.1 mg/L	0.02	10.1 mg/L	0.02	0.17%
	QC value within limits for Ba Recovery = 100.65%					
Ca†	385726.9	101 mg/L	0.3	101 mg/L	0.3	0.31%
	QC value within limits for Ca Recovery = 101.14%					
Co†	198665.1	10.2 mg/L	0.03	10.2 mg/L	0.03	0.25%
	QC value within limits for Co Recovery = 101.67%					
Cr†	609254.4	10.1 mg/L	0.00	10.1 mg/L	0.00	0.05%
	QC value within limits for Cr Recovery = 100.83%					
Cu†	3584785.3	10.1 mg/L	0.00	10.1 mg/L	0.00	0.00%
	QC value within limits for Cu Recovery = 101.37%					
Fe†	4521.2	10.1 mg/L	0.04	10.1 mg/L	0.04	0.44%
	QC value within limits for Fe Recovery = 101.31%					
K†	100520.9	97.2 mg/L	0.69	97.2 mg/L	0.69	0.71%
	QC value within limits for K Recovery = 97.16%					
Mg†	155428.8	100 mg/L	0.2	100 mg/L	0.2	0.24%
	QC value within limits for Mg Recovery = 100.17%					
Mn†	4421686.1	10.1 mg/L	0.03	10.1 mg/L	0.03	0.27%
	QC value within limits for Mn Recovery = 100.77%					
Mo†	101929.4	10.0 mg/L	0.02	10.0 mg/L	0.02	0.18%
	QC value within limits for Mo Recovery = 100.35%					
Na†	369186.5	98.8 mg/L	0.47	98.8 mg/L	0.47	0.47%
	QC value within limits for Na Recovery = 98.84%					
Zn†	373421.0	10.2 mg/L	0.00	10.2 mg/L	0.00	0.04%
	QC value within limits for Zn Recovery = 102.29%					

All analyte(s) passed QC.

Sequence No.: 8
 Sample ID: QCS3
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 1
 Date Collected: 11/21/2007 7:55:15 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QCS3

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: QCS3

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	381035.1	89.6 %	0.16			0.17%
Yr	196239.2	95.2 %	0.49			0.52%
Al†	39846.3	10.0 mg/L	0.03	10.0 mg/L	0.03	0.29%
	QC value within limits for Al Recovery = 100.44%					
B_†	122235.0	5.03 mg/L	0.001	5.03 mg/L	0.001	0.02%
	QC value within limits for B_ Recovery = 100.57%					
Ba†	562928.7	10.0 mg/L	0.01	10.0 mg/L	0.01	0.12%
	QC value within limits for Ba Recovery = 100.27%					
Ca†	387277.9	102 mg/L	0.1	102 mg/L	0.1	0.08%
	QC value within limits for Ca Recovery = 101.55%					
Co†	197915.3	10.1 mg/L	0.00	10.1 mg/L	0.00	0.02%
	QC value within limits for Co Recovery = 101.28%					
Cr†	606784.5	10.0 mg/L	0.00	10.0 mg/L	0.00	0.01%
	QC value within limits for Cr Recovery = 100.42%					
Cu†	3540201.0	10.0 mg/L	0.02	10.0 mg/L	0.02	0.18%
	QC value within limits for Cu Recovery = 100.11%					
Fe†	4538.3	10.2 mg/L	0.14	10.2 mg/L	0.14	1.40%
	QC value within limits for Fe Recovery = 101.70%					
K†	101792.7	98.4 mg/L	0.42	98.4 mg/L	0.42	0.43%
	QC value within limits for K Recovery = 98.39%					
Mg†	156113.5	101 mg/L	0.0	101 mg/L	0.0	0.03%
	QC value within limits for Mg Recovery = 100.61%					
Mn†	4390877.4	10.0 mg/L	0.01	10.0 mg/L	0.01	0.07%
	QC value within limits for Mn Recovery = 100.07%					
Mo†	101579.7	10.0 mg/L	0.00	10.0 mg/L	0.00	0.00%
	QC value within limits for Mo Recovery = 100.00%					
Na†	373285.3	99.9 mg/L	0.11	99.9 mg/L	0.11	0.11%
	QC value within limits for Na Recovery = 99.94%					
Zn†	371939.3	10.2 mg/L	0.01	10.2 mg/L	0.01	0.14%
	QC value within limits for Zn Recovery = 101.89%					

All analyte(s) passed QC.

Sequence No.: 9
 Sample ID: LINEARITY
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 9
 Date Collected: 11/21/2007 7:58:48 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	358138.4	84.2 %	%	0.47			0.56%
Yr	189009.2	91.7 %	%	0.26			0.28%
Al†	-75.5	-0.0190 mg/L	mg/L	0.00370	-0.0190 mg/L	0.00370	19.44%
	QC value within limits for Al Recovery = Not calculated						
B_†	481.7	0.0199 mg/L	mg/L	0.00080	0.0199 mg/L	0.00080	4.00%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	106.0	0.00189 mg/L	mg/L	0.000099	0.00189 mg/L	0.000099	5.27%
	QC value within limits for Ba Recovery = Not calculated						
Ca†	1139109.5	299 mg/L	mg/L	2.0	299 mg/L	2.0	0.66%
	QC value within limits for Ca Recovery = 99.56%						
Co†	57.2	0.00293 mg/L	mg/L	0.000000	0.00293 mg/L	0.000000	0.02%
	QC value within limits for Co Recovery = Not calculated						
Cr†	87.9	0.00145 mg/L	mg/L	0.000015	0.00145 mg/L	0.000015	1.07%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-3539.5	-0.01000 mg/L	mg/L	0.000043	-0.01000 mg/L	0.000043	0.43%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	43416.8	97.3 mg/L	mg/L	0.36	97.3 mg/L	0.36	0.37%
	QC value within limits for Fe Recovery = 97.29%						
K†	318238.3	308 mg/L	mg/L	1.5	308 mg/L	1.5	0.49%
	QC value within limits for K Recovery = 102.53%						
Mg†	298673.5	192 mg/L	mg/L	0.8	192 mg/L	0.8	0.41%
	QC value within limits for Mg Recovery = 96.24%						
Mn†	1562.1	0.00356 mg/L	mg/L	0.000040	0.00356 mg/L	0.000040	1.12%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	-5.0	-0.00050 mg/L	mg/L	0.000339	-0.00050 mg/L	0.000339	68.31%
	QC value within limits for Mo Recovery = Not calculated						
Na†	1122430.4	301 mg/L	mg/L	0.7	301 mg/L	0.7	0.23%
	QC value within limits for Na Recovery = 100.17%						
Zn†	923.7	0.0253 mg/L	mg/L	0.00021	0.0253 mg/L	0.00021	0.84%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 10
 Sample ID: ICSA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 11/21/2007 8:02:57 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte Back Pressure Flow
 All 226.0 kPa 0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	371165.9	87.3 %	0.77			0.88%
Yr	189973.1	92.2 %	0.42			0.45%
Al†	993368.7	250 mg/L	2.0	250 mg/L	2.0	0.80%
	QC value within limits for Al Recovery = 100.16%					
B_†	-62.8	-0.00260 mg/L	0.001246	-0.00260 mg/L	0.001246	48.00%
	QC value within limits for B_ Recovery = Not calculated					
Ba†	140.9	0.00251 mg/L	0.000063	0.00251 mg/L	0.000063	2.53%
	QC value within limits for Ba Recovery = Not calculated					
Ca†	978198.3	256 mg/L	0.5	256 mg/L	0.5	0.21%
	QC value within limits for Ca Recovery = 102.60%					
Co†	27.8	0.00142 mg/L	0.000142	0.00142 mg/L	0.000142	9.97%
	QC value within limits for Co Recovery = Not calculated					
Cr†	-44.6	-0.00074 mg/L	0.000359	-0.00074 mg/L	0.000359	48.63%
	QC value within limits for Cr Recovery = Not calculated					
Cu†	-3702.6	-0.0105 mg/L	0.00012	-0.0105 mg/L	0.00012	1.13%
	QC value within limits for Cu Recovery = Not calculated					
Fe†	43578.8	97.7 mg/L	0.84	97.7 mg/L	0.84	0.86%
	QC value within limits for Fe Recovery = 97.65%					
K†	443.6	0.429 mg/L	0.0159	0.429 mg/L	0.0159	3.71%
	QC value within limits for K Recovery = Not calculated					
Mg†	377844.9	244 mg/L	0.4	244 mg/L	0.4	0.17%
	QC value within limits for Mg Recovery = 97.40%					
Mn†	1735.4	0.00396 mg/L	0.000050	0.00396 mg/L	0.000050	1.27%
	QC value within limits for Mn Recovery = Not calculated					
Mo†	-50.3	-0.00495 mg/L	0.000193	-0.00495 mg/L	0.000193	3.89%
	QC value within limits for Mo Recovery = Not calculated					
Na†	424.3	0.114 mg/L	0.0048	0.114 mg/L	0.0048	4.24%
	QC value within limits for Na Recovery = Not calculated					
Zn†	721.5	0.0198 mg/L	0.00055	0.0198 mg/L	0.00055	2.79%
	QC value within limits for Zn Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 11
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 11
 Date Collected: 11/21/2007 8:07:06 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte Back Pressure Flow
 All 226.0 kPa 0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	371836.0	87.4 %		0.57			0.65%
Yr	194608.7	94.4 %		0.30			0.32%
Al†	984298.3	248 mg/L		0.1	248 mg/L	0.1	0.04%
	QC value within limits for Al Recovery = 99.24%						
B_†	-328.0	-0.0140 mg/L		0.00042	-0.0140 mg/L	0.00042	2.97%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	14598.7	0.260 mg/L		0.0013	0.260 mg/L	0.0013	0.52%
	QC value within limits for Ba Recovery = 104.02%						
Ca†	968774.5	254 mg/L		0.6	254 mg/L	0.6	0.25%
	QC value within limits for Ca Recovery = 101.61%						
Co†	4661.1	0.239 mg/L		0.0012	0.239 mg/L	0.0012	0.51%
	QC value within limits for Co Recovery = 95.41%						
Cr†	14866.2	0.246 mg/L		0.0013	0.246 mg/L	0.0013	0.53%
	QC value within limits for Cr Recovery = 98.41%						
Cu†	85795.9	0.243 mg/L		0.0015	0.243 mg/L	0.0015	0.63%
	QC value within limits for Cu Recovery = 97.04%						
Fe†	43748.7	98.0 mg/L		0.10	98.0 mg/L	0.10	0.10%
	QC value within limits for Fe Recovery = 98.03%						
K†	171.9	0.166 mg/L		0.0060	0.166 mg/L	0.0060	3.60%
	QC value within limits for K Recovery = Not calculated						
Mg†	373871.1	241 mg/L		0.3	241 mg/L	0.3	0.12%
	QC value within limits for Mg Recovery = 96.38%						
Mn†	113661.6	0.259 mg/L		0.0005	0.259 mg/L	0.0005	0.18%
	QC value within limits for Mn Recovery = 103.62%						
Mo†	-65.3	-0.00642 mg/L		0.000315	-0.00642 mg/L	0.000315	4.90%
	QC value within limits for Mo Recovery = Not calculated						
Na†	307.0	0.0822 mg/L		0.00283	0.0822 mg/L	0.00283	3.45%
	QC value within limits for Na Recovery = Not calculated						
Zn†	19659.2	0.539 mg/L		0.0032	0.539 mg/L	0.0032	0.59%
	QC value within limits for Zn Recovery = 107.71%						

All analyte(s) passed QC.

Sequence No.: 12
 Sample ID: WASH
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 0
 Date Collected: 11/21/2007 8:11:15 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: WASH

Analyte Back Pressure Flow
 All 225.0 kPa 0.65 L/min

Mean Data: WASH

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	422213.9	99.3	%	0.31			0.32%
Yr	207513.6	101	%	1.1			1.13%
Al†	-34.3	-0.00865	mg/L	0.007248	-0.00865 mg/L	0.007248	83.83%
	QC value within limits for Al Recovery = Not calculated						
B_†	-360.1	-0.0149	mg/L	0.00039	-0.0149 mg/L	0.00039	2.60%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-2.2	-0.00004	mg/L	0.000046	-0.00004 mg/L	0.000046	119.33%
	QC value within limits for Ba Recovery = Not calculated						
Ca†	-10.5	-0.00276	mg/L	0.001487	-0.00276 mg/L	0.001487	53.95%
	QC value within limits for Ca Recovery = Not calculated						
Co†	0.3	0.00002	mg/L	0.000297	0.00002 mg/L	0.000297	>999.9%
	QC value within limits for Co Recovery = Not calculated						
Cr†	1.0	0.00002	mg/L	0.000079	0.00002 mg/L	0.000079	468.11%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	114.6	0.00032	mg/L	0.000085	0.00032 mg/L	0.000085	26.28%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-0.4	-0.00084	mg/L	0.000804	-0.00084 mg/L	0.000804	95.68%
	QC value within limits for Fe Recovery = Not calculated						
K†	-20.7	-0.0200	mg/L	0.03757	-0.0200 mg/L	0.03757	187.44%
	QC value within limits for K Recovery = Not calculated						
Mg†	0.9	0.00058	mg/L	0.001042	0.00058 mg/L	0.001042	178.81%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-6.5	-0.00001	mg/L	0.000025	-0.00001 mg/L	0.000025	166.38%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	-66.6	-0.00656	mg/L	0.000032	-0.00656 mg/L	0.000032	0.48%
	QC value within limits for Mo Recovery = Not calculated						
Na†	-26.3	-0.00705	mg/L	0.002583	-0.00705 mg/L	0.002583	36.62%
	QC value within limits for Na Recovery = Not calculated						
Zn†	-4.0	-0.00011	mg/L	0.000115	-0.00011 mg/L	0.000115	104.78%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 13
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 11/21/2007 8:15:06 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	386803.8	91.0 %	0.22			0.24%
Yr	199906.2	97.0 %	2.20			2.26%
Al†	19888.7	5.01 mg/L	0.021	5.01 mg/L	0.021	0.43%
	QC value within limits for Al Recovery = 100.26%					
B_†	60032.9	2.47 mg/L	0.004	2.47 mg/L	0.004	0.18%
	QC value within limits for B_ Recovery = 98.77%					
Ba†	288201.8	5.13 mg/L	0.004	5.13 mg/L	0.004	0.08%
	QC value within limits for Ba Recovery = 102.67%					
Ca†	195807.6	51.3 mg/L	0.35	51.3 mg/L	0.35	0.69%
	QC value within limits for Ca Recovery = 102.69%					
Co†	101223.5	5.18 mg/L	0.000	5.18 mg/L	0.000	0.00%
	QC value within limits for Co Recovery = 103.60%					
Cr†	307481.3	5.09 mg/L	0.003	5.09 mg/L	0.003	0.06%
	QC value within limits for Cr Recovery = 101.77%					
Cu†	1787762.4	5.06 mg/L	0.023	5.06 mg/L	0.023	0.46%
	QC value within limits for Cu Recovery = 101.11%					
Fe†	2274.8	5.10 mg/L	0.112	5.10 mg/L	0.112	2.19%
	QC value within limits for Fe Recovery = 101.95%					
K†	50222.3	48.5 mg/L	0.14	48.5 mg/L	0.14	0.29%
	QC value within limits for K Recovery = 97.09%					
Mg†	79562.1	51.3 mg/L	0.22	51.3 mg/L	0.22	0.42%
	QC value within limits for Mg Recovery = 102.55%					
Mn†	2272251.5	5.18 mg/L	0.001	5.18 mg/L	0.001	0.02%
	QC value within limits for Mn Recovery = 103.57%					
Mo†	51373.7	5.06 mg/L	0.010	5.06 mg/L	0.010	0.20%
	QC value within limits for Mo Recovery = 101.15%					
Na†	185277.7	49.6 mg/L	0.04	49.6 mg/L	0.04	0.08%
	QC value within limits for Na Recovery = 99.21%					
Zn†	191060.6	5.23 mg/L	0.001	5.23 mg/L	0.001	0.02%
	QC value within limits for Zn Recovery = 104.68%					

All analyte(s) passed QC.

Sequence No.: 49
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 11/21/2007 10:47:02 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	400218.4	94.1 %		1.71			1.81%
Yr	199683.2	96.9 %		1.44			1.49%
Al†	10446.0	2.63 mg/L		0.010	2.63 mg/L	0.010	0.38%
	QC value within limits for Al Recovery = 105.32%						
B_†	30581.2	1.26 mg/L		0.001	1.26 mg/L	0.001	0.06%
	QC value within limits for B_ Recovery = 101.05%						
Ba†	149402.1	2.66 mg/L		0.012	2.66 mg/L	0.012	0.44%
	QC value within limits for Ba Recovery = 106.45%						
Ca†	101887.4	26.7 mg/L		0.06	26.7 mg/L	0.06	0.21%
	QC value within limits for Ca Recovery = 106.87%						
Cr†	159165.2	2.63 mg/L		0.006	2.63 mg/L	0.006	0.23%
	QC value within limits for Cr Recovery = 105.37%						
Cu†	909461.9	2.57 mg/L		0.004	2.57 mg/L	0.004	0.16%
	QC value within limits for Cu Recovery = 102.77%						
Fe†	1207.2	2.71 mg/L		0.039	2.71 mg/L	0.039	1.43%
	QC value within limits for Fe Recovery = 108.20%						
K†	26021.2	25.2 mg/L		0.10	25.2 mg/L	0.10	0.40%
	QC value within limits for K Recovery = 100.61%						
Mg†	41959.8	27.0 mg/L		0.05	27.0 mg/L	0.05	0.20%
	QC value within limits for Mg Recovery = 108.16%						
Mn†	1182887.8	2.70 mg/L		0.010	2.70 mg/L	0.010	0.36%
	QC value within limits for Mn Recovery = 107.84%						
Mo†	26400.9	2.60 mg/L		0.007	2.60 mg/L	0.007	0.28%
	QC value within limits for Mo Recovery = 103.96%						
Na†	95699.5	25.6 mg/L		0.05	25.6 mg/L	0.05	0.21%
	QC value within limits for Na Recovery = 102.48%						
Zn†	99079.4	2.71 mg/L		0.008	2.71 mg/L	0.008	0.31%
	QC value within limits for Zn Recovery = 108.57%						
All analyte(s) passed QC.							

Sequence No.: 50
 Sample ID: MBLANK
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 0
 Date Collected: 11/21/2007 10:51:12 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MBLANK

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: MBLANK

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	430193.6	101 %	%	0.3			0.31%
Yr	204442.8	99.2 %	%	1.30			1.31%
Al†	-24.8	-0.00625 mg/L	mg/L	0.011772	-0.00625 mg/L	0.011772	188.48%
	QC value within limits for Al Recovery = Not calculated						
B_†	-295.0	-0.0122 mg/L	mg/L	0.00021	-0.0122 mg/L	0.00021	1.70%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-1.2	-0.00002 mg/L	mg/L	0.000039	-0.00002 mg/L	0.000039	181.74%
	QC value within limits for Ba Recovery = Not calculated						
Ca†	-18.6	-0.00487 mg/L	mg/L	0.002974	-0.00487 mg/L	0.002974	61.10%
	QC value within limits for Ca Recovery = Not calculated						
Cr†	-2.4	-0.00004 mg/L	mg/L	0.000037	-0.00004 mg/L	0.000037	94.78%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-357.4	-0.00101 mg/L	mg/L	0.000092	-0.00101 mg/L	0.000092	9.13%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-1.0	-0.00234 mg/L	mg/L	0.005349	-0.00234 mg/L	0.005349	228.13%
	QC value within limits for Fe Recovery = Not calculated						
K†	-89.4	-0.0864 mg/L	mg/L	0.04971	-0.0864 mg/L	0.04971	57.52%
	QC value within limits for K Recovery = Not calculated						
Mg†	-3.0	-0.00197 mg/L	mg/L	0.001221	-0.00197 mg/L	0.001221	62.15%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-9.3	-0.00002 mg/L	mg/L	0.000035	-0.00002 mg/L	0.000035	163.32%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	-57.9	-0.00570 mg/L	mg/L	0.000599	-0.00570 mg/L	0.000599	10.51%
	QC value within limits for Mo Recovery = Not calculated						
Na†	-53.2	-0.0142 mg/L	mg/L	0.00268	-0.0142 mg/L	0.00268	18.84%
	QC value within limits for Na Recovery = Not calculated						
Zn†	-46.7	-0.00128 mg/L	mg/L	0.000028	-0.00128 mg/L	0.000028	2.20%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 51
 Sample ID: MRL
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 3
 Date Collected: 11/21/2007 10:55:02 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	423775.2	99.7 %		0.69			0.69%
Yr	204113.1	99.0 %		0.59			0.59%
Al†	170.1	0.0429 mg/L		0.00254	0.0429 mg/L	0.00254	5.92%
	QC value within limits for Al Recovery = 85.76%						
B_†	764.8	0.0316 mg/L		0.00065	0.0316 mg/L	0.00065	2.07%
	QC value within limits for B_ Recovery = 63.18%						
Ba†	1170.9	0.0209 mg/L		0.00017	0.0209 mg/L	0.00017	0.80%
	QC value within limits for Ba Recovery = 90.68%						
Ca†	4029.7	1.06 mg/L		0.000	1.06 mg/L	0.000	0.04%
	QC value within limits for Ca Recovery = 105.66%						
Cr†	576.7	0.00954 mg/L		0.000210	0.00954 mg/L	0.000210	2.20%
	QC value within limits for Cr Recovery = 95.44%						
Cu†	3086.7	0.00872 mg/L		0.000095	0.00872 mg/L	0.000095	1.09%
	QC value within limits for Cu Recovery = 87.20%						
Fe†	5.9	0.0133 mg/L		0.00068	0.0133 mg/L	0.00068	5.08%
	QC value within limits for Fe Recovery = 66.53%						
K†	952.2	0.920 mg/L		0.0391	0.920 mg/L	0.0391	4.25%
	QC value within limits for K Recovery = 92.04%						
Mg†	165.9	0.107 mg/L		0.0025	0.107 mg/L	0.0025	2.34%
	QC value within limits for Mg Recovery = 106.88%						
Mn†	944.0	0.00215 mg/L		0.000039	0.00215 mg/L	0.000039	1.81%
	QC value within limits for Mn Recovery = 107.57%						
Mo†	126.4	0.0124 mg/L		0.00010	0.0124 mg/L	0.00010	0.81%
	QC value within limits for Mo Recovery = 62.22%						
Na†	3532.2	0.946 mg/L		0.0041	0.946 mg/L	0.0041	0.43%
	QC value within limits for Na Recovery = 94.57%						
Zn†	771.5	0.0211 mg/L		0.00028	0.0211 mg/L	0.00028	1.32%
	QC value within limits for Zn Recovery = 105.68%						
All analyte(s) passed QC.							

Sequence No.: 52
 Sample ID: LCS
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 6
 Date Collected: 11/21/2007 10:59:05 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: LCS

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	406585.9	95.6 %		0.07			0.07%
Yr	200202.4	97.1 %		0.47			0.48%
Al†	8083.4	2.04 mg/L		0.005	2.04 mg/L	0.005	0.27%
	QC value within limits for Al Recovery = 101.88%						
B_†	11736.0	0.485 mg/L		0.0077	0.485 mg/L	0.0077	1.59%
	QC value within limits for B_ Recovery = 96.95%						
Ba†	57837.2	1.03 mg/L		0.011	1.03 mg/L	0.011	1.10%
	QC value within limits for Ba Recovery = 103.02%						
Ca†	200847.7	52.7 mg/L		0.16	52.7 mg/L	0.16	0.30%
	QC value within limits for Ca Recovery = 105.33%						
Cr†	62269.4	1.03 mg/L		0.011	1.03 mg/L	0.011	1.04%
	QC value within limits for Cr Recovery = 103.05%						
Cu†	361850.7	1.02 mg/L		0.009	1.02 mg/L	0.009	0.85%
	QC value within limits for Cu Recovery = 102.22%						
Fe†	2399.4	5.38 mg/L		0.031	5.38 mg/L	0.031	0.57%
	QC value within limits for Fe Recovery = 107.53%						
K†	20720.2	20.0 mg/L		0.09	20.0 mg/L	0.09	0.43%
	QC value within limits for K Recovery = 100.14%						
Mg†	32996.6	21.3 mg/L		0.11	21.3 mg/L	0.11	0.50%
	QC value within limits for Mg Recovery = 106.32%						
Mn†	234481.7	0.534 mg/L		0.0001	0.534 mg/L	0.0001	0.03%
	QC value within limits for Mn Recovery = 106.88%						
Mo†	10384.9	1.02 mg/L		0.009	1.02 mg/L	0.009	0.90%
	QC value within limits for Mo Recovery = 102.24%						
Na†	188852.2	50.6 mg/L		0.32	50.6 mg/L	0.32	0.64%
	QC value within limits for Na Recovery = 101.12%						
Zn†	39525.8	1.08 mg/L		0.011	1.08 mg/L	0.011	0.98%
	QC value within limits for Zn Recovery = 108.27%						

All analyte(s) passed QC.

Sequence No.: 53
 Sample ID: LCSD
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 7
 Date Collected: 11/21/2007 11:03:11 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: LCSD

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	401581.2	94.4	%	0.33				0.35%
Yr	198995.7	96.5	%	0.03				0.04%
Al†	8034.2	2.03	mg/L	0.009	2.03	mg/L	0.009	0.43%
	QC value within limits for Al Recovery = 101.26%							
B_†	11818.1	0.488	mg/L	0.0042	0.488	mg/L	0.0042	0.87%
	QC value within limits for B_ Recovery = 97.63%							
Ba†	58269.2	1.04	mg/L	0.011	1.04	mg/L	0.011	1.08%
	QC value within limits for Ba Recovery = 103.79%							
Ca†	202902.7	53.2	mg/L	0.09	53.2	mg/L	0.09	0.16%
	QC value within limits for Ca Recovery = 106.41%							
Cr†	62502.7	1.03	mg/L	0.012	1.03	mg/L	0.012	1.17%
	QC value within limits for Cr Recovery = 103.44%							
Cu†	361161.1	1.02	mg/L	0.004	1.02	mg/L	0.004	0.39%
	QC value within limits for Cu Recovery = 102.03%							
Fe†	2383.7	5.34	mg/L	0.002	5.34	mg/L	0.002	0.03%
	QC value within limits for Fe Recovery = 106.83%							
K†	20884.3	20.2	mg/L	0.08	20.2	mg/L	0.08	0.38%
	QC value within limits for K Recovery = 100.93%							
Mg†	33229.8	21.4	mg/L	0.01	21.4	mg/L	0.01	0.06%
	QC value within limits for Mg Recovery = 107.07%							
Mn†	233784.9	0.533	mg/L	0.0011	0.533	mg/L	0.0011	0.22%
	QC value within limits for Mn Recovery = 106.56%							
Mo†	10434.2	1.03	mg/L	0.011	1.03	mg/L	0.011	1.03%
	QC value within limits for Mo Recovery = 102.72%							
Na†	190937.3	51.1	mg/L	0.06	51.1	mg/L	0.06	0.12%
	QC value within limits for Na Recovery = 102.24%							
Zn†	39914.3	1.09	mg/L	0.010	1.09	mg/L	0.010	0.95%
	QC value within limits for Zn Recovery = 109.34%							
All analyte(s) passed QC.								

Sequence No.: 55
 Sample ID: 2711210438MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 63
 Date Collected: 11/21/2007 11:12:06 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210438MS

Analyte Back Pressure Flow
 All 226.0 kPa 0.65 L/min

Mean Data: 2711210438MS

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		RSD
	Intensity				Conc. Units	Std.Dev.	
Sca	376198.0		88.5 %	0.74			0.83%
Yr	187240.8		90.8 %	0.15			0.17%
Alf	8017.2		2.02 mg/L	0.007	2.02 mg/L	0.007	0.35%
B_t	14128.4		0.584 mg/L	0.0032	0.584 mg/L	0.0032	0.55%
Bat	57629.4		1.03 mg/L	0.005	1.03 mg/L	0.005	0.49%
Caf	265547.3		69.6 mg/L	0.29	69.6 mg/L	0.29	0.42%
Crt	60892.3		1.01 mg/L	0.006	1.01 mg/L	0.006	0.57%
Cuf	356118.3		1.01 mg/L	0.005	1.01 mg/L	0.005	0.51%
Fef	2392.2		5.36 mg/L	0.041	5.36 mg/L	0.041	0.76%
Kf	28454.7		27.5 mg/L	0.12	27.5 mg/L	0.12	0.42%
Mgf	78526.2		50.6 mg/L	0.03	50.6 mg/L	0.03	0.07%
Mnf	231041.5		0.527 mg/L	0.0021	0.527 mg/L	0.0021	0.40%
Mof	9860.8		0.971 mg/L	0.0037	0.971 mg/L	0.0037	0.38%
Naf	856217.5		229 mg/L	0.3	229 mg/L	0.3	0.11%
Znf	41329.7		1.13 mg/L	0.003	1.13 mg/L	0.003	0.24%

Sequence No.: 56
 Sample ID: 2711210438MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 64
 Date Collected: 11/21/2007 11:16:15 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210438MSD

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: 2711210438MSD

Analyte	Mean Corrected		Calib	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	374962.1	88.2 %	%	0.36			0.41%
Yr	190437.2	92.4 %	%	1.55			1.67%
Al†	7931.7	2.00 mg/L	mg/L	0.033	2.00 mg/L	0.033	1.66%
B†	14231.0	0.588 mg/L	mg/L	0.0042	0.588 mg/L	0.0042	0.71%
Ba†	57874.6	1.03 mg/L	mg/L	0.004	1.03 mg/L	0.004	0.39%
Ca†	264694.3	69.4 mg/L	mg/L	0.39	69.4 mg/L	0.39	0.56%
Cr†	61176.5	1.01 mg/L	mg/L	0.004	1.01 mg/L	0.004	0.41%
Cu†	354191.0	1.00 mg/L	mg/L	0.003	1.00 mg/L	0.003	0.33%
Fe†	2359.8	5.29 mg/L	mg/L	0.084	5.29 mg/L	0.084	1.59%
K†	28574.3	27.6 mg/L	mg/L	0.18	27.6 mg/L	0.18	0.65%
Mg†	78453.7	50.6 mg/L	mg/L	0.29	50.6 mg/L	0.29	0.58%
Mn†	230964.4	0.526 mg/L	mg/L	0.0009	0.526 mg/L	0.0009	0.17%
Mo†	9962.3	0.981 mg/L	mg/L	0.0028	0.981 mg/L	0.0028	0.28%
Ni†	857017.8	229 mg/L	mg/L	0.3	229 mg/L	0.3	0.11%
Zn†	41472.3	1.14 mg/L	mg/L	0.005	1.14 mg/L	0.005	0.48%

Sequence No.: 66
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 11/21/2007 11:59:43 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	380192.8	89.4 %	0.07			0.08%
Yr	190236.2	92.3 %	0.14			0.16%
Al†	20164.3	5.08 mg/L	0.019	5.08 mg/L	0.019	0.38%
	QC value within limits for Al Recovery = 101.65%					
B_†	59530.3	2.46 mg/L	0.002	2.46 mg/L	0.002	0.08%
	QC value within limits for B_ Recovery = 98.35%					
Ba†	284961.7	5.08 mg/L	0.001	5.08 mg/L	0.001	0.03%
	QC value within limits for Ba Recovery = 101.52%					
Ca†	196969.4	51.6 mg/L	0.15	51.6 mg/L	0.15	0.29%
	QC value within limits for Ca Recovery = 103.30%					
Crt	304642.4	5.04 mg/L	0.002	5.04 mg/L	0.002	0.05%
	QC value within limits for Cr Recovery = 100.84%					
Cu†	1742299.6	4.92 mg/L	0.080	4.92 mg/L	0.080	1.63%
	QC value within limits for Cu Recovery = 98.44%					
Fe†	2360.0	5.29 mg/L	0.009	5.29 mg/L	0.009	0.17%
	QC value within limits for Fe Recovery = 105.77%					
K†	50267.3	48.6 mg/L	0.29	48.6 mg/L	0.29	0.59%
	QC value within limits for K Recovery = 97.18%					
Mg†	81282.0	52.4 mg/L	0.10	52.4 mg/L	0.10	0.18%
	QC value within limits for Mg Recovery = 104.76%					
Mn†	2229753.2	5.08 mg/L	0.033	5.08 mg/L	0.033	0.66%
	QC value within limits for Mn Recovery = 101.64%					
Mo†	50735.9	4.99 mg/L	0.010	4.99 mg/L	0.010	0.20%
	QC value within limits for Mo Recovery = 99.90%					
Na†	184367.4	49.4 mg/L	0.12	49.4 mg/L	0.12	0.25%
	QC value within limits for Na Recovery = 98.72%					
Zn†	189126.9	5.18 mg/L	0.004	5.18 mg/L	0.004	0.07%
	QC value within limits for Zn Recovery = 103.62%					
All analyte(s) passed QC.						

Sequence No.: 67
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 0
 Date Collected: 11/22/2007 12:03:52 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	412724.0	97.1 %		0.66			0.68%
Yr	196809.8	95.5 %		0.58			0.61%
Al†	-21.6	-0.00544 mg/L		0.000699	-0.00544 mg/L	0.000699	12.85%
	QC value within limits for Al Recovery = Not calculated						
B_†	-99.4	-0.00411 mg/L		0.000400	-0.00411 mg/L	0.000400	9.75%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	2.4	0.00004 mg/L		0.000041	0.00004 mg/L	0.000041	96.50%
	QC value within limits for Ba Recovery = Not calculated						
Ca†	-20.6	-0.00540 mg/L		0.000591	-0.00540 mg/L	0.000591	10.94%
	QC value within limits for Ca Recovery = Not calculated						
Cr†	-4.4	-0.00007 mg/L		0.000036	-0.00007 mg/L	0.000036	49.34%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-346.4	-0.00098 mg/L		0.000027	-0.00098 mg/L	0.000027	2.79%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-3.0	-0.00675 mg/L		0.001845	-0.00675 mg/L	0.001845	27.36%
	QC value within limits for Fe Recovery = Not calculated						
K†	-54.3	-0.0525 mg/L		0.00449	-0.0525 mg/L	0.00449	8.56%
	QC value within limits for K Recovery = Not calculated						
Mg†	-1.2	-0.00076 mg/L		0.000589	-0.00076 mg/L	0.000589	77.45%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-18.8	-0.00004 mg/L		0.000008	-0.00004 mg/L	0.000008	18.22%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	-48.5	-0.00477 mg/L		0.000522	-0.00477 mg/L	0.000522	10.94%
	QC value within limits for Mo Recovery = Not calculated						
Na†	-98.8	-0.0265 mg/L		0.00492	-0.0265 mg/L	0.00492	18.58%
	QC value within limits for Na Recovery = Not calculated						
Zn†	-44.6	-0.00122 mg/L		0.000087	-0.00122 mg/L	0.000087	7.11%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 69
 Sample ID: 2711210440MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 75
 Date Collected: 11/22/2007 12:12:31 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210440MS

Analyte	Back Pressure	Flow
All	227.0 kPa	0.65 L/min

Mean Data: 2711210440MS

Analyte	Mean Corrected		Calib Conc. Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity				Conc. Units			
Sca	366210.3		86.1 %	0.72				0.84%
Yr	184312.0		89.4 %	0.24				0.27%
Alf	8429.4		2.12 mg/L	0.000	2.12 mg/L	0.000		0.02%
B ₁ f	14443.6		0.597 mg/L	0.0009	0.597 mg/L	0.0009		0.15%
Ba ₁ f	59627.3		1.06 mg/L	0.005	1.06 mg/L	0.005		0.45%
Ca ₁ f	254547.3		66.7 mg/L	0.11	66.7 mg/L	0.11		0.17%
Cr ₁ f	62669.2		1.04 mg/L	0.002	1.04 mg/L	0.002		0.21%
Cu ₁ f	365301.1		1.03 mg/L	0.004	1.03 mg/L	0.004		0.34%
Fe ₁ f	2474.8		5.55 mg/L	0.019	5.55 mg/L	0.019		0.35%
K ₁ f	28429.9		27.5 mg/L	0.07	27.5 mg/L	0.07		0.25%
Mg ₁ f	67220.4		43.3 mg/L	0.01	43.3 mg/L	0.01		0.02%
Mn ₁ f	243174.5		0.554 mg/L	0.0032	0.554 mg/L	0.0032		0.57%
Mo ₁ f	10234.7		1.01 mg/L	0.006	1.01 mg/L	0.006		0.57%
Na ₁ f	754676.0		202 mg/L	1.0	202 mg/L	1.0		0.49%
Zn ₁ f	43499.9		1.19 mg/L	0.003	1.19 mg/L	0.003		0.23%

Sequence No.: 70
 Sample ID: 2711210440MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 76
 Date Collected: 11/22/2007 12:16:40 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210440MSD

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: 2711210440MSD

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	375749.7	88.4 %	0.19			0.21%
Yr	189675.9	92.0 %	0.75			0.82%
Alf	8197.6	2.07 mg/L	0.025	2.07 mg/L	0.025	1.22%
B_f	14276.7	0.590 mg/L	0.0001	0.590 mg/L	0.0001	0.01%
Bat	58325.8	1.04 mg/L	0.002	1.04 mg/L	0.002	0.19%
Ca†	249085.8	65.3 mg/L	0.07	65.3 mg/L	0.07	0.11%
Crf	61473.5	1.02 mg/L	0.000	1.02 mg/L	0.000	0.04%
Cut	357623.4	1.01 mg/L	0.000	1.01 mg/L	0.000	0.00%
Fet	2439.6	5.47 mg/L	0.053	5.47 mg/L	0.053	0.96%
K†	27883.3	27.0 mg/L	0.16	27.0 mg/L	0.16	0.59%
Mgt	65906.4	42.5 mg/L	0.13	42.5 mg/L	0.13	0.31%
Mnt	236954.8	0.540 mg/L	0.0015	0.540 mg/L	0.0015	0.28%
Mot	10031.6	0.988 mg/L	0.0033	0.988 mg/L	0.0033	0.34%
Naf	741943.1	199 mg/L	0.5	199 mg/L	0.5	0.27%
Znt	42650.6	1.17 mg/L	0.001	1.17 mg/L	0.001	0.09%

Sequence No.: 79
 Sample ID: 2711200049
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 85
 Date Collected: 11/22/2007 12:58:21 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711200049

Analyte Back Pressure Flow
 All 227.0 kPa 0.65 L/min

Mean Data: 2711200049

Analyte	Mean Corrected Intensity	Conc. Units	Calib Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	389322.9	91.5 %		0.42			0.46%
Yr	189819.7	92.1 %		0.15			0.17%
Alf	107.2	0.0270 mg/L		0.00066	0.0270 mg/L	0.00066	2.45%
B_f	3233.8	0.134 mg/L		0.0006	0.134 mg/L	0.0006	0.47%
Bat	2313.5	0.0412 mg/L		0.00048	0.0412 mg/L	0.00048	1.16%
Cat	163510.1	42.9 mg/L		0.21	42.9 mg/L	0.21	0.48%
Crt	-53.1	-0.00088 mg/L		0.000122	-0.00088 mg/L	0.000122	13.85%
Cut	-254.7	-0.00072 mg/L		0.000074	-0.00072 mg/L	0.000074	10.33%
Fet	-2.2	-0.00484 mg/L		0.003682	-0.00484 mg/L	0.003682	76.13%
Kf	3127.0	3.02 mg/L		0.006	3.02 mg/L	0.006	0.18%
Mgt	25897.4	16.7 mg/L		0.11	16.7 mg/L	0.11	0.63%
Mnt	94.6	0.00022 mg/L		0.000017	0.00022 mg/L	0.000017	7.84%
Mof	-48.9	-0.00481 mg/L		0.000677	-0.00481 mg/L	0.000677	14.06%
Nat	202399.3	54.2 mg/L		0.08	54.2 mg/L	0.08	0.14%
Znt	95.2	0.00261 mg/L		0.000079	0.00261 mg/L	0.000079	3.01%

Sequence No.: 80
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 11/22/2007 1:03:06 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	228.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	374007.4	87.9 %	1.15			1.30%
Yr	182528.9	88.6 %	0.05			0.06%
Al†	20410.4	5.14 mg/L	0.009	5.14 mg/L	0.009	0.18%
	QC value within limits for Al Recovery = 102.89%					
B_†	60012.5	2.48 mg/L	0.003	2.48 mg/L	0.003	0.13%
	QC value within limits for B_ Recovery = 99.15%					
Ba†	286236.7	5.10 mg/L	0.011	5.10 mg/L	0.011	0.22%
	QC value within limits for Ba Recovery = 101.97%					
Ca†	199311.6	52.3 mg/L	0.01	52.3 mg/L	0.01	0.02%
	QC value within limits for Ca Recovery = 104.52%					
Cr†	307451.2	5.09 mg/L	0.002	5.09 mg/L	0.002	0.05%
	QC value within limits for Cr Recovery = 101.77%					
Cu†	1758230.9	4.97 mg/L	0.042	4.97 mg/L	0.042	0.85%
	QC value within limits for Cu Recovery = 99.34%					
Fe†	2409.0	5.40 mg/L	0.007	5.40 mg/L	0.007	0.13%
	QC value within limits for Fe Recovery = 107.96%					
K†	51405.3	49.7 mg/L	0.10	49.7 mg/L	0.10	0.20%
	QC value within limits for K Recovery = 99.37%					
Mg†	82363.3	53.1 mg/L	0.01	53.1 mg/L	0.01	0.02%
	QC value within limits for Mg Recovery = 106.16%					
Mn†	2272456.2	5.18 mg/L	0.021	5.18 mg/L	0.021	0.41%
	QC value within limits for Mn Recovery = 103.58%					
Mo†	51028.8	5.02 mg/L	0.006	5.02 mg/L	0.006	0.12%
	QC value within limits for Mo Recovery = 100.47%					
Na†	186891.3	50.0 mg/L	0.27	50.0 mg/L	0.27	0.54%
	QC value within limits for Na Recovery = 100.07%					
Zn†	190580.7	5.22 mg/L	0.010	5.22 mg/L	0.010	0.20%
	QC value within limits for Zn Recovery = 104.41%					

All analyte(s) passed QC.

Sequence No.: 81
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 0
 Date Collected: 11/22/2007 1:07:15 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	227.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Conc. Units	Calib	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	403197.7	94.8 %	%	0.25			0.26%
Yr	191909.7	93.1 %	%	0.93			1.00%
Al†	35.1	0.00885 mg/L	mg/L	0.008783	0.00885 mg/L	0.008783	99.26%
	QC value within limits for Al Recovery = Not calculated						
B_†	-28.5	-0.00118 mg/L	mg/L	0.000731	-0.00118 mg/L	0.000731	62.00%
	QC value within limits for B_ Recovery = Not calculated						
Ba†	-1.5	-0.00003 mg/L	mg/L	0.000026	-0.00003 mg/L	0.000026	96.67%
	QC value within limits for Ba Recovery = Not calculated						
Ca†	-24.3	-0.00636 mg/L	mg/L	0.001002	-0.00636 mg/L	0.001002	15.74%
	QC value within limits for Ca Recovery = Not calculated						
Cr†	6.3	0.00010 mg/L	mg/L	0.000041	0.00010 mg/L	0.000041	39.87%
	QC value within limits for Cr Recovery = Not calculated						
Cu†	-359.1	-0.00101 mg/L	mg/L	0.000022	-0.00101 mg/L	0.000022	2.15%
	QC value within limits for Cu Recovery = Not calculated						
Fe†	-3.0	-0.00669 mg/L	mg/L	0.007008	-0.00669 mg/L	0.007008	104.71%
	QC value within limits for Fe Recovery = Not calculated						
K†	-57.2	-0.0553 mg/L	mg/L	0.04114	-0.0553 mg/L	0.04114	74.39%
	QC value within limits for K Recovery = Not calculated						
Mg†	-2.7	-0.00175 mg/L	mg/L	0.001080	-0.00175 mg/L	0.001080	61.60%
	QC value within limits for Mg Recovery = Not calculated						
Mn†	-22.2	-0.00005 mg/L	mg/L	0.000008	-0.00005 mg/L	0.000008	14.96%
	QC value within limits for Mn Recovery = Not calculated						
Mo†	-46.8	-0.00460 mg/L	mg/L	0.000095	-0.00460 mg/L	0.000095	2.06%
	QC value within limits for Mo Recovery = Not calculated						
Na†	-36.9	-0.00987 mg/L	mg/L	0.001687	-0.00987 mg/L	0.001687	17.08%
	QC value within limits for Na Recovery = Not calculated						
Zn†	-47.5	-0.00130 mg/L	mg/L	0.000048	-0.00130 mg/L	0.000048	3.65%
	QC value within limits for Zn Recovery = Not calculated						

All analyte(s) passed QC.

Sequence No.: 82
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 11/22/2007 1:11:06 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	228.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Calib Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	389026.3	91.5 %	0.94			1.02%
Yr	187722.8	91.1 %	0.59			0.65%
Al†	10515.7	2.65 mg/L	0.002	2.65 mg/L	0.002	0.07%
	QC value within limits for Al Recovery = 106.03%					
B_†	30302.7	1.25 mg/L	0.015	1.25 mg/L	0.015	1.24%
	QC value within limits for B_ Recovery = 100.13%					
Ba†	147302.0	2.62 mg/L	0.036	2.62 mg/L	0.036	1.36%
	QC value within limits for Ba Recovery = 104.95%					
Ca†	103569.0	27.2 mg/L	0.26	27.2 mg/L	0.26	0.98%
	QC value within limits for Ca Recovery = 108.63%					
Cr†	157106.4	2.60 mg/L	0.039	2.60 mg/L	0.039	1.49%
	QC value within limits for Cr Recovery = 104.00%					
Cu†	892863.4	2.52 mg/L	0.008	2.52 mg/L	0.008	0.30%
	QC value within limits for Cu Recovery = 100.89%					
Fe†	1233.9	2.76 mg/L	0.014	2.76 mg/L	0.014	0.52%
	QC value greater than the upper limit for Fe Recovery = 110.60%					
K†	26123.6	25.3 mg/L	0.09	25.3 mg/L	0.09	0.35%
	QC value within limits for K Recovery = 101.00%					
Mg†	42944.8	27.7 mg/L	0.02	27.7 mg/L	0.02	0.08%
	QC value greater than the upper limit for Mg Recovery = 110.70%					
Mn†	1167748.7	2.66 mg/L	0.003	2.66 mg/L	0.003	0.11%
	QC value within limits for Mn Recovery = 106.46%					
Mo†	25942.1	2.55 mg/L	0.027	2.55 mg/L	0.027	1.07%
	QC value within limits for Mo Recovery = 102.16%					
Na†	95511.2	25.6 mg/L	0.09	25.6 mg/L	0.09	0.35%
	QC value within limits for Na Recovery = 102.28%					
Zn†	97636.9	2.67 mg/L	0.030	2.67 mg/L	0.030	1.11%
	QC value within limits for Zn Recovery = 106.98%					
QC Failed. Continue with analysis.						

Sequence No.: 54
 Sample ID: 2711210438
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 62
 Date Collected: 11/21/2007 11:07:17 PM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210438

Analyte	Back Pressure	Flow
All	225.0 kPa	0.65 L/min

Mean Data: 2711210438

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	389403.5	91.6	%	0.31				0.33%
Yr	195258.2	94.7	%	0.63				0.66%
Alf	26.6	0.00670	mg/L	0.000039	0.00670	mg/L	0.000039	0.58%
B_f	1757.5	0.0726	mg/L	0.00037	0.0726	mg/L	0.00037	0.50%
Baf	124.1	0.00221	mg/L	0.000025	0.00221	mg/L	0.000025	1.13%
Ca_f	69655.6	18.3	mg/L	0.01	18.3	mg/L	0.01	0.08%
Crt	-2.6	-0.00004	mg/L	0.000132	-0.00004	mg/L	0.000132	307.18%
Cuf	965.5	0.00273	mg/L	0.000058	0.00273	mg/L	0.000058	2.14%
Fef	11.4	0.0255	mg/L	0.00351	0.0255	mg/L	0.00351	13.76%
Kf	8120.2	7.85	mg/L	0.081	7.85	mg/L	0.081	1.03%
Mgf	47369.2	30.5	mg/L	0.06	30.5	mg/L	0.06	0.19%
Mnf	384.3	0.00088	mg/L	0.000002	0.00088	mg/L	0.000002	0.20%
Mof	-61.5	-0.00606	mg/L	0.000121	-0.00606	mg/L	0.000121	2.00%
Naf	701407.2	188	mg/L	1.2	188	mg/L	1.2	0.64%
Znf	325.7	0.00892	mg/L	0.000063	0.00892	mg/L	0.000063	0.70%

Sequence No.: 68
 Sample ID: 2711210440
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 74
 Date Collected: 11/22/2007 12:07:43 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711210440

Analyte	Back Pressure	Flow
All	226.0 kPa	0.65 L/min

Mean Data: 2711210440

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	389561.2	91.6	%	0.99				1.08%
Yr	193479.2	93.9	%	1.36				1.45%
Alf	137.2	0.0346	mg/L	0.00263	0.0346	mg/L	0.00263	7.61%
B_†	1702.8	0.0703	mg/L	0.00116	0.0703	mg/L	0.00116	1.64%
Bat	66.2	0.00118	mg/L	0.000099	0.00118	mg/L	0.000099	8.42%
Cat	51773.3	13.6	mg/L	0.17	13.6	mg/L	0.17	1.28%
Crt	-10.9	-0.00018	mg/L	0.000029	-0.00018	mg/L	0.000029	16.38%
Cut	-47.0	-0.00013	mg/L	0.000150	-0.00013	mg/L	0.000150	112.76%
Fef	44.0	0.0986	mg/L	0.00238	0.0986	mg/L	0.00238	2.41%
K†	7043.6	6.81	mg/L	0.177	6.81	mg/L	0.177	2.60%
Mgt	34389.7	22.2	mg/L	0.32	22.2	mg/L	0.32	1.42%
Mnt	4901.9	0.0112	mg/L	0.00011	0.0112	mg/L	0.00011	0.96%
Mot	-67.5	-0.00664	mg/L	0.000113	-0.00664	mg/L	0.000113	1.70%
Nat	570940.3	153	mg/L	2.0	153	mg/L	2.0	1.33%
Znt	1177.4	0.0323	mg/L	0.00007	0.0323	mg/L	0.00007	0.22%

Sequence No.: 78
 Sample ID: 2711200573_10X
 Analyst:
 Initial Sample Wt:
 Dilution: 10X

Autosampler Location: 84
 Date Collected: 11/22/2007 12:53:31 AM
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2711200573_10X
 Analyte Back Pressure Flow
 All 228.0 kPa 0.65 L/min

Mean Data: 2711200573_10X

Analyte	Mean Corrected		Calib Units	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc.			Conc.	Units		
Sca	381830.5	89.8	%	1.28				1.43%
Yr	185306.2	89.9	%	1.34				1.49%
Alf	-27.6	-0.00695	mg/L	0.004604	-0.0695	mg/L	0.04604	66.27%
B_t	7376.9	0.305	mg/L	0.0004	3.05	mg/L	0.004	0.14%
Bat	118.6	0.00211	mg/L	0.000007	0.0211	mg/L	0.00007	0.33%
Ca†	156246.5	41.0	mg/L	0.28	410	mg/L	2.8	0.69%
Crt	81.7	0.00135	mg/L	0.000029	0.0135	mg/L	0.00029	2.17%
Cuf	-333.9	-0.00094	mg/L	0.000042	-0.00943	mg/L	0.000416	4.41%
Fet	-1.8	-0.00405	mg/L	0.007177	-0.0405	mg/L	0.07177	177.37%
K†	2661.7	2.57	mg/L	0.042	25.7	mg/L	0.42	1.64%
Mgf	29381.4	18.9	mg/L	0.34	189	mg/L	3.4	1.77%
Mnt	39390.0	0.0898	mg/L	0.00037	0.898	mg/L	0.0037	0.41%
Mot	-5.7	-0.00056	mg/L	0.000329	-0.00558	mg/L	0.003286	58.87%
Na†	484447.5	130	mg/L	0.6	1300	mg/L	5.5	0.43%
Znt	-139.4	-0.00382	mg/L	0.000029	-0.0382	mg/L	0.00029	0.75%

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Lot #
 HNO3 R# 1004132 HCL R# 100427
 IS = Yttrium(ME0709008)0.75mL + Scandium (ME0710007)0.5mL to 1000mL w/ 2% HNO3

Standards	Lot #	Exp. Date	Dilution
Calibration	ME0704023	(05/01/08)	1:10 ME0704027
(Prepare daily)	ME0704024	(05/01/08)	1:10
			CCV/ECV MCV
CCV/MCV/ECV	ME0710008	(04/17/09)	1:20 ME0710008 1:40 ME0710008
(Prepare daily)			
Spike/LCS	ME0606004	(12/13/07)	1:100 ME0601006
(Prepare daily)	ME0710001	(04/02/08)	1:100
	ME0708001	(02/03/09)	1:200
MRL	ME0709021	(03/20/08)	1:100 ME0603015
(Prepare daily)			
ICSA	ME0709017	(03/20/08)	
ICSAB	ME0709019	(03/20/08)	
QCS	ME0610005	(04/10/08)	
Linearity	ME0711002	(05/06/08)	
Method Sr/Ti/Sn/SiO2			
Calibration	ME0708004	(02/07/08)	
CCV/ECV	ME0711003	(02/06/08)	
QCS	ME0711004	(02/06/08)	
Spike/LCS	ME0709014	(11/30/07)	1:100
(Prepare daily)			
MRL	ME0709013	(11/30/07)	1:100
(Prepare daily)			
Method Li			
Std/ICV/MRL	ME0711001	(05/01/08)	1:1000, 200, 40, 10
(Prepare daily)			
QCS	ME0707002	(01/10/08)	1:1000
(Prepare daily)			
LCS/Spike	ME0707002	(01/10/08)	1:50
(Prepare daily)			
ccv	ME0707002	(01/10/08)	1:40
(Prepare daily)			

From May 2005: the calibration std for ICP should be ME0505010,011 not ME0408010
 dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.

ME0704027

Initial:
Date:

wjh
4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration STD
Date Received/Prepped: Prep Daily
Date Expired: 5/1/2008
Manufacturer: MWH-wbh
Matrix: 2% HNO₃ + 5% HCl
Amount:

ME #: 07040027
By: wbh
Lot #:
Certificate: NO
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	1:10 ME0704024	10 ug/ml
Sb		10 ug/ml
Sn		10 ug/ml
Ti		10 ug/ml
B		5 ug/ml
Ca	1:10 ME0704023	100 ug/ml
K		100 ug/ml
Mg		100 ug/ml
Na		100 ug/ml
Al		10 ug/ml
As		10 ug/ml
Ba		10 ug/ml
Co		10 ug/ml
Cr		10 ug/ml
Cu		10 ug/ml
Fe		10 ug/ml
Mn		10 ug/ml
Ni		10 ug/ml
Pb		10 ug/ml
Se		10 ug/ml
Tl		10 ug/ml
V		10 ug/ml
Zn		10 ug/ml
Cd		5 ug/ml
Be		4 ug/ml
SR		3 ug/ml
Ag		2 ug/ml

ME0704024

Initial:
Date:

wbh
4/23/07

METALS STANDARD DOCUMENTATION

Standard:	ICPCalibration Stock Std #2	ME #: 0704024
Date Received/Prepped:	4/23/2007	By: wbh
Date Expired:	5/1/2008	Lot #: A2-MEB235011
Manufacturer:	Inorganic Ventures	Certificate: Y
Matrix:	5% Nitric Acid + Trace HF	NIST SRM: Varies
Amount:	500 mL	Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml

ME0704023

Initial:
Date:

wbh
4/23/07

METALS STANDARD DOCUMENTATION

Standard: ICPCalibration Stock Std #1 **ME #:** 0704023
Date Received/Prepped: 4/23/2007 **By:** wbh
Date Expired: 5/1/2008 **Lot #:** A2-MEB235010
Manufacturer: Inorganic Ventures **Certificate:** Y
Matrix: 5% Nitric Acid **NIST SRM:** Varies
Amount: 500 mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		100 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		50 ug/ml
Be		40 ug/ml
SR		30 ug/ml
Ag		20 ug/ml

19115 Highway Avenue, Suite 4
 Columbia, Maryland 21046
 410-301-1000

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-2
 Lot Number: **A2-MEB235011**
 Matrix: tr. HF, 5% HNO₃(abs)

M70704024

100.00 µg/mL each:
 Mo, Sb, Sn, Ti,
 50.00 µg/mL each:
 B

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	50.07 ± 0.28 µg/mL	Molybdenum, Mo	100.3 ± 0.3 µg/mL
Tin, Sn	100.3 ± 0.3 µg/mL	Titanium, Ti	100.5 ± 0.2 µg/mL		

Certified Density: 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

\bar{x} = mean
 x_i = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

$\sum s_i$ = The summation of all significant estimated errors (Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

"Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std' is specified.

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principals."

2.0 DESCRIPTION OF CRM Custom Solution
 Catalog No.: MWH-ICAP-CAL-1
 Lot Number: **A2-MEB235010**
 Matrix: 5% HNO₃(abs)

M80704023

1,000.00 µg/mL each:
 Ca, K, Mg, Na,
 100.00 µg/mL each:
 Al, As, Ba, Co, Cr₃, Cu, Fe, Mn, Ni, Pb, Se, Tl, V, Zn,
 50.00 µg/mL each:
 Cd,
 40.00 µg/mL each:
 Be,
 30.00 µg/mL each:
 Sr,
 20.00 µg/mL each:
 Ag

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.1 ± 0.5 µg/mL	Arsenic, As	100.2 ± 0.3 µg/mL	Barium, Ba	99.9 ± 0.2 µg/mL
Beryllium, Be	39.98 ± 0.08 µg/mL	Cadmium, Cd	50.05 ± 0.12 µg/mL	Calcium, Ca	997 ± 3 µg/mL
Chromium+3, Cr ₃	100.1 ± 0.4 µg/mL	Cobalt, Co	100.1 ± 0.2 µg/mL	Copper, Cu	100.1 ± 0.2 µg/mL
Iron, Fe	100.0 ± 0.2 µg/mL	Lead, Pb	100.1 ± 0.3 µg/mL	Magnesium, Mg	996 ± 3 µg/mL
Manganese, Mn	100.1 ± 0.3 µg/mL	Nickel, Ni	100.1 ± 0.2 µg/mL	Potassium, K	1,003 ± 2 µg/mL
Selenium, Se	100.1 ± 0.2 µg/mL	Silver, Ag	20.03 ± 0.06 µg/mL	Sodium, Na	997 ± 5 µg/mL
Strontium, Sr	29.92 ± 0.18 µg/mL	Thallium, Tl	100.0 ± 0.1 µg/mL	Vanadium, V	100.1 ± 0.3 µg/mL
Zinc, Zn	100.1 ± 0.4 µg/mL				

Certified Density: 1.056 g/mL (measured at 22° C)

Initial:

SE

Date:

10/19/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV/QCS Stock Standard
Date Received/Prepped: 10/19/2007
Date Expired: 4/17/2009
Manufacturer: CPI
Matrix: 5% HNO3 = tr HF
Amount: 100 mL x 10

ME #: 0710008
By: STE
Lot #: 07J154
Certificate: Y
NIST SRM: Various
Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	P/N 4400-061003RH01	20 ppm
Al		100 ppm
As		100 ppm
B		50 ppm
Ba		100 ppm
bE		40 ppm
Ca		1000 ppm
Cd		50 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		1000 ppm
Mg		1000 ppm
Mn		100 ppm
Mo		100 ppm
Na		1000 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		20 ppm
Sn		20 ppm
Ti		20 ppm



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 4/17/2009

Certificate of Analysis

Part Number: 4400-061003RH01
Lot Number: 07J154
Shelf Life: 18 months

ME 0710008

REC'D: 10/19/CA

STE

MWH
 Custom Multi
 5% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

STE

Date:

10/2/07

METALS STANDARD DOCUMENTATION

Standard: ICP Spike Solution
Date Received/Prepped: 10/2/2007
Date Expired: 4/2/2008
Manufacturer: MWH-STE
Matrix: H2O
Amount: 125mL

ME #: 0710001
By: STE
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
AS	8.0mL ME0611005/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

METALS STANDARD DOCUMENTATION

wbh
6/19/06

Standard:	ICP/ICPMS LCS/SPIKE Solution	ME #: 0606004
Date Received/Prepped:	6/19/2006	By: wbh
Date Expired:	12/13/2006	Lot #: 06F103
Manufacturer:	CPI	Certificate: Y
Matrix:	5% HNO ₃ + 0.1% HF	NIST SRM: 3100 Series
Amount:	10 x 100 mL	Storage: Room Temp

Component	Comment	Conc. Unit:
Iron	CPI P/N: 4400-050314RH01	500 mg/L
Aluminum		200 mg/L
Barium		100 mg/L
Cobalt		100 mg/L
Chromium		100 mg/L
Copper		100 mg/L
Molybdenum		100 mg/L
Strontium		100 mg/L
Titanium		100 mg/L
Vanadium		100 mg/L
Zinc		100 mg/L
Tin		100 mg/L
Silver		50 mg/L
Boron		50 mg/L
Manganese		50 mg/L
Nickel		50 mg/L
Antimony		50 mg/L
Arsenic		20 mg/L
Cadmium		20 mg/L
Lead		20 mg/L
Selenium		20 mg/L
Thallium		20 mg/L
Uraium		20 mg/L
Beryllium		5 mg/L
Tin		100 mg/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

Expiry: 12/13/2007

Certificate of Analysis

Part Number: 4400-050314RH01
Lot Number: 06F103
Shelf Life: 18 months

M70 606004

MWH Labs
 5% HNO₃ + 0.1% HF
 #REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

WBS
11/1/06

METALS STANDARD DOCUMENTATION

Standard: Arsenic Stock Std
Date Received/Prepped: 11/1/2006
Date Expired: 10/1/2007
Manufacturer: IV
Matrix: 1.4% HNO3
Amount: 100mL

ME #: 0611005
By: WBH
Lot #: Y-AS02029
Certificate: Y
NIST SRM:
Storage: Room Temp

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
AS	Cat # CGAS1-1	1004 ppm

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM **1000 µg/mL Arsenic in 1.4% (abs) HNO3**

Catalog Number: CGAS1-1, CGAS1-2, and CGAS1-5
 Lot Number: **Y -AS02029**
 Starting Material: As Polycrystalline lump
 Starting Material Purity (%): 99.999055
 Starting Material Lot No: 23115
 Matrix: 1.4% (abs) HNO3

MT0611005

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Concentration: 1004 ± 2 µg/mL 995 ± 2 µg/g

Certified Density: 1.009 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_1}{n}$$

(\bar{x}) = mean
 x_1 = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_1)^2]}{(n)}^{1/2}$$

$\sum s_1$ = The summation of all significant estimated errors
 (Most common are the errors from instrumental measurement weighting, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

• "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)

• This IV product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 Assay Method #1 **1004 ± 2 µg/mL 995 ± 2 µg/g (Avg 2 runs)**
 ICP Assay NIST SRM 3103a Lot Number: 010713

Assay Method #2 **1003 ± 5 µg/mL 994 ± 5 µg/g**
 Gravimetric NIST SRM Lot Number: See Sec. 4.2

Initial:

WBH

Date:

4/16/07

METALS STANDARD DOCUMENTATION

Standard: Lead Stock Standard
Date Received/Prepped: 4/16/2007
Date Expired: 10/11/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0704013
By: WBH
Lot #: 07A097
Certificate: Y
NIST SRM: 3128
Room temp. storage

Component	Comment	Conc. Unit:
Pb	P/N S4400-1000281	1000 ppm

Initial:

WBY

Date:

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Thallium 1000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0702006
By: WBH
Lot #: 06H213
Certificate: Y
NIST SRM: 3158
Room temp. storage

Component	Comment	Conc. Unit:
TI	P/N S4400-1000581	1000 ppm

Aug 16 08



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

M70702006

CERTIFICATE OF ANALYSIS

P/N 4400-1000581

P/N S4400-1000581

Single Element Thallium Standard

Tl in 2% HNO₃

1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal
Source Purity: 99.999%
Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>												
Al	13.3	0.1	Cu	9.3	0.1	Pb	41	0.1	K	ND	70	Tl	X	0.1
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1	Th	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1	Tm	ND	0.1
Ba	0.37	0.1	Eu	ND	0.1	Mg	1.7	0.2	Rh	ND	0.1	Sn	ND	0.1
Be	0.67	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1	Ti	0.45	0.1
Bi	0.12	0.1	Ga	ND	0.1	Hg	0.16	0.2	Ru	ND	0.1	W	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.21	0.1	Sm	ND	0.1	U	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6	V	ND	1
Cd	1.6	0.1	Hf	ND	0.1	Ni	1.1	0.1	Si	46	8	Yb	ND	0.1
Ca	51	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.3	0.1	Y	ND	0.1
Ce	ND	0.1	I	0.4	0.2	Os	ND	0.1	Na	3.3	1	Zn	14.7	2
Cs	0.24	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1	Zr	ND	0.1
Cr	ND	1	Fe	ND	30	P	20	10	Ta	ND	0.1			
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1			

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.



Initial:
Date:

W/BY
3/5/07

METALS STANDARD DOCUMENTATION

Standard: Selenium Stock Standard
Date Received/Prepped: 3/5/2007
Date Expired: 8/22/2008
Manufacturer: CPI
Matrix: 2% HNO₃
Amount: 100 mL

ME #: 0703001
By: wbh
Lot #: 6.00E+228
Certificate: Y
NIST SRM: 3148
Storage: Room Temp

Component	Comment	Conc. Unit:
Se	P/N # S4400-1000491	1000 ppm

Initial:
Date:

DYH
8/8/07

METALS STANDARD DOCUMENTATION

Standard: Dat MW Standard
Date Received/Prepped: 8/8/2007
Date Expired: 2/3/2009
Manufacturer: CPI International
Matrix: 5% HNO₃
Amount: 100mL

ME #: 0708001
By: DYH
Lot #: 07H025
Certificate:
NIST SRM:
Storage: Room Temp

Component	Comment	Conc. Unit:
Na	P/N 4400-130309	10000 ug/mL
Ca		10000 ug/mL
Mg		4000 ug/mL
K		4000 ug/mL



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

Innovative Solutions
in Analytical Science and
Technology

ME0708001

CERTIFICATE OF ANALYSIS

P/N: 4400-130309
Lot Number: 07H025
Shelf Life: 18 months
Expiration Date: 02/03/2009

MWH
Dat MW Standard
 $\mu\text{g/mL} \pm 0.5\%$ in 5% HNO_3

Na 10,000 Ca 10,000 Mg 4,000 K 4,000

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at $1000\mu\text{g/mL}$ by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: STE
 Date: 9/20/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSA^B) ME #: 0709019
 Date Received/Prepped: 9/20/2007 By: STE
 Date Expired: 3/20/2008 Lot #:
 Manufacturer: MWH-STE Certificate:
 Matrix: 5% HNO₃ NIST SRM:
 Amount: 500 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	250 ppm
Ca	P/N 4400-INTB1-100	250 ppm
Fe		100 ppm
Mg		250 ppm
Ag		0.5 ppm
Ba		0.25 ppm
Be		0.25 ppm
Cd		0.5 ppm
Component		0.25 ppm
Cr		0.25 ppm
Cu		0.25 ppm
Mn		0.25 ppm
Ni		0.5 ppm
Pb		0.5 ppm
V		0.25 ppm
Zn		0.5 ppm

Initial:
Date:

STE
9/20/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSA) **ME #:** 0709017
Date Received/Prepped: 9/20/2007 **By:** Ste
Date Expired: 3/20/2008 **Lot #:**
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	250 ppm
Ca		250 ppm
Fe		100 ppm
Mg		250 ppm

ME0705010

Initial:

STE

Date:

8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Analytes B Solution
Date Received/Prepped: 8/27/2007
Date Expired: 2/15/2009
Manufacturer: CPI International
Matrix: 5% HNO₃
Amount: 100 mL

ME #: 0708010
By: STE
Lot #: 07c256
Certificate:
NIST SRM:
Storage: Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ag	P/N 4400-INTB1-100	100 ug/L
Ba		50 ug/L
Be		50 ug/L
Cd		100 ug/L
Co		50 ug/L
Cr		50 ug/L
Cu		50 ug/L
Mn		50 ug/L
Ni		100 ug/L
Pb		100 ug/L
V		50 ug/L
Zn		100 ug/L



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com



*Innovative Solutions
 in Analytical Science and
 Technology*

ME 070 800

CERTIFICATE OF ANALYSIS

P/N 4400-INTB1-100

CLP Analytes B Solution
 in 5% HNO₃

Lot # 07c256

Material Source: Metals and Salts
 Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag	100	Ba	50	Be	50	Cd	100
Co	50	Cr	50	Cu	50	Mn	50
Ni	100	Pb	100	V	50	Zn	100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: STE
Date: 8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Interferents A Solution **ME #:** 0708009
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 2/15/2009 **Lot #:** 07E175
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL **Storage:** Room Temp.

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	5000 ug/mL
Ca		5000 ug/mL
Fe		2000 ug/mL
Mg		5000 ug/mL

FEB 15 09



USA
5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE
P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

*Innovative Solutions
in Analytical Science and
Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500
CLP Interferents A Solution
in 5% HNO₃

ME 0708009

Lot # 07E175

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: STE
Date: 11/6/07

METALS STANDARD DOCUMENTATION

Standard: ICP LINEARITY CHECK
Date Received/Prepped: 11/6/2007
Date Expired: ^{STE} ~~5/6/2008~~ ^{MIC/CP} 2/6/2008
Manufacturer: MWH-STE
Matrix: 5% HNO₃
Amount: 500 mL

ME #: 0711002
By: STE
Lot #: VARIOUS
Certificate:
NIST SRM:
Storage: Room Temp.

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	15.0 mL ME0702002/ 500 mL	300 ppm
K	15.0 mL ME0702005/ 500 mL	300 ppm
Mg	10.0 mL ME0702004/ 500 mL	200 ppm
Na	15 mL ME0702003/ 500 mL	300 ppm
Fe	5.0 mL ME0701008/ 500 mL	100 ppm

Initial:

WZU

Date:

1/27/07

METALS STANDARD DOCUMENTATION

Standard: FE 10000ppm Stock Std
Date Received/Prepped: 1/26/2007
Date Expired: 7/19/2008
Manufacturer: CPI
Matrix: 4% HNO3
Amount: 100 mL

ME #: 701008
By: WBH
Lot #: 061143
Certificate:
NIST SRM: 3126a
Storage: Room Temp

Component	Comment	Conc. Unit:
Fe	PN4400-10M261	10000 PPM



USA
 5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

Innovative Solutions
 JUL 19 1998 in Analytical Science and
 Technology

CERTIFICATE OF ANALYSIS

HTE090100X

P/N 4400-10M261
P/N S4400-10M261
 Single-Element Iron Standard
 Fe in 4% HNO₃
 10,000 ± 30 µg/mL

Lot # 06I143

Material Source: Iron Metal
 Source Purity: 99.999%
 Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>								
Al INT	0.1	Cu 6.4	0.1	Pb ND	0.1	K ND	70	Tl 0.18	0.1
Sb 0.35	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba ND	0.1	Eu ND	0.1	Mg 1.3	0.2	Rh ND	0.1	Sn 0.67	0.1
Be ND	0.1	Gd ND	0.1	Mn INT	1	Rb ND	0.1	Ti 0.21	0.1
Bi ND	0.1	Ga 0.41	0.1	Hg ND	0.2	Ru ND	0.1	W 0.13	0.1
B ND	4	Ge INT	0.1	Mo 4.9	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 9.3	0.1	Si INT	8	Yb ND	0.1
Ca 15	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.34	0.2	Os ND	0.1	Na 8	1	Zn 8.6	2
Cs 0.34	0.1	Ir ND	0.1	Pd ND	0.1	Sr ND	0.1	Zr ND	0.1
Cr 3.3	1	Fe X	30	P 28	10	Ta ND	0.1		
Co 12	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std **ME #:** 0702005
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B056
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO₃ **NIST SRM:** 3141
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

AUG 16 '08



USA

5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

M20702005

P/N 4400-10M411
P/N S4400-10M411
 Single-Element Potassium Standard
 K in 1% HNO₃
 10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)
 Source Purity: 99.999%
 Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>									
Al	0.39	0.1	Cu	0.16	0.1	Pb	ND	0.1	K	X	70
Sb	0.34	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.14	0.1	Eu	ND	0.1	Mg	2.6	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	0.93	1	Rb	9.5	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	20
Ca	82	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	19	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1
									Tl	ND	0.1
									Th	ND	0.1
									Tm	ND	0.1
									Sn	0.17	0.1
									Ti	ND	0.1
									W	ND	0.1
									U	ND	0.1
									V	ND	1
									Yb	ND	0.1
									Y	ND	0.1
									Zn	2.9	1
									Zr	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

72

Initial:
Date:

WBH
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std **ME #:** 0702004
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B058
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3131
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



USA
 5580 Skyline Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

APR 16 08

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-10M311

P/N S4400-10M311

Single-Element Magnesium Standard

Mg in 4% HNO₃

10,000 ± 30 µg/mL

Lot # 07B058

M70702004

Material Source: Magnesium Metal
 Source Purity: 99.99%
 Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>									
Al	28	0.1	Cu	1.6	0.1	Pb	7.7	0.7	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	0.28	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.28	0.1	Eu	ND	0.1	Mg	X	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	0.23	0.1	Mn	19.8	1	Rb	ND	0.1
Bi	ND	0.1	Ga	0.18	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	1.1	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	1	0.1	Si	64	20
Ca	ND	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.19	0.1
Ce	2.1	0.1	I	1	0.2	Os	ND	0.1	Na	7.2	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	0.19	0.1
Cr	ND	1	Fe	80	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	0.76	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

UBJ
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std
Date Received/Prepped: 2/20/2007
Date Expired: 8/16/2008
Manufacturer: CPI
Matrix: 1% HNO₃
Amount: 250 mL

ME #: 0702003
By: WBH
Lot #: 07B057
Certificate: Y
NIST SRM: 3152a
Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 15 03



USA
 5580 Skytane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
 www.cpiinternational.com Fax 707.545.7901

EUROPE
 P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

*Innovative Solutions
 in Analytical Science and
 Technology*

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

M70702003

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO₃)
 Source Purity: 99.99%
 Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.


The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

<u>ppb</u>	<u>DL</u>								
Al 1.5	0.1	Cu 0.45	0.1	Pb ND	0.1	K ND	70	Ti ND	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.13	0.1	Eu ND	0.1	Mg 2.3	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 0.4	0.1	Si 50	8	Yb ND	0.1
Ca 120	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I ND	0.2	Os ND	0.1	Na X	1	Zn 2.9	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 1	0.1	Zr ND	0.1
Cr ND	1	Fe ND	30	P 18	10	Ta ND	0.1		
Co ND	0.1	La ND	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.



Initial: WBH
Date: 2/20/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 10000ppm Stock Std **ME #:** 0702002
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B065
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3109a
Amount: 250 mL Room temp. storage

<u>Component</u>	<u>Comment</u>	<u>Conc. Unit:</u>
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

180702602

Single-Element Calcium Standard
Ca in 4% HNO₃
10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)
Source Purity: 99.997%
Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL	ppb	DL	ppb	DL	ppb	DL	ppb	DL
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Ti 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Tl ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 11.6	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.