

TABLE OF CONTENTS

COVER PAGE	1
QC CHECKLIST.....	2
SUMMARY SHEET	3
RUNLOG.....	5
INITIAL CALIBRATION.....	7
PERIODIC QC.....	24
QC: (MBLANK, LCS1, LCS2)	29
SAMPLE (2805280392).....	32
QC: (MS/MSD 2805280392)	33
PERIODIC QC.....	35
SAMPLE (2805270410).....	40
QC: (MS/MSD 2805270410)	41
PERIODIC QC.....	43
SAMPLE (2805280376).....	47
ANALYTICAL SEQUENCE	48
QC CHECKLIST.....	49
SUMMARY SHEET	50
RUNLOG.....	53
INITIAL CALIBRATION.....	55
PERIODIC QC.....	66
QC: (MBLANK, LCS1, LCS2)	72
SAMPLE (2805290427).....	79
QC: (MS/MSD 2805290427)	80
SAMPLE (2805310013).....	82
QC: (MS/MSD 2805310013)	83
PERIODIC QC.....	85
SAMPLE (2805280375).....	88
CLOSING QC	89
ANALYTICAL SEQUENCE	93
STANDARDS PREPARATION WORKSHEET AND CERTIFICATES OF ANALYSIS.	95

Level IV Data Package

MWH Group 242355

Method: EPA 200.7

2805280375

2805280376

EPA 200.7/6010B QC Check List

Analyst CGK Analysis Date 5-29-08 Reviewer/Date 5-30-08

Instrument Perkin Elmer Optima 4300DV

- All sample analyzed within 6 month holding time
 All sample raw concentration below the high standard or linear range or marked for dilution and rerun

Initial and closing QC

- ICV within +/- 5%
 Linearity check +/- 10%
 ICSAB +/- 20%
 QCS +/- 5%
 MRL +/- 50%

Middle, closing and batch QC

- FilterCheck < 1/2 MRL
 MBLANK < 1/2 MRL
 LCS +/- 15%
 MS/MSD +/- 30% (200.7) +/- 25% (6010B)
 CCV/MCV/ECV +/- 10%
 ICB/CCB/ECB < 1/2 MRL
 CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/- 20%
 RPD between LCS/LCSD is within +/- 20%
 Internal standards 60 TO 125%
 All pH of the samples are < 2

No more than 20 samples per batch

MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

NA QIR needed for failed QC

NA Special Det Code noted on the cover sheet

NA R value for multi point calibration is > 0.995

NP Proper MRL check ran for special low MRL samples

Reagent and Standards used for
Optima 4300 DV
Updated 03/31/08

Int: CGK
Date: 5-30-08

Method 200.7/6010

ICP SUMMARY SHEET

File ID: 080529A

Date Started: 5/29/08

Analyst ID: CSK

SAMPLE ID

LINEARITY	(20:47)	Wash	(20:59)	FILTERCHECK	(21:32)
2805280392	(21:45)	2805280378	(21:57)	2805290001	(22:01)
2805280348	(22:17)	2805280351	(22:20)	2805280489	(22:24)
2805280354	(22:28)	2805280395	(22:32)	2805280219	(22:36)
2805280272	(22:40)	2805270410	(22:45)	2805280273	(23:10)
2805280341	(23:14)	2805280217	(23:18)	2805280342	(23:21)
2805280241	(23:25)	2805280242	(23:29)	2805280213	(23:34)
2805280477	(23:38)	2805280376_2	(23:42)	2805270411	(0:09)
2805280243	(0:21)	2805280244	(0:25)	2805280245	(0:29)
2805280246	(0:33)	2805280247	(0:51)	2805280406	(0:55)
2805280413	(1:00)	2805280095	(1:04)	2805280096	(1:08)
2805280415	(1:11)	2805280097	(1:23)	2805280099	(1:27)
2805280100	(1:41)	2805280253	(1:44)		

VB-Me (4/08)

COMMENT:

Analyst: CSK
5-30-08

Approved By: WSR

BATCH NUMBER for 080529A

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2805280392

2805280392	2805280378	2805290001
2805280348	2805280351	2805280489
2805280354	2805280395	2805280219
2805280272	2805270410	2805280273
2805280341	2805280217	2805280342
2805280241	2805280242	2805280213
2805280477	2805280376_2X	

Batch ID: 2805270411

2805270411	2805280243	2805280244
2805280245	2805280246	2805280247
2805280406	2805280413	2805280095
2805280096	2805280415	2805280097
2805280099	2805280100	2805280253

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	5/29/08	20:44	1	10.053	10.1	95-105	100%
LINEARITY	5/29/08	20:47	1	0.0033	.0033		
ICSA	5/29/08	20:51	1	0.0007	0.0007	80-120	
ICSAB	5/29/08	20:55	1	.25386	.254	80-120	101%
Wash	5/29/08	20:59	1	0.0001	0.0001		
QC-25 1ppm	5/29/08	21:02	1	1.0103	1.0		
CCV	5/29/08	21:06	1	5.0282	5.03	90-110	100%
ICB	5/29/08	21:12	1	0.0000	0.0000		
MRL	5/29/08	21:16	1	0.0103	.0103	50-150	102%
MRL/2	5/29/08	21:19	1	0.0049	.0049		
MCV	5/29/08	21:24	1	2.4864	2.49	90-110	99.4%
FILTERCHECK	5/29/08	21:32	1	0.0000	0		
MBLANK	5/29/08	21:36	1	0.0000	0.0000		
LCS	5/29/08	21:39	1	.93172	.932✓	85-115	93.1%
LCSD	5/29/08	21:42	1	.91490	.915✓	85-115	91.4%
2805280392	5/29/08	21:45	1	-0.0002	ND		
2805280392MS	5/29/08	21:49	1	.94384	.944	✓ [0.944]	94.3%
2805280392MSD	5/29/08	21:53	1	.96400	.964	✓ [0.964]	96.4%
2805280392T	5/29/08	21:53	1		1.00	70 - 130	
2805280378	5/29/08	21:57	1	-0.0003	ND		
2805290001	5/29/08	22:01	1	0.0075	.0075		
CCV	5/29/08	22:04	1	5.0474	5.05	90-110	100%
CCB	5/29/08	22:13	1	0.0001	0.0001		
2805280348	5/29/08	22:17	1	-0.0002	ND		
2805280351	5/29/08	22:20	1	0.0001	0.0000		
2805280489	5/29/08	22:24	1	-0.0004	ND		
2805280354	5/29/08	22:28	1	0.0009	0.0009		
2805280395	5/29/08	22:32	1	-0.0003	ND		
2805280219	5/29/08	22:36	1	0.0013	.0013		
2805280272	5/29/08	22:40	1	0.0020	.002		
2805270410	5/29/08	22:45	1	-0.0003	ND		
2805270410MS	5/29/08	22:49	1	.94509	.945	✓ [0.945]	94.5%
2805270410MSD	5/29/08	22:53	1	.94593	.946	✓ [0.946]	94.5%
2805270410T	5/29/08	22:53	1		1.00	70 - 130	
CCV	5/29/08	22:56	1	4.9778	4.98	90-110	99.5%
CCB	5/29/08	23:03	1	0.0000	0.0000		
MCV	5/29/08	23:06	1	2.5050	2.51	90-110	100%
2805280273	5/29/08	23:10	1	0.0016	.0016		
2805280341	5/29/08	23:14	1	0.0024	.0024		
2805280217	5/29/08	23:18	1	-0.0001	ND		
2805280342	5/29/08	23:21	1	.01842	.018		
2805280241	5/29/08	23:25	1	0.0079	.0079		
2805280242	5/29/08	23:29	1	0.0040	.004		
2805280213	5/29/08	23:34	1	-0.0002	ND		
2805280477	5/29/08	23:38	1	-0.0000	ND		
2805280376_2X	5/29/08	23:42	2	.13753	.140✓		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
MBLANK	5/29/08	23:46	1	0.0003	0.0003		
CCV	5/29/08	23:50	1	5.0047	5.00	/ 90-110	100%
CCB	5/29/08	23:56	1	0.0001	0.0001		
MRL	5/29/08	23:59	1	0.0104	.0104	50-150	103%
LCS	5/30/08	0:03	1	.93335	.933	85-115	93.3%
LCSD	5/30/08	0:06	1	.93428	.934	85-115	93.4%
2805270411	5/30/08	0:09	1	-0.0002	ND		
2805270411MS	5/30/08	0:14	1	.94341	.943	[0.943]	94.3%
2805270411MSD	5/30/08	0:17	1	.94519	.945	[0.945]	94.5%
2805270411T	5/30/08	0:17	1		1.00	70 - 130	
2805280243	5/30/08	0:21	1	0.0010	0.0009		
2805280244	5/30/08	0:25	1	-0.0000	ND		
2805280245	5/30/08	0:29	1	0.0002	0.0001		
2805280246	5/30/08	0:33	1	0.0043	.0043		
CCV	5/30/08	0:38	1	4.9545	4.95	/ 90-110	99.0%
CCB	5/30/08	0:44	1	0.0000	0.0000		
MCV	5/30/08	0:47	1	2.4678	2.47	90-110	98.7%
2805280247	5/30/08	0:51	1	0.0011	.0011		
2805280406	5/30/08	0:55	1	-0.0003	ND		
2805280413	5/30/08	1:00	1	0.0020	.002		
2805280095	5/30/08	1:04	1	0.0008	0.0008		
2805280096	5/30/08	1:08	1	0.0000	0		
2805280415	5/30/08	1:11	1	0.0013	.0013		
2805280415MS	5/30/08	1:16	1	.93875	.939	[0.937]	93.7%
2805280415MSD	5/30/08	1:19	1	.93527	.935	[0.934]	93.3%
2805280415T	5/30/08	1:19	1		1.00	70 - 130	
2805280097	5/30/08	1:23	1	0.0010	0.0009		
2805280099	5/30/08	1:27	1	-0.0001	ND		
CCV	5/30/08	1:31	1	4.8077	4.81	/ 90-110	96.1%
CCB	5/30/08	1:37	1	0.0000	0.0000		
2805280100	5/30/08	1:41	1	0.0001	0.0001		
2805280253	5/30/08	1:44	1	-0.0001	ND		
ECV	5/30/08	1:48	1	4.7930	4.79	/ 90-110	95.8%
ECB	5/30/08	1:54	1	0.0001	0.0001		

Analysis Begun

Start Time: 5/29/2008 17:55:51 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Method Loaded

Method Name: 200.766010_080304
 IEC File: IEC080304.iec
 Method Description: 200.766010_080304

Method Last Saved: 5/2/2008 09:27:36
 MSF File:

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Mo	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a

Sequence No.: 1
 Sample ID: Calib Blank 1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 5/29/2008 17:56:09
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1
 Analyte Back Pressure Flow
 All 213.0 kPa 0.65 L/min

Analyte	Mean Corrected		RSD	Calib	
	Intensity	Std.Dev.		Conc.	Units
Sca	409823.9	5363.41	1.31%	100	%
Yr	505336.2	2250.10	0.45%	100	%
Agt	214.0	2.91	1.36%	[0.00]	mg/L
Alt	39.4	12.58	31.92%	[0.00]	mg/L
Ast	0.6	0.68	114.50%	[0.00]	mg/L
B _t	200.1	0.70	0.35%	[0.00]	mg/L
Bat	-26.6	1.29	4.84%	[0.00]	mg/L
Bet	-3428.0	53.19	1.55%	[0.00]	mg/L
Caf	3742.3	18.04	0.48%	[0.00]	mg/L

Cdt	34.7	0.15	0.42%	[0.00]	mg/L
Cot	-45.0	5.91	13.12%	[0.00]	mg/L
Crt	96.9	4.77	4.92%	[0.00]	mg/L
Cut	3513.1	16.40	0.47%	[0.00]	mg/L
Fet	-110.2	1.44	1.31%	[0.00]	mg/L
Kt	219.7	16.87	7.68%	[0.00]	mg/L
Mgt	-548.1	4.44	0.81%	[0.00]	mg/L
Mnt	125.5	12.64	10.07%	[0.00]	mg/L
Mof	13.6	1.00	7.34%	[0.00]	mg/L
Nat	-137.5	6.37	4.63%	[0.00]	mg/L
Nit	-57.6	6.74	11.69%	[0.00]	mg/L
Pbt	-21.3	2.64	12.38%	[0.00]	mg/L
Sbt	5.9	2.63	44.81%	[0.00]	mg/L
Set	0.6	0.63	106.39%	[0.00]	mg/L
Tlt	-24.6	4.29	17.43%	[0.00]	mg/L
Vt	147.3	8.50	5.77%	[0.00]	mg/L
Znt	80.7	17.87	22.14%	[0.00]	mg/L

Analysis Begun

Start Time: 5/29/2008 20:28:14 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Method Loaded

Method Name: 200.7&6010_080304
 IEC File: IEC080304.iec
 Method Description: 200.7/6010_080304

Method Last Saved: 5/2/2008 09:27:36
 MSF File:

Analyte	Calibration Equation	Processing	View	Internal Standard	IEC
Ag	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Al	Lin, Calc Int	Peak Area	Radial	Yr	Yes
As	Lin, Calc Int	Peak Area	Axial	Sca	Yes
B	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ba	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Be	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Ca	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Cd	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Co	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cr	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Cu	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Fe	Lin, Calc Int	Peak Area	Radial	Yr	Yes
K	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mg	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Mn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Mo	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Na	Lin, Calc Int	Peak Area	Radial	Yr	Yes
Ni	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Pb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sb	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Se	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Tl	Lin, Calc Int	Peak Area	Axial	Sca	Yes
V	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Zn	Lin, Calc Int	Peak Area	Axial	Sca	Yes
Sca	Lin, Calc Int	Peak Area	Axial	n/a	n/a
Yr	Lin, Calc Int	Peak Area	Radial	n/a	n/a

Sequence No.: 1
 Sample ID: Calib Blank 1
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 5/29/2008 20:28:41
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: Calib Blank 1

Analyte	Mean Corrected		RSD	Conc.	Units
	Intensity	Std.Dev.			
Sca	405411.9	3293.05	0.81%	100	%
Yr	495652.7	379.44	0.08%	100	%
Ag†	309.1	27.45	8.88%	[0.00]	mg/L
Al†	35.0	11.02	31.49%	[0.00]	mg/L
As†	1.3	0.63	47.02%	[0.00]	mg/L
B†	264.6	1.34	0.51%	[0.00]	mg/L
Ba†	-22.5	1.56	6.94%	[0.00]	mg/L
Be†	-3334.8	1.39	0.04%	[0.00]	mg/L
Cat	3605.3	14.40	0.40%	[0.00]	mg/L

Cdt	33.1	0.69	2.10%	[0.00] mg/L
Cof	-43.1	0.20	0.48%	[0.00] mg/L
Crt	102.2	2.88	2.81%	[0.00] mg/L
Cut	3373.7	38.30	1.14%	[0.00] mg/L
Fet	-100.7	4.07	4.04%	[0.00] mg/L
Kt	196.3	22.11	11.26%	[0.00] mg/L
Mgt	-531.3	1.72	0.32%	[0.00] mg/L
Mnt	229.3	2.91	1.27%	[0.00] mg/L
Mot	14.6	5.94	40.71%	[0.00] mg/L
Nat	-203.5	5.94	2.92%	[0.00] mg/L
Nit	-50.8	1.30	2.57%	[0.00] mg/L
Pbt	-18.6	4.49	24.12%	[0.00] mg/L
Sbt	4.3	2.14	49.75%	[0.00] mg/L
Set	-2.9	1.96	67.20%	[0.00] mg/L
Tlt	-27.4	0.75	2.72%	[0.00] mg/L
Vt	133.8	10.16	7.59%	[0.00] mg/L
Znt	84.1	5.11	6.08%	[0.00] mg/L

Analysis Begun

Start Time: 5/29/2008 20:38:50 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 1 Autosampler Location: 0
 Sample ID: Calib Blank 1 Date Collected: 5/29/2008 20:38:52
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: Calib Blank 1

Analyte	Mean Corrected	Calib			
	Intensity	Std.Dev.	RSD	Conc.	Units
Sca	404017.9	2306.37	0.57%	100	%
Yr	491376.8	8698.23	1.77%	100	%
Agt	239.2	54.78	22.90%	[0.00]	mg/L
Alt	49.3	13.11	26.56%	[0.00]	mg/L
Ast	1.4	6.66	464.52%	[0.00]	mg/L
B_f	267.7	5.35	2.00%	[0.00]	mg/L
Bat	-22.7	2.30	10.13%	[0.00]	mg/L
Bet	-3325.8	21.58	0.65%	[0.00]	mg/L
Cat	3603.3	45.40	1.26%	[0.00]	mg/L
Cdt	36.7	0.10	0.27%	[0.00]	mg/L
Cot	-45.5	4.15	9.12%	[0.00]	mg/L
Crt	102.1	1.09	1.07%	[0.00]	mg/L
Cut	3383.1	11.32	0.33%	[0.00]	mg/L
Fet	-101.8	5.83	5.73%	[0.00]	mg/L
Kt	210.5	13.47	6.40%	[0.00]	mg/L
Mgt	-538.8	14.26	2.65%	[0.00]	mg/L
Mnt	217.7	8.33	3.83%	[0.00]	mg/L
Mot	15.5	0.55	3.55%	[0.00]	mg/L
Nat	-207.1	62.66	30.27%	[0.00]	mg/L
Nit	-52.4	5.38	10.28%	[0.00]	mg/L
Pbt	-19.8	0.46	2.32%	[0.00]	mg/L
Sbt	2.9	2.35	80.93%	[0.00]	mg/L
Set	3.2	1.24	38.53%	[0.00]	mg/L
Tlt	-25.1	1.04	4.12%	[0.00]	mg/L
Vt	139.1	1.39	1.00%	[0.00]	mg/L
Znt	86.2	5.34	6.19%	[0.00]	mg/L

User canceled analysis.

Analysis Begun

Start Time: 5/29/2008 20:42:28 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 2 Autosampler Location: 15
 Sample ID: Standard 2 Date Collected: 5/29/2008 20:42:29
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: Standard 2
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected			Conc.	Calib
	Intensity	Std.Dev.	RSD		
Sca	367187.7	177.98	0.05%	90.9	%
Yr	467625.8	2697.82	0.58%	95.2	%
Ag _t	489793.1	529.24	0.11%	[2]	mg/L
Al _t	79580.5	13.65	0.02%	[10]	mg/L
Ast	15532.0	240.92	1.55%	[10]	mg/L
B _t	127683.8	507.67	0.40%	[5.02]	mg/L
Bat	385818.6	1007.57	0.26%	[10]	mg/L
Bet	10280184.7	25142.40	0.24%	[4.01]	mg/L
Cat	3144105.4	3435.28	0.11%	[100]	mg/L
Cdt	106763.5	257.78	0.24%	[5.01]	mg/L
Cot	148879.3	328.92	0.22%	[10]	mg/L
Crt	431407.9	198.41	0.05%	[9.97]	mg/L
Cut	2936559.7	4590.09	0.16%	[10]	mg/L
Fet	119656.4	230.88	0.19%	[9.98]	mg/L
Kt	123673.2	446.75	0.36%	[100]	mg/L
Mgt	2174312.1	3591.62	0.17%	[100]	mg/L
Mnt	4792614.9	9175.41	0.19%	[10]	mg/L
Mot	98476.6	59.46	0.06%	[9.98]	mg/L
Nat	464168.8	346.67	0.07%	[100]	mg/L
Nit	186333.3	224.52	0.12%	[10]	mg/L
Pbt	37972.5	494.83	1.30%	[10]	mg/L
Sbt	15439.4	226.00	1.46%	[10]	mg/L
Set	9645.8	122.24	1.27%	[10]	mg/L
Tlt	21385.5	281.76	1.32%	[10]	mg/L
Vt	1501467.2	3314.84	0.22%	[10]	mg/L
Znt	382382.3	604.38	0.16%	[10]	mg/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	244900	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	7958	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	1553	0.00000	1.000000	
B _t	1	Lin, Calc Int	-0.0	25440	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	38580	0.00000	1.000000	
Be	1	Lin, Calc Int	-0.0	2564000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	31440	0.00000	1.000000	
Cd	1	Lin, Calc Int	0.0	21310	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	14890	0.00000	1.000000	
Cr	1	Lin, Calc Int	-0.0	43270	0.00000	1.000000	
Cu	1	Lin, Calc Int	0.0	293700	0.00000	1.000000	

Fe	1	Lin, Calc Int	0.0	11990	0.00000	1.000000
K	1	Lin, Calc Int	0.0	1237	0.00000	1.000000
Mg	1	Lin, Calc Int	0.0	21740	0.00000	1.000000
Mn	1	Lin, Calc Int	0.0	479300	0.00000	1.000000
Mo	1	Lin, Calc Int	0.0	9867	0.00000	1.000000
Na	1	Lin, Calc Int	-0.0	4642	0.00000	1.000000
Ni	1	Lin, Calc Int	-0.0	18630	0.00000	1.000000
Pb	1	Lin, Calc Int	0.0	3797	0.00000	1.000000
Sb	1	Lin, Calc Int	-0.0	1544	0.00000	1.000000
Se	1	Lin, Calc Int	0.0	964.6	0.00000	1.000000
Tl	1	Lin, Calc Int	0.0	2139	0.00000	1.000000
V	1	Lin, Calc Int	0.0	150100	0.00000	1.000000
Zn	1	Lin, Calc Int	0.0	38240	0.00000	1.000000

=====

Analysis Begun

Start Time: 5/29/2008 20:44:30 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 1 Autosampler Location: 15
 Sample ID: ICV Date Collected: 5/29/2008 20:44:32
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected		Calib.	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc.	Units	
Sca	360399.7	89.2 %	0.13				0.15%
Yr	463599.2	94.3 %	1.11				1.18%
Agt	492892.4	2.02 mg/L	0.003		2.02 mg/L		0.17%
QC value within limits for Ag		Recovery = 100.81%					
Alt	79840.8	9.67 mg/L	0.039		9.67 mg/L		0.039 0.40%
QC value within limits for Al		Recovery = 96.66%					
Ast	15550.0	10.0 mg/L	0.03		10.0 mg/L		0.03 0.26%
QC value within limits for As		Recovery = 100.12%					
B _t	129854.9	5.02 mg/L	0.018		5.02 mg/L		0.018 0.35%
QC value within limits for B _t		Recovery = 100.34%					
Bat	388434.9	10.1 mg/L	0.01		10.1 mg/L		0.01 0.06%
QC value within limits for Ba		Recovery = 100.68%					
Bet	10369882.8	4.05 mg/L	0.007		4.05 mg/L		0.007 0.16%
QC value within limits for Be		Recovery = 101.16%					
Cat	3141706.9	99.9 mg/L	1.07		99.9 mg/L		1.07 1.07%
QC value within limits for Ca		Recovery = 99.92%					
Cdt	107581.8	4.94 mg/L	0.003		4.94 mg/L		0.003 0.06%
QC value within limits for Cd		Recovery = 98.82%					
Cot	150116.0	10.1 mg/L	0.01		10.1 mg/L		0.01 0.12%
QC value within limits for Co		Recovery = 100.83%					
Crt	435012.3	10.1 mg/L	0.01		10.1 mg/L		0.01 0.15%
QC value within limits for Cr		Recovery = 100.53%					
Cut	2953486.3	10.1 mg/L	0.00		10.1 mg/L		0.00 0.03%
QC value within limits for Cu		Recovery = 100.67%					
Fet	118528.8	9.89 mg/L	0.025		9.89 mg/L		0.025 0.25%
QC value within limits for Fe		Recovery = 98.86%					
Kt	124661.1	101 mg/L	0.8		101 mg/L		0.8 0.79%
QC value within limits for K		Recovery = 100.80%					
Mgt	2172003.0	99.9 mg/L	1.04		99.9 mg/L		1.04 1.04%
QC value within limits for Mg		Recovery = 99.91%					
Mnt	4820298.0	10.1 mg/L	0.00		10.1 mg/L		0.00 0.03%
QC value within limits for Mn		Recovery = 100.56%					
Mot	99098.0	10.0 mg/L	0.02		10.0 mg/L		0.02 0.16%
QC value within limits for Mo		Recovery = 100.43%					
Nat	470178.4	101 mg/L	0.6		101 mg/L		0.6 0.56%
QC value within limits for Na		Recovery = 101.29%					
Nit	187100.7	10.0 mg/L	0.01		10.0 mg/L		0.01 0.08%
QC value within limits for Ni		Recovery = 100.41%					
Pbt	38074.2	10.0 mg/L	0.05		10.0 mg/L		0.05 0.47%
QC value within limits for Pb		Recovery = 100.27%					
Sbt	15568.9	10.1 mg/L	0.03		10.1 mg/L		0.03 0.27%
QC value within limits for Sb		Recovery = 101.24%					
Set	9661.8	10.0 mg/L	0.06		10.0 mg/L		0.06 0.62%

	QC value within limits for Se	Recovery = 100.41%				
T1t	21621.8	10.2 mg/L	0.00	10.2 mg/L	0.00	0.03%
	QC value within limits for Tl	Recovery = 101.68%				
Vt	1509613.8	10.1 mg/L	0.01	10.1 mg/L	0.01	0.12%
	QC value within limits for V	Recovery = 101.07%				
Znt	384155.4	9.96 mg/L	0.005	9.96 mg/L	0.005	0.05%
	QC value within limits for Zn	Recovery = 99.60%				
All analyte(s) passed QC.						

Sequence No.: 2
 Sample ID: LINEARITY
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 9
 Date Collected: 5/29/2008 20:47:49
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Calib.	Sample	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	347401.6	86.0 %	0.20		0.23%
Yr	445689.2	90.7 %	0.37		0.40%
Ag†	-5191.2	0.0146 mg/L	0.00036	0.0146 mg/L	0.00036 2.49%
QC value within limits for Ag		Recovery = Not calculated			
Alt	1.4	0.00016 mg/L	0.005948	0.00016 mg/L	0.005948 >999.9%
QC value within limits for Al		Recovery = Not calculated			
Ast	-157.6	-0.101 mg/L	0.0047	-0.101 mg/L	0.0047 4.60%
QC value within limits for As		Recovery = Not calculated			
B†	1320.9	0.0520 mg/L	0.00601	0.0520 mg/L	0.00601 11.56%
QC value within limits for B		Recovery = Not calculated			
Bat	50.3	0.00130 mg/L	0.000132	0.00130 mg/L	0.000132 10.13%
QC value within limits for Ba		Recovery = Not calculated			
Bet	-725.5	-0.00028 mg/L	0.000031	-0.00028 mg/L	0.000031 10.91%
QC value within limits for Be		Recovery = Not calculated			
Cat	9248564.8	294 mg/L	0.5	294 mg/L	0.5 0.16%
QC value within limits for Ca		Recovery = 98.05%			
Cdt	-16.2	0.00065 mg/L	0.000037	0.00065 mg/L	0.000037 5.67%
QC value within limits for Cd		Recovery = Not calculated			
Cot	26.5	0.00178 mg/L	0.000280	0.00178 mg/L	0.000280 15.73%
QC value within limits for Co		Recovery = Not calculated			
Crt	142.0	0.00328 mg/L	0.000073	0.00328 mg/L	0.000073 2.22%
QC value within limits for Cr		Recovery = Not calculated			
Cut	-2907.9	-0.00990 mg/L	0.000268	-0.00990 mg/L	0.000268 2.71%
QC value within limits for Cu		Recovery = Not calculated			
Fet	1166633.7	97.3 mg/L	0.08	97.3 mg/L	0.08 0.08%
QC value within limits for Fe		Recovery = 97.30%			
K†	390829.1	316 mg/L	0.5	316 mg/L	0.5 0.15%
QC value within limits for K		Recovery = 105.34%			
Mgt	4073964.3	188 mg/L	0.2	188 mg/L	0.2 0.12%
QC value within limits for Mg		Recovery = Not calculated			
Mnt	-1076.8	-0.00646 mg/L	0.000081	-0.00646 mg/L	0.000081 1.25%
QC value within limits for Mn		Recovery = Not calculated			
Mot	3.8	0.00039 mg/L	0.000018	0.00039 mg/L	0.000018 4.79%
QC value within limits for Mo		Recovery = Not calculated			
Nat	1416014.8	305 mg/L	0.5	305 mg/L	0.5 0.16%
QC value within limits for Na		Recovery = 101.69%			
Nit	8.3	0.00044 mg/L	0.000212	0.00044 mg/L	0.000212 47.74%
QC value within limits for Ni		Recovery = Not calculated			
Pbt	-11.4	-0.00300 mg/L	0.002944	-0.00300 mg/L	0.002944 98.26%
QC value within limits for Pb		Recovery = Not calculated			
Sbt	31.7	0.0205 mg/L	0.00346	0.0205 mg/L	0.00346 16.87%
QC value within limits for Sb		Recovery = Not calculated			
Set	-117.6	0.119 mg/L	0.0183	0.119 mg/L	0.0183 15.35%
QC value within limits for Se		Recovery = Not calculated			
Tlt	31.6	0.0148 mg/L	0.00038	0.0148 mg/L	0.00038 2.58%
QC value within limits for Tl		Recovery = Not calculated			
V†	-426.4	0.00099 mg/L	0.000284	0.00099 mg/L	0.000284 28.60%
QC value within limits for V		Recovery = Not calculated			
Znf	914.6	0.0239 mg/L	0.00032	0.0239 mg/L	0.00032 1.36%
QC value within limits for Zn		Recovery = Not calculated			

All analyte(s) passed QC.

Sequence No.: 3
 Sample ID: ICSA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 5/29/2008 20:51:33
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	358106.9	88.6 %		0.48			0.54%
Yr	457856.8	93.2 %		0.47			0.50%
Agt	-5194.2	0.0144 mg/L	0.00028		0.0144 mg/L	0.00028	1.96%
QC value within limits for Ag		Recovery = Not calculated					
Alt	1975608.3	248 mg/L	0.3		248 mg/L	0.3	0.12%
QC value within limits for Al		Recovery = 99.30%					
Ast	-345.0	-0.222 mg/L	0.0074		-0.222 mg/L	0.0074	3.35%
QC value within limits for As		Recovery = Not calculated					
B_f	337.2	0.0133 mg/L	0.00102		0.0133 mg/L	0.00102	7.65%
QC value within limits for B		Recovery = Not calculated					
Bat	66.7	0.00173 mg/L	0.000030		0.00173 mg/L	0.000030	1.76%
QC value within limits for Ba		Recovery = Not calculated					
Bet	-854.9	-0.00033 mg/L	0.000010		-0.00033 mg/L	0.000010	3.09%
QC value within limits for Be		Recovery = Not calculated					
Cat	7850627.9	250 mg/L	0.2		250 mg/L	0.2	0.08%
QC value within limits for Ca		Recovery = 99.88%					
Cdt	-29.5	0.00169 mg/L	0.000154		0.00169 mg/L	0.000154	9.09%
QC value within limits for Cd		Recovery = Not calculated					
Cof	13.7	0.00092 mg/L	0.000150		0.00092 mg/L	0.000150	16.31%
QC value within limits for Co		Recovery = Not calculated					
Crt	30.6	0.00071 mg/L	0.000072		0.00071 mg/L	0.000072	10.20%
QC value within limits for Cr		Recovery = Not calculated					
Cut	-3287.4	-0.0112 mg/L	0.00008		-0.0112 mg/L	0.00008	0.71%
QC value within limits for Cu		Recovery = Not calculated					
Fet	1160391.0	96.8 mg/L	0.11		96.8 mg/L	0.11	0.12%
QC value within limits for Fe		Recovery = 96.78%					
Kt	280.2	0.227 mg/L	0.0248		0.227 mg/L	0.0248	10.97%
QC value within limits for K		Recovery = Not calculated					
Mgt	5091164.5	234 mg/L	0.0		234 mg/L	0.0	0.00%
QC value within limits for Mg		Recovery = 93.73%					
Mnt	-1127.3	-0.00761 mg/L	0.000032		-0.00761 mg/L	0.000032	0.42%
QC value within limits for Mn		Recovery = Not calculated					
Mot	3.9	0.00040 mg/L	0.000432		0.00040 mg/L	0.000432	108.23%
QC value within limits for Mo		Recovery = Not calculated					
Nat	555.0	0.120 mg/L	0.0013		0.120 mg/L	0.0013	1.05%
QC value within limits for Na		Recovery = Not calculated					
Nif	-9.4	-0.00051 mg/L	0.000210		-0.00051 mg/L	0.000210	41.59%
QC value within limits for Ni		Recovery = Not calculated					
Pbt	-148.7	-0.0392 mg/L	0.00035		-0.0392 mg/L	0.00035	0.90%
QC value within limits for Pb		Recovery = Not calculated					
Sbt	18.4	0.0119 mg/L	0.00064		0.0119 mg/L	0.00064	5.42%
QC value within limits for Sb		Recovery = Not calculated					
Set	-129.7	0.105 mg/L	0.0070		0.105 mg/L	0.0070	6.67%
QC value within limits for Se		Recovery = Not calculated					
Tlt	24.1	0.0112 mg/L	0.00254		0.0112 mg/L	0.00254	22.66%
QC value within limits for Tl		Recovery = Not calculated					
Vt	-469.0	0.00068 mg/L	0.000045		0.00068 mg/L	0.000045	6.71%
QC value within limits for V		Recovery = Not calculated					
Znt	713.0	0.0186 mg/L	0.00021		0.0186 mg/L	0.00021	1.15%
QC value within limits for Zn		Recovery = Not calculated					

All analyte(s) passed QC.

Sequence No.: 4
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:
 Dilution:
 Autosampler Location: 11
 Date Collected: 5/29/2008 20:55:17
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	357002.4	88.4 %	0.85				0.97%
Yr	458542.8	93.3 %	0.42				0.45%
Agt	12362.4	0.0864 mg/L	0.00024		0.0864 mg/L	0.00024	0.27%
	QC value less than the lower limit for Ag	Recovery = 17.29%					
Alt	1989866.6	250 mg/L	0.7		250 mg/L	0.7	0.29%
	QC value within limits for Al	Recovery = 100.02%					
Ast	-345.1	-0.222 mg/L	0.0015		-0.222 mg/L	0.0015	0.65%
	QC value less than the lower limit for As	Recovery = Not calculated					
B_f	-7.3	-0.00246 mg/L	0.000743		-0.00246 mg/L	0.000743	30.25%
	QC value within limits for B	Recovery = Not calculated					
Bat	10266.1	0.266 mg/L	0.0003		0.266 mg/L	0.0003	0.11%
	QC value within limits for Ba	Recovery = 106.43%					
Bet	641095.0	0.250 mg/L	0.0004		0.250 mg/L	0.0004	0.15%
	QC value within limits for Be	Recovery = 100.04%					
Cat	7901605.5	251 mg/L	0.4		251 mg/L	0.4	0.16%
	QC value within limits for Ca	Recovery = 100.53%					
Cdt	10765.9	0.509 mg/L	0.0073		0.509 mg/L	0.0073	1.43%
	QC value within limits for Cd	Recovery = 101.80%					
Cot	3601.7	0.242 mg/L	0.0028		0.242 mg/L	0.0028	1.16%
	QC value within limits for Co	Recovery = 96.77%					
Crt	10984.8	0.254 mg/L	0.0005		0.254 mg/L	0.0005	0.20%
	QC value within limits for Cr	Recovery = 101.55%					
Cut	73201.0	0.250 mg/L	0.0010		0.250 mg/L	0.0010	0.42%
	QC value within limits for Cu	Recovery = 99.80%					
Fet	1172023.8	97.8 mg/L	0.15		97.8 mg/L	0.15	0.15%
	QC value within limits for Fe	Recovery = 97.75%					
Kt	40.1	0.0324 mg/L	0.03081		0.0324 mg/L	0.03081	95.12%
	QC value within limits for K	Recovery = Not calculated					
Mgt	5136480.5	236 mg/L	0.1		236 mg/L	0.1	0.02%
	QC value within limits for Mg	Recovery = 94.56%					
Mnt	122068.6	0.249 mg/L	0.0013		0.249 mg/L	0.0013	0.52%
	QC value within limits for Mn	Recovery = 99.76%					
Mot	-0.4	-0.00004 mg/L	0.000349		-0.00004 mg/L	0.000349	902.60%
	QC value within limits for Mo	Recovery = Not calculated					
Nat	338.5	0.0729 mg/L	0.00770		0.0729 mg/L	0.00770	10.56%
	QC value within limits for Na	Recovery = Not calculated					
Nit	8825.0	0.474 mg/L	0.0064		0.474 mg/L	0.0064	1.36%
	QC value within limits for Ni	Recovery = 94.72%					
Pbt	1766.4	0.465 mg/L	0.0076		0.465 mg/L	0.0076	1.63%
	QC value within limits for Pb	Recovery = 93.04%					
Sbt	21.8	0.0151 mg/L	0.00583		0.0151 mg/L	0.00583	38.65%
	QC value within limits for Sb	Recovery = Not calculated					
Set	-142.6	0.0941 mg/L	0.01345		0.0941 mg/L	0.01345	14.29%
	QC value within limits for Se	Recovery = Not calculated					
Tlt	3.0	0.00361 mg/L	0.006274		0.00361 mg/L	0.006274	173.65%
	QC value within limits for Tl	Recovery = Not calculated					
Vt	37201.0	0.253 mg/L	0.0009		0.253 mg/L	0.0009	0.34%
	QC value within limits for V	Recovery = 101.13%					
Znt	20933.7	0.544 mg/L	0.0024		0.544 mg/L	0.0024	0.44%
	QC value within limits for Zn	Recovery = 108.76%					
	QC Failed. Continue with analysis.						

Sequence No.: 5
 Sample ID: Wash
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 5/29/2008 20:59:03
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: Wash

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: Wash

Analyte	Mean Corrected	Calib.		Sample		RSD	
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units	
Sca	401302.6	99.3	%	0.08			0.08%
Yr	491110.1	99.9	%	1.89			1.89%
Agt	42.1	0.00017	mg/L	0.000020	0.00017	mg/L	0.000020
QC value within limits for Ag		Recovery = Not calculated					
Alt	-1.5	-0.00019	mg/L	0.000934	-0.00019	mg/L	0.000934
QC value within limits for Al		Recovery = Not calculated					
Ast	2.8	0.00183	mg/L	0.001026	0.00183	mg/L	0.001026
QC value within limits for As		Recovery = Not calculated					
B _t	228.6	0.00898	mg/L	0.000103	0.00898	mg/L	0.000103
QC value within limits for B _t		Recovery = Not calculated					
Bat	-0.6	-0.00001	mg/L	0.000008	-0.00001	mg/L	0.000008
QC value within limits for Ba		Recovery = Not calculated					
Bet	-22.0	-0.00001	mg/L	0.000012	-0.00001	mg/L	0.000012
QC value within limits for Be		Recovery = Not calculated					
Cat	30.9	0.00098	mg/L	0.002047	0.00098	mg/L	0.002047
QC value within limits for Ca		Recovery = Not calculated					
Cdt	-0.8	-0.00006	mg/L	0.000013	-0.00006	mg/L	0.000013
QC value within limits for Cd		Recovery = Not calculated					
Cot	-0.2	-0.00001	mg/L	0.000174	-0.00001	mg/L	0.000174
QC value within limits for Co		Recovery = Not calculated					
Crt	4.4	0.00010	mg/L	0.000053	0.00010	mg/L	0.000053
QC value within limits for Cr		Recovery = Not calculated					
Cut	192.3	0.00065	mg/L	0.000136	0.00065	mg/L	0.000136
QC value within limits for Cu		Recovery = Not calculated					
Fet	32.9	0.00274	mg/L	0.000365	0.00274	mg/L	0.000365
QC value within limits for Fe		Recovery = Not calculated					
Kt	41.6	0.0336	mg/L	0.00735	0.0336	mg/L	0.00735
QC value within limits for K		Recovery = Not calculated					
Mgt	40.9	0.00188	mg/L	0.000557	0.00188	mg/L	0.000557
QC value within limits for Mg		Recovery = Not calculated					
Mnt	43.9	0.00009	mg/L	0.000004	0.00009	mg/L	0.000004
QC value within limits for Mn		Recovery = Not calculated					
Mot	-1.8	-0.00018	mg/L	0.000248	-0.00018	mg/L	0.000248
QC value within limits for Mo		Recovery = Not calculated					
Nat	143.5	0.0309	mg/L	0.00547	0.0309	mg/L	0.00547
QC value within limits for Na		Recovery = Not calculated					
Nit	-4.6	-0.00025	mg/L	0.000051	-0.00025	mg/L	0.000051
QC value within limits for Ni		Recovery = Not calculated					
Pbt	5.9	0.00156	mg/L	0.000497	0.00156	mg/L	0.000497
QC value within limits for Pb		Recovery = Not calculated					
Sbt	6.3	0.00406	mg/L	0.000859	0.00406	mg/L	0.000859
QC value within limits for Sb		Recovery = Not calculated					
Set	-6.8	-0.00700	mg/L	0.002852	-0.00700	mg/L	0.002852
QC value within limits for Se		Recovery = Not calculated					
Tlt	4.3	0.00201	mg/L	0.002221	0.00201	mg/L	0.002221
QC value within limits for Tl		Recovery = Not calculated					
Vt	2.7	0.00002	mg/L	0.000018	0.00002	mg/L	0.000018
QC value within limits for V		Recovery = Not calculated					
Znt	8.6	0.00023	mg/L	0.000088	0.00023	mg/L	0.000088
QC value within limits for Zn		Recovery = Not calculated					
All analyte(s) passed QC.							

Sequence No.: 6
 Sample ID: QC-25 1ppm
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 12
 Date Collected: 5/29/2008 21:02:26
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: QC-25 1ppm
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: QC-25 1ppm

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std.Dev.	RSD
	Conc. Units	Units	Conc. Units	Std.Dev.	
Sca	403452.1	99.9 %	0.13		0.13%
Yr	491471.7	100 %	0.7		0.69%
Agt	235867.1	0.964 mg/L	0.964 mg/L	0.0043	0.44%
QC value within limits for Ag	Recovery = 96.35%				
Alt	8027.6	0.974 mg/L	0.974 mg/L	0.0089	0.92%
QC value within limits for Al	Recovery = 97.38%				
Ast	1432.2	0.922 mg/L	0.922 mg/L	0.0028	0.30%
QC value within limits for As	Recovery = 92.21%				
B_f	24269.3	0.945 mg/L	0.945 mg/L	0.0007	0.08%
QC value within limits for B	Recovery = 94.52%				
Bat	39975.4	1.04 mg/L	1.04 mg/L	0.009	0.91%
QC value within limits for Ba	Recovery = 103.61%				
Bet	2547576.0	0.994 mg/L	0.994 mg/L	0.0045	0.45%
QC value within limits for Be	Recovery = 99.39%				
Cat	32573.3	1.04 mg/L	1.04 mg/L	0.010	0.95%
QC value within limits for Ca	Recovery = 103.60%				
Cdt	20663.8	0.960 mg/L	0.960 mg/L	0.0045	0.47%
QC value within limits for Cd	Recovery = 96.01%				
Cot	15550.9	1.04 mg/L	1.04 mg/L	0.001	0.07%
QC value within limits for Co	Recovery = 104.45%				
Crt	43719.5	1.01 mg/L	1.01 mg/L	0.001	0.11%
QC value within limits for Cr	Recovery = 101.04%				
Cut	295920.3	1.01 mg/L	1.01 mg/L	0.004	0.42%
QC value within limits for Cu	Recovery = 100.87%				
Fet	12347.1	1.03 mg/L	1.03 mg/L	0.011	1.09%
QC value within limits for Fe	Recovery = 102.98%				
Kt	12162.6	9.83 mg/L	9.83 mg/L	0.120	1.22%
QC value within limits for K	Recovery = 98.34%				
Mgt	23207.6	1.07 mg/L	1.07 mg/L	0.008	0.77%
QC value within limits for Mg	Recovery = 106.92%				
Mnt	501104.0	1.05 mg/L	1.05 mg/L	0.004	0.40%
QC value within limits for Mn	Recovery = 104.56%				
Mot	9449.4	0.958 mg/L	0.958 mg/L	0.0023	0.24%
QC value within limits for Mo	Recovery = 95.76%				
Nat	4846.8	1.04 mg/L	1.04 mg/L	0.028	2.71%
QC value within limits for Na	Recovery = 104.42%				
Nit	19901.6	1.07 mg/L	1.07 mg/L	0.003	0.31%
QC value within limits for Ni	Recovery = 106.81%				
Pbt	4000.5	1.05 mg/L	1.05 mg/L	0.005	0.46%
QC value within limits for Pb	Recovery = 105.35%				
Sbt	1461.2	0.950 mg/L	0.950 mg/L	0.0012	0.13%
QC value within limits for Sb	Recovery = 95.01%				
Set	913.1	0.949 mg/L	0.949 mg/L	0.0016	0.17%
QC value within limits for Se	Recovery = 94.91%				
Tlt	2294.4	1.08 mg/L	1.08 mg/L	0.001	0.05%
QC value within limits for Tl	Recovery = 107.88%				
Vt	143892.2	0.964 mg/L	0.964 mg/L	0.0031	0.32%
QC value within limits for V	Recovery = 96.36%				
Znt	39673.8	1.03 mg/L	1.03 mg/L	0.004	0.34%
QC value within limits for Zn	Recovery = 102.84%				

All analyte(s) passed QC.

Sequence No.: 7
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 5/29/2008 21:06:14
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected		Calib.	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	378379.7	93.7 %	0.00				0.00%
Yr	476200.5	96.9 %	0.46				0.47%
Agt	238545.4	0.976 mg/L	0.0001		0.976 mg/L	0.0001	0.01%
QC value within limits for Ag		Recovery = 97.59%					
Alt	39720.7	4.81 mg/L	0.004		4.81 mg/L	0.004	0.08%
QC value within limits for Al		Recovery = 96.19%					
Ast	7440.5	4.79 mg/L	0.021		4.79 mg/L	0.021	0.43%
QC value within limits for As		Recovery = 95.81%					
B_f	64059.0	2.47 mg/L	0.005		2.47 mg/L	0.005	0.20%
QC value within limits for B_f		Recovery = 98.99%					
Bat	194246.5	5.03 mg/L	0.008		5.03 mg/L	0.008	0.15%
QC value within limits for Ba		Recovery = 100.69%					
Bet	5286806.1	2.06 mg/L	0.010		2.06 mg/L	0.010	0.48%
QC value within limits for Be		Recovery = 103.15%					
Cat	1574251.4	50.1 mg/L	0.07		50.1 mg/L	0.07	0.14%
QC value within limits for Ca		Recovery = 100.14%					
Cdt	42626.9	1.95 mg/L	0.005		1.95 mg/L	0.005	0.28%
QC value within limits for Cd		Recovery = 97.49%					
Cot	75737.8	5.09 mg/L	0.048		5.09 mg/L	0.048	0.94%
QC value within limits for Co		Recovery = 101.74%					
Crt	217577.4	5.03 mg/L	0.028		5.03 mg/L	0.028	0.57%
QC value within limits for Cr		Recovery = 100.57%					
Cut	1457508.3	4.97 mg/L	0.011		4.97 mg/L	0.011	0.21%
QC value within limits for Cu		Recovery = 99.36%					
Fet	60255.2	5.03 mg/L	0.001		5.03 mg/L	0.001	0.02%
QC value within limits for Fe		Recovery = 100.51%					
Kt	60759.0	49.1 mg/L	0.09		49.1 mg/L	0.09	0.17%
QC value within limits for K		Recovery = 98.26%					
Mgt	1085292.9	49.9 mg/L	0.12		49.9 mg/L	0.12	0.25%
QC value within limits for Mg		Recovery = 99.85%					
Mnt	2413053.5	5.03 mg/L	0.013		5.03 mg/L	0.013	0.25%
QC value within limits for Mn		Recovery = 100.68%					
Mot	49123.6	4.98 mg/L	0.010		4.98 mg/L	0.010	0.20%
QC value within limits for Mo		Recovery = 99.57%					
Nat	229393.7	49.4 mg/L	0.13		49.4 mg/L	0.13	0.27%
QC value within limits for Na		Recovery = 98.84%					
Nit	95111.6	5.10 mg/L	0.003		5.10 mg/L	0.003	0.07%
QC value within limits for Ni		Recovery = 102.09%					
Pbt	19396.8	5.11 mg/L	0.006		5.11 mg/L	0.006	0.13%
QC value within limits for Pb		Recovery = 102.16%					
Sbt	7545.1	4.91 mg/L	0.023		4.91 mg/L	0.023	0.46%
QC value within limits for Sb		Recovery = 98.13%					
Set	4816.3	5.01 mg/L	0.006		5.01 mg/L	0.006	0.11%
QC value within limits for Se		Recovery = 100.11%					
Tlt	11220.9	5.28 mg/L	0.008		5.28 mg/L	0.008	0.16%
QC value within limits for Tl		Recovery = 105.51%					
Vt	743906.4	4.98 mg/L	0.001		4.98 mg/L	0.001	0.03%
QC value within limits for V		Recovery = 99.62%					
Znt	195982.9	5.08 mg/L	0.001		5.08 mg/L	0.001	0.01%
QC value within limits for Zn		Recovery = 101.63%					

All analyte(s) passed QC.

Sequence No.: 8
 Sample ID: ICB
 Analyst:
 Initial Sample Wt:
 Dilution:
 Autosampler Location: 0
 Date Collected: 5/29/2008 21:10:01
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	405515.7	100 %		0.3			0.33%
Yr	499108.3	102 %		2.4			2.38%
Agt	82.7	0.00034 mg/L	0.000015		0.00034 mg/L	0.000015	4.37%
Alt	6.1	0.00074 mg/L	0.000502		0.00074 mg/L	0.000502	67.57%
Ast	19.0	0.0122 mg/L	0.00293		0.0122 mg/L	0.00293	23.92%
B _t	692.9	0.0272 mg/L	0.00153		0.0272 mg/L	0.00153	5.63%
Bat	1.6	0.00004 mg/L	0.000055		0.00004 mg/L	0.000055	136.33%
Bet	98.0	0.00004 mg/L	0.000004		0.00004 mg/L	0.000004	11.72%
Cat	-43.3	-0.00138 mg/L	0.002420		-0.00138 mg/L	0.002420	175.77%
Cdt	4.3	0.00003 mg/L	0.000218		0.00003 mg/L	0.000218	698.45%
Cot	-0.6	-0.00004 mg/L	0.000148		-0.00004 mg/L	0.000148	348.20%
Crt	3.3	0.00008 mg/L	0.000136		0.00008 mg/L	0.000136	176.51%
Cut	93.2	0.00032 mg/L	0.000249		0.00032 mg/L	0.000249	78.56%
Fet	13.5	0.00113 mg/L	0.000482		0.00113 mg/L	0.000482	42.81%
Kt	49.9	0.0403 mg/L	0.00730		0.0403 mg/L	0.00730	18.11%
Mgt	31.4	0.00145 mg/L	0.000380		0.00145 mg/L	0.000380	26.31%
Mnt	62.5	0.00013 mg/L	0.000002		0.00013 mg/L	0.000002	1.62%
Mot	6.7	0.00068 mg/L	0.000316		0.00068 mg/L	0.000316	46.39%
Nat	129.5	0.0279 mg/L	0.00128		0.0279 mg/L	0.00128	4.59%
Nit	5.2	0.00028 mg/L	0.000071		0.00028 mg/L	0.000071	25.17%
Pbt	9.3	0.00245 mg/L	0.000023		0.00245 mg/L	0.000023	0.93%
Sbt	13.6	0.00884 mg/L	0.001614		0.00884 mg/L	0.001614	18.26%
Set	-3.8	-0.00397 mg/L	0.001941		-0.00397 mg/L	0.001941	48.95%
Tlt	9.6	0.00450 mg/L	0.003163		0.00450 mg/L	0.003163	70.22%
Vt	-0.1	0.00000 mg/L	0.000005		0.00000 mg/L	0.000005	>999.9%
Znt	-2.2	-0.00006 mg/L	0.000053		-0.00006 mg/L	0.000053	90.49%
		QC value within limits for Zn	Recovery = Not calculated				
		QC Failed. Retry.					

Sequence No.: 9
 Sample ID: ICB

Autosampler Location: 0
 Date Collected: 5/29/2008 21:12:37

Analyst:
 Initial Sample Wt:
 Dilution:

Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected	Calib.	Sample	
	Intensity	Conc. Units	Conc. Units	Std.Dev. RSD
Sca	407355.5	101 %	0.7	0.66%
Yr	500904.5	102 %	0.0	0.05%
Agt	77.7	0.00032 mg/L	0.000265 mg/L	0.000265 83.40%
QC value within limits for Ag		Recovery = Not calculated		
Alt	8.7	0.00106 mg/L	0.000584 mg/L	0.000584 55.03%
QC value within limits for Al		Recovery = Not calculated		
Ast	10.6	0.00682 mg/L	0.002700 mg/L	0.002700 39.58%
QC value within limits for As		Recovery = Not calculated		
B_t	463.7	0.0182 mg/L	0.00071 mg/L	0.00071 3.92%
QC value within limits for B		Recovery = Not calculated		
Bat	-0.2	-0.00001 mg/L	-0.000016 mg/L	0.000016 305.92%
QC value within limits for Ba		Recovery = Not calculated		
Bet	85.6	0.00003 mg/L	0.000004 mg/L	0.000004 11.17%
QC value within limits for Be		Recovery = Not calculated		
Cat	-50.4	-0.00160 mg/L	-0.00166 mg/L	0.000166 10.35%
QC value within limits for Ca		Recovery = Not calculated		
Cdt	1.9	0.00000 mg/L	0.000298 mg/L	0.000298 >999.9%
QC value within limits for Cd		Recovery = Not calculated		
Cot	3.8	0.00026 mg/L	0.000019 mg/L	0.000019 7.20%
QC value within limits for Co		Recovery = Not calculated		
Crt	1.9	0.00004 mg/L	0.000026 mg/L	0.000026 57.09%
QC value within limits for Cr		Recovery = Not calculated		
Cut	32.4	0.00011 mg/L	0.000024 mg/L	0.000024 21.54%
QC value within limits for Cu		Recovery = Not calculated		
Fet	5.8	0.00048 mg/L	0.000120 mg/L	0.000120 24.89%
QC value within limits for Fe		Recovery = Not calculated		
Kt	28.4	0.0230 mg/L	0.01149 mg/L	0.01149 50.06%
QC value within limits for K		Recovery = Not calculated		
Mgt	32.7	0.00151 mg/L	0.000067 mg/L	0.000067 4.48%
QC value within limits for Mg		Recovery = Not calculated		
Mnt	48.2	0.00010 mg/L	0.000017 mg/L	0.000017 16.49%
QC value within limits for Mn		Recovery = Not calculated		
Mot	7.2	0.00073 mg/L	0.000109 mg/L	0.000109 15.04%
QC value within limits for Mo		Recovery = Not calculated		
Naf	103.0	0.0222 mg/L	0.00311 mg/L	0.00311 14.03%
QC value within limits for Na		Recovery = Not calculated		
Nit	-0.6	-0.00003 mg/L	-0.000036 mg/L	0.000036 105.51%
QC value within limits for Ni		Recovery = Not calculated		
Pbt	8.4	0.00221 mg/L	0.000037 mg/L	0.000037 1.66%
QC value within limits for Pb		Recovery = Not calculated		
Sbt	9.7	0.00628 mg/L	0.000705 mg/L	0.000705 11.21%
QC value within limits for Sb		Recovery = Not calculated		
Set	-4.3	-0.00441 mg/L	-0.000972 mg/L	0.000972 22.03%
QC value within limits for Se		Recovery = Not calculated		
Tlt	6.9	0.00320 mg/L	0.000488 mg/L	0.000488 15.24%
QC value within limits for Tl		Recovery = Not calculated		
Vt	6.4	0.00004 mg/L	0.000089 mg/L	0.000089 205.56%
QC value within limits for V		Recovery = Not calculated		
Znf	0.3	0.00001 mg/L	0.000001 mg/L	0.000001 14.13%
QC value within limits for Zn		Recovery = Not calculated		

All analyte(s) passed QC.

Sequence No.: 11
 Sample ID: MRL/2
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 16
 Date Collected: 5/29/2008 21:19:38
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL/2

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: MRL/2

Analyte	Mean Corrected	Calib.			Sample			RSD
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units	Std.Dev.	
Sca	412762.0	102	%	1.3				1.24%
Yr	503650.2	102	%	1.1				1.04%
Agt	1198.0	0.00490	mg/L	0.000073	0.00490	mg/L	0.000073	1.50%
Alt	183.6	0.0227	mg/L	0.00116	0.0227	mg/L	0.00116	5.13%
Ast	73.2	0.0471	mg/L	0.00408	0.0471	mg/L	0.00408	8.67%
B_t	872.9	0.0342	mg/L	0.00006	0.0342	mg/L	0.00006	0.17%
Bat	392.4	0.0102	mg/L	0.00011	0.0102	mg/L	0.00011	1.07%
Bet	1358.8	0.00053	mg/L	0.000004	0.00053	mg/L	0.000004	0.70%
Cat	15643.7	0.498	mg/L	0.0060	0.498	mg/L	0.0060	1.20%
Cdt	64.8	0.00247	mg/L	0.000233	0.00247	mg/L	0.000233	9.44%
Cot	377.7	0.0254	mg/L	0.00006	0.0254	mg/L	0.00006	0.25%
Crt	213.0	0.00492	mg/L	0.000111	0.00492	mg/L	0.000111	2.24%
Cut	1487.3	0.00508	mg/L	0.000400	0.00508	mg/L	0.000400	7.87%
Fet	139.8	0.0117	mg/L	0.00010	0.0117	mg/L	0.00010	0.85%
Kt	616.0	0.498	mg/L	0.0038	0.498	mg/L	0.0038	0.77%
Mgt	1136.3	0.0523	mg/L	0.00005	0.0523	mg/L	0.00005	0.10%
Mnt	550.6	0.00115	mg/L	0.000014	0.00115	mg/L	0.000014	1.23%
Mot	94.1	0.00954	mg/L	0.000128	0.00954	mg/L	0.000128	1.34%
Nat	2377.2	0.512	mg/L	0.0080	0.512	mg/L	0.0080	1.56%
Nit	197.6	0.0106	mg/L	0.00006	0.0106	mg/L	0.00006	0.61%
Pbt	45.9	0.0121	mg/L	0.00022	0.0121	mg/L	0.00022	1.78%
Sbt	37.2	0.0241	mg/L	0.00042	0.0241	mg/L	0.00042	1.75%
Set	41.8	0.0433	mg/L	0.00446	0.0433	mg/L	0.00446	10.29%
Tlt	121.4	0.0567	mg/L	0.00220	0.0567	mg/L	0.00220	3.88%
Vt	159.7	0.00109	mg/L	0.000121	0.00109	mg/L	0.000121	11.05%
Znt	382.8	0.00992	mg/L	0.000104	0.00992	mg/L	0.000104	1.05%

Sequence No.: 10
 Sample ID: MRL
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 20
 Date Collected: 5/29/2008 21:16:01
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	404759.1	100 %	0.9			0.93%
Yr	498212.9	101 %	0.5			0.46%
Agt	2467.5	0.0101 mg/L	0.00024	0.0101 mg/L	0.00024	2.38%
QC value within limits for Ag		Recovery = 100.83%				
Alt	366.1	0.0453 mg/L	0.00098	0.0453 mg/L	0.00098	2.16%
QC value within limits for Al		Recovery = 90.56%				
Ast	141.7	0.0912 mg/L	0.00151	0.0912 mg/L	0.00151	1.65%
QC value within limits for As		Recovery = 91.22%				
B_f	1590.5	0.0623 mg/L	0.00018	0.0623 mg/L	0.00018	0.28%
QC value within limits for B		Recovery = 124.70%				
Bat	780.6	0.0202 mg/L	0.00020	0.0202 mg/L	0.00020	0.98%
QC value within limits for Ba		Recovery = 101.16%				
Bei	2690.6	0.00105 mg/L	0.000000	0.00105 mg/L	0.000000	0.04%
QC value within limits for Be		Recovery = 105.12%				
Cat	31372.1	0.998 mg/L	0.0019	0.998 mg/L	0.0019	0.19%
QC value within limits for Ca		Recovery = 99.78%				
Cdt	128.3	0.00492 mg/L	0.000024	0.00492 mg/L	0.000024	0.49%
QC value within limits for Cd		Recovery = 98.38%				
Cot	767.4	0.0515 mg/L	0.00056	0.0515 mg/L	0.00056	1.10%
QC value within limits for Co		Recovery = 103.09%				
Crt	445.3	0.0103 mg/L	0.00013	0.0103 mg/L	0.00013	1.23%
QC value within limits for Cr		Recovery = 102.91%				
Cut	2994.9	0.0102 mg/L	0.00011	0.0102 mg/L	0.00011	1.12%
QC value within limits for Cu		Recovery = 102.31%				
Fet	251.3	0.0210 mg/L	0.00004	0.0210 mg/L	0.00004	0.21%
QC value within limits for Fe		Recovery = 104.80%				
Kt	1214.4	0.982 mg/L	0.0080	0.982 mg/L	0.0080	0.82%
QC value within limits for K		Recovery = 98.20%				
Mgt	2251.9	0.104 mg/L	0.0001	0.104 mg/L	0.0001	0.06%
QC value within limits for Mg		Recovery = 103.61%				
Mnt	1078.2	0.00225 mg/L	0.000029	0.00225 mg/L	0.000029	1.27%
QC value within limits for Mn		Recovery = 112.37%				
Mot	194.1	0.0197 mg/L	0.00043	0.0197 mg/L	0.00043	2.21%
QC value within limits for Mo		Recovery = 98.34%				
Nat	4643.2	1.00 mg/L	0.009	1.00 mg/L	0.009	0.93%
QC value within limits for Na		Recovery = 100.03%				
Nit	396.3	0.0213 mg/L	0.00008	0.0213 mg/L	0.00008	0.36%
QC value within limits for Ni		Recovery = 106.34%				
Pbt	86.7	0.0228 mg/L	0.00048	0.0228 mg/L	0.00048	2.09%
QC value within limits for Pb		Recovery = 114.10%				
Sbt	72.7	0.0471 mg/L	0.00121	0.0471 mg/L	0.00121	2.58%
QC value within limits for Sb		Recovery = 94.10%				
Set	89.3	0.0926 mg/L	0.00346	0.0926 mg/L	0.00346	3.74%
QC value within limits for Se		Recovery = 92.64%				
Tlt	246.8	0.115 mg/L	0.0016	0.115 mg/L	0.0016	1.38%
QC value within limits for Tl		Recovery = 115.26%				
Vt	314.1	0.00215 mg/L	0.000152	0.00215 mg/L	0.000152	7.06%
QC value within limits for V		Recovery = 107.49%				
Znt	775.4	0.0201 mg/L	0.00021	0.0201 mg/L	0.00021	1.02%
QC value within limits for Zn		Recovery = 100.49%				

All analyte(s) passed QC.

User canceled analysis.

Analysis Begun

Start Time: 5/29/2008 21:24:06 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 35 Autosampler Location: 5
 Sample ID: MCV Date Collected: 5/29/2008 21:24:08
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	391599.9	96.9 %	1.18			1.21%
Yr	489238.4	99.6 %	0.03			0.03%
Agt	117865.7	0.482 mg/L	0.0023	0.482 mg/L	0.0023	0.47%
QC value within limits for Ag		Recovery = 96.44%				
Alt	19625.0	2.38 mg/L	0.002	2.38 mg/L	0.002	0.09%
QC value within limits for Al		Recovery = 95.03%				
Ast	3664.0	2.36 mg/L	0.023	2.36 mg/L	0.023	0.99%
QC value within limits for As		Recovery = 94.36%				
B_t	31181.8	1.20 mg/L	0.007	1.20 mg/L	0.007	0.61%
QC value within limits for B		Recovery = 96.33%				
Bat	96734.9	2.51 mg/L	0.008	2.51 mg/L	0.008	0.31%
QC value within limits for Ba		Recovery = 100.29%				
Bet	2598682.0	1.01 mg/L	0.015	1.01 mg/L	0.015	1.47%
QC value within limits for Be		Recovery = 101.40%				
Cat	783877.9	24.9 mg/L	0.01	24.9 mg/L	0.01	0.04%
QC value within limits for Ca		Recovery = 99.73%				
Cdt	21026.0	0.962 mg/L	0.0094	0.962 mg/L	0.0094	0.97%
QC value within limits for Cd		Recovery = 96.16%				
Cot	37690.8	2.53 mg/L	0.018	2.53 mg/L	0.018	0.72%
QC value within limits for Co		Recovery = 101.27%				
Crt	107589.7	2.49 mg/L	0.023	2.49 mg/L	0.023	0.91%
QC value within limits for Cr		Recovery = 99.46%				
Cut	724331.9	2.47 mg/L	0.019	2.47 mg/L	0.019	0.77%
QC value within limits for Cu		Recovery = 98.76%				
Fet	30244.5	2.52 mg/L	0.004	2.52 mg/L	0.004	0.14%
QC value within limits for Fe		Recovery = 100.90%				
Kt	29859.6	24.1 mg/L	0.04	24.1 mg/L	0.04	0.17%
QC value within limits for K		Recovery = 96.58%				
Mgt	545041.9	25.1 mg/L	0.01	25.1 mg/L	0.01	0.05%
QC value within limits for Mg		Recovery = 100.29%				
Mnt	1210600.3	2.53 mg/L	0.039	2.53 mg/L	0.039	1.56%
QC value within limits for Mn		Recovery = 101.02%				
Mot	24424.5	2.48 mg/L	0.026	2.48 mg/L	0.026	1.04%
QC value within limits for Mo		Recovery = 99.01%				
Nat	112789.5	24.3 mg/L	0.02	24.3 mg/L	0.02	0.07%
QC value within limits for Na		Recovery = 97.20%				
Nit	47689.9	2.56 mg/L	0.006	2.56 mg/L	0.006	0.23%
QC value within limits for Ni		Recovery = 102.38%				
Pbt	9791.8	2.58 mg/L	0.032	2.58 mg/L	0.032	1.26%
QC value within limits for Pb		Recovery = 103.15%				
Sbt	3741.6	2.43 mg/L	0.036	2.43 mg/L	0.036	1.48%
QC value within limits for Sb		Recovery = 97.32%				

Set	2368.4	2.46 mg/L	0.019	2.46 mg/L	0.019	0.78%
	QC value within limits for Se	Recovery = 98.47%				
Tl _f	5690.6	2.68 mg/L	0.031	2.68 mg/L	0.031	1.14%
	QC value within limits for Tl	Recovery = 107.00%				
V _f	365344.9	2.45 mg/L	0.006	2.45 mg/L	0.006	0.23%
	QC value within limits for V	Recovery = 97.85%				
Zn _f	97458.5	2.53 mg/L	0.007	2.53 mg/L	0.007	0.27%
	QC value within limits for Zn	Recovery = 101.07%				
All analyte(s) passed QC.						

User canceled analysis.

Analysis Begun

Start Time: 5/29/2008 21:32:22 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 12 Autosampler Location: 18
 Sample ID: FILTERCHECK Date Collected: 5/29/2008 21:32:24
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: 1X Sample Prep Vol:

Nebulizer Parameters: FILTERCHECK
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: FILTERCHECK

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	404055.7	100 %	0.3			0.31%
Yr	500232.9	102 %	0.5			0.51%
Agt	-6.9	-0.00003 mg/L	0.000161	-0.00003 mg/L	0.000161	571.90%
Alt	-16.5	-0.00208 mg/L	0.000229	-0.00208 mg/L	0.000229	11.00%
Ast	2.0	0.00127 mg/L	0.001013	0.00127 mg/L	0.001013	79.99%
B _t	253.9	0.00998 mg/L	0.000269	0.00998 mg/L	0.000269	2.69%
Bat	0.6	0.00002 mg/L	0.000116	0.00002 mg/L	0.000116	759.81%
Bet	95.4	0.00004 mg/L	0.000000	0.00004 mg/L	0.000000	0.05%
Cat	47.9	0.00152 mg/L	0.000393	0.00152 mg/L	0.000393	25.82%
Cdt	2.6	0.00011 mg/L	0.000019	0.00011 mg/L	0.000019	17.50%
Cot	1.2	0.00008 mg/L	0.000309	0.00008 mg/L	0.000309	391.33%
Crt	0.2	0.00000 mg/L	0.000001	0.00000 mg/L	0.000001	18.61%
Cut	-40.7	-0.00014 mg/L	0.000288	-0.00014 mg/L	0.000288	207.89%
Fet	5.3	0.00045 mg/L	0.000515	0.00045 mg/L	0.000515	115.44%
Kt	12.7	0.0102 mg/L	0.02409	0.0102 mg/L	0.02409	235.52%
Mgt	17.2	0.00079 mg/L	0.000359	0.00079 mg/L	0.000359	45.27%
Mnt	65.0	0.00014 mg/L	0.000024	0.00014 mg/L	0.000024	18.05%
Mot	0.8	0.00008 mg/L	0.000128	0.00008 mg/L	0.000128	157.95%
Nat	103.2	0.0222 mg/L	0.00626	0.0222 mg/L	0.00626	28.16%
Nit	1.6	0.00009 mg/L	0.000200	0.00009 mg/L	0.000200	226.69%
Pbt	6.5	0.00171 mg/L	0.001361	0.00171 mg/L	0.001361	79.44%
Sbt	7.0	0.00456 mg/L	0.000207	0.00456 mg/L	0.000207	4.54%
Set	-9.0	-0.00938 mg/L	0.002081	-0.00938 mg/L	0.002081	22.19%
Tlt	-0.0	-0.00001 mg/L	0.003885	-0.00001 mg/L	0.003885	>999.9%
Vt	10.2	0.00007 mg/L	0.000022	0.00007 mg/L	0.000022	32.24%
Znt	22.9	0.00060 mg/L	0.000223	0.00060 mg/L	0.000223	37.30%

Sequence No.: 13
 Sample ID: MBLANK
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 38
 Date Collected: 5/29/2008 21:36:00
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MBLANK

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: MBLANK

Analyte	Mean Corrected		Calib.	Sample			RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	Std.Dev.	
Sca	410599.2	102 %	0.1				0.15%
Yr	493938.6	101 %	0.5				0.50%
Ag†	-59.2	-0.00024 mg/L	0.000248	-0.00024 mg/L	0.000248	102.48%	
Alt	-15.6	-0.00196 mg/L	0.000390	-0.00196 mg/L	0.000390	19.89%	
Ast	5.1	0.00327 mg/L	0.001829	0.00327 mg/L	0.001829	55.94%	
B†	184.3	0.00725 mg/L	0.000247	0.00725 mg/L	0.000247	3.41%	
Bat	-0.2	-0.00001 mg/L	0.000042	-0.00001 mg/L	0.000042	768.66%	
Bet	78.0	0.00003 mg/L	0.000024	0.00003 mg/L	0.000024	80.29%	
Caf	259.6	0.00826 mg/L	0.001105	0.00826 mg/L	0.001105	13.39%	
Cdt	-2.4	-0.00016 mg/L	0.000080	-0.00016 mg/L	0.000080	50.91%	
Cot	-1.3	-0.00009 mg/L	0.000198	-0.00009 mg/L	0.000198	229.19%	
Crt	2.2	0.00005 mg/L	0.000016	0.00005 mg/L	0.000016	31.93%	
Cut	32.1	0.00011 mg/L	0.000409	0.00011 mg/L	0.000409	374.50%	
Fet	8.6	0.00072 mg/L	0.000631	0.00072 mg/L	0.000631	87.77%	
K†	7.3	0.00587 mg/L	0.001884	0.00587 mg/L	0.001884	32.10%	
Mgt	14.9	0.00069 mg/L	0.000317	0.00069 mg/L	0.000317	46.30%	
Mnt	53.5	0.00011 mg/L	0.000004	0.00011 mg/L	0.000004	3.92%	
Mot	0.4	0.00004 mg/L	0.000376	0.00004 mg/L	0.000376	>999.9%	
Nat	50.7	0.0109 mg/L	0.00230	0.0109 mg/L	0.00230	21.00%	
Nit	1.2	0.00007 mg/L	0.000488	0.00007 mg/L	0.000488	736.45%	
Pbt	-2.4	-0.00064 mg/L	0.000427	-0.00064 mg/L	0.000427	67.25%	
Sbt	5.7	0.00372 mg/L	0.001941	0.00372 mg/L	0.001941	52.16%	
Set	-6.1	-0.00636 mg/L	0.001100	-0.00636 mg/L	0.001100	17.30%	
Tlt	8.1	0.00380 mg/L	0.002156	0.00380 mg/L	0.002156	56.70%	
Vt	-4.6	-0.00003 mg/L	0.000002	-0.00003 mg/L	0.000002	7.64%	
Znt	25.6	0.00067 mg/L	0.000071	0.00067 mg/L	0.000071	10.55%	

Sequence No.: 14
 Sample ID: LCS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 39
 Date Collected: 5/29/2008 21:39:37
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: LCS

Analyte	Mean Corrected	Calib.			Sample			RSD
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units	Std.Dev.	
Sca	396103.8	98.0	%	0.10				0.11%
Yr	494053.2	101	%	1.4				1.36%
Agt	110814.3	0.454	mg/L	0.0006	0.454	mg/L	0.0006	0.13%
Alt	14423.4	1.78	mg/L	0.003	1.78	mg/L	0.003	0.16%
Ast	1425.4	0.918	mg/L	0.0001	0.918	mg/L	0.0001	0.01%
B _t	11776.4	0.455	mg/L	0.0018	0.455	mg/L	0.0018	0.41%
Bat	35924.6	0.931	mg/L	0.0054	0.931	mg/L	0.0054	0.58%
Bei	122774.6	0.0480	mg/L	0.00019	0.0480	mg/L	0.00019	0.40%
Cat	1482125.5	47.1	mg/L	0.16	47.1	mg/L	0.16	0.34%
Cdt	4067.9	0.181	mg/L	0.0004	0.181	mg/L	0.0004	0.24%
Cot	14170.9	0.952	mg/L	0.0026	0.952	mg/L	0.0026	0.28%
Crt	40316.4	0.932	mg/L	0.0020	0.932	mg/L	0.0020	0.22%
Cut	275230.5	0.938	mg/L	0.0022	0.938	mg/L	0.0022	0.24%
Fet	56049.6	4.67	mg/L	0.067	4.67	mg/L	0.067	1.42%
Kt	22765.0	18.4	mg/L	0.10	18.4	mg/L	0.10	0.55%
Mgt	413758.8	19.0	mg/L	0.15	19.0	mg/L	0.15	0.78%
Mnt	227097.4	0.473	mg/L	0.0000	0.473	mg/L	0.0000	0.00%
Mot	9136.1	0.926	mg/L	0.0026	0.926	mg/L	0.0026	0.28%
Nat	216875.0	46.7	mg/L	0.09	46.7	mg/L	0.09	0.19%
Nit	8931.4	0.479	mg/L	0.0014	0.479	mg/L	0.0014	0.29%
Pbt	3686.0	0.971	mg/L	0.0031	0.971	mg/L	0.0031	0.32%
Sbt	715.4	0.467	mg/L	0.0004	0.467	mg/L	0.0004	0.08%
Set	898.9	0.943	mg/L	0.0052	0.943	mg/L	0.0052	0.55%
Tlt	2158.1	1.01	mg/L	0.008	1.01	mg/L	0.008	0.77%
Vt	138577.8	0.928	mg/L	0.0001	0.928	mg/L	0.0001	0.01%
Znt	37162.4	0.967	mg/L	0.0017	0.967	mg/L	0.0017	0.18%

Sequence No.: 15
 Sample ID: LCSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 40
 Date Collected: 5/29/2008 21:42:39
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: LCSD

Analyte	Mean Corrected	Calib.	Sample	RSD		
	Intensity	Conc. Units	Std.Dev.			
Sca	393177.7	97.3 %	0.06		0.06%	
Yr	493309.5	100 %	0.7		0.75%	
Agt	108641.7	0.445 mg/L	0.0001	0.445 mg/L	0.0001	0.03%
Alt	14224.5	1.75 mg/L	0.005	1.75 mg/L	0.005	0.28%
Ast	1410.3	0.908 mg/L	0.0032	0.908 mg/L	0.0032	0.35%
B _t	11632.1	0.449 mg/L	0.0016	0.449 mg/L	0.0016	0.35%
Bat	35152.9	0.911 mg/L	0.0006	0.911 mg/L	0.0006	0.07%
Bet	120246.0	0.0470 mg/L	0.00009	0.0470 mg/L	0.00009	0.19%
Cat	1435421.1	45.7 mg/L	0.61	45.7 mg/L	0.61	1.33%
Cdt	4001.8	0.178 mg/L	0.0004	0.178 mg/L	0.0004	0.22%
Cot	14023.7	0.942 mg/L	0.0017	0.942 mg/L	0.0017	0.18%
Crt	39588.4	0.915 mg/L	0.0025	0.915 mg/L	0.0025	0.28%
Cut	269029.5	0.917 mg/L	0.0036	0.917 mg/L	0.0036	0.40%
Fet	54403.9	4.54 mg/L	0.004	4.54 mg/L	0.004	0.08%
Kt	22306.3	18.0 mg/L	0.10	18.0 mg/L	0.10	0.57%
Mgt	400939.6	18.4 mg/L	0.01	18.4 mg/L	0.01	0.04%
Mnt	222794.3	0.464 mg/L	0.0007	0.464 mg/L	0.0007	0.15%
Mot	8975.1	0.910 mg/L	0.0004	0.910 mg/L	0.0004	0.05%
Nat	212250.9	45.7 mg/L	0.15	45.7 mg/L	0.15	0.32%
Nit	8765.2	0.470 mg/L	0.0002	0.470 mg/L	0.0002	0.04%
Pbt	3627.6	0.955 mg/L	0.0012	0.955 mg/L	0.0012	0.13%
Sbt	705.2	0.460 mg/L	0.0000	0.460 mg/L	0.0000	0.01%
Set	883.0	0.927 mg/L	0.0080	0.927 mg/L	0.0080	0.86%
Tlt	2131.2	0.998 mg/L	0.0035	0.998 mg/L	0.0035	0.35%
Vt	135680.2	0.909 mg/L	0.0015	0.909 mg/L	0.0015	0.16%
Znt	36483.9	0.949 mg/L	0.0027	0.949 mg/L	0.0027	0.28%

Sequence No.: 16
 Sample ID: 2805280392
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 41
 Date Collected: 5/29/2008 21:45:41
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805280392

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: 2805280392

Analyte	Mean Corrected		Calib.	Sample			RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	Std.Dev.	
Sca	412964.1	102 %	1.0				0.93%
Yr	510865.4	104 %	0.1				0.05%
Agt	1.8	0.00001 mg/L	0.000251	0.00001 mg/L	0.000251	>999.9%	
Alt	184.5	0.0232 mg/L	0.00102	0.0232 mg/L	0.00102	4.39%	
Ast	-6.5	-0.00420 mg/L	0.002137	-0.00420 mg/L	0.002137	50.93%	
B_t	372.2	0.0146 mg/L	0.00066	0.0146 mg/L	0.00066	4.53%	
Bat	938.6	0.0243 mg/L	0.00016	0.0243 mg/L	0.00016	0.66%	
Bet	27.7	0.00001 mg/L	0.000017	0.00001 mg/L	0.000017	160.60%	
Cat	465207.0	14.8 mg/L	0.19	14.8 mg/L	0.19	1.31%	
Cdt	-7.9	-0.00031 mg/L	0.000026	-0.00031 mg/L	0.000026	8.18%	
Cot	0.9	0.00006 mg/L	0.000101	0.00006 mg/L	0.000101	175.98%	
Crt	-7.3	-0.00017 mg/L	0.000057	-0.00017 mg/L	0.000057	34.11%	
Cut	103.9	0.00035 mg/L	0.000061	0.00035 mg/L	0.000061	17.33%	
Fet	25.0	0.00208 mg/L	0.000283	0.00208 mg/L	0.000283	13.58%	
Kt	857.7	0.694 mg/L	0.0024	0.694 mg/L	0.0024	0.35%	
Mgt	22290.1	1.03 mg/L	0.004	1.03 mg/L	0.004	0.35%	
Mnt	360.0	0.00073 mg/L	0.000016	0.00073 mg/L	0.000016	2.26%	
Mot	6.0	0.00061 mg/L	0.000257	0.00061 mg/L	0.000257	42.34%	
Nat	15005.7	3.23 mg/L	0.033	3.23 mg/L	0.033	1.03%	
Nit	-1.9	-0.00010 mg/L	0.000136	-0.00010 mg/L	0.000136	134.14%	
Pbt	5.5	0.00144 mg/L	0.000213	0.00144 mg/L	0.000213	14.83%	
Sbt	2.5	0.00162 mg/L	0.001865	0.00162 mg/L	0.001865	115.39%	
Set	-4.9	-0.00509 mg/L	0.001663	-0.00509 mg/L	0.001663	32.67%	
Tt	17.7	0.00829 mg/L	0.000206	0.00829 mg/L	0.000206	2.48%	
Vt	56.2	0.00037 mg/L	0.000049	0.00037 mg/L	0.000049	13.11%	
Znt	29.1	0.00071 mg/L	0.000138	0.00071 mg/L	0.000138	19.44%	

Sequence No.: 17
 Sample ID: 2805280392MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 42
 Date Collected: 5/29/2008 21:49:18
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805280392MS

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: 2805280392MS

Analyte	Mean Corrected	Calib.	Sample	RSD
	Intensity	Conc. Units	Std.Dev.	
Sca	392991.1	97.3 %	0.58	0.59%
Yr	486395.4	99.0 %	0.58	0.58%
Agt	98116.7	0.402 mg/L	0.0006	0.15%
Alt	14974.4	1.85 mg/L	0.005	0.25%
Ast	1488.4	0.958 mg/L	0.0070	0.73%
B _t	12116.8	0.468 mg/L	0.0034	0.72%
Bat	37462.0	0.971 mg/L	0.0008	0.09%
Bet	122918.6	0.0481 mg/L	0.00003	0.06%
Cat	2051532.5	65.3 mg/L	0.47	0.73%
Cdt	4192.2	0.186 mg/L	0.0018	0.96%
Cot	14421.7	0.969 mg/L	0.0025	0.26%
Crt	40840.6	0.944 mg/L	0.0028	0.29%
Cut	266501.1	0.908 mg/L	0.0042	0.46%
Fet	56734.4	4.73 mg/L	0.049	1.03%
Kt	25281.0	20.4 mg/L	0.22	1.08%
Mgt	468285.4	21.5 mg/L	0.07	0.31%
Mnt	232002.5	0.484 mg/L	0.0001	0.01%
Mot	9073.6	0.920 mg/L	0.0012	0.13%
Nat	244076.0	52.6 mg/L	0.29	0.56%
Nit	8993.1	0.483 mg/L	0.0022	0.45%
Pbt	3804.3	1.00 mg/L	0.004	0.42%
Sbt	720.3	0.470 mg/L	0.0033	0.71%
Sei	943.0	0.989 mg/L	0.0031	0.31%
Tlt	2244.2	1.05 mg/L	0.001	0.001
Vt	139986.0	0.937 mg/L	0.0002	0.0002
Znt	39109.9	1.02 mg/L	0.004	0.41%

Method: 200.7&6010 080304

Page 2

Date: 5/29/2008 21:55:46

Tlt	2257.3	1.06 mg/L	0.003	1.06 mg/L	0.003	0.27%
Vt	142578.7	0.955 mg/L	0.0026	0.955 mg/L	0.0026	0.27%
Znt	39766.7	1.03 mg/L	0.000	1.03 mg/L	0.000	0.04%

Analysis Begun

Start Time: 5/29/2008 21:52:38 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 1 Autosampler Location: 15
 Sample ID: ICV Date Collected: 5/29/2008 21:52:39
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:
 User canceled analysis.

Analysis Begun

Start Time: 5/29/2008 21:53:39 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 17 Autosampler Location: 43
 Sample ID: 2805280392MSD Date Collected: 5/29/2008 21:53:39
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: 1X Sample Prep Vol:

Nebulizer Parameters: 2805280392MSD
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: 2805280392MSD

Analyte	Mean Corrected		Calib.	Std.Dev.	Sample		Std.Dev.	RSD
	Intensity	Conc. Units			Conc.	Units		
Sca	395372.5	97.9 %	0.46					0.47%
Yr	496714.8	101 %	0.3					0.31%
Agt	100611.5	0.413 mg/L	0.0065	0.413 mg/L	0.0065	mg/L	0.0065	1.56%
Alt	15208.3	1.88 mg/L	0.013	1.88 mg/L	0.013	mg/L	0.013	0.69%
Ast	1488.1	0.958 mg/L	0.0128	0.958 mg/L	0.0128	mg/L	0.0128	1.33%
B_t	12407.7	0.480 mg/L	0.0006	0.480 mg/L	0.0006	mg/L	0.0006	0.13%
Bat	38165.2	0.989 mg/L	0.0012	0.989 mg/L	0.0012	mg/L	0.0012	0.12%
Bet	125296.7	0.0490 mg/L	0.00006	0.0490 mg/L	0.00006	mg/L	0.00006	0.11%
Cat	2043858.4	65.0 mg/L	0.17	65.0 mg/L	0.17	mg/L	0.17	0.26%
Cdt	4253.1	0.189 mg/L	0.0008	0.189 mg/L	0.0008	mg/L	0.0008	0.44%
Cot	14696.9	0.987 mg/L	0.0012	0.987 mg/L	0.0012	mg/L	0.0012	0.12%
Crt	41713.2	0.964 mg/L	0.0013	0.964 mg/L	0.0013	mg/L	0.0013	0.13%
Cut	270583.2	0.922 mg/L	0.0025	0.922 mg/L	0.0025	mg/L	0.0025	0.27%
Fet	57894.6	4.83 mg/L	0.021	4.83 mg/L	0.021	mg/L	0.021	0.43%
Kt	25586.2	20.7 mg/L	0.06	20.7 mg/L	0.06	mg/L	0.06	0.31%
Mgt	470676.3	21.7 mg/L	0.01	21.7 mg/L	0.01	mg/L	0.01	0.06%
Mnt	236569.5	0.493 mg/L	0.0001	0.493 mg/L	0.0001	mg/L	0.0001	0.02%
Mot	9197.3	0.932 mg/L	0.0105	0.932 mg/L	0.0105	mg/L	0.0105	1.12%
Nat	246490.8	53.1 mg/L	0.04	53.1 mg/L	0.04	mg/L	0.04	0.08%
Nit	9154.3	0.491 mg/L	0.0024	0.491 mg/L	0.0024	mg/L	0.0024	0.50%
Pbt	3831.6	1.01 mg/L	0.005	1.01 mg/L	0.005	mg/L	0.005	0.52%
Sbt	731.6	0.478 mg/L	0.0077	0.478 mg/L	0.0077	mg/L	0.0077	1.60%
Set	953.0	1.000 mg/L	0.0127	1.000 mg/L	0.0127	mg/L	0.0127	1.27%

Sequence No.: 20
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 5/29/2008 22:04:43
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected		Calib.	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	380136.6	94.1 %	0.57				0.61%
Yr	481430.2	98.0 %	1.25				1.27%
Ag†	238292.0	0.975 mg/L	0.0025		0.975 mg/L	0.0025	0.26%
QC value within limits for Ag		Recovery = 97.49%					
Alt	39445.2	4.77 mg/L	0.037		4.77 mg/L	0.037	0.78%
QC value within limits for Al		Recovery = 95.50%					
Ast	7458.7	4.80 mg/L	0.026		4.80 mg/L	0.026	0.53%
QC value within limits for As		Recovery = 96.04%					
B†	63825.8	2.47 mg/L	0.022		2.47 mg/L	0.022	0.89%
QC value within limits for B		Recovery = 98.63%					
Bat	193388.8	5.01 mg/L	0.025		5.01 mg/L	0.025	0.50%
QC value within limits for Ba		Recovery = 100.25%					
Bet	5220902.6	2.04 mg/L	0.010		2.04 mg/L	0.010	0.48%
QC value within limits for Be		Recovery = 101.86%					
Cat	1566874.0	49.8 mg/L	0.24		49.8 mg/L	0.24	0.49%
QC value within limits for Ca		Recovery = 99.67%					
Cdt	42634.4	1.95 mg/L	0.005		1.95 mg/L	0.005	0.27%
QC value within limits for Cd		Recovery = 97.50%					
Cot	75999.1	5.10 mg/L	0.026		5.10 mg/L	0.026	0.50%
QC value within limits for Co		Recovery = 102.09%					
Crt	218405.0	5.05 mg/L	0.004		5.05 mg/L	0.004	0.08%
QC value within limits for Cr		Recovery = 100.95%					
Cut	1453916.7	4.96 mg/L	0.009		4.96 mg/L	0.009	0.18%
QC value within limits for Cu		Recovery = 99.12%					
Fet	59981.6	5.00 mg/L	0.029		5.00 mg/L	0.029	0.58%
QC value within limits for Fe		Recovery = 100.06%					
Kt	60252.8	48.7 mg/L	0.22		48.7 mg/L	0.22	0.45%
QC value within limits for K		Recovery = 97.44%					
Mgt	1080962.9	49.7 mg/L	0.18		49.7 mg/L	0.18	0.36%
QC value within limits for Mg		Recovery = 99.45%					
Mnt	2412404.3	5.03 mg/L	0.005		5.03 mg/L	0.005	0.11%
QC value within limits for Mn		Recovery = 100.65%					
Mot	49168.0	4.98 mg/L	0.011		4.98 mg/L	0.011	0.22%
QC value within limits for Mo		Recovery = 99.66%					
Nat	225761.2	48.6 mg/L	0.23		48.6 mg/L	0.23	0.48%
QC value within limits for Na		Recovery = 97.28%					
Nit	95101.5	5.10 mg/L	0.012		5.10 mg/L	0.012	0.24%
QC value within limits for Ni		Recovery = 102.08%					
Pbt	19419.1	5.11 mg/L	0.021		5.11 mg/L	0.021	0.41%
QC value within limits for Pb		Recovery = 102.28%					
Sbt	7538.9	4.90 mg/L	0.018		4.90 mg/L	0.018	0.37%
QC value within limits for Sb		Recovery = 98.05%					
Set	4836.1	5.03 mg/L	0.029		5.03 mg/L	0.029	0.57%
QC value within limits for Se		Recovery = 100.52%					
Tl†	11223.1	5.28 mg/L	0.031		5.28 mg/L	0.031	0.59%
QC value within limits for Tl		Recovery = 105.53%					
V†	742992.5	4.97 mg/L	0.010		4.97 mg/L	0.010	0.19%
QC value within limits for V		Recovery = 99.50%					
Znt	196108.9	5.08 mg/L	0.009		5.08 mg/L	0.009	0.18%
QC value within limits for Zn		Recovery = 101.70%					

All analyte(s) passed QC.

Sequence No.: 21
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 5/29/2008 22:08:31
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	399849.6	99.0 %	0.02		0.02%
Yr	494465.4	101 %	1.2		1.20%
Agt	-3.1	-0.00001 mg/L	0.000282	-0.00001 mg/L	0.000282 >999.9%
QC value within limits for Ag		Recovery = Not calculated			
Alt	0.5	0.00003 mg/L	0.000448	0.00003 mg/L	0.000448 >999.9%
QC value within limits for Al		Recovery = Not calculated			
Ast	13.2	0.00849 mg/L	0.000434	0.00849 mg/L	0.000434 5.11%
QC value within limits for As		Recovery = Not calculated			
B _t	544.6	0.0214 mg/L	0.00109	0.0214 mg/L	0.00109 5.10%
QC value greater than the upper limit for B		Recovery = Not calculated			
Bat	-1.3	-0.00003 mg/L	0.000049	-0.00003 mg/L	0.000049 139.94%
QC value within limits for Ba		Recovery = Not calculated			
Bet	51.5	0.00002 mg/L	0.000032	0.00002 mg/L	0.000032 157.56%
QC value within limits for Be		Recovery = Not calculated			
Cat	12.0	0.00038 mg/L	0.000819	0.00038 mg/L	0.000819 214.08%
QC value within limits for Ca		Recovery = Not calculated			
Cdt	3.5	0.00005 mg/L	0.000109	0.00005 mg/L	0.000109 240.43%
QC value within limits for Cd		Recovery = Not calculated			
Cot	-4.0	-0.00027 mg/L	0.000234	-0.00027 mg/L	0.000234 86.63%
QC value within limits for Co		Recovery = Not calculated			
Crt	7.6	0.00018 mg/L	0.000106	0.00018 mg/L	0.000106 60.09%
QC value within limits for Cr		Recovery = Not calculated			
Cut	97.5	0.00033 mg/L	0.000093	0.00033 mg/L	0.000093 28.03%
QC value within limits for Cu		Recovery = Not calculated			
Fet	13.8	0.00115 mg/L	0.000234	0.00115 mg/L	0.000234 20.34%
QC value within limits for Fe		Recovery = Not calculated			
Kt	19.6	0.0159 mg/L	0.00271	0.0159 mg/L	0.00271 17.09%
QC value within limits for K		Recovery = Not calculated			
Mgt	17.7	0.00082 mg/L	0.000108	0.00082 mg/L	0.000108 13.19%
QC value within limits for Mg		Recovery = Not calculated			
Mnt	59.2	0.00012 mg/L	0.000012	0.00012 mg/L	0.000012 9.55%
QC value within limits for Mn		Recovery = Not calculated			
Mot	9.6	0.00097 mg/L	0.000450	0.00097 mg/L	0.000450 46.18%
QC value within limits for Mo		Recovery = Not calculated			
Nat	127.9	0.0276 mg/L	0.00304	0.0276 mg/L	0.00304 11.02%
QC value within limits for Na		Recovery = Not calculated			
Nit	2.7	0.00015 mg/L	0.000142	0.00015 mg/L	0.000142 96.25%
QC value within limits for Ni		Recovery = Not calculated			
Pbt	7.6	0.00201 mg/L	0.000453	0.00201 mg/L	0.000453 22.54%
QC value within limits for Pb		Recovery = Not calculated			
Sbt	12.0	0.00778 mg/L	0.000341	0.00778 mg/L	0.000341 4.38%
QC value within limits for Sb		Recovery = Not calculated			
Set	-3.9	-0.00402 mg/L	0.002001	-0.00402 mg/L	0.002001 49.79%
QC value within limits for Se		Recovery = Not calculated			
Tl _t	10.4	0.00485 mg/L	0.000799	0.00485 mg/L	0.000799 16.48%
QC value within limits for Tl		Recovery = Not calculated			
V _t	13.0	0.00009 mg/L	0.000072	0.00009 mg/L	0.000072 81.70%
QC value within limits for V		Recovery = Not calculated			
Znf	3.3	0.00008 mg/L	0.000051	0.00008 mg/L	0.000051 60.37%
QC value within limits for Zn		Recovery = Not calculated			
QC Failed. Retry. User canceled analysis.					

Analysis Begun

Start Time: 5/29/2008 22:13:39 Plasma On Time: 5/29/2008 10:30:51
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080529A.sif
 Batch ID: 080529A
 Results Data Set: 080529A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 22 Autosampler Location: 0
 Sample ID: CCB Date Collected: 5/29/2008 22:13:41
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	Std.Dev.	RSD
Sca	404863.6	100 %	0.8			0.75%
Yr	490283.7	99.8 %	0.26			0.26%
Agt	30.8	0.00013 mg/L	0.000062	0.00013 mg/L	0.000062	48.98%
QC value within limits for Ag		Recovery = Not calculated				
Alt	-8.0	-0.00101 mg/L	0.001064	-0.00101 mg/L	0.001064	105.54%
QC value within limits for Al		Recovery = Not calculated				
Ast	4.8	0.00312 mg/L	0.002322	0.00312 mg/L	0.002322	74.45%
QC value within limits for As		Recovery = Not calculated				
B_t	265.3	0.0104 mg/L	0.00004	0.0104 mg/L	0.00004	0.41%
QC value within limits for B		Recovery = Not calculated				
Bat	-0.6	-0.00002 mg/L	0.000028	-0.00002 mg/L	0.000028	167.34%
QC value within limits for Ba		Recovery = Not calculated				
Bet	92.7	0.00004 mg/L	0.000001	0.00004 mg/L	0.000001	2.06%
QC value within limits for Be		Recovery = Not calculated				
Cat	7.9	0.00025 mg/L	0.000058	0.00025 mg/L	0.000058	23.17%
QC value within limits for Ca		Recovery = Not calculated				
Cdt	2.2	0.00006 mg/L	0.000141	0.00006 mg/L	0.000141	240.12%
QC value within limits for Cd		Recovery = Not calculated				
Cot	-3.5	-0.00024 mg/L	0.000105	-0.00024 mg/L	0.000105	44.68%
QC value within limits for Co		Recovery = Not calculated				
Crt	5.2	0.00012 mg/L	0.000151	0.00012 mg/L	0.000151	126.51%
QC value within limits for Cr		Recovery = Not calculated				
Cut	-1.0	0.00000 mg/L	0.000284	0.00000 mg/L	0.000284	>999.9%
QC value within limits for Cu		Recovery = Not calculated				
Fet	9.2	0.00076 mg/L	0.000094	0.00076 mg/L	0.000094	12.30%
QC value within limits for Fe		Recovery = Not calculated				
Kt	15.9	0.0129 mg/L	0.00129	0.0129 mg/L	0.00129	10.00%
QC value within limits for K		Recovery = Not calculated				
Mgt	-4.0	-0.00018 mg/L	0.000655	-0.00018 mg/L	0.000655	358.05%
QC value within limits for Mg		Recovery = Not calculated				
Mnt	35.0	0.00007 mg/L	0.000002	0.00007 mg/L	0.000002	2.48%
QC value within limits for Mn		Recovery = Not calculated				
Mot	2.5	0.00025 mg/L	0.000663	0.00025 mg/L	0.000663	262.75%
QC value within limits for Mo		Recovery = Not calculated				
Nat	56.5	0.0122 mg/L	0.00402	0.0122 mg/L	0.00402	32.98%
QC value within limits for Na		Recovery = Not calculated				
Nit	7.7	0.00041 mg/L	0.000466	0.00041 mg/L	0.000466	112.51%
QC value within limits for Ni		Recovery = Not calculated				
Pbt	-1.8	-0.00048 mg/L	0.000536	-0.00048 mg/L	0.000536	110.99%
QC value within limits for Pb		Recovery = Not calculated				
Sbt	7.2	0.00468 mg/L	0.003056	0.00468 mg/L	0.003056	65.26%
QC value within limits for Sb		Recovery = Not calculated				
Set	-4.6	-0.00475 mg/L	0.000317	-0.00475 mg/L	0.000317	6.67%
QC value within limits for Se		Recovery = Not calculated				
Tlt	2.9	0.00136 mg/L	0.000809	0.00136 mg/L	0.000809	59.50%
QC value within limits for Tl		Recovery = Not calculated				

Vt	3.0	0.00002 mg/L	0.000035	0.00002 mg/L	0.000035	165.50%
	QC value within limits for V Recovery = Not calculated					
Zn _t	3.5	0.00009 mg/L	0.000088	0.00009 mg/L	0.000088	98.02%
	QC value within limits for Zn Recovery = Not calculated					
All analyte(s) passed QC.						

Sequence No.: 30
 Sample ID: 2805270410
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 53
 Date Collected: 5/29/2008 22:45:12
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805270410

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: 2805270410

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	396052.0	98.0 %	0.78			0.80%
Yr	496773.6	101 %	0.7			0.68%
Agt	-14.3	-0.00006 mg/L	0.000043	-0.00006 mg/L	0.000043	74.56%
Alt	-19.0	-0.00243 mg/L	0.000929	-0.00243 mg/L	0.000929	38.26%
Ast	-25.1	-0.0162 mg/L	0.00384	-0.0162 mg/L	0.00384	23.73%
B_t	961.5	0.0378 mg/L	0.00045	0.0378 mg/L	0.00045	1.20%
Bat	3830.3	0.0993 mg/L	0.00108	0.0993 mg/L	0.00108	1.09%
Bet	-167.9	-0.00007 mg/L	0.000007	-0.00007 mg/L	0.000007	10.55%
Cat	1839079.4	58.5 mg/L	0.11	58.5 mg/L	0.11	0.18%
Cdt	-15.1	-0.00049 mg/L	0.000118	-0.00049 mg/L	0.000118	24.33%
Cot	-4.2	-0.00029 mg/L	0.000253	-0.00029 mg/L	0.000253	88.68%
Crt	-12.9	-0.00030 mg/L	0.000280	-0.00030 mg/L	0.000280	93.84%
Cut	236.6	0.00081 mg/L	0.000018	0.00081 mg/L	0.000018	2.18%
Fei	42.6	0.00355 mg/L	0.000221	0.00355 mg/L	0.000221	6.22%
Kt	3111.8	2.52 mg/L	0.033	2.52 mg/L	0.033	1.31%
Mgt	306871.5	14.1 mg/L	0.05	14.1 mg/L	0.05	0.32%
Mnt	-19.9	-0.00036 mg/L	0.000020	-0.00036 mg/L	0.000020	5.48%
Mot	9.7	0.00098 mg/L	0.000093	0.00098 mg/L	0.000093	9.50%
Nat	130644.1	28.1 mg/L	0.00	28.1 mg/L	0.00	0.00%
Nit	-5.4	-0.00029 mg/L	0.000352	-0.00029 mg/L	0.000352	120.55%
Pbt	-9.0	-0.00237 mg/L	0.002540	-0.00237 mg/L	0.002540	107.11%
Sbt	3.3	0.00211 mg/L	0.003360	0.00211 mg/L	0.003360	159.37%
Set	-18.5	-0.0192 mg/L	0.01070	-0.0192 mg/L	0.01070	55.82%
Tlt	23.3	0.0109 mg/L	0.00445	0.0109 mg/L	0.00445	40.88%
Vt	37.0	0.00025 mg/L	0.000133	0.00025 mg/L	0.000133	54.11%
Znt	-48.1	-0.00146 mg/L	0.000043	-0.00146 mg/L	0.000043	2.95%

Sequence No.: 31
 Sample ID: 2805270410MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 54
 Date Collected: 5/29/2008 22:49:30
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805270410MS
 Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: 2805270410MS

Analyte	Mean Corrected	Calib.			Sample			
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units	Std.Dev.	RSD
Sca	384984.1	95.3	%	0.63				0.66%
Yr	473314.9	96.3	%	1.02				1.06%
Agt	99791.3	0.409	mg/L	0.0019	0.409	mg/L	0.0019	0.46%
Alt	14916.3	1.84	mg/L	0.008	1.84	mg/L	0.008	0.46%
Ast	1486.5	0.957	mg/L	0.0024	0.957	mg/L	0.0024	0.25%
B_t	13077.5	0.506	mg/L	0.0025	0.506	mg/L	0.0025	0.50%
Bat	40324.2	1.05	mg/L	0.005	1.05	mg/L	0.005	0.45%
Bet	123987.6	0.0485	mg/L	0.00026	0.0485	mg/L	0.00026	0.54%
Cat	3440540.9	109	mg/L	0.9	109	mg/L	0.9	0.80%
Cdt	4211.1	0.187	mg/L	0.0004	0.187	mg/L	0.0004	0.23%
Cot	14237.7	0.956	mg/L	0.0113	0.956	mg/L	0.0113	1.18%
Crt	40894.7	0.945	mg/L	0.0086	0.945	mg/L	0.0086	0.91%
Cut	270829.5	0.923	mg/L	0.0018	0.923	mg/L	0.0018	0.20%
Fet	58052.7	4.84	mg/L	0.004	4.84	mg/L	0.004	0.09%
Kt	28026.3	22.7	mg/L	0.09	22.7	mg/L	0.09	0.42%
Mgt	759782.1	35.0	mg/L	0.01	35.0	mg/L	0.01	0.01%
Mnt	229773.7	0.479	mg/L	0.0023	0.479	mg/L	0.0023	0.47%
Mot	9197.5	0.932	mg/L	0.0049	0.932	mg/L	0.0049	0.53%
Nat	360116.5	77.6	mg/L	0.03	77.6	mg/L	0.03	0.04%
Nit	8800.8	0.472	mg/L	0.0006	0.472	mg/L	0.0006	0.14%
Pbt	3798.3	1.00	mg/L	0.002	1.00	mg/L	0.002	0.24%
Sbt	734.8	0.480	mg/L	0.0005	0.480	mg/L	0.0005	0.11%
Set	950.5	0.997	mg/L	0.0001	0.997	mg/L	0.0001	0.01%
Tlt	2209.8	1.04	mg/L	0.003	1.04	mg/L	0.003	0.34%
Vt	141158.5	0.945	mg/L	0.0034	0.945	mg/L	0.0034	0.36%
Znt	38738.8	1.01	mg/L	0.005	1.01	mg/L	0.005	0.53%

Sequence No.: 32
 Sample ID: 2805270410MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 55
 Date Collected: 5/29/2008 22:53:12
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805270410MSD

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: 2805270410MSD

Analyte	Mean Corrected	Calib.	Sample	RSD	
	Intensity	Conc. Units	Std.Dev.		
Sca	387696.8	96.0 %	0.10		0.10%
Yr	482826.6	98.3 %	0.47		0.48%
Agt	100667.9	0.413 mg/L	0.0008	0.413 mg/L	0.0008 0.20%
Alt	14789.3	1.82 mg/L	0.001	1.82 mg/L	0.001 0.06%
Asr	1470.5	0.947 mg/L	0.0013	0.947 mg/L	0.0013 0.14%
B _t	13125.4	0.508 mg/L	0.0024	0.508 mg/L	0.0024 0.47%
Bat	40296.7	1.04 mg/L	0.007	1.04 mg/L	0.007 0.64%
Bet	123735.7	0.0484 mg/L	0.00015	0.0484 mg/L	0.00015 0.30%
Cat	3349999.8	107 mg/L	2.2	107 mg/L	2.2 2.02%
Cdt	4176.8	0.186 mg/L	0.0005	0.186 mg/L	0.0005 0.26%
Cot	14146.3	0.950 mg/L	0.0004	0.950 mg/L	0.0004 0.04%
Crt	40931.2	0.946 mg/L	0.0048	0.946 mg/L	0.0048 0.51%
Cut	268643.9	0.916 mg/L	0.0005	0.916 mg/L	0.0005 0.05%
Fet	58307.4	4.86 mg/L	0.004	4.86 mg/L	0.004 0.08%
Kt	27212.5	22.0 mg/L	0.17	22.0 mg/L	0.17 0.75%
Mgt	747729.0	34.4 mg/L	0.04	34.4 mg/L	0.04 0.11%
Mnt	228682.9	0.476 mg/L	0.0015	0.476 mg/L	0.0015 0.32%
Mot	9141.8	0.926 mg/L	0.0011	0.926 mg/L	0.0011 0.12%
Nat	352688.6	76.0 mg/L	0.31	76.0 mg/L	0.31 0.41%
Nit	8727.6	0.468 mg/L	0.0013	0.468 mg/L	0.0013 0.28%
Pbt	3777.1	0.995 mg/L	0.0011	0.995 mg/L	0.0011 0.11%
Sbt	727.6	0.475 mg/L	0.0058	0.475 mg/L	0.0058 1.23%
Set	950.0	0.997 mg/L	0.0032	0.997 mg/L	0.0032 0.33%
Tlt	2192.3	1.03 mg/L	0.006	1.03 mg/L	0.006 0.62%
Vt	140688.2	0.942 mg/L	0.0023	0.942 mg/L	0.0023 0.24%
Znt	38687.4	1.01 mg/L	0.002	1.01 mg/L	0.002 0.16%

Sequence No.: 33
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 5/29/2008 22:56:54
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	RSD
Sca	380391.8	94.2 %	0.28		0.30%
Yr	481112.8	97.9 %	0.41		0.42%
Agt	235336.2	0.963 mg/L	0.0036	0.963 mg/L	0.0036 0.38%
QC value within limits for Ag		Recovery = 96.28%			
Alt	38825.2	4.70 mg/L	0.012	4.70 mg/L	0.012 0.26%
QC value within limits for Al		Recovery = 93.98%			
Ast	7399.7	4.76 mg/L	0.000	4.76 mg/L	0.000 0.01%
QC value within limits for As		Recovery = 95.28%			
B _t	63059.2	2.44 mg/L	0.022	2.44 mg/L	0.022 0.90%
QC value within limits for B _t		Recovery = 97.45%			
Bat	191578.9	4.97 mg/L	0.015	4.97 mg/L	0.015 0.29%
QC value within limits for Ba		Recovery = 99.31%			
Bet	5179774.9	2.02 mg/L	0.007	2.02 mg/L	0.007 0.35%
QC value within limits for Be		Recovery = 101.06%			
Cat	1558070.1	49.6 mg/L	0.03	49.6 mg/L	0.03 0.05%
QC value within limits for Ca		Recovery = 99.11%			
Cdt	42258.0	1.93 mg/L	0.011	1.93 mg/L	0.011 0.56%
QC value within limits for Cd		Recovery = 96.63%			
Cot	75294.2	5.06 mg/L	0.043	5.06 mg/L	0.043 0.86%
QC value within limits for Co		Recovery = 101.15%			
Crt	215395.7	4.98 mg/L	0.025	4.98 mg/L	0.025 0.50%
QC value within limits for Cr		Recovery = 99.56%			
Cut	1429473.2	4.87 mg/L	0.001	4.87 mg/L	0.001 0.01%
QC value within limits for Cu		Recovery = 97.45%			
Fet	60273.3	5.03 mg/L	0.016	5.03 mg/L	0.016 0.31%
QC value within limits for Fe		Recovery = 100.54%			
Kt	59409.5	48.0 mg/L	0.00	48.0 mg/L	0.00 0.01%
QC value within limits for K		Recovery = 96.07%			
Mgt	1079981.3	49.7 mg/L	0.10	49.7 mg/L	0.10 0.21%
QC value within limits for Mg		Recovery = 99.36%			
Mnt	2389151.5	4.98 mg/L	0.021	4.98 mg/L	0.021 0.42%
QC value within limits for Mn		Recovery = 99.68%			
Mof	48679.1	4.93 mg/L	0.020	4.93 mg/L	0.020 0.40%
QC value within limits for Mo		Recovery = 98.67%			
Naf	221624.0	47.7 mg/L	0.00	47.7 mg/L	0.00 0.01%
QC value within limits for Na		Recovery = 95.49%			
Nit	94275.6	5.06 mg/L	0.014	5.06 mg/L	0.014 0.28%
QC value within limits for Ni		Recovery = 101.19%			
Pbt	19285.9	5.08 mg/L	0.009	5.08 mg/L	0.009 0.18%
QC value within limits for Pb		Recovery = 101.58%			
Sbt	7486.1	4.87 mg/L	0.021	4.87 mg/L	0.021 0.42%
QC value within limits for Sb		Recovery = 97.36%			
Set	4793.2	4.98 mg/L	0.011	4.98 mg/L	0.011 0.22%
QC value within limits for Se		Recovery = 99.63%			
Tlt	11149.0	5.24 mg/L	0.023	5.24 mg/L	0.023 0.44%
QC value within limits for Tl		Recovery = 104.83%			
Vt	735072.7	4.92 mg/L	0.024	4.92 mg/L	0.024 0.48%
QC value within limits for V		Recovery = 98.44%			
Znf	194040.8	5.03 mg/L	0.018	5.03 mg/L	0.018 0.36%
QC value within limits for Zn		Recovery = 100.62%			

All analyte(s) passed QC.

Sequence No.: 34
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 5/29/2008 23:00:42
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	411150.6	102 %	0.6		0.55%
Yr	502139.4	102 %	1.1		1.11%
Agt	16.8	0.00007 mg/L	0.000253	0.00007 mg/L	0.000253 368.49%
QC value within limits for Ag		Recovery = Not calculated			
Alt	-16.6	-0.00213 mg/L	0.000889	-0.00213 mg/L	0.000889 41.72%
QC value within limits for Al		Recovery = Not calculated			
Ast	15.0	0.00968 mg/L	0.000533	0.00968 mg/L	0.000533 5.51%
QC value within limits for As		Recovery = Not calculated			
B _t	586.1	0.0230 mg/L	0.00121	0.0230 mg/L	0.00121 5.25%
QC value greater than the upper limit for B _t		Recovery = Not calculated			
Bat	0.6	0.00002 mg/L	0.000014	0.00002 mg/L	0.000014 89.70%
QC value within limits for Ba		Recovery = Not calculated			
Bet	151.1	0.00006 mg/L	0.000007	0.00006 mg/L	0.000007 12.32%
QC value within limits for Be		Recovery = Not calculated			
Cat	16.3	0.00052 mg/L	0.002172	0.00052 mg/L	0.002172 419.01%
QC value within limits for Ca		Recovery = Not calculated			
Cdt	1.8	-0.00005 mg/L	0.000206	-0.00005 mg/L	0.000206 415.45%
QC value within limits for Cd		Recovery = Not calculated			
Cot	-0.8	-0.00006 mg/L	0.000257	-0.00006 mg/L	0.000257 451.51%
QC value within limits for Co		Recovery = Not calculated			
Crt	3.1	0.00007 mg/L	0.000104	0.00007 mg/L	0.000104 143.93%
QC value within limits for Cr		Recovery = Not calculated			
Cut	-4.1	-0.00001 mg/L	0.000232	-0.00001 mg/L	0.000232 >999.9%
QC value within limits for Cu		Recovery = Not calculated			
Fet	5.2	0.00043 mg/L	0.000131	0.00043 mg/L	0.000131 30.26%
QC value within limits for Fe		Recovery = Not calculated			
Kt	27.3	0.0220 mg/L	0.01323	0.0220 mg/L	0.01323 60.01%
QC value within limits for K		Recovery = Not calculated			
Mgt	21.6	0.00099 mg/L	0.000338	0.00099 mg/L	0.000338 33.98%
QC value within limits for Mg		Recovery = Not calculated			
Mnt	52.7	0.00011 mg/L	0.000003	0.00011 mg/L	0.000003 2.70%
QC value within limits for Mn		Recovery = Not calculated			
Mot	11.4	0.00115 mg/L	0.000421	0.00115 mg/L	0.000421 36.48%
QC value within limits for Mo		Recovery = Not calculated			
Nat	95.9	0.0207 mg/L	0.00412	0.0207 mg/L	0.00412 19.96%
QC value within limits for Na		Recovery = Not calculated			
Nit	1.8	0.00010 mg/L	0.000087	0.00010 mg/L	0.000087 90.29%
QC value within limits for Ni		Recovery = Not calculated			
Pbt	7.7	0.00202 mg/L	0.002782	0.00202 mg/L	0.002782 137.96%
QC value within limits for Pb		Recovery = Not calculated			
Sbt	10.2	0.00658 mg/L	0.000392	0.00658 mg/L	0.000392 5.96%
QC value within limits for Sb		Recovery = Not calculated			
Set	-8.1	-0.00844 mg/L	0.002078	-0.00844 mg/L	0.002078 24.61%
QC value within limits for Se		Recovery = Not calculated			
Tlt	9.9	0.00464 mg/L	0.002014	0.00464 mg/L	0.002014 43.41%
QC value within limits for Tl		Recovery = Not calculated			
Vt	13.0	0.00009 mg/L	0.000098	0.00009 mg/L	0.000098 112.61%
QC value within limits for V		Recovery = Not calculated			
Znt	-1.4	-0.00004 mg/L	0.000019	-0.00004 mg/L	0.000019 49.56%
QC value within limits for Zn		Recovery = Not calculated			
QC Failed. Retry.					

Sequence No.: 35
 Sample ID: CCB

Autosampler Location: 0
 Date Collected: 5/29/2008 23:03:15

Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte Back Pressure Flow
 All 212.0 kPa 0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	409175.5	101 %	0.4		0.39%
Yr	494660.1	101 %	0.5		0.51%
Agt	-12.6	-0.00005 mg/L	0.000127	-0.00005 mg/L	0.000127 247.06%
QC value within limits for Ag		Recovery = Not calculated			
Alt	-3.1	-0.00041 mg/L	0.000031	-0.00041 mg/L	0.000031 7.56%
QC value within limits for Al		Recovery = Not calculated			
Ast	9.3	0.00601 mg/L	0.002095	0.00601 mg/L	0.002095 34.87%
QC value within limits for As		Recovery = Not calculated			
B_t	435.7	0.0171 mg/L	0.00044	0.0171 mg/L	0.00044 2.54%
QC value within limits for B		Recovery = Not calculated			
Bat	-2.1	-0.00006 mg/L	0.000089	-0.00006 mg/L	0.000089 161.49%
QC value within limits for Ba		Recovery = Not calculated			
Bef	56.1	0.00002 mg/L	0.000029	0.00002 mg/L	0.000029 132.33%
QC value within limits for Be		Recovery = Not calculated			
Cat	72.7	0.00231 mg/L	0.001429	0.00231 mg/L	0.001429 61.78%
QC value within limits for Ca		Recovery = Not calculated			
Cdt	1.3	-0.00002 mg/L	0.000100	-0.00002 mg/L	0.000100 455.03%
QC value within limits for Cd		Recovery = Not calculated			
Cot	0.5	0.00003 mg/L	0.000303	0.00003 mg/L	0.000303 931.83%
QC value within limits for Co		Recovery = Not calculated			
Crt	1.6	0.00004 mg/L	0.000150	0.00004 mg/L	0.000150 399.41%
QC value within limits for Cr		Recovery = Not calculated			
Cut	-63.5	-0.00022 mg/L	0.000139	-0.00022 mg/L	0.000139 64.26%
QC value within limits for Cu		Recovery = Not calculated			
Fet	3.1	0.00026 mg/L	0.000466	0.00026 mg/L	0.000466 179.20%
QC value within limits for Fe		Recovery = Not calculated			
Kt	5.2	0.00417 mg/L	0.012870	0.00417 mg/L	0.012870 308.88%
QC value within limits for K		Recovery = Not calculated			
Mgt	12.7	0.00058 mg/L	0.000490	0.00058 mg/L	0.000490 83.83%
QC value within limits for Mg		Recovery = Not calculated			
Mnt	38.0	0.00008 mg/L	0.000017	0.00008 mg/L	0.000017 21.11%
QC value within limits for Mn		Recovery = Not calculated			
Mot	5.5	0.00056 mg/L	0.000387	0.00056 mg/L	0.000387 69.67%
QC value within limits for Mo		Recovery = Not calculated			
Nat	66.6	0.0143 mg/L	0.00950	0.0143 mg/L	0.00950 66.27%
QC value within limits for Na		Recovery = Not calculated			
Nit	1.4	0.00008 mg/L	0.000038	0.00008 mg/L	0.000038 48.84%
QC value within limits for Ni		Recovery = Not calculated			
Pbt	7.3	0.00193 mg/L	0.000174	0.00193 mg/L	0.000174 9.01%
QC value within limits for Pb		Recovery = Not calculated			
Sbt	8.3	0.00536 mg/L	0.002249	0.00536 mg/L	0.002249 41.98%
QC value within limits for Sb		Recovery = Not calculated			
Set	-10.1	-0.0104 mg/L	0.00208	-0.0104 mg/L	0.00208 19.91%
QC value within limits for Se		Recovery = Not calculated			
Tlt	7.2	0.00338 mg/L	0.001592	0.00338 mg/L	0.001592 47.10%
QC value within limits for Tl		Recovery = Not calculated			
Vt	6.0	0.00004 mg/L	0.000076	0.00004 mg/L	0.000076 190.19%
QC value within limits for V		Recovery = Not calculated			
Znt	-1.7	-0.00004 mg/L	0.000020	-0.00004 mg/L	0.000020 44.70%
QC value within limits for Zn		Recovery = Not calculated			

All analyte(s) passed QC.

Sequence No.: 36
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 5/29/2008 23:06:39
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected		Calib.	Std.Dev.	Sample		RSD
	Intensity	Conc. Units			Conc. Units	Std.Dev.	
Sca	393039.7	97.3 %	0.62				0.63%
Yr	484713.7	98.6 %	0.73				0.74%
Agt	117851.9	0.482 mg/L	0.0015		0.482 mg/L	0.0015	0.30%
QC value within limits for Ag		Recovery = 96.43%					
Alt	19583.9	2.37 mg/L	0.006		2.37 mg/L	0.006	0.26%
QC value within limits for Al		Recovery = 94.83%					
Ast	3657.4	2.35 mg/L	0.010		2.35 mg/L	0.010	0.40%
QC value within limits for As		Recovery = 94.19%					
B_f	31443.8	1.21 mg/L	0.005		1.21 mg/L	0.005	0.38%
QC value within limits for B		Recovery = 97.15%					
Bat	96767.2	2.51 mg/L	0.006		2.51 mg/L	0.006	0.22%
QC value within limits for Ba		Recovery = 100.32%					
Bet	2615439.0	1.02 mg/L	0.008		1.02 mg/L	0.008	0.77%
QC value within limits for Be		Recovery = 102.06%					
Cat	788811.2	25.1 mg/L	0.04		25.1 mg/L	0.04	0.15%
QC value within limits for Ca		Recovery = 100.35%					
Cdt	20963.7	0.959 mg/L	0.0036		0.959 mg/L	0.0036	0.38%
QC value within limits for Cd		Recovery = 95.91%					
Cot	38107.6	2.56 mg/L	0.005		2.56 mg/L	0.005	0.19%
QC value within limits for Co		Recovery = 102.39%					
Crt	108393.4	2.51 mg/L	0.015		2.51 mg/L	0.015	0.59%
QC value within limits for Cr		Recovery = 100.20%					
Cut	724106.5	2.47 mg/L	0.019		2.47 mg/L	0.019	0.79%
QC value within limits for Cu		Recovery = 98.73%					
Fet	30392.0	2.53 mg/L	0.002		2.53 mg/L	0.002	0.07%
QC value within limits for Fe		Recovery = 101.39%					
Kt	29606.1	23.9 mg/L	0.10		23.9 mg/L	0.10	0.40%
QC value within limits for K		Recovery = 95.76%					
Mgt	549767.3	25.3 mg/L	0.02		25.3 mg/L	0.02	0.07%
QC value within limits for Mg		Recovery = 101.16%					
Mnt	1217628.5	2.54 mg/L	0.021		2.54 mg/L	0.021	0.82%
QC value within limits for Mn		Recovery = 101.60%					
Mot	24358.9	2.47 mg/L	0.011		2.47 mg/L	0.011	0.44%
QC value within limits for Mo		Recovery = 98.75%					
Nat	111277.2	24.0 mg/L	0.02		24.0 mg/L	0.02	0.08%
QC value within limits for Na		Recovery = 95.89%					
Nit	47940.4	2.57 mg/L	0.001		2.57 mg/L	0.001	0.03%
QC value within limits for Ni		Recovery = 102.91%					
Pbt	9794.0	2.58 mg/L	0.015		2.58 mg/L	0.015	0.59%
QC value within limits for Pb		Recovery = 103.17%					
Sbt	3712.6	2.41 mg/L	0.012		2.41 mg/L	0.012	0.51%
QC value within limits for Sb		Recovery = 96.57%					
Set	2362.2	2.46 mg/L	0.000		2.46 mg/L	0.000	0.01%
QC value within limits for Se		Recovery = 98.21%					
Tlt	5681.6	2.67 mg/L	0.014		2.67 mg/L	0.014	0.51%
QC value within limits for Tl		Recovery = 106.84%					
Vt	366990.6	2.46 mg/L	0.004		2.46 mg/L	0.004	0.18%
QC value within limits for V		Recovery = 98.30%					
Znt	97984.8	2.54 mg/L	0.002		2.54 mg/L	0.002	0.06%
QC value within limits for Zn		Recovery = 101.62%					

All analyte(s) passed QC.

Sequence No.: 45
 Sample ID: 2805280376_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 64
 Date Collected: 5/29/2008 23:42:53
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805280376_2X
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: 2805280376_2X

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	345464.5	85.5 %	0.13			0.15%
Yr	446944.3	91.0 %	0.21			0.23%
Agt	-133.6	-0.00048 mg/L	0.000005	-0.00096 mg/L	0.000010	1.09%
Alt	-82.7	-0.0119 mg/L	0.00625	-0.0239 mg/L	0.01250	52.37%
Ast	-2.6	-0.00164 mg/L	0.001492	-0.00328 mg/L	0.002984	90.87%
B _t	45063.6	1.77 mg/L	0.014	3.54 mg/L	0.028	0.78%
Bat	445.9	0.0116 mg/L	0.00009	0.0231 mg/L	0.00018	0.79%
Bet	-994.7	-0.00038 mg/L	0.000026	-0.00075 mg/L	0.000052	6.96%
Cat	6258824.4	199 mg/L	0.3	398 mg/L	0.6	0.15%
Cdt	-9.3	-0.00041 mg/L	0.000272	-0.00062 mg/L	0.000543	66.12%
Cot	19.0	0.00128 mg/L	0.000560	0.00256 mg/L	0.001119	43.76%
Crt	2975.7	0.0688 mg/L	0.00017	0.138 mg/L	0.0003	0.25%
Cut	899.3	0.00307 mg/L	0.000240	0.00614 mg/L	0.000481	7.83%
Fet	2137.3	0.178 mg/L	0.0025	0.357 mg/L	0.0049	1.38%
Kt	18277.9	14.8 mg/L	0.10	29.6 mg/L	0.21	0.70%
Mgt	2059510.9	94.7 mg/L	0.04	189 mg/L	0.1	0.04%
Mnt	226813.0	0.475 mg/L	0.0002	0.951 mg/L	0.0003	0.03%
Mot	416.9	0.0422 mg/L	0.00025	0.0845 mg/L	0.00051	0.60%
Nat	3146920.5	678 mg/L	2.1	1360 mg/L	4.2	0.31%
Nit	131.5	0.00705 mg/L	0.000662	0.0141 mg/L	0.00132	9.38%
Pbt	-33.2	-0.00875 mg/L	0.001496	-0.0175 mg/L	0.00299	17.09%
Sbt	11.2	0.00717 mg/L	0.003068	0.0143 mg/L	0.00614	42.80%
Set	-39.6	-0.0406 mg/L	0.00399	-0.0812 mg/L	0.00799	9.84%
Tlt	27.2	0.0161 mg/L	0.00322	0.0323 mg/L	0.00645	19.99%
Vt	2980.2	0.0202 mg/L	0.00011	0.0404 mg/L	0.00022	0.54%
Znt	-80.1	-0.00216 mg/L	0.000195	-0.00433 mg/L	0.000389	9.00%

Analytical Sequence

Method : 200.7&6010_080304

Seq.	Loc.	Sample ID	Status
1	15	K ⁺ qc	QC Passed
2	9	K ⁺ qc	QC Passed
3	10	K ⁺ qc	QC Passed
4	11	K ⁺ qc	QC Failed
5	0	K ⁺ qc	Wash
6	12	K ⁺ qc	QC-25 1ppm
7	4	K ⁺ qc	CCV
8	0	K ⁺ qc	ICB
9	0	K ⁺ qc	ICB
10	20	K ⁺ qc	MRL
11	16	K ⁺ qc	MRL/2
12	18	K ⁺ qc	FILTERCHECK
13	38	K ⁺ qc	MBLANK
14	39	K ⁺ qc	LCS
15	40	K ⁺ qc	LCSD
16	41	K ⁺ qc	2805280392
17	42	K ⁺ qc	2805280392MS
18	43	K ⁺ qc	2805280392MSD
19	44	K ⁺ qc	2805280378
20	45	K ⁺ qc	2805290001
21	4	qc	CCV
22	0	qc	CCB
23	46	K ⁺ qc	2805280348
24	47	K ⁺ qc	2805280351
25	48	K ⁺ qc	2805280489
26	49	K ⁺ qc	2805280354
27	50	K ⁺ qc	2805280395
28	51	K ⁺ qc	2805280219
29	52	K ⁺ qc	2805280272
30	53	K ⁺ qc	2805270410
31	54	K ⁺ qc	2805270410MS
32	55	K ⁺ qc	2805270410MSD
33	4	qc	CCV
34	0	qc	CCB
35	5	qc	MCV
36	56	K ⁺ qc	2805280273
37	57	K ⁺ qc	2805280341
38	58	K ⁺ qc	2805280217
39	59	K ⁺ qc	2805280342
40	60	K ⁺ qc	2805280241
41	61	K ⁺ qc	2805280242
42	62	K ⁺ qc	2805280213
43	63	K ⁺ qc	2805280477
44	64	K ⁺ qc	2805280376_2X
45	65	K ⁺ qc	MBLANK
46	4	qc	CCV
47	0	qc	CCB
48	21	K ⁺ qc	MRL
49	66	K ⁺ qc	LCS
50	67	K ⁺ qc	LCSD
51	68	K ⁺ qc	2805270411
52	69	K ⁺ qc	2805270411MS
53	70	K ⁺ qc	2805270411MSD
54	71	K ⁺ qc	2805280243
55	72	K ⁺ qc	2805280244
56	73	K ⁺ qc	2805280245

EPA 200.7/6010B QC Check List

Analyst CSK Analysis Date 6-4-08 Reviewer/Date 6-5-08

Instrument Perkin Elmer Optima 4300DV

- All sample analyzed within 6 month holding time
 All sample raw concentration below the high standard or linear range marked for dilution and rerun

Initial and closing QC

- ICV within +/- 5%
 Linearity check +/- 10%
 ICSAB +/- 20%
 QCS +/- 5%
 MRL +/- 50%

Middle, closing and batch QC

- FilterCheck < 1/2 MRL
 MBLANK < 1/2 MRL
 LCS +/- 15%
 MS/MSD +/- 30% (200.7) +/- 25% (6010B)
 CCV/MCV/ECV +/- 10%
 ICB/CCB/ECB < 1/2 MRL
 CCB ran after the CCV

General QC

- RPD between MS/MSD is within +/- 20%
 RPD between LCS/LCSD is within +/- 20%
 Internal standards 60 TO 125%
 All pH of the samples are < 2

No more than 20 samples per batch

MS is run at frequency of 1 every 10 samples and MSD is run at frequency of 1 every 20 samples

NA QIR needed for failed QC

NA Special Det Code noted on the cover sheet

NA R value for multi point calibration is > 0.995

NA Proper MRL check ran for special low MRL samples

Reagent and Standards used for

Optima 4300 DV
Updated 03/31/08

Int: CSK
Date: 6-5-08

Method 200.7/6010

ICP SUMMARY SHEET

File ID: 080604A
Date Started: 6/4/08
Analyst ID: CSK

SAMPLE ID

LINEARITY	(18:41)	Wash	(18:56)	LINEARITY	(20:52)
FILTERCHECK	(21:13)	2806030669	(21:25)	2806030741	(21:35)
2806030674	(21:39)	2806030542	(21:50)	2806030544	(21:53)
2806030545	(21:57)	2806030547	(22:01)	2806030844	(22:04)
2806030859	(22:09)	2806030867	(22:13)	2806030671	(22:17)
2806030660	(22:37)	2806030870	(22:42)	2806030871	(22:45)
2806030872	(22:49)	2806030570	(22:53)	2806030713	(22:56)
2806030780	(23:01)	2806030672	(23:05)	D805220919_2	(23:09)
2805290427	(23:36)	2805310013	(23:46)	2805280136	(0:05)
2805280137	(0:09)	2805280138	(0:13)	2805280139	(0:16)
2805280274	(0:20)	2805280347	(0:23)	2805280375_2	(0:26)
2805280487	(0:30)	2805280472	(0:33)	2805280473	(0:37)
2805280474	(0:47)	2805280475	(0:51)	2805280476	(0:54)
2805280478	(0:58)	2805280479	(1:02)	2805280480	(1:06)
2805280481	(1:11)	2805190003	(1:15)		

NO Mn < 10³

COMMENT:

The second batch (digested) LCSD fail for "Na"

Rerun @ 080605B see attached.

For closure CCSB, see RAW DATA 08/6/08

Analyst: CSK
6-5-08

Approved By: WBD

Peer Reviewed! BSR 6/5/08

Sequence No.: 14
 Sample ID: LCSD2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 25
 Date Collected: 6/5/2008 18:52:51
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD2007

Analyte	Back Pressure	Flow
All	213.0 kPa	0.65 L/min

Mean Data: LCSD2007

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Conc. Units	Std.Dev.	
Sca	385820.0	97.2 %	0.30			0.31%
Yr	422356.1	99.8 %	0.08			0.08%
Ag†	111574.2	0.441 mg/L	0.0002	0.441 mg/L	0.0002	0.04%
Alt	12045.4	1.79 mg/L	0.007	1.79 mg/L	0.007	0.40%
Ast	1405.8	0.900 mg/L	0.0000	0.900 mg/L	0.0000	0.00%
B ₋ †	11977.7	0.441 mg/L	0.0020	0.441 mg/L	0.0020	0.46%
Bat	34162.6	0.905 mg/L	0.0002	0.905 mg/L	0.0002	0.02%
Bei	125755.4	0.0473 mg/L	0.00001	0.0473 mg/L	0.00001	0.02%
Cat	1238306.5	46.3 mg/L	0.32	46.3 mg/L	0.32	0.68%
Cdt	4024.7	0.177 mg/L	0.0004	0.177 mg/L	0.0004	0.22%
Cot	15189.2	0.948 mg/L	0.0018	0.948 mg/L	0.0018	0.19%
Crt	42068.0	0.915 mg/L	0.0002	0.915 mg/L	0.0002	0.02%
Cut	271529.1	0.917 mg/L	0.0020	0.917 mg/L	0.0020	0.21%
Fet	48609.2	4.55 mg/L	0.013	4.55 mg/L	0.013	0.29%
Kt	18482.2	18.2 mg/L	0.06	18.2 mg/L	0.06	0.33%
Mgt	371042.9	18.9 mg/L	0.00	18.9 mg/L	0.00	0.01%
Mnt	230855.4	0.466 mg/L	0.0004	0.466 mg/L	0.0004	0.08%
Mot	9189.9	0.895 mg/L	0.0006	0.895 mg/L	0.0006	0.07%
Nat	244836.8	46.2 mg/L	0.04	46.2 mg/L	0.04	0.09%
Nit	8832.5	0.465 mg/L	0.0012	0.465 mg/L	0.0012	0.27%
Pbt	3678.2	0.947 mg/L	0.0012	0.947 mg/L	0.0012	0.13%
Sbt	703.4	0.451 mg/L	0.0006	0.451 mg/L	0.0006	0.13%
Set	906.4	0.919 mg/L	0.0025	0.919 mg/L	0.0025	0.27%
Tlt	2158.1	0.989 mg/L	0.0044	0.989 mg/L	0.0044	0.44%
Vt	139100.7	0.909 mg/L	0.0010	0.909 mg/L	0.0010	0.11%
Znt	37356.2	0.953 mg/L	0.0039	0.953 mg/L	0.0039	0.41%

BATCH NUMBER for 080604A

Test Parameter:

SCA YR AG AL AS B_ BA BE CA CD CO CR CU FE K MG MN MO NA NI

Batch ID: 2806030669

2806030669	2806030741	2806030674
2806030542	2806030544	2806030545
2806030547	2806030844	2806030859
2806030867	2806030671	2806030660
2806030870	2806030871	2806030872
2806030570	2806030713	2806030780
2806030672		

Batch ID: 2805290427

2805290427	2805310013	2805280136
2805280137	2805280138	2805280139
2805280274	2805280347	2805280375_2X
2805280487	2805280472	2805280473
2805280474	2805280475	2805280476
2805280478	2805280479	2805280480
2805280481	2805190003	

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
ICV	6/4/08	18:37	1	10.023	10	95-105	100%
LINEARITY	6/4/08	18:41	1	0.0034	.0034		
ICSA	6/4/08	18:48	1	0.0007	0.0006	80-120	
ICSAB	6/4/08	18:52	1	.24903	.249	80-120	99.6%
Wash	6/4/08	18:56	1	0.0000	0.0000		
QC-25 1ppm	6/4/08	19:02	1	.99582	1.0		
CCV	6/4/08	19:05	1	5.0574	5.06	90-110	101%
QC-25 1ppm	6/4/08	19:13	1	.99667	1.0		
ICV	6/4/08	20:40	1	10.125	10.1	95-105	101%
QC-25 1ppm	6/4/08	20:42	1	.98522	.990		
CCV	6/4/08	20:46	1	5.0730	5.07	90-110	101%
MCV	6/4/08	20:50	1	2.5589	2.56	90-110	102%
LINEARITY	6/4/08	20:52	1	0.0035	.0035		
ICSA	6/4/08	20:56	1	0.0006	0.0006	80-120	
ICSAB	6/4/08	20:59	1	.25489	.255	80-120	101%
ICB	6/4/08	21:02	1	-0.0001	ND		
MRL	6/4/08	21:05	1	0.0100	.01	50-150	99.7%
MRL/2	6/4/08	21:09	1	0.0051	.0051		
FILTERCHECK	6/4/08	21:13	1	-0.0002	ND		
MBLANK	6/4/08	21:16	1	0.0000	0.0000		
LCS	6/4/08	21:20	1	.93609	.936	85-115	93.6%
LCSD	6/4/08	21:22	1	.95388	.954	85-115	95.3%
2806030669	6/4/08	21:25	1	-0.0001	ND		
2806030669MS	6/4/08	21:29	1	.94774	.948	[0.948]	94.7%
2806030669MSD	6/4/08	21:32	1	.97932	.979	[0.979]	97.9%
2806030669T	6/4/08	21:32	1		1.00	70 - 130	
2806030741	6/4/08	21:35	1	-0.0002	ND		
2806030674	6/4/08	21:39	1	0.0087	.0087		
CCV	6/4/08	21:43	1	4.9646	4.96	90-110	99.2%
CCB	6/4/08	21:46	1	-0.0001	ND		
2806030542	6/4/08	21:50	1	0.0009	0.0009		
2806030544	6/4/08	21:53	1	-0.0004	ND		
2806030545	6/4/08	21:57	1	-0.0003	ND		
2806030547	6/4/08	22:01	1	-0.0003	ND		
2806030844	6/4/08	22:04	1	0.0033	.0033		
2806030859	6/4/08	22:09	1	0.0031	.0031		
2806030867	6/4/08	22:13	1	0.0014	.0014		
2806030671	6/4/08	22:17	1	-0.0002	ND		
2806030671MS	6/4/08	22:22	1	.96488	.965	[0.965]	96.4%
2806030671MSD	6/4/08	22:25	1	.95256	.953	[0.953]	95.2%
2806030671T	6/4/08	22:25	1		1.00	70 - 130	
CCV	6/4/08	22:28	1	5.0573	5.06	90-110	101%
CCB	6/4/08	22:31	1	0.0001	0.0000		
MCV	6/4/08	22:34	1	2.5195	2.52	90-110	100%
2806030660	6/4/08	22:37	1	-0.0002	ND		
2806030870	6/4/08	22:42	1	-0.0003	ND		

Sample ID	Date	Time	Dil	Raw	Rept.	Limit	Comment
2806030871	6/4/08	22:45	1	-0.0004	ND		
2806030872	6/4/08	22:49	1	-0.0006	ND		
2806030570	6/4/08	22:53	1	0.0129	.013		
2806030713	6/4/08	22:56	1	0.0023	.0023		
2806030780	6/4/08	23:01	1	0.0001	0.0000		
2806030672	6/4/08	23:05	1	-0.0002	ND		
D805220919_20X	6/4/08	23:09	20	0.0021	.0021		
MBLANK2007	6/4/08	23:13	1	0.0000	0.0000		
CCV	6/4/08	23:17	1	5.0903	5.09	90-110	101%
CCB	6/4/08	23:20	1	-0.0001	ND		
MRL	6/4/08	23:24	1	0.0103	.0103	50-150	103%
MRL2007	6/4/08	23:27	1	0.0103	.0103		
LCS2007	6/4/08	23:31	1	.94960	.95 ✓	85-115	94.9%
LCSD2007	6/4/08	23:33	1	.90854	.909 ✓	85-115	90.8%
2805290427	6/4/08	23:36	1	0.0026	.0026		
2805290427MS	6/4/08	23:39	1	.98198	.982	[0.979]	97.9%
2805290427MSD	6/4/08	23:43	1	.98733	.987	[0.985]	98.4%
2805290427T	6/4/08	23:43	1		1.00	70 - 130	
2805310013	6/4/08	23:46	1	0.0003	0.0003		
2805310013MS	6/4/08	23:49	1	.98937	.989	[0.989]	98.9%
2805310013MSD	6/4/08	23:52	1	.95237	.952	[0.952]	95.2%
2805310013T	6/4/08	23:52	1		1.00	70 - 130	
CCV	6/4/08	23:56	1	5.0555	5.06	90-110	101%
CCB	6/4/08	23:59	1	0.0000	0.0000		
MCV	6/5/08	0:02	1	2.5238	2.52	90-110	100%
2805280136	6/5/08	0:05	1	0.0008	0.0008		
2805280137	6/5/08	0:09	1	0.0021	.0021		
2805280138	6/5/08	0:13	1	0.0015	.0015		
2805280139	6/5/08	0:16	1	0.0020	.002		
2805280274	6/5/08	0:20	1	0.0014	.0014		
2805280347	6/5/08	0:23	1	0.0016	.0016		
2805280375_2X	6/5/08	0:26	2	0.0707	.071 ✓		
2805280487	6/5/08	0:30	1	0.0002	0.0002		
2805280472	6/5/08	0:33	1	0.0049	.0049		
2805280473	6/5/08	0:37	1	0.0007	0.0007		
CCV	6/5/08	0:40	1	5.0084	5.01	90-110	100%
CCB	6/5/08	0:44	1	0.0001	0.0000		
2805280474	6/5/08	0:47	1	0.0007	0.0007		
2805280475	6/5/08	0:51	1	0.0010	.001		
2805280476	6/5/08	0:54	1	0.0004	0.0004		
2805280478	6/5/08	0:58	1	0.0005	0.0004		
2805280479	6/5/08	1:02	1	0.0010	0.0009		
2805280480	6/5/08	1:06	1	0.0180	.018		
2805280481	6/5/08	1:11	1	0.0009	0.0009		
2805190003	6/5/08	1:15	1	0.0003	0.0002		
ECV	6/5/08	1:18	1	5.0365	5.04	90-110	100%

Nebulizer Parameters: Hg_ReAlign

Analyte	Back Pressure	Flow
All	151.0 kPa	0.54 L/min

6/4/2008 20:12:15 Hg ReAlign... Actual peak offset (nm): 0.003
Drift (nm): 0.000 Slit adjustment: -4

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
------------	------------	-----------

-2.0	15.0	343540.4
-1.6	15.0	460075.9
-1.2	15.0	588083.6
-0.8	15.0	692863.4
-0.4	15.0	743756.4
0.0	15.0	745589.8
0.4	15.0	672804.8
0.8	15.0	581350.3
1.2	15.0	462483.5
1.6	15.0	338482.0
2.0	15.0	254976.4
0.0	10.0	11563.8
0.0	10.5	34422.1
0.0	11.0	54899.6
0.0	11.5	83070.9
0.0	12.0	122164.8
0.0	12.5	246978.8
0.0	13.0	340929.3
0.0	13.5	440124.3
0.0	14.0	554772.1
0.0	14.5	702680.4
0.0	15.0	731146.7
0.0	15.5	714362.3
0.0	16.0	628403.9
0.0	16.5	414536.2
0.0	17.0	311998.3
0.0	17.5	225191.0
0.0	18.0	159424.4
0.0	18.5	106639.6
0.0	19.0	28758.3
0.0	19.5	12564.1
0.0	20.0	5792.6
-0.8	15.0	698245.6
-0.4	15.0	740561.0
0.0	15.0	737766.4
0.4	15.0	676450.2
0.8	15.0	578508.9
-0.4	13.0	345600.3
-0.4	13.5	440744.5
-0.4	14.0	564349.4
-0.4	14.5	738681.6
-0.4	15.0	753180.5
-0.4	15.5	715916.6
-0.4	16.0	623411.4
-0.4	16.5	430745.0
-0.4	17.0	313891.5

6/4/2008 20:16:11 aligned for analyte Mn 257.610

X viewing position set to -0.4 mm having Peak intensity 753180.5 for Axial viewing
Y viewing position set to 15.0 mm having Peak intensity 753180.5 for Axial viewing

Align View XY Axial for analyte Mn 257.610

X-position	Y-position	Intensity
------------	------------	-----------

-2.0	15.0	246200.1
-1.6	15.0	354529.6
-1.2	15.0	472939.5
-0.8	15.0	611661.8
-0.4	15.0	690813.0
0.0	15.0	769802.8

0.4	15.0	750486.8
0.8	15.0	703044.3
1.2	15.0	593096.0
1.6	15.0	477217.9
2.0	15.0	362495.3
0.0	10.0	11041.4
0.0	10.5	34465.0
0.0	11.0	54344.1
0.0	11.5	81783.1
0.0	12.0	121189.6
0.0	12.5	248373.4
0.0	13.0	338386.2
0.0	13.5	451937.9
0.0	14.0	559465.0
0.0	14.5	745083.6
0.0	15.0	761759.8
0.0	15.5	732451.3
0.0	16.0	649186.1
0.0	16.5	424593.4
0.0	17.0	315416.9
0.0	17.5	233928.6
0.0	18.0	164463.3
0.0	18.5	110343.1
0.0	19.0	27098.6
0.0	19.5	11594.5
0.0	20.0	5482.3
-0.8	15.0	616151.3
-0.4	15.0	698152.5
0.0	15.0	755641.2
0.4	15.0	753257.2
0.8	15.0	687496.7
0.0	13.0	357233.5
0.0	13.5	445726.6
0.0	14.0	572299.3
0.0	14.5	740107.5
0.0	15.0	769735.0
0.0	15.5	731468.4
0.0	16.0	633017.1
0.0	16.5	435511.9
0.0	17.0	323247.6

6/4/2008 20:18:50 aligned for analyte Mn 257.610

X viewing position set to 0.0 mm having Peak intensity 769735.0 for Axial viewing
Y viewing position set to 15.0 mm having Peak intensity 769735.0 for Axial viewing

Align View X Radial for analyte Mn 257.610

X-position	Y-position	Intensity
-7.0	15.0	4.0
-6.5	15.0	-2.1
-6.0	15.0	-2.1
-5.5	15.0	-11.4
-5.0	15.0	-16.9
-4.5	15.0	-31.1
-4.0	15.0	-45.5
-3.5	15.0	-59.2
-3.0	15.0	-59.3
-2.5	15.0	-65.2
-2.0	15.0	-73.5
-1.5	15.0	-70.4
-1.0	15.0	-63.2
-0.5	15.0	-85.6
0.0	15.0	-99.0
0.5	15.0	-83.9
1.0	15.0	-54.2
1.5	15.0	-57.7
2.0	15.0	-66.1
2.5	15.0	-64.1
3.0	15.0	-34.9
3.5	15.0	-30.6
4.0	15.0	-24.8
4.5	15.0	-8.7

5.0	15.0	-3.9
5.5	15.0	0.0
6.0	15.0	-0.1
6.5	15.0	3.2
7.0	15.0	-1.6

Align View X Radial for analyte Mn 257.610

X-position	Y-position	Intensity
-7.0	15.0	2224.6
-6.5	15.0	2194.8
-6.0	15.0	2209.5
-5.5	15.0	3540.3
-5.0	15.0	7847.5
-4.5	15.0	16826.7
-4.0	15.0	32788.3
-3.5	15.0	60115.8
-3.0	15.0	96730.8
-2.5	15.0	147339.1
-2.0	15.0	204120.2
-1.5	15.0	257376.2
-1.0	15.0	339902.4
-0.5	15.0	428367.3
0.0	15.0	566160.7
0.5	15.0	596093.9
1.0	15.0	485861.2
1.5	15.0	369733.7
2.0	15.0	253127.2
2.5	15.0	165343.9
3.0	15.0	74181.1
3.5	15.0	29860.7
4.0	15.0	14515.7
4.5	15.0	6843.6
5.0	15.0	3080.1
5.5	15.0	1668.8
6.0	15.0	1177.1
6.5	15.0	974.2
7.0	15.0	889.3

6/4/2008 20:27:20 aligned for analyte Mn 257.610

X viewing position set to 0.5 mm having Peak intensity 596093.9 for Radial viewing

Align View X Radial for analyte Mn 257.610

X-position	Y-position	Intensity
-7.0	15.0	2184.2
-6.5	15.0	2186.6
-6.0	15.0	3488.7
-5.5	15.0	7673.1
-5.0	15.0	16667.3
-4.5	15.0	32841.9
-4.0	15.0	57582.2
-3.5	15.0	97806.6
-3.0	15.0	150905.1
-2.5	15.0	207249.2
-2.0	15.0	256847.1
-1.5	15.0	340237.1
-1.0	15.0	430224.7
-0.5	15.0	562827.2
0.0	15.0	600778.7
0.5	15.0	489994.5
1.0	15.0	368452.6
1.5	15.0	249490.6
2.0	15.0	169169.8
2.5	15.0	74299.0
3.0	15.0	30353.5
3.5	15.0	14233.0
4.0	15.0	6917.0
4.5	15.0	3060.8
5.0	15.0	1619.9
5.5	15.0	1176.3
6.0	15.0	964.2
6.5	15.0	869.1

7.0 15.0 743.3

6/4/2008 20:29:53 aligned for analyte Mn 257.610
X viewing position set to 0.0 mm having Peak intensity 600778.7 for Radial viewing

=====

Analysis Begun

Start Time: 6/4/2008 20:34:55 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 1 Autosampler Location: 0
 Sample ID: Calib Blank 1 Date Collected: 6/4/2008 20:35:12
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: Calib Blank 1
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: Calib Blank 1

Analyte	Mean Corrected				Calib
	Intensity	Std.Dev.	RSD	Conc. Units	
Sca	422000.9	644.36	0.15%	100 %	
Yr	425859.0	345.18	0.08%	100 %	
Agt	295.3	34.01	11.52%	[0.00] mg/L	
Alt	36.0	16.62	46.21%	[0.00] mg/L	
Ast	-1.0	0.89	86.10%	[0.00] mg/L	
B _t	187.8	0.69	0.37%	[0.00] mg/L	
Bat	-22.0	2.98	13.57%	[0.00] mg/L	
Bet	-3825.8	6.46	0.17%	[0.00] mg/L	
Cat	2968.6	0.57	0.02%	[0.00] mg/L	
Cdt	37.3	1.59	4.25%	[0.00] mg/L	
Cot	-44.3	0.53	1.20%	[0.00] mg/L	
Crt	101.8	0.49	0.48%	[0.00] mg/L	
Cut	3446.0	26.33	0.76%	[0.00] mg/L	
Fet	-97.0	0.06	0.06%	[0.00] mg/L	
Kt	205.3	14.62	7.12%	[0.00] mg/L	
Mgt	-509.5	18.63	3.66%	[0.00] mg/L	
Mnt	297.3	5.73	1.93%	[0.00] mg/L	
Mot	15.8	1.19	7.55%	[0.00] mg/L	
Nat	-169.0	53.24	31.51%	[0.00] mg/L	
Nit	-45.7	0.53	1.16%	[0.00] mg/L	
Pbt	-20.7	0.96	4.64%	[0.00] mg/L	
Sbt	6.7	4.72	70.19%	[0.00] mg/L	
Set	-5.0	0.69	13.97%	[0.00] mg/L	
Tlf	-22.1	0.99	4.47%	[0.00] mg/L	
Vt	147.9	7.14	4.83%	[0.00] mg/L	
Znt	132.9	7.07	5.32%	[0.00] mg/L	

User canceled analysis.

Analysis Begun

Start Time: 6/4/2008 20:38:24 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 2 Autosampler Location: 15
 Sample ID: Standard 2 Date Collected: 6/4/2008 20:38:26
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: Standard 2
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: Standard 2

Analyte	Mean Corrected Intensity	Std. Dev.	RSD	Conc.	Calib Units
Sca	376984.2	1978.60	0.52%	89.3	%
Yr	405968.8	1854.89	0.46%	95.3	%
Agt	515840.8	112.71	0.02%	[2]	mg/L
Alt	68030.7	6.26	0.01%	[10]	mg/L
Ast	15999.6	182.44	1.14%	[10]	mg/L
B_f	138257.1	115.45	0.08%	[5.02]	mg/L
Bat	394426.0	423.74	0.11%	[10]	mg/L
Bet	10826990.0	96215.23	0.89%	[4.01]	mg/L
Cat	2614885.4	11951.72	0.46%	[100]	mg/L
Cdt	109603.9	272.08	0.25%	[5.01]	mg/L
Cot	155647.2	790.05	0.51%	[10]	mg/L
Crt	449776.5	1494.44	0.33%	[9.97]	mg/L
Cut	3090424.8	11855.00	0.38%	[10]	mg/L
Fet	107061.2	806.81	0.75%	[9.98]	mg/L
Kt	104928.5	143.55	0.14%	[100]	mg/L
Mgt	1966712.1	10191.14	0.52%	[100]	mg/L
Mnt	4977338.1	14003.55	0.28%	[10]	mg/L
Mot	102258.2	297.64	0.29%	[9.98]	mg/L
Nat	559866.5	468.08	0.08%	[100]	mg/L
Nit	191654.0	612.72	0.32%	[10]	mg/L
Pbt	39215.8	272.52	0.69%	[10]	mg/L
Sbt	16134.5	130.02	0.81%	[10]	mg/L
Set	10122.1	69.14	0.68%	[10]	mg/L
Tlt	22093.6	89.43	0.40%	[10]	mg/L
Vt	1551790.5	2953.74	0.19%	[10]	mg/L
Znt	390786.8	897.26	0.23%	[10]	mg/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag	1	Lin, Calc Int	0.0	257900	0.00000	1.000000	
Al	1	Lin, Calc Int	0.0	6803	0.00000	1.000000	
As	1	Lin, Calc Int	0.0	1600	0.00000	1.000000	
B_f	1	Lin, Calc Int	0.0	27540	0.00000	1.000000	
Ba	1	Lin, Calc Int	0.0	39440	0.00000	1.000000	
Be	1	Lin, Calc Int	-0.0	2700000	0.00000	1.000000	
Ca	1	Lin, Calc Int	0.0	26150	0.00000	1.000000	
Cd	1	Lin, Calc Int	0.0	21880	0.00000	1.000000	
Co	1	Lin, Calc Int	0.0	15560	0.00000	1.000000	
Cr	1	Lin, Calc Int	-0.0	45110	0.00000	1.000000	
Cu	1	Lin, Calc Int	0.0	309000	0.00000	1.000000	

Method: 200.746010 080304

Page 3

Date: 6/4/2008 20:40:05

Fe	1	Lin, Calc Int	0.0	10730	0.00000	1.000000
K	1	Lin, Calc Int	0.0	1049	0.00000	1.000000
Mg	1	Lin, Calc Int	0.0	19670	0.00000	1.000000
Mn	1	Lin, Calc Int	0.0	497700	0.00000	1.000000
Mo	1	Lin, Calc Int	0.0	10250	0.00000	1.000000
Na	1	Lin, Calc Int	0.0	5599	0.00000	1.000000
Ni	1	Lin, Calc Int	0.0	19170	0.00000	1.000000
Pb	1	Lin, Calc Int	0.0	3922	0.00000	1.000000
Sb	1	Lin, Calc Int	-0.0	1613	0.00000	1.000000
Se	1	Lin, Calc Int	0.0	1012	0.00000	1.000000
Tl	1	Lin, Calc Int	0.0	2209	0.00000	1.000000
V	1	Lin, Calc Int	0.0	155200	0.00000	1.000000
Zn	1	Lin, Calc Int	0.0	39080	0.00000	1.000000

Analysis Begun

Start Time: 6/4/2008 20:40:20 Plasma On Time: 6/4/2008 20:11:34
Logged In Analyst: Charley Kay Technique: ICP Continuous
Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
Batch ID: 080604A
Results Data Set: 080604A
Results Library: C:\pe\Charley Kay\Results\Results.mdb

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
All 211.0 kPa 0.65 L/min

Mean Data: ICV

Analyte	Mean Corrected		Calib.		Sample			
	Intensity	Conc.	Units	Std.Dev.	Conc.	Units	Std.Dev.	RSD
Sca	375464.5	89.0	%	0.37				0.41%
Yr	401929.5	94.4	%	1.52				1.61%
Agt	520838.9	2.02	mg/L	0.003	2.02	mg/L	0.003	0.16%
	QC value within limits for Ag	Recovery = 101.16%						
Alt	68148.7	9.65	mg/L	0.039	9.65	mg/L	0.039	0.41%
	QC value within limits for Al	Recovery = 96.48%						
Ast	16403.6	10.3	mg/L	0.02	10.3	mg/L	0.02	0.22%
	QC value within limits for As	Recovery = 102.53%						
B_+	141441.4	5.05	mg/L	0.004	5.05	mg/L	0.004	0.08%
	QC value within limits for B_	Recovery = 100.94%						
Bat	397024.2	10.1	mg/L	0.03	10.1	mg/L	0.03	0.27%
	QC value within limits for Ba	Recovery = 100.66%						
Bet	10914575.8	4.04	mg/L	0.002	4.04	mg/L	0.002	0.05%
	QC value within limits for Be	Recovery = 101.10%						
Cat	2698233.1	103	mg/L	0.7	103	mg/L	0.7	0.66%
	QC value within limits for Ca	Recovery = 103.19%						
Cdt	110915.1	4.96	mg/L	0.017	4.96	mg/L	0.017	0.33%
	QC value within limits for Cd	Recovery = 99.19%						
Cot	158105.3	10.2	mg/L	0.07	10.2	mg/L	0.07	0.71%
	QC value within limits for Co	Recovery = 101.58%						
CrI	456782.4	10.1	mg/L	0.06	10.1	mg/L	0.06	0.57%
	QC value within limits for Cr	Recovery = 101.25%						
Cut	3113662.5	10.1	mg/L	0.02	10.1	mg/L	0.02	0.16%
	QC value within limits for Cu	Recovery = 100.85%						
Fet	110231.3	10.3	mg/L	0.01	10.3	mg/L	0.01	0.06%
	QC value within limits for Fe	Recovery = 102.76%						
Kt	105182.4	100	mg/L	0.5	100	mg/L	0.5	0.50%
	QC value within limits for K	Recovery = 100.24%						
Mgt	1998672.0	102	mg/L	1.0	102	mg/L	1.0	0.96%
	QC value within limits for Mg	Recovery = 101.64%						
Mnt	5039187.8	10.1	mg/L	0.05	10.1	mg/L	0.05	0.47%
	QC value within limits for Mn	Recovery = 101.22%						
Mot	103704.7	10.1	mg/L	0.04	10.1	mg/L	0.04	0.37%
	QC value within limits for Mo	Recovery = 101.21%						
Nat	541850.8	96.8	mg/L	0.13	96.8	mg/L	0.13	0.13%
	QC value within limits for Na	Recovery = 96.78%						
Nit	194268.1	10.1	mg/L	0.04	10.1	mg/L	0.04	0.36%
	QC value within limits for Ni	Recovery = 101.36%						
Pbt	40106.4	10.2	mg/L	0.01	10.2	mg/L	0.01	0.12%
	QC value within limits for Pb	Recovery = 102.27%						
Sbt	16519.3	10.3	mg/L	0.02	10.3	mg/L	0.02	0.16%
	QC value within limits for Sb	Recovery = 102.79%						
Set	10305.0	10.2	mg/L	0.02	10.2	mg/L	0.02	0.24%

	QC value within limits for Se	Recovery = 102.06%					
T1f	22431.9	10.2 mg/L	0.01	10.2 mg/L	0.01	0.08%	
Vt	1572685.0	10.2 mg/L	0.02	10.2 mg/L	0.02	0.23%	
Znf	396771.7	10.1 mg/L	0.04	10.1 mg/L	0.04	0.40%	
QC value within limits for Zn Recovery = 100.66%							
All analyte(s) passed QC.							

User canceled analysis.

Analysis Begun

Start Time: 6/4/2008 20:42:44 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 6 Autosampler Location: 12
 Sample ID: QC-25 1ppm Date Collected: 6/4/2008 20:42:45
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: QC-25 1ppm

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: QC-25 1ppm

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	425106.3	101 %	0.9			0.93%
Yr	429250.9	101 %	0.2			0.24%
Ag†	239900.9	0.931 mg/L	0.0095	0.931 mg/L	0.0095	1.02%
QC value within limits for Ag		Recovery = 93.05%				
Alt†	6683.5	0.948 mg/L	0.0068	0.948 mg/L	0.0068	0.71%
QC value within limits for Al		Recovery = 94.77%				
Ast†	1461.5	0.913 mg/L	0.0125	0.913 mg/L	0.0125	1.37%
QC value within limits for As		Recovery = 91.35%				
B_t	27204.2	0.979 mg/L	0.0172	0.979 mg/L	0.0172	1.76%
QC value within limits for B		Recovery = 97.91%				
Bat	40141.4	1.02 mg/L	0.010	1.02 mg/L	0.010	0.97%
QC value within limits for Ba		Recovery = 101.77%				
Bet†	2601174.6	0.964 mg/L	0.0043	0.964 mg/L	0.0043	0.45%
QC value within limits for Be		Recovery = 96.35%				
Cat	27109.0	1.04 mg/L	0.001	1.04 mg/L	0.001	0.11%
QC value within limits for Ca		Recovery = 103.67%				
Cdt†	20766.5	0.940 mg/L	0.0080	0.940 mg/L	0.0080	0.85%
QC value within limits for Cd		Recovery = 93.98%				
Cot†	16149.7	1.04 mg/L	0.012	1.04 mg/L	0.012	1.15%
QC value within limits for Co		Recovery = 103.76%				
Crt†	44446.5	0.985 mg/L	0.0060	0.985 mg/L	0.0060	0.60%
QC value within limits for Cr		Recovery = 98.52%				
Cut†	299653.0	0.971 mg/L	0.0010	0.971 mg/L	0.0010	0.10%
QC value within limits for Cu		Recovery = 97.06%				
Fet†	11068.3	1.03 mg/L	0.002	1.03 mg/L	0.002	0.16%
QC value within limits for Fe		Recovery = 103.18%				
Kt†	9907.3	9.44 mg/L	0.035	9.44 mg/L	0.035	0.37%
QC value within limits for K		Recovery = 94.42%				
Mgt†	20580.9	1.05 mg/L	0.002	1.05 mg/L	0.002	0.23%
QC value within limits for Mg		Recovery = 104.83%				
Mnt†	507585.1	1.02 mg/L	0.006	1.02 mg/L	0.006	0.57%
QC value within limits for Mn		Recovery = 101.98%				
Mot†	9742.6	0.951 mg/L	0.0113	0.951 mg/L	0.0113	1.19%
QC value within limits for Mo		Recovery = 95.08%				
Nat†	5198.2	0.928 mg/L	0.0007	0.928 mg/L	0.0007	0.07%
QC value within limits for Na		Recovery = 92.85%				
Nit†	20110.1	1.05 mg/L	0.010	1.05 mg/L	0.010	0.93%
QC value within limits for Ni		Recovery = 104.93%				
Pbt†	4094.5	1.04 mg/L	0.015	1.04 mg/L	0.015	1.48%
QC value within limits for Pb		Recovery = 104.41%				
Sbt†	1487.1	0.925 mg/L	0.0142	0.925 mg/L	0.0142	1.54%
QC value within limits for Sb		Recovery = 92.53%				

Method: 200.7&6010 080304

Page 4

Date: 6/4/2008 20:44:55

Set	945.6	0.937 mg/L	0.0149	0.937 mg/L	0.0149	1.59%
QC value within limits for Se Recovery = 93.67%						
Tlt	2360.1	1.07 mg/L	0.014	1.07 mg/L	0.014	1.26%
QC value within limits for Tl Recovery = 107.39%						
Vt	145942.2	0.946 mg/L	0.0099	0.946 mg/L	0.0099	1.05%
QC value within limits for V Recovery = 94.56%						
Znt	39926.1	1.01 mg/L	0.010	1.01 mg/L	0.010	0.95%
QC value within limits for Zn Recovery = 101.27%						
All analyte(s) passed QC.						

Sequence No.: 7
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/4/2008 20:46:34
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	RSD
Sca	396065.7	93.9 %	0.27		0.29%
Yr	417666.2	98.1 %	0.97		0.99%
Agt	249187.4	0.968 mg/L	0.0048	0.968 mg/L	0.0048 0.50%
QC value within limits for Ag		Recovery = 96.81%			
Alt	33462.3	4.74 mg/L	0.009	4.74 mg/L	0.009 0.20%
QC value within limits for Al		Recovery = 94.72%			
Ast	7706.2	4.82 mg/L	0.008	4.82 mg/L	0.008 0.16%
QC value within limits for As		Recovery = 96.33%			
B _t	70125.9	2.50 mg/L	0.004	2.50 mg/L	0.004 0.16%
QC value within limits for B		Recovery = 100.11%			
Bat	198424.8	5.03 mg/L	0.038	5.03 mg/L	0.038 0.76%
QC value within limits for Ba		Recovery = 100.61%			
Bet	5537200.1	2.05 mg/L	0.000	2.05 mg/L	0.000 0.01%
QC value within limits for Be		Recovery = 102.58%			
Cat	1346840.9	51.5 mg/L	0.72	51.5 mg/L	0.72 1.40%
QC value within limits for Ca		Recovery = 103.01%			
Cdt	43939.3	1.96 mg/L	0.011	1.96 mg/L	0.011 0.55%
QC value within limits for Cd		Recovery = 97.89%			
Cot	80181.1	5.15 mg/L	0.041	5.15 mg/L	0.041 0.79%
QC value within limits for Co		Recovery = 103.03%			
Crt	228859.0	5.07 mg/L	0.063	5.07 mg/L	0.063 1.25%
QC value within limits for Cr		Recovery = 101.46%			
Cut	1521268.1	4.93 mg/L	0.015	4.93 mg/L	0.015 0.31%
QC value within limits for Cu		Recovery = 98.55%			
Fet	55672.4	5.19 mg/L	0.009	5.19 mg/L	0.009 0.17%
QC value within limits for Fe		Recovery = 103.79%			
Kt	50312.2	47.9 mg/L	0.07	47.9 mg/L	0.07 0.16%
QC value within limits for K		Recovery = 95.90%			
Mgt	1000966.9	50.9 mg/L	0.64	50.9 mg/L	0.64 1.26%
QC value within limits for Mg		Recovery = 101.81%			
Mnt	2524111.1	5.07 mg/L	0.039	5.07 mg/L	0.039 0.76%
QC value within limits for Mn		Recovery = 101.40%			
Mot	51316.2	5.01 mg/L	0.034	5.01 mg/L	0.034 0.68%
QC value within limits for Mo		Recovery = 100.17%			
Nat	257333.4	46.0 mg/L	0.16	46.0 mg/L	0.16 0.36%
QC value within limits for Na		Recovery = 91.93%			
Nit	98666.7	5.15 mg/L	0.036	5.15 mg/L	0.036 0.70%
QC value within limits for Ni		Recovery = 102.96%			
Pbt	20121.1	5.13 mg/L	0.011	5.13 mg/L	0.011 0.21%
QC value within limits for Pb		Recovery = 102.62%			
Sbt	7890.1	4.91 mg/L	0.010	4.91 mg/L	0.010 0.21%
QC value within limits for Sb		Recovery = 98.20%			
Set	5059.9	5.01 mg/L	0.033	5.01 mg/L	0.033 0.65%
QC value within limits for Se		Recovery = 100.23%			
Tlt	11572.7	5.27 mg/L	0.024	5.27 mg/L	0.024 0.45%
QC value within limits for Tl		Recovery = 105.33%			
Vt	772895.1	5.01 mg/L	0.033	5.01 mg/L	0.033 0.66%
QC value within limits for V		Recovery = 100.15%			
Znt	202414.8	5.14 mg/L	0.035	5.14 mg/L	0.035 0.68%
QC value within limits for Zn		Recovery = 102.71%			

All analyte(s) passed QC.

User canceled analysis.

=====

Analysis Begun

Start Time: 6/4/2008 20:50:17 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 34 Autosampler Location: 5
 Sample ID: MCV Date Collected: 6/4/2008 20:50:18
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

=====

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

=====

Mean Data: MCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std. Dev.	Sample Conc. Units	Std. Dev.	RSD
Sca	401185.4	95.1 %	0.09			0.09%
Yr	419131.2	98.4 %	0.12			0.12%
Agt	124418.4	0.483 mg/L	0.0007	0.483 mg/L	0.0007	0.15%
QC value within limits for Ag		Recovery = 96.67%				
B_t	35064.4	1.26 mg/L	0.003	1.26 mg/L	0.003	0.24%
QC value within limits for B_t		Recovery = 100.58%				
Bat	99884.6	2.53 mg/L	0.018	2.53 mg/L	0.018	0.73%
QC value within limits for Ba		Recovery = 101.30%				
Bet	2734664.6	1.01 mg/L	0.001	1.01 mg/L	0.001	0.10%
QC value within limits for Be		Recovery = 101.32%				
Cat	665980.3	25.5 mg/L	0.08	25.5 mg/L	0.08	0.31%
QC value within limits for Ca		Recovery = 101.88%				
Cdt	21946.8	1.00 mg/L	0.002	1.00 mg/L	0.002	0.21%
QC value within limits for Cd		Recovery = 100.32%				
Crt	115439.7	2.56 mg/L	0.001	2.56 mg/L	0.001	0.06%
QC value within limits for Cr		Recovery = 102.36%				
Cut	760263.8	2.46 mg/L	0.000	2.46 mg/L	0.000	0.01%
QC value within limits for Cu		Recovery = 98.44%				
Fet	27814.7	2.59 mg/L	0.000	2.59 mg/L	0.000	0.02%
QC value within limits for Fe		Recovery = 103.71%				
Kt	24685.3	23.5 mg/L	0.06	23.5 mg/L	0.06	0.24%
QC value within limits for K		Recovery = 94.10%				
Mgt	498303.1	25.3 mg/L	0.02	25.3 mg/L	0.02	0.09%
QC value within limits for Mg		Recovery = 101.37%				
Mnt	1272602.3	2.56 mg/L	0.003	2.56 mg/L	0.003	0.12%
QC value within limits for Mn		Recovery = 102.25%				
Mot	25728.0	2.51 mg/L	0.008	2.51 mg/L	0.008	0.30%
QC value within limits for Mo		Recovery = 100.44%				
Nat	126006.3	22.5 mg/L	0.01	22.5 mg/L	0.01	0.03%
QC value within limits for Na		Recovery = 90.03%				
Nit	49989.4	2.61 mg/L	0.010	2.61 mg/L	0.010	0.38%
QC value within limits for Ni		Recovery = 104.33%				
Vt	385900.1	2.50 mg/L	0.012	2.50 mg/L	0.012	0.49%
QC value within limits for V		Recovery = 100.01%				
Znt	102119.6	2.59 mg/L	0.015	2.59 mg/L	0.015	0.57%
QC value within limits for Zn		Recovery = 103.63%				

All analyte(s) passed QC.

User canceled analysis.

Analysis Begun

Start Time: 6/4/2008 20:52:16 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 2 Autosampler Location: 9
 Sample ID: LINEARITY Date Collected: 6/4/2008 20:52:18
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: LINEARITY

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: LINEARITY

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	359449.8	85.2 %	0.61			0.72%
Yr	385745.6	90.6 %	0.54			0.60%
Ag†	-5244.3	0.0161 mg/L	0.00003	0.0161 mg/L	0.00003	0.16%
B†	508.4	0.0185 mg/L	0.00401	0.0185 mg/L	0.00401	21.64%
B-†	QC value within limits for B-	Recovery = Not calculated				
Ba†	51.4	0.00130 mg/L	0.000032	0.00130 mg/L	0.000032	2.47%
Ba†	QC value within limits for Ba	Recovery = Not calculated				
Bet	-662.9	-0.00024 mg/L	0.000017	-0.00024 mg/L	0.000017	6.94%
Be†	QC value within limits for Be	Recovery = Not calculated				
Cat	7763202.2	297 mg/L	1.5	297 mg/L	1.5	0.51%
Ca†	QC value within limits for Ca	Recovery = 98.96%				
Cdt	-22.0	-0.00100 mg/L	0.000194	-0.00100 mg/L	0.000194	19.35%
Cd†	QC value within limits for Cd	Recovery = Not calculated				
Crt	156.6	0.00347 mg/L	0.000019	0.00347 mg/L	0.000019	0.56%
Cr†	QC value within limits for Cr	Recovery = Not calculated				
Cut	-3211.4	-0.0104 mg/L	0.00005	-0.0104 mg/L	0.00005	0.47%
Cu†	QC value within limits for Cu	Recovery = Not calculated				
Fet	1062042.9	99.0 mg/L	0.69	99.0 mg/L	0.69	0.69%
Fe†	QC value within limits for Fe	Recovery = 99.00%				
K†	316799.8	302 mg/L	0.8	302 mg/L	0.8	0.27%
K†	QC value within limits for K	Recovery = 100.64%				
Mgt	3664440.1	186 mg/L	0.3	186 mg/L	0.3	0.18%
Mg†	QC value within limits for Mg	Recovery = Not calculated				
Mnt	-1199.6	-0.00660 mg/L	0.000030	-0.00660 mg/L	0.000030	0.45%
Mn†	QC value within limits for Mn	Recovery = Not calculated				
Mot	26.2	0.00256 mg/L	0.001313	0.00256 mg/L	0.001313	51.41%
Mo†	QC value within limits for Mo	Recovery = Not calculated				
Nat	1548032.6	277 mg/L	0.2	277 mg/L	0.2	0.08%
Na†	QC value within limits for Na	Recovery = 92.17%				
Nit	10.9	0.00057 mg/L	0.000100	0.00057 mg/L	0.000100	17.62%
Ni†	QC value within limits for Ni	Recovery = Not calculated				
V†	-485.0	0.00077 mg/L	0.000068	0.00077 mg/L	0.000068	8.79%
V†	QC value within limits for V	Recovery = Not calculated				
Znt	944.9	0.0242 mg/L	0.00000	0.0242 mg/L	0.00000	0.01%
Zn†	QC value within limits for Zn	Recovery = Not calculated				

All analyte(s) passed QC.

Sequence No.: 3
 Sample ID: ICSA
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 10
 Date Collected: 6/4/2008 20:56:01
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSA

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICSA

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	RSD
Sca	375556.5	89.0 %	0.02		0.03%
Yr	399465.7	93.8 %	0.69		0.74%
Agt	-5132.4	0.0165 mg/L	0.00012	0.00012	0.75%
B _t	-11.2	Recovery = Not calculated			
B _t	-11.2	-0.00032 mg/L	0.000697	-0.00032 mg/L	0.000697 218.90%
Bat	68.0	Recovery = Not calculated			
Bat	68.0	0.00172 mg/L	0.000034	0.00172 mg/L	0.000034 1.98%
Bet	-731.7	Recovery = Not calculated			
Bet	-731.7	-0.00027 mg/L	0.000026	-0.00027 mg/L	0.000026 9.46%
Cat	6596915.5	252 mg/L	0.5	252 mg/L	0.5 0.18%
Cat	6596915.5	Recovery = 100.91%			
Cdt	-30.4	Recovery = Not calculated			
Cdt	-30.4	-0.00139 mg/L	0.000107	-0.00139 mg/L	0.000107 7.68%
Crt	27.8	Recovery = Not calculated			
Crt	27.8	0.00062 mg/L	0.000006	0.00062 mg/L	0.000006 0.92%
Cut	-3454.7	Recovery = Not calculated			
Cut	-3454.7	-0.0112 mg/L	0.00009	-0.0112 mg/L	0.00009 0.85%
Fet	1060128.8	Recovery = Not calculated			
Fet	1060128.8	98.8 mg/L	0.46	98.8 mg/L	0.46 0.47%
Kt	311.8	Recovery = 98.82%			
Kt	311.8	0.297 mg/L	0.0006	0.297 mg/L	0.0006 0.19%
Mgt	4549744.6	Recovery = Not calculated			
Mgt	4549744.6	232 mg/L	0.3	232 mg/L	0.3 0.12%
Mnt	1275.6	Recovery = 92.61%			
Mnt	1275.6	-0.00776 mg/L	0.000038	-0.00776 mg/L	0.000038 0.49%
Mot	1.2	Recovery = Not calculated			
Mot	1.2	0.00011 mg/L	0.000717	0.00011 mg/L	0.000717 628.43%
Nat	882.3	Recovery = Not calculated			
Nat	882.3	0.158 mg/L	0.0028	0.158 mg/L	0.0028 1.79%
Nit	-21.8	Recovery = Not calculated			
Nit	-21.8	-0.00114 mg/L	0.000142	-0.00114 mg/L	0.000142 12.45%
Vt	-546.2	Recovery = Not calculated			
Vt	-546.2	0.00036 mg/L	0.000138	0.00036 mg/L	0.000138 38.50%
Znt	723.8	Recovery = Not calculated			
Znt	723.8	0.0185 mg/L	0.00024	0.0185 mg/L	0.00024 1.32%
All analyte(s) passed QC.					

Sequence No.: 4
 Sample ID: ICSAB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 11
 Date Collected: 6/4/2008 20:59:45
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ICSAB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICSAB

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	RSD
Sca	373398.6	88.5 %	0.55		0.62%
Yr	399262.3	93.8 %	0.17		0.18%
Ag†	50826.2	0.234 mg/L	0.0102	0.234 mg/L	0.0102 4.35%
	QC value less than the lower limit for Ag	Recovery = 46.72%			
B_†	-149.7	-0.00701 mg/L	0.000677	-0.00701 mg/L	0.000677 9.66%
	QC value within limits for B_	Recovery = Not calculated			
Ba†	10279.7	0.261 mg/L	0.0017	0.261 mg/L	0.0017 0.66%
	QC value within limits for Ba	Recovery = 104.25%			
Bet	662643.3	0.245 mg/L	0.0019	0.245 mg/L	0.0019 0.79%
	QC value within limits for Be	Recovery = 98.18%			
Ca†	6669828.3	255 mg/L	0.1	255 mg/L	0.1 0.03%
	QC value within limits for Ca	Recovery = 102.03%			
Cdt	10903.3	0.498 mg/L	0.0033	0.498 mg/L	0.0033 0.65%
	QC value within limits for Cd	Recovery = 99.68%			
Cr†	11498.9	0.255 mg/L	0.0012	0.255 mg/L	0.0012 0.47%
	QC value within limits for Cr	Recovery = 101.96%			
Cu†	74923.6	0.243 mg/L	0.0007	0.243 mg/L	0.0007 0.30%
	QC value within limits for Cu	Recovery = 97.01%			
Fet	1065698.6	99.3 mg/L	0.22	99.3 mg/L	0.22 0.22%
	QC value within limits for Fe	Recovery = 99.34%			
K†	124.9	0.119 mg/L	0.0073	0.119 mg/L	0.0073 6.16%
	QC value within limits for K	Recovery = Not calculated			
Mg†	4588463.1	233 mg/L	0.1	233 mg/L	0.1 0.05%
	QC value within limits for Mg	Recovery = 93.39%			
Mn†	125575.4	0.247 mg/L	0.0006	0.247 mg/L	0.0006 0.26%
	QC value within limits for Mn	Recovery = 98.82%			
Mo†	-3.4	-0.00033 mg/L	0.000657	-0.00033 mg/L	0.000657 196.94%
	QC value within limits for Mo	Recovery = Not calculated			
Nat	549.8	0.0982 mg/L	0.01487	0.0982 mg/L	0.01487 15.14%
	QC value within limits for Na	Recovery = Not calculated			
Nit	9045.8	0.472 mg/L	0.0013	0.472 mg/L	0.0013 0.27%
	QC value within limits for Ni	Recovery = 94.40%			
V†	37979.2	0.250 mg/L	0.0013	0.250 mg/L	0.0013 0.52%
	QC value within limits for V	Recovery = 99.95%			
Zn†	21390.1	0.544 mg/L	0.0036	0.544 mg/L	0.0036 0.66%
	QC value within limits for Zn	Recovery = 108.75%			
	QC Failed. Continue with analysis.				

User canceled analysis.

Analysis Begun

Start Time: 6/4/2008 21:02:34 Plasma On Time: 6/4/2008 20:11:34
 Logged In Analyst: Charley Kay Technique: ICP Continuous
 Spectrometer Model: Optima 4300 DV, S/N 077N2121801 Autosampler Model: AS-93plus

Sample Information File: C:\pe\Charley Kay\Sample Information\080604A.sif
 Batch ID: 080604A
 Results Data Set: 080604A
 Results Library: C:\pe\Charley Kay\Results\Results.mdb

Sequence No.: 8 Autosampler Location: 0
 Sample ID: ICB Date Collected: 6/4/2008 21:02:35
 Analyst: Data Type: Original
 Initial Sample Wt: Initial Sample Vol:
 Dilution: Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: ICB

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	417360.6	98.9 %	0.82			0.83%
Yr	432815.3	102 %	0.8			0.81%
Agt	198.5	0.00077 mg/L	0.000117	0.00077 mg/L	0.000117	15.18%
QC value within limits for Ag		Recovery = Not calculated				
B_t	95.1	0.00345 mg/L	0.000122	0.00345 mg/L	0.000122	3.52%
QC value within limits for B_t		Recovery = Not calculated				
Bat	2.4	0.00006 mg/L	0.000105	0.00006 mg/L	0.000105	174.79%
QC value within limits for Ba		Recovery = Not calculated				
Bet	240.0	0.00009 mg/L	0.000029	0.00009 mg/L	0.000029	32.51%
QC value within limits for Be		Recovery = Not calculated				
Cat	284.0	0.0109 mg/L	0.00556	0.0109 mg/L	0.00556	51.17%
QC value within limits for Ca		Recovery = Not calculated				
Cdt	2.5	0.00011 mg/L	0.000025	0.00011 mg/L	0.000025	21.85%
QC value within limits for Cd		Recovery = Not calculated				
Crt	-3.4	-0.00008 mg/L	0.000143	-0.00008 mg/L	0.000143	188.16%
QC value within limits for Cr		Recovery = Not calculated				
Cut	106.0	0.00034 mg/L	0.000311	0.00034 mg/L	0.000311	90.69%
QC value within limits for Cu		Recovery = Not calculated				
Fet	51.5	0.00480 mg/L	0.002965	0.00480 mg/L	0.002965	61.77%
QC value within limits for Fe		Recovery = Not calculated				
Kt	116.1	0.111 mg/L	0.0145	0.111 mg/L	0.0145	13.15%
QC value within limits for K		Recovery = Not calculated				
Mgt	144.5	0.00736 mg/L	0.005156	0.00736 mg/L	0.005156	70.07%
QC value within limits for Mg		Recovery = Not calculated				
Mnt	294.5	0.00059 mg/L	0.000006	0.00059 mg/L	0.000006	1.07%
QC value within limits for Mn		Recovery = Not calculated				
Mot	1.6	0.00015 mg/L	0.000073	0.00015 mg/L	0.000073	47.76%
QC value within limits for Mo		Recovery = Not calculated				
Nat	337.8	0.0603 mg/L	0.00509	0.0603 mg/L	0.00509	8.43%
QC value within limits for Na		Recovery = Not calculated				
Nit	-0.9	-0.00005 mg/L	0.000041	-0.00005 mg/L	0.000041	87.58%
QC value within limits for Ni		Recovery = Not calculated				
Vt	6.0	0.00004 mg/L	0.000068	0.00004 mg/L	0.000068	176.21%
QC value within limits for V		Recovery = Not calculated				
Znt	2.5	0.00007 mg/L	0.000084	0.00007 mg/L	0.000084	128.12%
QC value within limits for Zn		Recovery = Not calculated				

All analyte(s) passed QC.

Sequence No.: 44
 Sample ID: MBLANK2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 65
 Date Collected: 6/4/2008 23:13:58
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MBLANK2007

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: MBLANK2007

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	420380.9	99.6 %	0.68			0.68%
Yr	433807.5	102 %	0.5			0.54%
Agt	-17.7	-0.00007 mg/L	0.000065	-0.00007 mg/L	0.000065	94.88%
B_t	-31.9	-0.00116 mg/L	0.000286	-0.00116 mg/L	0.000286	24.60%
Bat	1.3	0.00003 mg/L	0.000103	0.00003 mg/L	0.000103	309.86%
Bet	43.4	0.00002 mg/L	0.000033	0.00002 mg/L	0.000033	207.59%
Cat	663.1	0.0254 mg/L	0.00143	0.0254 mg/L	0.00143	5.66%
Cdt	0.8	0.00003 mg/L	0.000132	0.00003 mg/L	0.000132	385.65%
Crt	0.9	0.00002 mg/L	0.000136	0.00002 mg/L	0.000136	663.31%
Cut	186.4	0.00060 mg/L	0.000068	0.00060 mg/L	0.000068	11.33%
Fet	10.2	0.00095 mg/L	0.000071	0.00095 mg/L	0.000071	7.46%
Kt	49.5	0.0471 mg/L	0.00848	0.0471 mg/L	0.00848	17.99%
Mgt	59.2	0.00301 mg/L	0.000612	0.00301 mg/L	0.000612	20.35%
Mnt	-76.1	-0.00015 mg/L	0.000000	-0.00015 mg/L	0.000000	0.00%
Mot	-1.6	-0.00016 mg/L	0.000058	-0.00016 mg/L	0.000058	36.76%
Nat	531.1	0.0949 mg/L	0.00352	0.0949 mg/L	0.00352	3.72%
Nit	-5.7	-0.00030 mg/L	0.000316	-0.00030 mg/L	0.000316	106.60%
Vt	-17.7	-0.00011 mg/L	0.000060	-0.00011 mg/L	0.000060	52.42%
Znt	2.3	0.00006 mg/L	0.000123	0.00006 mg/L	0.000123	204.24%

Sequence No.: 45
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/4/2008 23:17:36
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std. Dev.	RSD
	Conc. Units	Units	Conc. Units	Std. Dev.	
Sca	380439.2	90.2 %	0.09		0.10%
Yr	408555.1	95.9 %	2.62		2.73%
Ag†	250532.6	0.973 mg/L	0.0045	0.973 mg/L	0.0045 0.46%
QC value within limits for Ag		Recovery = 97.32%			
B_†	68360.5	2.45 mg/L	0.016	2.45 mg/L	0.016 0.65%
QC value within limits for B_		Recovery = 97.98%			
Ba†	195997.4	4.97 mg/L	0.002	4.97 mg/L	0.002 0.04%
QC value within limits for Ba		Recovery = 99.38%			
Bet	5537787.7	2.05 mg/L	0.004	2.05 mg/L	0.004 0.17%
QC value within limits for Be		Recovery = 102.59%			
Cat	1297840.3	49.6 mg/L	1.25	49.6 mg/L	1.25 2.52%
QC value within limits for Ca		Recovery = 99.27%			
Cdt	44042.2	2.01 mg/L	0.007	2.01 mg/L	0.007 0.34%
QC value within limits for Cd		Recovery = 100.66%			
Crt	229640.5	5.09 mg/L	0.041	5.09 mg/L	0.041 0.80%
QC value within limits for Cr		Recovery = 101.81%			
Cut	1542182.5	4.99 mg/L	0.030	4.99 mg/L	0.030 0.61%
QC value within limits for Cu		Recovery = 99.84%			
Fet	54241.6	5.06 mg/L	0.012	5.06 mg/L	0.012 0.23%
QC value within limits for Fe		Recovery = 101.13%			
K†	49727.7	47.4 mg/L	0.57	47.4 mg/L	0.57 1.20%
QC value within limits for K		Recovery = 94.78%			
Mgt	971470.8	49.4 mg/L	1.33	49.4 mg/L	1.33 2.70%
QC value within limits for Mg		Recovery = 98.81%			
Mnt	2532469.0	5.09 mg/L	0.051	5.09 mg/L	0.051 1.01%
QC value within limits for Mn		Recovery = 101.74%			
Mot	51345.3	5.01 mg/L	0.028	5.01 mg/L	0.028 0.56%
QC value within limits for Mo		Recovery = 100.22%			
Naf	255991.7	45.7 mg/L	0.12	45.7 mg/L	0.12 0.26%
QC value within limits for Na		Recovery = 91.45%			
Nit	98298.0	5.13 mg/L	0.008	5.13 mg/L	0.008 0.16%
QC value within limits for Ni		Recovery = 102.58%			
V†	771993.7	5.00 mg/L	0.010	5.00 mg/L	0.010 0.19%
QC value within limits for V		Recovery = 100.03%			
Znt	202895.6	5.15 mg/L	0.028	5.15 mg/L	0.028 0.55%
QC value within limits for Zn		Recovery = 102.96%			

All analyte(s) passed QC.

Sequence No.: 46
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/4/2008 23:20:49
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	408655.8	96.8 %	0.71			0.73%
Yr	423561.5	99.5 %	0.42			0.42%
Agt	26.5	0.00010 mg/L	0.000066	0.00010 mg/L	0.000066	64.51%
B ₋	QC value within limits for Ag	Recovery = Not calculated				
B ₋	306.6	0.0111 mg/L	0.00123	0.0111 mg/L	0.00123	11.06%
B ₋	QC value within limits for B ₋	Recovery = Not calculated				
Bat	1.3	0.00003 mg/L	0.000002	0.00003 mg/L	0.000002	7.26%
Bet	QC value within limits for Ba	Recovery = Not calculated				
Bet	-9.0	0.00000 mg/L	0.000026	0.00000 mg/L	0.000026	763.92%
Cat	105.2	0.00402 mg/L	0.000334	0.00402 mg/L	0.000334	8.30%
Cdt	QC value within limits for Ca	Recovery = Not calculated				
Cdt	3.1	0.00014 mg/L	0.000007	0.00014 mg/L	0.000007	4.67%
Crt	QC value within limits for Cd	Recovery = Not calculated				
Crt	-4.0	-0.00009 mg/L	0.000136	-0.00009 mg/L	0.000136	151.81%
Cut	QC value within limits for Cr	Recovery = Not calculated				
Cut	127.1	0.00041 mg/L	0.000229	0.00041 mg/L	0.000229	55.74%
Fet	QC value within limits for Cu	Recovery = Not calculated				
Fet	5.8	0.00054 mg/L	0.000183	0.00054 mg/L	0.000183	33.79%
Kt	QC value within limits for Fe	Recovery = Not calculated				
Kt	51.3	0.0489 mg/L	0.00617	0.0489 mg/L	0.00617	12.62%
Mgt	QC value within limits for K	Recovery = Not calculated				
Mgt	40.0	0.00204 mg/L	0.000291	0.00204 mg/L	0.000291	14.32%
Mnt	QC value within limits for Mg	Recovery = Not calculated				
Mnt	86.1	0.00017 mg/L	0.000030	0.00017 mg/L	0.000030	17.09%
Mot	QC value within limits for Mn	Recovery = Not calculated				
Mot	11.7	0.00114 mg/L	0.000115	0.00114 mg/L	0.000115	10.08%
Nat	QC value within limits for Mo	Recovery = Not calculated				
Nat	139.7	0.0249 mg/L	0.00605	0.0249 mg/L	0.00605	24.26%
Nit	QC value within limits for Na	Recovery = Not calculated				
Nit	-6.7	-0.00035 mg/L	0.000288	-0.00035 mg/L	0.000288	82.75%
Vt	QC value within limits for Ni	Recovery = Not calculated				
Vt	8.7	0.00006 mg/L	0.000037	0.00006 mg/L	0.000037	66.32%
Znt	QC value within limits for V	Recovery = Not calculated				
Znt	-28.9	-0.00074 mg/L	0.000070	-0.00074 mg/L	0.000070	9.55%
	QC value within limits for Zn	Recovery = Not calculated				

All analyte(s) passed QC.

Sequence No.: 47
 Sample ID: MRL
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 21
 Date Collected: 6/4/2008 23:24:12
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL

Analyte	Back Pressure	Flow
All	212.0 kPa	0.65 L/min

Mean Data: MRL

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	408928.1	96.9 %	0.28			0.29%
Yr	421628.1	99.0 %	0.44			0.44%
Agt	2543.0	0.00987 mg/L	0.000201	0.00987 mg/L	0.000201	2.03%
B_f	1416.9	0.0514 mg/L	0.00041	0.0514 mg/L	0.00041	0.80%
Bat	805.5	0.0204 mg/L	0.00016	0.0204 mg/L	0.00016	0.77%
Bet	2814.8	0.00104 mg/L	0.000005	0.00104 mg/L	0.000005	0.45%
Cat	26338.4	1.01 mg/L	0.005	1.01 mg/L	0.005	0.48%
Cdt	133.7	0.00611 mg/L	0.000202	0.00611 mg/L	0.000202	3.30%
Cri	465.0	0.0103 mg/L	0.00022	0.0103 mg/L	0.00022	2.16%
Cut	3198.4	0.0104 mg/L	0.00029	0.0104 mg/L	0.00029	2.80%
Fet	219.2	0.0204 mg/L	0.00038	0.0204 mg/L	0.00038	1.84%
Kt	1007.0	0.960 mg/L	0.0085	0.960 mg/L	0.0085	0.88%
Mgt	2052.0	0.104 mg/L	0.0006	0.104 mg/L	0.0006	0.56%
Mnt	1160.8	0.00233 mg/L	0.000007	0.00233 mg/L	0.000007	0.30%
Mot	205.9	0.0201 mg/L	0.00015	0.0201 mg/L	0.00015	0.74%
Nat	5130.1	0.916 mg/L	0.0173	0.916 mg/L	0.0173	1.89%
Nit	408.7	0.0213 mg/L	0.00013	0.0213 mg/L	0.00013	0.59%
Vt	310.3	0.00206 mg/L	0.000032	0.00206 mg/L	0.000032	1.53%
Znt	765.1	0.0194 mg/L	0.00001	0.0194 mg/L	0.00001	0.07%

Sequence No.: 48
 Sample ID: MRL2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 24
 Date Collected: 6/4/2008 23:27:49
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MRL2007

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: MRL2007

Analyte	Mean Corrected	Calib.	Sample	RSD	
	Intensity	Conc. Units	Std.Dev.		
Sca	418876.7	99.3 %	0.87		0.87%
Yr	428357.9	101 %	1.6		1.61%
Agt	2497.1	0.00969 mg/L	0.000193	0.00969 mg/L	0.000193 1.99%
B _t	1306.3	0.0474 mg/L	0.00054	0.0474 mg/L	0.00054 1.13%
B _a	785.0	0.0199 mg/L	0.00043	0.0199 mg/L	0.00043 2.16%
B _e	2659.9	0.00099 mg/L	0.000013	0.00099 mg/L	0.000013 1.33%
Cat	27123.2	1.04 mg/L	0.019	1.04 mg/L	0.019 1.79%
Cdt	133.3	0.00609 mg/L	0.000093	0.00609 mg/L	0.000093 1.52%
Cr _t	462.6	0.0103 mg/L	0.00024	0.0103 mg/L	0.00024 2.38%
Cut	3249.4	0.0105 mg/L	0.00008	0.0105 mg/L	0.00008 0.72%
Fet	243.5	0.0227 mg/L	0.00020	0.0227 mg/L	0.00020 0.88%
K _t	993.0	0.946 mg/L	0.0256	0.946 mg/L	0.0256 2.71%
Mgt	2067.4	0.105 mg/L	0.0022	0.105 mg/L	0.0022 2.06%
Mnt	1020.2	0.00205 mg/L	0.000018	0.00205 mg/L	0.000018 0.89%
Mot	196.8	0.0192 mg/L	0.00021	0.0192 mg/L	0.00021 1.10%
Nat	5295.3	0.946 mg/L	0.0141	0.946 mg/L	0.0141 1.49%
Nit	403.8	0.0211 mg/L	0.00032	0.0211 mg/L	0.00032 1.53%
V _t	289.9	0.00193 mg/L	0.000080	0.00193 mg/L	0.000080 4.15%
Znt	910.5	0.0231 mg/L	0.00031	0.0231 mg/L	0.00031 1.33%

Sequence No.: 49
 Sample ID: LCS2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 66
 Date Collected: 6/4/2008 23:31:28
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCS2007

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: LCS2007

Analyte	Mean Corrected	Calib.	Sample	RSD	
	Intensity	Conc. Units	Std.Dev.		
Sca	394313.9	93.4 %	0.59		0.63%
Yr	422304.0	99.2 %	0.31		0.32%
Agt	115084.0	0.448 mg/L	0.0003	0.448 mg/L	0.0003 0.08%
B _t	12535.7	0.449 mg/L	0.0052	0.449 mg/L	0.0052 1.15%
Bat	36472.3	0.925 mg/L	0.0008	0.925 mg/L	0.0008 0.09%
Bet	126680.1	0.0471 mg/L	0.00017	0.0471 mg/L	0.00017 0.36%
Cat	1196798.0	45.8 mg/L	0.19	45.8 mg/L	0.19 0.42%
Cdt	4226.3	0.193 mg/L	0.0017	0.193 mg/L	0.0017 0.86%
Crt	42839.5	0.950 mg/L	0.0084	0.950 mg/L	0.0084 0.88%
Cut	287240.2	0.930 mg/L	0.0015	0.930 mg/L	0.0015 0.16%
Fet	49548.9	4.62 mg/L	0.001	4.62 mg/L	0.001 0.02%
Kt	18537.0	17.7 mg/L	0.19	17.7 mg/L	0.19 1.05%
Mgt	366773.6	18.7 mg/L	0.05	18.7 mg/L	0.05 0.25%
Mnt	234844.9	0.471 mg/L	0.0002	0.471 mg/L	0.0002 0.03%
Mot	9451.2	0.922 mg/L	0.0067	0.922 mg/L	0.0067 0.73%
Nat	239182.8	42.7 mg/L	0.26	42.7 mg/L	0.26 0.61%
Nit	9177.2	0.479 mg/L	0.0040	0.479 mg/L	0.0040 0.84%
Vt	141763.8	0.919 mg/L	0.0009	0.919 mg/L	0.0009 0.10%
Znt	38114.4	0.970 mg/L	0.0057	0.970 mg/L	0.0057 0.59%

Sequence No.: 50
 Sample ID: LCSD2007
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 67
 Date Collected: 6/4/2008 23:33:54
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: LCSD2007

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: LCSD2007

Analyte	Mean Corrected	Calib.	Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	403786.3	95.7 %	0.29			0.30%
Yr	423604.8	99.5 %	0.39			0.39%
Agt	112138.2	0.436 mg/L	0.0009	0.436 mg/L	0.0009	0.20%
B_t	12180.4	0.436 mg/L	0.0017	0.436 mg/L	0.0017	0.39%
Bat	35046.0	0.889 mg/L	0.0077	0.889 mg/L	0.0077	0.86%
Bef	123493.0	0.0459 mg/L	0.00007	0.0459 mg/L	0.00007	0.15%
Cat	1171588.6	44.8 mg/L	0.12	44.8 mg/L	0.12	0.27%
Cdt	4061.6	0.186 mg/L	0.0001	0.186 mg/L	0.0001	0.03%
Crt	40987.2	0.909 mg/L	0.0133	0.909 mg/L	0.0133	1.46%
Cut	279551.5	0.905 mg/L	0.0014	0.905 mg/L	0.0014	0.15%
Fet	48563.5	4.53 mg/L	0.014	4.53 mg/L	0.014	0.30%
Kt	18116.9	17.3 mg/L	0.14	17.3 mg/L	0.14	0.79%
Mgt	359466.5	18.3 mg/L	0.06	18.3 mg/L	0.06	0.31%
Mnt	228577.6	0.459 mg/L	0.0007	0.459 mg/L	0.0007	0.16%
Mot	9107.6	0.889 mg/L	0.0018	0.889 mg/L	0.0018	0.20%
Nat	234392.4	41.9 mg/L	0.23	41.9 mg/L	0.23	0.54%
Nit	8839.0	0.461 mg/L	0.0005	0.461 mg/L	0.0005	0.12%
Vt	138213.8	0.896 mg/L	0.0017	0.896 mg/L	0.0017	0.19%
Znt	36775.9	0.936 mg/L	0.0037	0.936 mg/L	0.0037	0.39%

Sequence No.: 51
 Sample ID: 2805290427
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 68
 Date Collected: 6/4/2008 23:36:20
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805290427

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: 2805290427

Analyte	Mean Corrected		Calib.	Sample		RSD
	Intensity	Conc. Units		Std.Dev.	Conc. Units	
Sca	410982.8	97.4 %	0.46			0.47%
Yr	426215.8	100 %	0.0			0.02%
Ag†	28.5	0.00014 mg/L	0.000110	0.00014 mg/L	0.000110	81.08%
B_t	401.6	0.0146 mg/L	0.00031	0.0146 mg/L	0.00031	2.10%
Bat	135.0	0.00342 mg/L	0.000016	0.00342 mg/L	0.000016	0.45%
Bet	-40.5	-0.00001 mg/L	0.000031	-0.00001 mg/L	0.000031	211.04%
Cat	247362.9	9.46 mg/L	0.060	9.46 mg/L	0.060	0.64%
Cdt	-1.8	-0.00008 mg/L	0.000107	-0.00008 mg/L	0.000107	133.05%
Crt	115.2	0.00255 mg/L	0.000243	0.00255 mg/L	0.000243	9.51%
Cut	803.7	0.00261 mg/L	0.000162	0.00261 mg/L	0.000162	6.23%
Fet	728.7	0.0679 mg/L	0.00067	0.0679 mg/L	0.00067	0.98%
K†	1310.1	1.25 mg/L	0.007	1.25 mg/L	0.007	0.56%
Mg†	143814.5	7.31 mg/L	0.028	7.31 mg/L	0.028	0.39%
Mnt	4070.6	0.00801 mg/L	0.000041	0.00801 mg/L	0.000041	0.51%
Mot	16.1	0.00157 mg/L	0.000456	0.00157 mg/L	0.000456	28.93%
Nat	57833.5	10.3 mg/L	0.00	10.3 mg/L	0.00	0.00%
Nit	2.0	0.00011 mg/L	0.000241	0.00011 mg/L	0.000241	228.56%
V†	2025.8	0.0131 mg/L	0.00003	0.0131 mg/L	0.00003	0.22%
Znt	163.2	0.00417 mg/L	0.000060	0.00417 mg/L	0.000060	1.43%

Sequence No.: 52
 Sample ID: 2805290427MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 69
 Date Collected: 6/4/2008 23:39:58
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805290427MS
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: 2805290427MS

Analyte	Mean Corrected	Calib.	Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	397619.9	94.2 %	0.14			0.15%
Yr	418051.7	98.2 %	0.93			0.94%
Agt	120708.8	0.470 mg/L	0.0016	0.470 mg/L	0.0016	0.35%
B _t	13512.9	0.484 mg/L	0.0028	0.484 mg/L	0.0028	0.58%
B _{at}	38087.1	0.966 mg/L	0.0042	0.966 mg/L	0.0042	0.44%
Bet	133289.9	0.0495 mg/L	0.00009	0.0495 mg/L	0.00009	0.18%
Cat	1519038.1	58.1 mg/L	0.26	58.1 mg/L	0.26	0.44%
Cdt	4319.1	0.197 mg/L	0.0008	0.197 mg/L	0.0008	0.41%
Crt	44300.3	0.982 mg/L	0.0042	0.982 mg/L	0.0042	0.42%
Cut	298912.1	0.968 mg/L	0.0067	0.968 mg/L	0.0067	0.69%
Fet	52832.1	4.92 mg/L	0.005	4.92 mg/L	0.005	0.11%
Kt	20905.2	19.9 mg/L	0.00	19.9 mg/L	0.00	0.02%
Mgt	522929.3	26.6 mg/L	0.01	26.6 mg/L	0.01	0.03%
Mnt	249289.1	0.500 mg/L	0.0001	0.500 mg/L	0.0001	0.02%
Mot	9702.6	0.947 mg/L	0.0010	0.947 mg/L	0.0010	0.11%
Nat	311708.8	55.7 mg/L	0.09	55.7 mg/L	0.09	0.15%
Nit	9262.4	0.483 mg/L	0.0023	0.483 mg/L	0.0023	0.47%
Vt	149893.5	0.971 mg/L	0.0008	0.971 mg/L	0.0008	0.09%
Znt	39382.3	1.00 mg/L	0.000	1.00 mg/L	0.000	0.04%

Sequence No.: 53
 Sample ID: 2805290427MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 70
 Date Collected: 6/4/2008 23:43:04
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805290427MSD

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: 2805290427MSD

Analyte	Mean Corrected	Calib.	Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	394681.5	93.5 %	0.30			0.32%
Yr	415705.6	97.6 %	0.92			0.95%
Agt	121615.2	0.473 mg/L	0.0013	0.473 mg/L	0.0013	0.27%
B_f	13765.5	0.493 mg/L	0.0013	0.493 mg/L	0.0013	0.27%
Bat	38584.0	0.978 mg/L	0.0013	0.978 mg/L	0.0013	0.13%
Bet	134191.4	0.0498 mg/L	0.00022	0.0498 mg/L	0.00022	0.44%
Cat	1506394.1	57.6 mg/L	0.43	57.6 mg/L	0.43	0.74%
Cdt	4397.3	0.201 mg/L	0.0002	0.201 mg/L	0.0002	0.11%
Crt	44541.4	0.987 mg/L	0.0077	0.987 mg/L	0.0077	0.78%
Cut	302031.5	0.978 mg/L	0.0003	0.978 mg/L	0.0003	0.03%
Fet	53954.7	5.03 mg/L	0.006	5.03 mg/L	0.006	0.12%
Kt	20862.8	19.9 mg/L	0.12	19.9 mg/L	0.12	0.60%
Mgt	525917.7	26.7 mg/L	0.02	26.7 mg/L	0.02	0.06%
Mnt	250853.6	0.503 mg/L	0.0011	0.503 mg/L	0.0011	0.21%
Mot	9862.8	0.963 mg/L	0.0004	0.963 mg/L	0.0004	0.04%
Nat	311775.0	55.7 mg/L	0.13	55.7 mg/L	0.13	0.23%
Nit	9387.0	0.490 mg/L	0.0033	0.490 mg/L	0.0033	0.67%
Vt	151157.8	0.979 mg/L	0.0025	0.979 mg/L	0.0025	0.25%
Znt	39789.2	1.01 mg/L	0.006	1.01 mg/L	0.006	0.60%

Sequence No.: 54
 Sample ID: 2805310013
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 71
 Date Collected: 6/4/2008 23:46:10
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805310013

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: 2805310013

Analyte	Mean Corrected	Calib.	Sample	RSD	
	Intensity	Conc. Units	Std.Dev.		
Sca	401263.5	95.1 %	0.19		0.21%
Yr	420218.7	98.7 %	0.13		0.14%
Ag†	-10.8	0.00007 mg/L	0.000199	0.00007 mg/L	0.000199 295.22%
B†	2770.3	0.101 mg/L	0.0001	0.101 mg/L	0.0001 0.05%
Bat	2153.8	0.0546 mg/L	0.00006	0.0546 mg/L	0.00006 0.11%
Bet	-202.4	-0.00007 mg/L	0.000002	-0.00007 mg/L	0.000002 2.97%
Cat	1401953.8	53.6 mg/L	0.04	53.6 mg/L	0.04 0.08%
Cdt	-11.5	-0.00053 mg/L	0.000049	-0.00053 mg/L	0.000049 9.39%
Crt	15.3	0.00034 mg/L	0.000026	0.00034 mg/L	0.000026 7.64%
Cut	1586.0	0.00513 mg/L	0.000235	0.00513 mg/L	0.000235 4.58%
Fet	3189.0	0.297 mg/L	0.0042	0.297 mg/L	0.0042 1.40%
K†	4111.8	3.92 mg/L	0.142	3.92 mg/L	0.142 3.61%
Mgt	239612.0	12.2 mg/L	0.05	12.2 mg/L	0.05 0.40%
Mnt	39576.5	0.0792 mg/L	0.00004	0.0792 mg/L	0.00004 0.06%
Mot	66.8	0.00652 mg/L	0.000509	0.00652 mg/L	0.000509 7.80%
Nat	247989.7	44.3 mg/L	0.20	44.3 mg/L	0.20 0.46%
Nit	13.7	0.00071 mg/L	0.000176	0.00071 mg/L	0.000176 24.74%
Vt	248.0	0.00161 mg/L	0.000120	0.00161 mg/L	0.000120 7.45%
Znt	675.8	0.0172 mg/L	0.00005	0.0172 mg/L	0.00005 0.27%

Sequence No.: 55
 Sample ID: 2805310013MS
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 72
 Date Collected: 6/4/2008 23:49:52
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805310013MS

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: 2805310013MS

Analyte	Mean Corrected	Calib.	Sample	RSD		
	Intensity	Conc. Units	Std.Dev.			
Sca	386327.5	91.5 %	0.24		0.26%	
Yr	409432.3	96.1 %	0.11		0.12%	
Agt	122742.6	0.478 mg/L	0.0007	0.478 mg/L	0.0007	0.15%
B _t	16200.8	0.582 mg/L	0.0006	0.582 mg/L	0.0006	0.11%
Baf	40843.4	1.04 mg/L	0.003	1.04 mg/L	0.003	0.32%
Bef	135269.6	0.0502 mg/L	0.00032	0.0502 mg/L	0.00032	0.65%
Cat	2640055.5	101 mg/L	0.1	101 mg/L	0.1	0.08%
Cdt	4450.6	0.203 mg/L	0.0003	0.203 mg/L	0.0003	0.14%
Cr _t	44633.6	0.989 mg/L	0.0054	0.989 mg/L	0.0054	0.54%
Cut	309610.0	1.00 mg/L	0.001	1.00 mg/L	0.001	0.14%
Fet	57059.4	5.32 mg/L	0.022	5.32 mg/L	0.022	0.42%
Kt	24253.3	23.1 mg/L	0.15	23.1 mg/L	0.15	0.65%
Mgt	623038.6	31.7 mg/L	0.05	31.7 mg/L	0.05	0.15%
Mnt	288457.7	0.579 mg/L	0.0020	0.579 mg/L	0.0020	0.35%
Mot	9969.4	0.973 mg/L	0.0051	0.973 mg/L	0.0051	0.52%
Nat	495875.1	88.6 mg/L	0.33	88.6 mg/L	0.33	0.37%
Nit	9554.3	0.499 mg/L	0.0016	0.499 mg/L	0.0016	0.32%
Vt	151351.8	0.981 mg/L	0.0036	0.981 mg/L	0.0036	0.37%
Znt	40816.8	1.04 mg/L	0.003	1.04 mg/L	0.003	0.33%

Sequence No.: 56
 Sample ID: 2805310013MSD
 Analyst:
 Initial Sample Wt:
 Dilution: 1X

Autosampler Location: 73
 Date Collected: 6/4/2008 23:52:59
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805310013MSD

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: 2805310013MSD

Analyte	Mean Corrected	Calib.	Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	390553.2	92.5 %	0.44			0.48%
Yr	412525.2	96.9 %	1.08			1.12%
Ag†	117750.7	0.458 mg/L	0.0003	0.458 mg/L	0.0003	0.06%
B†	15529.5	0.558 mg/L	0.0032	0.558 mg/L	0.0032	0.57%
Bat	39320.2	0.997 mg/L	0.0062	0.997 mg/L	0.0062	0.62%
Bet	129793.4	0.0482 mg/L	0.00014	0.0482 mg/L	0.00014	0.30%
Cat	2563359.2	98.0 mg/L	0.66	98.0 mg/L	0.66	0.67%
Cdt	4271.1	0.195 mg/L	0.0022	0.195 mg/L	0.0022	1.10%
Crt	42964.3	0.952 mg/L	0.0080	0.952 mg/L	0.0080	0.84%
Cut	296602.5	0.960 mg/L	0.0001	0.960 mg/L	0.0001	0.01%
Fet	54678.5	5.10 mg/L	0.019	5.10 mg/L	0.019	0.37%
K†	23501.5	22.4 mg/L	0.03	22.4 mg/L	0.03	0.12%
Mgt	602695.7	30.7 mg/L	0.04	30.7 mg/L	0.04	0.12%
Mnt	277413.3	0.557 mg/L	0.0012	0.557 mg/L	0.0012	0.22%
Mot	9603.7	0.937 mg/L	0.0061	0.937 mg/L	0.0061	0.65%
Nat	488690.5	87.3 mg/L	0.19	87.3 mg/L	0.19	0.22%
Nit	9201.2	0.480 mg/L	0.0031	0.480 mg/L	0.0031	0.64%
V†	144961.1	0.939 mg/L	0.0035	0.939 mg/L	0.0035	0.37%
Znt	39209.0	0.998 mg/L	0.0058	0.998 mg/L	0.0058	0.58%

Sequence No.: 57
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/4/2008 23:56:06
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Sca	384131.2	91.0 %	0.29			0.32%
Yr	407943.5	95.8 %	0.54			0.57%
Agt	250354.2	0.973 mg/L	0.0028	0.973 mg/L	0.0028	0.29%
B_t	68431.7	Recovery = 97.25% 2.45 mg/L	0.016	2.45 mg/L	0.016	0.64%
Bat	197339.5	Recovery = 98.09% 5.00 mg/L	0.002	5.00 mg/L	0.002	0.03%
Bet	5468325.0	Recovery = 100.06% 2.03 mg/L	0.010	2.03 mg/L	0.010	0.48%
Cat	1294371.1	Recovery = 101.30% 49.5 mg/L	0.01	49.5 mg/L	0.01	0.01%
Cdt	43992.3	Recovery = 99.00% 2.01 mg/L	0.002	2.01 mg/L	0.002	0.12%
Crt	228073.1	Recovery = 100.54% 5.06 mg/L	0.013	5.06 mg/L	0.013	0.25%
Cut	1540252.5	Recovery = 101.11% 4.99 mg/L	0.033	4.99 mg/L	0.033	0.66%
Fet	54099.2	Recovery = 99.71% 5.04 mg/L	0.042	5.04 mg/L	0.042	0.84%
Kt	50355.6	Recovery = 100.86% 48.0 mg/L	0.04	48.0 mg/L	0.04	0.08%
Mgt	983230.4	Recovery = 95.98% 50.0 mg/L	0.07	50.0 mg/L	0.07	0.13%
Mnt	2516686.5	Recovery = 100.01% 5.06 mg/L	0.001	5.06 mg/L	0.001	0.02%
Mot	51212.9	Recovery = 101.10% 5.00 mg/L	0.011	5.00 mg/L	0.011	0.23%
Nat	259227.8	Recovery = 99.96% 46.3 mg/L	0.01	46.3 mg/L	0.01	0.01%
Nit	98095.9	Recovery = 92.60% 5.12 mg/L	0.008	5.12 mg/L	0.008	0.15%
Vt	771001.7	Recovery = 102.37% 5.00 mg/L	0.012	5.00 mg/L	0.012	0.25%
Znt	202269.0	Recovery = 99.90% 5.13 mg/L	0.010	5.13 mg/L	0.010	0.19%
	QC value within limits for Zn	Recovery = 102.64%				

All analyte(s) passed QC.

Sequence No.: 58
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/4/2008 23:59:20
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std.Dev.	RSD
	Conc. Units	Units	Conc. Units	Std.Dev.	
Sca	407337.1	96.5 %	0.31		0.32%
Yr	420401.5	98.7 %	0.66		0.67%
Agt	8.5	0.00003 mg/L	0.000450	0.00003 mg/L	0.000450 >999.9%
QC value within limits for Ag		Recovery = Not calculated			
B_t	356.1	0.0129 mg/L	0.00119	0.0129 mg/L	0.00119 9.19%
QC value within limits for B		Recovery = Not calculated			
Bat	0.4	0.00001 mg/L	0.000003	0.00001 mg/L	0.000003 26.78%
QC value within limits for Ba		Recovery = Not calculated			
Bet	3.5	0.00000 mg/L	0.000017	0.00000 mg/L	0.000017 >999.9%
QC value within limits for Be		Recovery = Not calculated			
Cat	109.1	0.00417 mg/L	0.000617	0.00417 mg/L	0.000617 14.80%
QC value within limits for Ca		Recovery = Not calculated			
Cdt	-0.3	-0.00001 mg/L	0.000020	-0.00001 mg/L	0.000020 141.29%
QC value within limits for Cd		Recovery = Not calculated			
Crt	1.9	0.00004 mg/L	0.000091	0.00004 mg/L	0.000091 217.85%
QC value within limits for Cr		Recovery = Not calculated			
Cut	76.9	0.00025 mg/L	0.000121	0.00025 mg/L	0.000121 48.68%
QC value within limits for Cu		Recovery = Not calculated			
Fet	6.4	0.00060 mg/L	0.000442	0.00060 mg/L	0.000442 74.25%
QC value within limits for Fe		Recovery = Not calculated			
Kt	39.0	0.0372 mg/L	0.00215	0.0372 mg/L	0.00215 5.79%
QC value within limits for K		Recovery = Not calculated			
Mgt	24.9	0.00127 mg/L	0.000243	0.00127 mg/L	0.000243 19.13%
QC value within limits for Mg		Recovery = Not calculated			
Mnt	60.0	0.00012 mg/L	0.000006	0.00012 mg/L	0.000006 5.28%
QC value within limits for Mn		Recovery = Not calculated			
Mot	11.9	0.00116 mg/L	0.000003	0.00116 mg/L	0.000003 0.25%
QC value within limits for Mo		Recovery = Not calculated			
Nat	114.3	0.0204 mg/L	0.00081	0.0204 mg/L	0.00081 3.95%
QC value within limits for Na		Recovery = Not calculated			
Nif	-4.3	-0.00023 mg/L	0.000244	-0.00023 mg/L	0.000244 108.18%
QC value within limits for Ni		Recovery = Not calculated			
Vt	0.0	0.00000 mg/L	0.000006	0.00000 mg/L	0.000006 797.92%
QC value within limits for V		Recovery = Not calculated			
Znt	-27.4	-0.00070 mg/L	0.000038	-0.00070 mg/L	0.000038 5.46%
QC value within limits for Zn		Recovery = Not calculated			

All analyte(s) passed QC.

Sequence No.: 59
 Sample ID: MCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 5
 Date Collected: 6/5/2008 00:02:43
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: MCV

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: MCV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc. Units	Std.Dev.	RSD
Sca	393904.5	93.3 %	0.31		0.33%
Yr	415932.1	97.7 %	1.38		1.42%
Agt	124443.3	0.483 mg/L	0.0044	0.483 mg/L	0.0044 0.92%
B_t	33776.0	Recovery = 96.68% 1.21 mg/L	0.013	1.21 mg/L	0.013 1.04%
Bat	99036.6	Recovery = 96.79% 2.51 mg/L	0.036	2.51 mg/L	0.036 1.44%
Bet	2755579.8	QC value within limits for Ba Recovery = 100.44% 1.02 mg/L	0.000	1.02 mg/L	0.000 0.03%
Cat	655292.8	QC value within limits for Be Recovery = 102.09% 25.1 mg/L	0.09	25.1 mg/L	0.09 0.35%
Cdt	21904.2	QC value within limits for Ca Recovery = 100.24% 1.00 mg/L	0.009	1.00 mg/L	0.009 0.91%
Crt	113856.4	QC value within limits for Cd Recovery = 100.12% 2.52 mg/L	0.030	2.52 mg/L	0.030 1.17%
Cut	777739.8	QC value within limits for Cr Recovery = 100.95% 2.52 mg/L	0.005	2.52 mg/L	0.005 0.21%
Fet	27085.3	QC value within limits for Cu Recovery = 100.70% 2.52 mg/L	0.009	2.52 mg/L	0.009 0.36%
Kt	24569.2	QC value within limits for Fe Recovery = 100.99% 23.4 mg/L	0.02	23.4 mg/L	0.02 0.09%
Mgt	492520.8	QC value within limits for K Recovery = 93.66% 25.0 mg/L	0.08	25.0 mg/L	0.08 0.34%
Mnt	1283775.4	QC value within limits for Mg Recovery = 100.19% 2.58 mg/L	0.001	2.58 mg/L	0.001 0.05%
Mot	25497.9	QC value within limits for Mn Recovery = 103.15% 2.49 mg/L	0.032	2.49 mg/L	0.032 1.27%
Nat	127109.3	QC value within limits for Mo Recovery = 99.54% 22.7 mg/L	0.04	22.7 mg/L	0.04 0.17%
Nit	49743.7	QC value within limits for Na Recovery = 90.81% 2.60 mg/L	0.022	2.60 mg/L	0.022 0.86%
Vt	383820.7	QC value within limits for Ni Recovery = 103.82% 2.49 mg/L	0.022	2.49 mg/L	0.022 0.88%
Znt	101568.1	QC value within limits for V Recovery = 99.47% 2.58 mg/L	0.024	2.58 mg/L	0.024 0.94%
		QC value within limits for Zn Recovery = 103.08%			

All analyte(s) passed QC.

Sequence No.: 66
 Sample ID: 2805280375_2X
 Analyst:
 Initial Sample Wt:
 Dilution: 2X

Autosampler Location: 80
 Date Collected: 6/5/2008 00:26:43
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: 2805280375_2X
 Analyte Back Pressure Flow
 All 211.0 kPa 0.65 L/min

Mean Data: 2805280375_2X

Analyte	Mean Corrected	Calib.	Sample			RSD
	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	349022.0	82.7 %	0.08			0.09%
Yr	389056.7	91.4 %	0.53			0.58%
Agt	-148.2	-0.00049 mg/L	0.000167	-0.00099 mg/L	0.000334	33.82%
B _t	47737.1	1.73 mg/L	0.008	3.47 mg/L	0.016	0.46%
B _{at}	493.1	0.0125 mg/L	0.00002	0.0250 mg/L	0.00004	0.14%
Bet	-1250.4	-0.00046 mg/L	0.000011	-0.00092 mg/L	0.000023	2.46%
Cat	5212889.5	199 mg/L	0.1	399 mg/L	0.2	0.06%
Cdt	-12.6	-0.00058 mg/L	0.000136	-0.00115 mg/L	0.000272	23.60%
Crt	1595.8	0.0354 mg/L	0.00043	0.0707 mg/L	0.00087	1.23%
Cut	822.6	0.00267 mg/L	0.000378	0.00533 mg/L	0.000756	14.18%
Fet	2350.3	0.219 mg/L	0.0036	0.438 mg/L	0.0073	1.66%
Kt	15231.1	14.5 mg/L	0.08	29.0 mg/L	0.17	0.57%
Mgt	1848029.4	94.0 mg/L	0.08	188 mg/L	0.2	0.09%
Mnt	233423.1	0.467 mg/L	0.0015	0.934 mg/L	0.0030	0.33%
Mot	84.3	0.00823 mg/L	0.000135	0.0165 mg/L	0.00027	1.65%
Nat	3506006.2	626 mg/L	0.4	1250 mg/L	0.9	0.07%
Nit	76.7	0.00400 mg/L	0.000169	0.00800 mg/L	0.000339	4.24%
Vt	1635.3	0.0107 mg/L	0.00007	0.0214 mg/L	0.00014	0.67%
Zn _t	-51.2	-0.00136 mg/L	0.000196	-0.00273 mg/L	0.000392	14.36%

Sequence No.: 70
 Sample ID: CCV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/5/2008 00:40:55
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std.Dev.	RSD
	Conc. Units	Units	Conc. Units	Std.Dev.	
Sca	381962.0	90.5 %	0.39		0.43%
Yr	400068.5	93.9 %	0.09		0.09%
Agt	249682.1	0.970 mg/L	0.0011,	0.970 mg/L	0.0011 0.11%
B_f	68181.8	2.44 mg/L	0.001	2.44 mg/L	0.001 0.03%
Bat	196927.5	4.99 mg/L	0.010	4.99 mg/L	0.010 0.21%
Bet	5412042.4	2.01 mg/L	0.009	2.01 mg/L	0.009 0.42%
Cat	1322451.3	50.6 mg/L	0.19	50.6 mg/L	0.19 0.37%
Cdt	43821.8	2.00 mg/L	0.002	2.00 mg/L	0.002 0.11%
Crt	225944.3	5.01 mg/L	0.010	5.01 mg/L	0.010 0.20%
Cut	1536544.3	4.97 mg/L	0.005	4.97 mg/L	0.005 0.10%
Fef	55498.2	5.17 mg/L	0.029	5.17 mg/L	0.029 0.57%
Kt	50775.8	48.4 mg/L	0.08	48.4 mg/L	0.08 0.17%
Mgt	994763.7	50.6 mg/L	0.22	50.6 mg/L	0.22 0.44%
Mnt	2498617.7	5.02 mg/L	0.003	5.02 mg/L	0.003 0.05%
Mot	50811.7	4.96 mg/L	0.023	4.96 mg/L	0.023 0.46%
Nat	253826.8	45.3 mg/L	0.09	45.3 mg/L	0.09 0.21%
Nit	97296.5	5.08 mg/L	0.011	5.08 mg/L	0.011 0.22%
Vt	765881.2	4.96 mg/L	0.010	4.96 mg/L	0.010 0.21%
Znt	201092.0	5.10 mg/L	0.013	5.10 mg/L	0.013 0.25%
QC value within limits for Zn Recovery = 102.05%					

All analyte(s) passed QC.

Sequence No.: 71
 Sample ID: CCB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/5/2008 00:44:08
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: CCB

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	410180.1	97.2 %	1.25		1.29%
Yr	419929.4	98.6 %	0.41		0.42%
Agt	-60.4	-0.00023 mg/L	0.000195	-0.00023 mg/L	0.000195 83.48%
B_f	373.5	0.0136 mg/L	0.00089	0.0136 mg/L	0.00089 6.56%
Bat	1.8	0.00005 mg/L	0.000056	0.00005 mg/L	0.000056 120.90%
Bet	50.2	0.00002 mg/L	0.000003	0.00002 mg/L	0.000003 16.53%
Cat	140.3	0.00537 mg/L	0.000839	0.00537 mg/L	0.000839 15.64%
Cdt	2.7	0.00012 mg/L	0.000209	0.00012 mg/L	0.000209 168.98%
Crt	2.4	0.00005 mg/L	0.000085	0.00005 mg/L	0.000085 160.04%
Cut	-14.1	-0.00005 mg/L	0.000172	-0.00005 mg/L	0.000172 375.92%
Fet	3.1	0.00029 mg/L	0.000613	0.00029 mg/L	0.000613 209.13%
Kt	54.8	0.0522 mg/L	0.00600	0.0522 mg/L	0.00600 11.48%
Mgf	37.0	0.00188 mg/L	0.000043	0.00188 mg/L	0.000043 2.26%
Mnt	159.7	0.00032 mg/L	0.000004	0.00032 mg/L	0.000004 1.34%
Mot	14.4	0.00140 mg/L	0.000155	0.00140 mg/L	0.000155 11.01%
Nat	321.4	0.0574 mg/L	0.00058	0.0574 mg/L	0.00058 1.01%
Nit	-5.6	-0.00029 mg/L	0.000313	-0.00029 mg/L	0.000313 106.58%
Vf	-2.5	-0.00002 mg/L	0.000061	-0.00002 mg/L	0.000061 394.37%
Znt	-32.2	-0.00082 mg/L	0.000069	-0.00082 mg/L	0.000069 8.37%
		QC value within limits for Zn	Recovery = Not calculated		
			All analyte(s) passed QC.		

Sequence No.: 80
 Sample ID: ECV
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 4
 Date Collected: 6/5/2008 01:18:22
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ECV

Analyte	Back Pressure	Flow
All	210.0 kPa	0.65 L/min

Mean Data: ECV

Analyte	Mean Corrected Intensity	Calib.	Sample Conc.	Std.Dev.	RSD
	Conc. Units	Units	Conc. Units	Std.Dev.	
Sca	382954.6	90.7 %	0.76		0.83%
Yr	398536.9	93.6 %	0.84		0.90%
Agt	252646.4	0.981 mg/L	0.0029	0.981 mg/L	0.0029 0.29%
	QC value within limits for Ag	Recovery = 98.15%			
B _{-t}	69109.7	2.48 mg/L	0.005	2.48 mg/L	0.005 0.22%
	QC value within limits for B _{-t}	Recovery = 99.05%			
Bat	199978.7	5.07 mg/L	0.006	5.07 mg/L	0.006 0.11%
	QC value within limits for Ba	Recovery = 101.40%			
Bet	5476395.3	2.03 mg/L	0.029	2.03 mg/L	0.029 1.41%
	QC value within limits for Be	Recovery = 101.45%			
Cat	1350171.4	51.6 mg/L	0.02	51.6 mg/L	0.02 0.03%
	QC value within limits for Ca	Recovery = 103.27%			
Cdt	44327.0	2.03 mg/L	0.014	2.03 mg/L	0.014 0.70%
	QC value within limits for Cd	Recovery = 101.31%			
Crt	227213.4	5.04 mg/L	0.047	5.04 mg/L	0.047 0.93%
	QC value within limits for Cr	Recovery = 100.73%			
Cut	1561962.9	5.06 mg/L	0.072	5.06 mg/L	0.072 1.42%
	QC value within limits for Cu	Recovery = 101.12%			
Fet	56674.9	5.28 mg/L	0.020	5.28 mg/L	0.020 0.39%
	QC value within limits for Fe	Recovery = 105.66%			
Kt	51333.8	48.9 mg/L	0.43	48.9 mg/L	0.43 0.88%
	QC value within limits for K	Recovery = 97.85%			
Mgt	1021547.8	52.0 mg/L	0.06	52.0 mg/L	0.06 0.11%
	QC value within limits for Mg	Recovery = 103.90%			
Mnt	2516496.0	5.05 mg/L	0.069	5.05 mg/L	0.069 1.36%
	QC value within limits for Mn	Recovery = 101.09%			
Mot	51238.6	5.00 mg/L	0.035	5.00 mg/L	0.035 0.70%
	QC value within limits for Mo	Recovery = 100.01%			
Nat	252883.2	45.2 mg/L	0.22	45.2 mg/L	0.22 0.50%
	QC value within limits for Na	Recovery = 90.34%			
Nit	98415.2	5.14 mg/L	0.025	5.14 mg/L	0.025 0.48%
	QC value within limits for Ni	Recovery = 102.70%			
Vt	774712.4	5.02 mg/L	0.022	5.02 mg/L	0.022 0.44%
	QC value within limits for V	Recovery = 100.38%			
Znt	203057.7	5.15 mg/L	0.029	5.15 mg/L	0.029 0.57%
	QC value within limits for Zn	Recovery = 103.04%			

All analyte(s) passed QC.

Sequence No.: 81
 Sample ID: ECB
 Analyst:
 Initial Sample Wt:
 Dilution:

Autosampler Location: 0
 Date Collected: 6/5/2008 01:21:36
 Data Type: Original
 Initial Sample Vol:
 Sample Prep Vol:

Nebulizer Parameters: ECB

Analyte	Back Pressure	Flow
All	211.0 kPa	0.65 L/min

Mean Data: ECB

Analyte	Mean Corrected Intensity	Calib.	Sample	Std.Dev.	RSD
	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	
Sca	417048.9	98.8 %	0.37		0.38%
Yr	423496.4	99.4 %	1.34		1.35%
Agt	-61.7	-0.00024 mg/L	0.000120	-0.00024 mg/L	0.000120 50.01%
B _t	321.6	0.0117 mg/L	0.00059	0.0117 mg/L	0.00059 5.05%
B ₋	-	QC value within limits for B ₋	Recovery = Not calculated		
Bat	-0.2	0.00000 mg/L	0.000022	0.00000 mg/L	0.000022 583.52%
Bet	36.3	0.00001 mg/L	0.000022	0.00001 mg/L	0.000022 167.21%
Be	-	QC value within limits for Be	Recovery = Not calculated		
Cat	127.4	0.00487 mg/L	0.001245	0.00487 mg/L	0.001245 25.55%
Ca	-	QC value within limits for Ca	Recovery = Not calculated		
Cdt	1.9	0.00009 mg/L	0.000072	0.00009 mg/L	0.000072 83.85%
Crt	-0.7	-0.00001 mg/L	0.000056	-0.00001 mg/L	0.000056 376.40%
Cut	12.6	0.00004 mg/L	0.000025	0.00004 mg/L	0.000025 60.80%
Cu	-	QC value within limits for Cu	Recovery = Not calculated		
Fet	9.7	0.00091 mg/L	0.000283	0.00091 mg/L	0.000283 31.20%
Fe	-	QC value within limits for Fe	Recovery = Not calculated		
Kt	26.0	0.0248 mg/L	0.00730	0.0248 mg/L	0.00730 29.45%
K	-	QC value within limits for K	Recovery = Not calculated		
Mgt	33.0	0.00168 mg/L	0.000336	0.00168 mg/L	0.000336 20.02%
Mg	-	QC value within limits for Mg	Recovery = Not calculated		
Mnt	288.4	0.00058 mg/L	0.000010	0.00058 mg/L	0.000010 1.76%
Mn	-	QC value within limits for Mn	Recovery = Not calculated		
Mot	15.1	0.00147 mg/L	0.000513	0.00147 mg/L	0.000513 34.79%
Mo	-	QC value within limits for Mo	Recovery = Not calculated		
Nat	198.5	0.0355 mg/L	0.00377	0.0355 mg/L	0.00377 10.64%
Na	-	QC value within limits for Na	Recovery = Not calculated		
Nit	-5.6	-0.00029 mg/L	0.000182	-0.00029 mg/L	0.000182 61.72%
Ni	-	QC value within limits for Ni	Recovery = Not calculated		
Vt	-6.7	-0.00004 mg/L	0.000009	-0.00004 mg/L	0.000009 20.80%
V	-	QC value within limits for V	Recovery = Not calculated		
Znt	-31.7	-0.00081 mg/L	0.000074	-0.00081 mg/L	0.000074 9.20%
Zn	-	QC value within limits for Zn	Recovery = Not calculated		

All analyte(s) passed QC.

Analytical Sequence

Method: 200.7&6010_080304

Seq.	Loc.	ID	Status
1	15	ICV	QC Passed
2	9	LINEARITY	QC Passed
3	10	ICSA	QC Passed
4	11	ICSAB	QC Failed
5	0	Wash	
6	12	QC-25 1ppm	QC Passed
7	4	CCV	QC Passed
8	0	ICB	QC Passed
9	20	MRL	QC Passed
10	16	MRL/2	Analyzed
11	18	FILTERCHECK	Analyzed
12	38	MBLANK	Analyzed
13	39	LCS	Analyzed
14	40	LCSD	Analyzed
15	41	2806030669	Analyzed
16	42	2806030669MS	Analyzed
17	43	2806030669MSD	Analyzed
18	44	2806030741	Analyzed
19	45	2806030674	Analyzed
20	4	CCV	QC Passed
21	0	CCB	QC Passed
22	46	2806030542	Analyzed
23	47	2806030544	Analyzed
24	48	2806030545	Analyzed
25	49	2806030547	Analyzed
26	50	2806030844	Analyzed
27	51	2806030859	Analyzed
28	52	2806030867	Analyzed
29	53	2806030671	Analyzed
30	54	2806030671MS	Analyzed
31	55	2806030671MSD	Analyzed
32	4	CCV	QC Passed
33	0	CCB	QC Passed
34	5	MCV	QC Passed
35	56	2806030660	Analyzed
36	57	2806030870	Analyzed
37	58	2806030871	Analyzed
38	59	2806030872	Analyzed
39	60	2806030570	Analyzed
40	61	2806030713	Analyzed
41	62	2806030780	Analyzed
42	63	2806030672	Analyzed
43	64	D805220919_20X	Analyzed
44	65	MBLANK2007	Analyzed
45	4	CCV	QC Passed
46	0	CCB	QC Passed
47	21	MRL	Analyzed
48	24	MRL2007	Analyzed
49	66	LCS2007	Analyzed
50	67	LCSD2007	Analyzed
51	68	2805290427	Analyzed
52	69	2805290427MS	Analyzed
53	70	2805290427MSD	Analyzed
54	71	2805310013	Analyzed
55	72	2805310013MS	Analyzed
56	73	2805310013MSD	Analyzed
57	4	CCV	QC Passed
58	0	CCB	QC Passed
59	5	MCV	QC Passed
60	74	2805280136	Analyzed
61	75	2805280137	Analyzed
62	76	2805280138	Analyzed
63	77	2805280139	Analyzed
64	78	2805280274	Analyzed
65	79	2805280347	Analyzed
66	80	2805280375_2X	Analyzed
67	81	2805280487	Analyzed

58	82	2805280472	Analyzed
69	83	2805280473	Analyzed
70	4	CCV	QC Passed
71	0	CCB	QC Passed
72	84	2805280474	Analyzed
73	85	2805280475	Analyzed
74	86	2805280476	Analyzed
75	87	2805280478	Analyzed
76	88	2805280479	Analyzed
77	89	2805280480	Analyzed
78	90	2805280481	Analyzed
79	91	2805190003	Analyzed
80	4	ECV	QC Passed
81	0	ECB	QC Passed

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Lot #
HNO₃ R# 100450 HCL R# 100446
IS = Yttrium(ME0709008) 0.75mL + Scandium (ME0710007) 0.5mL to 1000mL w/ 2% HNO₃

Standards	Lot #	Exp. Date	Dilution	
Calibration (Prepare daily)	ME0712001 ME0712001	(12/01/08) (12/01/08)	1:10 ME0801001 1:10	
CCV/MCV/ECV (Prepare daily)	ME0710008	(04/17/09)	CCV/ECV 1:20 ME0801002	MCV 1:40 ME0801003
Spike/LCS (Prepare daily)	ME0709009 ME0801004 ME0803001	(03/11/09) (07/11/08) (08/13/08)	1:100 ME0801005 1:100 1:50	
MRL (Prepare daily)	ME0801007	(07/11/08)	1:100 ME0801008	
ICSA	ME0712003	(06/01/08)		
ICSAB	ME0712004	(06/01/08)		
QCS	ME0610005	(04/10/08)		
1ppm Check	ME0801010	(07/11/08)		
Linearity	ME0805001	(11/05/08)		

Method Sr/Ti/Sn/SiO₂

Calibration	ME0801012	(07/11/08)	
CCV/ECV	ME0803011	(09/30/08)	
QCS	ME0801012	(07/11/08)	
Spike/LCS (Prepare daily)	ME0803012	(09/30/08)	1:100
MRL (Prepare daily)	ME0801014	(07/11/08)	1:100

Method Li

Std/ICV/MRL (Prepare daily)	ME0801009	(07/11/08)	1:1000, 200, 40, 10
QCS (Prepare daily)	ME0801011	(07/11/08)	1:10
LCS/Spike (Prepare daily)	ME0801011	(07/11/08)	1:50
ccv (Prepare daily)	ME0801011	(07/11/08)	1:40

From May 2005: the calibration std for ICP should be ME0505010,011 not ME0408010

dilution should be 1:20 and 1:40 not 1:200 and 1:400. 1/10/2006.

Initial: WSJ
Date: 7/16/08

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration STD ME #: 0801001
Date Received/Prepped: Prep Daily By: wbh
Date Expired: 12/1/2008 Lot #:
Manufacturer: MWH-wbh Certificate: NO
Matrix: 2% HNO₃ + 5% HCl NIST SRM:
Amount: Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	1:10 ME0712002	10 ug/ml
Sb		10 ug/ml
Sn		10 ug/ml
Ti		10 ug/ml
B		5 ug/ml
Ca	1:10 ME0712001	100 ug/ml
K		100 ug/ml
Mg		100 ug/ml
Na		100 ug/ml
Al		10 ug/ml
As		10 ug/ml
Ba		10 ug/ml
Co		10 ug/ml
Cr		10 ug/ml
Cu		10 ug/ml
Fe		10 ug/ml
Mn		10 ug/ml
Ni		10 ug/ml
Pb		10 ug/ml
Se		10 ug/ml
Tl		10 ug/ml
V		10 ug/ml
Zn		10 ug/ml
Cd		5 ug/ml
Be		4 ug/ml
SR		3 ug/ml
Ag		2 ug/ml

Initial:

STE

Date:

10/01/03

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration Stock Std #1 **ME #:** 0712001
Date Received/Prepped: 12/1/2007 **By:** STE
Date Expired: 12/1/2008 **Lot #:** A2-MEB243151
Manufacturer: Inorganic Ventures **Certificate:** Y
Matrix: 5% Nitric Acid **NIST SRM:** Varies
Amount: 500 mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ca	(P/N MWH-ICAP-CAL-1)	1000 ug/ml
K		1000 ug/ml
Mg		1000 ug/ml
Na		1000 ug/ml
Al		100 ug/ml
As		100 ug/ml
Ba		100 ug/ml
Co		100 ug/ml
Cr		100 ug/ml
Cu		100 ug/ml
Fe		100 ug/ml
Mn		100 ug/ml
Ni		100 ug/ml
Pb		100 ug/ml
Se		100 ug/ml
Tl		100 ug/ml
V		100 ug/ml
Zn		100 ug/ml
Cd		50 ug/ml
Be		40 ug/ml
SR		30 ug/ml
Ag		20 ug/ml



CERTIFICATE OF ANALYSIS

1.0 **INORGANIC VENTURES** is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 36-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 **DESCRIPTION OF CRM** Custom Solution

Catalog No.: MWH-ICAP-CAL-1

ME 0712031

Lot Number: A2-MEB243151

Matrix: 5% HNO₃(abs)

1,000.00 µg/mL ea:

Ca, K, Mg, Na,

100.00 µg/mL ea:

Al, As, Ba, Co, Cr₃, Cu, Fe, Mn, Ni, Pb, Se, Tl, V, Zn,

50.00 µg/mL ea:

Cd,

40.00 µg/mL ea:

Be,

30.00 µg/mL ea:

Sr,

20.00 µg/mL ea:

Ag

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Aluminum, Al	100.4 ± 0.3 µg/mL	Arsenic, As	100.1 ± 0.3 µg/mL	Barium, Ba	99.6 ± 0.4 µg/mL
Beryllium, Be	40.04 ± 0.08 µg/mL	Cadmium, Cd	50.15 ± 0.12 µg/mL	Calcium, Ca	1,000 ± 2 µg/mL
Chromium+3, Cr ₃	100.0 ± 0.2 µg/mL	Cobalt, Co	99.9 ± 0.2 µg/mL	Copper, Cu	100.0 ± 0.2 µg/mL
Iron, Fe	99.6 ± 0.1 µg/mL	Lead, Pb	100.0 ± 0.3 µg/mL	Magnesium, Mg	1,000 ± 4 µg/mL
Manganese, Mn	100.0 ± 0.3 µg/mL	Nickel, Ni	100.0 ± 0.3 µg/mL	Potassium, K	1,001 ± 5 µg/mL
Selenium, Se	100.0 ± 0.2 µg/mL	Silver, Ag	20.04 ± 0.02 µg/mL	Sodium, Na	1,002 ± 5 µg/mL
Strontium, Sr	30.04 ± 0.18 µg/mL	Thallium, Tl	99.7 ± 0.1 µg/mL	Vanadium, V	100.0 ± 0.3 µg/mL
Zinc, Zn	100.0 ± 0.3 µg/mL				

Certified Density: 1.055 g/mL (measured at 22° C)

Initial: STE
Date: 10/01/03

METALS STANDARD DOCUMENTATION

Standard: ICP Calibration Stock Std #2
Date Received/Prepped: 12/1/2007
Date Expired: 12/1/2008
Manufacturer: Inorganic Ventures
Matrix: 5% Nitric Acid + Trace HF
Amount: 500 mL

ME #: 0712002
By: STE
Lot #: A2-MEB243152
Certificate: Y
NIST SRM: Varies
Storage: Room Temp

Component	Comment	Conc. Unit:
Mo	(P/N MWH-ICAP-CAL-2)	100 ug/ml
Sb		100 ug/ml
Sn		100 ug/ml
Tl		100 ug/ml
B		50 ug/ml
Mo		100 ug/ml

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM

Catalog No.:	Custom Solution
Lot Number:	MWH-ICAP-CAL-2
Matrix:	tr. HF, 5% HNO ₃ (abs)

ME 0712 CG2

100.00 µg/mL ea:
 Mo, Sb, Sn, Ti,
 50.00 µg/mL ea:
 B

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE	ELEMENT	CERTIFIED VALUE
Antimony, Sb	100.4 ± 0.3 µg/mL	Boron, B	49.85 ± 0.20 µg/mL	Molybdenum, Mo	100.2 ± 0.3 µg/mL
Tin, Sn	100.2 ± 0.3 µg/mL	Titanium, Ti	100.4 ± 0.2 µg/mL		

Certified Density: 1.037 g/mL (measured at 22° C)

The Certified Value is based upon the most precise method used to analyze this CRM. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean

x_i = individual results

n = number of measurements

$\sum s_i$ = The summation of all significant estimated errors

(Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

$$\text{Uncertainty } (\pm) = \frac{2(\sum s_i)^{1/2}}{(n)}$$

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

- "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993)
- This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term "in-house std." is specified.

Initial: SE
Date: 10/19/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV/QCS Stock Standard **ME #:** 0710008
Date Received/Prepped: 10/19/2007 **By:** STE
Date Expired: 4/17/2009 **Lot #:** 07J154
Manufacturer: CPI **Certificate:** Y
Matrix: 5% HNO₃ = tr HF **NIST SRM:** Various
Amount: 100 mL x 10 **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ag	P/N 4400-061003RH01	20 ppm
Al		100 ppm
As		100 ppm
B		50 ppm
Ba		100 ppm
bE		40 ppm
Ca		1000 ppm
Cd		50 ppm
Co		100 ppm
Cr		100 ppm
Cu		100 ppm
Fe		100 ppm
K		1000 ppm
Mg		1000 ppm
Mn		100 ppm
Mo		100 ppm
Na		1000 ppm
Ni		100 ppm
Pb		100 ppm
Sb		100 ppm
Se		100 ppm
Tl		100 ppm
V		100 ppm
Zn		100 ppm
Sr		20 ppm
Sn		20 ppm
Ti		20 ppm



Innovative Solutions
in Analytical Science and
Technology

Expiry: 4/17/2009

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com

EUROPE

P.O. Box 2704
1000 CS Amsterdam
The Netherlands
+31 20 638 05 97
Fax +31 20 420 28 36
www.cpiinternational.com

Certificate of Analysis

Part Number: **4400-061003RH01**
Lot Number: **07J154**
Shelf Life: **18 months**

ME 0710008

exp. 4/17/09

STE

MWH
Custom Multi
5% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Ag	20	K	1000	Sr	20
Al	100	Mg	1000	Sn	20
As	100	Mn	100	Ti	20
B	50	Mo	100		
Ba	100	Na	1000		
Be	40	Ni	100		
Ca	1000	Pb	100		
Cd	50	Sb	100		
Co	100	Se	100		
Cr	100	TL	100		
Cu	100	V	100		
Fe	100	Zn	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000ug/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial: W3y
Date: 11/10/08

METALS STANDARD DOCUMENTATION

Standard: ICP MCV Working Standard ME #: 0801003
Date Received/Prepped: Daily By: Wbh
Date Expired: 4/17/2009 Lot #:
Manufacturer: CPI Certificate: Y
Matrix: 2%HNO₃ + 5% HCL NIST SRM: Various
Amount: 100 mL Storage: Room Temp

Component	Comment	Conc. Unit:
Ag	5mL ME0710008 / 100mL	0.5 ppm
Al		2.5 ppm
As		2.5 ppm
B		1.25 ppm
Ba		2.5 ppm
bE		1 ppm
Ca		25 ppm
Cd		1 ppm
Co		2.5 ppm
Cr		2.5 ppm
Cu		2.5 ppm
Fe		2.5 ppm
K		25 ppm
Mg		25 ppm
Mn		2.5 ppm
Mo		2.5 ppm
Na		25 ppm
Ni		2.5 ppm
Pb		2.5 ppm
Sb		2.5 ppm
Se		2.5 ppm
Tl		2.5 ppm
V		2.5 ppm
Zn		2.5 ppm
Sr		0.5 ppm
Sn		0.5 ppm
Ti		0.5 ppm

Initial:

Date:

WZJ7/16/03

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/ECV Working Standard **ME #:** 0801002
Date Received/Prepped: Daily **By:** Wbh
Date Expired: 4/17/2009 **Lot #:**
Manufacturer: CPI **Certificate:** Y
Matrix: 2%HNO₃ + 5% HCL **NIST SRM:** Various
Amount: 100 mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ag	5ml ME0710008 / 100mL	1 ppm
Al		5 ppm
As		5 ppm
B		2.5 ppm
Ba		5 ppm
bE		2 ppm
Ca		50 ppm
Cd		2 ppm
Co		5 ppm
Cr		5 ppm
Cu		5 ppm
Fe		5 ppm
K		50 ppm
Mg		50 ppm
Mn		5 ppm
Mo		5 ppm
Na		50 ppm
Ni		5 ppm
Pb		5 ppm
Sb		5 ppm
Se		5 ppm
Tl		5 ppm
V		5 ppm
Zn		5 ppm
Sr		1 ppm
Sn		1 ppm
Ti		1 ppm

Initial:

W31

Date:

7/11/08

METALS STANDARD DOCUMENTATION

Standard: Spike and LCS std for ICP **ME #:** 0801005
Date Received/Prepped: prep daily **By:** WBH
Date Expired: 7/11/2008 **Lot #:**
Manufacturer: MWH-wbh **Certificate:** N
Matrix: 2% HNO₃ + 5% HCl **NIST SRM:**
Amount: **Storage:** Room Temp

Component	Comment	Conc. Unit:
CA	1:200 ME0709007	50 ppm
K		20 ppm
MG		20 ppm
NA		50 ppm
Iron	1:100 ME0709009	5 mg/L
Aluminum		2 mg/L
Barium		1 mg/L
Cobalt		1 mg/L
Chromium		1 mg/L
Copper		1 mg/L
Molybdenum		1 mg/L
Strontium		1 mg/L
Titanium		1 mg/L
Vanadium		1 mg/L
Zinc		1 mg/L
Tin		1 mg/L
Silver		0.5 mg/L
Boron		0.5 mg/L
Manganese		0.5 mg/L
Nickel		0.5 mg/L
Antimony		0.5 mg/L
Arsenic		0.2 mg/L
Cadmium		0.2 mg/L
Lead		0.2 mg/L
Selenium		0.2 mg/L
Thallium		0.2 mg/L
Uraium		0.2 mg/L
Beryllium		0.05 mg/L
AS	1:100 ME0801004	0.8 mg/L
PB		0.8 mg/L
SE		0.8 mg/L
TL		0.8 mg/L

Initial:

W36

Date:

5/16/08

METALS STANDARD DOCUMENTATION

Standard: ICP Spike Solution ME #: 0801004
Date Received/Prepped: 1/11/2008 By: Wbh
Date Expired: 7/11/2008 Lot #: VARIOUS
Manufacturer: MWH-wbh Certificate:
Matrix: 2%HNO3 NIST SRM:
Amount: 100mL Storage: Room Temp

Component	Comment	Conc. Unit:
AS	8.0mL ME07090023/100mL	80 ppm
PB	8.0mL ME0704013/100mL	80 ppm
SE	8.0mL ME0703001/100mL	80 ppm
TL	8.0mL ME0702006/100mL	80 ppm

Initial:

Date:

WBH

2/25/08

METALS STANDARD DOCUMENTATION

Standard: ICP LCS/SPIKE Solution **ME #:** 0802001
Date Received/Prepped: 2/25/2008 **By:** wbh
Date Expired: 7/25/2008 **Lot #:**
Manufacturer: MWH-wbh **Certificate:** NO
Matrix: 2% HNO₃ + 5% HCl **NIST SRM:**
Amount: 100mL **Storage:** Room Temp

Component	Comment	Conc. Unit:
Ca	1:4 ME0702002	5000 ug/ml
K	1:10 ME0702005	1000 ug/ml
Mg	1:10 ME0702004	1000 ug/ml
Na	1:4 ME0702003	5000 ug/ml

Initial: STE
Date: 9/13/07

METALS STANDARD DOCUMENTATION

Standard: ICP CCV/MCV Stock Standard **ME #:** 0709009
Date Received/Prepped: 9/13/2007 **By:** STE
Date Expired: 3/11/2009 **Lot #:** 07I040
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ AND 0.1% HF **NIST SRM:**
Amount: 100 mL **Storage:** Room Temp.

Component	Comment	Conc. Unit:
Fe	P/N 4400-050314RH01	500
Al	(10 bottles)	200
Ba		100
Co		100
Cr		100
Cu		100
Mo		100
Sr		100
Ti		100
V		100
Zn		50
Ag		50
Ba		50
Mn		50
Ni		50
Sb		50
As		20
Cd		20
Pb		20
Se		20
Tl		20
Sn		100
Be		5
U		20



Innovative Solutions
in Analytical Science and
Technology

Expiry: 3/11/2009

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com
707.545.7901

EUROPE

P.O. Box 2704
1000 CG Amsterdam
The Netherlands
+31 20 638 05 97
Fax +31 20 420 28 36
www.cpiinternational.com

Certificate of Analysis

Part Number: 4400-050314RH01
Lot Number: 07I040
Shelf Life: 18 months

ME C7 C9 C0 C9

MWH Labs
5% HNO₃ + 0.1% HF
#REF!

Concentrations in ug/mL ± 0.5%

Fe	500	B	50
Al	200	Mn	50
Ba	100	Ni	50
Co	100	Sb	50
Cr	100	As	20
Cu	100	Cd	20
Mo	100	Pb	20
Sr	100	Se	20
Ti	100	TL	20
V	100	Sn	100
Zn	100	Be	5
Ag	50	U	20

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

w3y

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 1000ppm Stock Std **ME #:** 0702005
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B056
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO3 **NIST SRM:** 3141
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

Aug 16 '08



Innovative Solutions
In Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

MB0702005

P/N 4400-10M411**P/N S4400-10M411**

Single-Element Potassium Standard

K in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)

Source Purity: 99.999%

Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	ppb	DL		ppb	DL		ppb	DL		ppb	DL
Al	0.39	0.1	Cu	0.16	0.1	Pb	ND	0.1	K	X	70
Sb	0.34	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.14	0.1	Eu	ND	0.1	Mg	2.6	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	0.93	1	Rb	9.5	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
C ²	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	20
Ca	82	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	19	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WB

2/07/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 10000ppm Stock Std **ME #:** 0702004
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B058
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3131
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



W61608
*Innovative Solutions
 in Analytical Science and
 Technology*

USA

5580 Skylane Boulevard 707.525.5788
 Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
 1000 CS Amsterdam Fax +31 20 420 28 36
 The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS*47070704***P/N 4400-10M311****P/N S4400-10M311**

Single-Element Magnesium Standard

Mg in 4% HNO₃

10,000 ± 30 µg/mL

Lot # 07B058

Material Source: Magnesium Metal

Source Purity: 99.99%

Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL								
Al 28	0.1	Cu 1.6	0.1	Pb 7.7	0.7	K ND	70	Tl 0.91	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr 0.28	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 0.28	0.1	Eu ND	0.1	Mg X	0.2	Rh ND	0.1	Sn 0.14	0.1
Be ND	0.1	Gd 0.23	0.1	Mn 19.8	1	Rb ND	0.1	Ti ND	0.1
Bi ND	0.1	Ga 0.18	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B ND	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd 1.1	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Ni 1	0.1	Si 64	20	Yb ND	0.1
Ca ND	7	Ho ND	0.1	Nb ND	0.1	Ag 0.19	0.1	Y 0.2	0.1
Ce 21	0.1	I 1	0.2	Os ND	0.1	Na 7.2	1	Zn ND	1
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 0.19	0.1	Zr 0.29	0.1
Cr ND	1	Fe 80	30	P ND	10	Ta ND	0.1		
Co ND	0.1	La 0.76	0.1	Pt ND	0.1	Te ND	0.1		

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

W31

2/24/07

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std **ME #:** 0702003
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B057
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO3 **NIST SRM:** 3152a
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 08



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

170702003

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO₃)

Source Purity: 99.99%

Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	ppb	DL									
Al	1.5	0.1	Cu	0.45	0.1	Pb	ND	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.13	0.1	Eu	ND	0.1	Mg	2.3	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	8
Ca	120	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	X	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial:
Date:

w37
2/10/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 1000ppm Stock Std **ME #:** 0702002
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B065
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3109a
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707 525 5788
Santa Rosa, CA 95403 800 878.7654
www.cpiinternational.com Fax 707.545.7901

AUG 16 08

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

180702602

Single-Element Calcium Standard

Ca in 4% HNO₃

10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)

Source Purity: 99.997%

Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	ppb	DL									
Al	7	0.1	Cu	1.7	0.1	Pb	0.23	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	1.5	0.1	Eu	ND	0.1	Mg	38	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	1.5	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	3	0.1	Si	47	8
Ca	X	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	0.27	0.2	Os	ND	0.1	Na	11.6	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	55	0.1
Cr	ND	1	Fe	INT	30	P	ND	10	Ta	ND	0.1
Co	INT	0.1	La	0.41	0.1	Pt	ND	0.1	Te	ND	0.1

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBY

2/16/07

METALS STANDARD DOCUMENTATION

Standard: Thallium 1000ppm Stock Std **ME #:** 0702006
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 06H213
Manufacturer: CPI **Certificate:** Y
Matrix: 2% HNO3 **NIST SRM:** 3158
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
TI	P/N S4400-1000581	1000 ppm



Nov 16 08
 Innovative Solutions
 in Analytical Science and
 Technology

USA

5580 Skylane Boulevard
 Santa Rosa, CA 95403
www.cpiinternational.com
 707.545.7901

EUROPE

P.O. Box 2704
 1000 CS Amsterdam
 The Netherlands
 +31 20 638 05 97
 Fax +31 20 420 28 36
www.cpiinternational.com

CERTIFICATE OF ANALYSIS

MF0702006

**P/N 4400-1000581
 P/N S4400-1000581**

Single Element Thallium Standard

Tl in 2% HNO₃

1000 ± 3 µg/mL

Lot # 06H213

Material Source: Thallium metal

Source Purity: 99.999%

Specific Gravity: 1.015 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3158. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	ppb	DL									
Al	13.3	0.1	Cu	9.3	0.1	Pb	41	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.37	0.1	Eu	ND	0.1	Mg	1.7	0.2	Rh	ND	0.1
Be	0.67	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	0.12	0.1	Ga	ND	0.1	Hg	0.16	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.21	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	1.6	0.1	Hf	ND	0.1	Ni	1.1	0.1	Si	46	8
Ca	51	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.3	0.1
Ce	ND	0.1	I	0.4	0.2	Os	ND	0.1	Na	3.3	1
Cs	0.24	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1
Cr	ND	1	Fe	ND	30	P	20	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654.

METALS STANDARD DOCUMENTATION

122

Standard:	Selenium Stock Standard	ME #: 0703001	Date Received/Prepped:	3/5/2007	By: Wdh	Lot #: 8/22/2008	Manufacturer:	CPI	Matrix:	2% HNO3	Amount:	100 mL	Conc. Unit:	1000 ppm	Comments:	P/N # S4400-1000491	See
-----------	-------------------------	---------------	------------------------	----------	---------	------------------	---------------	-----	---------	---------	---------	--------	-------------	----------	-----------	---------------------	-----

Initial: W39y Date: 3/17/07



Innovative Solutions
In Analytical Science and
Technology

AUG 22 08

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-1000491

P/N S4400-1000491

Single-Element Selenium Standard

Se in 2% HNO₃

1000 ± 3 µg/mL

MZ0703001

Lot # 06E228

Material Source: Selenium Metal

Source Purity: 99.99%

Specific Gravity: 1.011 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3148. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	ppb	DL									
Al	1.8	0.1	Cu	0.4	0.1	Pb	0.3	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	ND	0.1	Eu	ND	0.1	Mg	1.1	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.6	0.1	Sm	ND	0.1
Br	INT	10	Au	ND	0.1	Nd	ND	0.1	Se	X	6
Cd	0.4	0.1	Hf	ND	0.1	Ni	0.6	0.1	Si	40	8
Ca	5	7	Ho	ND	0.1	Nb	INT	0.1	Ag	0.8	0.1
Ce	ND	0.1	I	0.5	0.2	Os	ND	0.1	Na	3.8	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1
Cr	ND	1	Fe	ND	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

123

PEAK PERFORMANCE
EXTRACTION MANIFOLD

Accu-prep 7000™
Extraction Manifold

MOD BLOCK

Hy-Phase™
PE Block

Initial:

Date:

WBH

4/16/07

METALS STANDARD DOCUMENTATION

Standard: Lead Stock Standard **ME #:** 0704013
Date Received/Prepped: 4/16/2007 **By:** WBH
Date Expired: 10/11/2008 **Lot #:** 07A097
Manufacturer: CPI **Certificate:** Y
Matrix: 2% HNO3 **NIST SRM:** 3128
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Pb	P/N S4400-1000281	1000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com

EUROPE

P.O. Box 2704
1000 CS Amsterdam
The Netherlands

+31 20 638 05 97

Fax +31 20 420 28 36

www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N S4400-1000281

P/N 4400-1000281

Single-Element Lead Standard

Pb in 2% HNO₃

1000 ± 3 µg/mL

120704013

Lot # 07A097

Material Source: Lead Metal

Source Purity: 99.995 %

Specific Gravity: 1.009 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3128. Trace impurities of the 1000 µg/mL standard were analyzed by ICP-MS.

	ppb	DL									
Al	10.3	0.1	Cu	58	0.1	Pb	X	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	2	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.22	0.1	Eu	ND	0.1	Mg	1.4	0.2	Rh	IN	0.1
Be	0.58	0.1	Gd	ND	0.1	Mn	3.8	1	Rb	ND	0.1
Bi	0.7	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	0.17	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.9	0.1	Si	31	8
Ca	25	7	Ho	ND	0.1	Nb	ND	0.1	Ag	6.1	0.1
Ce	ND	0.1	I	0.1	0.2	Os	ND	0.1	Na	3.5	1
Cs	0.26	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1
Cr	ND	1	Fe	ND	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element DL=Detection Limit ND=None Detected

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial:

Date:

STE9/24/07

METALS STANDARD DOCUMENTATION

Standard: As Stock Standard **ME #:** 0709023
Date Received/Prepped: 9/24/2007 **By:** STE
Date Expired: 10/1/2008 **Lot #:** A2-AS02035
Manufacturer: Inorganic Ventures **Certificate:** Y
Matrix: 1.4% HNO₃ **NIST SRM:**
Amount: 100 mL X2 **Storage:** Room Temp

Component	Comment	Conc. Unit:
As	PN: CGAS1-1	1000 ug/ml



195 Lenigh Avenue, Suite 4
Lakewood, New Jersey 08701 USA
Inorganicventures.com

CERTIFICATE OF ANALYSIS

1000 µg/mL Arsenic in 1.4% (abs) HNO₃
Lot No. A2-AS02035
Date 07/09/023

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35:1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM 1000 µg/mL Arsenic in 1.4% (abs) HNO₃

Catalog Number: CGAS1-1, CGAS1-2, and CGAS1-5
Lot Number: **A2-AS02035**
Starting Material: As Polycrystalline lump
Starting Material Purity (%): 99.998288
Starting Material Lot No: 23444
Matrix: 1.4% (abs) HNO₃

ME 0709023

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Concentration: 1000 ± 6 µg/mL

Certified Density: 1.010 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean

x_i = individual results

n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

(\bar{x}) = The summation of all significant estimated errors

(Most common are the errors from instrumental measurement, weighing, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

- "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties." (ISO VIM, 2nd ed., 1993, definition 6.10)
- This product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

4.1 Assay Method #1 1000 ± 6 µg/mL
ICP Assay NIST SRM 3103a Lot Number: 010713

Assay Method #2 1001 ± 5 µg/mL
Gravimetric NIST SRM Lot Number: See Sec. 4.2

Date: Wby
7/11/03

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working Stock Solution **ME #:** 0801008
Date Received/Prepped: Daily **By:** Wbh
Date Expired: 7/11/2008 **Lot #:**
Manufacturer: MWH-wbh **Certificate:** Y
5% HNO₃ **NIST SRM:**
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	0.1mL ME0801007 / 10mL	0.05 ppm
Sb		0.05 ppm
As		0.1 ppm
Ba		0.02 ppm
Be		0.001 ppm
Ba		0.05 ppm
Cd		0.005 ppm
Ca		1 ppm
Cr		0.01 ppm
Co		0.05 ppm
Cu		0.01 ppm
Fe		0.02 ppm
Pb		0.02 ppm
Li		0.1 ppm
Mg		0.002 ppm
Mn		0.02 ppm
Mo		0.02 ppm
Ni		1 ppm
K		0.1 ppm
Se		0.01 ppm
Ag		1 ppm
Na		0.02 ppm
Tl		0.002 ppm
V		0.1 ppm
Zn		0.001 ppm
Ti		0.02 ppm
Sr		0.01 ppm
Sn		0.2 ppm

Date:

WZB
11/11/08

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Working Stock Solu: **ME #:** 0801007
Date Received/Prepped: 1/11/2008 **By:** Wbh
Date Expired: 7/11/2008 **Lot #:**
Manufacturer: MWH-wbh **Certificate:** Y
Matrix: 5% HNO3 **NIST SRM:**
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	10mL ME0709020 / 100mL	5 ppm
Sb		5 ppm
As		10 ppm
Ba		2 ppm
Be		0.1 ppm
Ba		5 ppm
Cd		0.5 ppm
Ca		100 ppm
Cr		1 ppm
Co		5 ppm
Cu		1 ppm
Fe		2 ppm
Pb		2 ppm
Li		10 ppm
Mg		0.2 ppm
Mn		2 ppm
Mo		2 ppm
Ni		100 ppm
K		10 ppm
Se		1 ppm
Ag		100 ppm
Na		2 ppm
Tl		0.2 ppm
V		10 ppm
Zn		0.1 ppm
Ti		2 ppm
Sr		1 ppm
Sn		20 ppm

Initial: SFB
Date: 9/26/07

METALS STANDARD DOCUMENTATION

Standard: ICP MRL Stock Standard **ME #:** 0709020
Date Received/Prepped: 9/20/2007 **By:**
Date Expired: 9/18/2008 **Lot #:** 06I162
Manufacturer: CPI **Certificate:** Y
Matrix: 2% HNO₃ + tr HF **NIST SRM:**
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-060915RHO1	50 ppm
Sb		50 ppm
As		100 ppm
Ba		20 ppm
Be		1 ppm
Ba		50 ppm
Cd		5 ppm
Ca		1000 ppm
Cr		10 ppm
Co		50 ppm
Cu		10 ppm
Fe		20 ppm
Pb		20 ppm
Li		1 ppm
Mg		100 ppm
Mn		2 ppm
Mo		20 ppm
Ni		20 ppm
K		1000 ppm
Se		100 ppm
Ag		10 ppm
Na		1000 ppm
Tl		100 ppm
V		2 ppm
Zn		20 ppm
Ti		20 ppm
Sr		10 ppm
Sn		200 ppm



Innovative Solutions
in Analytical Science and
Technology

Expiry 9/18/2008

ME0709020

Certificate of Analysis

Part Number: 4400-060915RH01

Lot Number: 06I162

Shelf Life: 12 months

MWH
Custom Standard
2% HNO₃ + tr HF

Concentrations in ug/mL ± 0.5%

Al	50	Pb	20	Zn	20
St	50	Li	1	Ti	20
As	100	Mg	100	Sr	10
Ba	20	Mn	2	Sn	200
Be	1	Mo	20		
B	50	Ni	20		
Cd	5	K	1000		
Ca	1000	Se	100		
Cr	10	Ag	10		
Co	50	Na	1000		
Cu	10	TL	100		
Fe	20	V	2		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000µg/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA, +31 20 638 05 97 in Europe or visit our web-site at www.cpiinternational.com.

Initial:

Date:

20
12/01/09

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSAB) **ME #:** 0712004
Date Received/Prepped: 12/1/2007 **By:** STE
Date Expired: 6/1/2008 **Lot #:**
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500 (25 mL)	250 ppm
Ca	P/N 4400-INTB1-100 (2.5 mL)	250 ppm
Fe		100 ppm
Mg		250 ppm
Ag		0.5 ppm
Ba		0.25 ppm
Be		0.25 ppm
Cd		0.5 ppm
Co		0.25 ppm
Cr		0.25 ppm
Cu		0.25 ppm
Mn		0.25 ppm
Ni		0.5 ppm
Pb		0.5 ppm
V		0.25 ppm
Zn		0.5 ppm

Initial: STE
Date: 12/1/07

METALS STANDARD DOCUMENTATION

Standard: Interference Check Standard (ICSA) **ME #:** 0712003
Date Received/Prepped: 12/1/2007 **By:** STE
Date Expired: 6/1/2008 **Lot #:**
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL Room temp. storage

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	250 ppm
Ca	25 mL in 500 mL	250 ppm
Fe		100 ppm
Mg		250 ppm

Initial:
Date:ST
8/29/07**METALS STANDARD DOCUMENTATION**

Standard: CLP Analytes B Solution **ME #:** 0708010
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 2/15/2009 **Lot #:** 07c256
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 100 mL **Storage:** Room Temp.

Component	Comment	Conc. Unit:
Ag	P/N 4400-INTB1-100	100 ug/L
Ba		50 ug/L
Be		50 ug/L
Cd		100 ug/L
Co		50 ug/L
Cr		50 ug/L
Cu		50 ug/L
Mn		50 ug/L
Ni		100 ug/L
Pb		100 ug/L
V		50 ug/L
Zn		100 ug/L



USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

FEB 15 09
*Innovative Solutions
in Analytical Science and
Technology*

ME 070 SAA

CERTIFICATE OF ANALYSIS

P/N 4400-INTB1-100
CLP Analytes B Solution
in 5% HNO₃

Lot # 07c256

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Ag	100	Ba	50	Be	50	Cd	100
Co	50	Cr	50	Cu	50	Mn	50
Ni	100	Pb	100	V	50	Zn	100

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

STE

8/27/07

METALS STANDARD DOCUMENTATION

Standard: CLP Interferents A Solution **ME #:** 0708009
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 2/15/2009 **Lot #:** 07E175
Manufacturer: CPI International **Certificate:**
Matrix: 5% HNO₃ **NIST SRM:**
Amount: 500 mL **Storage:** Room Temp.

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500	5000 ug/mL
Ca		5000 ug/mL
Fe		2000 ug/mL
Mg		5000 ug/mL

FEB 15 09



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-INTA1-500

CLP Interferents A Solution
in 5% HNO₃

ME 0708009

Lot # 07E175

Material Source: Metals and Salts
Source Purity: 99.99+%

Elements and Concentrations: µg/mL

Al 5000 Ca 5000 Fe 2000 Mg 5000

This standard solution was prepared using high-purity reference materials, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against an independent source traceable to the National Institute of Standards and Technology's SRM 3100 series.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: Wbh
 Date: 1/4/08

METALS STANDARD DOCUMENTATION

Standard:	ICP 1 PPM CHECK	ME #: 0801010
Date Received/Prepped	1/11/2008	By: Wbh
Date Expired:	7/11/2008	Lot #: VARIOUS
Manufacturer:	MWH-wbh	Certificate:
Matrix:	5% HNO3	NIST SRM:
Amount:	500 mL	Storage: Room Tem

Component	Comment	Conc. Unit:
Ag	5mL ME0708011 /500mL	1 mg/L
Al		1 mg/L
B		1 mg/L
Ba		1 mg/L
K		10 mg/L
Na		1 mg/L
Si		1 mg/L
As	5mL ME0708012 /500mL	1 mg/L
Be		1 mg/L
Ca		1 mg/L
Cd		1 mg/L
Co		1 mg/L
Cr		1 mg/L
Cu		1 mg/L
Fe		1 mg/L
Li		1 mg/L
Mg		1 mg/L
Mn		1 mg/L
Mo		1 mg/L
Ni		1 mg/L
Pb		1 mg/L
Sb		1 mg/L
Se		1 mg/L
Sr		1 mg/L
Ti		1 mg/L
Tl		1 mg/L
V		1 mg/L
Zn		1 mg/L

Initial:
Date:STE8/22/07**METALS STANDARD DOCUMENTATION**

Standard: QC Check Standard 21 **ME #:** 0708012
Date Received/Prepped: 8/27/2007 **By:** STE
Date Expired: 8/31/2008 **Lot #:** 074438H
Manufacturer: Crescent Chemical Co. Inc. **Certificate:**
Matrix: 5% HNO₃/tr. F/tr Tartaric Acid **NIST SRM:**
Amount: 100 mL Room temp. storage

Component	Comment	Conc. Unit:
As	Catalog No: QC-021.1	100 ug/mL
Be		100 ug/mL
Ca		100 ug/mL
Cd		100 ug/mL
Co		100 ug/mL
Cu		100 ug/mL
Fe		100 ug/mL
Li		100 ug/mL
Mg		100 ug/mL
Mn		100 ug/mL
Mo		100 ug/mL
Ni		100 ug/mL
Pb		100 ug/mL
Sb		100 ug/mL
Se		100 ug/mL
Sr		100 ug/mL
Ti		100 ug/mL
Tl		100 ug/mL
V		100 ug/mL
Zn		100 ug/mL

ME 0708012

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 21

CATALOG NO: QC-021.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F⁻/tr. Tartaric Acid

LOT NO.: 074438H

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3102a, 3103a, 3105a, 3108, 3109a, 3112a, 3113, 3114, 3126a, 3128, 3129a, 3131a, 3132, 3134, 3136, 3149, 3153a, 3158, 3162a, 3165, and 3168a.

Concentrations are given in µg/mL unless noted otherwise.

As 100	Be 100	Ca 100	Cd 100	Co 100
Cr 100	Cu 100	Fe 100	Li 100	Mg 100
Mn 100	Mo 100	Ni 100	Pb 100	Sb 100
Se 100	Sr 100	Ti 100	Tl 100	V 100
Zn 100				

Crescent Chemical Co. Inc.

Julie M. MacIntosh
QA Manager

EXPIRES: August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

*Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913*

ME0705011

Initial:

Date:

STE

8/27/07

METALS STANDARD DOCUMENTATION

Standard: QC Check Standard 7
Date Received/Prepped: 8/27/2007
Date Expired: 8/31/2007
Manufacturer: Crescent Chemical Co. Inc.
Matrix: 5% HNO3
Amount: 100

ME #: 0705011
By: STE
Lot #: 0744381
Certificate:
NIST SRM:
Storage: Room Temp.

Component	Comment	Conc. Unit:
Ag	Catalog No: QC-007.1	100 ug/mL
Al		100 ug/mL
B		100 ug/mL
Ba		100 ug/mL
K		1000 +/- 5 ug/mL
Na		100 ug/mL
Si		50 ug/mL

ME 07 08 011

Laboratory Report - Certificate of Analysis

Environmental Multielement Standard

QC Check Standard 7

CATALOG NO: QC-007.1

CONTENTS: See Below

MATRIX: 5% HNO₃/tr. F

LOT NO.: 074438I

This solution is intended for use as a calibration standard for plasma emission spectroscopy (ICP or DCP). It is a multielement solution, that was prepared gravimetrically to contain the elements/concentrations shown below.

In order to verify the concentration, the final solution was checked against NIST SRMS: 3101a, 3104a, 3107, 3141a, 3150, 3151, and 3152a.

Concentrations are given in µg/mL unless noted otherwise.

Ag 100	Al 100	B 100	Ba 100	K 1,000±5
Na 100	Si 50.0			

Crescent Chemical Co. Inc.

Julie M. MacIntosh
QA Manager

EXPIRES: August 2008

CRESCENT CHEMICAL CO, INC., waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.

*Crescent Chemical Co, Inc., 2 Oval Drive, Islandia, NY 11749
(516) 348-0333 - Fax (516) 348-0913*

Initial: CSK
Date: 05/05/08

METALS STANDARD DOCUMENTATION

Standard: ICP LINEARITY CHECK **ME #:** 0805001
Date Received/Prepped: 5/5/2008 **By:** CSK
Date Expired: 11/5/2008 **Lot #:** VARIOUS
Manufacturer: MWH-STE **Certificate:**
Matrix: 5% HNO3 **NIST SRM:**
Amount: 500ML **Storage:** Room Temp

Component	Comment	Conc. Unit:
CA	15.0ML ME0702002/500ML	300 ppm
K	15.0ML ME0702005/500ML	300 ppm
MG	10.0ML ME0702004/500ML	200 ppm
NA	15.0ML ME0702003/500ML	300 ppm
FE	5.0ML ME0701008/500ML	100 ppm

Initial:

Date:

WBY

2/20/07

METALS STANDARD DOCUMENTATION

Standard: Potassium 10000ppm Stock Std **ME #:** 0702005
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B056
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO3 **NIST SRM:** 3141
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
K	P/N 4400-10M411	10000 ppm

AUG 16 '08



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

ME070205

P/N 4400-10M411**P/N S4400-10M411**

Single-Element Potassium Standard

K in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B056

Material Source: Potassium Nitrate (KNO₃)

Source Purity: 99.999%

Specific Gravity: 1.019 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3141. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>									
Al	0.39	0.1	Cu	0.16	0.1	Pb	ND	0.1	K	X	70
Sb	0.34	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.14	0.1	Eu	ND	0.1	Mg	2.6	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	0.93	1	Rb	9.5	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cf	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	20
Ca	82	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	19	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

WBH

2/10/07

METALS STANDARD DOCUMENTATION

Standard: Magnesium 1000ppm Stock Std **ME #:** 0702004
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B058
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3131
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Mg	P/N 4400-10M311	10000 ppm



Aug 16 98
*Innovative Solutions
in Analytical Science and
Technology*

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com

EUROPE

P.O. Box 2704
1000 CS Amsterdam
The Netherlands
+31 20 638 05 97
Fax +31 20 420 28 36
www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M311

P/N S4400-10M311

Single-Element Magnesium Standard

Mg in 4% HNO₃
10,000 ± 30 µg/mL

47070204

Lot # 07B058

Material Source: Magnesium Metal

Source Purity: 99.99%

Specific Gravity: 1.056 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3131. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	ppb	DL									
Al	28	0.1	Cu	1.6	0.1	Pb	7.7	0.7	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	0.28	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.28	0.1	Eu	ND	0.1	Mg	X	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	0.23	0.1	Mn	19.8	1	Rb	ND	0.1
Bi	ND	0.1	Ga	0.18	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	1.1	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	1	0.1	Si	64	20
Ca	ND	7	Ho	ND	0.1	Nb	ND	0.1	Ag	0.19	0.1
Ce	2.1	0.1	I	1	0.2	Os	ND	0.1	Na	7.2	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	0.19	0.1
Cr	ND	1	Fe	80	30	P	ND	10	Ta	ND	0.1
Co	ND	0.1	La	0.76	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

W31

2/29/22

METALS STANDARD DOCUMENTATION

Standard: Sodium 10000ppm Stock Std **ME #:** 0702003
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B057
Manufacturer: CPI **Certificate:** Y
Matrix: 1% HNO3 **NIST SRM:** 3152a
Amount: 250 mL Room temp. storage

Component	Comment	Conc. Unit:
Na	P/N 4400-10M521	10000 ppm

AUG 16 1998



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard 707.525.5788
Santa Rosa, CA 95403 800.878.7654
www.cpiinternational.com Fax 707.545.7901

EUROPE

P.O. Box 2704 +31 20 638 05 97
1000 CS Amsterdam Fax +31 20 420 28 36
The Netherlands www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M521

170702003

P/N S4400-10M521

Single-Element Sodium Standard

Na in 1% HNO₃

10,000 ± 30 µg/mL

Lot # 07B057

Material Source: Sodium Nitrate (NaNO₃)

Source Purity: 99.99%

Specific Gravity: 1.053 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3152a. Trace impurities of the standard solution at 1000µg/mL were analyzed by ICP-MS.

	ppb	DL									
Al	1.5	0.1	Cu	0.45	0.1	Pb	ND	0.1	K	ND	70
Sb	ND	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	0.13	0.1	Eu	ND	0.1	Mg	2.3	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	ND	1	Rb	ND	0.1
Bi	ND	0.1	Ga	ND	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	ND	0.1	Mo	ND	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	0.4	0.1	Si	50	8
Ca	120	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	ND	0.2	Os	ND	0.1	Na	X	1
Cs	ND	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	1	0.1
Cr	ND	1	Fe	ND	30	P	18	10	Ta	ND	0.1
Co	ND	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

X=Major Element INT=Interference from Major Element ND=Not Detected DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the United States or +31 20 638 05 97 in Europe.

Initial:
Date:

W37
2/20/07

METALS STANDARD DOCUMENTATION

Standard: Calcium 1000ppm Stock Std **ME #:** 0702002
Date Received/Prepped: 2/20/2007 **By:** WBH
Date Expired: 8/16/2008 **Lot #:** 07B065
Manufacturer: CPI **Certificate:** Y
Matrix: 4% HNO3 **NIST SRM:** 3109a
Amount: 250 mL **Room temp. storage**

Component	Comment	Conc. Unit:
Ca	P/N 4400-10M91	10000 ppm



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com

AUG 16 08

EUROPE

P.O. Box 2704
1000 CS Amsterdam
The Netherlands
+31 20 638 05 97
Fax +31 20 420 28 36
www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M91
P/N S4400-10M91

180702602

Single-Element Calcium Standard

Ca in 4% HNO₃

10,000 ± 30 µg/mL

Lot # 07B065

Material Source: Calcium Carbonate (CaCO₃)

Source Purity: 99.997%

Specific Gravity: 1.035 @ 21 °C

This standard solution was prepared using high-purity salt, sub-boiled distilled nitric acid and 18-megaohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3109a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

ppb	DL								
Al 7	0.1	Cu 1.7	0.1	Pb 0.23	0.1	K ND	70	Tl 0.27	0.1
Sb ND	0.1	Dy ND	0.1	Li ND	0.4	Pr ND	0.1	Th ND	0.1
As ND	6	Er ND	0.1	Lu ND	1	Re ND	0.1	Tm ND	0.1
Ba 1.5	0.1	Eu ND	0.1	Mg 38	0.2	Rh ND	0.1	Sn ND	0.1
Be ND	0.1	Gd ND	0.1	Mn ND	1	Rb ND	0.1	Tl ND	0.1
Bi ND	0.1	Ga ND	0.1	Hg ND	0.2	Ru ND	0.1	W ND	0.1
B 1.5	4	Ge ND	0.1	Mo ND	0.1	Sm ND	0.1	U ND	0.1
Br ND	10	Au ND	0.1	Nd ND	0.1	Se ND	6	V ND	1
Cd ND	0.1	Hf ND	0.1	Nr 3	0.1	Si 47	8	Yb ND	0.1
Ca X	7	Ho ND	0.1	Nb ND	0.1	Ag ND	0.1	Y ND	0.1
Ce ND	0.1	I 0.27	0.2	Os ND	0.1	Na 116	1	Zn 3.5	2
Cs ND	0.1	Ir ND	0.1	Pd ND	0.1	Sr 55	0.1	Zr ND	0.1
Cr ND	1	Fe INT	30	P ND	10	Ta ND	0.1		
Co INT	0.1	La 0.41	0.1	Pt ND	0.1	Te ND	0.1		

INT=Interference from Major Element ND=None Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial: WZy
Date: 1/27/07

METALS STANDARD DOCUMENTATION

Standard: FE 10000ppm Stock Std ME #: 701008
Date Received/Prepped: 1/26/2007 By: WBH
Date Expired: 7/19/2008 Lot #: 06I143
Manufacturer: CPI Certificate:
Matrix: 4% HNO3 NIST SRM: 3126a
Amount: 100 mL Storage: Room Temp

Component	Comment	Conc. Unit:
Fe	PN4400-10M261	10000 PPM



Innovative Solutions
in Analytical Science and
Technology

USA

5580 Skylane Boulevard
Santa Rosa, CA 95403
www.cpiinternational.com

EUROPE

P.O. Box 2704
1000 CS Amsterdam
The Netherlands

+31 20 638 05 97

Fax +31 20 420 28 36

www.cpiinternational.com

CERTIFICATE OF ANALYSIS

P/N 4400-10M261

P/N S4400-10M261

Single-Element Iron Standard

Fe in 4% HNO₃

10,000 ± 30 µg/mL

HTE070100X

Lot # 06I143

Material Source: Iron Metal

Source Purity: 99.999%

Specific Gravity: 1.062 @ 21 °C

This standard solution was prepared using high-purity metal, sub-boiled distilled nitric acid and 18-megohm deionized water. The starting material was weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

The standard solution concentration was certified by ICP against the National Institute of Standards and Technology's SRM 3126a. Trace impurities of the standard solution at 1000 µg/mL were analyzed by ICP-MS.

	<u>ppb</u>	<u>DL</u>									
Al	INT	0.1	Cu	6.4	0.1	Pb	ND	0.1	K	ND	70
Sb	0.35	0.1	Dy	ND	0.1	Li	ND	0.4	Pr	ND	0.1
As	ND	6	Er	ND	0.1	Lu	ND	1	Re	ND	0.1
Ba	ND	0.1	Eu	ND	0.1	Mg	1.3	0.2	Rh	ND	0.1
Be	ND	0.1	Gd	ND	0.1	Mn	INT	1	Rb	ND	0.1
Bi	ND	0.1	Ga	0.41	0.1	Hg	ND	0.2	Ru	ND	0.1
B	ND	4	Ge	INT	0.1	Mo	4.9	0.1	Sm	ND	0.1
Br	ND	10	Au	ND	0.1	Nd	ND	0.1	Se	ND	6
Cd	ND	0.1	Hf	ND	0.1	Ni	9.3	0.1	Si	INT	8
Ca	15	7	Ho	ND	0.1	Nb	ND	0.1	Ag	ND	0.1
Ce	ND	0.1	I	0.34	0.2	Os	ND	0.1	Na	8	1
Cs	0.34	0.1	Ir	ND	0.1	Pd	ND	0.1	Sr	ND	0.1
Cr	3.3	1	Fe	X	30	P	28	10	Ta	ND	0.1
Co	12	0.1	La	ND	0.1	Pt	ND	0.1	Te	ND	0.1

INT=Interference from Major Element ND=Not Detected X=Major Element DL=Detection Limit

Accuracy and stability are guaranteed to within plus or minus 0.3% of the certified value for 18 months after the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-878-7654 in the USA or +31 20 638 05 97 in Europe.

Initial:

Date:

CSK6-1-08

METALS STANDARD DOCUMENTATION

Standard: INTERFERENCE CHECK STANDAF **ME #:** 0806002
Date Received/Prepped: 6/1/2008 **By:** CSK
Date Expired: 12/1/2008 **Lot #:** various
Manufacturer: MWH-CSK **Certificate:**
Matrix: 5% HNO3 **NIST SRM:**
Amount: 500 mL **Storage:** Room Temp.

Component	Comment	Conc. Unit:
Al	P/N 4400-INTA1-500(25ML)	250 ppm
Ca	P/N 4400-INTB1-100(2.5ML)	250 ppm
Fe		100 ppm
Mg		250 ppm
AG		0.5 ppm
BA		0.25 ppm
BE		0.25 ppm
CD		0.5 ppm
Co		0.25 ppm
Cr3		0.25 ppm
Cu		0.5 ppm
Mn		0.5 ppm
Ni		0.25 ppm
Pb		0.25 ppm
V		0.25 ppm
Zn		0.5 ppm

TI
 V
 Zn
 Si

100 ppm
 100 ppm
 100 ppm
 50 ppm

31-4-08

METALS STANDARD DOCUMENTATION

Standard: INTERFERENCE CHECK STANDAR **ME #:** 0806001
Date Received/Prepped: 6/1/2008 **By:** CSK
Date Expired: 12/1/2008 **Lot #:** various
Manufacturer: MWH-CSK **Certificate:**
Matrix: 5% HNO3 **NIST SRM:**
Amount: 500ML **Storage:** Room Temp

Component	Comment	Conc. Unit:
AL	P/N 4400-INTA1-500	250 ppm
CA	25ML IN 500ML	250 ppm
FE		100 ppm
MG		250 ppm
AG		
BA		
BE		
CD		
CO		
CR		
CU		
MN		
NI		
PB		
V		
ZN		