

LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Northgate Environmental Management, Inc.
1100 Quail Street Ste. 102
Newport Beach, CA 92660
ATTN: Ms. Cindy Arnold

June 30, 2010

SUBJECT: Tronox LLC Facility, PCS, Henderson, Nevada,
Data Validation

Dear Ms. Arnold,

Enclosed are revised data validation reports for the fractions listed below. Please replace the previously submitted reports with the enclosed revised reports.


LDC Project # 23308:

<u>SDG #</u>	<u>Fraction</u>
280-2500-2	Perchlorate
280-2699-5, 280-2995-1	Semivolatiles
280-2995-2	Arsenic

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Northgate Environmental Management, Inc.
1100 Quail Street Ste. 102
Newport Beach, CA 92660
ATTN: Ms. Cindy Arnold

June 22, 2010

SUBJECT: Tronox LLC Facility, PCS, Henderson, Nevada,
Data Validation

Dear Ms. Arnold,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on June 3, 2010. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 23308:

SDG #

Fraction

280-2500-2, 280-2500-8, 280-2699-5
280-2771-5, 280-2836-4, 280-2836-6
280-2836-8, 280-2836-9, 280-2879-6
280-2879-7, 280-2879-8, 280-2960-3
280-2960-5, 280-2960-6, 280-2960-7
280-2995-1, 280-2995-2, 280-2995-5
280-2995-6, 280-3059-3, 280-3059-4
280-3059-6, 280-3059-8, 280-3100-1
280-3100-4, 280-3100-6

Semivolatiles, Chlorinated Pesticides
Metals, Perchlorate

The data validation was performed under Stage 2B/4 guidelines. The analyses were validated using the following documents, as applicable to each method:

- Standard Operating Procedures (SOP) 40, Data Review/Validation, BRC 2009
- Quality Assurance Project Plan Tronox LLC Facility, Henderson Nevada, June 2009
- NDEP Guidance, May 2006
- USEPA, Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review, June 2008
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, October 2004

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

LDC #23308 (Tronox LLC-Northgate, Henderson NV / Tronox PCS)

LDC	SDG#	DATE REC'D	DATE DUE (3)	SVOA (8270C)		Pest. (8081A)		As (6020)		Co (6020)		Pb (6020)		Mn (6020)		Mg (6020)		CLO ₄ (314.0)		
				W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W
Matrix: Water/Soil																				
A	280-2500-2	06/03/10	06/24/10	1	0	1	0	2	0	2	0	2	0	2	0	2	0	2	0	17
A	280-2500-2	06/03/10	06/24/10	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	3	
B	280-2500-8	06/03/10	06/24/10	-	-	-	-	0	2	-	-	-	-	-	-	-	-	-	-	
B	280-2500-8	06/03/10	06/24/10	-	-	-	-	0	1	-	-	-	-	-	-	-	-	-	-	
C	280-2699-5	06/03/10	06/24/10	0	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
D	280-2771-5	06/03/10	06/24/10	-	-	0	1	-	-	-	-	-	-	-	-	-	-	-	-	
E	280-2836-4	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
F	280-2836-6	06/03/10	06/24/10	0	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
G	280-2836-8	06/03/10	06/24/10	0	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
H	280-2836-9	06/03/10	06/24/10	-	-	-	-	0	1	-	-	-	-	-	-	-	-	-	-	
H	280-2836-9	06/03/10	06/24/10	-	-	-	-	0	1	-	-	-	-	-	-	-	-	-	-	
I	280-2879-6	06/03/10	06/24/10	0	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
J	280-2879-7	06/03/10	06/24/10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0	2	
K	280-2879-8	06/03/10	06/24/10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0	1	
L	280-2960-3	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
M	280-2960-5	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
N	280-2960-6	06/03/10	06/24/10	0	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
N	280-2960-6	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
O	280-2960-7	06/03/10	06/24/10	-	-	-	-	0	1	-	-	-	-	-	-	-	-	-	-	
P	280-2995-1	06/03/10	06/24/10	0	7	-	-	0	13	-	-	-	-	0	5	-	-	0	2	
P	280-2995-1	06/03/10	06/24/10	0	1	-	-	0	1	-	-	-	-	0	0	-	-	0	0	
Q	280-2995-2	06/03/10	06/24/10	1	0	-	-	1	0	-	-	-	-	-	-	-	-	-	-	
R	280-2995-5	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
R	280-2995-5	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
S	280-2995-6	06/03/10	06/24/10	-	-	-	-	0	2	-	-	-	-	-	-	-	-	-	-	
T	280-3059-3	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
U	280-3059-4	06/03/10	06/24/10	0	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
V	280-3059-6	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
W	280-3059-8	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
X	280-3100-1	06/03/10	06/24/10	0	6	-	-	0	6	-	-	-	-	-	-	-	-	0	3	
Y	280-3100-4	06/03/10	06/24/10	0	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Z	280-3100-6	06/03/10	06/24/10	0	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Total	T/LR			2	40	1	1	3	28	2	0	2	0	2	5	2	0	28	0	0

Shaded cells indicate Stage 4 validation (all other cells are Stage 2B validation). These sample counts do not include MS/MSD, and DUPs

EDD CHECKLIST

LDC #: 23308

SDG #: 280-2500-2, 280-2500-8, 280-2699-5, 280-2771-5
280-2836-4, 280-2836-6, 280-2836-8, 280-2836-9
280-2879-6, 280-2879-7, 280-2879-8, 280-2960-3
280-2960-5, 280-2960-6, 280-2960-7, 280-2995-1
280-2995-2, 280-2995-5, 280-2995-6, 280-3059-3
280-3059-4, 280-3059-6, 280-3059-8, 280-3100-1
280-3100-4, 280-3100-6

Tronox Northgate Henderson Worksheet

EDD Area	Yes	No	NA	Findings/Comments
I. Completeness				
Is there an EDD for the associated Tronox validation report?	X			
II. EDD Qualifier Population				
Were all qualifiers from the validation report populated into the EDD?	X			
III. EDD Lab Anomalies				
Were EDD anomalies identified?		X		
If yes, were they corrected or documented for the client?			X	See EDD_discrepancy_form_LDC23308_062210.doc
IV. EDD Delivery				
Was the final EDD sent to the client?	X			

**Tronox LLC Facility, PCS, Henderson, Nevada
Data Validation Reports
LDC #23308**

Semivolatiles

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 16, 2010

LDC Report Date: June 21, 2010

Matrix: Water

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2500-2

Sample Identification

EB-04152010-1-RZD

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample EB-04152010-1-RZD was identified as an equipment blank. No semivolatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2500-2	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2500-2**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2500-2	EB-04152010-1-RZD	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308A2a
SDG #: 280-2500-2
Laboratory: Test America

Stage 2B

Date: 6/14/10
Page: 1 of 1
Reviewer: JVG
2nd Reviewer: CF

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/16/10</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	<u>2 RSD r2</u>
IV.	Continuing calibration/ICV	A	<u>CCW / CV ≤ 25%</u>
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	<u>client spec</u>
VIII.	Laboratory control samples	A	<u>LCS / D</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	<u>EB = 1</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet
ND = No compounds detected
R = Rinsate
FB = Field blank
D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

Water

<u>1</u>	<u>EB-04152010-1RZD</u>	11		21		31
<u>2</u>	<u>MB 280-11838/A-A</u>	12		22		32
3		13		23		33
4		14		24		34
5		15		25		35
6		16		26		36
7		17		27		37
8		18		28		38
9		19		29		39
10		20		30		40

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 21, 2010

LDC Report Date: June 29, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2699-5

Sample Identification

SSAK7-01-9BPC

SSAK7-01-10BPC

Introduction

This data review covers 2 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

***V. Blanks**

Method blanks were reviewed for each matrix as applicable. No semivolatiles were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-12911/1-A	4/28/10	Bis(2-ethylhexyl)phthalate	62.7 ug/Kg	All samples in SDG 280-2699-5

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
SSAK7-01-9BPC	Bis(2-ethylhexyl)phthalate	73 ug/Kg	73U ug/Kg
SSAK7-01-10BPC	Bis(2-ethylhexyl)phthalate	110 ug/Kg	110U ug/Kg

*Corrected sample ID in table above.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2699-5

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2699-5	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2699-5**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2699-5	SSAK7-01-9BPC SSAK7-01-10BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

***Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2699-5**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-2699-5	SSAK7-01-9BPC	Bis(2-ethylhexyl)phthalate	73U ug/Kg	A	bl
280-2699-5	SSAK7-01-10BPC	Bis(2-ethylhexyl)phthalate	110U ug/Kg	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2699-5**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2699-5**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B

LDC #: 23308C2a
 SDG #: 280-2699-5
 Laboratory: Test America

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JVG
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/21/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD r ²
IV.	Continuing calibration/ICV	A	CV/ICV = 25%
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB-04072010- R2D (280-2216-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet
 ND = No compounds detected
 R = Rinsate
 FB = Field blank
 D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: Soil

1	SSAK7-01-9BPC	11	MB 280-1291/1-A	21		31
2	SSAK7-01-10BPC	12		22		32
3		13		23		33
4		14		24		34
5		15		25		35
6		16		26		36
7		17		27		37
8		18		28		38
9		19		29		39
10		20		30		40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

LDC #: 23308 C24
 SDG #: Seaney

VALIDATION FINDINGS WORKSHEET
Blanks

Page: 1 of 1
 Reviewer: JVB
 2nd Reviewer: D

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Was a method blank analyzed for each matrix?
- Y N N/A Was a method blank analyzed for each concentration preparation level?
- Y N N/A Was a method blank associated with every sample?
- Y N N/A Was the blank contaminated? If yes, please see qualification below.

Blank extraction date: 4/28/06 Blank analysis date: 4/25/06 Associated Samples: All (62)

Conc. units: mg/kg

Compound	Blank ID	Sample Identification
EEF	Mb 280-12911 A-A 62.7	1 73/u 110/u

Blank extraction date: _____ Blank analysis date: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification

5x Phthalates
 2x all others

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 23, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2836-4

Sample Identification

SSAJ2-01-1BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-12911/1-A	4/28/10	Bis(2-ethylhexyl)phthalate	62.7 ug/Kg	All samples in SDG 280-2836-4

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
SSAJ2-01-1BPC	Bis(2-ethylhexyl)phthalate	89 ug/Kg	89U ug/Kg

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2836-4

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2836-4	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2836-4**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2836-4	SSAJ2-01-1BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2836-4**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-2836-4	SSAJ2-01-1BPC	Bis(2-ethylhexyl)phthalate	89U ug/Kg	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2836-4**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2836-4**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

Stage 2B

LDC #: 23308E2a
 SDG #: 280-2836-4
 Laboratory: Test America

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JN
 2nd Reviewer: JN

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/23/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	20 RSD NY
IV.	Continuing calibration/ICV	A	CO/ICV ≤ 25%
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB-04672010- RZD (280-2216-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: Soil

1	SSAJ2-01-1BPC	11	21	31
2	MB 280-12911/A-A	12	22	32
3		13	23	33
4		14	24	34
5		15	25	35
6		16	26	36
7		17	27	37
8		18	28	38
9		19	29	39
10		20	30	40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

LDC #: 23308 Ena
SDG #: Sa Carr

Blanks

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Was a method blank analyzed for each matrix?
- Y N N/A Was a method blank analyzed for each concentration preparation level?
- Y N N/A Was a method blank associated with every sample?
- Y N N/A Was the blank contaminated? If yes, please see qualification below.

Blank extraction date: 4/24/10 Blank analysis date: 4/30/10 Associated Samples: A1 (61)

Compound	Blank ID	Sample Identification
EE	MB 286-12911 1-A 62.7 89/4	

Blank extraction date: _____ Blank analysis date: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification

5x Phthalates
2x all others

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 23, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2836-6

Sample Identification

SSAI3-06-3BPC
SSAI3-06-4BPC

Introduction

This data review covers 2 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2836-6

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2836-6	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2836-6**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2836-6	SSAI3-06-3BPC SSAI3-06-4BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2836-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2836-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2836-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308F2a

SDG #: 280-2836-6

Laboratory: Test America

Stage 2B

Date: 6/14/10

Page: 1 of 1

Reviewer: JLG

2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 6/23/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2% RSD r ²
IV.	Continuing calibration/ICV	A	CV/ICV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client Spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB-04072010 - R2D (280-2216-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

571

1	SSAI3-06-3BPC	11		21		31
2	SSAI3-06-4BPC	12		22		32
3	MB 280-13178/-A	13		23		33
4		14		24		34
5		15		25		35
6		16		26		36
7		17		27		37
8		18		28		38
9		19		29		39
10		20		30		40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis(2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJ.J. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

Field Blanks

Page: 1 of 1
 Reviewer: N/A
 2nd Reviewer: Q

LDC #: 23308 F > A
 SDG #: See Copy

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)
 Y/N N/A Were field blanks identified in this SDG?
 Y/N N/A Were target compounds detected in the field blanks?
 Blank units: wg/l Associated sample units: wg/kg
 Sampling date: 4/6/10 Field Blank (circle one) Rinsate / Other: A11
 Field blank type: (circle one) Field Blank / Rinsate / Other: A11 Associated Samples:

Compound	Blank ID	Sample Identification				
	FB-0407010-RZD					
EE	2.2	(> 5 x FB)				
CRQL						

Blank units: _____ Associated sample units: _____

Sampling date: _____ Associated Samples: _____

Field blank type: (circle one) Field Blank / Rinsate / Other: _____

Compound	Blank ID	Sample Identification				
CRQL						

5x Phthalates
 2x All others

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 23, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2836-8

Sample Identification

SSAI3-05-8BPC
SSAI3-05-10BPC

Introduction

This data review covers 2 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2836-8

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2836-8	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2836-8**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2836-8	SSAI3-05-8BPC SSAI3-05-10BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2836-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2836-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2836-8**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308G2a
SDG #: 280-2836-8
Laboratory: Test America

Stage 2B

Date: 6/14/10
Page: 1 of 1
Reviewer: JVG
2nd Reviewer: 9

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

Validation Area		Comments
I.	Technical holding times	A Sampling dates: 4/23/10
II.	GC/MS Instrument performance check	A
III.	Initial calibration	A 20 RSD 12
IV.	Continuing calibration/ICV	A CV/ICV ≤ 25 %
V.	Blanks	A
VI.	Surrogate spikes	A
VII.	Matrix spike/Matrix spike duplicates	N Client spec
VIII.	Laboratory control samples	A ICS
IX.	Regional Quality Assurance and Quality Control	N
X.	Internal standards	A
XI.	Target compound identification	N
XII.	Compound quantitation/CRQLs	N
XIII.	Tentatively identified compounds (TICs)	N
XIV.	System performance	N
XV.	Overall assessment of data	A
XVI.	Field duplicates	N
XVII.	Field blanks	SW TB = FB-04072010- R2D (280-2216-2)

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: Soil

1	SSAI3-05-8BPC	11	21	31
2	SSAI3-05-10BPC	12	22	32
3	MB 280-13949/21-A	13	23	33
4		14	24	34
5		15	25	35
6		16	26	36
7		17	27	37
8		18	28	38
9		19	29	39
10		20	30	40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis(2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 26, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2879-6

Sample Identification

SSAJ2-02-3BPC
SSAJ2-02-4BPC
SSAJ2-02-5BPC

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2879-6

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
SSAJ2-02-3BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2879-6	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2879-6**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2879-6	SSAJ2-02-3BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-2879-6	SSAJ2-02-3BPC SSAJ2-02-4BPC SSAJ2-02-5BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2879-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2879-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET

LDC #: 2330812a
SDG #: 280-2879-6
Laboratory: Test America

Stage 2B

Date: 6/12/10
Page: 1 of 1
Reviewer: JYG
2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/26/10</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	<u>2 RSD 12</u>
IV.	Continuing calibration/ICV	A	<u>CO / W ≤ 252</u>
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	<u>Client Spec</u>
VIII.	Laboratory control samples	A	<u>LCS</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	SW	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	<u>FB = FB-0407 2010 - RZD (280-2216-2)</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

Soil

1	SSAJ2-02-3BPC	11	<u>MB 280-13949/21A-21</u>	21	31	
2	SSAJ2-02-4BPC	12		22	32	
3	SSAJ2-02-5BPC	13		23	33	
4		14		24	34	
5		15		25	35	
6		16		26	36	
7		17		27	37	
8		18		28	38	
9		19		29	39	
10		20		30	40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

Compound Quantitation and Reported CRQLs

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

Y N N/A
Y N N/A

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		1	GGG, HHH unresolved		J/MS/P (g)

Comments: See sample calculation verification worksheet for recalculations

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 27, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2960-3

Sample Identification

SSAR7-02-1BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Bis(2-ethylhexyl)phthalate	2.7 ug/L	All samples in SDG 280-2960-3

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2960-3	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2960-3**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2960-3	SSAR7-02-1BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2960-3**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2960-3**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308L2a
 SDG #: 280-2960-3
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JLG
 2nd Reviewer: 9

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/27/10</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	<u>2 RSD</u> <u>r²</u>
IV.	Continuing calibration/ICV	A	<u>COV/ICV ≤ 25 %</u>
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	<u>Client spec</u>
VIII.	Laboratory control samples	A	<u>LCS</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	<u>FB = FB 04062010 - RZB (280-2131-2)</u>

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

Soil

1	SSAR7-02-1BPC	11		21		31	
2	<u>MB 280-13357/1-A</u>	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 27, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2960-5

Sample Identification

SSAK8-05-1BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2960-5

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2960-5	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2960-5**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2960-5	SSAK8-05-1BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2960-5**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2960-5**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

LDC #: 23308M2a
 SDG #: 280-2960-5
 Laboratory: Test America

VALIDATION COMPLETENESS WORKSHEET
 Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JVG
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/27/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD 12
IV.	Continuing calibration/ICV	A	CV/ICV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB-04072010-RZD (280-2216-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAK8-05-1BPC	11		21		31	
2	MB 280 - 13949 / 21A	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis(2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF. ** = Calibration check compound (CCC) for %RSD.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 27, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B & 4

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2960-6

Sample Identification

SSAK8-04-4BPC

SSAK8-04-5BPC**

SSAK8-04-5BPC_FD

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/07/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-2960-6

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2960-6	All compounds reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples SSAK8-04-5BPC** and SSAK8-04-5BPC_FD were identified as field duplicates. No semivolatiles were detected in any of the samples.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2960-6**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2960-6	SSAK8-04-4BPC SSAK8-04-5BPC** SSAK8-04-5BPC_FD	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2960-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2960-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2960-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308N2a
 SDG #: 280-2960-6
 Laboratory: Test America

Stage 2B/4

Date: 6/15/10
 Page: 1 of 1
 Reviewer: JY
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/27/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD ✓
IV.	Continuing calibration/ICV	A	COV/ICV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	ND	D = 2, 3
XVII.	Field blanks	SW	FB = FB-04672010-RZD (280-2216-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:
 ** Stage 4 (Soil)

1	SSAK8-04-4BPC	11	MB 280-13949 21-A	21	31
2	SSAK8-04-5BPC ** D	12		22	32
3	SSAK8-04-5BPC_FD D	13		23	33
4		14		24	34
5		15		25	35
6		16		26	36
7		17		27	37
8		18		28	38
9		19		29	39
10		20		30	40

LDC #: 77768 N2a
 SDG #: See cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

Method: Semivolatiles (EPA SW 846 Method 8270C)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	/			
Cooler temperature criteria was met.	/			
II. GC/MS Instrument performance check				
Were the DFTPP performance results reviewed and found to be within the specified criteria?	/			
Were all samples analyzed within the 12 hour clock criteria?	/			
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	/			
Were all percent relative standard deviations (%RSD) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	/			
Was a curve fit used for evaluation?	/			
Did the initial calibration meet the curve fit acceptance criteria of ≥ 0.990 ?	/			
Were all percent relative standard deviations (%RSD) $\leq 30\%$ and relative response factors (RRF) > 0.05 ?	/			
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	/			
Were all percent differences (%D) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	/			
Were all percent differences (%D) $\leq 25\%$ and relative response factors (RRF) ≥ 0.05 ?	/			
V. Blanks				
Was a method blank associated with every sample in this SDG?	/			
Was a method blank analyzed for each matrix and concentration?	/			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.		/		
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	/			
If 2 or more base neutral or acid surrogates were outside QC limits, was a reanalysis performed to confirm %R?			/	
If any %R was less than 10 percent, was a reanalysis performed to confirm %R?			/	
VII. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD. Soil / Water.		/		
Was a MS/MSD analyzed every 20 samples of each matrix?		/		
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?			/	
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	/			

LDC #: 23308 N 2a
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

Validation Area	Yes	No	NA	Findings/Comments
Was an LCS analyzed per extraction batch?	<input checked="" type="checkbox"/>			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>			
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		<input checked="" type="checkbox"/>		
Were the performance evaluation (PE) samples within the acceptance limits?			<input checked="" type="checkbox"/>	
X. Internal standards				
Were internal standard area counts within -50% or +100% of the associated calibration standard?	<input checked="" type="checkbox"/>			
Were retention times within + 30 seconds from the associated calibration standard?	<input checked="" type="checkbox"/>			
XI. Target compound identification				
Were relative retention times (RRT's) within + 0.06 RRT units of the standard?	<input checked="" type="checkbox"/>			
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	<input checked="" type="checkbox"/>			
Were chromatogram peaks verified and accounted for?	<input checked="" type="checkbox"/>			
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	<input checked="" type="checkbox"/>			
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>			
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 10 percent relative intensity) in the reference spectrum evaluated in sample spectrum?			<input checked="" type="checkbox"/>	
Were relative intensities of the major ions within + 20% between the sample and the reference spectra?			<input checked="" type="checkbox"/>	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?			<input checked="" type="checkbox"/>	
XIV. System performance				
System performance was found to be acceptable.	<input checked="" type="checkbox"/>			
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>			
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	<input checked="" type="checkbox"/>			
Target compounds were detected in the field duplicates.		<input checked="" type="checkbox"/>		
XVII. Field blanks				
Field blanks were identified in this SDG.	<input checked="" type="checkbox"/>			
Target compounds were detected in the field blanks.	<input checked="" type="checkbox"/>			

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

LDC #: 23308 N 29

SDG #: See Com

VALIDATION FINDINGS WORKSHEET

Field Blanks

Page: 1 of 1

Reviewer: JVG

2nd Reviewer: Y

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Y/N N/A Were field blanks identified in this SDG?

Y/N N/A Were target compounds detected in the field blanks?

Blank units: ug/L Associated sample units: ug/kg

Sampling date: 4/07/10

Field blank type: (circle one) Field Blank (X) Rinsate / Other:

Associated Samples: All (ND)

Compound	Blank ID	Sample Identification
	<u>FB-0902010- R 2 B</u>	
<u>EEB</u>	<u>2, 2</u>	
CRQL		

Blank units: _____ Associated sample units: _____

Sampling date: _____

Field blank type: (circle one) Field Blank / Rinsate / Other:

Associated Samples: _____

Compound	Blank ID	Sample Identification
CRQL		

5x Phthalates
2x All others

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_{is}) / (A_{is})(C_x)$ A_x = Area of Compound A_{is} = Area of associated internal standard
 average RRF = sum of the RRFs/number of standards C_x = Concentration of compound, C_{is} = Concentration of internal standard
 %RSD = $100 * (S/X)$ S = Standard deviation of the RRFs, X = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (50 std)	RRF (50 std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD		
1	ICAL	5/4/2010	1,4-Dioxane (IS1)	0.6700	0.6700	0.6718	0.6718	4.8	4.8	4.84	4.84
	MSS Y		Naphthalene (IS2)	1.0419	1.0419	0.9990	0.9990	8.0	8.0	8.04	8.04
			Fluorene (IS3)	1.3468	1.3468	1.3058	1.3058	8.0	8.0	8.02	8.02
			Hexachlorobenzene (IS4)	0.1996	0.1996	0.1947	0.1946	2.8	2.8	2.82	2.82
			Chrysene (IS5)	1.0651	1.0651	1.0509	1.0509	7.9	7.9	7.92	7.92
			Benzo(a)pyrene (IS6)	1.1462	1.1462	1.1042	1.1042	3.3	3.3	3.28	3.28

Inc IS/Cpd	Area cpd	Area IS
40/50	290884	347342
40/50	1775212	1363095
40/50	1313767	780352
40/50	335135	1343097
40/50	1866391	1401828
40/50	1809781	1263104

Conc	1,4-Dioxane	Naphthalene	Fluorene	Hexachlorob	Chrysene	Benzo(a)py
4.00		1.1191	1.4337		1.1603	1.0624
10.00	0.7200	1.0315	1.4006	0.1914	1.1208	1.0548
20.00	0.7128	1.0652	1.3863	0.1995	1.1246	1.1138
50.00	0.6700	1.0419	1.3468	0.1996	1.0651	1.1462
80.00	0.6540	0.9915	1.2886	0.2021	1.0538	1.1452
120.00	0.6579	0.9413	1.2601	0.1897	0.9907	1.1311
160.00	0.6321	0.9206	1.1894	0.1905	0.9596	1.1027
200.00	0.6558	0.8811	1.1406	0.1897	0.9324	1.0775
X =	0.6718	0.9990	1.3058	0.1946	1.0509	1.1042
S =	0.0325	0.0803	0.1047	0.0055	0.0832	0.0362

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

Where:
 $\% \text{ Difference} = 100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$
 $\text{RRF} = (\text{Ax}) / (\text{Cis}) / (\text{Ais}) / (\text{Cx})$
 ave. RRF = initial calibration average RRF
 RRF = continuing calibration RRF
 Ax = Area of compound Ais = Area of associated internal standard
 Cx = Concentration of compound Cis = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference IS)	Average RRF (Initial RRF)	Reported (CC RRF)	Recalculated (CC RRF)	Reported %D	Recalculated %D
1	Y2041	05/05/10	1,4-Dioxane (IS1)	0.6718	0.6326	0.6326	5.8	5.8
			Naphthalene (IS2)	0.9990	0.9901	0.9901	0.9	0.9
			Fluorene (IS3)	1.3058	1.2805	1.2805	1.9	1.9
			Hexachlorobenzene (IS4)	0.1947	0.1995	0.1995	2.5	2.5
			Chrysene (IS5)	1.0509	1.0427	1.0427	0.8	0.8
			Benzo(a)pyrene (IS6)	1.1042	1.1500	1.1500	4.1	4.1

Compound (Reference IS)	Concentration (IS/Cpd)	Area Cpd	Area IS
1,4-Dioxane (IS1)	40/80	529970	418891
Naphthalene (IS2)	40/80	3249540	1640967
Fluorene (IS3)	40/80	2525517	986110
Hexachlorobenzene (IS4)	40/80	667483	1672491
Chrysene (IS5)	40/80	3658447	1754242
Benzo(a)pyrene (IS6)	40/80	3746932	1629142

LDC #: 23308 N29
 SDG #: See Cover

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: JV
 2nd reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: $SF/SS * 100$

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: # 2

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5	100	75.7	76	76	0
2-Fluorobiphenyl	↓	77.3	77	77	↓
Terphenyl-d14	↓	88.8	89	89	↓
Phenol-d5	150	120.975	81	81	↓
2-Fluorophenol	↓	117.6	78	78	↓
2,4,6-Tribromophenol	↓	118.5	79	79	↓
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SC/SA)$

Where: SSC = Spike concentration
SA = Spike added

RPD = $100 * (LCS - LCSD) / (LCS + LCSD)$

LCSC = Laboratory control sample concentration
LCSDC = Laboratory control sample duplicate concentration

LCS/LCSD samples: LCS 280 - 13949 / 22-A

Compound	Spike Added (ug/kg)		Spike Concentration (ug/kg)		LCS		LCSD		Percent Recovery		RPD	
	LCS	LCSD	LCS	LCSD	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
Phenol												
N-Nitroso-di-n-propylamine												
4-Chloro-3-methylphenol												
Acenaphthene	2570	NA	1770	NA	70	70						
Pentachlorophenol												
Pyrene	2570		1870		79	74						

Comments: Refer to Laboratory Control Sample/Laboratory Control Sample Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23308 N2A
SDG #: See Cover

VALIDATION FINDINGS WORKSHEET Sample Calculation Verification

Page: 1 of 1
Reviewer: JW
2nd reviewer: JW

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Y N N/A
Y N N/A

Were all reported results recalculated and verified for all level IV samples?

Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

Concentration = $\frac{(A_x)(I_s)(V_i)(DF)(2.0)}{(A_{is})(RRF)(V_o)(V_i)(\%S)}$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_{is} = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- V_o = Volume or weight of sample extract in milliliters (ml) or grams (g).
- V_i = Volume of extract injected in microliters (ul)
- V_t = Volume of the concentrated extract in microliters (ul)
- Df = Dilution Factor.
- %S = Percent solids, applicable to soil and solid matrices only.

- 2.0 = Factor of 2 to account for GPC cleanup

Example:

Sample I.D. _____, _____ ND

Conc. = $\frac{() () () () ()}{() () () () ()}$

=

#	Sample ID	Compound	Reported Concentration ()	Calculated Concentration ()	Qualification

Laboratory Data Consultants, Inc. Data Validation Report

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 28, 2010
LDC Report Date: June 29, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2995-1

Sample Identification

RSAQ3-1BPC
RSAQ3-2BPC
SA169-1BPC
SA169-2BPC
SA110-1BPC
SA110-2BPC
SSAO6-04-3BPC**
SSAO6-04-1BPC
RSAQ3-2BPCMS
RSAQ3-2BPCMSD

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 10 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample EB04282010-RZB (from SDG 280-2995-2) was identified as an equipment blank. No semivolatile contaminants were found in these blanks with the following exceptions:

Equipment Blank ID	Sampling Date	Compound	Concentration	Associated Samples
EB04282010-RZB	4/06/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	RSAQ3-1BPC RSAQ3-2BPC SA169-1BPC SA169-2BPC SA110-1BPC SA110-2BPC

Sample concentrations were compared to concentrations detected in the equipment blanks as required by the QAPP. No sample data was qualified.

Samples FB04062010-RZB (from SDG 280-2131-2) and FB-04072010-RZC (from SDG 280-2280-2) were identified as field blanks. No semivolatile contaminants were found in these blanks with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04062010-RZB	4/06/10	Bis(2-ethylhexyl)phthalate	2.7 ug/L	RSAQ3-1BPC RSAQ3-2BPC SA169-1BPC SA169-2BPC SA110-1BPC SA110-2BPC

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

*VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits with the following exceptions:

Sample	Surrogate	%R (Limits)	Compound	Flag	A or P
SSAO6-04-1BPC	Nitrobenzene-d5 Terphenyl-d14	40 (50-120) 44 (55-120)	All TCL compounds	J- (all detects) UJ (all non-detects)	P

*Removed Phenol-d5 and 2-Fluorophenol information since no reported compounds are associated to these surrogate.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
RSAQ3-1BPC RSAQ3-2BPC SA169-1BPC SA169-2BPC SA110-1BPC SA110-2BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-1	All compounds reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2995-1

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2995-1	SSAO6-04-1BPC	All TCL compounds	J- (all detects) UJ (all non-detects)	P	Surrogate spikes (%R) (s)
280-2995-1	RSAQ3-1BPC RSAQ3-2BPC SA169-1BPC SA169-2BPC SA110-1BPC SA110-2BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-2995-1	RSAQ3-1BPC RSAQ3-2BPC SA169-1BPC SA169-2BPC SA110-1BPC SA110-2BPC SSAO6-04-3BPC** SSAO6-04-1BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2995-1

No Sample Data Qualified in this SDG

Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2995-1

No Sample Data Qualified in this SDG

Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2995-1

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308P2a
 SDG #: 280-2995-1
 Laboratory: Test America

Stage 2B / 4

Date: 6/15/10
 Page: 1 of 1
 Reviewer: JVC
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/28/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD ✓
IV.	Continuing calibration/ICV	A	CCV/ICV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	SW	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	NA	
XII.	Compound quantitation/CRQLs	SW	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	NA	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB 04062010 - RZB (280-2191-2) ↓ = *FB-04072010 - RZC (280-2280-2) EB = FB 04282010 - RZB (280-2995-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

*ND = No compounds detected D = Duplicate
 R = Rinsate TB = Trip blank
 FB = Field blank EB = Equipment blank

Validated Samples:
 ** Stage 4 Soil

1	RSAQ3-1BPC	11	MB 280-13357/1-A	21		31	
2	RSAQ3-2BPC	12		22		32	
3	SA169-1BPC	13		23		33	
4	SA169-2BPC	14		24		34	
5	SA110-1BPC	15		25		35	
6	SA110-2BPC	16		26		36	
7	SSAO6-04-3BPC **	17		27		37	
8	SSAO6-04-1BPC	18		28		38	
9	RSAQ3-2BPCMS	19		29		39	
10	RSAQ3-2BPCMSD	20		30		40	

LDC #: 23308 P2A
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: NT
 2nd Reviewer: Q

Method: Semivolatiles (EPA SW 846 Method 8270C)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical Holding Times				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Cooler temperature criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
II. DFTPP Instrument Performance Check				
Were the DFTPP performance results reviewed and found to be within the specified criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all samples analyzed within the 12 hour clock criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
III. Initial Calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a curve fit used for evaluation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did the initial calibration meet the curve fit acceptance criteria of > 0.990?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) ≤ 30% and relative response factors (RRF) > 0.05?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IV. Continuing Calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) ≤ 25% and relative response factors (RRF) ≥ 0.05?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
V. Blanks				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a method blank analyzed for each matrix and concentration?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were all surrogate %R within QC limits?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
If 2 or more base neutral or acid surrogates were outside QC limits, was a reanalysis performed to confirm %R?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
If any %R was less than 10 percent, was a reanalysis performed to confirm %R?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
VI. Matrix Spike / Duplicate				
Were a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD. Soil / Water.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a MS/MSD analyzed every 20 samples of each matrix?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
VII. LCS / QA Samples				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

LDC #: 23308 P29
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: JR
 2nd Reviewer: C

Validation Area	Yes	No	NA	Findings/Comments
Was an LCS analyzed per extraction batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
X Regional Quality Assurance (RQA) - Laboratory Control				
Were performance evaluation (PE) samples performed?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
Were the performance evaluation (PE) samples within the acceptance limits?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	
X Internal Standard				
Were internal standard area counts within -50% or +100% of the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were retention times within + 30 seconds from the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
X Large compound identification				
Were relative retention times (RRT's) within + 0.06 RRT units of the standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were chromatogram peaks verified and accounted for?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XI Compound Quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XII Qualitative Identification - MS/MS TIC/2D				
Were the major ions (> 10 percent relative intensity) in the reference spectrum evaluated in sample spectrum?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Were relative intensities of the major ions within $\pm 20\%$ between the sample and the reference spectra?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
System performance was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Field duplicate pairs were identified in this SDG.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field duplicates.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
XIII Field Blanks				
Field blanks were identified in this SDG.	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field blanks.	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (*)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenyl ether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF. ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET
 Field Blanks

LDC #: 23308 P29
 SDG #: S4 C29

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Were field blanks identified in this SDG? Y/N N/A

Were target compounds detected in the field blanks? Y/N N/A

Blank units: w/ l Associated sample units: w/kg

Sampling date: 4/26/10 4/28/10

Field blank type: (circle one) Field Blank/Rinsate/Other:

Associated Samples: 1-6

Compound	Blank ID	Blank ID	(4/28)	Sample Identification
	FB	EP		
EEF	2.7	2.2		(either ND or > 5X FB/EP)
FB = FB 04062010 - RZB				
EP = EP 04282010 - RZB				
CRQL				

Blank units: Associated sample units:

Sampling date: Associated Samples:

Compound	Blank ID	Sample Identification
CRQL		

5x Phthalates
 2x All others

VALIDATION FINDINGS WORKSHEET
Surrogate Recovery

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)
 Please see qualification below for all questions answered "N". Not applicable questions are identified as "N/A".
 Were percent recoveries (%R) for surrogates within QC limits?
 If 2 or more base neutral or acid surrogates were outside QC limits, was a reanalysis performed to confirm %R?
 If any %R was less than 10 percent, was a reanalysis performed to confirm %R?

#	Date	Sample ID	Surrogate	%R (Limits)	Qualifications
		8 (5x)	PHL	38 (52-120)	J- / NJ / P (AU TOL) (S)
			2FP	37 (53-120)	
			NBZ	40 (50-120)	
			TPH	44 (55-120)	

* QC limits are advisory
 S1 (NBZ) = Nitrobenzene-d5
 S2 (FBP) = 2-Fluorobiphenyl
 S3 (TPH) = Terphenyl-d14
 S4 (PHL) = Phenol-d5

QC Limits (Soil)
 23-120
 30-115
 18-137
 24-113

QC Limits (Water)
 35-114
 43-116
 33-141
 10-94

S5 (2FP) = 2-Fluorophenol
 S6 (TBP) = 2,4,6-Tribromophenol
 S7 (2CP) = 2-Chlorophenol-d4
 S8 (DCB) = 1,2-Dichlorobenzene-d4

QC Limits (Soil)
 25-121
 19-122
 20-130*

QC Limits (Water)
 21-100
 10-123
 33-110*
 16-110*

VALIDATION FINDINGS WORKSHEET
Compound Quantitation and Reported CRQLs

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?
 Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		1, 2, 3, 4, 5, 6	GGG, HHH unresolved		J/MJ/P (g)

Comments: See sample calculation verification worksheet for recalculations

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$$RRF = (A_x)(C_{is}) / (A_{is})(C_x)$$

$$\text{average RRF} = \text{sum of the RRFs} / \text{number of standards}$$

$$\%RSD = 100 * (SIX)$$

#	Standard ID	Calibration Date	Compound (Internal Standard)	Reported	Recalculated	Reported	Recalculated	Reported	Recalculated
				RRF (50 std)	RRF (50 std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD
1	ICAL	4/20/2010	1,4-Dioxane (IS1)	0.6731	0.6731	0.6818	0.6818	5.4	5.44
	MSS D		Naphthalene (IS2)	1.1079	1.1079	1.1204	1.1204	4.7	4.70
			Fluorene (IS3)	1.3779	1.3779	1.3629	1.3629	8.9	8.89
			Hexachlorobenzene (IS4)	0.2590	0.2590	0.2705	0.2705	14.0	13.97
			Chrysene (IS5)	1.0611	1.0611	1.0324	1.0324	4.3	4.35
			Benzo(a)pyrene (IS6)	1.1960	1.1960	1.1835	1.1835	13.5	13.49

Inc IS/Cpd	Area cpd	Area IS
40/50	220464	262046
40/50	1381644	997667
40/50	1155733	671030
40/50	394826	1219394
40/50	2008107	1513952
40/50	1958223	1309806

Conc	1,4-Dioxane	Naphthalene	Fluorene	Hexachlorob	Chrysene	Benzo(a)py
4.00	0.6984	1.0908	1.2935		1.0330	0.9394
10.00	0.7499	1.0730	1.1667	0.2303	0.9982	1.0100
20.00	0.6512	1.0585	1.2453	0.2289	1.0104	1.0839
50.00	0.6731	1.1079	1.3779	0.2590	1.0611	1.1960
80.00	0.6228	1.1000	1.3843	0.2562	1.0602	1.2099
120.00	0.6766	1.1473	1.4242	0.2854	1.0752	1.3098
160.00	0.6887	1.1741	1.4888	0.3029	1.0741	1.3626
200.00	0.6937	1.2114	1.5224	0.3306	0.9470	1.3565
X =	0.6818	1.1204	1.3629	0.2705	1.0324	1.1835
S =	0.0371	0.0527	0.1212	0.0378	0.0449	0.1597

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORSHEET
Continuing Calibration Results Verification

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

Where:
 ave. RRF = initial calibration average RRF
 RRF = continuing calibration RRF
 Ax = Area of compound
 Cx = Concentration of compound
 Ais = Area of associated internal standard
 Cis = Concentration of internal standard

$$\% \text{ Difference} = 100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$$

$$\text{RRF} = (\text{Ax}) / (\text{Cis}) / (\text{Ais}) / (\text{Cx})$$

#	Standard ID	Calibration Date	Compound (Reference IS)	Average RRF (Initial RRF)	Reported (CC RRF)	Recalculated (CC RRF)	Reported %D	Recalculated %D
1	D4531	05/01/10	1,4-Dioxane (IS1)	0.6818	0.6135	0.6135	10.0	10.0
			Naphthalene (IS2)	1.1204	1.1479	1.1479	2.5	2.5
			Fluorene (IS3)	1.3629	1.4115	1.4115	3.6	3.6
			Hexachlorobenzene (IS4)	0.2705	0.2804	0.2804	3.7	3.7
			Chrysene (IS5)	1.0324	1.0668	1.0668	3.3	3.3
			Benzo(a)pyrene (IS6)	1.1835	1.2509	1.2509	5.7	5.7

Compound (Reference IS)	Concentration (IS/Cpd)	Area Cpd	Area IS
1,4-Dioxane (IS1)	40/80	384024	312973
Naphthalene (IS2)	40/80	2646759	1152826
Fluorene (IS3)	40/80	2305108	816564
Hexachlorobenzene (IS4)	40/80	809868	1444254
Chrysene (IS5)	40/80	3940883	1847115
Benzo(a)pyrene (IS6)	40/80	3862774	1543947

LDC #: 23308 p2a
 SDG #: See Cover

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: *[Signature]*
 2nd reviewer: *[Signature]*

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: $SF/SS * 100$

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: # 7 (5x)

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5	100	80	80	80	0
2-Fluorobiphenyl		80.5	81	81	
Terphenyl-d14		89.5	90	90	
Phenol-d5	150	119	79	79	
2-Fluorophenol		107.5	72	72	
2,4,6-Tribromophenol		122	88	88	
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates Results Verification

LDC #: 223 308 P2A
 SDG #: Sec Cover

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 \cdot (SSC - SC) / SA$ Where: SSC = Spiked sample concentration
 SA = Spike added SC = Sample concentration

RPD = $100 \cdot MSC - MSC / 2(MSC + MSDC)$ MSC = Matrix spike concentration
 MSDC = Matrix spike duplicate concentration

MS/MSD samples: 9/10

Compound	Spike Added (ug/kg)		Sample Concentration (ug/kg)	Spiked Sample Concentration (ug/kg)		Matrix Spike Percent Recovery		Matrix Spike Duplicate Percent Recovery		RPD	
	MS	MSD		MS	MSD	Reported	Recalc	Reported	Recalc	Reported	Recalculated
Phenol											
N-Nitroso-di-n-propylamine											
4-Chloro-3-methylphenol	2630	2650	6	1780	2050	68	68	77	77	14	14
Acenaphthene											
Pentachlorophenol	2630	2650	60	2100	2480	77	77	92	12	17	17
Pyrene											

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SC/SA)$ Where: SSC = Spike concentration
 SA = Spike added

RPD = $100 * (LCS - LCSDC) / (LCS + LCSDC)$ LCSC = Laboratory control sample concentration LCSDC = Laboratory control sample duplicate concentration

LCS/LCSD samples: LCS 280 - 13357/2-A

Compound	Spike Added (ng/kg)		Spike Concentration (ng/L)		LCS Percent Recovery		LCSD Percent Recovery		LCS/LCSD RPD	
	LCS	LCSD	LCS	LCSD	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
Phenol										
N-Nitroso-di-n-propylamine										
4-Chloro-3-methylphenol	2560	NA	2140	NA	84	84				
Acenaphthene										
Pentachlorophenol										
Pyrene	2560	NA	2570	NA	99	99				

Comments: Refer to Laboratory Control Sample/Laboratory Control Sample Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 28, 2010

LDC Report Date: June 16, 2010

Matrix: Water

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2995-2

Sample Identification

EB04282010-RZB

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-13927/1-A	5/4/10	Bis(2-ethylhexyl)phthalate	1.88 ug/L	All samples in SDG 280-2995-2

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
EB04282010-RZB	Bis(2-ethylhexyl)phthalate	2.2 ug/L	2.2U ug/L

Sample EB04282010-RZB was identified as an equipment blank. No semivolatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Sampling Date	Compound	Concentration	Associated Samples
EB04282010-RZB	4/28/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	No associated samples in this SDG

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-2	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2995-2**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2995-2	EB04282010-RZB	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2995-2**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-2995-2	EB04282010-RZB	Bis(2-ethylhexyl)phthalate	2.2U ug/L	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-2995-2**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308Q2a
 SDG #: 280-2995-2
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JVG
 2nd Reviewer: Q

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/28/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD ✓
IV.	Continuing calibration/ICV	A	CV/1CV ≤ 25 %
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	client spec
VIII.	Laboratory control samples	A	LCS/D
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	EB = 1

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet
 ND = No compounds detected
 R = Rinstate
 FB = Field blank
 D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: Water

1	EB0428/010-RZB	11		21		31	
2	MB 280-13927/A-A	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 28, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2995-5

Sample Identification

SSAN6-07-1BPC**
SSAN6-07-2BPC
SSAN6-07-1BPCMS
SSAN6-07-1BPCMSD

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 4 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-14738/9-A	5/10/10	Bis(2-ethylhexyl)phthalate	72.7 ug/Kg	All samples in SDG 280-2995-5

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
SSAN6-07-1BPC**	Bis(2-ethylhexyl)phthalate	98 ug/Kg	98U ug/Kg
SSAN6-07-2BPC	Bis(2-ethylhexyl)phthalate	120 ug/Kg	120U ug/Kg

Sample FB-04072010-RZC (from SDG 280-2280-2) was identified as a field blank. No semivolatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-5	All compounds reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable for samples on which a Stage 4 review was performed. Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-2995-5**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2995-5	SSAN6-07-1BPC** SSAN6-07-2BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-2995-5**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-2995-5	SSAN6-07-1BPC**	Bis(2-ethylhexyl)phthalate	98U ug/Kg	A	bl
280-2995-5	SSAN6-07-2BPC	Bis(2-ethylhexyl)phthalate	120U ug/Kg	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-2995-5**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308R2a
 SDG #: 280-2995-5
 Laboratory: Test America

Stage 2B / 4

Date: 6/15/10
 Page: 1 of 1
 Reviewer: OVG
 2nd Reviewer: 9

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/28/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	
IV.	Continuing calibration/ICV	A	CV/ICV ≤ 25?
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	* A	
XII.	Compound quantitation/CRQLs	* A	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	* A	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	FB = FB-04072010-RZC (280-2280-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

* * Stage 4 Soil

1	SSAN6-07-1BPC * * 11	MB 280-14738/9-A	21	31
2	SSAN6-07-2BPC		22	32
3	SSAN6-07-1BPCMS		23	33
4	SSAN6-07-1BPCMSD		24	34
5			25	35
6			26	36
7			27	37
8			28	38
9			29	39
10			30	40

LDC #: 23308 R2c
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: JV
 2nd Reviewer: Q

Method: Semivolatiles (EPA SW 846 Method 8270C)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Cooler temperature criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
II. GC/MS Instrument performance check				
Were the DFTPP performance results reviewed and found to be within the specified criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all samples analyzed within the 12 hour clock criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a curve fit used for evaluation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did the initial calibration meet the curve fit acceptance criteria of > 0.990 ?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) $\leq 30\%$ and relative response factors (RRF) > 0.05 ?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) and relative response factors (RRF) within method criteria for all CCCs and SPCCs?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) $\leq 25\%$ and relative response factors (RRF) ≥ 0.05 ?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
V. Blanks				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a method blank analyzed for each matrix and concentration?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
If 2 or more base neutral or acid surrogates were outside QC limits, was a reanalysis performed to confirm %R?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
If any %R was less than 10 percent, was a reanalysis performed to confirm %R?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
VII. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD. Soil / Water.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a MS/MSD analyzed every 20 samples of each matrix?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

LDC #: 23708 R29
 SDG #: See Cover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: JV
 2nd Reviewer: Q

Validation Area	Yes	No	NA	Findings/Comments
Was an LCS analyzed per extraction batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the performance evaluation (PE) samples within the acceptance limits?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
X. Internal standards				
Were internal standard area counts within -50% or +100% of the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were retention times within + 30 seconds from the associated calibration standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XI. Target compound identification				
Were relative retention times (RRT's) within + 0.06 RRT units of the standard?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were chromatogram peaks verified and accounted for?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 10 percent relative intensity) in the reference spectrum evaluated in sample spectrum?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Were relative intensities of the major ions within \pm 20% between the sample and the reference spectra?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
XIV. System performance				
System performance was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field duplicates.	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
XVII. Field blanks				
Field blanks were identified in this SDG.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Target compounds were detected in the field blanks.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

Blanks

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Was a method blank analyzed for each matrix?
- Y N N/A Was a method blank analyzed for each concentration preparation level?
- Y N N/A Was a method blank associated with every sample?
- Y N N/A Was the blank contaminated? If yes, please see qualification below.

Blank extraction date: 5/10/10 Blank analysis date: 5/12/10

Conc. units: ug/kg Associated Samples: All (b2)

Compound	Blank ID	Sample Identification
	MB 260-14738/4-A	✓
EEH	72.7	98/4 120/4

Blank extraction date: _____ Blank analysis date: _____
 Conc. units: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_{is}) / (A_{is})(C_x)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$
 A_x = Area of Compound
 A_{is} = Area of associated internal standard
 C_x = Concentration of compound,
 C_{is} = Concentration of internal standard
 S = Standard deviation of the RRFs,
 X = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Internal Standard)	Reported	Recalculated	Reported	Recalculated	Reported	Recalculated
				RRF (50 std)	RRF (50 std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD
1	ICAL	5/8/2010	1,4-Dioxane (IS1)	0.5851	0.5851	0.5791	0.5791	6.8	6.82
	MSS D		Naphthalene (IS2)	1.0818	1.0818	1.0917	1.0917	2.8	2.83
			Fluorene (IS3)	1.3573	1.3573	1.3205	1.3205	6.7	6.66
			Hexachlorobenzene (IS4)	0.2608	0.2608	0.2633	0.2633	9.5	9.48
			Chrysene (IS5)	1.0472	1.0472	1.0309	1.0309	3.1	3.14
			Benzo(a)pyrene (IS6)	1.0983	1.0983	1.0834	1.0834	13.3	13.28

Inc IS/Cpd	Area cpd	Area IS
40/50	223629	305776
40/50	1535127	1135207
40/50	1281626	755377
40/50	450249	1381076
40/50	2196814	1678262
40/50	2123114	1546473

Conc	1,4-Dioxane	Naphthalene	Fluorene	Hexachlorob	Chrysene	Benzo(a)py
4.00	0.5215	1.1002	1.1947		1.0331	0.8317
10.00	0.6382	1.0477	1.2120	0.2263	1.0128	0.9476
20.00	0.6300	1.0478	1.2639	0.2397	1.0223	1.0101
50.00	0.5851	1.0818	1.3573	0.2608	1.0472	1.0983
80.00	0.5842	1.0953	1.3253	0.2600	1.0539	1.1259
120.00	0.5666	1.1092	1.3832	0.2754	1.0748	1.1887
160.00	0.5529	1.1267	1.3955	0.2815	1.0367	1.2251
200.00	0.5544	1.1249	1.4319	0.2996	0.9663	1.2400
X =	0.5791	1.0917	1.3205	0.2633	1.0309	1.0834
S =	0.0395	0.0309	0.0879	0.0250	0.0324	0.1439

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

LDC # 23308 R2a
 SDG # See Cover

METHOD: GC/MS SVOA (EPA SW 846 Method 8270C)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRF-s) and the continuing calibration RRF-s were recalculated for the compounds identified below using the following calculation:

Where:
 ave. RRF = initial calibration average RRF
 RRF = continuing calibration RRF
 Ax = Area of compound Ais = Area of associated internal standard
 Cx = Concentration of compound Cis = Concentration of internal standard

$$\% \text{ Difference} = 100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$$

$$\text{RRF} = (Ax) / (Cis) / (Ais) / (Cx)$$

#	Standard ID	Calibration Date	Compound (Reference IS)	Average RRF (Initial RRF)	Reported (CC RRF)	Recalculated (CC RRF)	Reported %D	Recalculated %D
1	D4891	05/12/10	1,4-Dioxane (IS1)	0.5791	0.5249	0.5249	9.4	9.4
			Naphthalene (IS2)	1.0917	1.0945	1.0945	0.3	0.3
			Fluorene (IS3)	1.3205	1.3456	1.3456	1.9	1.9
			Hexachlorobenzene (IS4)	0.2633	0.2713	0.2713	3.0	3.1
			Chrysene (IS5)	1.0309	1.0161	1.0161	1.4	1.4
			Benzo(a)pyrene (IS6)	1.0834	1.1127	1.1127	2.7	2.7

Compound (Reference IS)	Concentration (IS/Cpd)	Area Cpd	Area IS
1,4-Dioxane (IS1)	40/80	383501	365302
Naphthalene (IS2)	40/80	2883187	1317169
Fluorene (IS3)	40/80	2383317	885592
Hexachlorobenzene (IS4)	40/80	850720	1567635
Chrysene (IS5)	40/80	4037721	1986787
Benzo(a)pyrene (IS6)	40/80	4197272	1886084

LDC #: 73308 R 2a
 SDG #: Sre Cover

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: JVB
 2nd reviewer: ↓

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: $SF/SS * 100$

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: # 1

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5	100	72.6	73	73	0
2-Fluorobiphenyl	↓	77.3	77	77	↓
Terphenyl-d14	↓	99.2	99	99	↓
Phenol-d5	150	121.3	81	81	↓
2-Fluorophenol	↓	119.4	80	80	↓
2,4,6-Tribromophenol	↓	121.7	81	81	↓
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery Reported	Percent Recovery Recalculated	Percent Difference
Nitrobenzene-d5					
2-Fluorobiphenyl					
Terphenyl-d14					
Phenol-d5					
2-Fluorophenol					
2,4,6-Tribromophenol					
2-Chlorophenol-d4					
1,2-Dichlorobenzene-d4					

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * ((SC - SC)/SA)$ Where: SSC = Spiked sample concentration
 SA = Spike added
 SC = Sample concentration
 RPD = $100 * |MSC - MSC| * 2 / (MSC + MSC)$ MSC = Matrix spike concentration
 MSC = Matrix spike duplicate concentration
 MSDC = Matrix spike duplicate concentration

MS/MSD samples: 3/4

Compound	Spike Added (ug/kg)		Sample Concentration (ug/kg)		Spiked Sample Concentration (ug/kg)		Matrix Spike Percent Recovery		Matrix Spike Duplicate Percent Recovery		MS/MSD RPD	
	MS	MSD	MS	MSD	MS	MSD	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
Phenol												
N-Nitroso-di-n-propylamine												
4-Chloro-3-methylphenol												
Acenaphthene	2860	2860	0		2290	2220	80	80	78	78	3	3
Pentachlorophenol	2860	2860	↓		2490	2460	87	87	86	86	1	1
Pyrene												

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample/Laboratory Control Sample Duplicates Results Verification

LDC #: 23308 R 29
 SDG #: See Cover

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SC/SA)$ Where: SSC = Spike concentration
 SA = Spike added

RPD = $100 * (LCS - LCSDC) / (LCS + LCSDC)$ LCSC = Laboratory control sample concentration LCSDC = Laboratory control sample duplicate concentration

LCS/LCSD samples: LCS 280 - 14738/16-A

Compound	Spike Added (ug/kg)		Spike Concentration (ug/kg)		LCS		LCSD		LCS/LCSD											
	LCS	LCSD	LCS	LCSD	Percent Recovery		Percent Recovery		RPD											
	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.										
Phenol																				
N-Nitroso-di-n-propylamine																				
4-Chloro-3-methylphenol																				
Acenaphthene	2570	NA	1990	NA	77	77														
Pentachlorophenol	2570	↓	2260	↓	88	88														
Pyrene																				

Comments: Refer to Laboratory Control Sample/Laboratory Control Sample Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23308 R29
 SDG #: See Cover

VALIDATION FINDINGS WORKSHEET

Sample Calculation Verification

Page: 1 of 1
 Reviewer: JVB
 2nd reviewer: [Signature]

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Y N N/A
 Y N N/A

Were all reported results recalculated and verified for all level IV samples?
 Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(V_i)(DF)(2.0)}{(A_{is})(RRF)(V_o)(V_j)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_{is} = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- V_o = Volume or weight of sample extract in milliliters (ml) or grams (g).
- V_i = Volume of extract injected in microliters (ul)
- V_j = Volume of the concentrated extract in microliters (ul)
- Df = Dilution Factor.
- %S = Percent solids, applicable to soil and solid matrices only.
- 2.0 = Factor of 2 to account for GPC cleanup

Example:

Sample I.D. # 1, SS

$$\text{Conc.} = \frac{(192455)(40)(1.0\text{ml})(1000)}{(395254)(0.2633)(3.50\text{g})(0.92)}$$

$$= 722.3$$

≈ 720 ug/kg

#	Sample ID	Compound	Reported Concentration ()	Calculated Concentration ()	Qualification

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 29, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3059-3

Sample Identification

SSA05-05-2BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZC (from SDG 280-2280-2) was identified as a field blank. No semivolatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
SSAO5-05-2BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3059-3	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3059-3**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3059-3	SSAO5-05-2BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-3059-3	EB04282010-RZB	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3059-3**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3059-3**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308T2a
 SDG #: 280-3059-3
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JVL
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/29/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD ✓
IV.	Continuing calibration/ICV	A	CCV/ICV < 25 %
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	SW	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	RSD	FB = FB-04072010-RZC (280-2 280-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

Soil

1	SSAO5-05-2BPC	11		21		31	
2	MB 280-14337/1-A	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

LDC #: 28308 729
SDG #: 826 GY

VALIDATION FINDINGS WORKSHEET

Compound Quantitation and Reported CRQLs

Page: 1 of 1
Reviewer: JVC
2nd Reviewer: g

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

Y N N/A
Y N N/A

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		1	GGG, HHH un resolved		J/MJ/p (g)

Comments: See sample calculation verification worksheet for recalculations

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 29, 2010

LDC Report Date: June 22, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3059-4

Sample Identification

SSAQ4-04-1BPC

SSAO4-05-1BPC

SSAO4-05-1BPC_FD

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Bis(2-ethylhexyl)phthalate	2.7 ug/L	All samples in SDG 280-3059-4

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
SSAQ4-04-1BPC SSAO4-05-1BPC SSAO4-05-1BPC_FD	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3059-4	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples SSAO4-05-1BPC and SSAO4-05-1BPC_FD were identified as field duplicates. No semivolatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAO4-05-1BPC	SSAO4-05-1BPC_FD				
Benzo(b)fluoranthene	32	38	-	6 (≤ 340)	-	-
Benzo(g,h,i)perylene	22	21	-	1 (≤ 340)	-	-
Hexachlorobenzene	46	64	-	18 (≤ 340)	-	-
Pyrene	13	15	-	2 (≤ 340)	-	-

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3059-4**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3059-4	SSAQ4-04-1BPC SSAO4-05-1BPC SSAO4-05-1BPC_FD	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-3059-4	SSAQ4-04-1BPC SSAO4-05-1BPC SSAO4-05-1BPC_FD	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3059-4**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3059-4**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308U2a
SDG #: 280-3059-4
Laboratory: Test America

Stage 2B

Date: 6/14/10
Page: 1 of 1
Reviewer: [Signature]
2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

Validation Area		Comments	
I.	Technical holding times	A	Sampling dates: <u>4/29/10</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	<u>2 KSD 12</u>
IV.	Continuing calibration/ICV	A	<u>CCV/ICV ≤ 25%</u>
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	<u>Client spec</u>
VIII.	Laboratory control samples	A	<u>LCS</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	<u>D = 2, 3</u>
XVII.	Field blanks	SW	<u>FB = FB0406 2010 - RZB (280-2131-2)</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: 501

1	SSAQ4-04-1BPC	11		21		31
2	⁰ SSAQ4-05-1BPC ^D	12		22		32
3	⁰ SSAQ4-05-1BPC FD ^D	13		23		33
4	<u>MB 280-13949/21A</u>	14		24		34
5		15		25		35
6		16		26		36
7		17		27		37
8		18		28		38
9		19		29		39
10		20		30		40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis(2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes:* = System performance check compound (SPCC) for RRF. ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET
Compound Quantitation and Reported CRQLs

LDC #: 23308 V29
SDG #: Sea Coney

Page: 1 of 1
Reviewer: NC
2nd Reviewer: G

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
Y N N/A Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?
Y N N/A Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		1, 2, 3	GGG, HHH unresolved		J/WJ/p (g)

Comments: See sample calculation verification worksheet for recalculations

LDC#: 23308U2a
SDG#: See cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: NG
2nd Reviewer: Y

METHOD: GC/MS PAH (EPA SW 846 Method 8270C)

Y / N / NA Were field duplicate pairs identified in this SDG?

Y / N / NA Were target analytes detected in the field duplicate pairs?

Compound Name	Conc (ug/Kg)		RPD (≤50%)	Diff	Diff Limits	Quals (Parent Only)
	2	3				
Benzo(b)fluoranthene	32	38		6	≤340	
Benzo(g,h,i)perylene	22	21		1	≤340	
Hexachlorobenzene	46	64		18	≤340	
Pyrene	13	15		2	≤340	

V:\FIELD DUPLICATES\23308U2a.wpd

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 29, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-3059-6

Sample Identification

SSAO4-03-3BPC
SSAO4-03-3BPCMS
SSAO4-03-3BPCMSD

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB-04072010-RZC (from SDG 280-2280-2) was identified as a field blank. No semivolatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3059-6	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3059-6**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3059-6	SSAO4-03-3BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3059-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3059-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308V2a
 SDG #: 280-3059-6
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/29/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD ✓
IV.	Continuing calibration/ICV	A	COV/AV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	#
XVII.	Field blanks	ND	FB = FB-04072010 - RZC (280-2280-2)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

Soil

1	SSAO4-03-3BPC	11		21		31	
2	SSAO4-03-3BPCMS	12		22		32	
3	SSAO4-03-3BPCMSD	13		23		33	
4	MB 280-14337/A	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 29, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3059-8

Sample Identification

SSAQ3-01-6BPC

SSAQ3-01-6BPCMS

SSAQ3-01-6BPCMSD

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Bis(2-ethylhexyl)phthalate	2.7 ug/L	All samples in SDG 280-3059-8

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
SSAQ3-01-6BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3059-8	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3059-8**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3059-8	SSAQ3-01-6BPC	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-3059-8	SSAQ3-01-6BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3059-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3059-8**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308W2a
SDG #: 280-3059-8
Laboratory: Test America

Stage 2B

Date: 6/11/10
Page: 1 of 1
Reviewer: JVG
2nd Reviewer: A

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/29/10</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	<u>2 RSD</u>
IV.	Continuing calibration/ICV	A	<u>CV/ICV ≤ 25%</u>
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	<u>LCS</u>
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	SW	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	<u>FB = FBD486 2010- RZB (280-2131-γ)</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	SSAQ3-01-6BPC	11		21		31	
2	SSAQ3-01-6BPCMS	12		22		32	
3	SSAQ3-01-6BPCMSD	13		23		33	
4	<u>MB 280-14919 / 4-A</u>	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET
Compound Quantitation and Reported CRQLs

LDC #: 23208 W 79
 SDG #: 52 Con

Page: 1 of 1
 Reviewer: JG
 2nd Reviewer: C

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 Y N N/A Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?
 Y N N/A Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		<u>1</u>	<u>GGG, HHH un resolved</u>		<u>J/W/P (g)</u>

Comments: See sample calculation verification worksheet for recalculations

Laboratory Data Consultants, Inc. Data Validation Report

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 30, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3100-1

Sample Identification

SSAQ5-01-9BPC
SSAQ5-01-7BPC
SSAQ5-01-5BPC
SSAQ5-01-3BPC
SSAQ5-01-1BPC
SSAQ5-01-1BPC-FD
SSAQ5-01-5BPCMS
SSAQ5-01-5BPCMSD

Introduction

This data review covers 8 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Bis(2-ethylhexyl)phthalate	2.7 ug/L	All samples in SDG 280-3100-1

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compound quantitation and CRQLs were within validation criteria with the following exceptions:

Sample	Compound	Finding	Criteria	Flag	A or P
SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	Benzo(b)fluoranthene Benzo(k)fluoranthene	Due to lack of resolution between these compounds in the samples, the laboratory performed the quantitation using the total peak area.	Target compounds must be properly resolved and quantitated as individual compounds.	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3100-1	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples SSAQ5-01-1BPC and SSAQ5-01-1BPC-FD were identified as field duplicates. No semivolatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAQ5-01-1BPC	SSAQ5-01-1BPC-FD				
Benzo(a)anthracene	22	340U	-	318 (\leq 340)	-	-
Benzo(b)fluoranthene	51	27	-	24 (\leq 360)	-	-
Benzo(g,h,i)perylene	20	340U	-	320 (\leq 340)	-	-
Chrysene	32	340U	-	308 (\leq 340)	-	-

Compound	Concentration (ug/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAQ5-01-1BPC	SSAQ5-01-1BPC-FD				
Hexachlorobenzene	3300	2200	40 (≤ 50)	-	-	-
Octachlorostyrene	660	460	-	200 (≤ 360)	-	-
Pyrene	25	18	-	7 (≤ 360)	-	-

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3100-1**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3100-1	SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	Benzo(b)fluoranthene Benzo(k)fluoranthene	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Project Quantitation Limit (q)
280-3100-1	SSAQ5-01-9BPC SSAQ5-01-7BPC SSAQ5-01-5BPC SSAQ5-01-3BPC SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308X2a
 SDG #: 280-3100-1
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/30/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD r _r
IV.	Continuing calibration/ICV	A	CV/ICV ≤ 25%
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	SW	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	d = 5, 6
XVII.	Field blanks	SW	FB = FB04062010 - R2B (280-2131-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAQ5-01-9BPC	11	MB 280-13952/7A	21	31
2	SSAQ5-01-7BPC	12		22	32
3	SSAQ5-01-5BPC	13		23	33
4	SSAQ5-01-3BPC	14		24	34
5	SSAQ5-01-1BPC <u>d</u>	15		25	35
6	SSAQ5-01-1BPC-FD <u>d</u>	16		26	36
7	SSAQ5-01-5BPCMS	17		27	37
8	SSAQ5-01-5BPCMSD	18		28	38
9		19		29	39
10		20		30	40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(e)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-dl-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N / N/A Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?

Y N / N/A Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?

#	Date	Sample ID	Finding	Associated Samples	Qualifications
		5, 6	GGG, HHH peaks unresolved		J/US/P (g)

Comments: See sample calculation verification worksheet for recalculations

LDC#: 23308X2a
SDG#: See cover

VALIDATION FINDINGS WORKSHEET

Field Duplicates

Page: 1 of 1
Reviewer: JVG
2nd Reviewer:

METHOD: GC/MS PAH (EPA SW 846 Method 8270C)

Y N NA Were field duplicate pairs identified in this SDG?

Y N NA Were target analytes detected in the field duplicate pairs?

Compound Name	Conc (ug/Kg)		RPD (≤50%)	Diff	Diff Limits	Quals (Parent Only)
	5	6				
Benzo(a)anthracene	22	340U		318	≤340	
Benzo(b)fluoranthene	51	27		24	≤360	
Benzo(g,h,i)perylene	20	340U		320	≤340	
Chrysene	32	340U		308	≤340	
Hexachlorobenzene	3300	2200	40			
Octachlorostyrene	660	460		200	≤360	
Pyrene	25	18		7	≤360	

V:\FIELD DUPLICATES\23308X2a.wpd

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 30, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Semivolatiles
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-3100-4

Sample Identification

SSAK7-05-1BPC
SSAK7-05-1BPC_FD

Introduction

This data review covers 2 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-14276/1-A	5/6/10	Bis(2-ethylhexyl)phthalate	60.1 ug/Kg	All samples in SDG 280-3100-4

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
SSAK7-05-1BPC	Bis(2-ethylhexyl)phthalate	94 ug/Kg	94U ug/Kg
SSAK7-05-1BPC_FD	Bis(2-ethylhexyl)phthalate	86 ug/Kg	86U ug/Kg

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-3100-4

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3100-4	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples SSAK7-05-1BPC and SSAK7-05-1BPC_FD were identified as field duplicates. No semivolatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAK7-05-1BPC	SSAK7-05-1BPC_FD				
Bis(2-ethylhexyl)phthalate	94	86	-	8 (≤ 350)	-	-
Dimethyl phthalate	40	350U	-	310 (≤ 350)	-	-

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3100-4**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3100-4	SSAK7-05-1BPC SSAK7-05-1BPC_FD	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3100-4**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-3100-4	SSAK7-05-1BPC	Bis(2-ethylhexyl)phthalate	94U ug/Kg	A	bl
280-3100-4	SSAK7-05-1BPC_FD	Bis(2-ethylhexyl)phthalate	86U ug/Kg	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3100-4**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308Y2a
 SDG #: 280-3100-4
 Laboratory: Test America

Stage 2B

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JUC
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 6/30/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	? RSD r/r
IV.	Continuing calibration/ICV	A	CCV/ICV = 25%
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 1, 2
XVII.	Field blanks	SW	FB = FB-07072010-RZD (280-2216-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAK7-05-1BPC	D	11	21	31
2	SSAK7-05-1BPC_FD	D	12	22	32
3 ⁺	MB 280-14276/1-A		13	23	33
4			14	24	34
5			15	25	35
6			16	26	36
7			17	27	37
8			18	28	38
9			19	29	39
10			20	30	40

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF. ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

Blanks

LDC #: 23308 724

SDG #: 5a

Page: 1 of 1

Reviewer: Ne

2nd Reviewer: Q

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A

Y N N/A

Y N N/A

Y N N/A

Was the blank contaminated? If yes, please see qualification below.

Blank extraction date: 5/15/02

Blank analysis date: 5/15/02

Conc. units: ug/kg Associated Samples: A11

(62)

Compound	Blank ID	Sample Identification										
	Mb 280-1427C/A											
	60.1							1		2		
EEE								94		86		

Blank extraction date: _____

Blank analysis date: _____

Conc. units: _____



Associated Samples: _____

Compound	Blank ID	Sample Identification										

5x Phthalates
2x all others

LDC#: 23308Y2a
SDG#: See cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: 
2nd Reviewer: 

METHOD: GC/MS PAH (EPA SW 846 Method 8270C)

Y N NA Were field duplicate pairs identified in this SDG?

Y N NA Were target analytes detected in the field duplicate pairs?

Compound Name	Conc (ug/Kg)		RPD (≤ 50%)	Diff	Diff Limits	Quals (Parent Only)
	1	2				
bis(2-ethylhexyl)phthalate	94	86		8	≤350	
Dimethyl phthalate	40	350U		310	≤350	

V:\FIELD DUPLICATES\23308Y2a.wpd

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 30, 2010

LDC Report Date: June 16, 2010

Matrix: Soil

Parameters: Semivolatiles

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3100-6

Sample Identification

SSAK6-03-3BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8270C for Semivolatiles.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 30.0% for all compounds.

In the case where the laboratory used a calibration curve to evaluate the compounds, all coefficients of determination (r^2) were greater than or equal to 0.990 .

Average relative response factors (RRF) for all compounds were within method and validation criteria.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

Percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were within the method criteria of less than or equal to 20.0% for calibration check compounds (CCCs) and 25.0% for all other compounds.

The percent differences (%D) of the second source calibration standard were less than or equal to 25.0% for all compounds.

All of the continuing calibration relative response factors (RRF) were within method and validation criteria.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No semivolatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Extraction Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
MB280-14475/1-A	5/7/10	Bis(2-ethylhexyl)phthalate	62.7 ug/Kg	All samples in SDG 280-3100-6

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
SSAK6-03-3BPC	Bis(2-ethylhexyl)phthalate	83 ug/Kg	83U ug/Kg

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No semivolatile contaminants were found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Compound	Concentration	Associated Samples
FB-04072010-RZD	4/7/10	Bis(2-ethylhexyl)phthalate	2.2 ug/L	All samples in SDG 280-3100-6

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3100-6	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Data Qualification Summary - SDG 280-3100-6**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-3100-6	SSAK6-03-3BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Laboratory Blank Data Qualification Summary - SDG 280-3100-6**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P	Code
280-3100-6	SSAK6-03-3BPC	Bis(2-ethylhexyl)phthalate	83U ug/Kg	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Equipment Blank Data Qualification Summary - SDG 280-3100-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Semivolatiles - Field Blank Data Qualification Summary - SDG 280-3100-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

Stage 2B

LDC #: 23308Z2a
 SDG #: 280-3100-6
 Laboratory: Test America

Date: 6/14/10
 Page: 1 of 1
 Reviewer: JVG
 2nd Reviewer: [Signature]

METHOD: GC/MS Semivolatiles (EPA SW 846 Method 8270C)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/30/10
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	2 RSD r ²
IV.	Continuing calibration/ICV	A	CV/AV = 25%
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	N	Client spec
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	FB = FB-04072010-RZD (280-2216-?)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: [soi]

1	SSAK6-03-3BPC	11		21		31	
2	MB 280-14475/1-A	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: GC/MS BNA (EPA SW 846 Method 8270)

A. Phenol**	P. Bis(2-chloroethoxy)methane	EE. 2,6-Dinitrotoluene	TT. Pentachlorophenol**	III. Benzo(a)pyrene**
B. Bis (2-chloroethyl) ether	Q. 2,4-Dichlorophenol**	FF. 3-Nitroaniline	UU. Phenanthrene	JJJ. Indeno(1,2,3-cd)pyrene
C. 2-Chlorophenol	R. 1,2,4-Trichlorobenzene	GG. Acenaphthene**	VV. Anthracene	KKK. Dibenz(a,h)anthracene
D. 1,3-Dichlorobenzene	S. Naphthalene	HH. 2,4-Dinitrophenol*	WW. Carbazole	LLL. Benzo(g,h,i)perylene
E. 1,4-Dichlorobenzene**	T. 4-Chloroaniline	II. 4-Nitrophenol*	XX. Di-n-butylphthalate	MMM. Bis(2-Chloroisopropyl)ether
F. 1,2-Dichlorobenzene	U. Hexachlorobutadiene**	JJ. Dibenzofuran	YY. Fluoranthene**	NNN. Aniline
G. 2-Methylphenol	V. 4-Chloro-3-methylphenol**	KK. 2,4-Dinitrotoluene	ZZ. Pyrene	OOO. N-Nitrosodimethylamine
H. 2,2'-Oxybis(1-chloropropane)	W. 2-Methylnaphthalene	LL. Diethylphthalate	AAA. Butylbenzylphthalate	PPP. Benzoic Acid
I. 4-Methylphenol	X. Hexachlorocyclopentadiene*	MM. 4-Chlorophenyl-phenyl ether	BBB. 3,3'-Dichlorobenzidine	QQQ. Benzyl alcohol
J. N-Nitroso-di-n-propylamine*	Y. 2,4,6-Trichlorophenol**	NN. Fluorene	CCC. Benzo(a)anthracene	RRR. Pyridine
K. Hexachloroethane	Z. 2,4,5-Trichlorophenol	OO. 4-Nitroaniline	DDD. Chrysene	SSS. Benzidine
L. Nitrobenzene	AA. 2-Chloronaphthalene	PP. 4,6-Dinitro-2-methylphenol	EEE. Bis(2-ethylhexyl)phthalate	TTT.
M. Isophorone	BB. 2-Nitroaniline	QQ. N-Nitrosodiphenylamine (1)**	FFF. Di-n-octylphthalate**	UUU
N. 2-Nitrophenol**	CC. Dimethylphthalate	RR. 4-Bromophenyl-phenylether	GGG. Benzo(b)fluoranthene	VVV.
O. 2,4-Dimethylphenol	DD. Acenaphthylene	SS. Hexachlorobenzene	HHH. Benzo(k)fluoranthene	WWW.

Notes: * = System performance check compound (SPCC) for RRF; ** = Calibration check compound (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

LDC #: 23308 Z2a Page: 1 of 1
 SDG #: Su Cmed Reviewer: MB
 2nd Reviewer: A

Blanks

METHOD: GC/MS BNA (EPA SW 846 Method 8270C)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Was a method blank analyzed for each matrix?
- Y N N/A Was a method blank analyzed for each concentration preparation level?
- Y N N/A Was a method blank associated with every sample?
- Y N N/A Was the blank contaminated? If yes, please see qualification below.

Blank extraction date: 5/07/06 Blank analysis date: 5/12/06

Conc. units: ug/kg Associated Samples: All (61)

Compound	Blank ID	Sample Identification
BTE	MB 280-14475/1-A 62.7	1 83/0

Blank extraction date: _____ Blank analysis date: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification

5x Phthalates
2x all others

**Tronox LLC Facility, PCS, Henderson, Nevada
Data Validation Reports
LDC #23308**

Chlorinated Pesticides

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 15, 2010

LDC Report Date: June 15, 2010

Matrix: Water

Parameters: Chlorinated Pesticides

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2500-2

Sample Identification

EB-04152010-RIG2-RZE

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8081A for Chlorinated Pesticides.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/ECD Instrument Performance Check

Instrument performance was acceptable unless noted otherwise under initial calibration and continuing calibration sections.

III. Initial Calibration

Initial calibration of single compounds were performed for the primary (quantitation) column and confirmation column as required by this method.

The percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at required frequencies.

The percent differences (%D) of calibration factors in continuing standard mixtures were within the 20.0% QC limits.

The percent difference (%D) of the second source calibration standard were less than or equal to 20.0% for all compounds.

The individual 4,4'-DDT and Endrin breakdowns (%BD) were less than or equal to 15.0%.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No chlorinated pesticide contaminants were found in the method blanks.

Sample FB-04132010-RIG2-RZE was identified as a field blank. No chlorinated pesticide contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits with the following exceptions:

Sample	Column	Surrogate	%R (Limits)	Compound	Flag	A or P
EB-04152010-RIG2-RZE	Col. 1 Col. 2	Decachlorobiphenyl Decachlorobiphenyl	55 (68-122) 57 (68-122)	All TCL compounds	J- (all detects) UJ (all non-detects)	P

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Although the LCS percent recovery (%R) was not within QC limits for one compound, the LCSD percent recovery (%R) was within QC limits and no data were qualified.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Pesticide Cleanup Checks

a. Florisil Cartridge Check

Florisil cleanup was not required and therefore not performed in this SDG.

b. GPC Calibration

GPC cleanup was not required and therefore not performed in this SDG.

XI. Target Compound Identification

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2500-2	All compounds reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Data Qualification Summary - SDG 280-2500-2**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2500-2	EB-04152010-RIG2-RZE	All TCL compounds	J- (all detects) UJ (all non-detects)	P	Surrogate spikes (%R) (s)
280-2500-2	EB-04152010-RIG2-RZE	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Laboratory Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Field Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308A3a
SDG #: 280-2500-2
Laboratory: Test America

Stage 2B

Date: 6/14/10
Page: 1 of 1
Reviewer: JV
2nd Reviewer: Q

METHOD: GC Chlorinated Pesticides (EPA SW 846 Method 8081A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/15/10</u>
II.	GC/ECD Instrument Performance Check	A	
III.	Initial calibration	A	✓
IV.	Continuing calibration/ICV	A	<u>CV / IW ≤ 20%</u>
V.	Blanks	A	
VI.	Surrogate spikes	SW	
VII.	Matrix spike/Matrix spike duplicates	N	<u>Client Spec</u>
VIII.	Laboratory control samples	SW	<u>LCS/D</u>
IX.	Regional quality assurance and quality control	N	
Xa.	Florisil cartridge check	N	
Xb.	GPC Calibration	N	
XI.	Target compound identification	N	
XII.	Compound quantitation and reported CRQLs	N	
XIII.	Overall assessment of data	A	
XIV.	Field duplicates	N	
XV.	Field blanks	ND	<u>EB = 1</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

water

1	EB-04152010-RIG2-RZE	11		21		31	
2	<u>MB 280 - 11897 / 2-A</u>	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

VALIDATION FINDINGS WORKSHEET

METHOD: Pesticide/PCBs (EPASW 846 Method 8081/8082)

A. alpha-BHC	I. Dieldrin	Q. Endrin ketone	Y. Aroclor-1242	GG.
B. beta-BHC	J. 4,4'-DDE	R. Endrin aldehyde	Z. Aroclor-1248	HH.
C. delta-BHC	K. Endrin	S. alpha-Chlordane	AA. Aroclor-1254	II.
D. gamma-BHC	L. Endosulfan II	T. gamma-Chlordane	BB. Aroclor-1260	JJ.
E. Heptachlor	M. 4,4'-DDD	U. Toxaphene	CC. DB 608	KK.
F. Aldrin	N. Endosulfan sulfate	V. Aroclor-1016	DD. DB 1701	LL.
G. Heptachlor epoxide	O. 4,4'-DDT	W. Aroclor-1221	EE.	MM.
H. Endosulfan I	P. Methoxychlor	X. Aroclor-1232	FF.	NN.

Notes:

LDC #: 23308 A 3a
 SDG #: Sea Cow

VALIDATION FINDINGS WORKSHEET
Surrogate Spikes

Page: 1 of 1
 Reviewer: JV6
 2nd Reviewer: Q

METHOD: GC Pesticides/PCBs (EPA SW 846 Method 8081/8082)

Please see qualification below for all questions answered "N". Not applicable questions are identified as "N/A".
 N N/A Were surrogates spiked into all samples, standards and blanks?
 N N/A Did all surrogate percent recoveries (%R) meet the QC limits?

#	Date	Sample ID	Column	Surrogate Compound	%R (Limits)	Qualifications
			<u>Col. 1</u>	<u>B</u>	<u>55</u>	<u>5-113/P (S)</u>
			<u>2</u>		<u>57</u>	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	

Letter Designation	Surrogate Compound	Recovery QC Limits (Soil)	Recovery QC Limits (Water)	Comments
A	Tetrachloro-m-xylene			
B	Decachlorobiphenyl			

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 22, 2010
LDC Report Date: June 15, 2010
Matrix: Soil
Parameters: Chlorinated Pesticides
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2771-5

Sample Identification

SSAM2-01-4BPC
SSAM2-01-4BPCMS
SSAM2-01-4BPCMSD

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8081A for Chlorinated Pesticides.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Superfund Organic Methods Data Review (June 2008).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met with the following exceptions:

Sample	Compound	Total Days From Sample Collection Until Extraction	Required Holding Time (in Days) From Sample Collection Until Extraction	Flag	A or P
All samples in SDG 280-2771-5	All TCL compounds	19	14	J- (all detects) UJ (all non-detects)	P

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/ECD Instrument Performance Check

Instrument performance was acceptable unless noted otherwise under initial calibration and continuing calibration sections.

III. Initial Calibration

Initial calibration of single compounds were performed for the primary (quantitation) column and confirmation column as required by this method.

The percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at required frequencies.

The percent differences (%D) of calibration factors in continuing standard mixtures were within the 20.0% QC limits.

The percent difference (%D) of the second source calibration standard were less than or equal to 20.0% for all compounds.

The individual 4,4'-DDT and Endrin breakdowns (%BD) were less than or equal to 15.0%.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No chlorinated pesticide contaminants were found in the method blanks.

Sample FB-04132010-RIG2-RZE (from SDG 280-2400-2) was identified as a field blank. No chlorinated pesticide contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. Surrogate recoveries (%R) were not within QC limits for sample SSAM2-01-4BPC. Since the sample was diluted out, no data were qualified.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were not within the QC limits. Since the samples were diluted out, no data were qualified.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Pesticide Cleanup Checks

a. Florisil Cartridge Check

Florisil cleanup was not required and therefore not performed in this SDG.

b. GPC Calibration

GPC cleanup was not required and therefore not performed in this SDG.

XI. Target Compound Identification

Raw data were not reviewed for this SDG.

XII. Project Quantitation Limit

All compounds reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2771-5	All compounds reported below the PQL	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Data Qualification Summary - SDG 280-2771-5**

SDG	Sample	Compound	Flag	A or P	Reason (Code)
280-2771-5	SSAM2-01-4BPC	All TCL compounds	J- (all detects) UJ (all non-detects)	P	Technical holding times (h)
280-2771-5	SSAM2-01-4BPC	All compounds reported below the PQL.	J (all detects)	A	Project Quantitation Limit (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Laboratory Blank Data Qualification Summary - SDG 280-2771-5**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Chlorinated Pesticides - Field Blank Data Qualification Summary - SDG 280-2771-5**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308D3a
SDG #: 280-2771-5
Laboratory: Test America

Stage 2B

Date: 6/15/10
Page: 1 of 1
Reviewer: [Signature]
2nd Reviewer: [Signature]

METHOD: GC Chlorinated Pesticides (EPA SW 846 Method 8081A)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

Validation Area		Comments	
I.	Technical holding times	SW	Sampling dates: 4/22/10
II.	GC/ECD Instrument Performance Check	A	
III.	Initial calibration	A	% RSD 12
IV.	Continuing calibration/ICV	A	CV/ICV ≤ 20%
V.	Blanks	A	
VI.	Surrogate spikes	SW	
VII.	Matrix spike/Matrix spike duplicates	SW	
VIII.	Laboratory control samples	A	LCS
IX.	Regional quality assurance and quality control	N	
Xa.	Florisil cartridge check	N	
Xb.	GPC Calibration	N	
XI.	Target compound identification	N	
XII.	Compound quantitation and reported CRQLs	N	
XIII.	Overall assessment of data	A	
XIV.	Field duplicates	N	
XV.	Field blanks	ND	FB = FB-04132010-RIG2-RZE (280-240-2)

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: Soil

1	SSAM2-01-4BPC	11	21	31
2	SSAM2-01-4BPCMS	12	22	32
3	SSAM2-01-4BPCMSD	13	23	33
4	<u>MB 280-14906/2-A</u>	14	24	34
5		15	25	35
6		16	26	36
7		17	27	37
8		18	28	38
9		19	29	39
10		20	30	40

LDC #: 23308 D3A
 SDG #: Sa Crer

VALIDATION FINDINGS WORKSHEET
 Technical Holding Times

Page: 1 of 1
 Reviewer: JY6
 2nd Reviewer: [Signature]

All circled dates have exceeded the technical holding times.
 Y/N N/A Were all cooler temperatures within validation criteria?

METHOD : GC Pesticides/PCBs (EPA SW 846 Method 8081/8082)

Sample ID	Matrix	Preserved	Sampling Date	Extraction date	Analysis date	Total # of Days	Qualifier
All	S	N	4/22/10	5/11/10	5/13/10	19	J-UJ/P (h)

TECHNICAL HOLDING TIME CRITERIA

Water: Extracted within 7 days, analyzed within 40 days.
 Soil: Extracted within 14 days, analyzed within 40 days.

LDC #: 23368 D34
 SDG #: sa cmv

VALIDATION FINDINGS WORKSHEET
Surrogate Spikes

Page: 1 of 1
 Reviewer: JG
 2nd Reviewer: _____

METHOD: GC Pesticides/PCBs (EPA SW 846 Method 8081/8082)

Please see qualification below for all questions answered "N". Not applicable questions are identified as "N/A".

N N/A Were surrogates spiked into all samples, standards and blanks?
 N N/A Did all surrogate percent recoveries (%R) meet the QC limits?

#	Date	Sample ID	Column	Surrogate Compound	%R (Limits)	Qualifications
		1 (100%)	Cd.1	A	1720 (57-115)	No qual
				B	0 (63-124)	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	
					()	

Letter Designation	Surrogate Compound	Recovery QC Limits (Soil)	Recovery QC Limits (Water)	Comments
A	Tetrachloro-m-xylene			
B	Decachlorobiphenyl			

**Tronox LLC Facility, PCS, Henderson, Nevada
Data Validation Reports
LDC #23308**

Metals

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 15 through April 16, 2010
LDC Report Date: June 21, 2010
Matrix: Water
Parameters: Metals
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2500-2

Sample Identification

EB-04152010-1-RZD
EB-04152010-RIG2-RZE
EB-04152010-1-RZDMS
EB-04152010-1-RZDMSD

Introduction

This data review covers 4 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Metals. The metals analyzed were Arsenic, Cobalt, Lead, Magnesium, and Manganese.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No metal contaminants were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Cobalt Manganese	0.0113 ug/L 0.441 ug/L	All samples in SDG 280-2500-2

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-04152010-1-RZD	Manganese	0.84 ug/L	1.0U ug/L
EB-04152010-RIG2-RZE	Cobalt	0.16 ug/L	1.0U ug/L

Samples EB-04152010-1-RZD and EB-04152010-RIG2-RZE were identified as equipment blanks. No metal contaminants were found in these blanks with the following exceptions:

Equipment Blank ID	Sampling Date	Analyte	Concentration	Associated Samples
EB-04152010-1-RZD	4/16/10	Manganese	0.84 ug/L	No associated samples in this SDG
EB-04152010-RIG2-RZE	4/15/10	Lead Cobalt Manganese Magnesium	0.18 ug/L 0.16 ug/L 9.4 ug/L 50 ug/L	No associated samples in this SDG

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2500-2	All analytes reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Metals - Data Qualification Summary - SDG 280-2500-2**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2500-2	EB-04152010-1-RZD EB-04152010-RIG2-RZE	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
 Metals - Laboratory Blank Data Qualification Summary - SDG 280-2500-2**

SDG	Sample	Analyte	Modified Final Concentration	A or P	Code
280-2500-2	EB-04152010-1-RZD	Manganese	1.0U ug/L	A	bl
280-2500-2	EB-04152010-RIG2-RZE	Cobalt	1.0U ug/L	A	bl

**Tronox LLC Facility, PCS, Henderson, Nevada
 Metals - Equipment Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

LDC #: 23308A4
 SDG #: 280-2500-2
 Laboratory: Test America

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B

Date: 5/15/10
 Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: Metals (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/15-16/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	SW	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	A	ms/D
VII.	Duplicate Sample Analysis	N	
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	N	
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	N	
XV.	Field Blanks	SW	EB=1, 2 (no associated samples)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Water

1	EB-04152010-1RZD	11	RBW	21		31	
2	EB-04152010-RIG2-RZE	12		22		32	
3	EB-04152010-1RZDMS	13		23		33	
4	EB-04152010-1RZDMSD	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC #: 23308A4
 SDG #: SOCOR

VALIDATION FINDINGS WORKSHEET
Sample Specific Element Reference

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

All circled elements are applicable to each sample.

Sample ID	Matrix	Target Analyte List (TAL)
1, 2		Al, Sb, (As) Ba, Be, Cd, Ca, Cr, (Co) Cu, Fe, (Pb) (Mg) (Mn) Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
QC-3, 4		Al, Sb, (As) Ba, Be, Cd, Ca, Cr, (Co) Cu, Fe, (Pb) (Mg) (Mn) Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
Analysis Method		
ICP		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
ICP Trace		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
ICP-MS		Al, Sb, (As) Ba, Be, Cd, Ca, Cr, (Co) Cu, Fe, (Pb) (Mg) (Mn) Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____
GFAA		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', _____

Comments: Mercury by CVAA if performed

LDC #: 23308AH
 SDG #: 8220202

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Reason: bl

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer:

Associated Samples: All

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: NA
 Sample Concentration units, unless otherwise noted: ug/L

Analyte	Maximum PB* (mg/Kg)	Maximum PB* (ug/L)	Maximum ICB/CCB* (ug/L)	Blank Action Limit	Sample Identification																				
					1	2																			
Al																									
Sb																									
As																									
Ba																									
Be																									
Cd																									
Ca																									
Cr																									
Cu																									
Fe																									
Pb																									
Mg																									
Mn																									
Hg																									
Ni																									
K																									
Se																									
Ag																									
Na																									
Tl																									
V																									
Zn																									
B																									
Mo																									
Sr																									

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 15, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Arsenic
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2500-8

Sample Identification

SA165-3BPC
SA131-6BPC
SA131-8BPC**

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 3 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Sample EB-04152010-RIG2-RZE (from SDG 280-2500-2) was identified as an equipment blank. No arsenic was found in this blank.

Sample FB-04132010-RIG2-RZE (from SDG 280-2460-2) was identified as a field blank. No arsenic was found in this blank.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

The laboratory has indicated that there were no matrix spike (MS) analyses specified for the samples in this SDG, and therefore matrix spike analyses were not performed for this SDG.

VII. Duplicate Sample Analysis

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution was not performed for this SDG.

XII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2500-8	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Data Qualification Summary - SDG 280-2500-8**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2500-8	SA165-3BPC SA131-6BPC SA131-8BPC**	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-2500-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Equipment Blank Data Qualification Summary - SDG 280-2500-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Field Blank Data Qualification Summary - SDG 280-2500-8**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308B4
SDG #: 280-2500-8
Laboratory: Test America

Stage 2B /4

Date: 6-15-10
Page: 1 of 1
Reviewer: [Signature]
2nd Reviewer: [Signature]

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/15/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	N	Client specified
VII.	Duplicate Sample Analysis	N	↓
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	N	Not performed
XII.	Sample Result Verification	A	Not reviewed for 2B
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	N	
XV.	Field Blanks	ND	EB = EB-04152010-RIG2-RZE, FB = FB-04132010-RIG2-RZE (280-2500-2) (280-2460-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
N = Not provided/applicable R = Rinsate TB = Trip blank
SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: soil ****Level 4**

1	SA165-3BPC	11	PBS	21		31	
2	SA131-6BPC	12		22		32	
3	SA131-8BPC **	13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

Method:Metals (EPA SW 846 Method 6010/7000/6020)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Cooler temperature criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
II. Calibration				
Were all isotopes in the tuning solution mass resolution within 0.1 amu?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were %RSD of isotopes in the tuning solution < 5%?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all instruments calibrated daily, each set-up time?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the proper number of standards used?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all initial calibration correlation coefficients > 0.995?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
III. Blanks				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
IV. ICP Interference Check Samples				
Were ICP interference check samples performed daily?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the AB solution percent recoveries (%R) with the 80-120% QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
IV. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of +/- RL (+/-2X RL for soil) was used for samples that were ≤ 5X the RL, including when only one of the duplicate sample values were < 5X the RL.	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
V. Laboratory control samples				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was an LCS analyzed per extraction batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

LDC #: 2330804
 SDG #: Seacover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: efr
 2nd Reviewer: v

Validation Area	Yes	No	NA	Findings/Comments
VI. Reference Material Absorption QC				
If MSA was performed, was the correlation coefficients > 0.995?			/	
Do all applicable analyses have duplicate injections? (Level IV only)			/	
For sample concentrations > RL, are applicable duplicate injection RSD values < 20%? (Level IV only)			/	
Were analytical spike recoveries within the 85-115% QC limits?			/	
VII. ICP Serial Dilution				
Was an ICP serial dilution analyzed if analyte concentrations were > 50X the IDL?		/		
Were all percent differences (%Ds) < 10%?			/	
Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data.			/	
VIII. Internal Standards (EPA SW-846 Method 6020)				
Were all the percent recoveries (%R) within the 30-120% of the intensity of the internal standard in the associated initial calibration?	/			
If the %Rs were outside the criteria, was a reanalysis performed?	/			
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		/		
Were the performance evaluation (PE) samples within the acceptance limits?			/	
X. Sample Result Verification				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	/			
XI. Overall assessment of data				
Overall assessment of data was found to be acceptable.	/			
XII. Field duplicates				
Field duplicate pairs were identified in this SDG.		/		
Target analytes were detected in the field duplicates.			/	
XIII. Field blanks				
Field blanks were identified in this SDG.	/			
Target analytes were detected in the field blanks.		/		

LDC #: 230804
 SDG #: seecover

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: GR
 2nd Reviewer: W

METHOD: Trace Metals (EPA SW 846 Method 6010/6020/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$
 Where, Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration (in ug/L) of each analyte in the ICV or CCV source

Standard ID	Type of Analysis	Element	Found (ug/L)	True (ug/L)	Recalculated		Reported		Acceptable (Y/N)
					%R		%R		
	ICP (Initial calibration)								
	GFAA (Initial calibration)								
	CVAA (Initial calibration)								
	ICP (Continuing calibration)								
	GFAA (Continuing calibration)								
	CVAA (Continuing calibration)								
ICV	ICP/MS (Initial calibration)	As	38.4	40.0	96		96		Y
CCV (03-16)	ICP/MS (Continuing calibration)	↓	48.8	50.0	98		98		Y

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 28808034
 SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: GR
 2nd Reviewer: LA

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$

Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation,
 Found = SSR (spiked sample result) - SR (sample result).
 True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$

Where, S = Original sample concentration
 D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$

Where, I = Initial Sample Result (mg/L)
 SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

Sample ID	Type of Analysis	Element	Found / S / I (units) (mg/L) (mg/L)	True / D / SDR (units) mg/L	Recalculated		Acceptable (Y/N)
					%R / RPD / %D	%R / RPD / %D	
ICSPS	ICP interference check	As	100 µg/L	100 µg/L	100	100	Y
LCS	Laboratory control sample	As	18.5	20	93	92	Y
N	Matrix spike		(SSR-SR)				
N	Duplicate						
N	ICP serial dilution						

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23308P1
SDG #: seedler

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
Reviewer: CR
2nd reviewer: W

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

<input checked="" type="checkbox"/> Y	<input type="checkbox"/> N	<input type="checkbox"/> N/A	Have results been reported and calculated correctly?
<input checked="" type="checkbox"/> Y	<input type="checkbox"/> N	<input type="checkbox"/> N/A	Are results within the calibrated range of the instruments and within the linear range of the ICP?
<input checked="" type="checkbox"/> Y	<input type="checkbox"/> N	<input type="checkbox"/> N/A	Are all detection limits below the CRDL?

Detected analyte results for AS were recalculated and verified using the following equation:

Concentration = $\frac{(RD)(FV)(Dil)}{(In. Vol.)(\%S)}$

RD = Raw data concentration
FV = Final volume (ml)
In. Vol. = Initial volume (ml) or weight (G)
Dil = Dilution factor
%S = Decimal percent solids

Recalculation:

$$\frac{(13.75 \mu g/L / 1000) (100 mL) (5)}{(0.896) (1.17g)} = 6.6 \text{ mg/kg}$$

Sample ID	Analyte	Reported Concentration (mg/kg)	Calculated Concentration (mg/kg)	Acceptable (Y/N)
<u>3</u>	<u>AS</u>	<u>6.6</u>	<u>6.6</u>	<u>Y</u>

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 23, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Arsenic
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2836-9

Sample Identification

SSAJ2-01-7BPC
SSAJ2-01-9BPC**

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 2 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No arsenic was found in this blank.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

The laboratory has indicated that there were no matrix spike (MS) analyses specified for the samples in this SDG, and therefore matrix spike analyses were not performed for this SDG.

VII. Duplicate Sample Analysis

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2836-9	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Data Qualification Summary - SDG 280-2836-9**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2836-9	SSAJ2-01-7BPC SSAJ2-01-9BPC**	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-2836-9**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Field Blank Data Qualification Summary - SDG 280-2836-9**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308H4
SDG #: 280-2836-9
Laboratory: Test America

Stage 2B 14

Date: 6-15-10
Page: 1 of 1
Reviewer: CR
2nd Reviewer: W

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

Validation Area			Comments
I.	Technical holding times	A	Sampling dates: <u>4/23/10</u>
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	N	Client specified
VII.	Duplicate Sample Analysis	N	↓
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	A	Not reviewed for 2B
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	N	
XV.	Field Blanks	ND	FB = FB-04072010 - R20 (280-2216-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
N = Not provided/applicable R = Rinstate TB = Trip blank
SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil **Level 4

1	SSAJ2-01-7BPC	11		21		31	
2	SSAJ2-01-9BPC **	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC #: 23308H4
 SDG #: See cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: CR
 2nd Reviewer: [Signature]

Method:Metals (EPA SW 846 Method 6010/7000/6020)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	/			
Cooler temperature criteria was met.	/			
II. Calibration				
Were all isotopes in the tuning solution mass resolution within 0.1 amu?	/			
Were %RSD of isotopes in the tuning solution < 5%?	/			
Were all instruments calibrated daily, each set-up time?	/			
Were the proper number of standards used?	/			
Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits?	/			
Were all initial calibration correlation coefficients > 0.995?	/			
III. Blanks				
Was a method blank associated with every sample in this SDG?	/			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.		/		
IV. ICP Interference Check Sample				
Were ICP interference check samples performed daily?	/			
Were the AB solution percent recoveries (%R) with the 80-120% QC limits?	/			
IV. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.		/		
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.			/	
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of +/- RL(+/-2X RL for soil) was used for samples that were ≤ 5X the RL, including when only one of the duplicate sample values were < 5X the RL.			/	
V. Laboratory control samples				
Was an LCS analyzed for this SDG?	/			
Was an LCS analyzed per extraction batch?	/			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils?	/			

LDC #: 23308414
 SDG #: Seecover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: er
 2nd Reviewer: W

Validation Area	Yes	No	NA	Findings/Comments
VI. Method: Atomic Absorption QC				
If MSA was performed, was the correlation coefficients > 0.995?			/	
Do all applicable analyses have duplicate injections? (Level IV only)			/	
For sample concentrations > RL, are applicable duplicate injection RSD values < 20%? (Level IV only)			/	
Were analytical spike recoveries within the 85-115% QC limits?			/	
VII. ICP: Serial Dilution				
Was an ICP serial dilution analyzed if analyte concentrations were > 50X the IDL?		/		
Were all percent differences (%Ds) < 10%?			/	
Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data.			/	
VIII. Method Standards (EPA 8160-G Method 6020)				
Were all the percent recoveries (%R) within the 30-120% of the intensity of the internal standard in the associated initial calibration?	/			
If the %Rs were outside the criteria, was a reanalysis performed?	/			
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		/		
Were the performance evaluation (PE) samples within the acceptance limits?			/	
X. Sample Result Verification				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	/			
XI. Overall assessment of data				
Overall assessment of data was found to be acceptable.	/			
XII. Field Duplicates				
Field duplicate pairs were identified in this SDG.		/		
Target analytes were detected in the field duplicates.			/	
XIII. Field Blanks				
Field blanks were identified in this SDG.	/			
Target analytes were detected in the field blanks.		/		

LDC #: 2330874

SDG #: See cover

VALIDATION FINDINGS WORKSHEET

Initial and Continuing Calibration Calculation Verification

Page: 1 of 1

Reviewer: GR

2nd Reviewer: VR

METHOD: Trace Metals (EPA SW 846 Method 6010/6020/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration (in ug/L) of each analyte in the ICV or CCV source

Standard ID	Type of Analysis	Element	Found (ug/L)	True (ug/L)	Recalculated		Reported		Acceptable (Y/N)
					%R		%R		
	ICP (Initial calibration)								
	GFAA (Initial calibration)								
	CVAA (Initial calibration)								
	ICP (Continuing calibration)								
	GFAA (Continuing calibration)								
	CVAA (Continuing calibration)								
ICV	ICP/MS (Initial calibration)	As	41.0	40.0	103		102		Y
CCV (03/3)	ICP/MS (Continuing calibration)	↓	51.0	50.0	102		102		Y

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 2330874
 SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: GS
 2nd Reviewer: _____

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$

Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation,
 Found = SSR (spiked sample result) - SR (sample result).
 True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$

Where, S = Original sample concentration
 D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$

Where, I = Initial Sample Result (mg/L)
 SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

Sample ID	Type of Analysis	Element	Found / S / I (units)	True / D / SDR (units)	Recalculated		Acceptable (Y/N)
					%R / RPD / %D	%R / RPD / %D	
ICSPB	ICP interference check	As	99.9 ug/L	100 ug/L	100	100	Y
LCS	Laboratory control sample	J	19.0 mg/kg	20.0 mg/kg	95	95	J
N	Matrix spike		(SSR-SR)				
N	Duplicate						
N	ICP serial dilution						

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23306H4
SDG #: seader

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
Reviewer: [Signature]
2nd reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Have results been reported and calculated correctly?
- Y N N/A Are results within the calibrated range of the instruments and within the linear range of the ICP?
- Y N N/A Are all detection limits below the CRDL?

Detected analyte results for As were recalculated and verified using the following equation:

Concentration = $\frac{(RD)(FV)(DI)}{(In. Vol.)(\%S)}$

- RD = Raw data concentration
- FV = Final volume (ml)
- In. Vol. = Initial volume (ml) or weight (G)
- Dil = Dilution factor
- %S = Decimal percent solids

Recalculation:

$$\frac{(65.8 \text{ mg/L}) (5) (100 \text{ mL})}{(0.928) (1.12 \text{ g})} = 32 \text{ mg/kg}$$

Sample ID	Analyte	Reported Concentration (mg/kg)	Calculated Concentration (mg/kg)	Acceptable (Y/N)
2	As	32	32	Y

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 27, 2010

LDC Report Date: June 21, 2010

Matrix: Soil

Parameters: Arsenic

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2960-7

Sample Identification

SSAK8-04-4BPC

Introduction

This data review covers one soil sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Sample FB-04072010-RZD (from SDG 280-2216-2) was identified as a field blank. No arsenic was found in this blank.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

The laboratory has indicated that there were no matrix spike (MS) analyses specified for the samples in this SDG, and therefore matrix spike analyses were not performed for this SDG.

VII. Duplicate Sample Analysis

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution was not performed for this SDG.

XII. Sample Result Verification and Project Quantitation Limit

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2960-7	All analytes reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Data Qualification Summary - SDG 280-2960-7**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2960-7	SSAK8-04-4BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-2960-7**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Field Blank Data Qualification Summary - SDG 280-2960-7**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 2330804
SDG #: 280-2960-7
Laboratory: Test America

Stage 2B

Date: 6-15-10
Page: 1 of 1
Reviewer: CTC
2nd Reviewer: W

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/27/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	N	Client specified
VII.	Duplicate Sample Analysis	N	↓
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	N	Not performed
XII.	Sample Result Verification	N	
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	N	
XV.	Field Blanks	ND	FB = FB-04072010-RZD (280-2216-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
N = Not provided/applicable R = Rinstate TB = Trip blank
SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAK8-04-4BPC	11	PBS	21		31	
2		12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

Laboratory Data Consultants, Inc. Data Validation Report

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 28, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Metals
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2995-1

Sample Identification

SSAN6-07-1BPC
SSAN6-07-5BPC
RSAQ3-1BPC
RSAQ3-2BPC
SA56-1BPC
SA56-2BPC
SA56-2BPC_FD
SA48-1BPC**
SA48-1BPC_FD
SA48-2BPC
SA09-1BPC
SA09-2BPC
SA188-1BPC
SA188-2BPC
RSAN6-1BPC
RSAN6-2BPC
SSAN6-07-1BPCMS
SSAN6-07-1BPCMSD
RSAQ3-2BPCMS
RSAQ3-2BPCMSD

**Indicates sample underwent Stage 4 review

Introduction

This data review covers 20 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Metals. The metals analyzed were Arsenic and Manganese.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No metal contaminants were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
ICB/CCB	Manganese	0.829 ug/L	SA56-1BPC SA56-2BPC SA56-2BPC_FD SA188-1BPC SA188-2BPC

Sample concentrations were compared to concentrations detected in the method blanks as required by the QAPP. No sample data was qualified.

Sample EB-04282010-RZB (from SDG 280-2995-2) was identified as an equipment blank. No metal contaminants were found in this blank.

Samples FB-04072010-RZC (from SDG 280-2280-2) and FB04062010-RZB (from SDG 280-2131-1) were identified as field blanks. No metal contaminants were found in these blanks.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-1	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

Samples SA56-2BPC and SA56-2BPC_FD and samples SA48-1BPC** and SA48-1BPC_FD were identified as field duplicates. No metal contaminants were detected in any of the samples with the following exceptions:

Compound	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SA56-2BPC	SA56-2BPC_FD				
Arsenic	3.4	3.3	3 (≤ 50)	-	-	-
Manganese	560	530	6 (≤ 50)	-	-	-

Compound	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SA48-1BPC**	SA48-1BPC_FD				
Arsenic	3.3	3.4	3 (≤ 50)	-	-	-

**Tronox LLC Facility, PCS, Henderson, Nevada
Metals - Data Qualification Summary - SDG 280-2995-1**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2995-1	SSAN6-07-1BPC SSAN6-07-5BPC RSAQ3-1BPC RSAQ3-2BPC SA56-1BPC SA56-2BPC SA56-2BPC_FD SA48-1BPC** SA48-1BPC_FD SA48-2BPC SA09-1BPC SA09-2BPC SA188-1BPC SA188-2BPC RSAN6-1BPC RSAN6-2BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Metals - Laboratory Blank Data Qualification Summary - SDG 280-2995-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Metals - Equipment Blank Data Qualification Summary - SDG 280-2995-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Metals - Field Blank Data Qualification Summary - SDG 280-2995-1**

No Sample Data Qualified in this SDG

LDC #: 23308P4
 SDG #: 280-2995-1
 Laboratory: Test America

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B /4

Date: 6-15-10
 Page: 1 of 1
 Reviewer: CR
 2nd Reviewer: ✓

METHOD: As & Mn (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/25/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	SW	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	A	MS/D
VII.	Duplicate Sample Analysis	N	
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	A	Not reviewed for 2B
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	SW	(6,7), (8,9)
XV.	Field Blanks	ND	FB = FB-04072010-RZC, FB04062010-RZB (280-2280-2) (280-2131-1) EB = FB04281010-RZB (280-2995-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAN6-07-1BPC	11	SA09-1BPC	21	PBS	31	
2	SSAN6-07-5BPC	12	SA09-2BPC	22		32	
3	RSAQ3-1BPC	13	SA188-1BPC	23		33	
4	RSAQ3-2BPC	14	SA188-2BPC	24		34	
5	SA56-1BPC	15	RSAN6-1BPC	25		35	
6	SA56-2BPC	16	RSAN6-2BPC	26		36	
7	SA56-2BPC_FD	17	SSAN6-07-1BPCMS	27		37	
8	SA48-1BPC **	18	SSAN6-07-1BPCMSD	28		38	
9	SA48-1BPC_FD	19	RSAQ3-2BPCMS	29		39	
10	SA48-2BPC	20	RSAQ3-2BPCMSD	30		40	

Notes: _____

Method:Metals (EPA SW 846 Method 6010/7000/6020)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	/			
Cooler temperature criteria was met.	/			
II. Calibration				
Were all isotopes in the tuning solution mass resolution within 0.1 amu?	/			
Were %RSD of isotopes in the tuning solution < 5%?	/			
Were all instruments calibrated daily, each set-up time?	/			
Were the proper number of standards used?	/			
Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits?	/			
Were all initial calibration correlation coefficients > 0.995?	/			
III. Blanks				
Was a method blank associated with every sample in this SDG?	/			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	/			
IV. ICP Interference Check Samples				
Were ICP interference check samples performed daily?	/			
Were the AB solution percent recoveries (%R) with the 80-120% QC limits?	/			
IV. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.	/			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.	/			
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of +/- RL(+/-2X RL for soil) was used for samples that were ≤ 5X the RL, including when only one of the duplicate sample values were < 5X the RL.	/			
V. Laboratory control samples				
Was an LCS analyzed for this SDG?	/			
Was an LCS analyzed per extraction batch?	/			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils?	/			

LDC #: 2332068P4
 SDG #: secaer

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: er
 2nd Reviewer: v

Validation Area	Yes	No	NA	Findings/Comments
VI. Matrix-Absorbance Absorption QC				
If MSA was performed, was the correlation coefficients > 0.995?			✓	
Do all applicable analyses have duplicate injections? (Level IV only)			✓	
For sample concentrations > RL, are applicable duplicate injection RSD values < 20%? (Level IV only)			✓	
Were analytical spike recoveries within the 85-115% QC limits?			✓	
VII. ICP Serial Dilution				
Was an ICP serial dilution analyzed if analyte concentrations were > 50X the IDL?	✓			
Were all percent differences (%Ds) < 10%?	✓			
Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data.		✓		
VIII. Internal Standards (EPA SW-846 Method 6020)				
Were all the percent recoveries (%R) within the 30-120% of the intensity of the internal standard in the associated initial calibration?	✓			
If the %Rs were outside the criteria, was a reanalysis performed?	✓			
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		✓		
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
X. Sample Result Verification				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
XI. Overall Assessment of Data				
Overall assessment of data was found to be acceptable.	✓			
XII. Field Duplicates				
Field duplicate pairs were identified in this SDG.	✓			
Target analytes were detected in the field duplicates.	✓			
XIII. Field Blanks				
Field blanks were identified in this SDG.	✓			
Target analytes were detected in the field blanks.		✓		

LDC #: 2330624
 SDG #: SEOCOR

VALIDATION FINDINGS WORKSHEET
Sample Specific Element Reference

Page: 1 of 1
 Reviewer: CR
 2nd reviewer: V

All circled elements are applicable to each sample.

Sample ID	Matrix	Target Analyte List (TAL)
1-4, 8, 12, 15, 16		Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
57		Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, (Mn), Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
13, 14		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, (Mn), Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
QC: 17, 18		Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
19, 20		Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
Analysis Method		
ICP		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
ICP Trace		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
ICP-MS		Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, (Mn), Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____
GFAA		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Tl, V, Zn, Mo, B, Si, CN', _____

Comments: Mercury by CVAA if performed

LDC #: 2330804
 SDG #: 28208-1

VALIDATION FINDINGS WORKSHEET
 PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

5-7, 13, 14

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: ~~1.0~~ 1.0X
 Sample Concentration units, unless otherwise noted: mg/kg Associated Samples:

Analyte	Maximum PB* (mg/Kg)	Maximum PB* (ug/L)	Maximum ICB/CCB* (ug/L)	Blank Action Limit	Sample Identification
Al					No Qual (all 7 RL)
Sb					
As					
Ba					
Be					
Cd					
Ca					
Cr					
Co					
Cu					
Fe					
Pb			0.829		
Mg					
Mn					
Hg					
Ni					
K					
Se					
Ag					
Na					
Ti					
V					
Zn					
B					
Mo					
Sr					

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "L".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

LDC#: 23308P4
 SDG#: See Cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: CS
 2nd Reviewer: [Signature]

METHOD: Metals (EPA Method 6020/7000)

Y N NA Were field duplicate pairs identified in this SDG?
Y N NA Were target analytes detected in the field duplicate pairs?

Compound	Concentration (mg/Kg)		(≤ 50)	(mg/Kg)	(mg/Kg)	Qualifications (Parent Only)
	6	7	RPD	Difference	Limits	
Arsenic	3.4	3.3	3			
Manganese	560	530	6			

V:\FIELD DUPLICATES\FD_inorganic\23308P4.wpd

Compound	Concentration (mg/Kg)		(≤ 50)	(mg/Kg)	(mg/Kg)	Qualifications (Parent Only)
	8	9	RPD	Difference	Limits	
Arsenic	3.3	3.4	3			

LDC #: 2330894
 SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: GR
 2nd Reviewer: GR

METHOD: Trace Metals (EPA SW 846 Method 6010/6020/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$\%R = \frac{\text{Found}}{\text{True}} \times 100$ Where, Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration (in ug/L) of each analyte in the ICV or CCV source

Standard ID	Type of Analysis	Element	Found (ug/L)	True (ug/L)	Recalculated		Reported		Acceptable (Y/N)
					%R		%R		
	ICP (Initial calibration)								
	GFAA (Initial calibration)								
	CVAA (Initial calibration)								
	ICP (Continuing calibration)								
	GFAA (Continuing calibration)								
	CVAA (Continuing calibration)								
ICV	ICPMS (Initial calibration)	As	38.4	40.0			96		Y
CCV (0.103)	ICPMS (Continuing calibration)	J	49.8	50.0			100		Y

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 2328894
 SDG #: See cover

Page: 1 of 1
 Reviewer: CS
 2nd Reviewer: [Signature]

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$
 Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).
 True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$
 Where, S = Original sample concentration
 D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$
 Where, I = Initial Sample Result (mg/L)
 SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

Sample ID	Type of Analysis	Element	Found / S / I (units)	True / D / SDR (units)	Recalculated		Reported		Acceptable (Y/N)
					%R / RPD / %D	%R / RPD / %D			
ICSFB	ICP interference check	As	100 ug/L	100 ug/L	100	100	100	100	Y
LCS	Laboratory control sample		19.2	20	96	96	96	96	Y
17	Matrix spike		19.6 (SSR-SR)	21.7	90	90	90	90	Y
17/16	Duplicate		23.4	23.8	2	2	2	2	Y
1	ICP serial dilution		3.8	4.13	8.7	8.9	8.9	8.9	Y

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23768PM
 SDG #: seedler

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
 Reviewer: CR
 2nd reviewer: R

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- N N/A Have results been reported and calculated correctly?
- N N/A Are results within the calibrated range of the instruments and within the linear range of the ICP?
- N N/A Are all detection limits below the CRDL?

Detected analyte results for AS were recalculated and verified using the following equation:

Concentration = $\frac{(RD)(FV)(DII)}{(In. Vol.)(\%S)}$

Recalculation:

$$\frac{(6.35 \text{ mg/L})(5)(100 \text{ mL})}{1000} = 3.3 \text{ mg/kg}$$

$$(0.942)(1.02\%)$$

- RD = Raw data concentration
- FV = Final volume (ml)
- In. Vol. = Initial volume (ml) or weight (G)
- DII = Dilution factor
- %S = Decimal percent solids

Sample ID	Analyte	Reported Concentration (mg/kg)	Calculated Concentration (mg/kg)	Acceptable (Y/N)
8	AS	3.3	3.3	T

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 28, 2010

LDC Report Date: June 23, 2010

Matrix: Water

Parameters: Arsenic

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2995-2

***Sample Identification**

EB04282010-RZB

EB04282010-RZBMS

EB04282010-RZBMSD

*Corrected all sample IDs from EB04281010 to EB04282010

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Sample EB04282010-RZB was identified as a field blank. No arsenic was found in this blank.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-2	All analytes reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Data Qualification Summary - SDG 280-2995-2**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2995-2	EB04282010-RZB	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-2995-2**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Arsenic - Equipment Blank Data Qualification Summary - SDG 280-2995-2**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308Q4
 SDG #: 280-2995-2
 Laboratory: Test America

Stage 2B

Date: 6-15-10
 Page: 1 of 1
 Reviewer: CA
 2nd Reviewer: W

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/28/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	A	MS/D
VII.	Duplicate Sample Analysis	N	
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	N	
XIII.	Overall Assessment of Data	N	
XIV.	Field Duplicates	N	
XV.	Field Blanks	ND	EB=1 (no associated samples)

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet
 ND = No compounds detected
 R = Rinsate
 FB = Field blank
 D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: water

1	EB0428 ¹ 010-RZB	11	<u>RBW</u>	21		31	
2	EB0428 ² 010-RZBMS	12		22		32	
3	EB0428 ³ 010-RZBMSD	13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 28, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Arsenic
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2995-6

Sample Identification

SSAN6-07-7BPC
SSAN6-07-9BPC
SSAN6-07-7BPCMS
SSAN6-07-7BPCMSD

Introduction

This data review covers 4 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Sample FB-04072010-RZC (from SDG 280-2280-2) was identified as a field blank. No arsenic was found in this blank.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-6	All analytes reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Data Qualification Summary - SDG 280-2995-6**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2995-6	SSAN6-07-7BPC SSAN6-07-9BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-2995-6**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Field Blank Data Qualification Summary - SDG 280-2995-6**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308S4
SDG #: 280-2995-6
Laboratory: Test America

Stage 2B

Date: 6-15-10
Page: 1 of 1
Reviewer: CR
2nd Reviewer: W

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/28/10</u>
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	A	<u>MS/D</u>
VII.	Duplicate Sample Analysis	N	
VIII.	Laboratory Control Samples (LCS)	A	<u>LCS</u>
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	<u>Not utilized</u>
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	N	
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	N	
XV.	Field Blanks	ND	<u>FB=FB 04072010-RZC</u> <u>(250-2280-2)</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: 50.1

1	SSAN6-07-7BPC	11	<u>PODS</u>	21		31	
2	SSAN6-07-9BPC	12		22		32	
3	SSAN6-07-7BPCMS	13		23		33	
4	SSAN6-07-7BPCMSD	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 30, 2010
LDC Report Date: June 16, 2010
Matrix: Soil
Parameters: Arsenic
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-3100-1

Sample Identification

SSAK7-05-5BPC
SSAK7-05-1BPC
SSAK7-05-1BPC_FD
SSAQ5-01-5BPC
SSAQ5-01-1BPC
SSAQ5-01-1BPC-FD
SSAK7-05-5BPCMS
SSAK7-05-5BPCMSD

Introduction

This data review covers 8 soil samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 6020 for Arsenic.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blanks are summarized in Section IV.

Field duplicates are summarized in Section XIV.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. ICPMS Tune

The mass calibration was within 0.1 AMU and the percent relative standard deviation (%RSD) was less than or equal to 5% .

III. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

IV. Blanks

Method blanks were reviewed for each matrix as applicable. No arsenic was found in the initial, continuing and preparation blanks.

Samples FB0406010-RZB (from SDG 280-2131-1) and FB-04072010-RZD (from SDG 280-2216-2) were identified as field blanks. No arsenic was found in these blanks.

V. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

VI. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

X. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

XI. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XII. Sample Result Verification and Project Quantitation Limit

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3100-1	All analytes reported below the PQL.	J (all detects)	A

Raw data were not reviewed for this SDG.

XIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIV. Field Duplicates

Samples SSAK7-05-1BPC and SSAK7-05-1BPC_FD and samples SSAQ5-01-1BPC and SSAQ5-01-1BPC-FD were identified as field duplicates. No arsenic was detected in any of the samples with the following exceptions:

Compound	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAK7-05-1BPC	SSAK7-05-1BPC_FD				
Arsenic	3.0	3.3	10 (≤ 50)	-	-	-

Compound	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAQ5-01-1BPC	SSAQ5-01-1BPC-FD				
Arsenic	14	24	53 (≤ 50)	-	J (all detects)	A

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Data Qualification Summary - SDG 280-3100-1**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-3100-1	SSAK7-05-5BPC SSAK7-05-1BPC SSAK7-05-1BPC_FD SSAQ5-01-5BPC SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (PQL) (sp)
280-3100-1	SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	Arsenic	J (all detects)	A	Field duplicates (RPD) (fd)

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Laboratory Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Arsenic - Field Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson

VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308X4

SDG #: 280-3100-1

Laboratory: Test America

Stage 2B

Date: 6-15-10

Page: 1 of 1

Reviewer: CRZ

2nd Reviewer: L

METHOD: As (EPA SW 846 Method 6020)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/30/10
II.	ICP/MS Tune	A	
III.	Calibration	A	
IV.	Blanks	A	
V.	ICP Interference Check Sample (ICS) Analysis	A	
VI.	Matrix Spike Analysis	A	MS/D
VII.	Duplicate Sample Analysis	N	
VIII.	Laboratory Control Samples (LCS)	A	LCS
IX.	Internal Standard (ICP-MS)	A	
X.	Furnace Atomic Absorption QC	N	Not utilized
XI.	ICP Serial Dilution	A	
XII.	Sample Result Verification	N	
XIII.	Overall Assessment of Data	A	
XIV.	Field Duplicates	SW	(2,3), (5,6)
XV.	Field Blanks	ND	FB = FB0406010-RZB, FB-04072010-RZD (280-2131-1) (280-2216-2)

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: Soil

1	SSAK7-05-5BPC	11	PBS	21		31	
2	SSAK7-05-1BPC	12		22		32	
3	SSAK7-05-1BPC_FD	13		23		33	
4	SSAQ5-01-5BPC	14		24		34	
5	SSAQ5-01-1BPC	15		25		35	
6	SSAQ5-01-1BPC-FD	16		26		36	
7	SSAK7-05-5BPCMS	17		27		37	
8	SSAK7-05-5BPCMSD	18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC#: 23308X4
 SDG#: See Cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: CR
 2nd Reviewer: V

METHOD: Metals (EPA Method 6020/7000)

Y N NA Were field duplicate pairs identified in this SDG?
Y N NA Were target analytes detected in the field duplicate pairs?

Compound	Concentration (mg/Kg)		(≤ 50)	(mg/Kg)	(mg/Kg)	Qualifications (Parent Only)
	2	3	RPD	Difference	Limits	
Arsenic	3.0	3.3	10			

V:\FIELD DUPLICATES\FD_inorganic\23308X4.wpd

Compound	Concentration (mg/Kg)		(≤ 50)	(mg/Kg)	(mg/Kg)	Qualifications (Parent Only)
	5	6	RPD	Difference	Limits	
Arsenic	14	24	53			Jdet/A (fd)

**Tronox LLC Facility, PCS, Henderson, Nevada
Data Validation Reports
LDC #23308**

Perchlorate

LDC

Laboratory Data Consultants, Inc. Data Validation Report

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 15, 2010
LDC Report Date: June 23, 2010
Matrix: Soil
Parameters: Perchlorate
Validation Level: Stage 2B & 4
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2500-2

*Sample Identification

SA129-4BPC	SSAM4-02-2BPCMS
SA129-6BPC	SSAM4-02-2BPCMSD
SA129-8BPC	SSAM4-02-2BPCDUP
SA129-9BPC**	
SSAM5-02-2BPC	
SSAM5-02-4BPC	
SSAM5-02-6BPC**	
SSAM5-02-8BPC	
SSAM5-02-10BPC	
SSAM5-02-6BPC_FD	
SSAM4-02-2BPC**	
SSAM4-02-4BPC	
SSAM4-02-6BPC	
SSAM4-02-8BPC	
SSAM4-02-10BPC	
SSAN4-01-2BPC	
SSAN4-01-4BPC	
SSAN4-01-6BPC	
SSAN4-01-8BPC	
SSAN4-01-10BPC	

**Indicates sample underwent Stage 4 review

*Corrected first SSAN4-01-8BPC ID to SSAN4-01-6BPC

Introduction

This data review covers 22 soil samples listed on the cover sheet. The analyses were per EPA Method 314.0 for Perchlorate.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a Stage 4 review. A Stage 2B review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Stage 2B criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the initial, continuing and preparation blanks.

Samples FB-04132010-RIG2-RZE (from SDG 280-2400-2) and FB-04072010-RZC (from SDG 280-2280-2) were identified as field blanks. No perchlorate were found in these blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2500-2	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

VIII. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples SSAM5-02-6BPC** and SSAM5-02-6BPC_FD were identified as field duplicates. No perchlorate was detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAM5-02-6BPC**	SSAM5-02-6BPC_FD				
Perchlorate	440	520	17 (≤ 50)	-	-	-

***Tronox LLC Facility, PCS, Henderson, Nevada
Perchlorate - Data Qualification Summary - SDG 280-2500-2**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
*280-2500-2	SA129-4BPC SA129-6BPC SA129-8BPC SA129-9BPC** SSAM5-02-2BPC SSAM5-02-4BPC SSAM5-02-6BPC** SSAM5-02-8BPC SSAM5-02-10BPC SSAM5-02-6BPC_FD SSAM4-02-2BPC** SSAM4-02-4BPC SSAM4-02-6BPC SSAM4-02-8BPC SSAM4-02-10BPC SSAN4-01-2BPC SSAN4-01-4BPC SSAN4-01-8BPC SSAN4-01-6BPC SSAN4-01-10BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (sp)

**Tronox LLC Facility, 2009 Phase B Investigation, Henderson, Nevada
Perchlorate - Laboratory Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
Perchlorate - Field Blank Data Qualification Summary - SDG 280-2500-2**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B / 4

LDC #: 23308A6
 SDG #: 280-2500-2
 Laboratory: Test America

Date: 6-15-10
 Page: 1 of 1
 Reviewer: CR
 2nd Reviewer: W

METHOD: (Analyte) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

Validation Area		Comments	
I.	Technical holding times	A	Sampling dates: 4/15/10
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	MS/D
V	Duplicates	A	DUP
VI.	Laboratory control samples	A	LCS/D
VII.	Sample result verification	A	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	SW	(7,10)
X	Field blanks	ND	FB = FB-64132010-RIG2-RZE, FB-04072010-RZC (280-2400-2) (280-2280-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinsate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: soil ** Level 4

1	SA129-4BPC	11	SSAM4-02-2BPC **	21	SSAM4-02-2BPCMS	31	
2	SA129-6BPC	12	SSAM4-02-4BPC	22	SSAM4-02-2BPCMSD	32	
3	SA129-8BPC	13	SSAM4-02-6BPC	23	↓ DUP	33	
4	SA129-9BPC **	14	SSAM4-02-8BPC	24		34	
5	SSAM5-02-2BPC	15	SSAM4-02-10BPC	25		35	
6	SSAM5-02-4BPC	16	SSAN4-01-2BPC	26		36	
7	SSAM5-02-6BPC **	17	SSAN4-01-4BPC	27		37	
8	SSAM5-02-8BPC	18	SSAN4-01-6BPC	28		38	
9	SSAM5-02-10BPC	19	SSAN4-01-8BPC	29		39	
10	SSAM5-02-6BPC_FD	20	SSAN4-01-10BPC	30		40	

Notes: _____

LDC #: 23308A
 SDG #: see cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: CR
 2nd Reviewer: W

Method: inorganics (EPA Method 8000)

Validation Area	Yes	No	NA	Findings/Comments
All technical holding times were met.	/			
Cooler temperature criteria was met.	/			
Were all instruments calibrated daily, each set-up time?	/			
Were the proper number of standards used?	/			
Were all initial calibration correlation coefficients > 0.995?	/			
Were all initial and continuing calibration verification %Rs within the 90-110% QC limits?	/			
Were titrant checks performed as required? (Level IV only)			/	
Were balance checks performed as required? (Level IV only)			/	
Was a method blank associated with every sample in this SDG?	/			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.		/		
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.	/			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.	/			
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of ≤ CRDL (≤ 2X CRDL for soil) was used for samples that were ≤ 5X the CRDL, including when only one of the duplicate sample values were < 5X the CRDL.	/			
Was an LCS analyzed for this SDG?	/			
Was an LCS analyzed per extraction batch?	/			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% (85-115% for Method 300.0) QC limits?	/			
Were performance evaluation (PE) samples performed?		/	/	
Were the performance evaluation (PE) samples within the acceptance limits?		/	/	

LDC #: 23308A6
 SDG #: see cover

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: ER
 2nd Reviewer: W

Validation Area	Yes	No	NA	Findings/Comments
VII. Sample Result Validation				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	/			
Were detection limits < RL?	/			
VIII. Overall Assessment of Data				
Overall assessment of data was found to be acceptable.	/			
IX. Field Duplicates				
Field duplicate pairs were identified in this SDG.	/			
Target analytes were detected in the field duplicates.	/			
X. Field Blanks				
Field blanks were identified in this SDG.	/			
Target analytes were detected in the field blanks.		/		

LDC#: 23308A6
SDG#: See Cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: CR
2nd Reviewer: W

Inorganics, Method: See Cover

- Y N NA Were field duplicate pairs identified in this SDG?
- Y N NA Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (mg/Kg)		RPD (≤ 50)	Difference	Limits	Qualification (Parent only)
	7	10				
Perchlorate	440	520	17			

V:\FIELD DUPLICATES\FD_inorganic\23308A6.wpd

LDC #: 23308A6
 SDG #: seecaer

Validatin Findings Worksheet
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

Method: Inorganics, Method 34.0
 The correlation coefficient (r) for the calibration of ClO₄ was recalculated. Calibration date: 4/21/10

An initial or continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$\%R = \frac{\text{Found} \times 100}{\text{True}}$ Where, Found = concentration of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration of each analyte in the ICV or CCV source

Type of analysis	Analyte	Standard	Conc. (ug/l)	Area	Recalculated		Reported		Acceptable (Y/N)
					r	r ²	r	r ²	
Initial calibration	ClO ₄	s1	1	0.0025	0.998765	0.998771			
		s2	2.5	0.00841					
		s3	5	0.01661					
		s4	10	0.03291					
		s5	20	0.06345					
		s6	40	0.14097					
Calibration verification		ICV	20	Found 18.840	94	-	-		
Calibration verification		CCV	30	28.260	94	-	-		
Calibration verification		CCV	10	10.510	105	-	-		

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 2330898
 SDG #: seecover

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: ERS
 2nd Reviewer: LS

METHOD: Inorganics, Method Seecover

Percent recoveries (%R) for a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$\%R = \frac{\text{Found} \times 100}{\text{True}}$ Where, Found = concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).
 $\text{True} =$ concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$RPD = \frac{|S-D|}{(S+D)/2} \times 100$ Where, S = Original sample concentration
 D = Duplicate sample concentration

Sample ID	Type of Analysis	Element	Found / S (units) g/kg	True / D (units) g/kg	Recalculated		Reported %R / RPD	Acceptable (Y/N)
					%R	RPD		
LCS	Laboratory control sample	ClO ₄	0.107	0.0998	107	107	107	Y
22	Matrix spike sample	↓	(SSR-SR) 640	536	119	119	119	Y
23	Duplicate sample	↓	630	623	1	2	2	Y

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 23308A6
 SDG #: see over

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
 Reviewer: CF
 2nd reviewer: [Signature]

METHOD: Inorganics, Method see over

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Have results been reported and calculated correctly?
- Y N N/A Are results within the calibrated range of the instruments?
- Y N N/A Are all detection limits below the CRQL?

Compound (analyte) results for ClO₄ reported with a positive detect were recalculated and verified using the following equation:

Concentration = $\frac{(\text{Area} - \text{offset})}{\text{slope}} \times \text{Prep Factor (DF)} \times \frac{1}{\% \text{ Solid}}$

Recalculation: $\left(\frac{0.06901 + 0.0008}{0.0034} \right) / 1000 \times (1000) \times (10) = 230 \text{ mg/kg}$

(0.896)

#	Sample ID	Analyte	Reported Concentration (mg/kg)	Calculated Concentration (mg/kg)	Acceptable (Y/N)
	4	ClO ₄	230	230	Y

Note: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada

Collection Date: April 26, 2010

LDC Report Date: June 15, 2010

Matrix: Soil

Parameters: Perchlorate

Validation Level: Stage 2B

Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2879-7

Sample Identification

SSAR6-04-7BPC

SSAR6-04-9BPC

Introduction

This data review covers 2 soil samples listed on the cover sheet. The analyses were per EPA Method 314.0 for Perchlorate.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the initial, continuing and preparation blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No perchlorate was found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Analyte	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Perchlorate	92 ug/L	All samples in SDG 280-2879-7

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

IV. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) analyses specified for the samples in this SDG, and therefore matrix spike analyses were not performed for this SDG.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2879-7	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

VIII. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Data Qualification Summary - SDG 280-2879-7**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2879-7	SSAR6-04-7BPC SSAR6-04-9BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (sp)

**Tronox LLC Facility, 2009 Phase B Investigation, Henderson, Nevada
 Perchlorate - Laboratory Blank Data Qualification Summary - SDG 280-2879-7**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Field Blank Data Qualification Summary - SDG 280-2879-7**

No Sample Data Qualified in this SDG

**Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET**

LDC #: 23308J6
SDG #: 280-2879-7
Laboratory: Test America

Stage 2B

Date: 6-15-10
Page: 1 of 1
Reviewer: CR
2nd Reviewer: V

METHOD: (Analyte) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/26/10</u>
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	N	Client specified
V	Duplicates	N	
VI.	Laboratory control samples	A	LCS/D
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	N	
X.	Field blanks	SW	FB = FB04062010-RZB (280-231-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
N = Not provided/applicable R = Rinsate TB = Trip blank
SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAR6-04-7BPC	11	<u>PBS</u>	21		31	
2	SSAR6-04-9BPC	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 26, 2010
LDC Report Date: June 15, 2010
Matrix: Soil
Parameters: Perchlorate
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.
Sample Delivery Group (SDG): 280-2879-8

Sample Identification

SSAR6-04-10BPC
SSAR6-04-10BPCMS
SSAR6-04-10BPCMSD
SSAR6-04-10BPCDUP

Introduction

This data review covers 4 soil samples listed on the cover sheet. The analyses were per EPA Method 314.0 for Perchlorate.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the initial, continuing and preparation blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No perchlorate was found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Analyte	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Perchlorate	92 ug/L	All samples in SDG 280-2879-8

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2879-8	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

VIII. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Data Qualification Summary - SDG 280-2879-8**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2879-8	SSAR6-04-10BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (sp)

**Tronox LLC Facility, 2009 Phase B Investigation, Henderson, Nevada
 Perchlorate - Laboratory Blank Data Qualification Summary - SDG 280-2879-8**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Field Blank Data Qualification Summary - SDG 280-2879-8**

No Sample Data Qualified in this SDG

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET

LDC #: 23308K6
SDG #: 280-2879-8
Laboratory: Test America

Stage 2B

Date: 6-15-10
Page: 1 of 1
Reviewer: QC
2nd Reviewer: W

METHOD: (Analyte) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/26/10</u>
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	<u>MS/D</u>
V	Duplicates	A	<u>DUP</u>
VI.	Laboratory control samples	A	<u>LCS/D</u>
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	N	
X	Field blanks	SW	<u>FB = FB04062010 - RZB</u>

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

(280-2879-2)
D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: soil

1	SSAR6-04-10BPC	11	<u>PBS</u>	21		31	
2	SSAR6-04-10BPCMS	12		22		32	
3	SSAR6-04-10BPCMSD	13		23		33	
4	SSAR6-04-10BPCDUP	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 28, 2010
LDC Report Date: June 15, 2010
Matrix: Soil
Parameters: Perchlorate
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-2995-1

Sample Identification

SSAN6-07-1BPC
SSAN6-07-5BPC
SSAN6-07-1BPCMS
SSAN6-07-1BPCMSD
SSAN6-07-1BPCDUP

Introduction

This data review covers 5 soil samples listed on the cover sheet. The analyses were per EPA Method 314.0 for Perchlorate.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the initial, continuing and preparation blanks.

Sample FB-04072010-RZC (from SDG 280-2280-2) was identified as a field blank. No perchlorate was found in this blank.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-2995-1	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

VIII. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Data Qualification Summary - SDG 280-2995-1**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-2995-1	SSAN6-07-1BPC SSAN6-07-5BPC	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (sp)

**Tronox LLC Facility, 2009 Phase B Investigation, Henderson, Nevada
 Perchlorate - Laboratory Blank Data Qualification Summary - SDG 280-2995-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Field Blank Data Qualification Summary - SDG 280-2995-1**

No Sample Data Qualified in this SDG

LDC #: 23308P6
 SDG #: 280-2995-1
 Laboratory: Test America

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B

Date: 6-15-0
 Page: 1 of 1
 Reviewer: CR
 2nd Reviewer: V

METHOD: (Analyte) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>4/28/10</u>
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV.	Matrix Spike/Matrix Spike Duplicates	A	<u>MS/D</u>
V.	Duplicates	A	<u>DUP</u>
VI.	Laboratory control samples	A	<u>LCS/D</u>
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	N	
X.	Field blanks	NO	<u>FB = FB-04072010-RZC</u>

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

(280-2280-2)
 D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: SOIL

1	SSAN6-07-1BPC	11		21		31	
2	SSAN6-07-5BPC	12		22		32	
3	SSAN6-07-1BPCMS	13		23		33	
4	SSAN6-07-1BPCMSD	14		24		34	
5	SSAN6-07-1BPCDUP	15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: Tronox LLC Facility, PCS, Henderson, Nevada
Collection Date: April 30, 2010
LDC Report Date: June 15, 2010
Matrix: Soil
Parameters: Perchlorate
Validation Level: Stage 2B
Laboratory: TestAmerica, Inc.

Sample Delivery Group (SDG): 280-3100-1

Sample Identification

SSAQ5-01-5BPC
SSAQ5-01-1BPC
SSAQ5-01-1BPC-FD

Introduction

This data review covers 3 soil samples listed on the cover sheet. The analyses were per EPA Method 314.0 for Perchlorate.

This review follows the Standard Operating Procedures (SOP) 40, Data Review/Validation (BRC 2009), the Quality Assurance Project Plan Tronox LLC Facility, Henderson, Nevada (June 2009), NDEP guidance (May 2006), and a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004).

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- J+ Data are qualified as estimated, with a high bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J- Data are qualified as estimated, with a low bias likely to occur. False positives or false negatives are unlikely to have been reported.
- J Data are qualified as estimated; it is not possible to assess the direction of the potential bias. False positives or false negatives are unlikely to have been reported.
- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- R Data are qualified as rejected. There is a significant potential for the reporting of false negatives or false positives.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- B The analytical result may be a false positive totally attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JB The analytical result may be biased high and partially attributable to blank contamination. This qualifier is applicable to radiochemistry analysis only.
- JK The analytical result is an estimated maximum possible concentration (EMPC).
- X The analytical result is not used for reporting because a more accurate and precise result is reported in its place.
- J-TDS The analytical result is estimated based on failure of the Total Dissolved Solids (TDS) correctness check performed in accordance with the Standard Method 1030E.
- J-CAB The analytical result is estimated based on failure of the cation-anion balance correctness check performed in accordance with Standard Method 1030E.
- J-TDS & CAB The analytical result is unreliable based on the failure of the cation-anion balance and TDS correctness check performed in accordance with standard Method 1030E.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the initial, continuing and preparation blanks.

Sample FB04062010-RZB (from SDG 280-2131-2) was identified as a field blank. No perchlorate was found in this blank with the following exceptions:

Field Blank ID	Sampling Date	Analyte	Concentration	Associated Samples
FB04062010-RZB	4/6/10	Perchlorate	92 ug/L	All samples in SDG 280-3100-1

Sample concentrations were compared to concentrations detected in the field blanks as required by the QAPP. No sample data was qualified.

IV. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) analyses specified for the samples in this SDG, and therefore matrix spike analyses were not performed for this SDG.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification and Project Quantitation Limit

All sample result verifications were acceptable for samples on which a Stage 4 review was performed.

All analytes reported below the PQL were qualified as follows:

Sample	Finding	Flag	A or P
All samples in SDG 280-3100-1	All analytes reported below the PQL.	J (all detects)	A

Raw data were not evaluated for the samples reviewed by Stage 2B criteria.

VIII. Overall Assessment

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples SSAQ5-01-1BPC and SSAQ5-01-1BPC-FD were identified as field duplicates. No perchlorate was detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/Kg)		RPD (Limits)	Difference (Limits)	Flags	A or P
	SSAQ5-01-1BPC	SSAQ5-01-1BPC-FD				
Perchlorate	210	190	10 (≤ 50)	-	-	-

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Data Qualification Summary - SDG 280-3100-1**

SDG	Sample	Analyte	Flag	A or P	Reason (Code)
280-3100-1	SSAQ5-01-5BPC SSAQ5-01-1BPC SSAQ5-01-1BPC-FD	All analytes reported below the PQL.	J (all detects)	A	Sample result verification (sp)

**Tronox LLC Facility, 2009 Phase B Investigation, Henderson, Nevada
 Perchlorate - Laboratory Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

**Tronox LLC Facility, PCS, Henderson, Nevada
 Perchlorate - Field Blank Data Qualification Summary - SDG 280-3100-1**

No Sample Data Qualified in this SDG

LDC #: 23308X6
 SDG #: 280-3100-1
 Laboratory: Test America

Tronox Northgate Henderson
VALIDATION COMPLETENESS WORKSHEET
 Stage 2B

Date: 6-15-10
 Page: 1 of 1
 Reviewer: OR
 2nd Reviewer: W

METHOD: (Analyte) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/30/10
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	N	Client specified
V	Duplicates	N	↓
VI.	Laboratory control samples	A	LCS/D
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	SW	(2,3)
X	Field blanks	SW	FB = FB04062010-RZB (280-2131-2)

Note: A = Acceptable ND = No compounds detected D = Duplicate
 N = Not provided/applicable R = Rinstate TB = Trip blank
 SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: Soil

1	SSAQ5-01-5BPC	11	PPS	21		31	
2	SSAQ5-01-1BPC	12		22		32	
3	SSAQ5-01-1BPC-FD	13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: _____

LDC#: 23308X6
SDG#: See Cover

VALIDATION FINDINGS WORKSHEET

Field Duplicates

Page: 1 of 1
Reviewer: CR
2nd Reviewer: _____

Inorganics, Method: See Cover

- Y N NA Were field duplicate pairs identified in this SDG?
 Y N NA Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (mg/Kg)		RPD (≤ 50)	Difference	Limits	Qualification (Parent only)
	2	3				
Perchlorate	210	190	10			

V:\FIELD DUPLICATES\FD_inorganic\23308X6.wpd