

TABLE 2-2
Data Validation Qualifier Reason Codes

Reason Code	Explanation
a	qualified due to low abundance (radiochemical activity)
be	qualified due to equipment blank contamination
bf	qualified due to field blank contamination
bl	qualified due to lab blank contamination
bt	qualified due to trip blank contamination
bp	qualified due to pump blank contamination (wells w/o dedicated pumps, when contamination is detected in the Pump Blk)
br	qualified due to filter blank contamination (aqueous Hexavalent Chromium and Dissolved sample fractions)
c	qualified due to calibration problems
cp	qualified due to insufficient ingrowth (radiochemical only)
dc	duel column confirmation %D exceeded
e	concentration exceeded the calibration range
fd	qualified due to field duplicate imprecision
h	qualified due to holding time exceedance
i	qualified due to internal standard areas
k	qualified as Estimated Maximum Possible Concentrations (dioxins and PCB congeners)
l	qualified due to LCS recoveries
ld	qualified due to lab duplicate imprecision (matrix duplicate, MSD, LCSD)
m	qualified due to matrix spike recoveries
nb	qualified due to negative lab blank contamination (nondetect results only)
o	other
p	qualified as a false positive due to contamination during shipping
pH	sample preservation not within acceptance range
q	qualified due to quantitation problem
s	qualified due to surrogate recoveries
sd	serial dilution did not meet control criteria
sp	detected value reported >SQL <PQL
st	sample receipt temperature exceeded
t	qualified due to elevated helium tracer concentrations
vh	volatile headspace detected in aqueous sample containers submitted for VOC analysis
x	qualified due to low % solids
z	qualified due to ICS results