

Table E-5
Qualifications Based on Calibration Criteria Exceedances
Phase B Source Area Soil Gas Investigation
Tronox Facility - Henderson, Nevada

Sample ID	SDG	Method	Matrix	Analyte	Result	Qualifiers	Units	Reason	DQI	DQI Result	DQI Limit
SG38B-20	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.16	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG40B-05	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.16	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG40B-05D	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.16	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG41B-20	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.11	J	ug/m3	c	ICAL %RSD	30.54%	<30%
SG41B-20D	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.16	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG43B-05	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.19	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG64B-05	P0801385	TO-15	GS	1,2-Dichlorobenzene	0.20	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG83B-05-1	P0801342	TO-15	GS	1,2-Dichlorobenzene	16	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG83B-05-3	P0801342	TO-15	GS	1,2-Dichlorobenzene	17	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%
SG83B-05-7	P0801342	TO-15	GS	1,2-Dichlorobenzene	17	UJ	ug/m3	c	ICAL %RSD	30.54%	<30%

Note:
Reason codes are defined in Table E-2
Data qualifiers are defined in Table E-1