

Data Validation Summary Report
Additional Pre-Confirmation Sampling
Nevada Environmental Response Trust Site
Henderson, Nevada

April 14, 2011

Prepared For:


ENVIRON International Corporation
6001 Shellmound Street, Suite 700
Emeryville, California 94608

Prepared By:

Northgate Environmental Management, Inc.
300 Frank H. Ogawa Plaza, Suite 510
Oakland, California 94612



Derrick Willis
Principal



Cynthia Arnold
Senior Project Chemist



**Data Validation Summary Report
Additional Pre-Confirmation Sampling
Nevada Environmental Response Trust Site
Henderson, Nevada**

Responsible Certified Environmental Manager (CEM) for this project

I hereby certify that all laboratory analytical data was generated by a laboratory certified by the NDEP for each constituent and media presented herein.

I hereby certify that I am responsible for the services described in this document and for the preparation of this document. The services described in this document have been provided in a manner consistent with the current standards of the profession and, to the best of my knowledge, comply with all applicable federal, state and local statutes, regulations and ordinances.



Derrick Willis, CEM 2252 Exp.: 9/12/12
Northgate Environmental Management, Inc.



TABLE OF CONTENTS

| | |
|---|------------|
| ACRONYMS AND ABBREVIATIONS..... | iii |
| 1.0 INTRODUCTION..... | 1 |
| 2.0 DATA VALIDATION PROCESS..... | 2 |
| 2.1 Data Deliverables..... | 2 |
| 2.2 Validation of Analytical Deliverables | 3 |
| 3.0 DATA VALIDATION RESULTS..... | 4 |
| 3.1 Holding Times and Sample Preservation..... | 5 |
| 3.2 Initial and Continuing Calibration | 5 |
| 3.3 Serial Dilution | 5 |
| 3.4 Laboratory Blanks/Equipment Blanks/Field Blanks | 5 |
| 3.5 LCS/LCSD Results | 6 |
| 3.6 MS/MSD Results | 6 |
| 3.7 Surrogate Recoveries | 6 |
| 3.8 Internal Standard Performance | 7 |
| 3.9 Field Duplicate Results | 7 |
| 3.10 Quantitation Problems | 8 |
| 3.11 Professional Judgment | 8 |
| 4.0 EVALUATION OF QUALITY INDICATORS..... | 9 |
| 4.1 Precision..... | 9 |
| 4.2 Accuracy | 9 |
| 4.3 Representativeness | 10 |
| 4.4 Completeness | 10 |
| 4.5 Comparability | 11 |
| 4.6 Sensitivity | 11 |
| 5.0 CONCLUSIONS | 12 |
| 6.0 REFERENCES..... | 13 |



TABLES – Provided on DVD

- 1-1 Sample Analysis Summary
- 1-2 Field Sample IDs and Laboratory SDGs
- 1-3 Sample Delivery Groups and LDC Validation Reports
- 2-1 Data Validation Qualifiers
- 2-2 Data Validation Qualifier Reason Codes
- 3-1 Qualifications Based on Hold Times and Sample Preservation Exceedances
- 3-2 Qualifications Based on Calibration Criteria Exceedances
- 3-3 Qualifications Based on Serial Dilution Results
- 3-4 Qualifications Based on Blank Contamination
- 3-5 Qualifications Based on Laboratory Control Sample Recoveries
- 3-6 Qualifications Based on Matrix Spike Recoveries
- 3-7 Qualifications Based on Surrogate Recoveries
- 3-8 Qualifications Based on Internal Standard Performance
- 3-9 Qualifications Based on Field Duplicate Precision
- 3-10 Qualifications Based on Quantitation Problems
- 3-11 Qualifications Based on Professional Judgment
- 3-12 Summary of Rejected Data

APPENDICES – Provided on DVD

- A Laboratory Reports
- B Validation Reports
- C Electronic Data Deliverable



ACRONYMS AND ABBREVIATIONS

| Acronym | Meaning |
|---------|---|
| %D | Percent Difference |
| BEC | Basic Environmental Company |
| BRC | Basic Remediation Company |
| CEM | Certified Environmental Manager |
| CLP | Contract Laboratory Program |
| DOE | Department of Energy |
| DQI | Data Quality Indicator |
| DUP | Duplicate |
| EDD | Electronic Data Deliverable |
| EDXA | Energy Dispersive X-ray Analysis |
| EPA | U.S. Environmental Protection Agency |
| GC/MS | Gas Chromatograph/Mass Spectrometer |
| ICP | Inductively Coupled Plasma |
| LCS | Laboratory Control Sample |
| LCSD | Laboratory Control Sample Duplicate |
| LDC | Laboratory Data Consultants |
| MDL | Method Detection Limit |
| MS/MSD | Matrix Spike/Matrix Spike Duplicate |
| NDEP | Nevada Division of Environmental Protection |
| PAH | Polynuclear Aromatic Hydrocarbons |
| PARCCS | Precision, Accuracy, Representativeness, Comparability, Completeness, and Sensitivity |
| PCB | Polychlorinated Biphenyl |
| PQL | Practical Quantitation Limit |
| QAPP | Quality Assurance Project Plan |
| QC | Quality Control |
| R | Rejected |
| RPD | Relative Percent Difference |
| SAED | Selected Area Electron Diffraction |
| SAP | Sampling and Analysis Plan |
| SDG | Sample Delivery Group |
| SOP | Standard Operating Procedure |
| SQL | Sample Quantitation Limit |
| SVOC | Semivolatile Organic Compound |
| TEM | Transmission Electron Microscope |
| Tronox | Tronox LLC |



1.0 INTRODUCTION

On behalf of the Nevada Environmental Response Trust, Northgate Environmental Management, Inc. (Northgate) is submitting this Data Validation Summary Report to ENVIRON International Corporation. The purpose of the report is to assess the validity (based on data validation) and usability (based on project objectives) of the Additional Pre-Confirmation sampling at the Nevada Environmental Response Trust Site, Henderson, NV (Site), conducted by Northgate between April and November 12, 2010.

Additional Pre-Confirmation samples were collected and analyzed in accordance with the *Revised Pre-Confirmation Sampling Work Plan, Remediation Zones RZ-A through RZ-E, Phase B Investigation, Henderson Nevada, March 2010* (Northgate 2010). Additional Pre-Confirmation soil samples collected from Remediation Zones RZ-B, RZ-C, RZ-D and RZ-E resulted in the analysis of 2,584 environmental and 559 quality control (QC) samples (trip blank, field blank, equipment blank, field duplicate, and matrix spike [MS]/MS duplicate [MSD] analysis). The sampling and analysis summary of the environmental and associated field QC samples is presented in Table 1-1. The sampling and analysis strategy to delineate the Site prior to remediation is detailed in the *Revised Pre-Confirmation Soil Sampling Work Plan* (Northgate 2010).

All Additional Pre-Confirmation samples were logged into the laboratories in Sample Delivery Groups (SDGs). The Additional Pre-Confirmation soil data are contained in 249 SDGs presented in Table 1-2. Analytical services were provided by four laboratories for the analytical groups summarized below.

| Laboratory | Location | Analytical Group(s) |
|------------------------|---------------------|------------------------------|
| EMS Laboratories, Inc. | Pasadena, CA | Asbestos |
| EMSL Analytical | Westmont, NJ | Asbestos |
| Test America | Denver, CO | SVOC, Metals and Perchlorate |
| Test America | West Sacramento, CA | Dioxin/Furans |

The analytical data were validated by Laboratory Data Consultants, Inc. (LDC) in accordance with procedures described in the Nevada Division of Environmental Protection (NDEP) *Data Verification and Validation Requirements – Supplement, Henderson, Nevada, April 13, 2009*, established for the BMI Plant Sites and Common Areas Projects. The association between the laboratory SDGs and LDC validation reports is presented in Table 1-3.



2.0 DATA VALIDATION PROCESS

A formal validation of the Additional Pre-Confirmation sample analytical results was performed to assess the extent of remediation required at the Site. Consistent with the *Revised Pre-Confirmation Sampling Work Plan*, the *Tronox Quality Assurance Project Plan* (QAPP; AECOM/Northgate 2009), and *NDEP Supplemental Guidance on Data Validation for the BMI Plant Sites and Common Areas Projects* (NDEP Supplemental Guidance; NDEP 2009d), all of the Additional Pre-Confirmation soil data were validated. Approximately 90% of the analytical data were validated as Stage 2B and approximately 10% were validated by Stage 4 data validation procedures. EPA Stage 2B (EPA 2009) validation evaluates the following QC criteria:

- Completeness of deliverable;
- Technical holding times and sample preservation;
- Sample integrity and cooler/sample temperature at the time of laboratory receipt;
- Laboratory and field blank contamination;
- Surrogate spike recoveries;
- MS/MSD recoveries and relative percent differences (RPDs);
- Laboratory duplicate RPDs;
- Laboratory control sample (LCS) recoveries; and
- Initial and continuing calibrations.

The comprehensive validation, consistent with EPA designation of Stage 4 (EPA 2009), involves in-depth review of compound identification and quantification, spot-checks of calculations, and verification of summary data against the raw data. Table 1-3 is a cross-reference of laboratory SDG and associated validation reports. Field samples presented with shading were validated as Stage 4 (EPA 2009).

2.1 Data Deliverables

Analytical data deliverables were provided as an electronic data deliverable (EDD) version of the full data package, equivalent to a Contract Laboratory Program (CLP) deliverable (i.e., consisting of all the information required in a CLP package, including CLP-like summary forms). The electronic data packages were presented in PDF format with embedded text wherever possible and include complete bookmarking for all forms, tables, and sections. Each data package was also delivered as an EDD.



Asbestos deliverables included sample results, a case narrative, chain-of-custody, QC summary data, sample prep data, transmission electron microscope (TEM) calibration data (chrysotile beam dose sensitivity, camera constant calibrations, crocidolite spectrum Na sensitivity, Mg-Si K-alpha peak resolvability, K factors, and detector resolution of the Mn K-alpha peak), one energy dispersive x-ray analysis (EDXA) and one selected area electron diffraction (SAED) image per asbestos type per sample, filter blank lot data (4%), lab blanks, method blanks, equipment blanks, and all analyst worksheets. The analytical reports for the Additional Pre-Confirmation soil data collected at the Site are presented in Appendix A.

In addition to the laboratory deliverables, field information was provided to the validation staff in order to associate the field QC samples (field blanks, equipment blanks, and field duplicates) with the primary field samples prior to validation.

2.2 Validation of Analytical Deliverables

Validation of the Additional Pre-Confirmation sample data was performed by LDC using the appropriate EPA guidelines (EPA 1999, 2004, 2008, 2009) or equivalent regional EPA validation guidelines such as Region 9 Superfund Data Evaluation/Validation Guidance, R9QA/006.1 (EPA 2001), the *NDEP Supplemental Guidance* (NDEP 2009b, 2009c, 2009d, 2009e) and the *Basic Remediation Company (BRC) SOP 40, Data Review/Validation* (BRC SOP; BRC 2009). The federal EPA guidelines, prepared for CLP data, were adapted to reflect the analytical methods and measurement quality objectives established for the Additional Pre-Confirmation soil methods and the guidance provided by NDEP. LDC validation reports for the Additional Pre-Confirmation soil data are presented in Appendix B.

Analytical data deficiencies were qualified using the data validation qualifiers in Table 2-1 and project-specific reason codes shown in Table 2-2. The finalized EDD, prepared in accordance with NDEP requirements (NDEP 2010) for Additional Pre-Confirmation sample data is presented in Appendix C.



3.0 DATA VALIDATION RESULTS

The data validation qualifiers and reason codes were used to indicate all the data in the database where results were qualified as a result of validation. This information was sorted by the QC review elements listed below:

- Holding times and sample preservation;
- Initial and continuing calibrations;
- Serial dilution;
- Laboratory blanks/equipment blanks/field blanks;
- LCS/Laboratory Control Sample Duplicate (LCSD) results;
- MS/MSD results;
- Surrogate recoveries;
- Internal standard performance;
- Laboratory duplicate results;
- Field duplicate results; and
- Quantitation problems.

Tables 3-1 through 3-12 present the qualified results based on QC deficiencies identified during the validation process. Reason codes for each qualifier assignment have been provided in the tables. Where available, a numerical data quality indicator (DQI) result value and acceptance criteria for that value have been added to the table in columns to the right of the reason codes per NDEP's request. No QC problems were identified that resulted in qualification of results based on mass spectrometer tuning, gas chromatograph/mass spectrometer (GC/MS) performance checks, compound identification, or peak integration. A summary of rejected results is presented in Table 3-12. The data validation summary tables are sorted by Remediation Zone, Sample ID and SDG to assist the data user in locating the associated data validation memorandum. The data validation memoranda presented in Appendix B discuss the application of qualifiers in detail. Tables 3-1 through 3-12 are provided to NDEP on CD as an Excel spreadsheet that can be sorted to assist the data user in locating validation information for any particular sample, SDG, method, or analyte.



3.1 Holding Times and Sample Preservation

Sample preservation and analytical holding times are evaluated to assure that the sample integrity is intact for accurate sample preparation and analysis. Sample preservation and analytical hold time are presented for each method of analysis in Table B-1 of the QAPP. Holding time exceedances can cause loss of sample constituents due to biodegradation, precipitation, volatilization, and chemical degradation. In accordance with EPA guidance (USEPA 2004, 2008), sample results for organic and non-metal analyses that were performed after the method holding time but less than two times the method holding time are qualified as estimated (J- or UJ) and results for analyses performed after two times the method holding time are qualified as rejected (R). Inorganic hold time exceedances are qualified as estimated J- or R.

Sample results were qualified as estimated (J/UJ) due to hold time exceedances. Less than 1% (0.13%) of the Additional Pre-Confirmation sample data were qualified due to hold time and preservation exceedances, as presented in Table 3-1. No data were rejected.

3.2 Initial and Continuing Calibration

Instrument performance was evaluated during the review of initial and continuing calibration for each method analyzed. The following target analytes exhibited poor performance: Method 8081A 4,4'-DDD, endrin ketone and endosulfan sulfate, Method 8260B tert-butyl alcohol and dichlorodifluoromethane, Method 8270C bis(2-ethylhexyl) phthalate, and Method 8290 OCDD and OCDF. Less than 1% (0.39%) of the Additional Pre-Confirmation sample data were qualified as (J/UJ) due to calibration deficiencies, as presented in Table 3-2. No data were rejected.

3.3 Serial Dilution

Sample matrix interference was exhibited by several target analytes. Arsenic, cobalt and nickel resulted in serial dilution exceedances greater than 2X the acceptance limit of 10% Difference (%D). In accordance with EPA guidance (EPA 2004), the associated results were qualified as estimated (J). Less than 2 % (1.46%) of the Additional Pre-Confirmation sample data were qualified due to serial dilution exceedances, as presented in Table 3-3. No data were rejected.

3.4 Laboratory Blanks/Equipment Blanks/Field Blanks

The Additional Pre-Confirmation soil data were assessed using the following blanks: field blanks, equipment blanks, trip blanks, and laboratory method blanks. Equipment blanks were collected at a frequency of 5% during the Additional Pre-Confirmation sampling, and one field



blank was collected for each investigative remediation zone per matrix. Data were evaluated and qualified in accordance with EPA guidance (EPA 2004, 2008), NDEP Supplemental Guidance on Data Validation for the BMI Plant Sites and Common Areas Projects (NDEP 2009c,d,e), and the BRC SOP 40, Data Review Validation, May 7, 2009 (BRC 2009). Approximately 2% (1.52%) of the Additional Pre-Confirmation sample data were qualified based on blank contamination, as presented in Table 3-4.

3.5 LCS/LCSD Results

Laboratory control samples and laboratory control sample duplicates were used to assess laboratory accuracy. The Additional Pre-Confirmation soil data were evaluated in accordance with the BRC SOP 40, Data Review Validation, May 7, 2009. All data exceedances were qualified as estimated (J/UJ). Less than 1% (0.03%) of the Additional Pre-Confirmation sample data were qualified due to LCS/LCSD precision and accuracy exceedances, as presented in Table 3-5. No data were rejected.

3.6 MS/MSD Results

Matrix spike and matrix spike duplicate samples consist of aliquots of environmental samples spiked with a subset of target compounds. MS/MSD samples monitor potential interference from the site-specific sample matrix and its effect on target compounds. Additional field sample aliquots were collected at a frequency of 5% during the Additional Pre-Confirmation sampling to evaluate site-specific matrix interference. Samples were evaluated using the EPA guidance (EPA 2004, 2008), NDEP Supplemental Guidance on Data Validation for the BMI Plant Sites and Common Areas Projects (NDEP 2009c,d,e), the BRC SOP (BRC 2009), and professional judgment.

Data were qualified as estimated (J/UJ) for MS and/or MSD precision and accuracy failure outside of the acceptance limit criteria. Less than 1% (0.26%) of the Additional Pre-Confirmation sample data were qualified due to MS/MSD exceedances, as presented in Table 3-6. No data were rejected.

3.7 Surrogate Recoveries

Surrogate recoveries were reviewed for organic methods: organochlorine pesticides (OCP) Method 8081A, volatile organic compounds (VOCs) Method 8260B, semi-volatile organic compounds (SVOCs) Method 8270C and dioxin/furan Method 8290. Organic data were evaluated using the EPA guidance (EPA 2004, 2008), NDEP Supplemental Guidance on Data



Validation for the BMI Plant Sites and Common Areas Projects (NDEP 2009c,d,e), and the BRC SOP (BRC 2009). All data were usable and qualified as estimated (J/UJ) for surrogate recovery exceedances. Approximately 2% (1.86%) of the Additional Pre-Confirmation sample data were qualified due to surrogate recovery exceedances, as presented in Table 3-7. No data were rejected.

3.8 Internal Standard Performance

Internal standards were prepared for certain organic and inductively coupled plasma (ICP)/MS analyses by adding compounds similar to target compounds of interest to sample aliquots. Internal standards are used in the quantitation of target compounds in the sample or sample extract. Internal standards were reviewed using the EPA guidance (EPA 2008), NDEP Supplemental Guidance (NDEP 2009c,d,e), and the *BRC SOP*. All data were usable and qualified as estimated (J/UJ) for internal standard exceedances with the exception of six rejected (R) SVOC analytes in sample SSAO8-05-0BPC where area counts were less than 25% of the associated CCV. Approximately 4% (4.06%) of the Additional Pre-Confirmation sample data were qualified due to internal standard performance exceedances, as presented in Table 3-8. A summary of rejected (R) Additional Pre-Confirmation data are presented in Table 3-12.

3.9 Field Duplicate Results

Field duplicates are used to evaluate sampling technique precision and homogeneity of the sample matrix. Field duplicates were collected at a frequency of 5% during the Additional Pre-Confirmation sampling. In accordance with the QAPP, NDEP Supplemental Guidance (NDEP 2009c,d,e), and the BRC SOP, the precision goal for field duplicate analyses was ± 50 percent RPD. If the field duplicate RPD exceeds the 50 percent limit, non-detected sample results shall be qualified as estimated (UJ) at the sample quantitation limit (SQL) and detected results shall be qualified as estimated (J). The RPD will be calculated using the reporting limit for non-detected sample results. Similar to analytical duplicates, this limit does not apply when the result for either the sample or its duplicate is less than five times the practical quantitation limit (PQL). For this situation, the absolute value of the PQL is to be used as the control limit.

Field duplicate exceedances were qualified as estimated (J/UJ). Less than 2% (1.20%) of the Additional Pre-Confirmation sample data were qualified due to field duplicate exceedances, as presented in Table 3-9. No data were rejected.



3.10 Quantitation Problems

During Stage 4 evaluation, all raw data were reviewed to confirm target analyte identification and quantitation. Results were qualified using method-specific criteria and EPA guidance (EPA 2004, 2008). Data for Method 8081A OCP, Method 8270C SVOC and Method 8290 dioxin/furan were qualified as estimated (J) for greater than 40 percent difference during second column confirmation, coeluting isomers, or an exceedance of the calibration range. Approximately 4% (4.01%) of the data were qualified due to sample quantitation issues, as presented in Table 3-10. No data were rejected.

3.11 Professional Judgment

Professional judgment was used to evaluate and qualify Method 8260 results for the 19 field samples listed in Table 3-11. The associated laboratory narrative documented that the acceptable freezer temperature (-20.5 to -10.5 degrees C) was compromised for an 11 hour period, rising to 4 degrees C during an outage. The associated Method 8260 sample data were qualified as estimated (J/UJ). Approximately 3% (3.16%) of the data were qualified using professional judgment, as presented in Table 3-11. No data were rejected.



4.0 EVALUATION OF QUALITY INDICATORS

Data quality indicators of precision, accuracy, representativeness, comparability, completeness, and sensitivity (PARCCS) were used to verify that sampling and analytical systems used in support of project activities are effective and that the quality of the data generated for the project is appropriate for making decisions affecting future activities. This section discusses the DQIs for the Additional Pre-Confirmation soil dataset. DQIs address the field and analytical data quality aspects as they affect uncertainties in the data collected for site characterization and risk assessment. The PARCCS parameters definition and assessment are presented in the *Tronox Revised Phase B QAPP* (Revised QAPP; AECOM/Northgate 2009), and the *Revised Pre-Confirmation Sampling Work Plan* (Northgate 2010). All data not meeting the established PARCCS criteria were qualified during the validation process using the guidelines presented in the QAPP, *National Functional Guidelines* (EPA 2004, 2005, 2008), BRC SOP, each analytical method employed, and professional judgment.

4.1 Precision

Precision is the measure of agreement among repeated measurements of the same property under identical or substantially similar conditions. Field precision was assessed through the collection and measurement of field duplicates and expressed as the RPD of the sample and field duplicate pair results. The assessment of field duplicate precision is discussed in Section 3.9 of this report, and is listed in Table 3-9. In general, field duplicate precision was acceptable for all analytes. No data were rejected.

Laboratory precision evaluates DQIs such as calibration, surrogates, MS/MSD, duplicate (DUP), LCS/LCSD and interference check samples previously discussed in Section 3 of this report. All laboratory precision was acceptable with exception of those noted in Sections 3.2, 3.5, 3.6, 3.7 and 3.9. No data were rejected.

4.2 Accuracy

Accuracy is the degree of agreement between an observed value and an accepted reference or true value. Laboratory accuracy was assessed during the validation using the recoveries of following QC parameters:

- Holding times and sample temperatures;
- Calibration;
- Serial dilution recovery (inorganics);



- Blank sample results;
- LCS percent recovery;
- MS/MSD percent recovery (organics); and
- Surrogate spike recovery.

Accuracy was evaluated for each of the DQIs in Sections 3.1 through 3.7. Evaluation of the Stage 4 QC elements that contribute to accuracy – such as mass spectrometer tuning, compound or element identification, peak integration and mass spectral matches, and calculation/transcription verifications did not result in the qualification or rejection of any data during validation.

4.3 Representativeness

Representativeness is a qualitative parameter defined by the degree to which data accurately and precisely represents a characteristic of a population, parameter variations at a sampling point, or a process or environmental condition. There is no formula for evaluating representativeness. Aspects of representativeness addressed during validation include the review of sample collection information in the chain-of-custody documentation, conformity of laboratory analyses to Work Plan intentions, adherence of the documented laboratory procedures to method requirements, and completeness of the laboratory data packages. All representativeness deficiencies were resolved during the actual field sampling event and/or data validation process. No qualification was necessary based on representativeness.

4.4 Completeness

Completeness is a measure of the amount of valid data obtained from a measurement system, compared to the amount expected under normal conditions. “Normal conditions” are defined as the conditions expected if the program specific work plan was implemented as proposed.

Field completeness is defined as the percentage of samples actually collected versus those intended to be collected per the Work Plan. The field completeness goal established in the QAPP is 90%. The Additional Pre-Confirmation dataset was collected beyond the locations proposed in the *Revised Pre-Confirmation Sampling Work Plan* (Northgate 2010) with the database sample IDs indicates that actual field completeness was 100%, exceeding the goal established for the project. Field completeness was assessed using the total sample locations scheduled in the *Revised Pre-Confirmation Sampling Work Plan* (Northgate 2010) compared to actual number submitted for analysis.



Laboratory completeness is defined as percentage of valid data points versus the total expected from the laboratory analyses. Valid data are defined as all the data points judged to be usable (i.e., not rejected as a result of the validation process). The laboratory completeness goal established in the QAPP is 95%. Actual laboratory completeness was 100% on the basis of sample analysis (i.e., all requested analyses were performed and reported by the laboratories), and 99.99% completeness based on valid data, with 0.01% of the data qualified as rejected (R) as described in Section 3.8 and summarized in Table 3-12.

4.5 Comparability

Comparability is a qualitative expression of the measure of confidence that two or more data sets may contribute to a common analysis. Comparability of data within the Additional Pre-Confirmation soil dataset was maximized by using standard methods for sampling and analysis, reporting data, and data validation.

4.6 Sensitivity

Sensitivity is the capability of a method to discriminate an actual deflection or response above instrument noise. For the EPA methods employed in this project, sensitivity is measured by the Method Detection Limit (MDL) and PQL. Both nominal MDLs and PQLs were provided by the laboratories in the laboratory data packages and were verified during validation. MDLs in general were adjusted for each Additional Pre-Confirmation soil sample to include the necessary dilution factors, preparation factors, and dry-weight factors of an individual sample as the SQL. The sensitivity requirements were based on the laboratory's ability to detect and report consistent and reliable limits.



5.0 CONCLUSIONS

One hundred percent of the laboratory data for the Additional Pre-Confirmation soil samples were validated using standardized guidelines and procedures recommended by EPA and NDEP. Based on the validated data, 99.99% of the results for Additional Pre-Confirmation soil data were determined usable and considered valid for all decision-making purposes.

A subset of the laboratory results was qualified during validation and those results are summarized in Tables 3-1 through 3-11. Qualified data are grouped by QC deficiency. A summary of the rejected Additional Pre-Confirmation soil data is presented in Table 3-12. Approximately 0.01% of the data were rejected. Data qualifiers and qualifier reason codes are presented in Table 2-1 and 2-2, respectively.

All the qualified results were evaluated with respect to the data quality indicators and compared to the QAPP and *Revised Pre-Confirmation Sampling Work Plan* (Northgate 2010). Details of this evaluation are discussed in Section 4 of this report. Based on the results of data validation, actual laboratory completeness was 100% on the basis of sample analysis, and 99.99% completeness based on valid data. The overall goals for data quality were achieved for the Additional Pre-Confirmation soil dataset.



6.0 REFERENCES

- AECOM, Inc. *Phase B Source Area Investigation Work Plan, Area I (Northern LOUs)*, Tronox LLC Facility. Henderson, Nevada. April 2008.
- AECOM, Inc. *Revised Phase B Site Investigation Work Plan, Text, Tables and Figures*. Tronox LLC Facility. Henderson, Nevada. December 2008.
- AECOM, Inc.; Northgate Environmental Management, Inc.; *Revised Phase B Quality Assurance Project Plan Tronox LLC Facility*. Henderson, Nevada (QAPP). Revision, July 2009.
- Basic Remediation Company (BRC); ERM-West (ERM) and MWH. *BRC Field Sampling and Standard Operating Procedures*. BMI Common Areas. Clark County, Nevada. Revision 3. December 2008.
- Basic Remediation Company (BRC). *BRC Standard Operating Procedure (SOP) 40. Data Review Validation*. Revision 4. May 7, 2009.
- Nevada Division of Environmental Protection (NDEP). *NDEP Detection Limits and Data Reporting for the BMI Plant Sites and Common Areas Projects*. Henderson, Nevada. December 2008.
- Nevada Division of Environmental Protection (NDEP). *NDEP Guidance on Uniform Electronic Deliverables for the BMI Plant Sites and Common Areas Projects*. Henderson, Nevada. February 27, 2009(a).
- Nevada Division of Environmental Protection (NDEP). *NDEP Supplemental Guidance on Data Validation for the BMI Plant Sites and Common Areas Projects*. Henderson, Nevada. March 19, 2009(c).
- Nevada Division of Environmental Protection (NDEP). *NDEP Supplemental Guidance on Data Validation for the BMI Plant Sites and Common Areas Project.*, Henderson, Nevada. April 13, 2009(d).
- Nevada Division of Environmental Protection (NDEP). *Unification of Electronic Data Deliverables (EDD), NDEP- Required EDD Format*. Henderson, Nevada. May 11, 2009(f).
- Nevada Division of Environmental Protection (NDEP). *BMI Plant Sites and Common Areas Projects. NDEP-Required Electronic Data Deliverables (EDD) Format*. Henderson, Nevada. July 7, 2010.
- Northgate Environmental Management (Northgate), *Revised Pre-Confirmation Sampling Work Plan, Remediation Zones RZ-A through RZ-E, Phase B Investigation*. Henderson, Nevada. March 19, 2010



- U.S. Environmental Protection Agency (EPA). Office of Solid Waste and Emergency Response. *Contract Laboratory Program, National Functional Guidelines for Organic Data Review*. October 1999.
- U.S. Environmental Protection Agency (EPA). *Region 9 Superfund Data Evaluation/Validation Guidance*. EPA R9QA/006.1. 2001.
- U.S. Environmental Protection Agency (EPA). Office of Solid Waste and Emergency Response. *Contract Laboratory Program, National Functional Guidelines for Inorganic Data Review*. October 2004.
- U.S. Environmental Protection Agency (EPA). *National Functional Guidelines for Chlorinated Dibenzo-p-Dioxins (CDDs) and Chlorinated Dibenzofurans (CDFs) Data Review*. 2005.
- U.S. Environmental Protection Agency (EPA). Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW-846. Third Edition. *update I, July 1992; update IVA, August 1993; update IV, September 1994; update IVB, January 1995; update IV, December 1996; update IV, February 2007*.
- U.S. Environmental Protection Agency (EPA). Office of Solid Waste and Emergency Response. *Contract Laboratory Program, National Functional Guidelines for Organic Data Review*. June 2008.
- U.S. Environmental Protection Agency (EPA). Office of Solid Waste and Emergency Response. *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use*. January 2009.



TABLES
Provided on DVD



APPENDICES
Provided on DVD

APPENDIX A
LABORATORY REPORTS

APPENDIX B
VALIDATION REPORTS

APPENDIX C
ELECTRONIC DATA DELIVERABLE

