

TABLE OF CONTENTS

Cover page	2
QC Checklist	3
Summary Sheet	4
Sequence Log	6
Initial Calibration	10
Periodic QC.....	19
QC: (MRL Check; Method Blank; LCS1/LCS2; Matrix Spike 2705020877 (MS/MSD)	21
2705020860.....	29
2705020862.....	30
Periodic QC.....	33
2705020864	35
Closing QC	36
Standard Preparation Worksheet and Certificates of Analysis	38

Level IV Data Package

MWH Group 203332

Method: SW 846 9056 Chlorate

Sample No.:

2705020860
2705020862
2705020864

DBP QC Checklist

Analysis Date: 05-04-07 Analyst: Raja

QC'd by m Date 9 May 07

Instrument: IC12

Batch # 2

Calibration including QCS(Secondary Source)

Correlation Coefficient of calibration curve for linear curve is 0.995 or better. (0.99 for quadratic)

CLO2 CLO3 N/A BR

Initial QC Check Samples (MCV, CCB, MBLANK, MRL) to be analyzed with every batch (up to 20 samples) or part thereof

MCV is analyzed before samples:

CLO2/CLO3: 90-110% (180-220ppb) N/A BR: 90-110% (90-110ppb)

CCB is analyzed before samples and after MCV and HCV

MLBANK is analyzed before samples. CLO2/CLO3 N/A BR, if present, is < or = half of the MRL.

CLO2: MRL at 10ppb is within 75%-125% (7.5-12.5ppb)

CLO3: MRL at 10ppb is within 75%-125% (7.5-12.5ppb)

N/A BR: MRL at 5.0ppb is within 75%-125% (3.75-6.25ppb)

LCS/LCSD: Accepted criteria are between 90-110% recovery

CLO2: 180-220ppb for 200ppb

CLO3: 180-220ppb for 200ppb

N/A BR: 90-110ppb for 100ppb

One pair analyzed per batch (up to 20 samples) or part thereof

MS/MSD: Acceptance criteria are between 75%-125% recovery.

CLO2: 75-125ppb for 100ppb spike

CLO3: 75-125ppb for 100ppb spike

N/A BR: 37.5-62.5ppb for 50ppb spike

RPD between MS/MSD is within 15%

One pair, and one MS is analyzed per batch (up to 20 samples) or part thereof

Continuing Calibration Verification (MCV and HCV) are required

MCV recovery is between 90-110%

CLO2 (180-220ppb) CLO3 (180-220ppb) N/A BR (90-110ppb)

HCV recovery is between 90-110%

CLO2 (720-880ppb) CLO3 (720-880ppb) N/A BR (360-440ppb)

Samples

All samples for CLO3 and BR are analyzed within 28 days of collection.

All samples for CLO2 are analyzed within 14 days of collection.

QIR

N/A QIR needed for failed QC

N/A QIR needed for samples analyzed outside of hold time

No.	Sample Name	Comment	Time	Dil.Fac.	Amount ppb CLO3 CD 1	
1,	autocal1,		05/01/07 10:54,	1.0,	n.a.	
2,	autocal2,	RAJA060520-3	05/01/07 11:19,	1.0,	10.1329	
3,	autocal3,	RAJA060520-4	05/01/07 11:45,	1.0,	19.4092	
4,	autocal4,	RAJA060520-5	05/01/07 12:10,	1.0,	100.8718	
5,	autocal5,	RAJA060520-6	05/01/07 12:36,	1.0,	199.5885	
6,	autocal6,	RAJA060520-7	05/01/07 13:01,	1.0,	399.981	
7,	autocal7,	RAJA060520-8	05/01/07 13:26,	1.0,	800.0166	
8,	WASH,		05/03/07 10:57,	1.0,	n.a.	
9,	MCV,		05/03/07 11:23,	1.0,	203.1106	102%
10,	CCB,		05/03/07 11:48,	1.0,	n.a.	
11,	MRLCHK,		05/03/07 12:13,	1.0,	11.1499	111% ✓
12,	MBLK,		05/03/07 12:39,	1.0,	n.a.	
13,	LCS1,		05/03/07 13:04,	1.0,	198.9166	99.5% ✓
14,	LCS2,		05/03/07 13:30,	1.0,	202.4428	101% ✓
15,	2705010113,	CLO2	05/03/07 13:55,	1.0,	n.a.	
16,	2705010113MS,		05/03/07 14:20,	1.0,	92.4719	92.5% ✓
17,	2705010113MSD,		05/03/07 14:46,	1.0,	94.1458	94.1% ✓
18,	2704270078_1/500-DN	BR	05/03/07 15:11,	500.0,	n.a.	
19,	2705010089_1/5000-D	CLO3	05/03/07 15:37,	5000.0,	n.a.	
20,	2704270164_1/5,	CLO3	05/03/07 16:02,	5.0,	201.2699	✓
21,	2705010137_1/10,	CLO39056	05/03/07 16:27,	10.0,	3537.747	✓
22,	2705010418-DNR,	BR	05/03/07 16:53,	1.0,	254.7534	
23,	2705010420-DNR,	BR	05/03/07 17:18,	1.0,	14.5761	
24,	2705020217-DNR,	BR	05/03/07 17:44,	1.0,	3282.976	
25,	2705020227-DNR,	BR	05/03/07 18:09,	1.0,	n.a.	
26,	2705020258-DNR,	BR	05/03/07 18:34,	1.0,	n.a.	
27,	MCV,		05/03/07 19:00,	1.0,	200.5498	100%
28,	CCB,		05/03/07 19:25,	1.0,	n.a.	
29,	2705020090,	CLO2/CLO3	05/03/07 19:51,	1.0,	56.9938	✓
30,	2705020137,	CLO2/CLO3	05/03/07 20:16,	1.0,	93.6337	✓
31,	2705020139,	CLO2/CLO3	05/03/07 20:41,	1.0,	26.4124	✓
32,	2705020139MS,		05/03/07 21:07,	1.0,	119.0162	92.6% ✓
33,	2705020139MSD,		05/03/07 21:32,	1.0,	119.9611	93.5% ✓
34,	2705020867,	CLO2/CLO3	05/03/07 21:58,	1.0,	33.6172	✓
35,	2705020868,	CLO2/CLO3	05/03/07 22:23,	1.0,	n.a.	✓
36,	2705020869,	CLO2/CLO3	05/03/07 22:48,	1.0,	242.2078	✓
37,	2705020778,	CLO2	05/03/07 23:14,	1.0,	n.a.	
38,	2705010140_1/1000,	CLO39056	05/03/07 23:39,	1000.0,	67162.06	✓
39,	2705010702_1/50-DNF	CLO39056	05/04/07 00:05,	50.0,	n.a.	
40,	2705010703_1/100,	CLO39056	05/04/07 00:30,	100.0,	68728.66	✓
41,	HCV,		05/04/07 00:55,	1.0,	804.0787	101% ✓
42,	CCB,		05/04/07 01:21,	1.0,	n.a.	
43,	MCV,		05/04/07 01:46,	1.0,	200.4611	100% ✓
44,	CCB,		05/04/07 02:12,	1.0,	n.a.	
45,	MRLCHK,		05/04/07 02:37,	1.0,	9.915	99.2% ✓
46,	MBLK,		05/04/07 03:02,	1.0,	n.a.	
47,	LCS1,		05/04/07 03:28,	1.0,	203.9949	102% ✓
48,	LCS2,		05/04/07 03:53,	1.0,	195.5742	97.8% ✓

No.	Sample Name	Comment	Time	Dil.Fac.	Amount ppb CLO3 CD_1	
49,	2705020877,	CLO2/CLO3	05/04/07 04:19,	1.0,	16.8778 ✓	
50,	2705020877MS,		05/04/07 04:44,	1.0,	109.6442 ✓	92.8% ✓
51,	2705020877MSD,		05/04/07 05:09,	1.0,	108.79 ✓	91.9% ✓
52,	2705010709_1/10,	CLO39056	05/04/07 05:35,	10.0,	176.3369 ✓	
53,	2705010712_1/10-DNF	CLO39056	05/04/07 06:00,	10.0,	1184.061 ✓	
54,	2705010716_1/1000,	CLO39056	05/04/07 06:26,	1000.0,	111212.2 ✓	
55,	2705010717_1/10-DNF	CLO39056	05/04/07 06:51,	10.0,	74.9489 ✓	
56,	2705020800_1/100,	CLO39056	05/04/07 07:16,	100.0,	18993.47 ✓	
57,	2705020801_1/100,	CLO39056	05/04/07 07:42,	100.0,	18191.32 ✓	
58,	2705020813_1/10-DNF	CLO39056	05/04/07 08:07,	10.0,	181.0154 ✓	
59,	2705020860_1/10000,	CLO39056	05/04/07 08:33,	10000.0,	3435387 ✓	
60,	2705020862_1/1000,	CLO39056	05/04/07 08:58,	1000.0,	220145 ✓	
61,	MCV,		05/04/07 09:23,	1.0,	199.5306 ✓	99.8% ✓
62,	CCB,		05/04/07 09:49,	1.0,	n.a.	
63,	2705020864_1/500,	CLO39056	05/04/07 10:14,	500.0,	19227.05 ✓	
64,	2705030182_1/10000,	CLO39056	05/04/07 10:40,	10000.0,	1384236 ✓	
65,	2705030202,	CLO2/CLO3	05/04/07 11:05,	1.0,	n.a. ✓	
66,	2705030202MS,		05/04/07 11:30,	1.0,	95.0056 ✓	95.0% ✓
67,	2705030202MSD,		05/04/07 11:56,	1.0,	91.2357 ✓	91.2% ✓
68,	2705030087,	CLO2	05/04/07 12:21,	1.0,	128.5606 ✓	
69,	2705030088,	CLO2	05/04/07 12:46,	1.0,	129.831 ✓	
70,	2705030089,	CLO2	05/04/07 13:12,	1.0,	132.5726 ✓	
71,	2705030090,	CLO2	05/04/07 13:37,	1.0,	134.7359 ✓	
72,	2705030103,	CLO2	05/04/07 14:03,	1.0,	119.6479 ✓	
73,	2705030105,	CLO2	05/04/07 14:28,	1.0,	383.6674 ✓	
74,	2705030106,	CLO2	05/04/07 14:53,	1.0,	107.339 ✓	
75,	HCV,		05/04/07 15:23,	1.0,	791.923 ✓	99.0% ✓
76,	CCB,		05/04/07 15:48,	1.0,	n.a.	

VB: NMS/10/07

Sequence: 050307-DBP-IC12
Operator: raja

Page 1 of 4
Printed: 5/7/2007 6:16:04 PM

Title:
Datasource: IC-SERVER_local
Location: IC12\2007MAY
Timebase: IC12
#Samples: 76

Created: 5/2/2007 1:10:43 PM by raja
Last Update: 5/4/2007 12:07:15 PM by raja

No.	Name	Comment	Dil. Factor	Sample ID	Type	Program
1	autocal1		1.0000		Standard	IC12 test Program
2	autocal2	RAJA060520-3	1.0000		Standard	IC12 test Program
3	autocal3	RAJA060520-4	1.0000		Standard	IC12 test Program
4	autocal4	RAJA060520-5	1.0000		Standard	IC12 test Program
5	autocal5	RAJA060520-6	1.0000		Standard	IC12 test Program
6	autocal6	RAJA060520-7	1.0000		Standard	IC12 test Program
7	autocal7	RAJA060520-8	1.0000		Standard	IC12 test Program
8	WASH		1.0000		Unknown	IC12 test Program
9	MCV		1.0000		Unknown	IC12 test Program
10	CCB		1.0000		Unknown	IC12 test Program
11	MRLCHK		1.0000		Unknown	IC12 test Program
12	MBLK		1.0000		Unknown	IC12 test Program
13	LCS1		1.0000		Unknown	IC12 test Program
14	LCS2		1.0000		Unknown	IC12 test Program
15	2705010113	CLO2	1.0000		Unknown	IC12 test Program
16	2705010113MS		1.0000		Unknown	IC12 test Program
17	2705010113MSD		1.0000		Unknown	IC12 test Program
18	2704270078_1/500-DNR	BR	500.0000		Unknown	IC12 test Program
19	2705010089_1/5000-DNR	CLO3	5000.0000		Unknown	IC12 test Program
20	2704270164_1/5	CLO3	5.0000		Unknown	IC12 test Program
21	2705010137_1/10	CLO39056	10.0000		Unknown	IC12 test Program
22	2705010418-DNR	BR	1.0000		Unknown	IC12 test Program
23	2705010420-DNR	BR	1.0000		Unknown	IC12 test Program
24	2705020217-DNR	BR	1.0000		Unknown	IC12 test Program
25	2705020227-DNR	BR	1.0000		Unknown	IC12 test Program
26	2705020258-DNR	BR	1.0000		Unknown	IC12 test Program
27	MCV		1.0000		Unknown	IC12 test Program
28	CCB		1.0000		Unknown	IC12 test Program
29	2705020090	CLO2/CLO3	1.0000		Unknown	IC12 test Program
30	2705020137	CLO2/CLO3	1.0000		Unknown	IC12 test Program
31	2705020139	CLO2/CLO3	1.0000		Unknown	IC12 test Program
32	2705020139MS		1.0000		Unknown	IC12 test Program
33	2705020139MSD		1.0000		Unknown	IC12 test Program
34	2705020867	CLO2/CLO3	1.0000		Unknown	IC12 test Program
35	2705020868	CLO2/CLO3	1.0000		Unknown	IC12 test Program
36	2705020869	CLO2/CLO3	1.0000		Unknown	IC12 test Program
37	2705020778	CLO2	1.0000		Unknown	IC12 test Program
38	2705010140_1/1000	CLO39056	1000.0000		Unknown	IC12 test Program
39	2705010702_1/50-DNR	CLO39056	50.0000		Unknown	IC12 test Program
40	2705010703_1/100	CLO39056	100.0000		Unknown	IC12 test Program
41	HCV		1.0000		Unknown	IC12 test Program
42	CCB		1.0000		Unknown	IC12 test Program

Sequence: 050307-DBP-IC12
Operator: raja

Page 2 of 4
Printed: 5/7/2007 6:16:04 PM

Title:
Datasource: IC-SERVER_local
Location: IC12\2007MAY
Timebase: IC12
#Samples: 76

Created: 5/2/2007 1:10:43 PM by raja
Last Update: 5/4/2007 12:07:15 PM by raja

No.	Name	Method	Status	Inj. Date/Time	*Analyst
1	autocal1	DBP-Method	Finished	5/1/2007 10:54:27 AM	raja
2	autocal2	DBP-Method	Finished	5/1/2007 11:19:51 AM	raja
3	autocal3	DBP-Method	Finished	5/1/2007 11:45:15 AM	raja
4	autocal4	DBP-Method	Finished	5/1/2007 12:10:39 PM	raja
5	autocal5	DBP-Method	Finished	5/1/2007 12:36:03 PM	raja
6	autocal6	DBP-Method	Finished	5/1/2007 1:01:26 PM	raja
7	autocal7	DBP-Method	Finished	5/1/2007 1:26:50 PM	raja
8	WASH	DBP-Method	Finished	5/3/2007 10:57:48 AM	raja
9	MCV	DBP-Method	Finished	5/3/2007 11:23:11 AM	raja
10	CCB	DBP-Method	Finished	5/3/2007 11:48:35 AM	raja
11	MRLCHK	DBP-Method	Finished	5/3/2007 12:13:59 PM	raja
12	MBLK	DBP-Method	Finished	5/3/2007 12:39:23 PM	raja
13	LCS1	DBP-Method	Finished	5/3/2007 1:04:47 PM	raja
14	LCS2	DBP-Method	Finished	5/3/2007 1:30:10 PM	raja
15	2705010113	DBP-Method	Finished	5/3/2007 1:55:34 PM	raja
16	2705010113MS	DBP-Method	Finished	5/3/2007 2:20:58 PM	raja
17	2705010113MSD	DBP-Method	Finished	5/3/2007 2:46:22 PM	raja
18	2704270078_1/500-DNR	DBP-Method	Finished	5/3/2007 3:11:46 PM	raja
19	2705010089_1/5000-DNR	DBP-Method	Finished	5/3/2007 3:37:09 PM	raja
20	2704270164_1/5	DBP-Method	Finished	5/3/2007 4:02:33 PM	raja
21	2705010137_1/10	DBP-Method	Finished	5/3/2007 4:27:57 PM	raja
22	2705010418-DNR	DBP-Method	Finished	5/3/2007 4:53:20 PM	raja
23	2705010420-DNR	DBP-Method	Finished	5/3/2007 5:18:44 PM	raja
24	2705020217-DNR	DBP-Method	Finished	5/3/2007 5:44:08 PM	raja
25	2705020227-DNR	DBP-Method	Finished	5/3/2007 6:09:32 PM	raja
26	2705020258-DNR	DBP-Method	Finished	5/3/2007 6:34:56 PM	raja
27	MCV	DBP-Method	Finished	5/3/2007 7:00:20 PM	raja
28	CCB	DBP-Method	Finished	5/3/2007 7:25:44 PM	raja
29	2705020090	DBP-Method	Finished	5/3/2007 7:51:08 PM	raja
30	2705020137	DBP-Method	Finished	5/3/2007 8:16:32 PM	raja
31	2705020139	DBP-Method	Finished	5/3/2007 8:41:56 PM	raja
32	2705020139MS	DBP-Method	Finished	5/3/2007 9:07:20 PM	raja
33	2705020139MSD	DBP-Method	Finished	5/3/2007 9:32:44 PM	raja
34	2705020867	DBP-Method	Finished	5/3/2007 9:58:08 PM	raja
35	2705020868	DBP-Method	Finished	5/3/2007 10:23:32 PM	raja
36	2705020869	DBP-Method	Finished	5/3/2007 10:48:55 PM	raja
37	2705020778	DBP-Method	Finished	5/3/2007 11:14:19 PM	raja
38	2705010140_1/1000	DBP-Method	Finished	5/3/2007 11:39:43 PM	raja
39	2705010702_1/50-DNR	DBP-Method	Finished	5/4/2007 12:05:07 AM	raja
40	2705010703_1/100	DBP-Method	Finished	5/4/2007 12:30:31 AM	raja
41	HCV	DBP-Method	Finished	5/4/2007 12:55:55 AM	raja
42	CCB	DBP-Method	Finished	5/4/2007 1:21:19 AM	raja

Sequence: 050307-DBP-IC12
Operator: raja

Page 3 of 4
Printed: 5/7/2007 6:16:04 PM

Title:
Datasource: IC-SERVER_local
Location: IC12\2007MAY
Timebase: IC12
#Samples: 76

Created: 5/2/2007 1:10:43 PM by raja
Last Update: 5/4/2007 12:07:15 PM by raja

No.	Name	Comment	Dil. Factor	Sample ID	Type	Program
43	MCV		1.0000		Unknown	IC12 test Program
44	CCB		1.0000		Unknown	IC12 test Program
45	MRLCHK		1.0000		Unknown	IC12 test Program
46	MBLK		1.0000		Unknown	IC12 test Program
47	LCS1		1.0000		Unknown	IC12 test Program
48	LCS2		1.0000		Unknown	IC12 test Program
49	2705020877	CLO2/CLO3	1.0000		Unknown	IC12 test Program
50	2705020877MS		1.0000		Unknown	IC12 test Program
51	2705020877MSD		1.0000		Unknown	IC12 test Program
52	2705010709_1/10	CLO39056	10.0000		Unknown	IC12 test Program
53	2705010712_1/10-DNR	CLO39056	10.0000		Unknown	IC12 test Program
54	2705010716_1/1000	CLO39056	1000.0000		Unknown	IC12 test Program
55	2705010717_1/10-DNR	CLO39056	10.0000		Unknown	IC12 test Program
56	2705020800_1/100	CLO39056	100.0000		Unknown	IC12 test Program
57	2705020801_1/100	CLO39056	100.0000		Unknown	IC12 test Program
58	2705020813_1/10-DNR	CLO39056	10.0000		Unknown	IC12 test Program
59	2705020860_1/10000	CLO39056	10000.0000		Unknown	IC12 test Program
60	2705020862_1/1000	CLO39056	1000.0000		Unknown	IC12 test Program
61	MCV		1.0000		Unknown	IC12 test Program
62	CCB		1.0000		Unknown	IC12 test Program
63	2705020864_1/500	CLO39056	500.0000		Unknown	IC12 test Program
64	2705030182_1/10000	CLO39056	10000.0000		Unknown	IC12 test Program
65	2705030202	CLO2/CLO3	1.0000		Unknown	IC12 test Program
66	2705030202MS		1.0000		Unknown	IC12 test Program
67	2705030202MSD		1.0000		Unknown	IC12 test Program
68	2705030087	CLO2	1.0000		Unknown	IC12 test Program
69	2705030088	CLO2	1.0000		Unknown	IC12 test Program
70	2705030089	CLO2	1.0000		Unknown	IC12 test Program
71	2705030090	CLO2	1.0000		Unknown	IC12 test Program
72	2705030103	CLO2	1.0000		Unknown	IC12 test Program
73	2705030105	CLO2	1.0000		Unknown	IC12 test Program
74	2705030106	CLO2	1.0000		Unknown	IC12 test Program
75	HCV		1.0000		Unknown	IC12 test Program
76	CCB		1.0000		Unknown	IC12 test Program

Sequence: 050307-DBP-IC12
Operator: raja


Page 4 of 4
Printed: 5/7/2007 6:16:04 PM

Title:
Datasource: IC-SERVER_local
Location: IC12\2007MAY
Timebase: IC12
#Samples: 76

Created: 5/2/2007 1:10:43 PM by raja
Last Update: 5/4/2007 12:07:15 PM by raja


No.	Name	Method	Status	Inj. Date/Time	*Analyst
43	MCV	DBP-Method	Finished	5/4/2007 1:46:43 AM	raja
44	CCB	DBP-Method	Finished	5/4/2007 2:12:07 AM	raja
45	MRLCHK	DBP-Method	Finished	5/4/2007 2:37:31 AM	raja
46	MBLK	DBP-Method	Finished	5/4/2007 3:02:55 AM	raja
47	LCS1	DBP-Method	Finished	5/4/2007 3:28:19 AM	raja
48	LCS2	DBP-Method	Finished	5/4/2007 3:53:42 AM	raja
49	2705020877	DBP-Method	Finished	5/4/2007 4:19:06 AM	raja
50	2705020877MS	DBP-Method	Finished	5/4/2007 4:44:30 AM	raja
51	2705020877MSD	DBP-Method	Finished	5/4/2007 5:09:54 AM	raja
52	2705010709_1/10	DBP-Method	Finished	5/4/2007 5:35:18 AM	raja
53	2705010712_1/10-DNR	DBP-Method	Finished	5/4/2007 6:00:42 AM	raja
54	2705010716_1/1000	DBP-Method	Finished	5/4/2007 6:26:06 AM	raja
55	2705010717_1/10-DNR	DBP-Method	Finished	5/4/2007 6:51:30 AM	raja
56	2705020800_1/100	DBP-Method	Finished	5/4/2007 7:16:53 AM	raja
57	2705020801_1/100	DBP-Method	Finished	5/4/2007 7:42:17 AM	raja
58	2705020813_1/10-DNR	DBP-Method	Finished	5/4/2007 8:07:40 AM	raja
59	2705020860_1/10000	DBP-Method	Finished	5/4/2007 8:33:03 AM	raja
60	2705020862_1/1000	DBP-Method	Finished	5/4/2007 8:58:27 AM	raja
61	MCV	DBP-Method	Finished	5/4/2007 9:23:50 AM	raja
62	CCB	DBP-Method	Finished	5/4/2007 9:49:14 AM	raja
63	2705020864_1/500	DBP-Method	Finished	5/4/2007 10:14:37 AM	raja
64	2705030182_1/10000	DBP-Method	Finished	5/4/2007 10:40:01 AM	raja
65	2705030202	DBP-Method	Finished	5/4/2007 11:05:25 AM	raja
66	2705030202MS	DBP-Method	Finished	5/4/2007 11:30:48 AM	raja
67	2705030202MSD	DBP-Method	Finished	5/4/2007 11:56:12 AM	raja
68	2705030087	DBP-Method	Finished	5/4/2007 12:21:35 PM	raja
69	2705030088	DBP-Method	Finished	5/4/2007 12:46:59 PM	raja
70	2705030089	DBP-Method	Finished	5/4/2007 1:12:23 PM	raja
71	2705030090	DBP-Method	Finished	5/4/2007 1:37:46 PM	raja
72	2705030103	DBP-Method	Finished	5/4/2007 2:03:09 PM	raja
73	2705030105	DBP-Method	Finished	5/4/2007 2:28:33 PM	raja
74	2705030106	DBP-Method	Finished	5/4/2007 2:53:56 PM	raja
75	HCV	DBP-Method	Finished	5/4/2007 3:23:03 PM	raja
76	CCB	DBP-Method	Finished	5/4/2007 3:48:27 PM	raja

1 autocal1			
Sample Name:	autocal1	Injection Volume:	1000.0
Vial Number:	334	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 10:54	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	

2 autocal2			
RAJA060520-3			
Sample Name:	autocal2	Injection Volume:	1000.0
Vial Number:	335	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 11:19	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	0.064	0.010	42.54	9.926	BMB
2	11.98	Br	0.016	0.005	20.78	5.721	BMB*
3	13.41	ClO3	0.025	0.009	36.68	10.133	BMB*
Total:			0.105	0.024	100.00	25.780	

3 autocal3			
RAJA060520-4			
Sample Name:	autocal3	Injection Volume:	1000.0
Vial Number:	336	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 11:45	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	0.126	0.020	29.66	20.367	BMB
2	12.00	Br	0.031	0.009	13.08	9.977	BMB*
3	13.42	ClO3	0.048	0.017	24.43	19.409	BMB*
4	14.51	n.a.	0.062	0.022	32.84	n.a.	BMB*
Total:			0.267	0.068	100.00	49.753	

4 autocal4			
RAJA060520-5			
Sample Name:	autocal4	Injection Volume:	1000.0
Vial Number:	337	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 12:10	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height µS	Area µS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	0.628	0.097	42.52	99.774	BMB
2	11.99	Br	0.155	0.045	19.79	49.048	BMB*
3	13.41	ClO3	0.245	0.086	37.69	100.872	BMB*
Total:			1.027	0.228	100.00	249.694	

5 autocal5			
RAJA060520-6			
Sample Name:	autocal5	Injection Volume:	1000.0
Vial Number:	338	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 12:36	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	1.273	0.195	42.58	199.693	BMB
2	11.99	Br	0.314	0.092	20.19	99.881	BMB
3	13.40	ClO3	0.495	0.171	37.23	199.588	BMB
Total:			2.082	0.458	100.00	499.163	


6 autocal6			
RAJA060520-7			
Sample Name:	autocal6	Injection Volume:	1000.0
Vial Number:	334	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 13:01	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	2.606	0.397	42.77	400.286	BMB
2	11.99	Br	0.634	0.187	20.13	200.466	BMB
3	13.40	ClO3	1.002	0.345	37.10	399.981	BMB
Total:			4.242	0.929	100.00	1000.733	





7 autocal7**RAJA060520-8**

Sample Name:	autocal7	Injection Volume:	1000.0
Vial Number:	334	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 13:26	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.49	ClO2	5.423	0.820	43.22	799.953	BMB
2	11.99	Br	1.278	0.377	19.86	399.907	BMB
3	13.39	ClO3	2.032	0.701	36.92	800.017	BMB
Total:			8.733	1.898	100.00	1999.877	

7 autocal7			
RAJA060520-8			
Sample Name:	autocal7	Injection Volume:	1000.0
Vial Number:	334	Channel:	CD_1
Sample Type:	standard	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/1/2007 13:26	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Cal.Type	Points	Corr.Coeff. %	Offset	Slope	Curve
1	5.49	ClO2	QOff	6	99.9837	0.0005	0.0010	0.0000
2	11.99	Br	QOff	6	99.9979	-0.0004	0.0009	0.0000
3	13.39	ClO3	QOff	6	99.9959	0.0000	0.0008	0.0000
Average:					99.9925	0.0000	0.0009	0.0000

8 WASH			
Sample Name:	WASH	Injection Volume:	1000.0
Vial Number:	335	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/3/2007 10:57	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	


9 MCV			
Sample Name:	MCV	Injection Volume:	1000.0
Vial Number:	336	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/3/2007 11:23	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.46	ClO2	1.300	0.204	43.31	208.386	BMB
2	11.88	Br	0.319	0.093	19.77	100.389	BMB
3	13.26	ClO3	0.508	0.174	36.92	203.111	BMB
Total:			2.127	0.470	100.00	511.886	


10 CCB

Sample Name:	CCB	Injection Volume:	1000.0
Vial Number:	336	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/3/2007 11:48	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	


11 MRLCHK			
Sample Name:	MRLCHK	Injection Volume:	1000.0
Vial Number:	336	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/3/2007 12:13	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height µS	Area µS*min	Rel.Area %	Amount ppb	Type
1	5.46	ClO2	0.065	0.010	29.08	10.130	BMB*
2	11.88	Br	0.053	0.015	43.87	17.065	BMB
3	13.26	ClO3	0.025	0.009	27.05	11.150	BMB
Total:			0.144	0.035	100.00	38.344	


45 MRLCHK

Sample Name:	MRLCHK	Injection Volume:	1000.0
Vial Number:	741	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 2:37	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.45	ClO2	0.064	0.010	44.72	9.670	BMB
2	11.83	Br	0.014	0.004	16.61	4.356	BMB*
3	13.21	ClO3	0.025	0.008	38.67	9.915	BMB
Total:			0.102	0.022	100.00	23.941	

46 MBLK			
Sample Name:	MBLK	Injection Volume:	1000.0
Vial Number:	742	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 3:02	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area $\mu\text{S}\cdot\text{min}$	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	


47 LCS1			
Sample Name:	LCS1	Injection Volume:	1000.0
Vial Number:	743	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 3:28	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.46	ClO2	1.324	0.204	43.32	208.974	BMB*
2	11.85	Br	0.318	0.093	19.70	100.332	BMB
3	13.23	ClO3	0.501	0.174	36.98	203.995	BMB*
Total:			2.143	0.472	100.00	513.301	


48 LCS2

Sample Name:	LCS2	Injection Volume:	1000.0
Vial Number:	744	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 3:53	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height µS	Area µS*min	Rel.Area %	Amount ppb	Type
1	5.45	ClO2	1.316	0.200	43.77	205.101	BMB*
2	11.84	Br	0.310	0.090	19.72	97.552	BMB
3	13.22	ClO3	0.488	0.167	36.50	195.574	BMB
Total:			2.113	0.458	100.00	498.228	


49 2705020877			
CLO2/CLO3			
Sample Name:	2705020877	Injection Volume:	1000.0
Vial Number:	745	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 4:19	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	11.84	Br	0.056	0.016	0.56	17.261	BMB
2	13.21	ClO3	0.045	0.014	0.51	16.878	BMB
3	14.27	n.a.	7.819	2.764	98.93	n.a.	BMB
Total:			7.920	2.794	100.00	34.139	

50 2705020877MS


Sample Name:	2705020877MS	Injection Volume:	1000.0
Vial Number:	746	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 4:44	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.44	ClO2	0.485	0.083	2.78	85.926	BMB
2	11.85	Br	0.205	0.060	1.99	64.702	BMB
3	13.22	ClO3	0.276	0.093	3.12	109.644	BM
4	14.27	n.a.	7.815	2.759	92.11	n.a.	MB
Total:			8.781	2.995	100.00	260.272	

51 2705020877MSD

Sample Name:	2705020877MSD	Injection Volume:	1000.0
Vial Number:	747	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 5:09	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.45	ClO2	0.486	0.084	2.80	86.263	BMB
2	11.85	Br	0.205	0.059	1.98	64.347	BMB
3	13.22	ClO3	0.277	0.093	3.09	108.790	BMB
4	14.27	n.a.	7.812	2.758	92.13	n.a.	BMB
Total:			8.780	2.994	100.00	259.400	

59 2705020860_1/10000


CLO39056

Sample Name:	2705020860_1/10000	Injection Volume:	1000.0
Vial Number:	755	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	#####
Recording Time:	5/4/2007 8:33	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	13.21	ClO3	0.843	0.295	94.55	3435387.350	BM
2	14.31	n.a.	0.045	0.017	5.45	n.a.	MB
Total:			0.889	0.312	100.00	3435387.350	

60 2705020862_1/1000			
CLO39056			
Sample Name:	2705020862_1/1000	Injection Volume:	1000.0
Vial Number:	756	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1000.0000
Recording Time:	5/4/2007 8:58	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	13.21	ClO3	0.543	0.188	90.97	220145.014	BMB*
2	14.29	n.a.	0.054	0.019	9.03	n.a.	BMB*
Total:			0.597	0.207	100.00	220145.014	

60 2705020862_1/1000			
CLO39056			
Sample Name:	2705020862_1/1000	Injection Volume:	1000.0
Vial Number:	756	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1000.0000
Recording Time:	5/4/2007 8:58	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height µS	Area µS*min	Rel.Area %	Amount ppb	Type
1	13.21	ClO3	0.546	0.196	91.28	229231.410	BMB
2	14.29	n.a.	0.054	0.019	8.72	n.a.	Rd
Total:			0.600	0.215	100.00	229231.410	


60 2705020862_1/1000			
CLO39056			
Sample Name:	2705020862_1/1000	Injection Volume:	1000.0
Vial Number:	756	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1000.0000
Recording Time:	5/4/2007 8:58	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret. Time min	Peak Name	Height µS	Area µS*min	Rel. Area %	Amount ppb	Type
1	13.21	ClO3	0.543	0.188	90.97	220145.014	BMB*
2	14.29	n.a.	0.054	0.019	9.03	n.a.	BMB*
Total:			0.597	0.207	100.00	220145.014	


61 MCV

Sample Name:	MCV	Injection Volume:	1000.0
Vial Number:	757	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 9:23	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height µS	Area µS*min	Rel.Area %	Amount ppb	Type
1	5.45	ClO2	1.276	0.197	42.34	201.224	BMB
2	11.84	Br	0.313	0.092	19.74	98.973	BMB
3	13.22	ClO3	0.497	0.171	36.73	199.531	BMB
4	14.31	n.a.	0.016	0.006	1.20	n.a.	BMB
Total:			2.103	0.464	100.00	499.728	

62 CCB			
Sample Name:	CCB	Injection Volume:	1000.0
Vial Number:	758	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 9:49	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000




No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	

63 2705020864_1/500			
CLO39056			
Sample Name:	2705020864_1/500	Injection Volume:	1000.0
Vial Number:	759	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	500.0000
Recording Time:	5/4/2007 10:14	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	13.20	ClO3	0.096	0.033	3.74	19227.054	BMB
2	14.28	n.a.	2.348	0.841	96.26	n.a.	BMB
Total:			2.444	0.873	100.00	19227.054	


75 HCV			
Sample Name:	HCV	Injection Volume:	1000.0
Vial Number:	771	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 15:23	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
1	5.45	ClO2	5.359	0.813	43.21	793.246	BMB
2	11.84	Br	1.272	0.375	19.93	397.916	BMB
3	13.20	ClO3	2.031	0.693	36.85	791.923	BMB
Total:			8.661	1.882	100.00	1983.085	

76 CCB

Sample Name:	CCB	Injection Volume:	1000.0
Vial Number:	772	Channel:	CD_1
Sample Type:	unknown	Wavelength:	n.a.
Control Program:	IC12 test Program	Bandwidth:	n.a.
Quantif. Method:	DBP-Method	Dilution Factor:	1.0000
Recording Time:	5/4/2007 15:48	Sample Weight:	1.0000
Run Time (min):	23.00	Sample Amount:	1.0000



No.	Ret.Time min	Peak Name	Height μS	Area μS*min	Rel.Area %	Amount ppb	Type
Total:			0.000	0.000	0.00	0.000	

**Standard
Preparation
Worksheet
&
Certificate of
Analysis**

Reagent Preparation Documentation

Reagent: DBP Calibration stock ^{PR} solution
 Date Received/Prepped: 052006/060606/062206/071306/072506/080806
 Date Expired: 062006/060606 Raja/072206/081306/082506/090806
 Manufacturer: _____
 Storage Condition: _____

MW #: Raja060520-2
 By: Raja
 Matrix: aq
 Amount: 100ml
 Lot #: _____

Component	Comment	Standard	Concentration
Bromide 100ppm	500 ML ? add 50 ML EDA solution	R 201373	5ppm
Exp:- 06/07	(LMR 060129-12) then		
Chlorate 100ppm	1000 ML dilute to 100ml with D.I	R 201374	10ppm
Exp:- 06/07	water		
Chlorite 100ppm	1000 ML	R 201375	10ppm
Exp:- 06/07			

Comment: 082706/100106/110106/111606/122006/010807/020807/
Exp:- 092706/110106/120106/121606/012007/020807/030807/
110106 Raja/120106

Reagent: DBP Calibration standard #1
 Date Received/Prepped: 052006/060606/062206/071306/072506/080806
 Date Expired: 062006/070606/072206/081306/082506/090806
 Manufacturer: _____
 Storage Condition: _____

MW #: Raja060520-3
 By: Raja
 Matrix: aq
 Amount: 100ml
 Lot #: _____

Component	Comment	Standard	Concentration
DBP calib. stock soln	100 ML ? Dilute with D.I. water	Raja060520-2	Br- 5ppb
	to 100ml		ClO ₂ - 10ppb
EDA solution	50 ML	LMR060129-12	ClO ₃ - 10ppb

Comment: Used as MRL-check Prep:- 082706/100106/110106/111606/122006/010807
Exp:- 092706/110106/120106/121606/012007/020807

Reagent: DBP Calibration Standard #2
 Date Received/Prepped: 052006/060606/062206/071306/072506/080806
 Date Expired: 062006/070606/072206/081306/082506/090806
 Manufacturer: _____
 Storage Condition: _____

MW #: Raja060520-4
 By: Raja
 Matrix: aq
 Amount: 100ml
 Lot #: _____

Component	Comment	Standard	Concentration
DBP calib. stock soln	200 ML ? Dilute with D.I. water	Raja060520-2	Br- 10ppb
	to 100ml		Br
EDA solution	50 ML	LMR060129-12	ClO ₂ - 20ppb
			ClO ₃ - 20ppb

Comment: Prep:- 082706/100106/110106/111606/122006/010807/020807/
Exp:- 092706/110106/120106/121606/012007/020807/030807/

Reagent Preparation Documentation

Reagent: DBP Calibration Standard #3
Date Received/Prepped: 052006/060606/062206/071306/072506/080806
Date Expired: 062006/070606/072206/081306/082506/090806
Manufacturer: _____
Storage Condition: _____

MW #: Raja060520-5 ✓
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
DBP calib. stock soln	1ml } Dilute with D.I. water to 100ml	Raja060520-2	Br-50ppb
EDA solution	50ML }	LMR060129-12	ClO ₂ -100ppb ClO ₃ -100ppb

Comment: Prep:- 082706/100106/110406/110106/111606/122006/010807/020607/
 Exp:- 092706/110106/120406/120106/121606/012007/020807/030607/

Reagent: DBP Calibration Standard #4
Date Received/Prepped: 052006/060606/062206/071306/072506/080806
Date Expired: 062006/070606/072206/081306/082506/090806
Manufacturer: _____
Storage Condition: _____

MW #: Raja060520-6 ✓
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
DBP calib. stock soln	2ml } Dilute with D.I. water to 100ml	Raja060520-2	Br-100ppb
EDA solution	50ML }	LMR060129-12	ClO ₂ 200ppb 250 ClO ₃ 200ppb Raja

Comment: It is also used as MCV Prep:- 082706/100106/110406/110106/111606/122006
 Exp:- 092706/110106/120406/120106/121606/012007

Reagent: DBP calibration Standard #5
Date Received/Prepped: 052006/060606/062206/071306/072506/080806
Date Expired: 062006/070606/072206/081306/082506/090806
Manufacturer: _____
Storage Condition: _____

MW #: Raja060520-7 ✓
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
DBP calib. stock soln	4ml } Dilute to 100ml with D.I. water	Raja060520-2	Br-20ppb
EDA solution	50ML }	LMR060129-12	ClO ₂ -400ppb ClO ₃ -400ppb

Comment: Prep:- 082706/100106/110406/110106/111606/122006/010807/020607
 Exp:- 092706/110106/120406/120106/121606/012007/020807/030607

Reagent Preparation Documentation

Page: 10

Reagent: DBP Calibration Standard #6
Date Received/Prepped: 052006/060606/062206/071306/072506/080806
Date Expired: 062006/070606/072206/081306/082506/090806
Manufacturer: _____
Storage Condition: _____

MW #: Raja060520-8
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
DBP calib stock soln	8ml } Dilute to 100ml with D.I. water	Raja060520-2	Br-400ppb
EDA	50ML }	LMR060129-12	ClO ₂ -800ppb
			ClO ₃ -800ppb

Comment: used as HCV Prep:- 082706 / 100106 / 110406 / 110106 / 111606 / 122006 / 010807 / 020607
Exp:- 092706 / 110106 / 120406 / 120106 / 121606 / 012007 / 020807 / 030607

Reagent: IPCCV Solution for ClO₄ 25ppb
Date Received/Prepped: 062406/071206/072206/080706/ 1
Date Expired: 092406/101206/102206/110706/ 1
Manufacturer: _____
Storage Condition: _____

MW #: Raja060624-1
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
Na ₂ CO ₃ 10,000ppm	4ml ?	LYL050510-06	400ppm
NaCl 10,000ppm	4ml Dilute to 100ml with	LYL050510-04	400ppm
Na ₂ SO ₄ 10,000ppm	4ml D.I. water	LY1050510-05	400ppm
ClO ₄ calib stock	2.5ml		25ppb

Comment: _____

Reagent: Na₂SO₄ Solution - ClO₄
Date Received/Prepped: 081406/110906/011707/032307/052107/
111406/020707/011707/062307/082107/
020407 062307
Manufacturer: _____
Storage Condition: _____

MW #: Raja060814-1
By: Raja
Matrix: aq
Amount: 100ml
Lot #: _____

Component	Comment	Standard	Concentration
Na ₂ SO ₄ Exp:- 03/08	1.48g → Dilute to 100ml with D.I. water	R200651	10,000ppm

Comment: _____

Reagent Preparation Documentation

Reagent: 3001 Multi-element Calibration Std - 7
Date Received/Prepped: 29 Jan 06 / 19 Feb 06 / 25 Apr 06 / 16 May 06 / 16 Apr 06 / 1 May 06
Date Expired: / / / / /
Manufacturer:
Storage Condition:

MW #: LMR060129-11
By: LMR
Matrix: ng
Amount: 100ml
Lot #: -

Component	Comment	Standard	Concentration
BrO ₃ - 1000 ppb	5000 µl	LMR050109-3	50
ClO ₂ /ClO ₂ - 10 ppm	2000 µl	LMR050825-1	200 ea
Aristo Cal Mix	2000 µl w/ DI H ₂ O	LMR060129-4	200 ea
EDA - 100000 ppm	50 µl	LMR060129-12	50 ppm

Comment:

Reagent: Ethylenediamine - 100,000 ppm (EDA)
Date Received/Prepped: 29 Jan 06 / 10 Jul 06 / / / /
Date Expired: / / / / /
Manufacturer:
Storage Condition:

MW #: LMR060129-12
By: LMR
Matrix: ng
Amount: 100 ml
Lot #: -

Component	Comment	Standard	Concentration
Ethylenediamine	11.2 ml diluted to 100ml w/ DI H ₂ O	R200653	100,000 ppm
	use 50 µl per 100 ml		

Comment:

Reagent: BrO₃ Color Reagent
Date Received/Prepped: 1/29/06 / 2/2/06 / 2/16/06 / 2/23/06 / 2/14/06 / 2/20/06 / 2/24/06 / 3/1/06
Date Expired: / / / / /
Manufacturer:
Storage Condition:

MW #: LMR060129-13
By: LMR
Matrix: ng
Amount: 2-1
Lot #: -

Component	Comment	Standard	Concentration
o-dianisidine (ODA)	1g dissolved in 400ml HPLC/pest grade Methanol	R201258	
KBr	10g dissolved in ~14L H ₂ O (DI)	R201284	
Ultra HNO ₃	100 ml	R201319	
		R201320 (2/18 - 3/1/06), 3/19/06, 3/14/06, 3/20/06	
		R201374 (1/19/06, 2/2/06, 2/23/06, 3/1/06, 3/14/06, 3/20/06)	
		R201393 (2/16/06, 2/23/06, 3/1/06, 3/14/06, 3/20/06)	

Comment: Add ODA to MeOH and dissolve. Using a 27.5ml flask 42.5ml KBr in DI H₂O. Add HNO₃ to KBr solution. Add ODA to KBr/HNO₃ soln and bring to mark w/ DI H₂O. Soln must be clear in 30 min. Soln must stand minimum 16 hrs. before using. Best to stand overnight.

Reagent Documentation

Reagent: Conductivity Std - 1000µmhos/cm
Date Received: 20 Mar 03
Date Expired: 29 Feb 04
Manufacturer: VWR
Storage Condition: room temp

Reagent #: 200652
By: LMR
Matrix: aq
Amount: 4-L
Lot #: 3049

Component	Comment	Standard	Concentration

Comment: _____

Reagent: Ethylenediamine
Date Received: 21 Mar 03
Date Expired: -
Manufacturer: J.T. Baker
Storage Condition: room temp

Reagent #: 200653 ✓
By: LMR
Matrix: neat
Amount: 500 ml
Lot #: X18582

Component	Comment	Standard	Concentration
	<u>CAS# 107-15-3</u>		

Comment: _____

Reagent: Sodium Hydroxide Pellets
Date Received: 21 Mar 03
Date Expired: -
Manufacturer: J.T. Baker
Storage Condition: room temp

Reagent #: 200654
By: LMR
Matrix: solid
Amount: 4 x 500g
Lot #: X21468

Component	Comment	Standard	Concentration

Comment: _____

Reagent Documentation

Reagent: Conductivity Std. - 1000 ppm
 Date Received: 3 May 06
 Date Expired: Jan 07
 Manufacturer: Ricca Chemical
 Storage Condition: room temp

Reagent #: 201372
 By: LMR
 Matrix: ag
 Amount: 4-L
 Lot #: 1601439

Component	Comment	Standard	Concentration
	VWOR# RC 2243-1		

Comment:

Reagent: Bromide - 1000 ppm Std
 Date Received: 4 May 06
 Date Expired: 1 Jun 07
 Manufacturer: Inorganic Ventures
 Storage Condition: room temp

Reagent #: 201373 ✓
 By: LMR
 Matrix: ag
 Amount: 125 ml
 Lot #: Y-BR01057

Component	Comment	Standard	Concentration
	IN# ICBR1-1		

Comment:

Reagent: Chlorate - 1000 ppm std
 Date Received: 4 May 06
 Date Expired: 1 Jun 07
 Manufacturer: Inorganic Ventures
 Storage Condition: refrigerate 4±2°C

Reagent #: 201374 ✓
 By: LMR
 Matrix: ag
 Amount: 125 ml
 Lot #: Y-CLOX01034

Component	Comment	Standard	Concentration
	IN# ICCL031-1		

Comment:

Reagent: Chlorite - 1000 ppm std
 Date Received: 4 May 06
 Date Expired: 1 Jun 07
 Manufacturer: Inorganic Ventures
 Storage Condition: refrigerate 4±2°C

Reagent #: 201375 ✓
 By: LMR
 Matrix: ag
 Amount: 125 ml
 Lot #: Y-CLOX01036

Component	Comment	Standard	Concentration
	N # ICCL021-1		

Comment:

Reagent: Ammonium Std - 1000 ppm as NH₄
 Date Received: 4 May 06
 Date Expired: 3 Oct 07
 Manufacturer: CPI
 Storage Condition: refrigerate 4±2°C

Reagent #: 201376
 By: LMR
 Matrix: ag
 Amount: 100 ml
 Lot #: 06D002

Component	Comment	Standard	Concentration
	CPI # 4400-010010		

Comment:

Reagent: Methylene Blue 1% w/v solution
 Date Received: 10 May 06
 Date Expired: 30 Nov 07
 Manufacturer: VWR
 Storage Condition: room temp

Reagent #: 201377
 By: LMR
 Matrix: ag
 Amount: 100 ml
 Lot #: 5319

Component	Comment	Standard	Concentration
	VWR # VWR 276-0		

Comment:

1.0 INORGANIC VENTURES is an ISO Guide 34:2000 registered Certified Reference Material (CRM) Manufacturer (Certificate #883-02). The certificate is designed and the data is determined in accordance with ISO Guide 31:2000 (Reference Materials-Contents of Certificates and Labels), ISO Guide 34:2000 "Quality System Guidelines for the Production of Reference Materials," and ISO Guide 35-1989 "Certification of Reference Materials - General and Statistical Principles."

2.0 DESCRIPTION OF CRM **1000 µg/mL Bromide in Water**

Catalog Number: ICBR1-1 and ICBR1-5

Lot Number: **Y-BR01057**

Starting Material: Potassium Bromide

Starting Material Purity (%): 99.0000

Starting Material Lot No.: 09014BY

Matrix: Water

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Concentration: 999 ± 3 µg/mL

Certified Density: 0.998 g/mL (measured at 22° C)

The Certified Value is the instrument analysis value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_1}{n}$$

(\bar{x}) = mean
 x1 = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = \frac{2(\sum s_1^2)}{(n)^{1/2}}^{1/2}$$

$\sum s_1$ = The summation of all significant estimated errors
 (Most common are the errors from instrumental measurement weighting, dilution to volume, and the fixed error reported on the NIST SRM certificate of analysis.)

The independent samples t-test was used to determine if there is agreement between the above assay methods at the 95% confidence interval. Both methods were compared and showed agreement within the stated uncertainties. This agreement is a confirmation of the accuracy of this CRM.

4.0 TRACEABILITY TO NIST AND VALUES OBTAINED BY INDEPENDENT METHODS

· "Property of the result of a measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties."
 (ISO VIM, 2nd ed., 1993, definition 6.10)

· This IV product is Traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRMs are available, the term 'in-house std.' is specified.

Assay Method #1 999 ± 3 µg/mL
 IC Assay NIST SRM 3184 Lot Number: 020701

Assay Method #2 997 ± 3 µg/mL
 Volhard NIST SRM 999a Lot Number: 999a

- 4.2 **BALANCE CALIBRATION** - All balances are checked daily using in-house procedure 6-IMM-001. The weights used for testing are annually compared to Gerhart Scale Corporation's master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 692476 - Class 1 and 692476A - Class 2. The NIST test number is 822/260017-98. All analytical balances are calibrated every 4 months by Gerhart Scale Corp. of South Amboy. The balances are calibrated with a class 1 and/or class 2 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is 822/260017-98.
- 4.3 **THERMOMETER CALIBRATION** - The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. The in-house procedure is 2-QC-001. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification Nos. 92564, 119016, 471047 and NIST test report Nos. 811/258522, 811/2557078, and 236090.
- 4.4 **GLASSWARE CALIBRATION** - In-house procedure 3-QC-002 is used to calibrate all Class A Glassware used in the manufacture and quality control of CRM's.

R201373

5.0 **Chromatogram - N/A**

6.0 **INTENDED USE**

For the calibration of analytical instruments including but not limited to the following:
HPLC, IC, TLC, ISE, IR, NMR, UV/VIS, MS, Capillary Electrophoresis, Potentiometry, Wet Chemistry and Voltammetry
For the validation of analytical methods
For the preparation of "working reference samples"
For interference studies and the determination of correction coefficients
For detection limit and linearity studies
For additional intended uses, contact IV Technical Support

7.0 **INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL**

Storage & Handling - Keep **Tightly** sealed when not in use. Store and use at $20 \pm 4^\circ\text{C}$. **Do Not** pipette from the container. **Do Not** return portions removed from pipetting to container.

- **Elements Specific Information** - For specific information regarding the elements in your blend:
Go on line at www.inorganicventures.com/extras/pertable <<http://www.inorganicventures.com/extras/pertable>> Take advantage of our interactive Periodic Table.
- **E-mail Technical Support** at ivtech@inorganicventures.com <<mailto:ivtech@inorganicventures.com>> at any time and a technical representative will respond within 24 business hours.
- Call our toll free Technical Support line at (800) 569-6799 or dial (732) 901-1900, Monday through Friday, from 8:00AM to 6:00PM Eastern Standard Time.

8.0 **HAZARDOUS INFORMATION** - Please refer to the enclosed Material Safety Data sheet for information regarding this CRM.

9.0 **HOMOGENEITY** - This solution was mixed according to an in-house procedure IV-MPM-004 and is guaranteed to be homogeneous.

10.0 QUALITY STANDARD DOCUMENTATION



R# 201373

10.1 ISO 9001:2000 Quality Management System Registration - QMI Certificate Number 010105

- Recognized by:
- Registrar Accreditation Board (ANSI-RAB)
- Standards Council of Canada (SCC)
- Dutch Council for Accreditation (RVA)
- Entidad Mexicana de Acreditacion, a.c.(EMA)

Members of IQ Net International Certification Network:

Argentina (IRAM), Australia (QAS), Austria (OQS), Belgium (Avinter), Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)

10.2 ISO/IEC 17025 - 1999 "General Requirements for the Competence of Testing and Calibration" - Chemical Testing - Accredited A2LA Certificate Number 883.01

10.3 ISO/IEC Guide 34 - 2000 "General Requirements for the Competence of Reference Material Producers" - Reference Materials Production - Accredited A2LA Certificate Number 883.02

A2LA Mutual Recognition Agreement Partners:

Australia (NATA), Austria (BmWA), Belgium (BELTEST) (BKO-OBE), Canada (SCC), Chinese Taipei (CNLA), Czech Republic (NAO), Denmark (DANAK), Finland (FINAS), France (COFRAC), Germany (DAR), Hong Kong (HKAS), Ireland (NAB), Italy (SIT) (SINAL), Japan (JAB) (JNLA), Republic of Korea (KOLAS), The Netherlands (RvA), New Zealand (IANZ), Norway (NA), Portugal (IPQ), Singapore (SAC-SINGLAS), Spain (ENAC), Sweden (SWEDAC), Switzerland (SAS), United Kingdom (UKAS) and United States (NVLAP) (ICBO ES)

10.4 10CFR50 Appendix B - Nuclear Regulatory Commission - Domestic Licensing of Production and Utilization Facilities

10.5 10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance

10.6 MIL-STD-45662A (Obsolete/Observed)

11.0 DATE OF CERTIFICATION AND PERIOD OF VALIDITY

11.1 IV Shelf Life - The period of time during which the concentration of the analyte(s) in a properly packaged, unopened and unused standard stored under environmentally controlled and monitored conditions will remain within the specified uncertainty range. Shelf life is limited primarily by transpiration (loss of water from the solution) and infrequently, by chemical instability. Transpiration studies (P-SP01020) of chemically-stable solutions performed at the Manufacture's facility shows a CRM shelf-life of four years for solutions packaged in 500-mL low density polyethylene bottles. When stored under special environmental controls that minimize transpiration and instability, the shelf life can be extended past this limit.

11.2 Expiration Date - The date after which a CRM should not be used. Routine laboratory use of a CRM increases transpiration losses and the chance of contamination which affect the integrity of the CRM and limit its useful life. The manufacturer concurs with state and federal regulatory agencies' recommendations that solution standards be assigned a one-year expiration date.

Certification Date: February 28, 2006

Expiration Date: EXPIRES 12/2007

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By: Nick Maida, QA Administrator


Certificate Approved By: Katalin Le, QC Manager

Certifying Officer: Paul Gaines, PhD., Technical Director

R# 201373

ATTACHMENT 1

CHROMATOGRAM FOR: 1000 $\mu\text{g/mL}$ Bromide Y-BR01057



DIONEX DX-120 Ion Chromatograph
Anal. Column: IonPac AS14 4 x 250 mm
Guard Column: IonPac AG14 4 x 50 mm
Anion Self Regenerating Suppressor:
ASRS-ULTRA II 4mm
Suppressor Current: 100 mA
Eluent: 3.5 mM Na_2CO_3
0.8mM NaHCO_3
Eluent Flow Rate: 1.00 mL/min
Cell Temp.: 35 °C
Scale: Y-axis = 20 μS scale
X-axis = minutes
Concentration: approx. 20 $\mu\text{g/g}$

I-CAL ION CHROMATOGRAPHY SOLUTION 1000 µg/mL Chlorate in H₂O
 Catalog No: ICCLO31-1 and ICCLO31-5

 Lot Number: **Y-CLOX01034**

 Starting Material: Potassium Chlorate
 Starting Material Lot No: 02407TF

CERTIFIED CONCENTRATION: 1008 ± 3 µg/mL

The Certified Value is the Inductively Coupled Plasma Spectroscopy (ICP) value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

(\bar{x}) = mean x_i = individual results n = number of measurements $\sum s_i$ = The summation of all significant estimated errors.

Calculated Value: 1001 ± 5 µg/mL


Method: Calculated, based on starting material.

Instrument Value: 1006 ± 2 µg/mL

Method: Ion Chromatography vs an in-house standard.

Instrument Value: 1008 ± 3 µg/mL

Method: Analysis by Inductively Coupled Plasma Spectroscopy (ICP) vs NIST SRM 3182 Lot number: 990506.

ClO₃⁻ Y-CLOX01034


DIONEX DX-120 Ion Chromatograph
 Anal. Column: IonPac AS9-HC 4 x 250mm
 Guard Column: IonPac AG9-HC 4 x 50mm
 Anion self Generating Suppressor:
 ASRS-ULTRA II 4mm
 Suppressor Current: 100mA
 Eluent: 9.0 mM Na₂CO₃
 Eluent Flow Rate: 1.0 mL/min
 Cell Temp.: 35 °C
 Scale: Y-axis = 15µS scale
 X-axis = minutes
 Concentration: 20µg/g

ANALYZED DENSITY OF SOLUTION (measured at 22°C): 0.998 g/mL

QA:MG

Rev 082405ntm



 Expires: **EXPIRES**
 12/2007

QUALITY STANDARD DOCUMENTATION



1. ISO 9001:2000 QMI Registered Quality System (Certificate Number 010105)
 Members of IQ Net : Argentina (IRAM), Australia (QAS), Austria (ÖQS), Belgium (Avinter) , Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)
 2. ISO/IEC Guide 34-2000 "General Requirements for the Competence of Reference Material Producers" - Reference Materials Production - Accredited A2LA Certificate 883.02
 3. ISO/IEC 17025-1999 "General Requirements for the Competence of Testing and Calibration" - Chemical Testing - Accredited A2LA Certificate 883.01
 4. MIL-STD-45662A
 5. 10CFR50 Appendix B - Nuclear Regulatory Commission - Domestic Licencing of Production and Utilization Facilities
 6. 10CFR21 - Nuclear Regulatory Commission - Reporting Defects and Non-Compliance
- Please contact our Quality Assurance Department for further information and copies of documents pertaining to our Quality Standard certifications.

STABILITY/ EXPIRATION DOCUMENTATION

- Shelf Life -** The length of time that a properly stored and packaged standard will remain within the specified uncertainty. Shelf life is affected by chemical stability and transpiration issues. Inorganic Ventures' Standard Solutions are chemically stable indefinitely. Transpiration loss is linear with time and limits the time a standard can be used with confidence. The smaller the bottle the higher the rate of transpiration. Inorganic Ventures' studies indicate that the shelf life of our 500 mL bottle is 4 years and the shelf life of our 125 mL bottle is 21 months.
- Expiration Date -** The date after which a standard solution should not be used. A one year expiration date is recommended by most state and federal regulatory agencies. Transpiration issues and repeated use of solutions over a one year period may adversely affect the integrity of the standard.

PACKAGING DOCUMENTATION

Purified acid, 18 megohm double deionized water that has been filtered through a 0.2 µm filter and in-house procedure IV-PACK-001 is used to clean all bottles. Contact us for technical information relating to contamination issues in packaging materials.

GLASSWARE CALIBRATION

In-house procedure 3-QC-002 is used to calibrate all Class A Glassware used in the manufacture and quality control of Custom Grade Standards.

BALANCE CALIBRATION

All balances are checked daily using in-house procedure number 6-IMM-001. The weights used for testing are annually compared to Gerhart Scale Corporation's master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 428359B and 454678. The NIST test number is 822/260017-98.

All analytical balances are calibrated every 4 months by Gerhart Scale Corp. of South Amboy. The balances are calibrated with a class 1 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is 822/260017-8.

THERMOMETER CALIBRATION

The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. The in-house procedure No. is 2-QC-001. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification Nos. 92564, 119016, 471047 and NIST test report Nos. 811/258522, 811/2557078, and 236090.

TECHNICAL SUPPORT

All customers are encouraged to contact us for technical support for the proper use of our products.

I-CAL ION CHROMATOGRAPHY SOLUTION 1000 $\mu\text{g/mL}$ Chlorite in H_2O

Catalog No: ICCL021-1 and ICCL021-5

 Lot Number: **Y-CLOX01036**

Starting Material:

Sodium Chlorite

Starting Material Lot No:

E02F39

CERTIFIED CONCENTRATION: 998 \pm 3 $\mu\text{g/mL}$

* The Certified Concentration for Lot No. Y-CLOX01036 is only the ClO_2^- . The value of Cl^- is $6 \pm 1 \mu\text{g/mL}$, and the value of ClO_3^- is $12 \pm 1 \mu\text{g/mL}$. This was determined by Ion Chromatography vs an in-house standard solutions traceable to NIST SRM 3182.

The Certified Value is based upon the wet assay value. The following equations are used in the calculation of the certified value and the uncertainty:

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

$$\text{Uncertainty } (\pm) = \frac{2[(\sum s_i)^2]^{1/2}}{(n)^{1/2}}$$

 (\bar{x}) = mean

 x_i = individual results

 n = number of measurements


 $\sum s_i$ = The summation of all significant estimated errors.

Instrument Value: 1002 \pm 2 $\mu\text{g/mL}$

Method: Ion Chromatography vs NIST SRM 136e Lot number 980702.

Wet Analysis: 998 \pm 3 $\mu\text{g/mL}$

Method: Iodometric Titration NIST SRM 136e Lot number 980702.

 ClO_2^- Y-CLOX01036


DIONEX DX-120 Ion Chromatograph
 Anal. Column: IonPac AS9-HC 4 x 250mm
 Guard Column: IonPac AG9-HC 4 x 50mm
 Anion self Generating Suppressor:
 ASRS-ULTRA II 4mm
 Suppressor Current: 100mA
 Eluent: 9 mM Na_2CO_3
 Eluent Flow Rate: 1.00 mL/min
 Cell Temp.: 35 °C
 Scale: Y-axis = 30 μS scale
 X-axis = minutes
 Concentration: 20 $\mu\text{g/g}$

ANALYZED DENSITY OF SOLUTION (measured at 22°C): 0.998 g/mL


QA:KL Rev 032006NTM



52

Quality Assurance Manager

Expires:



QUALITY STANDARD DOCUMENTATION

1. ISO 9001 QMI Registered Quality System (Certificate Number 010105)

Members of IQ Net : Argentina (IRAM), Australia (QAS), Austria (ÖQS), Belgium (Avinter) , Brazil (FCAV), Canada (QMI), Hong Kong (HKQAA), Columbia (ICONTEC), Czech Republic (CQS), Denmark (DS), Finland (SFS), France (AFAQ), Germany (DQS), Greece (ELOT), Hungary (MSZT), Ireland (NSAI), Israel (SII), Italy (CISQ), Japan (JQA), Korea (KSA-QA), Netherlands (KEMA), Norway (NCS), Poland(PCBC), Portugal (APCER), Singapore (PSB), Slovenia (SIQ), Spain (AENOR), Switzerland (SQS)



2. ISO/IEC 17025-1999 - Chemical Testing - Accredited A2LA (Certificate Number 883.01)
3. ISO/IEC Guide 34-2000 - Reference Materials Production - Accredited A2LA (Certificate Number 883.02)
4. 10CFR50 Appendix B
5. 10CFR21

Please contact our Quality Assurance Department for further information and copies of documents pertaining to our Quality Standard certifications.

STABILITY/ EXPIRATION DOCUMENTATION

Shelf Life - The length of time that a properly stored and packaged standard will remain within the specified uncertainty. Shelf life is affected by chemical stability and transpiration issues. Inorganic Ventures' Standard Solutions are chemically stable indefinitely. Transpiration loss is linear with time and limits the time a standard can be used with confidence. The smaller the bottle the higher the rate of transpiration. Inorganic Ventures' studies indicate that the shelf life of our 500 mL bottle is 4 years and the shelf life of our 125 mL bottle is 21 months.

Expiration Date - The date after which a standard solution should not be used. A one year expiration date is recommended by most state and federal regulatory agencies. Transpiration issues and repeated use of solutions over a one year period may adversely affect the integrity of the standard.

PACKAGING DOCUMENTATION

Purified acid, 18 megohm double deionized water that has been filtered through a 0.2 μm filter and in-house procedure IV-PACK-001 is used to clean all bottles. Contact us for technical information relating to contamination issues in packaging materials.

GLASSWARE CALIBRATION

In-house procedure 3-QC-002 is used to calibrate all Class A Glassware used in the manufacture and quality control of Custom Grade Standards.

BALANCE CALIBRATION

All balances are checked daily using in-house procedure number 6-IMM-001. The weights used for testing are annually compared to Gerhart Scale Corporation's master weights and are traceable to the National Institute of Standards and Technology (NIST). The NIST Traceability numbers are 428359B and 454678. The NIST test number is 822/260017-98.

All analytical balances are calibrated every 4 months by Gerhart Scale Corp. of South Amboy. The balances are calibrated with a class 1 analytical weight set. These weights are tested annually by a NIST / NVLAP accredited calibration lab. The NIST test number is 822/260017-8.

THERMOMETER CALIBRATION

The thermometers used in the determination of the final densities are calibrated vs standard thermometer No. 903-2680 which was certified in accordance with the procedures outlined by ASTM E77-87 and NIST Monograph 150 using NIST Test Nos. and Std Nos.: 769543, 217368/769543, 217368/P14452, 176240/P14452, 176240. The in-house procedure No. is 2-QC-001. Thermometers which are not calibrated vs standard thermometer No. 903-2680 are traceable to NIST Identification Nos. 92564, 119016, 471047 and NIST test report Nos. 811/258522, 811/2557078, and 236090.

TECHNICAL SUPPORT

All customers are encouraged to contact us for technical support for the proper use of our products.