

No.	Sample Name	Time	Dil.Fac.	Amount	Comment	Analyst:	Imr
				CRVI-LOW			
				UV VIS 1		Criteria	
1,	Standard 1 - 0.1 ppb,	03/22/06 15:39,	1.0,	0.116	Standard Stock		
2,	Standard 2 - 0.2ppb,	03/22/06 15:47,	1.0,	0.226	LMR050904-1		
3,	Standard 3 - 2.0 ppb,	03/22/06 15:56,	1.0,	2.000	exp 20 Apr 06		
4,	Standard 4 - 10 ppb,	03/22/06 16:04,	1.0,	10.007			
5,	Standard 5 - 20 ppb,	03/22/06 16:12,	1.0,	20.019			
6,	Standard 6 - 50ppb,	03/22/06 16:20,	1.0,	49.991			
7,	IPC 20,	03/25/06 10:17,	1.0,	20.055	20.1 100%		
8,	IPC 10,	03/25/06 10:25,	1.0,	10.083	10.1 101%	MRL 50%-150%	
9,	LRB,	03/25/06 10:33,	1.0,	n.a.		0.05 - 0.15ppb	
10,	LRB + Buffer,	03/25/06 10:41,	1.0,	0.014		LCS 90%-110%	
11,	MRL 0.1ppb,	03/25/06 10:49,	1.0,	0.095	0.095 95%	Range: 1.80 - 2.20	
12,	LCS 2.0ppb,	03/25/06 10:57,	1.0,	2.028	2.03 102%	MS/MSD 90%-110%	
13,	LCSD 2.0ppb,	03/25/06 11:06,	1.0,	2.025	2.03 102%	True Value = 2.0	
14,	2603250005,	03/25/06 11:14,	1.0,	0.014		IPC 95%-105%	
15,	2603250005_MS,	03/25/06 11:22,	1.0,	2.120	2.11 - 105% recovery	20ppb - 19-21ppb	
16,	2603250005_MSD,	03/25/06 11:30,	1.0,	2.256	2.24 - over range	10ppb - 9.5-10.5 ppb	
17,	LRB,	03/25/06 11:38,	1.0,	n.a.			
18,	2603250005_MSD,	03/25/06 11:46,	1.0,	2.152	2.14 - 107% recovery		
19,	IPC 20,	03/25/06 11:54,	1.0,	20.046			